



SGS-CSTC Standards Technical Services (Suzhou) Co., Ltd.

Report No.: SUCR250400032311

Rev.: 01

Page: 1 of 15

Appendix A

Detailed System Check Results

1. System Performance Check
System Performance Check 13 MHz
System Performance Check 750 MHz
System Performance Check 835 MHz
System Performance Check 1750 MHz
System Performance Check 1950 MHz
System Performance Check 2300 MHz
System Performance Check 2450 MHz
System Performance Check 2600 MHz
System Performance Check 3500 MHz
System Performance Check 2900 MHz
System Performance Check 5250 MHz
System Performance Check 5600 MHz
System Performance Check 5750 MHz

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone:(86-755) 8307 1443, or email: CN.Doccheck@sgs.com

SGS-CSTC Standards Technical Services (Suzhou) Co., Ltd.
Wireless Laboratory

South of No. 6 Plant, No. 1, RunSheng Road, Suzhou Industrial Park,
Suzhou Area, China (Jiangsu) Pilot Free Trade Zone 215000

t (86-512) 6229 2980
www.sgsgroup.com.cn


System Performance Check 13MHz

Communication System: Custom Band; Frequency: 13.000
Medium: HSL. Medium parameters used: $f=13.000$ MHz; $\sigma=0.747$ S/m; $\epsilon_r=55.7$

- DASY8 Configuration:
- Probe: EX3DV4 - SN7735; ConvF(13.68, 13.51, 13.73); Calibrated: 2025-01-29
 - Sensor-Surface: 1.4 mm
 - Electronics: DAE4ip Sn1826; Calibrated: 2025-02-17
 - Phantom: ELI V8.0 (20deg probe tilt); Serial: 2217
 - Measurement Software: cDASY8 V16.4.0.5005

Area Scan (80.0 mm x 100.0 mm): Measurement Grid: 10.0 mm x 15.0 mm
SAR (1g) = 0.114 W/kg; SAR (10g) = 0.093 W/kg;

Zoom Scan (30.0 mm x 30.0 mm x 30.0 mm): Measurement Grid: 6.0 mm x 6.0 mm x 1.5 mm
Power Drift = 0.09 dB
SAR (1g) = 0.106 W/kg; SAR (10g) = 0.067 W/kg;
M2/M1 [%] 68.6
Dist 3dB Peak [mm] 14.1




System Performance Check 750MHz

Communication System: ; Frequency: 750.000
Medium: HSL. Medium parameters used: $f=750.000$ MHz; $\sigma=0.909$ S/m; $\epsilon_r=41.7$

- DASY8 Configuration:
- Probe: EX3DV4 - SN7767; ConvF(10.76, 10.76, 10.76); Calibrated: 2024-12-31
 - Sensor-Surface: 1.4 mm
 - Electronics: DAE4ip Sn1826; Calibrated: 2025-02-17
 - Phantom: Twin-SAM V8.0 (30deg probe tilt); Serial: 2155
 - Measurement Software: cDASY8 V16.4.0.5005

Area Scan (40.0 mm x 90.0 mm): Measurement Grid: 10.0 mm x 15.0 mm
SAR (1g) = 2.29 W/kg; SAR (10g) = 1.54 W/kg;

Zoom Scan (30.0 mm x 30.0 mm x 30.0 mm): Measurement Grid: 6.0 mm x 6.0 mm x 1.5 mm
Power Drift = 0.06 dB
SAR (1g) = 2.26 W/kg; SAR (10g) = 1.53 W/kg;
M2/M1 [%] 91.3
Dist 3dB Peak [mm] > 15.0




System Performance Check 835MHz

Communication System: ; Frequency: 835.000
Medium: HSL. Medium parameters used: $f=835.000\text{ MHz}$; $\sigma=0.920\text{ S/m}$; $\epsilon_r=41.1$

- DASY8 Configuration:
- Probe: EX3DV4 - SN7767; ConvF(10.34, 10.34, 10.34); Calibrated: 2024-12-31
 - Sensor-Surface: 1.4 mm
 - Electronics: DAE4ip Sn1826; Calibrated: 2025-02-17
 - Phantom: Twin-SAM V8.0 (30deg probe tilt); Serial: 2155
 - Measurement Software: cDASY8 V16.4.0.5005

Area Scan (40.0 mm x 90.0 mm): Measurement Grid: 10.0 mm x 15.0 mm
SAR (1g) = 2.46 W/kg; SAR (10g) = 1.63 W/kg;

Zoom Scan (30.0 mm x 30.0 mm x 30.0 mm): Measurement Grid: 6.0 mm x 6.0 mm x 1.5 mm
Power Drift = 0.09 dB
SAR (1g) = 2.44 W/kg; SAR (10g) = 1.61 W/kg;
M2/M1 [%] 91.2
Dist 3dB Peak [mm] 19.0




System Performance Check 1750MHz

Communication System: ; Frequency: 1750.000
Medium: HSL. Medium parameters used: $f=1750.000\text{ MHz}$; $\sigma=1.38\text{ S/m}$; $\epsilon_r=40.7$

- DASY8 Configuration:
- Probe: EX3DV4 - SN7767; ConvF(8.95, 8.95, 8.95); Calibrated: 2024-12-31
 - Sensor-Surface: 1.4 mm
 - Electronics: DAE4ip Sn1826; Calibrated: 2025-02-17
 - Phantom: Twin-SAM V8.0 (30deg probe tilt); Serial: 2155
 - Measurement Software: cDASY8 V16.4.0.5005

Area Scan (40.0 mm x 90.0 mm): Measurement Grid: 10.0 mm x 15.0 mm
SAR (1g) = 9.41 W/kg; SAR (10g) = 5.17 W/kg;

Zoom Scan (30.0 mm x 30.0 mm x 30.0 mm): Measurement Grid: 6.0 mm x 6.0 mm x 1.5 mm
Power Drift = -0.07 dB
SAR (1g) = 9.36 W/kg; SAR (10g) = 5.14 W/kg;
M2/M1 [%] 85.5
Dist 3dB Peak [mm] 10.9




System Performance Check 1950MHz

Communication System: ; Frequency: 1950.000
Medium: HSL. Medium parameters used: $f= 1950.000$ MHz; $\sigma= 1.41$ S/m; $\epsilon_r = 40.4$

- DASY8 Configuration:
- Probe: EX3DV4 - SN7767; ConvF(8.6, 8.6, 8.6); Calibrated: 2024-12-31
 - Sensor-Surface: 1.4 mm
 - Electronics: DAE4ip Sn1826; Calibrated: 2025-02-17
 - Phantom: Twin-SAM V8.0 (30deg probe tilt); Serial: 2155
 - Measurement Software: cDASY8 V16.4.0.5005

Area Scan (40.0 mm x 90.0 mm): Measurement Grid: 10.0 mm x 15.0 mm
SAR (1g) = 9.88 W/kg; SAR (10g) = 5.21 W/kg;

Zoom Scan (30.0 mm x 30.0 mm x 30.0 mm): Measurement Grid: 6.0 mm x 6.0 mm x 1.5 mm
Power Drift = -0.04 dB
SAR (1g) = 9.84 W/kg; SAR (10g) = 5.16 W/kg;
M2/M1 [%] 85.4
Dist 3dB Peak [mm] 10.9




System Performance Check 2300MHz

Communication System: ; Frequency: 2300.000
Medium: HSL. Medium parameters used: $f= 2300.000$ MHz; $\sigma= 1.69$ S/m; $\epsilon_r = 39.0$

- DASY8 Configuration:
- Probe: EX3DV4 - SN7767; ConvF(8.33, 8.33, 8.33); Calibrated: 2024-12-31
 - Sensor-Surface: 1.4 mm
 - Electronics: DAE4ip Sn1826; Calibrated: 2025-02-17
 - Phantom: Twin-SAM V8.0 (30deg probe tilt); Serial: 2155
 - Measurement Software: cDASY8 V16.4.0.5005

Area Scan (40.0 mm x 80.0 mm): Measurement Grid: 10.0 mm x 10.0 mm
SAR (1g) = 12.2 W/kg; SAR (10g) = 5.86 W/kg;

Zoom Scan (30.0 mm x 30.0 mm x 30.0 mm): Measurement Grid: 5.0 mm x 5.0 mm x 1.5 mm
Power Drift = -0.09 dB
SAR (1g) = 12.3 W/kg; SAR (10g) = 5.94 W/kg;
M2/M1 [%] 83.3
Dist 3dB Peak [mm] 9.5




System Performance Check 2450MHz

Communication System: ; Frequency: 2450.000
Medium: HSL. Medium parameters used: $f= 2450.000$ MHz; $\sigma= 1.83$ S/m; $\epsilon_r = 39.3$

- DASY8 Configuration:
- Probe: EX3DV4 - SN7767; ConvF(8.03, 8.03, 8.03); Calibrated: 2024-12-31
 - Sensor-Surface: 1.4 mm
 - Electronics: DAE4ip Sn1826; Calibrated: 2025-02-17
 - Phantom: Twin-SAM V8.0 (30deg probe tilt); Serial: 2155
 - Measurement Software: cDASY8 V16.4.0.5005

Area Scan (40.0 mm x 80.0 mm): Measurement Grid: 10.0 mm x 10.0 mm
SAR (1g) = 13.5 W/kg; SAR (10g) = 6.26 W/kg;

Zoom Scan (30.0 mm x 30.0 mm x 30.0 mm): Measurement Grid: 5.0 mm x 5.0 mm x 1.5 mm
Power Drift = -0.01 dB
SAR (1g) = 13.3 W/kg; SAR (10g) = 6.23 W/kg;
M2/M1 [%] 82.4
Dist 3dB Peak [mm] 9.3




System Performance Check 2600MHz

Communication System: ; Frequency: 2600.000
Medium: HSL. Medium parameters used: $f= 2600.000$ MHz; $\sigma= 1.93$ S/m; $\epsilon_r = 38.1$

- DASY8 Configuration:
- Probe: EX3DV4 - SN7767; ConvF(8.03, 8.03, 8.03); Calibrated: 2024-12-31
 - Sensor-Surface: 1.4 mm
 - Electronics: DAE4ip Sn1826; Calibrated: 2025-02-17
 - Phantom: Twin-SAM V8.0 (30deg probe tilt); Serial: 2155
 - Measurement Software: cDASY8 V16.4.0.5005

Area Scan (40.0 mm x 80.0 mm): Measurement Grid: 10.0 mm x 10.0 mm
SAR (1g) = 13.6 W/kg; SAR (10g) = 6.24 W/kg;

Zoom Scan (30.0 mm x 30.0 mm x 30.0 mm): Measurement Grid: 5.0 mm x 5.0 mm x 1.5 mm
Power Drift = 0.06 dB
SAR (1g) = 13.7 W/kg; SAR (10g) = 6.26 W/kg;
M2/M1 [%] 81.4
Dist 3dB Peak [mm] 9.4




System Performance Check 3500MHz

Communication System: ; Frequency: 3500.000
Medium: HSL. Medium parameters used: $f= 3500.000$ MHz; $\sigma= 2.95$ S/m; $\epsilon_r = 37.9$

- DASY8 Configuration:
- Probe: EX3DV4 - SN7767; ConvF(7.23, 7.23, 7.23); Calibrated: 2024-12-31
 - Sensor-Surface: 1.4 mm
 - Electronics: DAE4ip Sn1826; Calibrated: 2025-02-17
 - Phantom: Twin-SAM V8.0 (30deg probe tilt); Serial: 2155
 - Measurement Software: cDASY8 V16.4.0.5005

Area Scan (40.0 mm x 80.0 mm): Measurement Grid: 10.0 mm x 10.0 mm
SAR (1g) = 6.53 W/kg; SAR (10g) = 2.50 W/kg;

Zoom Scan (28.0 mm x 28.0 mm x 28.0 mm): Measurement Grid: 5.0 mm x 5.0 mm x 1.4 mm
Power Drift = 0.06 dB
SAR (1g) = 6.54 W/kg; SAR (10g) = 2.49 W/kg;
M2/M1 [%] 77.6
Dist 3dB Peak [mm] 8.0




System Performance Check 3900MHz

Communication System: ; Frequency: 3900.000
Medium: HSL. Medium parameters used: $f=3900.000\text{ MHz}$; $\sigma=3.35\text{ S/m}$; $\epsilon_r=37.0$

- DASY8 Configuration:
- Probe: EX3DV4 - SN7767; ConvF(6.95, 6.95, 6.95); Calibrated: 2024-12-31
 - Sensor-Surface: 1.4 mm
 - Electronics: DAE4ip Sn1826; Calibrated: 2025-02-17
 - Phantom: Twin-SAM V8.0 (30deg probe tilt); Serial: 2155
 - Measurement Software: cDASY8 V16.4.0.5005

Area Scan (40.0 mm x 80.0 mm): Measurement Grid: 10.0 mm x 10.0 mm
SAR (1g) = 6.73 W/kg; SAR (10g) = 2.35 W/kg;

Zoom Scan (28.0 mm x 28.0 mm x 28.0 mm): Measurement Grid: 5.0 mm x 5.0 mm x 1.4 mm
Power Drift = -0.09 dB
SAR (1g) = 6.72 W/kg; SAR (10g) = 2.33 W/kg;
M2/M1 [%] 77.9
Dist 3dB Peak [mm] 8.5




System Performance Check 5250MHz

Communication System: ; Frequency: 5250.000
Medium: HSL. Medium parameters used: $f= 5250.000$ MHz; $\sigma= 4.68$ S/m; $\epsilon_r = 36.5$

- DASY8 Configuration:
- Probe: EX3DV4 - SN7767; ConvF(5.73, 5.73, 5.73); Calibrated: 2024-12-31
 - Sensor-Surface: 1.4 mm
 - Electronics: DAE4ip Sn1826; Calibrated: 2025-02-17
 - Phantom: Twin-SAM V8.0 (30deg probe tilt); Serial: 2155
 - Measurement Software: cDASY8 V16.4.0.5005

Area Scan (40.0 mm x 80.0 mm): Measurement Grid: 10.0 mm x 10.0 mm
SAR (1g) = 7.46 W/kg; SAR (10g) = 2.06 W/kg;

Zoom Scan (22.0 mm x 22.0 mm x 22.0 mm): Measurement Grid: 4.0 mm x 4.0 mm x 1.4 mm
Power Drift = 0.07 dB
SAR (1g) = 7.48 W/kg; SAR (10g) = 2.06 W/kg;
M2/M1 [%] 66.3
Dist 3dB Peak [mm] 6.6




System Performance Check 5600MHz

Communication System: ; Frequency: 5600.000
Medium: HSL. Medium parameters used: $f= 5600.000$ MHz; $\sigma= 5.06$ S/m; $\epsilon_r = 35.6$

- DASY8 Configuration:
- Probe: EX3DV4 - SN7767; ConvF(5.12, 5.12, 5.12); Calibrated: 2024-12-31
 - Sensor-Surface: 1.4 mm
 - Electronics: DAE4ip Sn1826; Calibrated: 2025-02-17
 - Phantom: Twin-SAM V8.0 (30deg probe tilt); Serial: 2155
 - Measurement Software: cDASY8 V16.4.0.5005

Area Scan (40.0 mm x 80.0 mm): Measurement Grid: 10.0 mm x 10.0 mm
SAR (1g) = 7.93 W/kg; SAR (10g) = 2.210 W/kg;

Zoom Scan (22.0 mm x 22.0 mm x 22.0 mm): Measurement Grid: 4.0 mm x 4.0 mm x 1.4 mm
Power Drift = 0.06 dB
SAR (1g) = 7.94 W/kg; SAR (10g) = 2.23 W/kg;
M2/M1 [%] 63.7
Dist 3dB Peak [mm] 6.6




System Performance Check 5750MHz

Communication System: ; Frequency: 5750.000
Medium: HSL. Medium parameters used: $f= 5750.000$ MHz; $\sigma= 5.25$ S/m; $\epsilon_r = 35.4$

- DASY8 Configuration:
- Probe: EX3DV4 - SN7767; ConvF(5.21, 5.21, 5.21); Calibrated: 2024-12-31
 - Sensor-Surface: 1.4 mm
 - Electronics: DAE4ip Sn1826; Calibrated: 2025-02-17
 - Phantom: Twin-SAM V8.0 (30deg probe tilt); Serial: 2155
 - Measurement Software: cDASY8 V16.4.0.5005

Area Scan (40.0 mm x 80.0 mm): Measurement Grid: 10.0 mm x 10.0 mm
SAR (1g) = 7.53 W/kg; SAR (10g) = 2.06 W/kg;

Zoom Scan (22.0 mm x 22.0 mm x 22.0 mm): Measurement Grid: 4.0 mm x 4.0 mm x 1.4 mm
Power Drift = 0.03 dB
SAR (1g) = 7.56 W/kg; SAR (10g) = 2.09 W/kg;
M2/M1 [%] 63.1
Dist 3dB Peak [mm] 6.3





SGS-CSTC Standards Technical Services (Suzhou) Co., Ltd.

Report No.: SUCR250400032311

Rev.: 01

Page: 15 of 15

- End of the Appendix -