

TEST REPORT

Applicant: Huizhou Jinghao Medical Technology Co., Ltd.
Address: No. 207, South SiHuan Rd, ZhongKai Hi-tech
Distric, Huizhou, China
Equipment Type: Hearings Aids
Model Name: JH-A40C (refer to section 2.3)
Brand Name: jinghao
FCC ID: 2A39N-ZHJHHA
Test Standard: 47 CFR Part 2.1093
KDB 447498 D04 v01
Sample Arrival Date: Apr. 11, 2024
Test Date: Apr. 19, 2024 - Apr. 21, 2024
Date of Issue: May 08, 2024

ISSUED BY:

Shenzhen BALUN Technology Co., Ltd.

**Tested by:** Xiong Lining**Checked by:** Xu Rui**Approved by:** Tolan Tu
(Testing Director)

Revision History

<u>Version</u>	<u>Issue Date</u>	<u>Revisions Content</u>
<u>Rev. 01</u>	<u>May 08, 2024</u>	<u>Initial Issue</u>

TABLE OF CONTENTS

1	GENERAL INFORMATION	3
1.1	Test Laboratory	3
1.2	Test Location	3
2	PRODUCT INFORMATION	4
2.1	Applicant Information	4
2.2	Manufacturer Information	4
2.3	General Description for Equipment under Test (EUT)	4
2.4	Technical Information	4
3	SUMMARY OF TEST RESULT	5
3.1	Test Standards	5
4	DEVICE CATEGORY AND LEVELS LIMITS	6
5	ASSESSMENT RESULT	8
5.1	Output Power	8
5.2	Tune-up power	8
5.3	RF Exposure Evaluation Result	8
5.4	Conclusion	8

1 GENERAL INFORMATION

1.1 Test Laboratory

Name	Shenzhen BALUN Technology Co., Ltd.
Address	Block B, 1/F, Baisha Science and Technology Park, Shahe Xi Road, Nanshan District, Shenzhen, Guangdong Province, P. R. China
Phone Number	+86 755 6685 0100

1.2 Test Location

Name	Shenzhen BALUN Technology Co., Ltd.
Location	☒ Block B, 1/F, Baisha Science and Technology Park, Shahe Xi Road, Nanshan District, Shenzhen, Guangdong Province, P. R. China
	☐ 1/F, Building B, Ganghongji High-tech Intelligent Industrial Park, No. 1008, Songbai Road, Yangguang Community, Xili Sub-district, Nanshan District, Shenzhen, Guangdong Province, P. R. China
Accreditation Certificate	The laboratory is a testing organization accredited by FCC as a accredited testing laboratory. The designation number is CN1196.

2 PRODUCT INFORMATION

2.1 Applicant Information

Applicant	Huizhou Jinghao Medical Technology Co., Ltd.
Address	No. 207, South SiHuan Rd, ZhongKai Hi-tech Distric, Huizhou, China

2.2 Manufacturer Information

Manufacturer	Huizhou Jinghao Medical Technology Co., Ltd.
Address	No. 207, South SiHuan Rd, ZhongKai Hi-tech Distric, Huizhou, China

2.3 General Description for Equipment under Test (EUT)

EUT Name	Hearings Aids
Model Name Under Test	JH-A40C
Series Model Name	JH-DW2ET, JH-DW3E, JH-DW1E, JH-A40B, JH-A41C, JH-A42C, JH-A43C, JH-A39H, JH-DW1A, JH-DW1B, JH-DW4B, JH-DW5A, JH-DW5B, JH-DW2C, JH-DW4C
Description of Model name differentiation	The model and brand name changed for different market and customer, the others are the same. (this information provided by the applicant)
Hardware Version	V1.0
Software Version	V1.0
Dimensions (Approx.)	N/A
Weight (Approx.)	N/A

2.4 Technical Information

Network and Wireless connectivity	Bluetooth (BR+EDR+BLE)
-----------------------------------	------------------------

The requirement for the following technical information of the EUT was tested in this report:

Operating Mode	Bluetooth	
Frequency Range	Bluetooth	2400 ~ 2483.5 MHz
Antenna Type	Bluetooth	Copper Wire Antenna
Exposure Category	General Population/Uncontrolled Exposure	
Product Type	Portable Device	

3 SUMMARY OF TEST RESULT

3.1 Test Standards

No.	Identity	Document Title
1	47 CFR Part 2.1093	Radiofrequency radiation exposure evaluation: portable devices
2	KDB 447498 D04 v01	KDB 447498 D04 Interim General RF Exposure Guidance v01

4 DEVICE CATEGORY AND LEVELS LIMITS

Portable Devices:

CFR Title 47 §2.1093(b)

(b) For purposes of this section, a portable device is defined as a transmitting device designed to be used so that the radiating structure(s) of the device is/are within 20 centimeters of the body of the user.

FCC KDB 447498 Devices:

According with FCC KDB 447498 D04, Appendix B, The SAR-based exemption formula applies for single fixed, mobile, and portable RF sources with available maximum time-averaged power or effective radiated power (ERP), whichever is greater, of less than or equal to the threshold P_{th} (mW).

This method shall only be used at separation distances from 0.5 cm to 40 cm and at frequencies from 0.3 GHz to 6 GHz (inclusive). The following table shows the power threshold from 5mm to 50mm.

Power Thresholds (mW)					
Frequency (MHz)	At separation distance of ≤ 5 mm	At separation distance of 10 mm	At separation distance of 15 mm	At separation distance of 20 mm	At separation distance of 25 mm
300	39 mW	65 mW	88 mW	110 mW	129 mW
450	22 mW	44 mW	67 mW	89 mW	112 mW
835	9 mW	25 mW	44 mW	66 mW	90 mW
1900	3 mW	12 mW	26 mW	44 mW	66 mW
2450	3 mW	10 mW	22 mW	38 mW	59 mW
3600	2 mW	8 mW	18 mW	32 mW	49 mW
5800	1 mW	6 mW	14 mW	25 mW	40 mW
Frequency (MHz)	At separation distance of 30 mm	At separation distance of 35 mm	At separation distance of 40 mm	At separation distance of 45 mm	At separation distance of 50 mm
300	148 mW	166 mW	184 mW	201 mW	217 mW
450	135 mW	158 mW	180 mW	203 mW	226 mW
835	116 mW	145 mW	175 mW	207 mW	240 mW
1900	92 mW	122 mW	157 mW	195 mW	236 mW
2450	83 mW	111 mW	143 mW	179 mW	219 mW
3600	71 mW	96 mW	125 mW	158 mW	195 mW
5800	58 mW	80 mW	106 mW	136 mW	169 mW

Note:

1. Maximum power is the source-based time-average power and represents the maximum RF output power including tune-up tolerance among production units
2. Per KDB 447498 D04, for larger devices, the test separation distance of adjacent edge configuration is determined by the closest separation between the antenna and the user.
3. Per KDB 447498 D04, standalone SAR test exclusion threshold is applied; If the distance of the antenna to the user is < 5mm, 5mm is used to determine SAR exclusion threshold
4. Per KDB 447498 D04, for separation distances from 0.5 cm to 40 cm and at frequencies from 0.3 GHz to 6 GHz (inclusive), the threshold Pth (mW) is given by Following:

$$P_{th}(mW) = \begin{cases} ERP_{20cm}(d/20cm)^x & d \leq 20cm \\ ERP_{20cm} & 20cm < d \leq 40cm \end{cases}$$

where

$$x = -\log_{10} \left(\frac{60}{ERP_{20cm}\sqrt{f}} \right)$$

- a. f(GHz) is the RF channel transmit frequency in GHz
- b. d is the separation distance (cm), The result is rounded to one decimal place for comparison
- c. ERP_{20cm} are determined by:

$$ERP_{20cm}(mW) = f(x) = \begin{cases} 2040f & 0.3GHz \leq f < 1.5GHz \\ 3060 & 1.5GHz \leq f \leq 6GHz \end{cases}$$

5 ASSESSMENT RESULT

5.1 Output Power

Bluetooth		
Mode	BR+EDR	BLE
Peak Power (dBm)	1.26	2.69
Antenna Gain (dBi)	0	
EIRP (dBm)	1.26	2.69
Note: This report listed the worst case power value, please refer to BL-SZ2441097-601~602 report for more details.		

5.2 Tune-up power

Mode	Conducted Power Range (dBm)	EIRP Range (dBm)	ERP Range (dBm)
Bluetooth	[1.00, 3.00]	[1.00, 3.00]	[-1.15, 0.85]
Note 1: ERP= EIRP -2.15dB			
Note 2: According KDB 447498 D04, used the greater of maximum conducted power or ERP to compare with the threshold value Pth.			

5.3 RF Exposure Evaluation Result

Mode	Distance (mm)	Calculation Frequency (GHz)	Tune-up limit power (dBm)	Tune-up limit power (mW)	Threshold Value(mW)	Verdict
Bluetooth	5	2.4	3.00	2.00	2.79	Compliance

5.4 Conclusion

This EUT is deemed to comply with the reference level limits, therefore the basic restrictions are compliant with human exposure limits.

Statement

1. The laboratory guarantees the scientificity, accuracy and impartiality of the test, and is responsible for all the information in the report, except the information provided by the customer. The customer is responsible for the impact of the information provided on the validity of the results.
2. The report without China inspection body and laboratory Mandatory Approval (CMA) mark has no effect of proving to the society.
3. For the report with CNAS mark or A2LA mark, the items marked with "☆" are not within the accredited scope.
4. This report is invalid if it is altered, without the signature of the testing and approval personnel, or without the "inspection and testing dedicated stamp" or test report stamp.
5. The test data and results are only valid for the tested samples provided by the customer.
6. This report shall not be partially reproduced without the written permission of the laboratory.
7. Any objection shall be raised to the laboratory within 30 days after receiving the report.

--END OF REPORT--