

MAGPy-8H3D+E3DV2 SN:3075
MAGPy-DASV2 SN:3111

May 05, 2026

Dynamic Range, H-field, Channel 4

H-field/(A/m) Applied			H-field/(A/m) Reading			Difference/(dB)			Tolerance/(dB)
x	y	z	x	y	z	x	y	z	
0.390	0.390	0.390	0.390	0.400	0.400	0.00	0.22	0.22	±1.00
0.530	0.530	0.530	0.540	0.540	0.540	0.16	0.16	0.16	±1.00
0.730	0.730	0.720	0.720	0.720	0.720	-0.12	-0.12	0.00	±1.00
0.950	0.950	0.940	0.950	0.950	0.950	0.00	0.00	0.09	±1.00
1.29	1.28	1.27	1.30	1.29	1.28	0.07	0.07	0.07	±1.00
1.77	1.76	1.74	1.77	1.76	1.75	0.00	0.00	0.05	±1.00
2.36	2.34	2.32	2.36	2.35	2.33	0.00	0.04	0.04	±0.20
3.15	3.14	3.11	3.16	3.14	3.11	0.03	0.00	0.00	±0.20
4.29	4.26	4.21	4.29	4.26	4.21	0.00	0.00	0.00	±0.20
5.79	5.76	5.69	5.79	5.78	5.70	0.00	0.03	0.02	±0.20
7.80	7.76	7.67	7.79	7.77	7.66	-0.01	0.01	-0.01	±0.20
10.4	10.3	10.2	10.4	10.4	10.3	0.00	0.08	0.08	±0.20
14.0	14.0	13.8	14.0	14.0	13.8	0.00	0.00	0.00	±0.20
18.9	18.9	18.6	18.9	18.9	18.6	0.00	0.00	0.00	±0.20
25.5	25.5	25.1	25.5	25.6	25.2	0.00	0.03	0.03	±0.20
34.1	34.1	33.5	34.3	34.2	33.7	0.05	0.03	0.05	±0.20
46.0	46.0	45.3	46.2	46.2	45.5	0.04	0.04	0.04	±0.20
62.3	62.4	61.3	62.7	62.8	61.7	0.06	0.06	0.06	±0.20
85.6	85.9	84.3	85.2	85.5	84.0	-0.04	-0.04	-0.03	±0.20
112	113	110	112	112	110	0.00	-0.08	0.00	±0.20
154	155	151	153	154	151	-0.06	-0.06	0.00	±0.20
213	215	210	213	214	209	0.00	-0.04	-0.04	±0.20
295	298	290	297	292	292	0.06	-0.18	0.06	±0.20
436	440	429	428	434	422	-0.16	-0.12	-0.14	±0.20
601	607	593	594	604	587	-0.10	-0.04	-0.09	±0.20
897	908	885	897	911	886	0.00	0.03	0.01	±0.20
1360	1380	1340	1380	1400	1360	0.13	0.12	0.13	±0.30
1860	1880	1830	1900	1930	1880	0.18	0.23	0.23	±0.30
3000	3040	2970	3110	3160	3080	0.31	0.34	0.32	±0.50
3550	3580	3510	3690	3740	3650	0.34	0.38	0.34	±0.50

SPEAG H-field linearity tolerance criteria¹:

- ±1.0dB for applied H-fields < 2.0A/m
- ±0.2dB for applied H-fields ≥ 2.0A/m and < 1000 A/m
- ±0.3dB for applied H-fields ≥ 1000 A/m and < 2000 A/m
- ±0.4dB for applied H-fields ≥ 2000 A/m and < 3000 A/m
- ±0.5dB for applied H-fields ≥ 3000 A/m

¹Calibration uncertainty not taken into account (shared risk 50%).

MAGPy-8H3D+E3DV2 SN:3075
MAGPy-DASV2 SN:3111

May 05, 2026

Dynamic Range, H-field, Channel 5

H-field/(A/m) Applied			H-field/(A/m) Reading			Difference/(dB)			Tolerance/(dB)
x	y	z	x	y	z	x	y	z	
0.400	0.390	0.400	0.400	0.420	0.420	0.00	0.64	0.42	±1.00
0.540	0.540	0.540	0.540	0.560	0.540	0.00	0.32	0.00	±1.00
0.740	0.730	0.740	0.740	0.730	0.730	0.00	0.00	-0.12	±1.00
0.960	0.960	0.970	0.960	0.950	0.960	0.00	-0.09	-0.09	±1.00
1.30	1.29	1.31	1.31	1.30	1.32	0.07	0.07	0.07	±1.00
1.79	1.78	1.79	1.79	1.79	1.79	0.00	0.05	0.00	±1.00
2.38	2.37	2.39	2.39	2.38	2.41	0.04	0.04	0.07	±0.20
3.18	3.17	3.19	3.19	3.17	3.20	0.03	0.00	0.03	±0.20
4.32	4.30	4.33	4.33	4.31	4.34	0.02	0.02	0.02	±0.20
5.84	5.81	5.85	5.86	5.83	5.87	0.03	0.03	0.03	±0.20
7.87	7.84	7.89	7.86	7.85	7.88	-0.01	0.01	-0.01	±0.20
10.5	10.4	10.5	10.5	10.5	10.5	0.00	0.08	0.00	±0.20
14.2	14.1	14.2	14.2	14.2	14.2	0.00	0.06	0.00	±0.20
19.1	19.1	19.1	19.1	19.1	19.1	0.00	0.00	0.00	±0.20
25.7	25.7	25.8	25.8	25.8	25.8	0.03	0.03	0.00	±0.20
34.4	34.4	34.5	34.6	34.6	34.6	0.05	0.05	0.03	±0.20
46.4	46.5	46.5	46.6	46.7	46.8	0.04	0.04	0.06	±0.20
62.8	63.0	63.0	63.3	63.4	63.4	0.07	0.05	0.05	±0.20
86.4	86.7	86.7	86.0	86.4	86.3	-0.04	-0.03	-0.04	±0.20
113	114	113	112	113	113	-0.08	-0.08	0.00	±0.20
155	156	156	155	156	155	0.00	0.00	-0.06	±0.20
215	217	216	214	216	215	-0.04	-0.04	-0.04	±0.20
298	301	299	299	295	300	0.03	-0.17	0.03	±0.20
439	444	441	431	438	434	-0.16	-0.12	-0.14	±0.20
606	613	609	599	610	603	-0.10	-0.04	-0.09	±0.20
905	917	910	904	921	910	-0.01	0.04	0.00	±0.20
1370	1390	1380	1390	1420	1400	0.13	0.19	0.12	±0.30
1870	1900	1880	1910	1950	1930	0.18	0.23	0.23	±0.30
3030	3070	3050	3130	3200	3160	0.28	0.36	0.31	±0.50
3580	3620	3610	3720	3790	3750	0.33	0.40	0.33	±0.50

SPEAG H-field linearity tolerance criteria¹:

- ±1.0dB for applied H-fields < 2.0 A/m
- ±0.2dB for applied H-fields ≥ 2.0 A/m and < 1000 A/m
- ±0.3dB for applied H-fields ≥ 1000 A/m and < 2000 A/m
- ±0.4dB for applied H-fields ≥ 2000 A/m and < 3000 A/m
- ±0.5dB for applied H-fields ≥ 3000 A/m

¹ Calibration uncertainty not taken into account (shared risk 50%).

MAGPy-8H3D+E3DV2 SN:3075
MAGPy-DASV2 SN:3111

May 05, 2026

Dynamic Range, H-field, Channel 6

H-field/(A/m) Applied			H-field/(A/m) Reading			Difference/(dB)			Tolerance/(dB)
x	y	z	x	y	z	x	y	z	
0.400	0.400	0.390	0.420	0.400	0.400	0.42	0.00	0.22	±1.00
0.550	0.540	0.530	0.560	0.520	0.540	0.16	-0.33	0.16	±1.00
0.750	0.740	0.730	0.750	0.720	0.720	0.00	-0.24	-0.12	±1.00
0.970	0.960	0.950	0.970	0.960	0.950	0.00	0.00	0.00	±1.00
1.32	1.30	1.29	1.32	1.31	1.28	0.00	0.07	-0.07	±1.00
1.81	1.79	1.77	1.82	1.78	1.77	0.05	-0.05	0.00	±1.00
2.41	2.38	2.35	2.41	2.38	2.35	0.00	0.00	0.00	±0.20
3.22	3.18	3.15	3.22	3.19	3.11	0.00	0.03	-0.11	±0.20
4.38	4.32	4.27	4.38	4.32	4.23	0.00	0.00	-0.08	±0.20
5.92	5.85	5.77	5.92	5.85	5.74	0.00	0.00	-0.05	±0.20
7.98	7.88	7.78	7.96	7.87	7.74	-0.02	-0.01	-0.04	±0.20
10.6	10.5	10.4	10.6	10.6	10.4	0.00	0.08	0.00	±0.20
14.3	14.2	14.0	14.3	14.3	14.0	0.00	0.06	0.00	±0.20
19.3	19.2	18.9	19.3	19.2	18.9	0.00	0.00	0.00	±0.20
26.1	25.9	25.4	26.1	26.0	25.5	0.00	0.03	0.03	±0.20
34.8	34.6	34.0	35.0	34.8	34.1	0.05	0.05	0.03	±0.20
47.0	46.7	45.9	47.2	47.0	46.1	0.04	0.06	0.04	±0.20
63.7	63.4	62.2	64.1	63.7	62.5	0.05	0.04	0.04	±0.20
87.5	87.2	85.5	87.2	86.9	85.1	-0.03	-0.03	-0.04	±0.20
114	114	112	114	114	111	0.00	0.00	-0.08	±0.20
157	157	153	157	157	153	0.00	0.00	0.00	±0.20
218	218	213	217	218	212	-0.04	0.00	-0.04	±0.20
302	302	294	303	297	296	0.03	-0.15	0.06	±0.20
445	447	435	437	441	428	-0.16	-0.12	-0.14	±0.20
614	617	601	608	613	595	-0.09	-0.06	-0.09	±0.20
917	922	897	917	926	899	0.00	0.04	0.02	±0.20
1390	1400	1360	1410	1420	1380	0.12	0.12	0.13	±0.30
1900	1910	1860	1940	1960	1900	0.18	0.22	0.18	±0.30
3070	3090	3010	3180	3210	3120	0.31	0.33	0.31	±0.50
3620	3640	3560	3770	3810	3710	0.35	0.40	0.36	±0.50

SPEAG H-field linearity tolerance criteria¹:

- ±1.0dB for applied H-fields < 2.0A/m
- ±0.2dB for applied H-fields ≥ 2.0A/m and < 1000A/m
- ±0.3dB for applied H-fields ≥ 1000A/m and < 2000A/m
- ±0.4dB for applied H-fields ≥ 2000A/m and < 3000A/m
- ±0.5dB for applied H-fields ≥ 3000A/m

¹ Calibration uncertainty not taken into account (shared risk 50%).

MAGPy-8H3D+E3DV2 SN:3075
MAGPy-DASV2 SN:3111

May 05, 2026

Dynamic Range, H-field, Channel 7

H-field/(A/m) Applied			H-field/(A/m) Reading			Difference/(dB)			Tolerance/(dB)
x	y	z	x	y	z	x	y	z	
0.400	0.390	0.380	0.410	0.390	0.380	0.21	0.00	0.00	±1.00
0.540	0.530	0.520	0.540	0.530	0.500	0.00	0.00	-0.34	±1.00
0.740	0.720	0.710	0.720	0.720	0.700	-0.24	0.00	-0.12	±1.00
0.970	0.940	0.920	0.950	0.950	0.920	-0.18	0.09	0.00	±1.00
1.31	1.27	1.24	1.30	1.28	1.26	-0.07	0.07	0.14	±1.00
1.79	1.75	1.71	1.81	1.76	1.72	0.10	0.05	0.05	±1.00
2.39	2.33	2.28	2.38	2.35	2.28	-0.04	0.07	0.00	±0.20
3.20	3.12	3.04	3.18	3.13	3.06	-0.05	0.03	0.06	±0.20
4.34	4.23	4.13	4.33	4.25	4.13	-0.02	0.04	0.00	±0.20
5.87	5.72	5.58	5.87	5.75	5.58	0.00	0.05	0.00	±0.20
7.91	7.72	7.52	7.89	7.75	7.52	-0.02	0.03	0.00	±0.20
10.5	10.3	10.0	10.5	10.3	10.1	0.00	0.00	0.09	±0.20
14.2	13.9	13.5	14.2	13.9	13.6	0.00	0.00	0.06	±0.20
19.2	18.8	18.2	19.2	18.8	18.3	0.00	0.00	0.05	±0.20
25.9	25.3	24.6	25.8	25.4	24.7	-0.03	0.03	0.04	±0.20
34.5	33.9	32.9	34.7	34.0	33.0	0.05	0.03	0.03	±0.20
46.6	45.8	44.4	46.8	46.0	44.6	0.04	0.04	0.04	±0.20
63.1	62.0	60.1	63.5	62.4	60.5	0.05	0.06	0.06	±0.20
86.8	85.4	82.7	86.4	85.1	82.3	-0.04	-0.03	-0.04	±0.20
113	112	108	113	111	108	0.00	-0.08	0.00	±0.20
156	154	148	155	153	148	-0.06	-0.06	0.00	±0.20
216	213	206	215	213	205	-0.04	0.00	-0.04	±0.20
299	296	285	300	291	286	0.03	-0.15	0.03	±0.20
441	438	421	433	432	414	-0.16	-0.12	-0.15	±0.20
609	604	581	602	601	575	-0.10	-0.04	-0.09	±0.20
909	903	868	909	907	868	0.00	0.04	0.00	±0.20
1380	1370	1320	1390	1390	1330	0.06	0.13	0.07	±0.30
1880	1870	1800	1920	1920	1840	0.18	0.23	0.19	±0.30
3040	3020	2910	3150	3150	3020	0.31	0.37	0.32	±0.50
3590	3570	3440	3730	3730	3580	0.33	0.38	0.35	±0.50

SPEAG H-field linearity tolerance criteria¹:

- ±1.0dB for applied H-fields < 2.0A/m
- ±0.2dB for applied H-fields ≥ 2.0A/m and < 1000 A/m
- ±0.3dB for applied H-fields ≥ 1000 A/m and < 2000 A/m
- ±0.4dB for applied H-fields ≥ 2000 A/m and < 3000 A/m
- ±0.5dB for applied H-fields ≥ 3000 A/m

¹ Calibration uncertainty not taken into account (shared risk 50%).

MAGPy-8H3D+E3DV2 SN:3075
MAGPy-DASV2 SN:3111

May 05, 2026

Dynamic Range, E-field, Channel 0

E-field/(V/m) Applied			E-field/(V/m) Reading			Difference/(dB)			Tolerance/(dB)		
x	y	z	x	y	z	x	y	z	x	y	z
0.360	0.220	0.130	0.360	0.230	0.150	0.00	0.39	1.24	±5.00	±5.00	±5.00
0.500	0.300	0.170	0.500	0.300	0.190	0.00	0.00	0.97	±5.00	±5.00	±5.00
0.680	0.410	0.240	0.690	0.400	0.230	0.13	-0.21	-0.37	±5.00	±5.00	±5.00
0.890	0.530	0.310	0.900	0.540	0.290	0.10	0.16	-0.58	±5.00	±5.00	±5.00
1.20	0.720	0.420	1.21	0.740	0.420	0.07	0.24	0.00	±5.00	±5.00	±5.00
1.64	0.990	0.570	1.67	0.990	0.580	0.16	0.00	0.15	±5.00	±5.00	±5.00
2.19	1.31	0.760	2.22	1.31	0.730	0.12	0.00	-0.35	±1.00	±5.00	±5.00
2.93	1.76	1.02	2.96	1.74	1.03	0.09	-0.10	0.08	±1.00	±5.00	±5.00
3.98	2.39	1.39	4.00	2.39	1.36	0.04	0.00	-0.19	±1.00	±1.00	±5.00
5.38	3.23	1.87	5.43	3.22	1.87	0.08	-0.03	0.00	±1.00	±1.00	±5.00
7.26	4.35	2.53	7.31	4.35	2.50	0.06	0.00	-0.10	±1.00	±1.00	±1.00
9.68	5.80	3.37	9.74	5.81	3.37	0.05	0.01	0.00	±1.00	±1.00	±1.00
13.1	7.84	4.54	13.2	7.84	4.50	0.07	0.00	-0.08	±1.00	±1.00	±1.00
17.7	10.6	6.12	17.7	10.6	6.05	0.00	0.00	-0.10	±1.00	±1.00	±1.00
23.8	14.3	8.26	24.0	14.3	8.19	0.07	0.00	-0.07	±1.00	±1.00	±1.00
31.9	19.1	11.0	32.2	19.1	11.0	0.08	0.00	0.00	±1.00	±1.00	±1.00
43.1	25.8	14.9	43.5	25.8	14.8	0.08	0.00	-0.06	±1.00	±1.00	±1.00
58.4	35.0	20.2	58.9	35.1	20.1	0.07	0.02	-0.04	±1.00	±1.00	±1.00
80.3	48.1	27.7	80.3	47.8	27.3	0.00	-0.05	-0.13	±1.00	±1.00	±1.00
105	63.1	36.2	105	62.6	35.7	0.00	-0.07	-0.12	±1.00	±1.00	±1.00
145	86.7	49.8	145	86.1	49.1	0.00	-0.06	-0.12	±1.00	±1.00	±1.00
201	120	69.0	201	120	68.1	0.00	0.00	-0.11	±1.00	±1.00	±1.00
278	167	95.5	281	167	95.0	0.09	0.00	-0.05	±1.00	±1.00	±1.00
412	247	141	396	235	141	-0.34	-0.43	0.00	±1.00	±1.00	±1.00
568	341	194	551	327	196	-0.26	-0.36	0.09	±1.00	±1.00	±1.00
849	509	290	832	494	296	-0.18	-0.26	0.18	±1.00	±1.00	±1.00
1290	772	440	1280	760	455	-0.07	-0.14	0.29	±1.00	±1.00	±1.00
1760	1050	601	1770	1050	627	0.05	0.00	0.37	±1.00	±1.00	±1.00
2850	1710	972	2900	1720	975	0.15	0.05	0.03	±1.00	±1.00	±1.00
3350	2010	1150	3430	2040	1160	0.20	0.13	0.08	±1.00	±1.00	±1.00

SPEAG E-field linearity tolerance criteria¹:

±5.0dB for applied E-field < 2V/m

±1.0dB for applied E-field ≥ 2V/m

¹ Calibration uncertainty not taken into account (shared risk 50%).

MAGPy-8H3D+E3DV2 SN:3075
MAGPy-DASV2 SN:3111

May 05, 2026

Frequency Response

Frequency Response, H-field, Channel 0

f/(Hz)	H-field/(A/m) Applied			H-field/(A/m) Reading			Difference/(dB)			Tolerance/(dB)
	x	y	z	x	y	z	x	y	z	
3000	3.23	3.23	3.23	3.24	3.24	3.23	0.03	0.03	0.00	±0.3
3200	3.25	3.25	3.25	3.28	3.26	3.27	0.08	0.03	0.05	±0.3
4000	3.25	3.25	3.25	3.26	3.26	3.25	0.03	0.03	0.00	±0.3
5200	3.23	3.23	3.23	3.24	3.21	3.22	0.03	-0.05	-0.03	±0.3
6600	3.20	3.20	3.20	3.21	3.21	3.20	0.03	0.03	0.00	±0.3
8200	3.16	3.16	3.16	3.20	3.18	3.17	0.11	0.05	0.03	±0.3
9000	3.15	3.15	3.15	3.15	3.15	3.15	0.00	0.00	0.00	±0.3
10600	4.37	4.29	4.28	4.37	4.30	4.28	0.00	0.02	0.00	±0.3
13400	4.37	4.31	4.30	4.37	4.32	4.29	0.00	0.02	-0.02	±0.3
17000	4.37	4.31	4.31	4.38	4.33	4.31	0.02	0.04	0.00	±0.3
21400	4.40	4.34	4.33	4.43	4.34	4.34	0.06	0.00	0.02	±0.3
27200	4.40	4.34	4.33	4.41	4.34	4.33	0.02	0.00	0.00	±0.3
34400	4.40	4.36	4.34	4.41	4.36	4.35	0.02	0.00	0.02	±0.3
40000	4.39	4.34	4.34	4.41	4.35	4.34	0.04	0.02	0.00	±0.3
43600	4.39	4.34	4.33	4.39	4.36	4.34	0.00	0.04	0.02	±0.3
55400	4.38	4.34	4.33	4.39	4.34	4.33	0.02	0.00	0.00	±0.3
70000	4.37	4.32	4.31	4.37	4.33	4.32	0.00	0.02	0.02	±0.3
88800	4.35	4.31	4.30	4.35	4.32	4.30	0.00	0.02	0.00	±0.3
112400	4.33	4.30	4.29	4.34	4.30	4.29	0.02	0.00	0.00	±0.3
142400	4.31	4.27	4.27	4.31	4.28	4.27	0.00	0.02	0.00	±0.3
161750	4.30	4.26	4.26	4.30	4.27	4.26	0.00	0.02	0.00	±0.3
180400	4.29	4.25	4.24	4.29	4.26	4.24	0.00	0.02	0.00	±0.3
228400	4.26	4.22	4.21	4.26	4.22	4.21	0.00	0.00	0.00	±0.3
289400	4.21	4.18	4.17	4.21	4.18	4.17	0.00	0.00	0.00	±0.3
366400	4.17	4.14	4.13	4.17	4.14	4.13	0.00	0.00	0.00	±0.3
400000	4.15	4.12	4.11	4.15	4.12	4.11	0.00	0.00	0.00	±0.3
464000	4.12	4.09	4.08	4.12	4.10	4.09	0.00	0.02	0.02	±0.3
587800	4.07	4.04	4.04	4.07	4.05	4.03	0.00	0.02	-0.02	±0.3
744200	4.02	4.00	4.00	4.03	4.01	4.00	0.02	0.02	0.00	±0.3
942600	4.00	3.98	3.98	4.00	3.99	3.97	0.00	0.02	-0.02	±0.3
1193600	3.98	3.96	3.95	3.98	3.96	3.95	0.00	0.00	0.00	±0.3
1511600	3.97	3.96	3.95	3.98	3.96	3.95	0.02	0.00	0.00	±0.3
1914400	3.96	3.94	3.93	3.96	3.95	3.93	0.00	0.02	0.00	±0.3
2424400	3.93	3.91	3.90	3.93	3.91	3.90	0.00	0.00	0.00	±0.3
3070200	3.89	3.87	3.87	3.89	3.88	3.86	0.00	0.02	-0.02	±0.3
3888000	3.83	3.82	3.81	3.84	3.82	3.81	0.02	0.00	0.00	±0.3
4000000	3.82	3.81	3.80	3.83	3.81	3.80	0.02	0.00	0.00	±0.3
4923800	3.75	3.73	3.72	3.75	3.73	3.73	0.00	0.00	0.02	±0.3
6235400	3.64	3.63	3.62	3.65	3.63	3.62	0.02	0.00	0.00	±0.3
7896400	3.50	3.49	3.48	3.51	3.50	3.48	0.02	0.02	0.00	±0.3
10000000	3.35	3.34	3.33	3.35	3.33	3.33	0.00	-0.03	0.00	±0.3

SPEAG H-field frequency response tolerance criteria¹:
±0.3dB for applied H-fields at calibration points from 3kHz to 10MHz

¹ Calibration uncertainty not taken into account (shared risk 50%).

MAGPy-8H3D+E3DV2 SN:3075
MAGPy-DASV2 SN:3111

May 05, 2026

Frequency Response, H-field, Channel 1

f/(Hz)	H-field/(A/m) Applied			H-field/(A/m) Reading			Difference/(dB)			Tolerance/(dB)
	x	y	z	x	y	z	x	y	z	
3000	3.23	3.23	3.23	3.23	3.24	3.23	0.00	0.03	0.00	±0.3
3200	3.25	3.25	3.25	3.27	3.25	3.28	0.05	0.00	0.08	±0.3
4000	3.25	3.25	3.25	3.26	3.26	3.26	0.03	0.03	0.03	±0.3
5200	3.23	3.23	3.23	3.24	3.21	3.22	0.03	-0.05	-0.03	±0.3
6600	3.20	3.20	3.20	3.21	3.21	3.20	0.03	0.03	0.00	±0.3
8200	3.16	3.16	3.16	3.19	3.17	3.17	0.08	0.03	0.03	±0.3
9000	3.15	3.15	3.15	3.15	3.15	3.15	0.00	0.00	0.00	±0.3
10600	4.37	4.29	4.28	4.38	4.29	4.28	0.02	0.00	0.00	±0.3
13400	4.37	4.31	4.30	4.38	4.32	4.30	0.02	0.02	0.00	±0.3
17000	4.37	4.31	4.31	4.39	4.32	4.30	0.04	0.02	-0.02	±0.3
21400	4.40	4.34	4.33	4.39	4.34	4.33	-0.02	0.00	0.00	±0.3
27200	4.40	4.34	4.33	4.40	4.33	4.33	0.00	-0.02	0.00	±0.3
34400	4.40	4.36	4.34	4.41	4.36	4.35	0.02	0.00	0.02	±0.3
40000	4.39	4.34	4.34	4.40	4.35	4.35	0.02	0.02	0.02	±0.3
43600	4.39	4.34	4.33	4.38	4.34	4.34	-0.02	0.00	0.02	±0.3
55400	4.38	4.34	4.33	4.38	4.34	4.33	0.00	0.00	0.00	±0.3
70000	4.37	4.32	4.31	4.37	4.33	4.32	0.00	0.02	0.02	±0.3
88800	4.35	4.31	4.30	4.35	4.31	4.30	0.00	0.00	0.00	±0.3
112400	4.33	4.30	4.29	4.34	4.30	4.29	0.02	0.00	0.00	±0.3
142400	4.31	4.27	4.27	4.32	4.28	4.27	0.02	0.02	0.00	±0.3
161750	4.30	4.26	4.26	4.30	4.27	4.26	0.00	0.02	0.00	±0.3
180400	4.29	4.25	4.24	4.28	4.25	4.25	-0.02	0.00	0.02	±0.3
228400	4.26	4.22	4.21	4.26	4.22	4.21	0.00	0.00	0.00	±0.3
289400	4.21	4.18	4.17	4.20	4.18	4.17	-0.02	0.00	0.00	±0.3
366400	4.17	4.14	4.13	4.17	4.13	4.13	0.00	-0.02	0.00	±0.3
400000	4.15	4.12	4.11	4.16	4.12	4.11	0.02	0.00	0.00	±0.3
464000	4.12	4.09	4.08	4.12	4.09	4.09	0.00	0.00	0.02	±0.3
587800	4.07	4.04	4.04	4.07	4.04	4.03	0.00	0.00	-0.02	±0.3
744200	4.02	4.00	4.00	4.03	4.00	4.00	0.02	0.00	0.00	±0.3
942600	4.00	3.98	3.98	4.00	3.98	3.98	0.00	0.00	0.00	±0.3
1193600	3.98	3.96	3.95	3.98	3.96	3.95	0.00	0.00	0.00	±0.3
1511600	3.97	3.96	3.95	3.98	3.96	3.95	0.02	0.00	0.00	±0.3
1914400	3.96	3.94	3.93	3.96	3.94	3.94	0.00	0.00	0.02	±0.3
2424400	3.93	3.91	3.90	3.93	3.91	3.90	0.00	0.00	0.00	±0.3
3070200	3.89	3.87	3.87	3.90	3.87	3.86	0.02	0.00	-0.02	±0.3
3888000	3.83	3.82	3.81	3.84	3.82	3.81	0.02	0.00	0.00	±0.3
4000000	3.82	3.81	3.80	3.82	3.81	3.80	0.00	0.00	0.00	±0.3
4923800	3.75	3.73	3.72	3.75	3.73	3.72	0.00	0.00	0.00	±0.3
6235400	3.64	3.63	3.62	3.65	3.63	3.62	0.02	0.00	0.00	±0.3
7896400	3.50	3.49	3.48	3.51	3.50	3.49	0.02	0.02	0.02	±0.3
10000000	3.35	3.34	3.33	3.35	3.33	3.32	0.00	-0.03	-0.03	±0.3

SPEAG H-field frequency response tolerance criteria¹:
±0.3dB for applied H-fields at calibration points from 3kHz to 10MHz

¹ Calibration uncertainty not taken into account (shared risk 50%).

MAGPy-8H3D+E3DV2 SN:3075
MAGPy-DASV2 SN:3111

May 05, 2026

Frequency Response, H-field, Channel 2

f/(Hz)	H-field/(A/m) Applied			H-field/(A/m) Reading			Difference/(dB)			Tolerance/(dB)
	x	y	z	x	y	z	x	y	z	
3000	3.23	3.23	3.23	3.23	3.24	3.23	0.00	0.03	0.00	±0.3
3200	3.25	3.25	3.25	3.27	3.26	3.27	0.05	0.03	0.05	±0.3
4000	3.25	3.25	3.25	3.25	3.26	3.25	0.00	0.03	0.00	±0.3
5200	3.23	3.23	3.23	3.25	3.21	3.22	0.05	-0.05	-0.03	±0.3
6600	3.20	3.20	3.20	3.20	3.21	3.21	0.00	0.03	0.03	±0.3
8200	3.16	3.16	3.16	3.19	3.17	3.17	0.08	0.03	0.03	±0.3
9000	3.15	3.15	3.15	3.15	3.15	3.15	0.00	0.00	0.00	±0.3
10600	4.37	4.29	4.28	4.38	4.29	4.28	0.02	0.00	0.00	±0.3
13400	4.37	4.31	4.30	4.35	4.31	4.30	-0.04	0.00	0.00	±0.3
17000	4.37	4.31	4.31	4.38	4.32	4.30	0.02	0.02	-0.02	±0.3
21400	4.40	4.34	4.33	4.39	4.34	4.32	-0.02	0.00	-0.02	±0.3
27200	4.40	4.34	4.33	4.39	4.35	4.33	-0.02	0.02	0.00	±0.3
34400	4.40	4.36	4.34	4.41	4.35	4.35	0.02	-0.02	0.02	±0.3
40000	4.39	4.34	4.34	4.40	4.35	4.34	0.02	0.02	0.00	±0.3
43600	4.39	4.34	4.33	4.38	4.34	4.33	-0.02	0.00	0.00	±0.3
55400	4.38	4.34	4.33	4.39	4.34	4.33	0.02	0.00	0.00	±0.3
70000	4.37	4.32	4.31	4.37	4.32	4.31	0.00	0.00	0.00	±0.3
88800	4.35	4.31	4.30	4.35	4.31	4.30	0.00	0.00	0.00	±0.3
112400	4.33	4.30	4.29	4.34	4.30	4.29	0.02	0.00	0.00	±0.3
142400	4.31	4.27	4.27	4.32	4.28	4.27	0.02	0.02	0.00	±0.3
161750	4.30	4.26	4.26	4.30	4.27	4.26	0.00	0.02	0.00	±0.3
180400	4.29	4.25	4.24	4.29	4.25	4.24	0.00	0.00	0.00	±0.3
228400	4.26	4.22	4.21	4.26	4.23	4.21	0.00	0.02	0.00	±0.3
289400	4.21	4.18	4.17	4.21	4.18	4.17	0.00	0.00	0.00	±0.3
366400	4.17	4.14	4.13	4.17	4.14	4.13	0.00	0.00	0.00	±0.3
400000	4.15	4.12	4.11	4.15	4.12	4.11	0.00	0.00	0.00	±0.3
464000	4.12	4.09	4.08	4.12	4.10	4.09	0.00	0.02	0.02	±0.3
587800	4.07	4.04	4.04	4.07	4.04	4.04	0.00	0.00	0.00	±0.3
744200	4.02	4.00	4.00	4.02	4.00	4.00	0.00	0.00	0.00	±0.3
942600	4.00	3.98	3.98	4.00	3.99	3.98	0.00	0.02	0.00	±0.3
1193600	3.98	3.96	3.95	3.98	3.96	3.95	0.00	0.00	0.00	±0.3
1511600	3.97	3.96	3.95	3.98	3.96	3.94	0.02	0.00	-0.02	±0.3
1914400	3.96	3.94	3.93	3.96	3.94	3.93	0.00	0.00	0.00	±0.3
2424400	3.93	3.91	3.90	3.93	3.91	3.90	0.00	0.00	0.00	±0.3
3070200	3.89	3.87	3.87	3.89	3.88	3.86	0.00	0.02	-0.02	±0.3
3888000	3.83	3.82	3.81	3.84	3.82	3.81	0.02	0.00	0.00	±0.3
4000000	3.82	3.81	3.80	3.83	3.81	3.80	0.02	0.00	0.00	±0.3
4923800	3.75	3.73	3.72	3.75	3.73	3.72	0.00	0.00	0.00	±0.3
6235400	3.64	3.63	3.62	3.65	3.63	3.62	0.02	0.00	0.00	±0.3
7896400	3.50	3.49	3.48	3.51	3.50	3.48	0.02	0.02	0.00	±0.3
10000000	3.35	3.34	3.33	3.34	3.34	3.34	-0.03	0.00	0.03	±0.3

SPEAG H-field frequency response tolerance criteria¹:
±0.3dB for applied H-fields at calibration points from 3kHz to 10MHz

¹ Calibration uncertainty not taken into account (shared risk 50%).

MAGPy-8H3D+E3DV2 SN:3075
MAGPy-DASV2 SN:3111

May 05, 2026

Frequency Response, H-field, Channel 3

f/(Hz)	H-field/(A/m) Applied			H-field/(A/m) Reading			Difference/(dB)			Tolerance/(dB)
	x	y	z	x	y	z	x	y	z	
3000	3.23	3.23	3.23	3.23	3.25	3.23	0.00	0.05	0.00	±0.3
3200	3.25	3.25	3.25	3.27	3.26	3.27	0.05	0.03	0.05	±0.3
4000	3.25	3.25	3.25	3.25	3.26	3.25	0.00	0.03	0.00	±0.3
5200	3.23	3.23	3.23	3.23	3.21	3.22	0.00	-0.05	-0.03	±0.3
6600	3.20	3.20	3.20	3.21	3.21	3.21	0.03	0.03	0.03	±0.3
8200	3.16	3.16	3.16	3.18	3.18	3.17	0.05	0.05	0.03	±0.3
9000	3.15	3.15	3.15	3.15	3.15	3.15	0.00	0.00	0.00	±0.3
10600	4.37	4.29	4.28	4.38	4.29	4.28	0.02	0.00	0.00	±0.3
13400	4.37	4.31	4.30	4.38	4.32	4.31	0.02	0.02	0.02	±0.3
17000	4.37	4.31	4.31	4.39	4.31	4.31	0.04	0.00	0.00	±0.3
21400	4.40	4.34	4.33	4.41	4.35	4.34	0.02	0.02	0.02	±0.3
27200	4.40	4.34	4.33	4.42	4.34	4.34	0.04	0.00	0.02	±0.3
34400	4.40	4.36	4.34	4.41	4.36	4.35	0.02	0.00	0.02	±0.3
40000	4.39	4.34	4.34	4.40	4.35	4.34	0.02	0.02	0.00	±0.3
43600	4.39	4.34	4.33	4.39	4.34	4.34	0.00	0.00	0.02	±0.3
55400	4.38	4.34	4.33	4.38	4.34	4.33	0.00	0.00	0.00	±0.3
70000	4.37	4.32	4.31	4.37	4.33	4.31	0.00	0.02	0.00	±0.3
88800	4.35	4.31	4.30	4.35	4.32	4.30	0.00	0.02	0.00	±0.3
112400	4.33	4.30	4.29	4.34	4.30	4.29	0.02	0.00	0.00	±0.3
142400	4.31	4.27	4.27	4.31	4.28	4.27	0.00	0.02	0.00	±0.3
161750	4.30	4.26	4.26	4.31	4.27	4.26	0.02	0.02	0.00	±0.3
180400	4.29	4.25	4.24	4.29	4.25	4.25	0.00	0.00	0.02	±0.3
228400	4.26	4.22	4.21	4.26	4.22	4.22	0.00	0.00	0.02	±0.3
289400	4.21	4.18	4.17	4.21	4.18	4.17	0.00	0.00	0.00	±0.3
366400	4.17	4.14	4.13	4.18	4.13	4.13	0.02	-0.02	0.00	±0.3
400000	4.15	4.12	4.11	4.16	4.12	4.12	0.02	0.00	0.02	±0.3
464000	4.12	4.09	4.08	4.12	4.09	4.09	0.00	0.00	0.02	±0.3
587800	4.07	4.04	4.04	4.07	4.05	4.04	0.00	0.02	0.00	±0.3
744200	4.02	4.00	4.00	4.03	4.01	4.00	0.02	0.02	0.00	±0.3
942600	4.00	3.98	3.98	4.00	3.99	3.98	0.00	0.02	0.00	±0.3
1193600	3.98	3.96	3.95	3.98	3.96	3.95	0.00	0.00	0.00	±0.3
1511600	3.97	3.96	3.95	3.98	3.96	3.95	0.02	0.00	0.00	±0.3
1914400	3.96	3.94	3.93	3.96	3.94	3.93	0.00	0.00	0.00	±0.3
2424400	3.93	3.91	3.90	3.93	3.91	3.91	0.00	0.00	0.02	±0.3
3070200	3.89	3.87	3.87	3.90	3.88	3.87	0.02	0.02	0.00	±0.3
3888000	3.83	3.82	3.81	3.84	3.82	3.81	0.02	0.00	0.00	±0.3
4000000	3.82	3.81	3.80	3.83	3.81	3.80	0.02	0.00	0.00	±0.3
4923800	3.75	3.73	3.72	3.75	3.74	3.73	0.00	0.02	0.02	±0.3
6235400	3.64	3.63	3.62	3.65	3.63	3.62	0.02	0.00	0.00	±0.3
7896400	3.50	3.49	3.48	3.51	3.49	3.49	0.02	0.00	0.02	±0.3
10000000	3.35	3.34	3.33	3.35	3.33	3.33	0.00	-0.03	0.00	±0.3

SPEAG H-field frequency response tolerance criteria¹:
±0.3dB for applied H-fields at calibration points from 3kHz to 10MHz

¹ Calibration uncertainty not taken into account (shared risk 50%).

MAGPy-8H3D+E3DV2 SN:3075
MAGPy-DASV2 SN:3111

May 05, 2026

Frequency Response, H-field, Channel 4

f/(Hz)	H-field/(A/m) Applied			H-field/(A/m) Reading			Difference/(dB)			Tolerance/(dB)
	x	y	z	x	y	z	x	y	z	
3000	3.23	3.23	3.23	3.23	3.24	3.23	0.00	0.03	0.00	±0.3
3200	3.25	3.25	3.25	3.27	3.25	3.27	0.05	0.00	0.05	±0.3
4000	3.25	3.25	3.25	3.25	3.26	3.25	0.00	0.03	0.00	±0.3
5200	3.23	3.23	3.23	3.24	3.21	3.22	0.03	-0.05	-0.03	±0.3
6600	3.20	3.20	3.20	3.21	3.21	3.20	0.03	0.03	0.00	±0.3
8200	3.16	3.16	3.16	3.19	3.18	3.17	0.08	0.05	0.03	±0.3
9000	3.15	3.15	3.15	3.15	3.15	3.15	0.00	0.00	0.00	±0.3
10600	4.37	4.29	4.28	4.37	4.29	4.28	0.00	0.00	0.00	±0.3
13400	4.37	4.31	4.30	4.37	4.32	4.30	0.00	0.02	0.00	±0.3
17000	4.37	4.31	4.31	4.38	4.31	4.31	0.02	0.00	0.00	±0.3
21400	4.40	4.34	4.33	4.40	4.34	4.34	0.00	0.00	0.02	±0.3
27200	4.40	4.34	4.33	4.40	4.34	4.33	0.00	0.00	0.00	±0.3
34400	4.40	4.36	4.34	4.41	4.36	4.35	0.02	0.00	0.02	±0.3
40000	4.39	4.34	4.34	4.41	4.35	4.34	0.04	0.02	0.00	±0.3
43600	4.39	4.34	4.33	4.40	4.34	4.33	0.02	0.00	0.00	±0.3
55400	4.38	4.34	4.33	4.38	4.34	4.32	0.00	0.00	-0.02	±0.3
70000	4.37	4.32	4.31	4.36	4.32	4.31	-0.02	0.00	0.00	±0.3
88800	4.35	4.31	4.30	4.35	4.31	4.30	0.00	0.00	0.00	±0.3
112400	4.33	4.30	4.29	4.33	4.29	4.30	0.00	-0.02	0.02	±0.3
142400	4.31	4.27	4.27	4.31	4.27	4.26	0.00	0.00	-0.02	±0.3
161750	4.30	4.26	4.26	4.30	4.26	4.26	0.00	0.00	0.00	±0.3
180400	4.29	4.25	4.24	4.29	4.25	4.24	0.00	0.00	0.00	±0.3
228400	4.26	4.22	4.21	4.25	4.22	4.21	-0.02	0.00	0.00	±0.3
289400	4.21	4.18	4.17	4.21	4.18	4.17	0.00	0.00	0.00	±0.3
366400	4.17	4.14	4.13	4.17	4.13	4.13	0.00	-0.02	0.00	±0.3
400000	4.15	4.12	4.11	4.15	4.12	4.11	0.00	0.00	0.00	±0.3
464000	4.12	4.09	4.08	4.12	4.09	4.08	0.00	0.00	0.00	±0.3
587800	4.07	4.04	4.04	4.07	4.05	4.04	0.00	0.02	0.00	±0.3
744200	4.02	4.00	4.00	4.02	4.01	4.00	0.00	0.02	0.00	±0.3
942600	4.00	3.98	3.98	4.00	3.98	3.97	0.00	0.00	-0.02	±0.3
1193600	3.98	3.96	3.95	3.97	3.96	3.95	-0.02	0.00	0.00	±0.3
1511600	3.97	3.96	3.95	3.98	3.96	3.95	0.02	0.00	0.00	±0.3
1914400	3.96	3.94	3.93	3.96	3.94	3.93	0.00	0.00	0.00	±0.3
2424400	3.93	3.91	3.90	3.93	3.91	3.90	0.00	0.00	0.00	±0.3
3070200	3.89	3.87	3.87	3.89	3.88	3.87	0.00	0.02	0.00	±0.3
3888000	3.83	3.82	3.81	3.84	3.82	3.81	0.02	0.00	0.00	±0.3
4000000	3.82	3.81	3.80	3.83	3.81	3.80	0.02	0.00	0.00	±0.3
4923800	3.75	3.73	3.72	3.75	3.73	3.72	0.00	0.00	0.00	±0.3
6235400	3.64	3.63	3.62	3.64	3.63	3.62	0.00	0.00	0.00	±0.3
7896400	3.50	3.49	3.48	3.50	3.50	3.48	0.00	0.02	0.00	±0.3
10000000	3.35	3.34	3.33	3.35	3.34	3.33	0.00	0.00	0.00	±0.3

SPEAG H-field frequency response tolerance criteria¹:
±0.3dB for applied H-fields at calibration points from 3kHz to 10MHz

¹Calibration uncertainty not taken into account (shared risk 50%).

MAGPy-8H3D+E3DV2 SN:3075
MAGPy-DASV2 SN:3111

May 05, 2026

Frequency Response, H-field, Channel 5

f/(Hz)	H-field/(A/m) Applied			H-field/(A/m) Reading			Difference/(dB)			Tolerance/(dB)
	x	y	z	x	y	z	x	y	z	
3000	3.23	3.23	3.23	3.23	3.23	3.23	0.00	0.00	0.00	±0.3
3200	3.25	3.25	3.25	3.27	3.26	3.27	0.05	0.03	0.05	±0.3
4000	3.25	3.25	3.25	3.25	3.25	3.25	0.00	0.00	0.00	±0.3
5200	3.23	3.23	3.23	3.24	3.21	3.22	0.03	-0.05	-0.03	±0.3
6600	3.20	3.20	3.20	3.21	3.21	3.21	0.03	0.03	0.03	±0.3
8200	3.16	3.16	3.16	3.18	3.18	3.17	0.05	0.05	0.03	±0.3
9000	3.15	3.15	3.15	3.15	3.15	3.15	0.00	0.00	0.00	±0.3
10600	4.37	4.29	4.28	4.34	4.31	4.29	-0.06	0.04	0.02	±0.3
13400	4.37	4.31	4.30	4.37	4.33	4.30	0.00	0.04	0.00	±0.3
17000	4.37	4.31	4.31	4.38	4.33	4.31	0.02	0.04	0.00	±0.3
21400	4.40	4.34	4.33	4.39	4.35	4.34	-0.02	0.02	0.02	±0.3
27200	4.40	4.34	4.33	4.40	4.34	4.34	0.00	0.00	0.02	±0.3
34400	4.40	4.36	4.34	4.41	4.36	4.34	0.02	0.00	0.00	±0.3
40000	4.39	4.34	4.34	4.40	4.35	4.34	0.02	0.02	0.00	±0.3
43600	4.39	4.34	4.33	4.40	4.36	4.33	0.02	0.04	0.00	±0.3
55400	4.38	4.34	4.33	4.38	4.34	4.33	0.00	0.00	0.00	±0.3
70000	4.37	4.32	4.31	4.37	4.33	4.32	0.00	0.02	0.02	±0.3
88800	4.35	4.31	4.30	4.35	4.32	4.30	0.00	0.02	0.00	±0.3
112400	4.33	4.30	4.29	4.33	4.30	4.29	0.00	0.00	0.00	±0.3
142400	4.31	4.27	4.27	4.31	4.28	4.27	0.00	0.02	0.00	±0.3
161750	4.30	4.26	4.26	4.30	4.27	4.26	0.00	0.02	0.00	±0.3
180400	4.29	4.25	4.24	4.29	4.25	4.24	0.00	0.00	0.00	±0.3
228400	4.26	4.22	4.21	4.25	4.22	4.21	-0.02	0.00	0.00	±0.3
289400	4.21	4.18	4.17	4.21	4.18	4.17	0.00	0.00	0.00	±0.3
366400	4.17	4.14	4.13	4.17	4.14	4.13	0.00	0.00	0.00	±0.3
400000	4.15	4.12	4.11	4.15	4.12	4.12	0.00	0.00	0.02	±0.3
464000	4.12	4.09	4.08	4.12	4.09	4.09	0.00	0.00	0.02	±0.3
587800	4.07	4.04	4.04	4.07	4.05	4.04	0.00	0.02	0.00	±0.3
744200	4.02	4.00	4.00	4.03	4.01	4.00	0.02	0.02	0.00	±0.3
942600	4.00	3.98	3.98	4.00	3.98	3.98	0.00	0.00	0.00	±0.3
1193600	3.98	3.96	3.95	3.98	3.96	3.95	0.00	0.00	0.00	±0.3
1511600	3.97	3.96	3.95	3.98	3.96	3.95	0.02	0.00	0.00	±0.3
1914400	3.96	3.94	3.93	3.96	3.94	3.94	0.00	0.00	0.02	±0.3
2424400	3.93	3.91	3.90	3.93	3.91	3.91	0.00	0.00	0.02	±0.3
3070200	3.89	3.87	3.87	3.89	3.87	3.87	0.00	0.00	0.00	±0.3
3888000	3.83	3.82	3.81	3.84	3.82	3.81	0.02	0.00	0.00	±0.3
4000000	3.82	3.81	3.80	3.83	3.81	3.80	0.02	0.00	0.00	±0.3
4923800	3.75	3.73	3.72	3.75	3.73	3.72	0.00	0.00	0.00	±0.3
6235400	3.64	3.63	3.62	3.65	3.63	3.62	0.02	0.00	0.00	±0.3
7896400	3.50	3.49	3.48	3.50	3.50	3.50	0.00	0.02	0.05	±0.3
10000000	3.35	3.34	3.33	3.35	3.33	3.34	0.00	-0.03	0.03	±0.3

SPEAG H-field frequency response tolerance criteria¹:
±0.3dB for applied H-fields at calibration points from 3kHz to 10MHz

¹ Calibration uncertainty not taken into account (shared risk 50%).

MAGPy-8H3D+E3DV2 SN:3075
MAGPy-DASV2 SN:3111

May 05, 2026

Frequency Response, H-field, Channel 6

f/(Hz)	H-field/(A/m) Applied			H-field/(A/m) Reading			Difference/(dB)			Tolerance/(dB)
	x	y	z	x	y	z	x	y	z	
3000	3.23	3.23	3.23	3.23	3.24	3.23	0.00	0.03	0.00	±0.3
3200	3.25	3.25	3.25	3.27	3.26	3.26	0.05	0.03	0.03	±0.3
4000	3.25	3.25	3.25	3.26	3.26	3.25	0.03	0.03	0.00	±0.3
5200	3.23	3.23	3.23	3.24	3.21	3.22	0.03	-0.05	-0.03	±0.3
6600	3.20	3.20	3.20	3.21	3.21	3.20	0.03	0.03	0.00	±0.3
8200	3.16	3.16	3.16	3.19	3.18	3.17	0.08	0.05	0.03	±0.3
9000	3.15	3.15	3.15	3.15	3.15	3.14	0.00	0.00	-0.03	±0.3
10600	4.37	4.29	4.28	4.37	4.33	4.29	0.00	0.08	0.02	±0.3
13400	4.37	4.31	4.30	4.36	4.31	4.31	0.02	0.00	0.02	±0.3
17000	4.37	4.31	4.31	4.38	4.32	4.31	0.02	0.02	0.00	±0.3
21400	4.40	4.34	4.33	4.41	4.33	4.34	0.02	-0.02	0.02	±0.3
27200	4.40	4.34	4.33	4.41	4.36	4.34	0.02	0.04	0.02	±0.3
34400	4.40	4.36	4.34	4.40	4.36	4.35	0.00	0.00	0.02	±0.3
40000	4.39	4.34	4.34	4.40	4.34	4.34	0.02	0.00	0.00	±0.3
43600	4.39	4.34	4.33	4.39	4.35	4.34	0.00	0.02	0.02	±0.3
55400	4.38	4.34	4.33	4.38	4.34	4.33	0.00	0.00	0.00	±0.3
70000	4.37	4.32	4.31	4.37	4.33	4.32	0.00	0.02	0.02	±0.3
88800	4.35	4.31	4.30	4.34	4.32	4.31	-0.02	0.02	0.02	±0.3
112400	4.33	4.30	4.29	4.33	4.30	4.29	0.00	0.00	0.00	±0.3
142400	4.31	4.27	4.27	4.32	4.29	4.27	0.02	0.04	0.00	±0.3
161750	4.30	4.26	4.26	4.30	4.27	4.26	0.00	0.02	0.00	±0.3
180400	4.29	4.25	4.24	4.28	4.26	4.25	-0.02	0.02	0.02	±0.3
228400	4.26	4.22	4.21	4.25	4.22	4.21	-0.02	0.00	0.00	±0.3
289400	4.21	4.18	4.17	4.21	4.18	4.17	0.00	0.00	0.00	±0.3
366400	4.17	4.14	4.13	4.17	4.13	4.13	0.00	-0.02	0.00	±0.3
400000	4.15	4.12	4.11	4.15	4.12	4.11	0.00	0.00	0.00	±0.3
464000	4.12	4.09	4.08	4.12	4.09	4.08	0.00	0.00	0.00	±0.3
587800	4.07	4.04	4.04	4.07	4.05	4.04	0.00	0.02	0.00	±0.3
744200	4.02	4.00	4.00	4.02	4.01	4.00	0.00	0.02	0.00	±0.3
942600	4.00	3.98	3.98	4.00	3.98	3.98	0.00	0.00	0.00	±0.3
1193600	3.98	3.96	3.95	3.98	3.96	3.95	0.00	0.00	0.00	±0.3
1511600	3.97	3.96	3.95	3.98	3.96	3.95	0.02	0.00	0.00	±0.3
1914400	3.96	3.94	3.93	3.96	3.94	3.93	0.00	0.00	0.00	±0.3
2424400	3.93	3.91	3.90	3.92	3.91	3.90	-0.02	0.00	0.00	±0.3
3070200	3.89	3.87	3.87	3.89	3.87	3.86	0.00	0.00	-0.02	±0.3
3888000	3.83	3.82	3.81	3.84	3.82	3.81	0.02	0.00	0.00	±0.3
4000000	3.82	3.81	3.80	3.83	3.81	3.80	0.02	0.00	0.00	±0.3
4923800	3.75	3.73	3.72	3.75	3.74	3.72	0.00	0.02	0.00	±0.3
6235400	3.64	3.63	3.62	3.64	3.63	3.62	0.00	0.00	0.00	±0.3
7896400	3.50	3.49	3.48	3.51	3.50	3.48	0.02	0.02	0.00	±0.3
10000000	3.35	3.34	3.33	3.35	3.33	3.33	0.00	-0.03	0.00	±0.3

SPEAG H-field frequency response tolerance criteria¹:
±0.3dB for applied H-fields at calibration points from 3kHz to 10MHz

¹ Calibration uncertainty not taken into account (shared risk 50%).