



Date: 3/12/2024

Class Permissive Change request letter

Type of Equipment: IOT module

FCC ID: 2A76T00NRF9160

Model: 9000121

To whom it may concern:

We,

Abbott Medical CRM
15900 Valley View Court,
Sylmar, CA 91342, USA

Pursuant FCC §2.1043, hereby request the evaluation of a Class 2 Permissive Change as described below for the device mentioned above.

Integrate Module 9000121 in a Host (LSRM01) and evaluate RF-Exposure conditions. Therefore, the approach followed according to the KDB 996369 Module Certification.

A C2PC is performed on the module to change the RF Exposure conditions once the module is integrated into the host evaluated. Complete SAR testing have been performed on the Host.

Furthermore, no changes have been made to circuitry/transmitter. However, to make sure there is no degradation in performance into the Host, radiated spurious emission checks have been performed.

We request to change mobile to portable conditions.

In addition, we request to change from single modular to limited modular approval.

Sincerely,

By:

A handwritten signature in blue ink, appearing to read 'PerryLi', with a stylized flourish at the end.

PerryLi

Title: AssociateResearchFellow

Telephone: (818)493-3793

e-mail: perry.li@abbott.com