

Federal Communications Commission
7435 Oakland Mills Road
Columbia, Maryland 21046
USA

Date: September 8, 2022

Ref: FCC Class II Permissive change for FCC ID: 2A77600001

Dear Examiner,

This is to request a Class II permissive change for module approval
FCC ID: **2A77600001**, originally granted on **09/07/2022**.

Change#1: Additional Chassis is added, Product name: **Wireless/ Wired X-Ray Flat Panel Detectors**, Brand: **Allengers**
Model number: **G4343RWG, G4343RWC**.

Change #2: Addition a new antenna, the antenna type is different from the original application.

Change #3: Output power is reduced through firmware, and SAR was measured.
(Only reduce Wi-Fi 5G Output Power, Wi-Fi 2.4G Output Power haven't changes).

Change #4: Turn off BT and BLE function through firmware.

Change #5: Original product name **Intel® Dual Band Wireless-AC 8265** changed to **Wireless/ Wired X-Ray Flat Panel Detectors**

Sincerely,



By: _____
Sanjeev K. Marjara
Title: Additional Managing Director
On behalf of: Allengers Medical Systems Limited
e-mail: rnd@allengers.net

ALLENGERS MEDICAL SYSTEMS LTD.

