

	8	7	6	5	4	3	2	1
--	---	---	---	---	---	---	---	---

NOTES:
1. Use 3M adhesive 467MP or 9667MP.
2. Subsurface printing only.
3. Background is white.
4. Match colors of graphics with provided graphic file.
5. One label per lot is to be inspected.

Boarder Key
Kiss-cut tab
Label edge

All material must be RoHS compliant

REVISION HISTORY		
REV	DESCRIPTION	DATE
1	INITIAL RELEASE	10/27/2020
2	UPDATED REF NUMBER	2/5/2021
3	REMOVED DUPLICATE VOLTAGE INFO, ADDED IP RATING NUMBER, UPDATE TOLERANCES	4/7/2021
4	REMOVED CE MARK AND EMERGO EUROPE INFORMATION	6/29/2021
5	ADDED FCC MARK AND ID	--
6	ADDED CE MARK AND EMERGO EUROPE INFORMATION	--

Seno MEDICAL Imagio®
Seno Medical Instruments, Inc.
8023 Vantage Drive, Suite 1000
San Antonio, TX 78230 USA
www.senommedical.com
REF 9100
CISPR11 Group 1 Class A
Rx Only
Complies with 21 CFR 1040.10 and 1040.11 except for conformance with IEC 60825-1 Ed. 3 and IEC 60601-2-22 Ed. 3.1, as described in Laser Notice No. 56, dated May 8, 2019
EMERGO EUROPE Prinsessegracht 20 2514 AP The Hague The Netherlands
FCC ID: 2A7QE-9100
IP20
500 lb / 227 kg
CE 2460 +

ETL CLASSIFIED CONFORMS TO AAMI STD ES60601-1 IEC STD 60825-1 IEC STD 60601-2-22 IEC STD 60601-2-37 Intertek 5013988

! 208-240 VAC 50/60Hz 10A

UNLESS OTHERWISE SPECIFIED		
Drawn By V. Pacheco	DIMENSIONAL UNITS: INCHES TOLERANCES FRACTIONAL +/- 1/32 ANGULAR: MACH +/- ½ DEGREE	 TITLE Label, Imagio Nameplate DWG. NO. LBLG-00000115 REV. 6 SCALE: N/A DRAWN IN VISIO Sheet 1 of 1
MATERIAL 0.007" Thk Polycarbonate	ONE DECIMAL PLACE +/- 0.1 TWO DECIMAL PLACES +/- 0.06 THREE DECIMAL PLACES +/- 0.005	
FINISH N/A	PROPRIETARY AND CONFIDENTIAL	
Made From Part No. N/A	THE INFORMATION CONTAINED IN THIS DRAWING IS THE SOLE PROPERTY OF SENO MEDICAL INSTRUMENTS. ANY REPRODUCTION IN PART OR AS A WHOLE WITHOUT THE WRITTEN PERMISSION OF SENO MEDICAL INSTRUMENTS IS PROHIBITED.	
Previous number N/A		

	8	7	6	5	4	3	2	1
--	---	---	---	---	---	---	---	---