



Imagio® Breast Imaging System

Model 9100

Operator/User Manual

for use with probe models 8101 and 8001

I About Seno Medical Instruments™

Seno Medical Instruments™ is a medical device company focused on the detection of cancer using its patented Opto-acoustic (OA) technology. Opto-acoustic (OA) imaging combines light and sound to produce functional images.

Legal Manufacturer

Seno Medical Instruments, Inc.

8023 Vantage Dr. Ste 1000

San Antonio, TX 78230-4726

Tel: 210-615-6501

Seno Medical Service

Tel: 888-978-8835

Email: service@senomedical.com



Caution: Rx Only in the United States (Federal law restricts this device to sale by or on the order of a physician).

www.senomedical.com

<https://senomedical.com/patents/>

All materials contained in this document are protected by United States copyright law and may not be reproduced, distributed, transmitted, displayed, published, or broadcast without the prior written permission of Seno Medical Instruments, Inc., or in the case of third-party materials, the owner of that content. You may not alter or remove any trademark, copyright, or other notice from copies of the content.

Imagio® System, Imagio® Breast Imaging System, Imagio® OA/US Breast Imaging System, SenoGram® and Seno Medical® are registered trademarks of ©2022 Seno Medical Instruments, Inc.

©2022 Hologic Inc., All rights reserved. Hologic, SuperSonic, MACH, ShearWave and associated logos are trademarks and /or registered trademarks of Supersonic Imagine, Hologic, Inc., and /or its subsidiaries in the United States, Europe and other countries. This information is intended for medical professionals in the U.S. and other markets and is not intended as a product solicitation or promotion where such activities are prohibited.

Table of Contents

I About Seno Medical Instruments™	2
II Statements	7
1 System Overview	9
1.1 Documentation Access	9
1.2 Overview	9
1.3 Who Should Use This Manual?	9
1.4 Risk Management To Medical Devices	10
1.5 Safety, EMC, and Regulatory Compliance	10
2 Imagio® Breast Imaging System	19
2.1 Device Description	19
2.2 Indications for Use	19
2.3 Intended Use	19
2.4 Probe Intended Use and Indications	20
2.5 Components	22
2.6 Compatible Accessories and Kits	25
3 Scan Room and the Imagio® Breast	26
3.1 The Imagio® Breast Imaging System Heating	26
3.2 Positioning the Imagio® Breast Imaging System	26
3.3 Probe Storage/Transportation	26
3.4 Moving the Imagio® Breast Imaging System	27
3.5 Positions of Operator, Patient, and Individuals in the Scan Room	27
3.6 Confirm Laser Safety Items	27
3.7 Ensure Proper Environmental Operating Conditions	28
4 Probes	38
4.1 Probe/Clinical Applications Compatibility	38
4.2 Probe Handling	39
4.3 Probe Sheaths for Use With US L18-5 Linear Probe Only	40
4.4 Ultrasound Transmission Gel	41
4.5 Gel Holder/Warmer	41
5 Cleaning System Components	43
5.1 Cleaning and Disinfecting Agents	43
5.2 Probe Inspecting Cleaning and Disinfecting	46
6 System Power On Off	49
6.1 Circuit Breaker	49

6.2	Power On Procedures	49
6.3	To Power Off the Imagio® Breast Imaging System	49
7	System Console	50
7.1	Controls.....	50
7.2	Monitor and Articulated Arm	54
7.3	How to Move the Console	55
8	OA Mode	56
8.1	OA Mode Selection.....	56
8.2	Laser Heating	57
8.3	Laser Sleep	57
8.4	Laser Standby	57
8.5	To Place the Imagio® Breast Imaging System in Laser Standby	58
8.6	To Enable the Laser - Authorizing the Laser	59
8.7	Laser Adjustment Period	60
8.8	To Enter Laser Ready.....	60
8.9	Probe Contact Sensor.....	62
8.10	To Disable the Laser	63
8.11	OA Display	63
8.12	Description of OA Mode Imaging Parameters.....	65
8.13	SenoGram®	65
8.14	Region of Interest (ROI).....	66
9	Imaging Artifacts	68
9.1	OA Artifacts	68
9.2	Ultrasound Artifacts	72
10	System Controls.....	73
10.1	Common Imaging Controls.....	73
11	Ultrasound Imaging Modes	78
11.1	Common Imaging Parameters.....	78
11.2	B-mode Imaging	80
11.3	ShearWave™ Elastography.....	86
11.4	Color Mode Imaging.....	90
11.5	Pulsed Wave Doppler Mode.....	94
12	Operator Activities	99
12.1	Adding Distilled Water to the Laser Coolant Reservoir	99
12.2	Laser Energy Calibration.....	100
12.3	Machine Check.....	105
13	Managing Patient Data	107
13.1	Saving Data.....	107
13.2	Performing measurements	108

13.3	Using annotations	113
13.4	Using body markers.....	114
13.5	Reviewing an Exam.....	114
13.6	Deleting exams and images.....	117
13.7	Image Query and Retrieve.....	118
13.8	Ending an Exam	120
13.9	Continuing an Exam	120
13.10	BI-RADS Analysis.....	121
13.11	Description of the report feature	122
13.12	Report builder	122
13.13	Exporting reports	124
14	Laser Emitting Alerts, Laser State	125
14.1	Notifications Window.....	125
14.2	No Probe Contact Message	125
14.3	Laser Emitting Alerts	126
14.4	Laser State Messages	126
14.5	Status Codes.....	127
14.6	Information Message	128
14.7	System Messages	128
14.8	General System Messages.....	129
14.9	Main Display Notification Icons.....	130
15	Connecting and Disconnecting.....	133
15.1	Removing and Connecting Probes.....	133
15.2	Cabinet Storage Areas	134
15.3	Input/output Connectors	135
15.4	Additional Video Devices.....	135
16	System Configuration.....	137
16.1	OA Settings.....	138
16.2	System/Display Tab	140
16.3	Device Settings.....	144
16.4	Administration	148
16.5	Presets	155
16.6	Measurements	161
16.7	Online Services	164
16.8	System Diagnostics.....	164
16.9	Remote Support.....	165
A	Troubleshooting.....	167
A.1	Troubleshooting.....	167
B	Laser Classification and Warning	169
B.1	Laser Classification	169
B.2	Electromagnetic Compatibility Compliance.....	176

C Product Specifications	179
D Probe Ranges, Accuracies, and	182
E Acoustic Output Tables	186
E.1 Maximal Temperature Data	186
E.2 OA/US L18-1 & US L18-5 Linear Probe B-MODE.....	186
E.3 US L18-5 Linear Probe Color Doppler	187
E.4 US L18-5 Linear Probe Shear Wave Elastography.....	188
E.5 US L18-5 Linear Probe Pulsed Wave Doppler	189
E.6 OA/US L18-1 B-MODE Elastography	190
E.7 OA/US L18-1 B-MODE Color Doppler	192
E.8 OA/US L18-1 B-MODE Pulsed Wave Doppler	194
F Risk Benefit Analysis	195
G Guidelines for OA US Scanning.....	196
G.1 OA/US INSTRUMENTATION AND SCAN TECHNIQUE	196
G.2 OA/US Breast Imaging	197
G.3 Axillary Lymph Node OA/US Scanning.....	200
H Tissue Phantom Study	201
I Clinical Study Summary	205

II Statements

Service Statement

Service and/or troubleshooting to be performed only by Seno Medical personnel and/or their designated third-party service companies referred to as the local service provider throughout the manual.

Seno Medical ServiceTel: 888-978-8835

Email: service@senomedical.com

Training Statement

Seno Medical provides a training program accessible online at Seno University <https://senouniversity.com/> for new operators. Operators can arrange to participate by notifying Seno Medical. In addition, Seno Medical provides on-site Imagio® Breast Imaging System training by a Clinical Applications Specialist.

Warranty Statement

Subject to the terms below, Seno Medical Instruments, Inc. ("Seno") provides the following limited warranty for its products. The Imagio® Breast Imaging System is a medical diagnostic device that should only be operated by trained operators.

The warranty period commences on the date of shipment. Seno's sole obligation for this limited warranty is to repair or replace a defective product at no charge to Customer, or to credit Customer's account for the purchase price paid for the defective product, at Seno's discretion.

Under no circumstances should anyone other than authorized Seno Service personnel make any modifications to the Imagio® Breast Imaging System. Each of the system's safety circuits has been designed to minimize potential adverse situations. These safety circuits must never be bypassed, altered, or disabled. The manufacturer and distributor assume no responsibility or liability for personal injury or property damage resulting from misuse of the systems.

Seno warrants the equipment to be free from defects in workmanship and materials for the contracted warranty period. Seno has robustly tested and can reliably promote a functional life of at least 1 year for the system. Seno recommends a six (6) month preventive maintenance schedule to maintain optimal performance of the Imagio® Breast Imaging System. The limited warranty does not apply if the defective product (i) has been damaged due to abuse, misuse, neglect, accident, unusual physical or electrical stress, or tampering, (ii) has not been used in accordance with Seno's written instructions for use (IFU), (iii) was not purchased from Seno or an authorized dealer of Seno, or (iv) was modified from its original configuration or repaired or altered by anyone other than Seno or a person authorized by Seno.

To make a warranty claim, Seno's Customer Care Department must be contacted within five days of discovery of the defect to obtain a return authorization number. Seno will be responsible for shipping costs on defective products that are under warranty which are returned by customer to Seno with a return authorization number. Replaced or repaired product will be shipped to customer's site and installed by Seno or an authorized representative of Seno at Seno's expense.

Seno aims for the use of the Imagio® Breast Imaging System to be fully operational and trouble-free throughout the useful life of the system. However, periodic repairs may be needed to be performed by an authorized Seno Service provider. All parts that are replaced become the property of Seno. The operator is responsible for backing up and maintaining any data stored on the Imagio® system.

Due to the complex nature of the system and the importance of safety to patients and operators, all service work must be performed by an authorized Seno Service representative. Call Seno Service for assistance at (888) 978-8835.

Safety Descriptors

This manual uses the following safety descriptors:



Warning: Warning statements inform the operator of safety alerts that must be followed carefully to avoid bodily injury.



Caution: Caution statements inform the operator of safety alerts that must be observed to avoid damage to the equipment.

1 System Overview

1.1 Documentation Access

To access the latest version of the Imagio® Breast Imaging System Operator Manual, please go to: **Error! Hyperlink reference not valid.** senomedical.com/manuals/current

1.2 Overview

The purpose of this manual is to describe the Ultrasound and Opto-acoustic (OA) functionality of the Imagio® Breast Imaging System and to provide usage instructions.

This manual contains:

- Intended Use
- Online Access to Imagio® Breast Imaging System Operator Manual.
- Device description
- Important safety information
- An introduction to the Imagio® Breast Imaging System
- Detailed instructions for operating the Imagio® Breast Imaging System
- How to save scanned images and upload them to a facility's Picture Archiving and Communication System (PACS)
- How to perform laser energy calibration maintenance and cleaning

1.3 Who Should Use This Manual?

This operator manual is a reference for operators of the Imagio® Breast Imaging System. It is designed for a reader familiar with ultrasound imaging techniques; it does not provide training in sonography or clinical procedure or practices. This manual is written for the individual who has attended and received Imagio® Breast Imaging System training from Seno personnel or their agents. This manual is intended to assist with the safe and effective operation of the Imagio® Breast Imaging System.

Formal System training from Seno personnel or their agents combined with the information in this manual provides sufficient information to operate the Imagio® Breast Imaging System safely. Read and understand all instructions in this manual before attempting to use the Imagio® Breast Imaging System and strictly observe all warnings and cautions.

1.3.1 Permissible/Impermissible Impairments for operators of the system

Permissible impairments include:

- Latex allergy

Impermissible impairments include:

- Vision impaired persons without corrections.

- Persons unable to firmly and steadily grip an OA/US L18-1 Probe.
- Persons unable or unwilling to wear laser safety protective eye wear.
- Color blindness.
- Persons with difficulty standing.

1.4 Risk Management To Medical Devices

Seno follows the requirements for risk management described in ISO 14971:2019 Medical devices –Application of risk management to medical devices.

1.5 Safety, EMC, and Regulatory Compliance

1.5.1 Laser Safety

For the Imagio® Breast Imaging System, Seno Medical recommends the facility should check and abide by any country/state regulations as pertaining to our laser classifications noted below.

1.5.2 Laser Classification

The Imagio® Breast Imaging System uses an Nd:YAG 1064 nm 5-25 ns pulse laser and an Alexandrite 757 nm 50-100 ns pulse laser. Both lasers operate at 85 mJ +/-20% that correspond to a max energy out of the probe of 102 mJ (0.51W) per laser (wavelength). Nominal operational energy is at 85 mJ, and the maximum energy is 102mJ, corresponding to 20% over the operational energy output.

The Class 3B laser beams emitted by the Imagio® Breast Imaging System can cause loss of sight.

The lasers operate at two different wavelengths. Any energy transmitted by the Imagio® Breast Imaging System that enters the eye may be focused directly on the retina. Direct absorption of laser energy by the retina may result in temporary clouded vision, retina lesion, long term scotoma and long term photophobia.

The potential for loss of sight exists in any case of:

- Direct laser radiation
- Reflected laser radiation
- Diffused laser radiation

1.5.3 Essential Performance

List of ESSENTIAL PERFORMANCE functions

- 1) The system shall display co-registered structural and functional images typically used to assist in diagnosing breast abnormalities.
- 2) The system shall produce images that enable the operator to assess vessels in the internal zone of a breast anomaly.
- 3) The system shall produce images that enable the operator to assess blush in the internal zone of a breast anomaly.
- 4) The system shall produce images that enable the operator to assess total hemoglobin in the internal zone of a breast anomaly.
- 5) The system shall produce images that enable the operator to assess vessels in the external peripheral zone of a breast anomaly.

- 6) The system shall produce images that enable the operator to assess vessels in the external boundary zone of a breast anomaly.
- 7) The system shall allow the operator to adjust the following imaging parameters for differences in anatomy: gain, TGC, focus zones, fundamental frequency, tissue harmonic frequency, gray scale, dynamic range, persistence, and depth.
- 8) The system shall have an auto-optimization capability for gain and TGC for US modes with and without OA.
- 9) The system shall provide the ability to assign OA/US and US imaging feature scores and provide a suggested POM and BI-RADS category via separately packaged software (e.g. Online form).
- 10) The system shall be Free from noise on a waveform or artifact or distortion in an image or error of displayed numerical value which cannot be attributed to a physiological effect, and which may alter the diagnosis.
- 11) The system shall be Free from the display of incorrect numerical values associated with the diagnosis to be performed [see **NOTE 1**].
- 12) The system shall be Free from the display of incorrect safety-related indications [see **NOTE 1**].
- 13) The system shall be Free from the production of unintended or excessive ultrasound output.
- 14) The system shall be Free from the production of unintended or excessive TRANSDUCER ASSEMBLY surface temperature.

Supplementary Information:

ESSENTIAL PERFORMANCE is performance, the absence or degradation of which, would result in an unacceptable risk.

NOTE 1: “incorrect” in the sense that the displayed value differs from what is calculated (having been altered during data transfer), or the calculation itself is incorrect.

1.5.4 Output Specifications

Specified Parameter	Nd:YAG Flashlamp Pumped, QS SSL	Alexandrite Flashlamp Pumped, QS SSL
Wavelength (nm) (1)	1064	757
Beam Shape (cm ²) (2)	Rectangular 5.5 x 1.25	Rectangular 5.5 x 1.25
Laser Mode	Pulsed	Pulsed
Repetition Rate (Hz)	5	5
Pulse Energy (mJ) (3)	102	102
Average Power (W)	0.51	0.51
Pulse Width (ns)	7	80

1. (+/-10nm)
2. Contiguous output windows: Total Area 4.3cm x 1.6cm.
3. Maximum Energy Probe Output (85mJ+/-20%): 102mJ
4. Skin Exposure at the output of the Probe with Uniform Energy Distribution

Note: the actual area of the probe face is 4.3 cm x 3.3 cm

Ocular Exposure		
Specified Parameter	YAG Laser (1064nm)	Alexandrite Laser (757nm)
Radiant Exposure @ 10cm (mW/cm ²)	100	3.59
MPE Repetitive-Pulse (μJ/cm ²)	6.6	0.83
Energy/Pulse Thru 7mm Aperture(J)	1.38 x 10 ⁻³	1.38 x 10 ⁻³
Class 3B AEL (J)	150 x 10 ⁻³	150 x 10 ⁻³
Laser Class	3B	3B
Viewing Condition	Unaided Viewing	Unaided Viewing
Necessary OD @ 10cm Viewer Distance	2.73	3.64
Nominal Ocular Hazard Distance (m)	2.66	7.48
Skin Exposure		
Specified Parameter	YAG Laser (1064nm)	Alexandrite Laser (757nm)
MPE Skin (mJ/cm ²)	100	25.2
Seno Level Radiant Exposure @ 0cm (mJ/cm ²) (4)	14.8	14.8
Necessary OD	None	None

1.5.5 Least Favorable Working Conditions

The “least favorable working condition” (i.e. the most hazardous scenario) has been determined to be when the Imagio® Breast Imaging System laser is emitting.

1.5.6 Laser Cooling System

The laser cooling system is filled with distilled water. The laser cooling system must contain a sufficient amount of distilled water at all times to ensure proper operation.

The Add Coolant message will appear on the lower right of the Display Monitor if the amount of distilled water in the laser coolant reservoir is insufficient.

1.5.7 Device Information Standards:

The Imagio® Breast Imaging System is designed to meet the following standards:

IEC/EN/ANSI AAMI ES 60601-1	Medical electrical equipment - Part 1: General requirements for basic safety and essential performance
IEC/EN 60601-1-2	Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic compatibility - Requirements and tests

IEC/EN 60601-1-6	Medical electrical equipment – Part 1-6: General Requirements for Basic Safety and Essential Performance – Collateral Standard: Usability
IEC/EN 60601-2-22	Medical electrical equipment - Part 2-22: Particular requirements for basic safety and essential performance of surgical, therapeutic and diagnostic laser equipment
IEC/EN 60601-2-37	Medical electrical equipment - Part 2-37: Particular requirements for the basic safety and essential performance of ultrasonic medical diagnostic and monitoring equipment
IEC/EN 60825-1	Safety of laser products - Part 1: Equipment classification and requirements
EN ISO 10993-1	Biological evaluation of medical devices -- Part 1: Evaluation and testing within a risk management process

1.5.8 Device Information Classifications:

The Imagio® Breast Imaging System is designed to the following classifications: IEC/EN 60601-1-2, Medical Equipment, Electromagnetic Compatibility Classification:

- Group 1 Class A equipment (CISPR11)
- Suitable for use in all establishments other than domestic and those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes

IEC/EN 60601-1, Medical Electrical Equipment: General Requirements for Basic Safety and Essential Performance Classification:

- Class I Equipment
- Type BF Applied Part: OA/US L18-1 Probe IPX1
- Type BF Applied Part: US L18-5 Linear Probe IPX1
- Electrical Ratings: 208-240V, 50/60Hz, 10A
- OA duty cycle: 5 minutes ON / 10 minutes OFF
- Laser Output energy accuracy: +/- 20%
- Ingress Protection: The Imagio® Breast Imaging System IPX0
- Ingress Protection: Foot Switch: IPX0

Isolation of the Imagio® Breast Imaging System from supply mains is accomplished internally by means of isolation transformers.

Note: The Imagio® Breast Imaging System is a medical device used for diagnostic purposes only and is not intended for life critical applications. Therefore the Imagio® Breast Imaging System was not designed to respond to interruptions to the main supply voltage of greater than 500 milliseconds in duration.

1.5.9 Environmental Temperature for safe operation and storage of the Imagio® Breast Imaging System

- Operating Temperature Range 10 °C to 30 °C (50 °F to 86 °F)

- Storage/Transport Temperature Range 4 °C to 50 °C (39 °F to 122 °F)
- Operating Humidity Range 30% to 70%, non-condensing
- Storage/Transport Humidity Range 30% to 80%, non-condensing

1.5.10 Receiving and Storing the Imagio® Breast Imaging System

The Imagio® Breast Imaging System must be installed only by Seno authorized personnel.

Upon delivery and initial installation, please allow Seno authorized personnel to open the system packaging. If the Imagio® Breast Imaging System needs to be returned back to Seno for service, please do not pack it. Call Seno Service for assistance at (888) 978-8835.

1.5.11 Storage of the Device

The Imagio® Breast Imaging System may be stored indefinitely if it is done so under ambient conditions which do not exceed the storage environmental limits.

It is also prudent to have an authorized Seno Service representative re-start the system for the first time after prolonged storage.

1.5.12 Regulatory Requirements

Regulatory requirements for imaging facilities vary by country/province/state.

If your region has laser safety requirements, the facility's Laser Safety Officer is responsible for verifying and complying with such regulatory requirements concerning the use of lasers in the facility.

1.5.13 Emissions and Immunity Information

Please refer to "Declaration of Electromagnetic Emissions" on page 174 for the Imagio® Breast Imaging System emissions and immunity tables.

1.5.14 Cybersecurity



Caution: When connected to a network, the Imagio® Breast Imaging System is potentially able to be compromised by unauthorized access to data, extraction of data, loss of data and corruption of data.

- Application parameters can only be configured by a trained and authorized operator.
- Do not operate the Imagio® Breast Imaging System without appropriate network security measures. Organizations that elect to configure/use the networking functionality provided by Seno are assuming all liabilities and risks associated with that decision.
- We recommend operating the Imagio® Breast Imaging System with a firewall. Seno designed the Imagio® Breast Imaging System to be compatible with PACS systems.
- Do not install software or hardware on the Imagio® Breast Imaging System.
- Always call Seno Service for assistance at (888) 978-8835 in the event of a cybersecurity compromise on the Imagio® Breast Imaging System.
- Always maintain configuration and data backups for the Imagio® Breast Imaging System.

1.5.15 Security Management

All unnecessary communication ports and all remote logging services are disabled. Only the DICOM port remains opened (TCP/IP Network Port 11112 by default).

The Imagio® Breast Imaging System is designed according to the principle of least functionality. Only Administrator-configured servers can push objects to the Imagio® Breast Imaging System.

1. All operating System services that are not used by the Imagio® Breast Imaging System application software are disabled to minimize the sources of security vulnerabilities. The Imagio® Breast Imaging System provides only essential capabilities. The use of non-essential functions, ports, protocols, and services is prohibited or restricted to prevent unauthorized connection of devices, unauthorized transfer of information, or unauthorized tunneling.
2. Healthcare delivery organizations can utilize network scanning tools, intrusion detection and prevention systems to identify and prevent the use of prohibited functions, ports, protocols, and services.

The operator interface does not allow operators to interact with the operating system. There is no access for an operator to the underlying operating system. The operator has only access to the ultrasound interface controls software.

The “auto run” feature is disabled on the Imagio® Breast Imaging System. The Imagio® Breast Imaging System will not open or run a program available on USB memory stick inserted into the Imagio® Breast Imaging System. Seno Medical is performing monitoring of potential cybersecurity vulnerabilities. Should a security problem be discovered; the customer will be notified, and a cybersecurity patch will be provided. For security problems discovered by the facility, please contact cybersecurity@senomedical.com.

In order to minimize the consequences of any security event, Seno Medical strongly recommends routinely transferring sending images to a PACS.

To ensure patient and operator data protection, it is necessary to use secured tools for any data sharing containing Protected Health Information (PHI). E-mails and non-secured data hosting services do not guarantee an effective protection of data.

Recommendations regarding use of USB devices

To mitigate the threat of a potentially infected thumb drive plugged the Imagio® Breast Imaging System, the following measures should be taken at a minimum:

- Ensure the use of approved, trusted vendors for hardware purchases.
- Scan all hardware, especially removable storage media, on an external computer prior to its insertion into Imagio® Breast Imaging System.
- Operators should protect themselves and organizations by practicing good browsing habits, ensuring they do not respond to or select unsolicited email, and to not plug unknown USB devices into their work stations.
- If operators don't have the expertise to properly handle or identify potential cyber threats, please seek out an expert who can provide the expertise needed to secure your organization.

1.5.16 Anonymization

It is possible to anonymize images before export, however anonymization of image pixel data is ultimately the responsibility of the site. even if applications are used to anonymize the data, Anonymization of data burned into the image itself (i.e., image pixel data) is notoriously difficult. Review of all such images for accurate anonymization is strongly recommended when such images are batch anonymized by computer application.

1.5.17 Symbols and Markings Glossary

Symbol / Marking	Definition
	Manufacturer
	Date of Manufacture
	Authorized Representative in the European Community
	Catalogue number
	Serial number
	General warning sign
	General caution sign
	Warning; Laser beam
	Refer to instruction manual/booklet
	Wear eye protection
	Type BF applied part
	Optical applicator
	Foot switch

Symbol / Marking	Definition
	Remote interlock connector
	Filling
	Draining; emptying
	Valve; shut-off element
	Mass; weight
IPn1n2	Ingress Protection Mark
	WEEE Mark This symbol indicates that waste electrical and electronic equipment must not be disposed of as unsorted municipal waste and must be collected separately. Please contact an authorized representative of the manufacturer for information concerning the decommissioning of the equipment. <i>Note: Only relevant for EU</i>
Rx Only	Rx Only Mark
CE	CE Conformity Mark
	ETL Mark
IO	On the power switch, represents Imagio® Breast Imaging System power ON and OFF
	On the console represents Imagio® Breast Imaging System ON/Standby
	USB input/output port
	Ethernet connection

Symbol / Marking	Definition
	Equipotentiality ground
	DVI out
	Humidity limitation Indicates the range of humidity to which the medical device can be safely exposed.
	Atmospheric pressure limitation Indicates the range of atmospheric pressure to which the medical device can be safely exposed.
	Temperature limitation Indicates the range of temperature to which the medical device can be safely exposed.
	Probe connector Indicates the state of the probe connector lock.
	FCC identification number unique to Seno Imagio Breast Imaging System, Model 9100 is 2A7QE-9100 ¹ . This device complies with part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) This device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

¹ NOTE: This equipment has been tested and found to comply with the limits for a Class A digital device, pursuant to part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference when the equipment is operated in a commercial environment. This equipment generates, uses, and can radiate radio frequency energy and, if not installed and used in accordance with the instruction manual, may cause harmful interference to radio communications. Operation of this equipment in a residential area is likely to cause harmful interference in which case the user will be required to correct the interference at his own expense.

2 Imagio® Breast Imaging System

2.1 Device Description

The Imagio® Breast Imaging System is a diagnostic functional and morphologic imaging system that combines Opto-acoustic (OA) imaging with conventional diagnostic ultrasound (US) imaging to aid in diagnostic evaluation of breast lesions. The Imagio® Breast Imaging System acquires, processes, and displays co-registered OA and US images to provide structural and functional information about breast abnormalities in real-time. OA/US imaging technology combines the high contrast resolution of optical imaging with the high spatial resolution and deep penetration of ultrasound imaging.

In OA imaging (also known as Photo-acoustic imaging in the literature), short nanosecond pulses of laser light are delivered into tissue. When that light is absorbed, for instance by hemoglobin molecules found in blood, those molecules undergo thermoelastic expansion and generate acoustic pressure signals that can be detected at the tissue surface using a hand-held probe. The probe has optical windows that deliver light to tissue and an array of piezoelectric transducers that detect acoustic signals. The received OA signal increases with both the absorption strength and concentration of light-absorbing molecules. Two strong absorbers in tissue are oxygenated and de-oxygenated hemoglobin found in blood vessels. By performing OA imaging using two near-infrared optical wavelengths (757 nm and 1064 nm) for which oxygenated and de-oxygenated hemoglobin absorb light differently, the Imagio® Breast Imaging System can estimate the relative oxygen saturation of hemoglobin in tissue.

2.2 Indications for Use

The Imagio® Breast Imaging System is indicated for use by a trained and qualified healthcare provider for evaluation of palpable and non-palpable breast abnormalities in adult patients who are referred for a diagnostic imaging breast work-up, following clinical presentation or either screening or diagnostic mammography or other imaging examinations. The ultrasound mode should be initially used in a targeted fashion, to assess any focal area(s) of clinical or imaging concerns. In ultrasound mode, the device can be used to assign a BI-RADS category to either breast tissue or a mass that is causing clinical or imaging concerns. Masses that are classified as BI-RADS categories 3 through 5 can then be assessed using the Opto-Acoustic (OA) mode. In the OA mode, the Imagio® Breast Imaging System provides information about the central nidus, boundary, and peripheral zones, based on vascularity and blood oxygen saturation of the imaged tissues, to assist in the diagnosis of the benign or malignant mass(es) of interest. For ultrasound BI-RADS 3-5 masses, using the OA features of the mass allows for improved classification of the mass of interest as compared to ultrasound alone. The OA mode is not indicated for ultrasound BI-RADS 1 and 2 findings. The Imagio® Breast Imaging System includes an artificial intelligence (AI) based software function to assist the operators in assessing BI-RADS classifications.

This device is not intended to be used as a replacement for mammographic screening or for definitive pathologic diagnosis.

2.3 Intended Use

The Imagio® Breast Imaging System is intended to be used in the evaluation of breast abnormalities, regardless of the method or mechanism of initial detection.

The Imagio® Breast Imaging System is a diagnostic functional and morphologic imaging system intended for the opto-acoustic (OA) evaluation of breasts in women who have findings of suspicious masses (including both palpable and non-palpable masses) or imaging findings (such as architectural distortion, asymmetry, or suspicious calcifications).

The Imagio® Breast Imaging System utilizes real-time opto-acoustic and ultrasound technologies to acquire, process, display, and log co-registered structural and functional information about breast

abnormalities. This information is presented to the operator as fused and temporally interleaved opto-acoustic and ultrasonic images. The Imagio® Breast Imaging System also includes real-time diagnostic ultrasound with color and pulsed Doppler. The grayscale ultrasound includes image enhancing features, and the ability to measure and annotate images. The system is DICOM compliant and has PACS connectivity.

The Imagio® Breast Imaging System is intended to support qualified healthcare professionals in the differentiation of benign from malignant breast lesions and to support them in their decisions.

The Imagio® Breast Imaging System is not intended to be used as a screening device or for definitive pathologic diagnosis.

2.4 Probe Intended Use and Indications

System	Imagio® Breast Imaging System							
Probe	OA/US L18-1 Probe							
Intended Use	Real-time conventional diagnostic ultrasound with color and pulsed wave Doppler for fluid flow analysis							
Indications as follows:								
Clinical Application	Modes of Operation							
Specific (Track 3)	B	SW	PWD	CWD	Color/Power Doppler (CD/PD)	Combined	OA	
Small Organ: Breast	Yes	Yes	Yes	No	Yes	Yes	Yes	
Probe	US L18-5 Linear Probe							
Intended Use	Real-time conventional diagnostic ultrasound with color and pulsed wave Doppler for fluid flow analysis							
Indications as follows:								
Clinical Application	Modes of Operation							
Specific (Track 3)	B	SW	PWD	CWD	Color/Power Doppler (CD/PD)	Combined	OA	
Small Organ: Breast	Yes	Yes	Yes	No	Yes	Yes	No	

This device has not been evaluated for use on an individual who:

- Has a condition or impediment which could interfere with the intended field of view.
- Has or has had cancer in the ipsilateral breast within the same quadrant(s) as the mass(es) to be biopsied.
- Has had prior benign excisional breast biopsy within the immediate vicinity of the currently evaluated mass within the past 18 months.
- Has greater than three lesions.
- Has lesion(s) of interest greater than 4 cm.
- Currently has mastitis, is known to be pregnant or lactating.
- Has focal pain without thickening or mass.
- Has open sores including insect bites, rash, poison ivy, and/or chafing on the skin of the ipsilateral breast.
- Has an acute or a chronic hematoma and/or acute ecchymosis of the ipsilateral breast.
- Is experiencing photo-toxicity or photo-sensitivity or is undergoing treatment or taking medication for a photosensitive condition such as porphyria or lupus erythematosus.
- Has concurrent neoadjuvant therapy at the time of the Imagio® Breast Imaging System evaluation or the biopsy.
- Has previously had image guided core biopsy, image guided DVAB, or surgical biopsy of the mass of interest.
- Has a nipple ring and is unwilling to remove for the exam when the mass is <4cm from the nipple.
- Has applied fragrance or lotion on the skin where the probe will be applied.

The benefits and risks of use of the system are always to be weighed by the treating physician.

2.4.1 Contraindications

Absolute contraindications are as follows:

- Is pregnant.
- Has open sores including insect bites, rash, poison ivy, and/or chafing on the skin of the ipsilateral breast.
- Is experiencing photo-toxicity associated with currently taking, or having taken, photosensitizing agents within the previous 72 hours such as sulfonamides, ampicillin, tetracycline.
- Is currently undergoing phototherapy.
- Has a history of any photosensitive disease (e.g., porphyria, lupus erythematosus) or is undergoing treatment for a photosensitive disease and is experiencing photosensitivity.

2.5 Components

Opto-acoustic is an imaging modality that uses pulses of laser light to produce imaging data in a medical environment. The Imagio® Breast Imaging System utilizes this new modality to create images of tissue and then present those images on a display as well as record them for future reference. The Imagio® Breast Imaging System provides operators additional information about not only the presence of lesions within the tissue of interest but also whether those lesions may be malignant or benign.

The displays of the Opto-acoustic images include the addition of colors to the images to show where the laser light energy has been absorbed by structures within the tissue of interest. The relative amounts of absorption of the laser light by the tissue components can be related to the amount of blood content in and around a lesion and can even be related to the oxygenation level of that blood. This information can be used to assist in determining the disease state of such lesions.

The Imagio® Breast Imaging System uses specialty components designed to perform Opto-acoustic (OA) imaging exams.

The Imagio® Breast Imaging System



2.5.1 Cabinet

The Imagio® Breast Imaging System cabinet contains computers, the laser, an Ultrasound module, and power supplies.

2.5.2 System Console

The System Console is the Operator interface for entering patient data, exam settings, and for performing the exam. The Console, mounted on an adjustable arm, includes physical buttons and dials, a Touch Screen, TouchPad, probe holders, gel holder/warmer, and a cleaner holder bracket for cleaning and disinfectant agent

containers.

2.5.3 Display Monitor

The Display Monitor displays the Opto-acoustic images detected by the Imagio® Breast Imaging System probes during the exam. The Display Monitor is mounted on an adjustable arm.

2.5.4 Front Probe Cable Storage Tray

A front storage tray is available to hold excess probe cable length off the floor when, for example, not using a probe. This reduces excess probe cable length of a probe which is in use or keeping stored probe cables off the floor when moving the Imagio® Breast Imaging System.

2.5.5 Flat Breast Phantom

The Phantom Box contains an Opto-acoustic Flat Breast Phantom for use during, for example, in assessment of image quality or when performing the Machine Check optional procedure.

Contact Service for questions about image quality evaluation with the phantom. Seno Service will determine whether suboptimal performance or functional life is related to phantom degradation and a new phantom is necessary, or if the suboptimal performance is due to the Imagio® Breast Imaging System.

2.5.6 Foot Switch

The laser foot switch controls the emission of the Imagio® Breast Imaging System laser. OA images will be displayed if the laser is enabled, and the foot switch pressed when the probe is in contact with a patient or phantom.

2.5.7 Power Cord

A 4.5-meter (15 foot) power cord is supplied with the Imagio® Breast Imaging System.

2.5.8 Ethernet Cable

A 4.5-meter (15 foot) Ethernet cable is provided for connection to a site-provided RJ45 network connection jack to enable Internet access for uploading the Imagio® Breast Imaging System images to the site's Picture Archiving and Communication System (PACS).

2.5.9 Laser Protective Eyewear



Warning: *Invisible laser light is emitted from the OA/US L18-1 Probe. Reflected and/or misdirected laser light may damage eyes.*



- Do not use any controls or adjustments or performance settings other than those specified herein.
- The Patient, the Imagio® Breast Imaging System Operator, and everyone else in the scanroom must always wear Seno approved laser protective eyewear when the Imagio® Breast Imaging System is powered on and the laser is enabled.
- Wearing Seno approved laser protective eyewear is a laser safety requirement. It is necessary to wear laser protective eyewear when, for example; scanning, inspecting, and cleaning the OA/US L18-1 Probe, and calibrating laser energies or any other time the Imagio® Breast Imaging System is powered on and the laser is enabled.
- Do not look directly at the laser beam even when laser protective eyewear is being worn.
- Do not look at a laser beam emitted from the Imagio® Breast Imaging System that is visible. The Imagio® Breast Imaging System uses lasers that are classified as invisible; however, some individuals can see light beyond the visible spectrum.
- Always point the OA/US L18-1 Probe away from everyone's eyes.
- Only use your foot to press the foot switch.
- Always release the foot switch before removing the OA/US L18-1 Probe from the skin or calibration port.
- Always follow the laser safety precautions prior to laser emitting to protect the Patient, Operator, and Observers.
- Always place the OA/US L18-1 Probe in the Probe holder when not in use.
- The Seno laser protective eyewear has been specifically selected for two main reasons:
 - The laser protective eyewear is designed to protect the Operator, Patient, and Observers from the specific wavelengths of the Imagio® Breast Imaging System laser
 - The laser protective eyewear is designed to prevent interference with reading images

There are specific styles of eyewear only for the operator and observer. There is a specific style of eyewear only for the patient. Optical density, or OD, is a logarithmic function that corresponds to the amount of light that a lens transmits at a specific wavelength. OD is specified for specific wavelengths or ranges of wavelengths. Laser protective eyewear of OD4 or greater provides protection for the 1064 nm wavelength and OD5 or greater for the 757 nm wavelength.

For best comfort when wearing goggle style frames, adjust the strap so there is just enough tension to create a seal that prevents observation of visible light around the entire goggle seal.

After the laser has been authorized, a message appears on the display to remind the operators and everyone in the scan room wear laser protective eyewear.

- Seno has designated the exact laser protective eyewear required and this eyewear is the only Seno approved laser protective eyewear to be used with the Imagio® Breast Imaging System. They must be available for installation, service, training, and patient scans.

2.5.9.1 Laser protective eyewear for the Operator and Observer

The three styles of laser protective eyewear are only for the Operator and Observers. There are three different frame styles to accommodate different facial dimensions.

Inspect the eyewear prior to use or if they are dropped. Do not use the eyewear if damage to the eyewear is observed. If damage is discovered during a visual inspection, discontinue use and call Seno Service for assistance at (888) 978-8835. It is recommended operators always wear eyewear with clean lenses.

2.5.9.1.0.1 Modern Frame

For Operator and Observer only.



2.5.9.2 Laser protective eyewear for the Patient, Operator, or Observer

Laser protective eyewear for the patient provides protection against stray laser light during an Imagio® Breast Imaging System scan and must be worn by each patient. The direct wear goggle style laser protective eyewear is the only style permitted to be worn by the Patient.

2.5.9.1.0.2 Goggle-Style Frame

Important: Patient laser protective eyewear can also be worn by the Operator or Observer.



2.6 Compatible Accessories and Kits

Under some use conditions, additional materials provided by third-party vendors may be used by clinicians performing exams with the system. The decision to use the system and probes with third-party products is completely at the discretion of the clinician. Use of 3rd party product outside of that specified by Seno Medical Instruments, Inc is considered off-label and is done at risk to the Imagio® Breast Imaging System, and potentially, the patient.

3 Scan Room and the Imagio® Breast Imaging System Preparation

3.1 The Imagio® Breast Imaging System Heating

It is recommended the operator prepare the scan room and Imagio® Breast Imaging System before the first scan of the day because once powered on, the Imagio® Breast Imaging System lasers may take up to 15 minutes to reach operating temperature.

3.2 Positioning the Imagio® Breast Imaging System

A wheel locking mechanism is incorporated into the front of the Imagio® Breast Imaging System frame to secure the Imagio® Breast Imaging System once positioned in the scan room. The foot activated wheel lock locks/unlocks the front wheels (swivel casters) allowing the Imagio® Breast Imaging System to be swiveled but not moved.



Caution: The Imagio® Breast Imaging System requires uninterrupted airflow through the rear cooling intake. Any obstruction to airflow may cause overheating and Imagio® Breast Imaging System power off.

- Always position the Imagio® Breast Imaging System with a minimal distance of 31 cm (12") between the Imagio® Breast Imaging System and the wall.

3.2.1 Rear Cabinet Handle

A large rear cabinet handle allows firm and safe grasp of the Imagio® Breast Imaging System cabinet when steering/moving in the scan room. The front wheels swivel. The rear wheels are fixed and do not swivel to steer. To steer the cabinet right or left, push the rear handle forward on the opposite side (e.g. push the rear handle forward on the right to move cabinet to the left).

Important: Pushing the system is much easier and safer than pulling it. Grasp the handles firmly and maintain an upright posture when pushing. When moving the system over long distances, make sure to use the rear handle to push the system to avoid damaging it.

3.2.2 Brake and Wheel Locks

Wheel locks are provided to stabilize the system when stationary, such as when scanning or storing.

A wheel lock foot pedal is located at the front bottom middle and controls the wheel lock on or off.

3.3 Probe Storage/Transportation

A box is always provided for the probes as part of the entire Imagio® Breast Imaging System. Retain this box for storage, carrying, and transport/shipping. Failure to properly secure the probe during storage, transport and shipping may result in damage that would be the responsibility of the facility.

3.4 Moving the Imagio® Breast Imaging System

3.4.1 Preparing to Move the Imagio® Breast Imaging System

- Step: 1** Put the console back in the centered and lowered position
- Step: 2** Turn the system power off.
- Step: 3** Unplug the main power cord, and disconnect any accessory cords (network, etc.).
- Step: 4** Use the probe cable management system to raise probe cables up from wheel level. Do not allow the wheels to roll over accessory cables.
- Step: 5** Remove or ensure that all peripherals and items in the storage area are secure.
- Step: 6** Lock the articulating monitor arm in its lowest position.
- Step: 7** Unlock the wheel locks.

3.4.2 Moving the System

- Step: 1** Grasp the handle at the rear of the cabinet firmly.
- Step: 2** Keeping an upright posture, push the system forward. Do not push the system from either side with excessive force.
- Step: 3** Use ramps or elevators where possible and do not attempt to lift the system manually. Move system to the intended location.

Note: If moving this system between facilities by truck or other vehicle, please call Seno Service for assistance at (888) 978-8835.

3.5 Positions of Operator, Patient, and Individuals in the Scan Room

Clinical operators of the Imagio® Breast Imaging System should be in the standing position and have clear view of the Touch Screen and Display Monitor and unobstructed movement to the controls, probe, and foot switch. Patients are in the supine semi right or left lateral decubitus position with their contra-lateral arm raised above and their head, so their breast lies uniformly flat.

3.6 Confirm Laser Safety Items

3.6.1 Signs, Safety Labeling, Notices, and Laser Protective Eyewear Requirement and Warning

- The facility/hospital Laser Safety Officer (LSO) needs to reference state and or local requirements for laser safety signage. The LSO is responsible for understanding and implementing the state and or local and hospital regulations for laser safety for the Imagio® Breast Imaging System.
- Everyone in the scan room must wear the Seno provided laser protective eyewear when the Imagio® Breast Imaging System is powered on and the laser is enabled.

- Verify that the Laser Warning Sign, provided by your LSO is affixed to the outside of each door leading into the scan room. This is to warn people not to enter the room without wearing laser protective eyewear. Scattered laser radiation may be emitted from the OA/US L18-1 Probe during an exam and cause potential harm to anyone not wearing designated laser protective eyewear.
- Example of a Laser Warning Sign below:



- Verify that a Laser Warning Sign is also displayed near the Imagio® Breast Imaging System and is visible to the System Operator and the Patient. This is to remind the Patient, Operator, and Observers of the hazards of lasers.
- Verify safety labeling and notices are in place on the Imagio® Breast Imaging System.

3.7 Ensure Proper Environmental Operating Conditions



Caution: *The Imagio® Breast Imaging System is sensitive to temperature and may be damaged if used after exposure to low temperatures.*

- Do not expose the Imagio® Breast Imaging System to freezing temperatures (at or below 4° Cor 39° F).
- In case of exposure, do not use the Imagio® Breast Imaging System. Call Seno Service for assistance at (888) 978-8835.

Important: *The following environmental requirements are recommended to ensure optimal performance of the Imagio® Breast Imaging System. Please notify your Facility Engineer if there are environmental concerns.*

Operating Temperature Range 15 °C to 30 °C (59 °F to 86 °F)

Operating Humidity Range 30% to 70%, non-condensing.

Dust/Dirt Contamination Minimized, not in construction area.

Chemical Contamination Minimized; not subjected to chemical fumes.

Operating Magnetic Flux Density must be less than 5 gauss.



Caution: The OA/US L18-1 and/or US L18-5 Probes may be damaged if exposed to temperatures outside of the stated range.

- Do not store the OA/US L18-1 and/or US L18-5 Probes in environments beyond -20° C to 50° C or 30% to 85%relative humidity, as this was the range tested and verified by Seno Medical.
- Do not operate the OA/US L18-1 and/or US L18-5 Probes in environments beyond 15° C to 30° C, or 30% to 70%relative humidity, as this was the range tested and verified by Seno Medical.
- In case of exposure to excessive humidity or temperature, do not use the Imagio® Breast Imaging System. Call the Seno Service for assistance at (888) 978-8835.

3.7.1 Potential Equalization Conductor Terminal



() To use another medical device in combination with this system, an equipotential wire for connecting to an equipotential bus must be supplied. For more information, call the Seno Service for assistance at (888) 978-8835.



Caution: The Imagio® Breast Imaging System is sensitive and may be damaged if adjacent equipment is placed in close proximity to the Imagio® Breast Imaging System.

- The Imagio® Breast Imaging System should not be used adjacent to or stacked with other equipment. If adjacent or stacked use is necessary, the Imagio® Breast Imaging System should be observed to verify normal operation in the configuration in which it will be used.

3.7.2 Login and Lock/Logout

3.7.2.1 Logging Into A Session

The login new feature aims at reinforcing the Imagio® Breast Imaging System's security by means of operator profiles and logins.

Below is a table which gives an overview of each access given to each operator profile:


Operator Profiles	Access To		
	Exam (measurement, save image...)	System Configuration	
		Operator Settings	Administration
Emergency Access	YES	NO	NO
Sonographer	YES	YES	MINIMUM(1)
Admin	YES	YES	YES

(1) A minimum set of administrative settings is accessible to the operator.

At Imagio® Breast Imaging System startup, the main login screen will be displayed.



Select the desired profile.




Enter your password and select Login.

Note: Enable the auto-login feature to skip the login screen displayed at the Imagio® Breast Imaging System startup. For more information about the auto-login feature, please refer to “Operator Management” on page 154.

Note: When logged in, check which profile is in use by pressing the Touch Screen top right most operator icon button.

3.7.2.2 Locking Your Session And Logging Out

Select the round operator icon located in the top right-hand corner of the Touch Screen.



Either Lock your session or Log out.



To log out, select the dedicated icon, the main login screen is displayed.

Note: *Logging out is equivalent to an End Exam.*

To lock your session, select the dedicated icon. The system applications, patient lists, and exams are hidden until the session is unlocked.

Step: 1 To unlock your session, first Select any system console key to stop the screen saver

Step: 2 The unlock screen is now displayed



Step: 3 Enter your password and Select Unlock to go back to the current exam.

Note: Log out and/or Start an Emergency exam through the unlock screen by selecting the Emergency Access icon, which will automatically end the current exam.

3.7.3 Setting the Time and Date

The Imagio® Breast Imaging System includes a clock/calendar function, which maintains the time and date even when the Imagio® Breast Imaging System is turned off and disconnected from power.

Refer to [“Regional parameters” on page 141](#) for instructions on setting up time and date.

3.7.4 Setting the Language

Your authorized Seno Service provider will configure the operator interface language and regional parameters in the System Configuration during installation.

Refer to [“Regional parameters” on page 141](#) for instructions on setting up the language.

3.7.5 Connecting to a Network

The Imagio® Breast Imaging System supports standard network functions, which include sending exam data to a DICOM PACS.

Refer to “Ethernet Configuration” on page 150 for instructions on network configuration.

3.7.6 Selecting a Probe

The Imagio® Breast Imaging System uses four ports for connecting probes. All probe ports can be occupied at the same time, but only one probe can be active at a time.

Always use the cable management System to prevent cables from being stepped on or run over by the cabinet wheels.

3.7.6.1 Connecting a Probe


Refer to “Connecting and Disconnecting OA/US Probe, Ultrasound Probe, and Foot Switch” on page 133.

3.7.6.2 Selecting a Probe and Preset for an Exam

When the Imagio® Breast Imaging System is turned on, the Imagio® Breast Imaging System defaults to the last used probe, application and preset.

Follow the steps below to select among the probes during Imagio® Breast Imaging System operation.

Step: 1 Select Probe on the Touch Screen.



The left side of The Touch Screen will display the applications that are compatible with the probes that are connected to the Imagio® Breast Imaging System

Step: 2 Touch the tab corresponding to the desired clinical application. Ex: Breast. The Touch Screen will display the probes Use for the selected application, and the associated Presets.

Step: 3 Select the desired Preset. Ex: Breast. The Probe Touch Screen closes and the Imagio® Breast Imaging System is in live B-mode. The factory presets appear in darker gray, and the operator-created presets appear in lighter gray.

Note: It is possible to customize the displayed list of presets in the System Configuration. Refer to "System Configuration" on page 137.

The selected probe, application and preset appear on the header of the imaging main display.

3.7.7 Beginning an Exam

3.7.7.1 Getting Ready

Before acquiring images, optionally create a patient exam or retrieve a patient already created in the Imagio® Breast Imaging System.

Before scanning a new patient, make sure the previous exam has been ended, by doing one of the following:

- Select End Exam on the Touch Screen.


Note: Pressing End Exam automatically disables the laser.

- Select New Patient on the Touch Screen.

3.7.7.2 Creating a New Patient

3.7.7.2.1 Patient Data Entry

Step: 1 Select Patient on the Touch Screen. The Patient Data Entry is displayed on the main screen.



The General Information tab of the Patient Data Entry gathers the fields related to the patient identity.

The Worklist Information tab of the Patient Data Entry displays information coming from the ModalityWorklist (if associated and configured).

Eight other tabs are available, each of them displaying specific clinical application information: (for example, Breast). For any licensed application, this information will be included in a patient report.

- Thyroid
- Abdominal
- Genitourinary
- MSK

To navigate through the tabs, move the cursor to the desired tab to open and press the Touchpad. To enter new data, place the cursor in the desired field. The active field is automatically highlighted.

Type the desired text with the keyboard. Text appears when typing directly on the Touch Screen, in the preview area.

Patient Data

General Information **Imagio** Thyroid Abdominal OB-GYN Genito Urinary MSK Vascular Cardiac Worklist Information

Clinical Indications:

Pain: ☐ 1 ☐ 2 ☐ 3 ☐ 4 ☐ 5 ☐ 6 ☐ 7 ☐ 8 ☐ 9 ☐ 10 ☐ 11 ☐ 12 ☐ 13 ☐ 14

Palpable mass: ☐ 1 ☐ 2 ☐ 3 ☐ 4 ☐ 5 ☐ 6 ☐ 7 ☐ 8 ☐ 9 ☐ 10 ☐ 11 ☐ 12 ☐ 13 ☐ 14

Nipple discharge: ☐

Other Studies:

Ultrasound: Performed: ☐ Yes ☐ No Date: MM / DD / YYYY

Mammogram: Performed: ☐ Yes ☐ No Date: MM / DD / YYYY ACR Density: Findings: Microcalcification:

MRI: Performed: ☐ Yes ☐ No Date: MM / DD / YYYY

Other exam: Performed: ☐ Yes ☐ No Date: MM / DD / YYYY

Patient Information:

Prior Procedure: Breast surgery: ☐ Yes ☐ No Core biopsy: ☐ Yes ☐ No Fine needle aspiration: ☐ Yes ☐ No Date of last procedure: MM / DD / YYYY

Menstrual Status: Pregnant: LMP: MM / DD / YYYY Lactating: ☐ Yes ☐ No Prior pregnancy: ☐ Yes ☐ No Post menopausal: ☐ Yes ☐ No HRT: ☐ Yes ☐ No

Clinical History:

Comments:

IMAGIO™ BREAST IMAGING SYSTEM

PATIENT PROBE IMAGE REVIEW REPORT **END EXAM**

abc imagio

ABC

Symbols

1 2 3 4 5 6 7 8 9 0 - / Delete

! @ # \$ % & ' () * + , - . : ; Backspace

Q W E R T Y U I O P ' Enter

A S D F G H J K L ; : Shift

Z X C V B N M , . ? ↑

← ↓ →

Note: Patient ID: The Imagio® Breast Imaging System uses a unique ID to identify each patient. You may enter your own ID, or the Imagio® Breast Imaging System will create one automatically.

Images and reports are stored based on the patient ID. The name and ID appear at the top of all image displays.

3.7.7.2.1.1 To choose a Value in a List

- Move the cursor to the arrow on the right of the drop-down list.
- Press the Touchpad to display the list.
- Move the cursor down to the selected value and press the Touchpad again to validate your choice.

3.7.7.3 Loading a Patient from the Modality Worklist

3.7.7.3.1 Finding Exams in the Worklist

This feature is available if the Imagio® Breast Imaging System is connected to the network and the DICOM Worklist feature is enabled.

Step: 1 Select Patient on the Touch Screen.

- The Patient Data Entry is displayed on the main screen.
- The Patient Touch Screen is now displayed.

Note: If a Modality Worklist is configured and associated, pressing End Exam will automatically display the Modality Worklist (if configured).

Step: 1 Touch Modality Worklist on the Touch Screen.

- The modality worklist is displayed on the main screen.

There are three ways to locate a patient:

1. Check the boxes above the list to filter it.
2. Place the cursor over a column header and then press the Touchpad to sort the list by the selected column.
3. Use the search box to find a given patient.

Step: 1 Move the cursor over the desired patient file to open.

Step: 2 Press the Touchpad.

The Patient Data Entry is displayed on the main screen and pre-populated. Edit some of the information for this patient. Refer to “System Configuration” on page 137 for more information on how to configure the Modality Worklist.

3.7.7.3.2 Broad Query, Patient Query

Imagio® Breast Imaging System offers the two types of queries defined by the IHE (Integrating the Healthcare Enterprise): broad queries and patient queries.

By default, the Worklist performs broad queries.

Patient Query limits the number of answers that are returned, and therefore increases the confidentiality, limits the risk to select by mistake the wrong patient in the worklist, and limits the quantity of data to be transferred to the Imagio® Breast Imaging System.

To use Patient Query:

Step: 1 Select Patient Query on the Touch Screen to perform a query on a specific patient only. Use this mode especially if the network is slow. A window pops up.

Step: 2 Enter data for the specific patient to retrieve on the worklist. At least one mandatory field must be filled in to perform the query.

Step: 3 Select Query to perform the query on the specific patient.

Select Broad Query on the Touch Screen to switch to the Broad Worklist query mode. The Modality Worklist can be used when the Imagio® Breast Imaging System is disconnected from the network. Refer to “System Configuration” on page 137 for more details.

3.7.7.4 Editing Patient Data

3.7.7.4.1 Editable Data

Data may be edited for the current patient. Some information can be edited at any time during an exam, if not imported from the Modality Worklist.

3.7.7.4.1.1 To Edit Patient Data

Step: 1 Place the text cursor in the desired field. The active field is highlighted.

Step: 2 Type the new text with the keyboard.

If editing one of the identifier fields, a pop-up message is displayed. Select Yes to confirm or No to cancel. The identifier fields are the following:

- Patient Last Name
- Patient First Name
- Patient Middle Name
- Patient ID
- Accession Number
- All the fields in the DICOM tab of the Patient Data Entry

If editing one of the fields listed above AFTER the exam was sent to a server, the DICOM store may start a new exam with the new patient information.

Entering the patient's height and weight automatically calculates the Body Mass Index (BMI).

3.7.8 Ending an Exam

Make sure all the images required are saved. After the exam is complete, end the exam as follows: Press End Exam on the touch screen.

Note: Pressing End Exam automatically disables the laser.

4 Probes

4.1 Probe/Clinical Applications Compatibility

The following table lists all clinical applications and probe compatibility available with Imagio® Breast Imaging System.

Clinical Application	Probes	
	US L18-5 Linear Probe	OA/US L18-1 Probe
Breast	X	X

4.1.1 OA/US L18-1 Probe



The hand-held OA/US L18-1 Probe is used to obtain images of tissue. The OA/US L18-1 Probe consists of two internally sheathed cables: one cable (Optic Bundle) plugs into the Imagio® Breast Imaging System Laser Output Aperture and the second cable plugs into the Ultrasound Probe Port.

The OA/US L18-1 Probe is the only part that contacts the patient.

Refer to [“Product Specifications” on page 179](#).

4.1.2 US L18-5 Linear Probe



A US L18-5 Linear Probe or US Probe is available should your clinical needs call for a high resolution, small footprint ultrasound probe. The US Probe is a high-resolution broad bandwidth linear probe for use with the Imagio® Breast Imaging System. Refer to [“Product Specifications” on page 179](#).

4.1.3 Probe Indicator and Orientation



All probes have a raise marker on their housing indicating channel 1 or element one. It is recommended the operator align this cephalic and/or to the patients right when they are in the supine position. For example: when scanning a patient in anatomic position, this indicator is oriented cephalic (longitudinal scan plane) or towards the patients right (transverse scan plane).

4.2 Probe Handling



Caution: The OA/US L18-1 Probe and OA/US L18-1 Probe cable contain delicate optic cables that can be easily damaged if not handled carefully. If the optic cables are damaged, the OA/US L18-1 Probe may not emit the proper amount of laser light and the exam results may be degraded.

- Always store the OA/US L18-1 Probe in the Probe holder when not in use.
- Do not allow the OA/US L18-1 Probe cable to be crushed (for example: between equipment or stepped on.)
- Do not allow the cable to become crimped.
- The minimum bend radius for the optic cable is 5 cm (2"); therefore, the cable should not be wrapped tighter than an 11 cm (4") diameter.



Probe damage may be mitigated or eliminated by closely following recommendations provided below. Adhering to these "Do's and Do Not's" will help prolong the life of the Seno Medical probes.

Caution: Failure to follow these recommendations can result in damage to the equipment.

DO:

- Use care when handling the probes.
- Only connect or disconnect the probes by holding onto the connectors.

- Routinely inspect the entire probe assembly; optic ferrule, ultrasound connector, cabling, and the probe for any damage prior to connecting it to the Imagio® Breast Imaging System. If damage is discovered during a visual inspection, discontinue use and call Seno Service for assistance at (888)978-8835.

Possible damage may include, but is not limited to:

- Probe pins that are bent or broken within the ITT Cannon ZIF Connector
- Split or cut cabling or shielding
- Surface cracks, gouges, or rough spots on the housing
- Surface cracks on the face of the probe
- Damage/separation of the acoustic lens
- Damage/cracked/chipped on OA/US L18-1 Probe Optic Ferrule
- Debris on the light sensor pads that are adjacent to the Optic Ferrule
- Chips or cracks on glass windows
- Ensure that the OA/US L18-1 Probe glass windows and calibration port shield are clean and dry prior to performing laser energy calibration.
- Follow procedures contained in this manual regarding probe inspecting, cleaning and disinfection.

DO NOT:

- Drop or hit the probe or probe face. Even moderate blunt force to the probe may damage the crystal elements and lead to product failure.
- Do not allow probe cabling to hang where it could be run over, caught, or pulled by the Imagio® Breast Imaging System cabinet wheels/casters while stationary or during movement.
- Do not apply excess force when removing or inserting the OA Probe Optic ferrule from or into the Laser Output Aperture or the removing or inserting the ITT Cannon ZIF connector.

4.3 Probe Sheaths for Use with US L18-5 Linear Probe Only

Under certain conditions where the probe may encounter mucous membranes, blood or other bodily fluids, the use of a probe sheath is recommended.

- After use, the single-use sheath should be removed and discarded.

The following tables list basic information regarding sterile probe sheaths which will physically fit our ultrasound probes. Please check whether the recommended products in the tables below are approved for use. Call the Seno Service for assistance at (888) 978-8835 for an equivalent.

4.3.0.1 US L18-5 Linear Probe Dimensions and Compatible Sheaths (for reference only)

Probe width 6.58 cm (2.6 in.), cable length 2.1 m (82.7 in.)

For reference only. Always refer to each CIVCO sheaths current instruction for use.

Manufacturer	Sterility	Latex/Latex Free	Description	Product Number	Dimensions
CIVCO	Sterile	Latex Free	CIV-Flex/ Telescopically Folded	610-637	8.9 x 91.5cm
CIVCO	Sterile	Latex Free	CIV-Flex/ Telescopically Folded	610-1000	10.2 x 147cm
CIVCO	Sterile	Latex Free	CIV-Flex/Flat Folded	610-001	8.9 x 61cm



Warning: When using a sheath with a probe, only non-latex sheaths should be used due to the prevalence of latex allergies.

4.4 Ultrasound Transmission Gel

Only use Seno approved gel.

Aquasonic Clear® Ultrasound Gel is required because of its optical compatibility with our probes.

4.5 Gel Holder/Warmer

An on-cabinet gel holder has been provided as a convenience. The gel holder is located on the right side of the Console.




Important: Place gel bottles in the holder with the nozzle down so that the gel is always ready to flow.

A gel warmer is built into the System Console for your convenience. It is directly integrated onto the gelholder. It heats gel to 86°F (30°C) in an hour and up to 98.6°F (37°) in 4 hours.

Note: If it is activated, please note that the gel warmer will alternatively and automatically activate and deactivate in order to regulate the gel temperature.

To activate or deactivate the gel warmer, press the gel warmer icon located on the top right of the TouchScreen.



5 Cleaning System Components

5.1 Cleaning and Disinfecting Agents

The following cleaning and disinfectant agents have been tested by Seno. They are the only cleaner/disinfectants approved by Seno for cleaning the probe holder, calibration port, OA/US calibration probe adapter, and to clean and disinfect the Ultrasound only probe as well as the OA/US probe. Always reference the manufacturer's directions for use labeling affixed to each cleaner/disinfectant agent. Please reference the websites immediately below for availability to specific regions/countries for cleaning and disinfection purposes.

- Sani-Cloth® AF3 is approved for use on both L18-5 US Probe and L18-1 OA/US Probe <http://pdihc.com>
- CaviWipes™ US is approved for use only on L18-1 OA/US Probe. <http://www.metrex.com>.
- CaviWipes™ EU is approved for use only on L18-1 OA/US Probe. <http://www.metrex.com/EU/>.
- Sani-Cloth® Bleach is approved for use (globally) with the L18-5 US probe. <http://pdihc.com>
- Sani-Cloth® Super is approved for use (globally) with the L18-5 US probe. <http://pdihc.com>

5.1.1 Cleaning the Laser Protective Eyewear



Caution: *The Imagio® Breast Imaging System laser protective eyewear has delicate lenses which may be easily damaged by misuse of cleaning agents and solvents.*

- Do not use cleaning solvents on the laser protective eyewear.
- Only use water and a soft cloth or approved lens cleaning towelettes.



5.1.2 Cleaning the Cabinet, System Console, Handles, Power Cord, Monitor Housing and Gel Warmer

Be sure to clean any unused gel from the surface of the system console after Imagio® Breast Imaging System use.

Power off the Imagio® Breast Imaging System and Unplug the power cord from mains power outlet and the Imagio® Breast Imaging System before cleaning the power cord.

Apply a small amount of one of the following recommended cleaning solutions to a soft, non-abrasive cloth and wipe down the bezel:

- Water
- Mild detergent (PH level at or near 7) and water solution.
- Use warm water and a soft bristle brush to remove stubborn dried gel.

5.1.3 Cleaning the Monitor Display Screen Touch Screen Display

A clean, soft cotton cloth dampened with an ammonia-based window cleaner should be used to clean the Monitor display. Do not use paper towels, as these may scratch the display monitor. The cleaning agent should be sprayed onto the cloth, not directly onto the display surface. Take care not to get any cleaning agent into the housing of the Imagio® Breast Imaging System. This may lead to Imagio® Breast Imaging System damage.

5.1.4 Cleaning the Probe Holder, OA/US L18-1 Probe Calibration Adapter, Calibration Port - Energy Sensor Shield, and Flat Breast Phantom

Daily or as needed. Clean or disinfect with Seno approved cleaner disinfectants. Refer to [“Performing Laser Energy Calibration” on page 101](#).

5.1.5 Cleaning the Foot Switch

Apply a small amount of one of the following recommended cleaning solutions to a soft, non-abrasive cloth:

- Water or 70% isopropyl alcohol

5.1.6 Cleaning the Air Filter

The Imagio® Breast Imaging System is equipped with an air filter. The Imagio® Breast Imaging System displays an icon on the display left bottom as a reminder to check the air filter. The air filter is located at the back of the Imagio® Breast Imaging System.

It is recommended that the air filter be cleaned monthly or more frequently if a significant accumulation of dust or debris is noted.

The air filter should be checked at least monthly and cleaned as needed to keep the system functioning properly and avoid overheating. The air filter is located on the rear of the cabinet. To check the filter, perform the following steps.

Step: 1 Power off the Imagio® Breast Imaging System.

Step: 2 Remove the air filter using the built-in hand hold on the air filter cover.

Step: 3 Inspect the filter and clean as needed using one of the following methods. Replace the air filter if it is damaged. Additional air filters can be ordered. Call Seno Service for assistance at (888) 978-8835

a) Method 1 Vacuum Clean: A few passes of a vacuum cleaner will remove accumulated dust and dirt quickly.

b) Method 2 Blow with Compressed Air: Point compressed air nozzle in opposite direction of operating air flow (blow from exhaust side toward intake side).

c) Method 3 Water: Under normal conditions, collected dirt is washed away quickly and easily using a standard hose nozzle with plain water. If stubborn air-borne dirt is present, the filter may be dipped in a solution of warm water and mild detergent. Then simply rinse in clear water, let stand until completely dry and free of moisture, and return to service.

5.1.7 Cleaning Frequency

Perform the maintenance operations as described in this chapter. Cleaning of the Imagio® Breast Imaging System should occur at regular intervals, or more frequently as needed.

Seno provides the following recommended cleaning intervals as guidelines:

Part		Recommended Cleaning Frequency
	Probes:	Inspect, clean, and disinfect between patients.
	Laser Protective Eyewear	Inspect the eyewear prior to use. Clean if visibly dirty. It is recommended operators always wear eyewear with clean lenses.
	Display Monitor	Clean weekly, or as needed.
	System Console	Clean daily, or between patients.
	Touchscreen Display	Clean daily, or between patients.
	Exterior: chassis and Handles	Clean weekly, or as needed.
	Air Filter	Clean monthly.

It is prudent to continually survey the Imagio® Breast Imaging System for maintenance needs. Call Seno Service for assistance at (888) 978-8835.

5.2 Probe Inspecting Cleaning and Disinfecting

5.2.1 Probe Inspecting Cleaning and Disinfection Guidance

Use Seno approved cleaning and disinfecting agents only. Probes are sensitive and may easily be damaged by harsh cleaning agents, corrosive chemicals, or mishandling. Always use care when handling the probes.

The following generalized instructions for inspecting, cleaning, and disinfection are indicated for use with the probes.

The probe must be cleaned after each use for further disinfection.


The functional life of the probe(s) which are accessories to the Imagio® Breast Imaging System have been robustly tested and their functionality will be verified by the service technician as part of the recommended Seno 6-month Service and Maintenance program. However, repeated misuse, improper cleaning and disinfection, or careless handling may cause damage to the probe.

5.2.1.1 Inspecting the Probe


Step: 1 Disable the laser when inspecting the probe, cleaning the probe or when the Imagio® Breast Imaging System will be unattended. *It is not necessary to disconnect the probe from the cabinet*

Step: 2 As a precaution, visually inspect the probe face from an acute angle. Look for damage, especially cracks, or breaks, in the glass windows or separation of the lens of the probe, as well as gouges or breaks in the housing or gasket material that could scratch the patient or that could be a void for contaminants to accumulate. Inspect the probe cable for damage, especially tears and signs of it being crushed. Call Seno Service for assistance at (888) 978-8835 if damage is observed.

Cracked Glass



Lens Separation



Step: 3 After the probe and probe cable passes visual inspection, place into the Imagio® Breast Imaging System probe holder and proceed to the instructions for cleaning the probe.

5.2.1.2 Cleaning and Disinfecting the Probe with any approved Cleaning/Disinfecting Agents

To prepare to clean and disinfect the probe(s) it is important to check the directions provided with the cleaner/disinfectant as well as to follow those steps which have been verified effective. With all cleaning and disinfecting products called out below, it is important to:

- wear gloves,
- it is recommended user wear chemically protective eyewear,
- prepare the cannister or packet holding the cleaning/disinfecting agent per the directions for

- use on the packaging,
- check the cleaner/disinfecting agent directions for use, and
- visually inspect the probe in a well-lit area prior to cleaning, following cleaning pre-disinfection, and post disinfection.

A) Using Sani-Cloth® Super Cleaner/Disinfectant on L18-5 US Probe

5.2.1.3 Step: 1 Following the generic preparations cited above, wipe the internal surface of the probe holder, and continue until all visible soil is removed using the Sani-Cloth® Super [Disposable Wipe](#).

Step: 2 Use Sani-Cloth® Super Disposable Wipes to thoroughly clean all surfaces of the L18-5 US probe, paying special attention to crevices and hard-to-clean-areas. Include the cable but avoid applying liquid to the electrical connection end. Replace soiled wipe as needed and use additional wipes to ensure that all surfaces are uniformly cleaned and that no visible soil remains.

- **A1)** Disinfect with Sani-Cloth® Super Cleaner/Disinfectant on L18-5 US Probe www.metrex.com.

Step: 1 As long as there is no visible soil remaining on the L18-5 US Probe, disinfection can begin by wetting the surface of the probe face for two (2) minutes.

Step: 2 Allow the surface to air dry.

B) Using Cavi-Wipe® Cleaner/Disinfectant on L18-1 OA/US Probe

Step: 1 Following the generic preparations cited above, Wipe the internal surface of the probe holder, and continue until all visible soil is removed using the.

Step: 2 Use Cavi-Wipe® Disposable Wipe to thoroughly clean all surfaces of the L18-5 OA/US probe, paying special attention to crevices and hard-to-clean-areas. Include the cable but avoid applying liquid to the electrical connection end. Replace soiled wipe as needed and use additional wipes to ensure that all surfaces are uniformly cleaned and that no visible soil remains.

B1) Disinfecting using the Cavi-Wipe® Disposable Wipe on L18-1 OA/US Probe

Step: 1 As long as there is no visible soil remaining on the L18-1 OA/US Probe, disinfection can begin by wetting the surface of the probe face for two (2) minutes.

Step: 2 Allow the surface to air dry.

C) Using Sani-Cloth® Bleach Cleaner/Disinfectant on the L18-5 US Probe

Step: 1 Following the generic preparations cited above, wipe the internal surface of the probe holder, and continue until all visible soil is removed using the Sani-Cloth® Bleach Disposable Wipe.

Step: 2 Use Sani-Cloth® Bleach Disposable Wipes to thoroughly clean all surfaces of the L18-5 US probe, paying special attention to crevices and hard-to-clean-areas. Include the cable but avoid applying liquid to the electrical connection end. Replace soiled wipe as needed and use additional wipes to ensure that all surfaces are uniformly cleaned.

C1) Disinfect with Sani-Cloth® Bleach Cleaner/Disinfectant

Step: 1 Using a second Sani-Cloth® Bleach Wipe, wet the external surface of the probe face for a minimum of

four (4) minutes. Use additional wipes as needed to keep the surface of the probe face wet for the full four (4) minutes. At the end of the four (4) minutes, wipe until all visible soil is removed.

Step: 2 Visually inspect the probe in a well-lit area. If any soil remains, clean and disinfect until all visible soil is removed.

D) Using Sani-Cloth® AF3 Cleaner/Disinfectant on the L18-5 US and L18-1 OA/US Probes

Step: 1 Following the generic preparations cited above, wipe the internal surface of the probe holder, and continue until all visible soil is removed using the Sani-Cloth® AF3 Disposable Wipe.

Step: 2 Use Sani-Cloth® AF3 Disposable Wipes to thoroughly clean all surfaces of the probe, paying special attention to crevices and hard-to-clean-areas. Include the cable but avoid applying liquid to the electrical connection end. Replace soiled wipe as needed and use additional wipes to ensure that all surfaces are uniformly cleaned.

D2) Disinfect with Sani-Cloth® AF3 Cleaner/Disinfectant

Step: 1 Using another Sani-Cloth® AF3 Wipe, wet the external surface of the probe face for a minimum of three (3) minutes. Use additional wipes as needed to keep the surface of the probe face wet for the full three (3) minutes. At the end of the three (3) minutes, allow to air dry.

Step: 2 Visually inspect the probe in a well-lit area. If any soil remains, clean and disinfect until all visible soil is removed.

6 System Power On Off

6.1 Circuit Breaker

The Imagio® Breast Imaging System Circuit Breaker is located on the lower, back of the cabinet. Always leave the circuit breaker in the ON position for standard, daily operation.

6.2 Power On Procedures

Important: Wait 30 seconds after power off if an immediate power on is required.

Important: Performing the steps in the exact order is important to control electrical power distribution through the Imagio® Breast Imaging System.

Step: 1 Ensure the power cord is plugged in to the Imagio® Breast Imaging System.

Step: 2 Ensure that the Circuit Breaker located at the rear of the cabinet is switched to the OFF position.

Step: 3 Ensure the Imagio® Breast Imaging System is positioned at least 12" (30 cm) away from the wall and there are no air circulation restrictions.

Step: 4 Plug the Imagio® Breast Imaging System power cord into the proper power outlet.


Step: 5 Switch the Circuit Breaker to the ON position.

Step: 6 Momentarily press the Power button located on the Console to begin the power on sequence.

Step: 7 Ensure the Imagio® Breast Imaging System powers on to display the Login Screen before using the Imagio® Breast Imaging System.

Important: If the Imagio® Breast Imaging System is started without a probe attached, the message, "Please connect a transducer" is displayed. Connect and select the transducer on the Probe tab.

Important: If the Imagio® Breast Imaging System was just turned on, it may take up to 14 minutes for the lasers to reach operating temperature and before an OA Scan can be performed. The following Laser Icon will be displayed on the bottom right of the display monitor.



6.3 To Power Off the Imagio® Breast Imaging System

Step: 1 Press and release the Power button located on the Console. The power off sequence can take up to 5 minutes and will be complete when the cooling fans stop.



Caution: The Imagio® Breast Imaging System is a sensitive electronic instrument and may be damaged or data may be lost by improper or misuse.

- Do not disconnect the OA/US L18-1 Probe during live imaging.
- Do not use the Circuit breaker or main power for regular power off.
- Always perform an orderly power off of the Imagio® Breast Imaging System.

7 System Console

7.1 Controls

7.1.1 General Information

The system console includes the Touchpad and controls such as buttons, dials, and dial-buttons.

Press a button to activate or deactivate its function.

Turn a dial to change the selected setting.

Press a dial-button to activate its function. When a dial-button is associated to a value, turn it clockwise to increase the value of the settings or turn it counter-clockwise to decrease the value of the settings (unless configured differently in System Configuration).

Note: Pressing on buttons related to options that are not available will have no effect.

7.1.2 Touchpad

The Touchpad allows movement and navigation of the cursor in the operator interface of Imagio® Breast Imaging System. It also enables the operator to perform basic actions and measurements during exams by means of gestures.

Note: The Touchpad is designed to be used while wearing barrier protective gloves and while being covered during an interventional procedure.


The following table indicates how to perform the different gestures mentioned in this Operator Manual (unless configured differently in the System Configuration):

Gesture name	How to perform the gesture
Press	 Press on the Touchpad
Swipe	 Move your finger(s) vertically or horizontally

Gesture name	How to perform the gesture
Pinch/Spread	 Do a pinch or spread gesture using 2 fingers

The following table lists the main gestures available by default (unless configured differently in System Configuration):


Controls	Associated gesture
Select/Press	 Press with 1 finger
Freeze	 Press with 2 fingers
Clip navigation	 When in frozen mode: Swipe right with 2 fingers to navigate the clip forward Swipe left with 2 fingers to navigate the clip backward
Save image	 Press with 3 fingers
Save clip	 Press with 4 fingers

Controls	Associated gesture
Depth	 Swipe up with 3 fingers to decrease the depth Swipe down with 3 fingers to increase the gain
Gain	 Swipe right with 3 fingers to increase the gain Swipe left with 3 fingers to decrease the gain
Zoom	 Pinch with 2 fingers to decrease the zoom Spread 2 fingers to increase the zoom
Move a box	 Swipe 1 finger in the desired direction
Arbitration (Update)	 Press the Touchpad with 4 fingers to change between Box Move (position of the box) and Box size (resize the box). Swipe in the direction needed to reposition or resize the box. Press the Touchpad with 4 fingers to switch between Img1 and Img2 in dual displays.

Note: Make sure your fingers are placed on the Touchpad before pressing.

A gesture reminder area is available by pressing the gesture icon on the top right of the Touch Screen

The gesture reminder area is now displayed on the display monitor bottom right.









These gestures are entirely customizable in System Configuration.

Please refer to “Common Imaging Controls” on page 73 for more information about each control/gesture combination.

7.1.3 Touch Screen

The Touch Screen contains several types of controls, depending on the function to perform.

Touching this type of control...	... triggers this event
	Touch a tab to display a different imaging mode menu.
	Touch to open a page or function. Touching the Keyboard control in this example displays the keyboard.

Touching this type of control...	... triggers this event
	Touch to turn a function On or Off. The LED button is blue when the function is on, and dark blue when it is Off.
	Touch to change the value displayed on the button. Repeatedly pressing this type of button cycles through all the values. The value displayed is indicated by the position of the mark.
	Rotate the dial located below the Touch Screen label to change the value.
	Touch to change the value displayed on the button. The value displayed is indicated in the box.
	Touch the appropriate part of the button to change its value.

Note: Adjust the Touch Screen brightness in System Configuration, System sub-tab.

7.2 Monitor and Articulated Arm

The monitor is mounted on an articulated arm that permits it to be positioned vertically and horizontally. Adjust the position of the monitor to suit different operating positions and operator heights.

When it is released from its locked transport position, the monitor can be tilted up and down, swiveled left and right, and moved from side to side.

To release the monitor from its locked transportation position, rotate the knob on the arm counter-clockwise (blue arrow).

For transportation, put the monitor in its lower position and rotate the knob clockwise (red arrow) in order to lock the monitor in this position. If the system needs to be packed for transportation the monitor can be put in flat position.

To adjust the monitor, grasp it by its handle and tilt, swivel or move it.



Configure the brightness of the screen using the -/+ buttons located at the bottom right hand corner of the monitor.

7.3 How to Move the Console

The Imagio® Breast Imaging System must be powered on to move the console.

Press and hold the console release button located on the handle and move the console to the desired left-right, up-down position and release the button to lock the consoles position.

7.3.1 Speakers and Audio Adjustment

The Audio Volume may be adjusted in System Configuration.

Note: Either mute system sounds such as Save Image and Save Clip (in System Configuration, System/ Display tab, System sub-tab) or mute spectral audio volume when in PW by pressing on the Audio volume button.

8 OA Mode

8.1 OA Mode Selection

Press the OA button on the System Console to enter OA Mode. All OA parameters are automatically optimized. Only trained and authorized operators are permitted to perform OA scans with the Imagio® Breast Imaging System. Safeguards have been designed into the Imagio® Breast Imaging System to protect the Operator and the Patient from laser energy; however, it is the responsibility of the Operator to follow all Warnings and Cautions outlined in this manual. Refer to “Safety Descriptors” on page 8.

Once the Imagio® Breast Imaging System completes a power on it defaults to OA Mode and the OA/US L18-1 Probe is in laser standby not emitting laser energy. To enter OA mode at any time, press the Console OA button.

Important: Laser Icons and Status Codes appear on the bottom right of the display monitor to communicate the laser state to the Operator.



Warning: Invisible laser light is emitted from the OA/US L18-1 Probe. Reflected and/or misdirected laser light may damage eyes.



- Do not use any controls or adjustments or performance settings other than those specified herein.
- The Patient, the Imagio® Breast Imaging System Operator, and everyone else in the scanroom must always wear Seno approved laser protective eyewear when the Imagio® Breast Imaging System is powered on and the laser is enabled.
- Wearing Seno approved laser protective eyewear is a laser safety requirement. It is necessary to wear laser protective eyewear when, for example; scanning, inspecting, and cleaning the OA/US L18-1 Probe, and calibrating laser energies or any other time the Imagio® Breast Imaging System is powered on and the laser is enabled.
- Do not look directly at the laser beam even when laser protective eyewear is being worn.
- Do not look at a laser beam emitted from the Imagio® Breast Imaging System that is visible. The Imagio® Breast Imaging System uses lasers that are classified as invisible; however, some individuals can see light beyond the visible spectrum.
- Always point the OA/US L18-1 Probe away from everyone’s eyes.
- Only use your foot to press the foot switch.
- Always release the foot switch before removing the OA/US L18-1 Probe from the skin or calibration port.
- Always follow the laser safety precautions prior to laser emitting to protect the Patient, Operator, and Observers.
- Always place the OA/US L18-1 Probe in the Probe holder when not in use.



Warning: Operation of the Imagio® Breast Imaging System by untrained, unauthorized individuals may result in harm to patient and/or operator.

- Do not allow the Patient to touch the Console.
- Do not share your password or Key Card with anyone. Only trained operators can use the Imagio® Breast Imaging System.
- Always disable the laser by selecting the Display Laser Control button when the Imagio® Breast Imaging System will be left unattended.




Warning: Laser emission in the presence of flammable materials may cause fire or explosion.

- Do not emit laser in the presence of flammable materials, solutions, or gases; or in an oxygen enriched environment to include:
 - Gases containing more than 21% oxygen
 - A breathing atmosphere containing more than 21% oxygen
 - A liquid with greater than 21% oxygen (for example, RL (rich liquid) is usually 35-40% oxygen)
 - The use of nasal prongs is not considered an oxygen rich environment.

8.2 Laser Heating

Laser is warming up and has not reached operating temperature. Warmup.. xx% status code will also be displayed. This laser status icon will change, and the status code will disappear when the laser reaches operating temperature.



Important: If the Imagio® Breast Imaging System was just turned on, it may take up to 14 minutes for the lasers to reach operating temperature and before an OA Scan can be performed. The following Laser Icon will be displayed on the bottom right of the display monitor.

Laser Heating will take up to 15 minutes from initial power on or laser sleep.

Laser Heating will occur upon power on or after the laser is enabled and the freeze button is pressed twice freeze then unfreeze.

8.3 Laser Sleep


Laser will discontinue heating and enter laser sleep if no operator input is detected within 60 minutes by default. The time may be adjusted in the System Configuration.

To exit sleep, Laser Authorization must be performed to enter Laser Status Enabled

8.4 Laser Standby

The Imagio® Breast Imaging System defaults to Laser Standby. However, if the Laser is Enabled and in Laser

Ready, pressing the Console Freeze button in OA-mode commands the Imagio® Breast Imaging System Laser to stop all laser energy generation. The Laser Status icon displays Laser standby.



8.5 To Place the Imagio® Breast Imaging System in Laser Standby

Step: 1 Press the Console Freeze button in OA-mode.

The following occurs:

The laser state message icon changes to Laser Standby.


8.6 To Enable the Laser - Authorizing the Laser

Step: 1 Confirm that the Display Monitor is displaying the message Laser Status Disabled on the imaging screen.




Step: 2 Place the laser authorization card over the NFC Reader located on left of TouchPad of the Imagio® Breast Imaging Console.

- A status message is generated and displayed on the monitor bottom to indicate who authorized laser and is logged to notifications list. *Operator "tag 1" in this example.*



- The laser status will change to Laser Enabled and the switch will slide to the right and the background color will be green.



- If the imaging is not frozen and laser is heated the Laser Ready Icon is displayed,  Laser Ready.
- If the laser is not heated the icon will indicate Laser Heating,  Laser Heating.

A message appears on the monitor bottom left reminding the operators to check that everyone in the scan room is wearing laser protective eyewear.

The Imagio® will remain in Laser Status Enabled if the operator selects B-mode and the operator keeps the Imagio® Breast Imaging System active. This allows the operator to toggle back and forth into other imaging modes, the laser will not emit until OA mode is selected and foot switch is pressed.


Important: *Laser will be disabled after 15 minutes of inactivity, even if NFC card is left on console.*

8.7 Laser Adjustment Period

The Laser Adjustment Period is designed to ensure that the laser has time to compensate for variations prior to OA/US L18-1 Probe laser emission. The Laser Adjustment Period occurs when the thermal state of the Imagio® Breast Imaging System laser has changed, for example, after being powered off or allowed to sleep.

The Imagio® Breast Imaging System will begin a Laser Adjustment Period after the laser is first powered on, laser energy calibration, or laser sleep. During the adjustment period the Imagio® Breast Imaging System will emit laser with the Imagio® Breast Imaging System internal laser shutter closed so there is no laser emission from the OA/US L18-1 Probe. The adjustment period generally requires 3-10 seconds and a maximum of 60 seconds.

- No laser energy will emit from the OA/US L18-1 Probe during the Laser Adjustment Period; however, **it is required that everyone in the Imagio® Breast Imaging System suite wear laser protective eyewear and still take all necessary laser safety precautions**
- The Laser Adjustment Period is started when an operator unfreezes the Imagio® Breast Imaging System (with or without the foot switch pressed)
- The laser status icon Preparing is shown on the bottom right of the display
- No OA images will be displayed on the monitor until adjustment is complete




8.8 To Enter Laser Ready

Step: 1 Press the Console Freeze button on the Console.

The following occurs:


- The laser state message icon changes to Laser Ready if the laser is enabled and up to operating temperature. The laser needs to be up to operating temperature for this to occur.

A message appears on the display monitor bottom to remind the operator to check that everyone in the scan room is wearing laser protective eyewear.



8.8.1 Foot Switch Operation

The foot switch controls laser light emission from the OA/US L18-1 Probe.



Important: *If OA mode has been exited, the foot switch must be released before re-entering OA mode for proper Imagio® Breast Imaging System operation.*

8.8.2 How to Use the Foot Switch

Step: 1 Enter OA mode.

Step: 2 Authorize the Laser. Ensure Laser Ready.

Step: 3 Place the OA/US L18-1 Probe in contact with the surface. Be sure Laser Ready is indicated. If laser ready is not indicated and laser standby is indicated, press the system console freeze button.

Step: 4 Press the foot switch down and hold it to emit laser. The laser state icon indicates laser emitting.

Step: 5 Release the foot switch (or press the freeze button on the system console), to stop laser emission. The laser status icon will transition from emitting to laser ready if the foot switch is released. Pressing freeze will stop laser emission and the laser status icon will indicate laser standby.

8.9 Probe Contact Sensor

The OA/US L18-1 Probe is equipped with surface detection sensors which disable laser emission when the probe face is not in direct contact with a reflecting surface in order to help prevent accidental laser emission. The (4) four sensors for this application are in each corner of the probe face.

They visually indicate to the operator that the probe will emit laser.

- Laser emission will occur when the operator completes the following steps:

1. The Imagio® Breast Imaging System is in OA Mode. Press the OA button on the System Console.
2. The Laser has been authorized indicated by the message “Laser Enabled” and the slide switch positioned right and displaying green. A message appears on the display to remind the operators and everyone in the scan room wear laser protective eyewear.
3. The system is unfrozen, and the laser status icon indicates “Laser Ready”. *Press Freeze/Unfreeze if the image is frozen and the laser status icon is indicating “Laser Standby”.*
4. The probe face is in contact with the surface.
5. The foot switch is pressed.

8.9.1 No Probe Contact and No Laser Emission Remedy

When the Imagio® Breast Imaging System is in OA mode, the laser has been authorized and the Imagio® Breast Imaging System is in a live imaging state or (not frozen) - indicated by the green “Laser Ready” messaged icon and then the foot switch is pressed, if the probe was not in contact with the surface, the probe contact sensor trips and laser will **not** be emitted. A message will also flash on the bottom right of the display stating “No Probe Contact” reminding the operator the probe was not in contact with the surface.



- To correct for this and emit laser:

1. If continuing to press the foot switch, place the probe in contact with the surface. Laser will be emitted.

- If the foot switch has been released, ensure laser ready, and place the OA/US L18-1 Probe in contact with the surface first, then press the foot switch to emit laser.

Slight variations in probe surface contact won't stop laser emission. This gives the operator the ability to heel-toe the probe or to slightly angle the probe when attempting, for instance, to scan over and around the areola or on irregular or uneven surfaces at the axilla without interrupting laser emission.

Important: Any probe face skin and surface air gaps or gel stand-off may negatively affect OA image quality. Every attempt should be made to keep the entire OA/US L18-1 Probe face in full contact with skin to obtain maximum light transmission from the probe into the skin and tissue. If full contact is unavoidable, for example around the areola or axilla, the operator can attempt to scan in different scan planes or use slight probe angulation to maximize contact and light transmission.

8.10 To Disable the Laser

Step: 1 Confirm that the Display Monitor is displaying the message Laser Enabled on the imaging screen.

Step: 2 Press the cursor button to activate the cursor and click the green slider to move to the laser Enabled slide switch on the display to .



- The message Laser Disabled appears, and the switch moves to the left revealing a gray background.

Note: Pressing End Exam automatically disables the laser.




8.11 OA Display

When the Imagio® Breast Imaging System power on initialization is complete, the Imagio® Breast Imaging System automatically enters OA Mode as the default mode

If the Patient tab is active, press the Image tab to enter the live imaging mode.


8.11.1 Display Description



Areas of the Display

1	Gray-scale	The gray-scale applies only to the B-Mode image.
2	Relative Absorption Color Scale	Displays the order in which colors are used to depict relative levels of laser light absorption.
3	Image area	Displays gray-scale and co-registered and temporally interleaved OA images.
4	B-Mode Ultrasound Parameters	Current B-Mode Ultrasound parameters.
5	Laser Status Bar	Displays laser status such as: Laser state (e.g. Standby, Ready, Emitting) Provides laser enable control, calibration control, and a history of laser notifications. Software patents and version information are located within the "About" section of the Notifications control window.
6	Exam Information	Facility ID + Patient ID, Patient Initials, Probe Name-Imaging Mode, Operator ID, Date of Exam, Time of Exam
7	Data Store	Still Images and Clips
8	Cine Status	Retrospective recording memory buffer

8.11.2 Display Monitor Screen (During Live OA Scan)



Description of the OA Display

B-Mode	Displays grayscale ultrasound to identify the morphology of the imaging area. OA
Total	The OA Total image represents the features that appear prominently in both the
	OA Short and OA Long image.
OA Relative	The OA Relative image presents features that appear in the OA Short but not the OA
	Long image as a Red indication and features that appear in the OA Long but not the
	Short as a Green indication.
OA Combined	The combined image represents features in the OA Relative image that appear
	prominently in the OA Total image.
OA Short	Represents the Opto-acoustic (OA) images for the short laser wavelength, before the
	colorization process
OA Long	Represents the Opto-acoustic (OA) images for the long laser wavelength, before the
	colorization process

8.12 Description of OA Mode Imaging Parameters

B-mode imaging parameters are available in OA Mode. There are no specific OA Mode imaging controls available to the operator. See Sections [All Modes Common Parameters \(Section 11.1.1\)](#), [About B-Mode Touch Screen \(Section 12.2.4\)](#), and [Description of B-mode Imaging Parameters \(Section 11.2.3\)](#).

8.13 SenoGram®

The SenoGram® is clinical decision support software with artificial intelligence (AI) designed to provide a frame of reference for a radiologist's reading of OA/US images. The SenoGram® helps translate the Imagio® Breast


Imaging System OA/US feature score input by the Radiologist into a continuous POM and subsequently the probability of cancer. Feature scores and various patient data are entered into the SenoGram® in order to aid the physician in determining the final POM, subsequent BI-RADS category, and ultimately the decision to biopsy or not.

(Refer to the didactic SenoGram® training resources for detailed instructions on how to use the SenoGram®.)

8.14 Region of Interest (ROI)

The Operator can “trace” multiple regions of interest (ROI) on both B-Mode and OA images while the images are frozen or after they have been recorded. Free-form ROI loops can be traced around an object of interest, and graphic arrows and text can be applied.

The ROI loop is pink when being drawn or modified and white when the ROI loop is closed. An ROI loop may be traced on the B-mode grayscale image also. Refer to “Trace Region of Interest (ROI)” on page 110.



8.13.1 To Draw a Free-form Loop around a Region of Interest (ROI)

There are two ways to draw an ROI:

8.13.1.1 Touchpad ROI

Step: 1 Select the Measure button.

Step: 2 Select the Trace button.

a) A pink cross appears or an alternate icon indicator if there is more than one ROI.

Step: 3 Drag your finger to position the start location and press Touchpad to set start location.

b) **Step: 4** Drag the cursor to outline the shape with your finger on the Touchpad.

Step: 5 End the ROI at the start location by lifting your finger off the Touchpad or end by lifting your finger off the Touchpad in close proximity to the start location and press the End button to close the ROI.

a) The perimeter and distance is indicated on the display.

8.13.1.2 Touch Screen ROI

Step: 1 Select the Measure button

Step: 2 Select the Trace button

a) A pink cross appears or an alternate icon indicator if there is more than one ROI

Step: 3 On the touch screen grayscale display, press and hold anywhere on the Touch Screen then drag your finger to the shape of the ROI

Step: 4 End the ROI at the start location by lifting your finger or end by lifting your finger off the touch screen in close proximity to the start location and press the End button to close the ROI.

a) The perimeter and distance is indicated on the display.

Note: In OA Mode, closing the ROI populates all 6 up displays.

Note: Press Touch Screen Clear Last ROI or Clear All ROI button to remove ROIs from the image on the display.

9 Imaging Artifacts


9.1 OA Artifacts

The Imagio® Breast Imaging System displays co-registered Opto-acoustic and Ultrasound (US) images in real-time. As with any imaging modality, consideration must be given to the physical principles by which it works and how these principles affect the output image.

Only trained and authorized operators may interpret OA images and video for medical diagnosis. The following potential combination Opto-acoustic and US only image artifacts are noted:

9.1.1 Gel Standoff, Air Trapping, and Air Gap


Gel standoff occurs when there is too much gel applied to the skin causing the probe face to not be in close contact with the skin. On grayscale it appears as an anechoic line between the probe face and skin surface. On OA it prevents light from penetrating tissue. It is resolved by spreading out the gel so that only a wet surface exists.



Air trapping occurs when the gel contains trapped air bubbles. On grayscale it may appear as a ringdown artifact. On OA it may appear as a drop out or cause interference lines. It is resolved by spreading out the gel so that only a wet surface exists.


Air gap occurs when the probe face is lifted off the surface. On both grayscale and OA there is a drop out of signal. It is resolved by placing the entire probe face in contact with the surface.

Air Trapping and Air Gap have a similar appearance:



9.1.2 Temporal Mis-registration

Because the OA and US images are temporally interleaved, there can be a slight difference in the location of the probe during the time between OA image acquisition and the US image acquisition. This will manifest itself in the OA Combined, OA Relative and OA Total regions of the display showing colorization on areas of the B-Mode image that are displaced by the amount of probe movement. To minimize, operators are trained to use a slow scanning technique of no more than 5 mm/sec. Moving the probe faster, could result in potential mis-registration of the OA colorization and gray-scale images and can affect the diagnostic quality of the image.



Opto-Acoustic reflection artifact should be directly under the tumor, but there is a 3mm offset.


Since the probe was in motion along the long axis, the delay between the ultrasound frame capture and the opto-acoustic frame capture resulted in a 3mm horizontal misalignment between the OA/US and Ultrasound.

9.1.3 Sound Velocity Errors

Both the OA and US reconstruction algorithms use a speed of sound parameter to account for the propagation of acoustic signal through the tissue. This can cause mis-registration between the US and OA due to the bidirectional nature of the US acoustic signal compared to the unidirectional nature of the OA acoustic signal. The operator is advised to adjust the speed of sound parameter on the console to obtain the best co-registration and the best (most focused) appearance of targets in the OA short and long wavelength images.

Additionally, speed of sound variations within the imaging plane can cause mis-registration that must be accounted for. In US, this can cause a speed displacement artifact. In OA, this can cause both a speed displacement artifact and a co-registration displacement. To correct for this, ensure the speed of sound for breast imaging is set to the average speed of sound through breast tissue.

9.1.4 Non-Uniformity




In the OA images (OA Short and Long, Total, Relative, and Combined), it is important to consider that the two wavelengths of light can penetrate the tissue in different ways and can be absorbed in different amounts. A key component of the operation of the device is the preferential absorption of the 757 nm wavelength by de-oxygenated blood. This absorption means that there is less light that penetrates beyond the area of high absorption. In the alternate wavelength (1064 nm), the absorption in the same area of the image may not be affected in the same degree and hence colorization accuracy may be affected beyond the strong absorbing area, e.g. it may be colored with more green than it would normally be. Conversely, high absorption in the 1064 nm wavelength can yield the opposite effect and cause redder colorization beyond a strong absorbing target. To compensate for this effect, operators are advised to scan around dominant absorbers whenever possible by taking multiple video sweeps in multiple different scan planes to avoid and minimize the non-uniformity artifact.

9.1.4.1 Out-Of-Plane Artifacts

OA imaging is not focused in the same way that Ultrasound imaging is. As such, signals may appear in images that are outside of the imaging plane. This is referred to as 'out of plane artifact'. Out of plane artifacts often appear in the image below a target lesion or vessel in the OA short and long wavelength images and result in colorization in the colorized images that does not represent the tissue immediately under the probe. This is real signal and can cause the colorization of items in plane to be affected. To correct for this, scan the target lesion from multiple directions and scanning planes to create a good understanding of the morphology of the lesion.


large vessel crossing plane at downward angle



blue dot shows
where vessel
crosses imaging
plane

9.1.5 Reflections

The OA return signal may reflect off tissue boundaries and cause the appearance of phantom targets typically in the lower part of the image. Operators are advised to scan the target lesion from multiple directions and scanning planes to create a good understanding of the morphology of the lesion.



9.2 Ultrasound Artifacts

Because the OA return signal is an ultrasound response, the same issues that affect B-mode ultrasound can affect the OA image. These issues include but are not limited to propagation and attenuation artifacts.

Propagation Artifacts:

Reverberation
Comet Tail
Ring-down
Refraction Slice
thickness
Range ambiguity
Grating lobe Mirror
Image Speed error
Speckle

Attenuation Artifacts:

Enhancement
Edge shadow
Shadowing

For future reference, the RSNA supplies training on US artifacts at <https://pubs.rsna.org/doi/10.1148/r.2017160175>


10 System Controls

10.0.1 About the Imaging Touch Screen

The imaging Touch Screen varies according to the scanning mode.

It contains:

- permanent controls available in all imaging modes
- imaging parameters that are related to the current active mode(s)
- rotary controls that can be used with their corresponding knobs, located under the Touch Screen
- Additional rotary controls may be available at the bottom of the Touch Screen by swiping with one finger.



1. System Configuration (Select Gear)
2. Workflow Menus
3. Imaging mode tabs
4. Imaging Parameters
5. Permanent Controls
6. Slidable Bar

10.1 Common Imaging Controls

Some controls are common to all imaging modes. These controls are detailed below.

For a description of controls specific to each imaging mode, refer to the detailed section on each imaging mode.

10.1.1 Freeze

When in Live imaging, press Freeze on the system console to freeze the image. A snowflake appears on the image to indicate the image is frozen. When the image is frozen, press Freeze to unfreeze the image and return to live imaging.

Note: Freeze and unfreeze the image by using the Touchpad. Press with two fingers on the Touchpad (unless configured differently in System Configuration) to freeze the image.

10.1.2 Depth

Depth controls the distance over which the B-mode displays anatomy.

When the depth is increased, echoes are captured from deeper in the body. Therefore, it takes longer for the probe to receive all the signal. The receive time is increased and there is more information to process. The time between two ultrasound beams is increased.

As a result, the frame rate is decreased.

Depth is located on the system console and adjustable only in live imaging.

Rotate Depth:

- clockwise to increase the depth and visualize deeper structures
- counterclockwise to reduce the depth and focus on shallower structures.

Note: Increase or decrease the depth by using the Touchpad. Swipe up with three fingers to decrease the depth or swipe down to increase the depth (unless configured differently in System Configuration).

10.1.3 Gain

Overall gain increases or decreases the amount of echo information displayed in an image.

It may have the effect of brightening or darkening the image if sufficient echo information is generated.

Gain is adjustable in live imaging and on a frozen image.

Gain is adjustable in all modes with each mode knob.

Rotate B, CM, SWE, COL, or PW according to the active mode:

- clockwise to increase the overall gain in order to obtain a brighter image
- counterclockwise to decrease the overall gain in order to obtain a darker image.

Note: Increase or decrease the gain by using the Touchpad. Swipe right with three fingers to increase the gain, or swipe left to decrease the gain (unless configured differently in System Configuration).

10.1.4 Time Gain Compensation (TGC)

10.1.4.1 Auto

Auto automatically adjusts the B-mode gain at different depths and for different tissue attenuations. Overall brightness is automatically adjusted depending on setting of Auto Offset, refer to “Auto Offset” on page 83.

Auto is located on the system console.

Press Auto. If needed, make independent gain adjustments, refer to Gain above.

Note: While in PW, press Auto to automatically adjust the Scale and Baseline. Refer to “Setting Up PW” on page 95.

Turning this mode on enables the automatic TGC adjustments when changing the imaging controls. To activate or deactivate it, refer to “Exam” on page 142.

10.1.4.2 ManualTouchTGC™

Selectively adjust the uniformity of brightness throughout the image.

Press TGC on the Touch Screen.

The B-mode image is simultaneously displayed behind the sliders.

Draw the TGC curve as needed with a finger.

Change the position of each virtual control on its line by touching another place on the same line.

10.1.5 Focal Zone Management

The focal zone is the depth at which the lateral resolution is the best on the image. The focal position is the depth at which the transmitted ultrasound energy is focused.

Choose the Focal Zone Management to be Manual or Automatic.

Control this setting in the Imagio® Breast Imaging System/Display sub-tab Imagio® Breast Imaging System.

10.1.5.1 Manual Focal Zone Management

By selecting manual, the Focal zone and position are adjustable.

Focus is adjustable only on live imaging.

Focus is located on the system console.

Rotate Focus:

- clockwise to place the focal zone deeper
- counterclockwise to place the focal zone at a shallower depth

Press Focus and then rotate it:

- clockwise to increase the size of the focal zone
- counterclockwise to decrease the size of the focal zone

Press Focus again to move the focal zone.

- Changing the size of the Focal Zone affects the frame rate.
- The larger the Focal Zone, the slower the frame rate.

10.1.5.2 Automatic Focal Zone

By using the automatic option, the Focal Zone will be automatically managed to provide the most appropriated focal zone size and position. The ratio between the Focal Zone size and depth will be maintained and the Focal Zone position will be adjusted within any depth change.

10.1.6 Zoom

A region of interest in imaging modes can be magnified for closer examination. Two types of zoom are available on the Imagio® Breast Imaging System: HD Zoom and Digital Zoom*.

** These are developmental features for use in modes other than B. Do not change the scaling of the image during the prospective clip capture, especially the following settings when not in B-Mode:*

- digital zoom
- display format
- dual top/bottom
- wide image mode

10.1.6.1 HD Zoom

HD Zoom concentrates the Imagio® Breast Imaging System acquisition and processing ability on a specific region of interest in the real-time image display.

The result is a reduced field of view overall, but greater detail and higher frame rate, which contribute to better

image quality over the region of interest.

HD Zoom is adjustable only in live imaging.

Step: 1 Press Zoom on the system console.

- A box appears on the image.

Step: 2 Use the Touchpad to move the zoom box location.

Step: 3 Press on the Touchpad to alternate box location to size.

Step: 4 Use the Touchpad to resize the zoom box width and height.

Step: 5 Press Zoom again to have an enlarged image

Step: 6 Press Zoom again to exit zoom

10.1.6.2 Digital Zoom

Digital Zoom is a magnifying post-processing function. As such, it has no effect on the frame rate. Digital Zoom is adjustable in live imaging and on a frozen image.

Rotate Zoom to change its value.

Note: In live imaging, both HD Zoom and Digital zoom may be used at the same time.

Press Reset Zoom on the Touch Screen to set the zoom back to the preset default value.

Note: Increase or decrease the zoom by using the Touchpad. Do a spread gesture with two fingers to increase the zoom, or pinch to decrease the zoom (unless configured differently in System Configuration).

10.1.7 Left/Right

Left/Right flips the image along the left-right axis.

Touch Left/Right to flip the image.

10.1.8 Top/Bottom

Top/Bottom flips the image along the up-down axis.

Touch Top/Bottom to flip the image.

10.1.9 Dual

Touch Dual to activate or deactivate dual mode. The left image will be active.

Press the Touchpad to freeze the left side and make the right side active.

Press Freeze to have both sides frozen.

- the active image has an orange pink S as the orientation marker
- the inactive image has a white S as the orientation marker

Switch which of the dual images is active by pressing the Touchpad. Imaging settings will always be applied to the active image in dual. Imaging settings will be maintained independently for the dual images.

Pressing freeze before pressing the Touchpad, changes image settings and modes on the active side.

10.1.10 Dual Top/Bottom

Dual Top/Bottom is the same feature as Dual Top/Bottom but split the screen into a top image and a bottom image.

- Press Dual to enable Dual Top/Bottom.
- Press Dual Top/Bottom to activate or deactivate the Dual Top/Bottom display format.

10.1.11 Hide Patient Name

Hide Patient Name temporarily hides the patient information from the imaging screen (screen displays the star character, "*" in place of patient information). However, the hidden information remains in the database.

10.1.12 Arbitration (Update)

If several functions are available, they are displayed at the bottom of the image. The active function is displayed in pink.

Swipe left or right with 4 fingers the active Touchpad function.

Press the Touchpad to arbitrate/change between upper and lower functions. Examples of functions that are available via arbitration are Box Move (position of the box) and Box size (resize the box) or switching between Img1 and Img2 in dual displays.

10.1.13 Play Clip

When the image is frozen, in all modes, a Play Clip button appears on the Touch Screen.

Play Clip enables viewing clips. Refer to Image Capture and Data Management for more information on viewing clips.

11 Ultrasound Imaging Modes

11.0.1 Entering/Exiting Modes

The Imagio® Breast Imaging System offers a set of imaging modes via the mode selection buttons. On the system console, the mode buttons are located above the Touchpad.



- Press OA to scan in OA/US-mode
- Press B to scan in gray-scale B-mode.
- Press SWE™ to scan in gray-scale B-mode with a real-time color elasticity map superimposed.
- Press COL to use Color Flow Imaging (CFI), Color Power Imaging (CPI), or Directional Color Power imaging (dCPI)
- Press PW to use Pulsed Wave Doppler

Note: In any mode other than B-mode, whenever the B key is pressed, the current mode is exited, and B-mode is displayed. The previous settings are restored.

11.1 Common Imaging Parameters

The following imaging parameters are available in multiple modes.

11.1.1 All Modes Common Parameters

The following imaging parameters are available in all modes unless otherwise specified.

Parameter	Description	Usage
Acoustic Power	Acoustic Power regulates the output power of the Imagio® Breast Imaging System. Refer to Acoustic Output Tables for acoustic power guidance.	Acoustic Power is available only in live imaging. Rotate the knob located under Acoustic Power: • clockwise to increase acoustic power output • counterclockwise to decrease acoustic power output

Dynamic Range	Dynamic Range enables changing the range over which the amplitudes of returning ultrasound signals are displayed. Dynamic Range is available in live imaging and on a frozen image.	Rotate the rotary knob located under Dynamic Range: • clockwise to increase the dynamic range • counterclockwise to decrease the dynamic range
Optimization	Optimization enables changing of the center frequency to obtain increased resolution or increased penetration. Optimization is available in live imaging in all modes except PW.	Rotate the knob located under Optimization to change value: • Resolution • General • Penetration In SWE™, both B and SWE optimization settings are available. The SWE Optimization setting has the following values (see SWE™ Scanning TIPS for more information): • Res • Std • Pen
Persistence	Persistence is a frame averaging technique that reduces noise on the image. Persistence is available in live imaging for all modes except PW.	Rotate the rotary knob located under Persistence: • clockwise to increase persistence • counterclockwise to decrease persistence
Sector Size	Sector Size widens or narrows the size of the sector angle to maximize the image's region of interest. Sector size is available in B and SWE modes.	Changing the sector size affects the frame rate. The narrower the sector size, the faster the frame rate.
Smoothing	Smoothing is a spatial filter used to smooth the image in order to provide a more homogeneous appearance. Smoothing is available only in live imaging. Available in SWE, COL, and Panoramic modes.	Rotate the knob located under Smoothing: • clockwise to increase smoothing • counterclockwise to decrease smoothing

11.1.2 Doppler Mode Common Parameters (COL and PW)

The following imaging parameters are available in all Doppler modes unless otherwise specified.

Parameter	Description	Usage
Baseline	Baseline is a horizontal graphical line that represents the level of zero velocity. By changing the baseline position, the maximum positive velocity is increased, and the maximum negative velocity is decreased, or vice versa. In COL, the Baseline control is available only in CFI mode. In CPI, it is replaced by Dynamic Range. Dynamic Range in COL enables changing of the range over which the amplitude of returning ultrasound signals are displayed.	Rotate the knob located under Baseline: • clockwise to increase the maximum positive velocities • counterclockwise to increase the maximum negative velocities
Invert	In COL, invert enables viewing of blood flow from an inverse perspective, e.g. red away (negative velocities) and blue toward (positive velocities) the probe (available in CFI mode). In PW, invert vertically flips the spectral display without affecting the baseline position. Typically, positive velocities, (velocities toward the probe) are shown above the baseline while negative velocities, (velocities away from the probe) are shown below the baseline.	Touch Invert to activate and deactivate.
Quick Steer	Quick Steer enables quick steering of the insonifying beam.	Press Quick Steer to change its value. Use the Steering knob for finer settings.
Steering	The insonifying beam can be steered in a direction as close as possible to the expected direction of motion, within limitations imposed by the probe and the image plane geometry. Steering is available only in live imaging. To maximize the sensitivity and accuracy of color Doppler modes, it is usually recommended that clinical Doppler examinations be conducted with a Doppler angle equal to or less than 60 degrees between the ultrasound beam and the vessel direction.	Rotate the knob located under Steering to steer the Doppler line.
Wall Filter	This parameter filters out low velocity signals. It helps to decrease motion artifacts caused by patient or probe motion. Wall Filter is available only in live imaging.	Touch Wall Filter repeatedly to select the desired value.

11.2 B-mode Imaging


B-Mode is an ultrasound imaging mode which enables assessment of two-dimensional anatomy by displaying the echoes returning from tissues of different density in various shades of gray.

11.2.1 Accessing B-mode

B-mode is available on the OA/US L18-1 and the US L18-5 linear probes.

11.2.2 About B-Mode Main Display

Press B on the system console from any other modes.




- | | |
|---------------------|-----------------------|
| 1. Institution Name | 2. Date And Time |
| 3. Patient Name | 4. Patient Id |
| 5. Probe/Preset | 6. Imaging Parameters |
| 7. Ultrasound Image | 8. Depth Scale |
| 9. Focal Zone | 10. Gray-Scale |

11.2.3 About B-Mode Touch Screen

When in B-mode, the Touch Screen displays various image controls which optimizes the B-mode image.

These settings are controlled by touching the control on the Touch Screen, or by turning the rotary knob directly below the Touch Screen.



Additional rotary controls are available at the bottom of the Touch Screen by swiping left or right with one finger.

11.2.4 Description of B-mode Imaging Parameters

B-mode imaging parameters not already described in the Common Imaging Parameters section are listed below.

Parameter	Description	Usage
Harmonic Imaging	Harmonic Imaging available on the OA/US L18-1 and the US L18-5 linear probes. Harmonic imaging mode is used to reduce image clutter and reverberation and enhance borders. Harmonic imaging is available only in live imaging.	Press Harmonic Imaging to activate/deactivate harmonic imaging.
TissueTuner™	TissueTuner™ is available on the OA/US L18-1 and the US L18-5 linear probes. TissueTuner™ enables adjustment of the receive parameters associated with the assumed speed of sound of ultrasound in the body. Adjusting the speed of sound parameter to match the type of tissue being interrogated results in increased spatial and lateral resolution. TissueTuner™ is available only in live imaging. TissueTuner™ range is set by the selected application.	Rotate the knob located under TissueTuner™: - clockwise for a denser tissue (higher speed of sound values) - counterclockwise for a less dense tissue (lower speed of sound values)

2D Map	2D Map is available on the OA/US L18-1 and the US L18-5 linear probes. 2D Map assigns the display of echo amplitudes to a range of gray-scale or chromatic colors. 2D Map is available in live imaging and on a frozen image.	Rotate the rotary knob located under 2D Map: - clockwise to view next map - counterclockwise to view previous map
SuperCompound™	SuperCompound™ is available on the OA/US L18-1 and the US L18-5 linear probes. SuperCompound™ combines the returning ultrasound from a large number of steered echo lines to create an image which provides superior texture, enhanced edge delineation and reduces shadowing. SuperCompound™ is available in live imaging only.	Press SuperCompound™ to activate or deactivate the SuperCompound™ mode.
Pulse Repetition Frequency (PRF)	PRF is available on the OA/US L18-1 and the US L18-5 linear probes. PRF enables changing the pulse repetition frequency for the B-mode image. Decreasing the PRF may correct the reverberation artifacts.	Rotate the knob located under PRF to change its value.
Auto Offset	Auto offset enables programming automatic overall gain when using the Auto. Auto Offset is located on the second page of parameters.	Rotate the knob located under Auto offset: - clockwise for a brighter image - counterclockwise for a darker image
SuperRes™	SuperRes™ is an image processing feature which reduces speckle, improves image texture, and enhances borders without an impact on frame rate. SuperRes™ is available in live imaging and on a frozen image. SuperRes™ values depend on whether SuperCompound™ is On or Off.	Rotate the rotary knob located under SuperRes™: - clockwise to increase SuperRes™ - counterclockwise to decrease SuperRes™

11.2.5 B-Mode Scanning

Use the appropriate Preset for each organ being scanned by the OA/US L18-1 and the US L18-5 linear probes.

The OA/US L18-1 probe is only indicated for breast imaging.

Set depth to achieve desired field of view.

Place focal zone at the depth of the area of interest or slightly below.

Press Auto TGC as the tissue profile changes to optimize spatial brightness uniformity of OA/US image.

Press the touch screen TGC button then press or slide your finger on the touchscreen TGC curve to increase or decrease brightness at specific depths.

Adjust overall gain to obtain the appropriate brightness.

Adjust TissueTuner™ to achieve the best resolution: 1480 m/s for breast imaging or decrease to 1420 m/s for fatty tissue.

- Increase focal zone region for a larger focal area and increased resolution.

Use Harmonic Imaging to clear fluid-filled structures.

Use Resolution setting of Optimization control for increased image resolution

11.2.6 Panoramic Imaging

Panoramic imaging registers individual frames as the probe is moved along a structure of interest and superimposes the registered frames to generate a composite image, allowing the visualization and measurement of large structures of interest that would not fit within the probe's field of view.

Important: *Use adequate quantities of gel along the entire segment to be scanned to ensure smooth motion during the panoramic acquisition.*

Panoramic imaging is available for the US L18-5 Linear Probe.

11.2.6.1 Acquiring panoramic images

Step: 1 Adjust the available imaging parameters to optimize image quality and frame rate.

Step: 2 Press Panoramic on the Touch Screen.

- The Imagio® Breast Imaging System enters the Panoramic Stand-by stage.

Step: 3 Press on the Touchpad to start the acquisition.

Step: 4 Move the probe in a smooth and continuous manner along the structure of interest, making sure that the scanning motion is as much as possible parallel to the scan plane.

- The Imagio® Breast Imaging System automatically detects the probe's motion direction and the composite image is created accordingly. If needed, scan backwards to erase unwanted parts of the panoramic composite image, then resume scanning along the original motion direction.

Step: 5 Press Freeze.

- The Imagio® Breast Imaging System displays the whole composite image.

11.2.6.2 Scanning Recommendations

- Ensure that the scanned surface is flat or gently curved and avoid tissue areas dominated by noise or lacking structural details.
- Use a slow and steady scanning motion without wobbling, twisting or abrupt changes in the scanning direction, staying within the same scan plane as much as possible.

Important: *Panoramic registration errors are manifested by the presence of artifacts such as image gaps, irregular skin line contours and jagged edges, clearly visible "seams" between successive input image frames and blurred or poorly defined areas. If such artifacts are encountered, it is strongly recommended to discard the current Panoramic composite image and repeat the Panoramic acquisition.*

Note: All measurements derived from a Panoramic composite image are marked by an asterisk, to indicate that they represent estimated measurement results and may have been affected by Panoramic registration errors.

Note: When the structure to be measured can fit within the probe's field of view, it is strongly recommended to use a standard B-mode image frame and not the Panoramic composite image.

11.2.6.3 Manipulating panoramic images

Use the Zoom functionality as for regular B-mode. Use the Pan functionality as for regular B-mode.

Use the Rotate button or the Touchpad to rotate the panoramic composite image.

Use the Smoothing knob to soften abrupt transitions between input frames in the Panoramic composite image.

11.1.6.3.1 Trim

Trim enables discarding of frames from the start or end of the panoramic acquisition, so that they don't contribute to the panoramic composite image.

Step: 1 Press Trim on the Touch Screen to start trimming

Step: 2 The Touchpad is changed to Trim Start.

Step: 3 Move the Touchpad towards the end of the panoramic acquisition to discard the required frames from the start of the acquisition.

Step: 4 Press the Touchpad to switch to Trim End.

Step: 5 Move the Touchpad towards the start of the panoramic acquisition to discard the required frames from the end of the acquisition.

Step: 6 Turn Trim off to cut the trimmed frames out.

The remaining panoramic composite image will be automatically zoomed to fit on the screen.

Press Trim Reset to return the Trim Start and Trim End frames to the start and end of the panoramic acquisition, respectively.

11.1.6.3.2 Skinline Tick-Marks

Skinline Tick-Marks enables displaying or hiding of the tick marks along the skinline.

Press Skinline Tick-Marks to activate/deactivate the skinline tick marks.

11.2.6.4 Performing measurements on panoramic images

Press Meas. on the system console to access measurements available in panoramic imaging.

Important: *Do not perform measurements on Panoramic composite images acquired with substantial out-of-plane motion components (such as when following a tortuous structure).*

Important: *Do not perform measurements across structures which appear as strongly hypoechoic or very noisy in the Panoramic composite image.*

Important: *Do not perform measurements on Panoramic composite images of highly curved objects such as transverse sections through the arm or leg.*

Basic measurements (distance, ellipse, trace, ...) on a Panoramic composite image are performed as in standard B-mode.

11.1.6.4.1 Curved Distance Measurements

The Curved Distance measurement tool is similar to the Trace measurement tool, but the curved distance trace is not closed. The Curved Distance measurement tool provides a distance measurement result.

To perform a Curved distance measurement:

Step: 1 Acquire the panoramic image

Step: 2 Press Meas. on the system console

Step: 3 Press Curved Distance on the Touch Screen

- A first caliper is displayed.

Step: 4 Move the first caliper at the desired location

Step: 5 Press the Touchpad to anchor the first caliper

Step: 6 Move the Touchpad to start tracing the curve

Step: 7 Use the Trace knob to erase the trace if needed

Step: 8 Press the Touchpad to complete the measurement

- The Imagio® Breast Imaging System displays the curved distance measurement result in the measurement result area.

11.2.6.5 Saving panoramic images

Press Save Image to store the current Panoramic composite image so that it can be reviewed, manipulated, trimmed, and measured at a later time.

11.2.6.6 Exiting panoramic imaging

Press the Panoramic button on the Touch Screen to turn it off and exit panoramic imaging.

11.3 ShearWave™ Elastography

The ShearWave™ Elastography (SWE™) mode displays information about tissue elasticity in the form of an easy to interpret color-coded image. ShearWave™ Elastography (SWE™) is available on both the OA/US L18-1 and US L18-5 linear probes, however SWE™ is not available while in OA Mode.


The elasticity information is also quantitative. Local estimation of the tissue stiffness is displayed per pixel and can be expressed in kPa or in m/s over a wide range of values.

Note: Please note that the SWE™ algorithms and sequences are optimized accordingly for each clinical application need.

11.3.1 Accessing ShearWave™ Elastography

Press SWE™ on the system console from any other mode.

11.3.2 About the SWE™ Main Display



1. Exam and patient information
2. Probe/Preset/ MI/TI
3. Ultrasound image
4. Elasticity map
5. Clip buffer
6. Gray-scale and elasticity bars

11.3.3 About SWE™ Touch Screen

When in SWE™, the Touch Screen displays various image settings which help optimize the elasticity image. These settings are controlled by touching the control on the Touch Screen or turning the rotary knob directly below it.



Additional rotary controls are available at the bottom of the Touch Screen by swiping with one finger.

11.3.4 Description of SWE™ Parameters

SWE™ imaging parameters not already described in the Common Imaging Parameters section are listed below.

Parameter	Description	Usage
SWE Unit	This control enables changing of the color bar scale for SWE™ from kPa to m/s. SWE Unit is available in live mode and on a frozen image.	Touch SWE Unit to change value.

Parameter	Description	Usage
Elasticity Range	The Elasticity Range in SWE™ mode is similar to the dynamic range in B-mode. It changes the maximum elastic value displayed. It is used to compress the colored image for better visualization of tissues which have a variety of stiffnesses. Changing the Elasticity Range does not affect the actual values of elasticity, as long as they are within the elasticity range. Elasticity Range is displayed in kPa; the values depend on the selected application. Stiff tissue should appear yellow, orange or red. Increase the Elasticity Range if there is predominantly dark red in the color box. Elasticity Range is available in live imaging or on a frozen image.	Rotate the knob located under Elasticity Range: - clockwise to increase the elasticity range - counterclockwise to decrease the elasticity range
Velocity Range	The Velocity Range works in the same manner as the Elasticity Range. Velocity Range appears and replaces Elasticity Range when the Display Unit is set to m/s. It changes the maximum velocity value displayed. It is used to compress the colored image for better visualization of tissues which have a variety of stiffnesses. Changing the Velocity Range does not affect the actual values of velocity, as long as they are within the velocity range.	Rotate the knob located under Velocity Range: - clockwise to increase the velocity range - counterclockwise to decrease the velocity range
Display Format	This control changes the format in which B-mode and SWE™ images are presented on the main SWE™ display. Display Format is available in live imaging and on a frozen image.	Press Display Format repeatedly until the desired format is shown. The top/bottom format is the default setting. Three formats are available: - top/bottom - side by side - single
Opacity	Opacity enables changing of the transparency of the elasticity/viscosity map over the B-mode image. Opacity is available in live imaging and on a frozen image.	Rotate the knob located under Opacity: - clockwise to increase the prominence of the color image over the B-mode image - counterclockwise to decrease the prominence of the color image over the B-mode image

Parameter	Description	Usage
Elasticity Map	Elasticity Map determines how elasticity values are displayed in terms of color graduations. Elasticity Map is available in live imaging and on a frozen image.	Rotate the knob located under Elasticity Map: - clockwise to go to the next map - counterclockwise to go to the previous map

11.3.5 SWE™ Scanning TIPS

11.3.5.1 Notes and Recommendations

As is the case with all other ultrasound imaging modes - the SWE™ mode is associated with a significant learning curve and may exhibit operator dependence if suboptimal scanning techniques are employed.

It is recommended to carefully review [“Elastography measurements” on page 184](#) in order to become aware of the SWE™ capabilities and limitations with regards to penetration depth, spatial resolution, estimation bias and estimation precision.

When SWE™ is used to document the stiffness of a lesion, be aware that the bias and precision of shearwave velocity estimates vary as a function of lesion size and probe resolution.

The manual pressure used while scanning in SWE™ should be minimal. Please keep in mind that compression with the probe may change the velocity of shear waves in the tissue being imaged.

Recommended Optimization Settings:

Optimization	Rotary Knob Setting	Use Case
Resolution	Res	Use this setting when imaging small shallow areas of interest. This setting will also help to clear any erroneous visco-elastography signal out of areas which are suspected to be comprised of fluid.
Standard	Std	This setting is balanced between resolution and penetration. Use this setting as the default starting point for the stiffness and viscosity evaluation of an object of interest.
Penetration	Pen	Use this setting when imaging deeper or larger areas (anechoic or hypoechoic) which may show posterior dropout in the echo image. These areas are suspected to be stiff and viscous and require a penetration mode optimized for high velocity shear waves.

11.3.5.2 SWE™ Scanning

Remember the 3 S's:

Scan Softly, Smoothly and Slowly

Scan Softly: no manual compression is needed (with the exception of intercostal liver scanning), just alight touch and plenty of gel between skin and probe

Scan Smoothly: probe automatically induces vibrations in tissues so no need to shake or push on the probe

Scan Slowly: Adapt your speed of scanning to the frame rate. Once imaging the area of interest, avoid any movements and wait for image stabilization

Always start with the default setting of Standard for Optimization, and then go to Resolution or Penetration to

fine tune the image.

Controls that can be used in live or frozen mode such as Elasticity Range, display format, etc. should be done after freezing to make it easier for the operator since it is imperative to hold a steady hand during elastography.

11.4 Color Mode Imaging

Color imaging modes use Doppler principles to generate a color image of the mean flow velocity or flow power. On the Imagio® Breast Imaging System, three color modes are available: Color Flow Imaging (CFI), ColorPower Imaging (CPI), and Directional Color Power Imaging (dCPI).

11.4.1 Color Flow imaging (CFI)

Color Flow Imaging is a Doppler mode intended to add color-coded quantitative information concerning the relative velocity and direction of fluid motion within the B-mode image.

11.4.2 Color Power Imaging (CPI)

Color Power Imaging is a color flow mapping technique used to map the strength of the Doppler signal coming from the flow rather than the frequency shift of the signal. Using this technique, the ultrasound Imagio® Breast Imaging System displays the magnitude of the flow based on the number of reflectors that are moving, regardless of their velocity. CPI does not map flow velocity.

11.4.3 Directional Color Power Imaging (dCPI)


Directional Color Power Imaging (dCPI) displays the mean Doppler frequency of each pixel inside the Color box, including the direction of flow relative to the probe.

The color maps used in dCPI are separated into parts which typically contain red and blue hues, respectively, with positive Doppler frequencies (flow towards the probe) displayed using the top part of the dCPI color map, and negative Doppler frequencies (flow away from the probe) displayed using the bottom part of the dCPI color map. dCPI is useful in indicating the flow direction within the vessels of interest, as well as identifying areas of high flow velocities (aliasing), flow reversals, etc.

11.4.4 Accessing Color Modes

Press COL on the system console from any other modes.


11.4.5 About Color Modes Main Display



1. Focal Zone
2. Gray-scale
3. Color bar
4. Color box

11.4.6 About Color Modes Touch Screen

When in color modes, the Touch Screen displays various image settings which helps optimize the color image. These settings are controlled by touching the control on the Touch Screen or turning the rotary knob directly below it.



Additional rotary controls are available at the bottom of the Touch Screen by swiping with one finger.

11.4.7 Description of Color Imaging Parameters

Color imaging and frozen mode:

- clockwise to increase the gain
- counterclockwise to decrease the gain

11.4.7.2 Color Mode

Color Mode enables switching between CFI, dCPI and CPI. Color Mode is available only in live imaging. Touch Color Mode to select the desired mode.

11.4.7.3 Dual

Refer to Dual [“Dual” on page 76.](#)

11.4.7.4 Optimization

Optimization enables changing of the center frequency to obtain increased resolution or increased penetration. Optimization is available only in live imaging. Rotate the knob located under Optimization to change value:

- Resolution
- General
- Penetration

11.4.7.5 Boost

Boost enables adjustment of the frame rate and line density to achieve increased spatial or temporal resolution.

There are three options to choose from:

- High Definition
- Medium
- High Frame Rate

Rotate the knob located under Boost to change its value.

11.4.7.6 Invert

Invert enables viewing of blood flow from an inverse perspective, e.g. red away (negative velocities) and blue toward (positive velocities) the probe.

Invert is available in live imaging and on a frozen image. Invert is available only in CFI mode.

Touch Invert to activate and deactivate it.

11.4.7.7 Scale

Scale enables adjustment of the maximum displayed velocity. Scale is available only in live imaging.

Rotate the knob located under Scale:

- clockwise to increase the scale
- counterclockwise to decrease the scale

11.4.7.8 Hide Color

Hide Color hides the color information over the B-mode image. Hide Color is available in live imaging and on a frozen image.

Touch Hide Color to activate and deactivate it.

11.4.7.9 Wall Filter

This parameter filters out low velocity signals.

It helps to decrease motion artifacts caused by patient or probe motion. Wall Filter is available only in live

imaging.

Touch Wall Filter repeatedly to select the desired value.

11.4.7.10 Velocity Optimization

Velocity Opt. enables quick optimization of multiple parameters. Common Imaging Parameters section are listed below.

Parameter	Description	Usage
Color Gain	Changes the gain of the Color Mode Imaging overlay.	Rotate the COL button in order to change the gain both in live imaging and frozen mode: clockwise to increase the gain counterclockwise to decrease the gain
Modes	The Modes rotary control enables switching between CFI, dCPI and CPI. Available only in live imaging.	Rotate the knob located under Mode to select the desired Color Imaging mode.
Boost	Boost enables adjustment of the frame rate and line density to achieve increased spatial or temporal resolution.	Rotate the knob located under Boost to change its value. - High Definition - Medium - High Frame Rate
Scale	Scale enables adjustment of the maximum displayed velocity. Scale is available only in live imaging.	Rotate the knob located under Scale: - clockwise to increase the scale - counterclockwise to decrease the scale
Hide Color	Hide Color hides the color information over the B-mode image. Hide Color is available in live imaging and on a frozen image.	Touch Hide Color to activate and deactivate it.
Velocity Flow Optimization (Velocity Flow Opt.)	Velocity Flow Opt. enables quick optimization of multiple parameters with one control to achieve a desired flow optimization. Velocity Flow Opt. is available only in live imaging. Velocity Flow Opt. changes the following parameters: - Scale - Wall filter - Resolution/frame rate - Persistence	Touch Velocity Flow Opt. repeatedly to select the desired value: - Off - Low Flow - Med Flow - High Flow
Side by Side	Side by Side enables splitting of the screen in two live parts: - the left side displays the B-mode only, for reference - the right side displays the B-mode and the color image	Touch Side by Side to change the screen display.
Color Priority	Color Priority enables selecting a level beyond which color data are not displayed over the gray map. Color Priority is available in live imaging and on a frozen image. The range and default value depend on the Color Imaging mode.	Rotate the knob located under Color Priority: - clockwise to increase the color priority - counterclockwise to decrease color priority

Parameter	Description	Usage
Color Map	Color Map determines how flow values are displayed in terms of color graduations. Color Map is available in live imaging and on a frozen image.	Rotate the knob located under Color Map: - clockwise to go to next map - counterclockwise to go to previous map
Blending	This button controls the level of transparency of the color image that is superimposed on the B-mode image. When Blending is on, the color image is smoothly overlaid on the gray-scale image. When Blending is Off, the transition between B-mode and color is sharper. Blending is available in live imaging and on a frozen image.	Press Blending to turn it On or Off.
Flash Suppression	Flash Suppression utilizes an adaptive wall filtering scheme to suppress the flash artifacts due to tissue motion. When Flash Suppression is On, the frame rate is reduced. Flash Suppression is available only in live imaging.	Press Flash Suppression to turn it On or Off.

11.4.8 COL Scanning

Scan to obtain best angle to maximize flow sensitivity.

Use the Touchpad to change size of color box if needed. A larger box will cause a decrease in temporal resolution (frame rate).

Increase Color Gain until noise is visible and then reduce the gain just to eliminate noise. This will ensure the best color flow sensitivity possible.

If scanning smaller vessels (i.e. breast, thyroid), center steer is typically the best choice. In larger vessels (i.e. carotid artery, axillary artery) steer color box to obtain the best (least) angle to flow. A change in sensitivity will be observed when the appropriate steering angle is chosen.

Select the appropriate Velocity Optimization level according to the type of vessel being interrogated.

Low is typically used for smaller vessels with lower velocities and venous flows.

High is used for larger vessels with higher velocities and for arteries.

To increase flow sensitivity, lower the Scale or lower the Wall Filter. To eliminate aliasing (CFI only), increase the Scale.

11.5 Pulsed Wave Doppler Mode

Pulsed Wave Doppler (PW) is a Doppler mode that measures velocity of blood flow within a small region called the Doppler sample volume.

This information is presented as a sweeping display of velocity (or Doppler frequency) over time. This display is often called a spectral Doppler trace, or Doppler trace. The information may also be heard as a stereo audio output.

In addition to the Doppler velocity display, an image (gray-scale only or gray-scale and color) is also presented for proper positioning of the Doppler sample volume.

11.5.1 Accessing PW

Press PW on the system console from any other modes.

Swipe left and right on the Touchpad to move the Doppler line in the 2D image.

11.5.4.2 Sample Volume Position

Swipe up and down on the Touchpad to move the sample volume on the Doppler line.

11.5.4.3 Sample Volume Size

Slide your finger on the Touchpad to change the size of the sample volume.

11.5.5 Description of PW Imaging Parameters

PW imaging parameters are listed below.

Parameter	Description	Usage
Doppler Coarse Angle	Provides a coarse angle control for rapid PW angle adjustment. Doppler Coarse Angle is available in live imaging and on a frozen image.	Touch Doppler Coarse Angle repeatedly to select the desired value.
Fine Angle Correct	Provides a fine angle control for fine PW angle adjustment. Fine Angle Correct is available in live imaging and on a frozen image.	Rotate the knob located under Fine Angle Correct.
Display Format	Changes the layout of the PW display. Display Format is available in live imaging and on a frozen image.	Press Display Format to change its value. • 2D: 2D only • Side by side: B-mode and spectral side by side display • ½-1/2: ½ B-mode and ½ spectral display • 2/3-1/3: 2/3 B-mode and 1/3 spectral display • 1/3-2/3: 1/3 B-mode and 2/3 spectral display
PW Map	Changes the PW Map applied to the trace. PW Map is available in live imaging and on a frozen image.	Rotate the knob located under PW Map: • clockwise to go to previous map • counterclockwise to go to next map
Scale	Scale enables adjustment of the maximum velocities displayed on the Doppler spectrum. Scale is available only in live imaging.	Rotate the knob located under Scale: • clockwise to increase the scale • counterclockwise to decrease the scale
Sweep Speed	Sweep Speed enables changing the speed that the spectral columns are updated on the display. Sweep Speed is available in live imaging and on a frozen image.	Touch Sweep Speed repeatedly to reach the desired value.
Simultaneous (Duplex and Triplex)	Simultaneous allows live display of combinations of PW Doppler trace, Color Doppler, and B-mode imaging. Duplex imaging is active if Simultaneous is enabled during PW scanning (B-mode and PW trace are live). Triplex imaging is active if Simultaneous is enabled during PW scanning with Color Imaging (B-mode, Color Imaging, and PW trace are live). Simultaneous is available only in live imaging.	Touch Simultaneous to turn it On or Off.

Parameter	Description	Usage
High PRF	High PRF enables measuring flows of higher velocities than those that can be measured with the maximum conventional available PRF, by increasing the PRF above the Nyquist limit and avoiding aliasing. The maximum scale available is increased as a function of depth. Available in live imaging.	Touch High PRF to turn it on. See High PRF Scanning Tips for more information.

11.5.5.15.1 High PRF Scanning Tips

High PRF extends the PW Doppler velocity scale beyond the maximum velocity limit that can be measured for a given sample volume (SV) in a non-ambiguous manner. Beyond this limit, the Pulse Repetition Frequency (PRF) of the transmitted PW Doppler pulses corresponds to a Pulse Repetition Interval (PRI) which is shorter than the round-trip propagation time from the probe to the current sample volume depth, resulting in range ambiguity (the analyzed Doppler signals emanate not only from the primary sample volume of interest but also from one or two secondary sample volumes along the Doppler beam).

- Position the primary SV in the area of interest, in a manner similar to conventional PW Doppler.
- Avoid, as much as possible, placing secondary SV's in areas dominated by noise artifacts, shadowing, or areas of flow.
- Tips to assist in positioning secondary High PRF SV's to optimize the signal and avoid interference with the primary SV:
 - Vary the velocity range to change the secondary sample volume position.
 - Vary the depth of the image to change the secondary sample volume position.
 - Reduce the velocity range to remove the secondary sample volume.
 - Reduce gain.
 - Vary 2D image orientation.
 - Change the acoustic window.

11.5.5.1 PW Doppler Auto Trace Parameters

- The parameters for PW Auto are listed below.

Parameter	Description	Usage
Doppler Auto Trace	This control enables automatic PW trace measurements. Doppler Auto Trace is available in live imaging and on a frozen image. Doppler Auto Trace enables access to the additional settings in this table.	Press Doppler Auto Trace to display the envelope of the peak velocity of the flow as a function of time.
Mean Trace	Displays the mean flow velocity as a function of time.	Press Mean Trace to toggle on or off.
Velocity Points	Displays the PSV and EDV on the spectrum.	Press Velocity Points to toggle on or off.
Trace Detection	Enables detection of the various velocity traces.	Press Trace Detection to change its value. • Above: the velocities above the baseline • Below: the velocities below the baseline • Full: the velocities both above and below the baseline

Parameter	Description	Usage
Sensitivity	Defines the detection threshold for the Doppler Auto Trace.	Press Sensitivity to change its value. • Low • Medium • High • Maximum

11.5.6 PW Scanning Tips

As in color Doppler, optimize the scan plane for the best angle to flow.

Activate the PW cursor by pressing PW. Optimize the steering angle and sample volume size before initiating the PW spectral trace. If Color is activated before pressing PW, then the cursor will be at the same steer angle as the color box.

Optimize Doppler gain. Decrease gain to eliminate background noise. Increase scale or shift baseline to eliminate aliasing.


Shift baseline, change display format, PW map and sweep speed are available after freezing.

Use simultaneous mode when vessels are difficult to follow due to respiration, motion, or small vessel size.

12 Operator Activities

12.1 Adding Distilled Water to the Laser Coolant Reservoir

The Imagio® Breast Imaging System has a laser coolant reservoir containing distilled water used to cool the lasers. **Operators add distilled water to the laser coolant reservoir as needed.** Additionally, the “Add Coolant” message appears on the bottom right of the display to alert the Operator of the need to add distilled water to the laser coolant reservoir.



Note: Seno Service will top off distilled water during preventative maintenance.

12.2.1 Distilled Water

Distilled Water Specifications

Conductivity $\leq 5 \mu\text{S}/\text{cm}$ (less than or equal to 2 micro Siemens per centimeter)



Caution: *The Imagio® Breast Imaging System is a sensitive electronic instrument and may be damaged by using the incorrect fluid and the incorrect or improper laser coolant reservoir filling technique.*

- Do not place the refill bottle or overflow bottle on any surface of the Imagio® Breast Imaging System.
- Using anything other than distilled water may result in microbial contamination of the cooling system. Use only distilled water when filling the laser coolant reservoir.

12.2.2 To Add Distilled Water to the Laser Coolant Reservoir

Step: 1 On the rear of the Imagio® Breast Imaging System cabinet, connect the Overflow Bottle spout to the distilled water outlet. Push the connector onto the outlet until an audible click is heard which indicates the connector is locked to the outlet.



Step: 2 On the rear of the Imagio® Breast Imaging System cabinet, connect the Refill Bottle spout to the distilled water inlet. Push the connector onto the inlet until an audible click is heard which indicates the connector is locked to the inlet.



Add water, replace cap, and ensure securely tightened

- Step: 3** Squeeze the bottle to feed distilled water into the reservoir until water is visually seen flowing into the Overflow Bottle.
- Step: 4** Remove the Refill Bottle by pressing the silver buttons on the refill connector hose and gently pulling the hose away from the outlet.
- Step: 5** Remove the Overflow Bottle by pressing the silver buttons on the overflow connector hose and gently pulling the hose away from the outlet.


Important: *If required to record the Imagio® Breast Imaging System usage by your local laser authority, record “Distilled Water Added” and the date in the Imagio® Breast Imaging System Log with details such as (e.g. regular maintenance.)*

12.2 Laser Energy Calibration

Calibrating the Imagio® Breast Imaging System lasers maximizes image quality. Laser Energy Calibration is to be performed whenever the Imagio® Breast Imaging System displays a message to the Operator that a laser energy calibration is due. (The Imagio® Breast Imaging System may require additional calibration checks based on the Imagio® Breast Imaging System usage.) After 25 days, the Imagio® Breast Imaging System will recommend that calibration be performed by displaying messages in the bottom left of screen. After 28 days, the Imagio® Breast Imaging System will require calibration before using OA. The laser status code “Cal. Required” will appear on the bottom right screen.

The Imagio® Breast Imaging System Calibration routine guides the Operator through the laser energy calibration procedure used to check the OA/US Probe energy output. Upon completion of the Imagio® Breast Imaging System Calibration, the Imagio® Breast Imaging System software uses the data collected during the procedure to calibrate the laser output energy for optimum image quality and performance.

Calibration Required Example



If calibration is needed, the Operator is informed in a pop-up message.

Important: *If calibration is determined to be required, the Operator is allowed 60 minutes of use to finish the current exam.*

Please allow 5 minutes to complete the calibration procedure.

The Laser Authorization Card must be used to perform the calibration.

Important: *Perform Imagio® Breast Imaging System Calibration and the Machine Check if an OA/US L18-1 Probe is replaced with a new OA/US L18-1 Probe.*

Important: *If a calibration error is detected at any time during the procedure that requires the Imagio® Breast*

Imaging System Calibration routine to be terminated, a termination message will appear on the Display. Call Seno Service for assistance at (888) 978-8835. Minor error messages may appear on the Display Monitor that will not terminate the Calibration routine. If these messages appear, continue the Calibration through to completion. It is not necessary to dismiss these error messages software.

12.3.1 Performing Laser Energy Calibration

It is necessary to reach the end of the Imagio® Breast Imaging System Calibration routine (Exit screen) to enable the software to analyze the data and determine the success of the calibration. If the calibration is canceled before the Exit screen is reached, or an error occurs that causes the Imagio® Breast Imaging System Calibration routine to power off early, the OA/US Probe calibration has failed. Call Seno Service for assistance at (888) 978-8835.

Important: *The Imagio® Breast Imaging System will inhibit calibration if the laser is not at operating temperature as indicated by the Laser Ready icon.*

Important: *A dirty OA/US L18-1 Probe face or Calibration Port Shield during calibration may cause erroneous calibration errors.*

- Step: 1** Inspect and clean the face of the OA/US L18-1 Probe with the Seno approved wipes ensuring all gel (wet and dry) and other debris are completely removed. Refer to Cleaning System Components.
- Step: 2** The OA/US L18-1 Probe Calibration Adapter is a device used to guide and properly seat the OA/US L18-1 Probe in the Calibration Port. The OA/US L18-1 Probe Calibration Adapter is stored in the Calibration port. The Energy Sensor Shield is a clear window barrier that protects the light sensor.
- Remove the OA/US L18-1 Probe Calibration Adapter.
 - Look into the Calibration Port and perform a visual check of the Energy Sensor Shield to ensure the Energy Sensor Shield is free of gel and visible contamination.



Step: 3 If the Energy Sensor Shield appears dirty: Thoroughly clean the Energy Sensor Shield with Seno approved wipes ensuring all gel (wet and dry) and other debris are completely removed.

Step: 4 Place the OA/US L18-1 Probe Calibration Adapter over the OA/US L18-1 Probe.




Step: 5 Verify the Calibration Port is dry before inserting the OA/US L18-1 Probe and equipped Calibration Adapter.

Step: 6 Insert the OA/US L18-1 Probe with the equipped Calibration Adapter into the Calibration Port until it is seated firmly against the Energy Sensor Shield.



Step: 7 Select the Probe option on the touchscreen.




Step: 8 Select the OA/US L18-1 probe.

Step: 9 Enter OA mode by pressing the OA mode button on the system console.

Step: 10 Wear the Seno Medical-approved protective eyewear and confirm everyone in the room is properly wearing Seno Medical-approved eyewear.

Step: 11 Authorize the laser by placing the Laser Authorization Card over the NFC reader on the console.

Step: 12 If the cursor is not displayed, display the cursor by pressing the system console cursor button. move the cursor over the Calibrate button and select the button by pressing the Touchpad.



Note: When "Calibrate" is selected, review the Laser Calibration Check Box List displayed on the Laser Energy Calibration window and complete any actions required to check all boxes.

Laser Energy Calibration Check Box List:

Laser Energy Calibration

Operator: Seno Operator

- ☒ OA Probe Selected on Console
- ☒ OA Probe Seated in Calibration Box
- ☒ Probe Light Connector Seated
- ☒ System Conditions Met
- ☒ Laser Enabled

- OA Probe Selected on Console
- OA Probe Seated in Calibration Box
- Probe Light Connector Seated
- “System Conditions Met” means coolant is at temperature and interlocks are in place (Remote, Foot Switch Released)
- Laser Enabled

Step: 13 Select the next button.

Note: The “Next” button is not enabled until all pre-conditions are met.

Laser Calibration In Progress

Press and hold footswitch

0%

Cancel

- **Important:** *The foot switch should remain pressed during calibration to complete the process. Interlocks must remain in place throughout calibration.*

Step: 14 Press the foot switch down completely to begin. The calibration screens will update as the calibration proceeds to 100%.

Laser Calibration In Progress



Press and hold footswitch .

22%


Cancel

Note: The % complete will jump to 50% upon transitioning from Alex to YAG.

Step: 15 Release the foot switch when the following message appears. The finalization message appears.



Note: The following message appears after successful calibration.



Step: 16 On the Calibration Completed screen click the Exit button when prompted.


Step: 17 Remove the OA/US L18-1 Probe from the Calibration Port.

Step: 18 Remove the Calibration adapter from the probe and place the Calibration adapter in the Calibration port. Place the OA/US L18-1 Probe in the Probe Holder.

12.3.1.1 Unsuccessful Calibration

Important: The following is an example of a message that appears after unsuccessful calibration. Any calibration where the Complete screen does not show the message, "Calibration successful. Laser is ready for use" is an unsuccessful calibration.

Important: The following message appears after unsuccessful calibration.



12.3.1.1.1 Laser Adjustment Period

- The Imagio® Breast Imaging System may begin a Laser Adjustment Period after a period of inactivity. The adjustment period generally requires 3 - 60 seconds.
- The laser emit tone will sound during the Laser Adjustment Period
- No OA images will be displayed on the monitor until adjustment is complete
- The word “Preparing” is shown below the laser “Ready” icon on the bottom right of the display

12.3 Machine Check

An operator may be asked to perform the machine check procedure by Seno Service if there are any changes in image quality between service intervals. Optionally, an operator may also perform the machine check to practice their OA imaging training. Seno Service should be contacted for questions about changes in machine check results as these may relate to phantom degradation or suboptimal Imagio® machine performance.

12.4.1 To Initiate the Machine Check Procedure

Ensure everyone in the scan room is wearing laser protective eyewear. This is a laser safety requirement.

Warning: Invisible laser light is emitted from the OA/US L18-1 Probe. Reflected and/or misdirected laser light may damage eyes.



- Do not use any controls or adjustments or performance settings other than those specified herein.
- The Patient, the Imagio® Breast Imaging System Operator, and everyone else in the scan room must always wear Seno approved laser protective eyewear when the Imagio® Breast Imaging System is powered on and the laser is enabled.
- Wearing Seno approved laser protective eyewear is a laser safety requirement. It is necessary to wear laser protective eyewear when, for example; scanning, inspecting, and cleaning the OA/US L18-1 Probe, and calibrating laser energies or any other time the Imagio® Breast Imaging System is powered on and the laser is enabled.
- Do not look directly at the laser beam even when laser protective eyewear is being worn.
- Do not look at a laser beam emitted from the Imagio® Breast Imaging System that is visible. The Imagio® Breast Imaging System uses lasers that are classified as invisible; however, some individuals can see light beyond the visible spectrum.
- Always point the OA/US L18-1 Probe away from everyone’s eyes.
- Only use your foot to press the foot switch.
- Always release the foot switch before removing the OA/US L18-1 Probe from the skin or calibration port.
- Always follow the laser safety precautions prior to laser emitting to protect the Patient, Operator, and Observers.
- Always place the OA/US L18-1 Probe in the Probe holder when not in use.

Important: A dirty OA/US L18-1 Probe face or calibration port shield during the Machine Check may cause imaging artifacts.

Step: 1 Ensure the face of the OA/US L18-1 Probe is completely free of dried gel and other debris. If necessary, clean the shield of the OA/US L18-1 Probe with the Seno approved wipes. Power on the Imagio® Breast Imaging System and enable the laser if needed.

Step: 2 Place the OA Flat Breast Phantom within reach of probe and apply Seno approved gel.

Step: 3 Press the Patient button on the Touch Screen.

Step: 4 Enter Machine Check Identifier in the Patient ID field


Step: 5 Enter your name in the Performing physician/sonographer field.

Step: 6 Press the Image tab to enter live imaging and if needed, press the System Console OA button to enter OA mode. Press freeze if the image is frozen.

Step: 7 Wear laser Protective eyewear then enable the laser.

Step: 8 Observing the Display Monitor, press the foot switch completely and fully to the floor, and ensure that a clear, complete image is displayed.

Step: 9 Move/rotate the OA/US L18-1 Probe on the Breast Flat Phantom and attempt to duplicate the image below. Call Seno Service for assistance at (888) 978-8835 if this image cannot be duplicated or artifacts appear that prohibit a diagnostic quality image.



Step: 10 Press the Freeze button.

Step: 11 Press the Save Image button.

Step: 12 Release the foot switch.

Step: 13 Press End Exam button to save and close out the Machine Check exam.

Note: *Pressing End Exam automatically disables the laser.*

Step: 14 Remove laser protective eyewear.

Step: 15 Clean the OA/US L18-1 Probe with the Seno approved wipes, and then securely store the OA/US L18-1 Probe in the probe holder.

Step: 16 Clean the gel from the Flat Breast Phantom using the Seno approved wipes.

Step: 17 Cover the OA Flat Breast Phantom and return it to the designated storage location.

13 Managing Patient Data

13.1 Saving Data

13.1.1 Capturing Images and Clips

Capture and save a single image or a clip. Clips can be captured retrospectively or prospectively. The captured frame or clip is saved in the patient exam, and a thumbnail is available in the live imaging display and the Review display. When the capture is complete, a thumbnail of the image is displayed.

13.1.1.1 Capturing a single Image

Step: 1 Press the Freeze button.

Step: 2 Press the Save Image button.

- The image appears in the thumbnail strip on the right side of the screen.

13.1.1.2 Capturing a Retrospective Clip

Step: 1 Press Freeze.

Step: 2 Press Save Clip when frozen to save the Retrospective buffer as a clip.

- The clip appears in the thumbnail strip on the right side of the screen.

13.1.1.3 Play Clip

When the image is frozen, in all modes, a Play Clip button appears on the Touch Screen. Play Clip enables viewing of clips.

13.1.1.4 Capturing a Prospective Clip

The Imagio® Breast Imaging System enables capturing of prospective clips in live imaging. Operators can set the duration of the prospective clip from 3-15 seconds, refer to “Laser Enable Duration” on page 139 “Clip Preferences”.

Step: 1 Press Save Clip during live imaging to initiate a prospective clip capture.

- Prospective clip capture will automatically complete after the duration specified in the System Configuration.

Step: 2 Press Save Clip again or Freeze to stop the capture. The clip appears in the thumbnail strip on the right side of the screen.

Note: *Do not change the scaling of the image during the clip capture, especially the following settings:*

- Digital Zoom
- Display Format
- Dual Top/Bottom
- Sector

Note: Pressing Save Image while acquiring a clip does not stop the clip acquisition.

13.1.1.5 Trimming Clips

Trim enables discarding of frames from the start or end of a prospective or retrospective clip.

Step: 1 Touch Trim on the Touch Screen.

- The Touchpad is changed to the pink active Trim End.

Step: 2 Slide your finger on the Touchpad towards the beginning of the clip to discard the required frames from the end of the clip.

Step: 3 Press the Touchpad to set the end position and activate the pink Trim Start.

- The Touchpad is changed to Trim Start

Step: 4 Move the Touchpad towards the end of the clip to discard the non-desired frames from the start of the clip. Press the Touchpad to set the beginning position.

Step: 5 Press Save Clip to save a new clip without the non-desired frames.

- If needed, press Trim Reset to return the original clip Start and End frames.
- Press Trim Clip, Image, or the Mode button to return to live imaging.

13.2 Performing measurements

13.2.1 Accessing measurements

Perform measurements to assess the dimension, area, perimeter, or volume of anatomical features. Two types of measurements are available on Imagio® Breast Imaging System:

- Basic measurements, which can be done with a set of measurement tools
- Labeled measurements, which correspond to the actual anatomy of the body. Labeled measurements are not linked to an image. They appear in the report even if the image on which they were performed is deleted.

Important: Measurements are replicated across the 6 on 1 display on OA Mode.

Important: Make sure the image is correctly optimized before performing any measurements.

Important: Press the “Back to Image parameters” button to permit scrolling back in the timeline.

Press Measures on the system console to enter the measurement mode. A default caliper appears on the main display.

The image area is duplicated on the Touch Screen to facilitate stylus measurements.

Measurements that appear with a star (*), ellipse for example, are estimated measurement results.

13.2.2 Basic measurements

Press the Measurements button on the system console.

Touch Measures on the Touch Screen.

The Imagio® Breast Imaging System displays the measurement tools that are available for the selected mode. Measurement results are displayed in the Measurement ResultBox which is located to the lower right of the image, below the gray-scale bar, on the main display. This ResultBox can be moved and placed all over the image.

In More Meas. Tools, press the Reset ResultBox position button to automatically replace the ResultBox to its initial place.

Touch the desired measurement tool:

13.2.2.1 Distance

Distance is available in all operational modes.

Method 1 (with the Touchpad):

Step: 1 Press Measure.

Step: 2 Use the Touchpad to move the active cursor to the first measurement point.

Step: 3 Press the Touchpad to anchor the first point. The Imagio® Breast Imaging System anchors the first cursor and displays a second active cursor. The Imagio® Breast Imaging System displays the value of distance in the measurement result area to the right of the image on the main display.

Step: 4 Touching End will end the measurement.

Method 2 (with the Touch Screen):

Step: 1 Press Measure.

- The active cursor appears.

Step: 2 Touch the image on the Touch Screen to anchor the first cursor location. The Imagio® Breast Imaging System moves the cursor to the position indicated, anchors the first cursor, and displays a second active cursor.

Step: 3 Touch the image on the Touch Screen to position the second cursor location.

Step: 4 Touch End to anchor the second cursor.

It is possible to combine the approaches of Method 1 and Method 2 to optimize the measurement workflow.

13.2.2.2 Ellipse

Ellipse is available in all modes.

Method 1 (with the Touchpad):

Step: 1 Press Measure

Step: 2 Touch Ellipse

Step: 3 Press the Touchpad to anchor the first point. The Imagio® Breast Imaging System anchors the first cursor and displays a second active cursor.

Step: 4 Use the Touchpad to move the active cursor to the second point of measurement.

Step: 5 Press the Touchpad to anchor the second cursor. The Imagio® Breast Imaging System displays an ellipsoid trace between the two points.

Step: 6 Use the Touchpad to adjust the shape of the ellipse.

Step: 7 Pressing the Touchpad will sequentially switch control to the first cursor, second cursor and ellipse trace.

Step: 8 Press End to complete the measurement.

Method 2 (with the Touch Screen):

Step: 1 Press Measure

Step: 2 Touch Ellipse. The active cursor appears.

Step: 3 Touch the image on the Touch Screen to anchor the first cursor location. A second cursor will automatically appear.

Step: 4 Without lifting the finger or stylus, drag the second active cursor along the axis of the ellipse. The Imagio® Breast Imaging System displays an ellipsoid trace between the two points. Lifting the stylus or

finger will anchor the second cursor.

Step: 5 Touch the image near the opposite axis of the ellipse.

Step: 6 Without lifting the finger or stylus, drag the ellipse trace to the desired position.

Step: 7 Touch End to complete the measurement.

It is possible to combine the approaches of Method 1 and Method 2 to optimize the measurement workflow.

13.2.2.3 Trace Region of Interest (ROI)

Trace is available in all operational modes. Trace is a drawing of the ROI.

Method 1 (with the Touchpad):

Step: 1 Press Measure

Step: 2 Touch Trace button.

Step: 3 Use the Touchpad to move the active cursor to the first measurement point.

Step: 4 Press the Touchpad to anchor the first point. The Imagio® Breast Imaging System anchors the first cursor and displays a second active cursor directly on top of the first.

Step: 5 Use the Touchpad to move the active cursor to the second measurement point. The path will be displayed.

Step: 6 Press on the Touchpad to anchor the second cursor. The Imagio® Breast Imaging System will automatically connect the end points of the trace.

Step: 7 Alternately, touch End to complete the measurement.

If the start and end points of the trace become very near to each other, the Imagio® Breast Imaging System will automatically close the trace.

The knob under Trace can be used on an active trace to incrementally erase the trace in reverse and redo the trace.

Method 2 (with the Touch Screen):

Step: 1 Press Measure

Step: 2 Touch Trace

- The active cursor appears.

Step: 3 Touch the image on the Touch Screen to anchor the first cursor location.

- A second cursor will automatically appear directly on top of the first.

Step: 4 Without lifting the finger or stylus, drag the second active cursor along the path of the desired trace.

- The path will be displayed.
- Lifting the finger will pause the trace.
- Touching the cursor will reactivate the trace.

Step: 5 Touch End to complete the measurement. If the start and end points of the trace become very near to each other, the Imagio® Breast Imaging System will automatically close the trace.

13.2.2.4 Depth

This tool enables defining of the depth on the image at a particular point of interest.

Step: 1 Press Measure.

Step: 2 Touch More Meas. Tools.

Step: 3 Touch Depth.

Step: 4 Press the Touchpad to anchor it at the desired depth.

Step: 5 The depth value will be displayed.

13.2.2.5 Volume

A three-distance volume enables using three independent distances in the same or in orthogonal images to create a volume calculation.

Step: 1 Press Measure.

Step: 1 Press More Meas. Tools.

Step: 2 Touch Volume.

Step: 3 Use your preferred method (refer to above, Distance measurements) to draw three distances.

Step: 4 The Imagio® Breast Imaging System calculates a volume from the three distances.

Calculate a volume with labeled measurements is available. Refer to “Labeled measurements” on page 112.

13.2.2.6 Ellipse distance volume

Step: 1 Press Measure.

Step: 2 Touch More Meas. Tools.

Step: 3 Touch Volume (Ellipse + Distance).

Step: 4 Use your preferred method to draw an ellipse on the plane of interest Once the ellipse is drawn, the Imagio® Breast Imaging System displays the first caliper of a distance measurement.

Step: 5 Choose an orthogonal plane to draw the distance Once the distance is drawn, the Imagio® Breast Imaging System displays the values of the volume in the measurement result area.

13.2.2.7 Q-Box™

The Quantification Box (Q-Box™) enables accurate quantification of the stiffness of an area.

Q-Box™ is available only in SWE™ mode on a frozen image.

Step: 1 Press Measure

Step: 2 Touch Q-Box™

- Q-Box™ displays a circle that can be resized or moved, and that is duplicated on the B-mode image (in side by side and top bottom formats), for reference purposes.

Step: 3 Select to anchor it.

13.2.2.8 Q-Box™ Ratio

Q-Box™ Ratio enables comparison of the stiffness of two areas on the same image.

Q-Box™ Ratio is available only in SWE™ mode on a frozen image.

Step: 1 Press Measure

Step: 2 Touch Q-Box™ Ratio.

- Q-Box™ Ratio displays a circle that can be resized or moved, and that is duplicated on the B-mode image (in side by side and top bottom formats), for reference purposes.

Step: 3 Place the Q-Box™ within the stiffest visualized area

Step: 4 Press the Touchpad to anchor it

Step: 5 Another Q-Box™ is displayed.

Step: 6 Place it on soft tissue .

Step: 7 Press the Touchpad to anchor it.

Note: The values for ShearWave speed and tissue modulus are relative indices intended only for the purpose of comparison with other measurements performed using the Imagio® Breast Imaging System, and absolute values for these measurements should not be compared to shear wave elastography measurements from other manufacturers' ultrasound imaging systems.

Please note that breast lesions on mammography and sonography may contain blue coded areas. This may be due to soft focal areas in the tumor or noisy data that triggers errors in Shear Wave speed estimation. Place the ROI in the stiffest part of either the lesion or tissue immediately adjacent to the lesion to gain the most relevant diagnostic information.

13.2.2.9 Q-Box™ trace

Q-Box™ Trace enables manually tracing a Q-Box in SWE™ imaging mode in order to obtain some elasticity values within the traced area.

Q-Box™ Trace is available only in SWE™ mode, on a frozen image.

Step: 1 Press Measure

Step: 2 Touch Q-Box™ Trace

Step: 3 Use your preferred method (with the Touchpad, or with the Touch Screen)

Step: 4 Press the Touchpad to end the measurement

13.2.2.10 Multi Q-Box™

Multi Q-Box™ enables automatic calculation of the average of several Q-Box measurements.

Multi Q-Box™ is available only in SWE™ mode on a frozen image.

Step: 1 Press Measure

Step: 2 Touch Multi Q-Box™

Step: 3 Resize and move the Q-Box as needed

Step: 4 Press the Touchpad to validate

Step: 5 To add another Q-box, repeat the operation the Imagio® Breast Imaging System calculates the average of each Q-Box result. Optionally, continue the Multi Q-Box measurement cycle on several images.

Step: 6 Press End to end the Multi Q-Box measurement cycle

13.2.3 Other Touch Screen measurement functionalities

The measurements Touch Screen has some specialty controls to facilitate measurement work flow.

13.2.3.1 Erase all

Erase All enables erasing all of the measurements displayed on the frozen image.

13.2.3.2 Erase last

Erase Last enables erasing only the last caliper anchored.

13.2.3.3 End

End will end the active measurement.

13.2.3.4 Undo/Redo

The knob under Undo/Redo can be used on an active trace to incrementally erase the trace in reverse and redo the trace.

13.2.3.5 Zoom

Zoom enables enlarging of the size of the Touchscreen measurement window.

Rotate the knob located under Zoom:

- clockwise to increase the zoom factor
- counterclockwise to decrease the zoom factor

13.2.4 Labeled measurements

Press Meas. on the system console.

A list of labeled measurements is displayed on the right side of the image.

The available labels depend on the selected application and on the current active mode.

The label packages can be configured in the System Configuration. Refer to “Packages” on page 162. Press Cursor.

To scroll the list, place the cursor over it, and scroll on the Touchpad. Pick the appropriate label from the list by using the Touchpad.

The appropriate measurement tool is launched (distance...). Perform the measurement as for a basic

measurement.

Make up to 5 measurements for the same label. They appear with their corresponding caliper in the list.

Note: Labeled measurements may be hidden, depending on the selected option from the System Configuration.

13.2.4.1 Assign last

Assign Last enables assignment of the last measurement performed to a label.

Step: 1 Perform a measurement

Step: 2 Touch Assign Last on the Touch Screen

- The Imagio® Breast Imaging System displays the list of available labels for your measurement

Step: 3 Select the desired label

For more information on labeled measurements, refer to “Labeled measurements” on page 112.

13.3 Using annotations

13.3.1 How to use annotations

Add text annotations and arrows on an image to denote anatomical structures and locations. Annotate is located on the system console.

The annotation feature is available in live imaging and on frozen image. To annotate an image, press the Annotate button on the system console.

A list of annotations appears on the Touch Screen and an annotation cursor appears on the main monitor display.

Rotate the Page knob to access various pages of annotations. To add an annotation, choose from the following actions:

- Choose one of the pre-programmed annotations in the library of annotations. Touch the desired annotation to add it to the main display.
- Touch Keyboard to display a keyboard and manually enter an annotation.
- Touch Arrow on the Touch Screen to add an arrow graphic to the main display. Use the Touchpad to position the arrow.

To remove annotations from the main display, use the following functions:

- Clear All removes all the annotations and arrows from the main display.
- Delete Last Annot. removes the annotation last text annotation entry.
- Del. Last Arrow removes the last arrow marker applied to the image.

The annotations feature has some specialty controls to facilitate annotation work flow.

13.3.2 Annotation mode

When adding an annotation as a title, it will remain when the image is unfrozen.

When adding an annotation as a free annotation, it will be removed when the image is unfrozen. Press Mode to change the annotation mode.

13.3.3 Group of annotations

Annotations can be grouped by color. If they have a specific color, they belong to the same group. When selecting a first annotation from a group, it will be added to the screen. When selecting an annotation from the same group, it will replace the one that is already on the screen, if it is from the same group.

Annotations on transparent buttons do not belong to a given group. Therefore, they will not be replaced. Annotations all appear in white when the image is saved or printed.

13.3.4 Moving annotations

Step: 1 Select an annotation from the Touch Screen to add it

Step: 2 Move the cursor over the newly added annotation

Step: 3 Press the Touchpad

- The annotation is selected

Step: 4 Move the cursor to move the new annotation location

Step: 5 Press the Touchpad to release the annotation

Note: If moving an annotation over an existing annotation, the existing annotation will be replaced.

13.3.5 Annotation Library

To edit the library of pre-set annotations manually. Touch EDIT LIB. (edit library). to edit the library of annotations.

This opens the System Configuration page, the System/Display tab, and Annotation sub-tab. To manage the annotations for the desired clinical application.

Refer to “System Configuration” on page 137 for more details.

13.3.6 Home

Touch Home to move the annotations cursor to the home position.

Touch Set Home to define the current position of the cursor as the Home position for the current display format.

13.3.7 Exit

Touch Back To Image Parameters or press the Freeze button twice to exit annotations.

13.4 Using body markers

13.4.1 How to use body markers

Body markers is available in live imaging and on frozen image.

Press Annotate on the system console then Body markers on the Touch Screen to add a body marker to an image.

Note: To customize the access to Body markers. Please refer to “Bodymarker” on page 157

- The default body marker will appear on the on the Touch Screen and on the main monitor display to the lower right of the image.

Touch PICTO to change the body marker pictogram.

A selection of body markers appears on the Touch Screen.

Touch the desired body marker to select it.

The probe orientation can be indicated directly on the pictogram.

To add a probe orientation to the pictogram, simply touch the pictogram to indicate the edge of the probe that corresponds to the orientation marker on the probe.

Then touch the pictogram again to indicate the edge of the probe opposite the orientation marker.

The probe orientation appears. Use the Rotate knob to rotate the probe orientation on the body marker. Press

Exit on the Touch Screen to close the Body Marker page.

Touch HIDE to hide the picto on the image.

To restore a body marker to the main display, touch SHOW.

13.5 Reviewing an Exam

During or after an exam, use Review to examine and compare images acquired in the exam. Multiple exams for one patient may be reviewed. In Review, operators look at the images or clips that have been stored. Stored

images can be viewed, sent, printed, searched, and backed up. Image analysis can also be performed in Review. Images that are stored on the ultrasound Imagio® Breast Imaging System hard-drive can be sent to a USB media, or on DICOM-compatible devices on a network.

13.5.1 General Information on Review

In Review mode, a variety of tasks may be completed, including performing measurements and playing clips.

13.5.2 Measurements in Review

In Review mode, measurements on images in the current exam can be performed. An image must first be displayed in full size. Measurements made in Review on the current exam can be saved in the report.

Measurements made in Review on prior exams cannot be saved. To display the measurement controls on the Review Exam Touch Screen, press Meas. Clear all measurements from an image by touching Erase all on the Review Exam Touch Screen.

13.5.3 Starting Review

Press Review to enter Review mode.

The display that appears depends on whether an exam is active on the Imagio® Breast Imaging System.

If an exam is in progress, pressing Review opens the Review Exam display and a reminder is displayed on each thumbnail indicating “Not from Current Exam”.

If there is not an exam currently in progress when Review is selected, the Patient Directory will be displayed.

If an automatic deletion is set up in the System Configuration, locking some exams to prevent them from being automatically deleted is available.

To return to live imaging, press Review again.

13.5.3.1 Patient Directory

The Patient Directory is a list of exams that are stored on the selected disk drive. The Patient Directory includes options that allow the operator to sort, view, and transfer exams. If no current exam is in progress when Review is started, Patient Directory is displayed.

13.5.3.1.1 About patient directory headers

In the Patient Directory, exams are organized in a table containing several columns. Each column header describes the contents of that column using either text or an icon. Press a column header to sort the list by this column. The Exam Status column displays the number of times an exam has been continued (if any).

13.5.4 Selecting and Loading Exams

If an exam is currently in progress, pressing Review opens the Review Exam display with images loaded from the current exam.

If no exam is in progress, pressing Review opens the Patient Directory display, which lists previous exams. To review previous exams, first select an exam in the Patient Directory.

Multiple exams of the same patient may be selected and loaded for review.

Select, Select All to include all exams.

Select Display Selected to load the selected exams for viewing

13.5.4.1 Viewing images

The Review Exam display is used for viewing and comparing exam images in the selected layout. Thumbnail images for the current or selected exam appear on the right side of the display. The TouchScreen display provides access to other review functions.

13.5.4.2 Navigating thumbnails of images

In Review small images, called thumbnails, can be viewed. Thumbnails are located on the right side of the Review Exam display.

Touch Previous thumbnail and Next thumbnail on the Touch Screen to navigate through thumbnails of images. This will display the previous or the next image in full screen review mode.

Putting the cursor over a thumbnail makes it expand to a larger dimension. Selecting the thumbnail image will make it expand to its real size.

Selecting the garbage icon over a thumbnail deletes that image.

13.5.4.3 Comparing images

In Review, two images from different exams of the same patient may be compared.

Step: 1 Select the two exams in the Patient Directory

Step: 2 Select Display Selected.

Step: 3 Select the patient name on the left side Thumbnails of the two exams will appear.

Step: 4 Select the two images to be reviewed and touch Compare.

The two images appear side by side. They can be reviewed at the same time and measurements and annotations can be performed.

13.5.5 Sending an Exam

13.5.5.1 Export Formats

In Review, specific images can be pushed to DICOM printers (optional) and servers on a network or to a USB device. The format can be chosen before sending images or clips. To send an enlarged image to a printer or USB device, refer to “System Configuration” on page 137 “Enlarge Mode”.

13.5.5.2 Sending Images

In the Review Exam display, images can be pushed from the Imagio® Breast Imaging System to DICOM compatible printers and servers on a network.

Step: 1 In the Review Exam display, select one or more images

Step: 2 Touch Export to DICOM or Export as JPEG/AVI

Step: 3 In the Send To dialog box, select a destination

In the Send To dialog box, the total space, as well as the free space available on the device are displayed for a USB device.

Note: In accordance with DICOM PS3.15 2019a, operators can check the Image de-identification check box in the Send To dialog box to send the image(s) with identification fields cleared from the database and the image.

Step: 1 Select OK to send the selected images

- The exams will be exported into folders organized by patient name. To eject a USB device:

Step: 2 Wait until the  symbol (exporting data to USB) is replaced by  in the notification icons area

Step: 3 Remove the USB device

At any time, operators may check the export status, by selecting the appropriate notification icon on the left side of the screen. For information on icons, refer to “Main Display Notification Icons” on page 130.

13.5.5.3 DICOM Structures Reporting (DICOM SR)

The DICOM Structured Report is a document architecture designed for encoding and exchanging clinical information and findings in the context of Radiology imaging.

By default, the Imagio® Breast Imaging System will not generate DICOM structured Reports at the end of the exam.

13.5.5.3.1 Configuration

13.5.5.3.1.1 Structured Report (SR) option

In order to activate the DICOM SR feature:

Step: 1 Press Imagio® Breast Imaging System Configuration “Imagio O gear” top left touch screen button.

Step: 2 Open the Device Settings tab

Step: 3 Open the Imagio® Breast Imaging System DICOM options sub-tab

Step: 4 Check the Generate Structured Reports check-box

13.5.5.4.1.2 DICOM Store

Step: 1 In the System Configuration, press the Devices sub-tab of the Administration tab

Step: 2 Select Edit Selected

Step: 3 Select the Advanced options of the DICOM store device

Step: 4 Select the Enable Export of SR check-box of your choice

13.5.5.4.1.3 DICOM Removable media

Step: 1 In the System Configuration, select the Devices sub-tab of the Administration tab

Step: 2 Select the DICOM Removable media

Step: 3 Select on Edit Selected

Step: 4 Select the Enable Export of SR check-box of your choice

13.5.5.3.2 Sending Data Methods

13.5.5.3.2.1 Automatic Send of DICOM SR Reports

When the export of SR is enabled for one or several DICOM Store devices, a DICOM SR file will be sent automatically at the end of the exam to this/these configured DICOM Store devices.

13.5.5.3.2.2 Send On Demand of DICOM SR Reports

During a live examination, and if the export of SR is enabled for one or several DICOM Store devices, sending on demand updated versions of the DICOM SR report to this/these configured DICOM Store devices is an option.

Step: 1 Press Review or Report

Step: 2 Press Send SR on the Touch Screen

13.5.5.4.2.3 Manual Send of DICOM SR Reports

Step: 1 When an exam is ended, press Review

Step: 2 Select the exam to send the SR, and press Display Selected

Step: 3 In the Review Image List, select the DICOM SR icon

Step: 4 Press Export to DICOM on the Touch Screen

Step: 5 Select the device and select OK

Note: Operators can configure all the DICOM exportation settings. Refer to “Exporting reports” on page

124.

13.6 Deleting exams and images

The Patient Directory enables deleting of exams from the Imagio® Breast Imaging System. In the Review Exam display, stored images from an exam can be deleted.

This is only possible, however, when Send Images/Clips in Print/Network setups is set to At End of Exam. When an image is deleted, it remains temporarily in the Review Exam display with an X marked through it.

13.6.1 To delete one or more images

- Step: 1** Select the images to be deleted
Step: 2 Touch Delete Selected on the Touch Screen
Step: 3 Press Delete in the confirmation box

13.6.2 To delete exams

- Step: 1** In the Patient Directory, select one or more exams. To select all exams, touch Select All
Step: 2 Press Delete Selected
Step: 3 Press Delete in the confirmation box

13.7 Image Query and Retrieve

Prior to, or during an exam, a query may be performed to find patient exams located on a remote PACS or external USB media.

Qualified images (for modality types specified in the DICOM Conformance Statement) can be extracted from these exams and retrieved by the Imagio® Breast Imaging System.

Once retrieved, the images can be browsed, and an appropriate image can be selected for Review.

A retrieved image may also be displayed in a side-by-side manner next to a live image in the current exam.

Important: *The Query and Retrieve feature requires the Imagio® Breast Imaging System is connected to a network hosting a DICOM compliant image archive workstation or external USB media.*

13.7.1 General Information about Query-Retrieve

If there is no current exam, a query may be performed for any number of exams from the Q/R source.

These exams can be retrieved by the Imagio® Breast Imaging System in preparation for upcoming new exams.

If there is a current exam in progress, the query will attempt to find prior exams of the current patient (if the automatic Q/R was configured).

Valid types to be retrieved are as follows:

- DICOM Ultrasound images
- DICOM Ultrasound images extracted from ultrasound clips
- DICOM images from a variety of modalities: Computed Radiography (CR), Computed Tomography (CT), Magnetic Resonance (MR), Nuclear Medicine (NM), Secondary Capture (SC), X-Ray Angiographic (XA), Radio Fluoroscopy (RF), Mammography (MG)

Please consult the PACS operator's manual, DICOM conformance statement or the PACS administrator for assistance with secondary captures.

Any retrieved exam can be displayed along with a live image in a current exam.

However, if there are mismatches in patient identification characteristics between the current exam and the retrieved exam, a warning will be displayed.

Refer to “Main Display Notification Icons” on page 130 for Query-Retrieve notification icon.

Important: *Displaying a patient with a different patient identification to that in the current exam must be done with caution.*

- This feature is provided in the event of a patient name change, or hospital ID change.
- The operator assumes all risk when performing query retrieve operations on Imagio® Breast Imaging System.

- Retrieved images can be very helpful when used for reference comparison to the same feature of interest in the live ultrasound image.
- Use caution when reviewing retrieved images on Imagio® Breast Imaging System, as they may be compressed or displayed at a lower level of image quality than observed on the PACS workstation.
- Images retrieved and displayed on the Imagio® Breast Imaging System are not intended to be used for sole diagnosis.

13.7.2 Performing a Query

13.7.2.1 Automatic Query

If configured, the Imagio® Breast Imaging System can perform an automatic query of various types of exams for a given patient.

When the exam is started for a given patient, press Save Image to save a first image for this patient. The Imagio® Breast Imaging System automatically performs a query on the selected patient.

Press the Q/R button on the system console to view the queried exams. Refer to “System Configuration” on page 137 for more details.

13.7.2.2 Manual Query

Press the Q/R button to initiate an exam query. The query dialog box will appear.

Select the tab to query:

- Patient Name (Name)
- Patient ID (ID)
- Accession Number (Accession)

Use the appropriate filters to manage the query then press Query.

- A list of matching exams will be displayed.

13.7.2.3 Import from Removable Media

Step: 1 Press the Q/R button to initiate an exam query.

Step: 2 Press the Import from Media button

Step: 3 A browser window appears

Step: 4 Select the location of data to be imported

Note: *You may select a folder containing DICOM data, a specific DICOM file or a DICOMDIR file to import.*

Step: 5 Press OK to confirm

13.7.3 Performing a Retrieve

13.7.3.1 Automatic Retrieve

If configured, the Imagio® Breast Imaging System performs an automatic retrieve of the last acquired series for each of the selected modalities.

13.7.3.2 Manual Retrieve

After a query is performed, the Imagio® Breast Imaging System will return a list of matching exams.

Exams are displayed in a table with the following attributes:

- Patient Name
- Patient ID
- Date of Birth

- Modality
- Number of Images
- Date/Time

To see more information about a particular exam, place the cursor over an exam and a tool-tip box will appear with additional information.

Use the Touchpad to choose exams from the query list. A check-mark will appear next to selected exams.

Choose the New Query button at the bottom of the menu, to query again. Press the Retrieve button to retrieve the selected exams.

Note: These exams will be transferred to the Imagio® Breast Imaging System. Exams containing large numbers of images or long ultrasound loops may take a long time. It is best to perform a query and retrieve operation prior to the start of an exam.

When the selected exams have been retrieved, RETRIEVED appears next to each line.

13.7.4 Opening pushed DICOM data

You may send DICOM data from a PACS or a DICOM modality to the Imagio® Breast Imaging System.

The Q/R notification icon will be displayed to indicate data have been pushed on the Imagio® Breast Imaging System.

Press the Q/R notification icon or press the Q/R button on the system console to see the received exams.

13.7.5 Displaying and Navigating Retrieved Images

Double-Press the Touchpad on an image to display it in full screen. The retrieve image will be displayed in the right pane.

Navigate through the retrieved images by using the buttons at the top of the pane:

- Previous will go to the previous image
- Next will advance to the next image.

Use the Touchpad to navigate through the retrieved images.

The full screen icon will enlarge the retrieved image to the side of the full screen, temporarily hiding the live or frozen image from the current exam.

All Imagio® Breast Imaging System controls will operate as usual on the current live or frozen image.

Imagio® Breast Imaging System controls will not affect the retrieved image.

13.7.6 Exiting Query and Retrieve

To quit viewing retrieved images, press the Q/R button on the system console.

13.8 Ending an Exam

Make sure all the images in the exam are saved. After the exam is complete, end the exam as follows:

Press End Exam on the Touch Screen.

Note: Pressing End Exam automatically disables the laser.

13.9 Continuing an Exam

You have the possibility to continue an exam that was ended.

Step: 1 Touch End Exam to make sure there is no exam in progress

Step: 2 Touch Review

Step: 3 From the Patient Directory, select the exam to continue from

Step: 4 Touch Continue on the Touch Screen

The Imagio® Breast Imaging System has re-opened the closed exam and is in B-mode. The thumbnails of the previous session(s) of the same exam are displayed on the right side.

The time limit to re-open an exam may be configured in the Imagio® Breast Imaging System/Display tab of the System Configuration.

Press End Exam to close the exam.

13.10 BI-RADS Analysis

The Breast Imaging Reporting and Data System (BI-RADS®), developed by the American College of Radiology, provides a standardized classification for ultrasound studies of the breast. It is composed of a series of descriptors, from which an assessment linked to a category can be made by the physician.

13.10.1 BI-RADS® assessment categories

BI-RADS® score	Assessment
0	Assessment incomplete: need additional imaging evaluation
1	Negative
2	Benign finding
3	Probably benign finding
4	Suspicious malignancy
4a	Low suspicion for malignancy
4b	Moderate suspicion for malignancy
4c	High suspicion for malignancy
5	Highly suggestive of malignancy
6	Known biopsy-proven malignancy

Note: All BI-RADS criteria are displayed in English on the Imagio® Breast Imaging System, regardless of the selected language.

Operators can easily perform BI-RADS® classification using the integrated BI-RADS® Lexicon Classification form. Up to twelve lesions may be classified in a single exam.

13.10.2 Performing BI-RADS® lexicon classification

Step: 1 In any mode, press freeze to freeze the image.

Step: 2 Touch BI-RADS.

Step: 3 Touch Add New Lesion. An arrow appears labeled “A” appears (first arrow, multiple sequential arrows B, C, D, with each press of Add New Lesion).,

Step: 4 Using the Touchpad, position the lesion arrow marker to indicate the lesion of interest on the main display.

Step: 5 Press on the Touchpad to anchor the arrow marker.

- The BI-RADS® menu will appear to the left of the image on the main display.

Step: 6 Select the lesion features in the BI-RADS® lexicon classification form using the Cursor and the Touchpad

Step: 7 Press the NEXT button in the menu or the Lesion Page control on the Touch Screen to move to the next page of the BI-RADS® menu.

Step: 8 Press Save Image.

Step: 9 Select the “X” to close the BI-RADS® menu.

Important: You must press Save Image after filling out the BI-RADS® classification form and exiting the menu or the data will not be saved.

The BI-RADS® menu can be accessed as many times as needed to complete the assessment for a particular lesion. Measurements performed while in the BI-RADS® menu will be directly associated with the Lesion selected (e.g. Lesion A) in the final report. Measurements performed in the BI-RADS® menu will appear on page 4 of the report.

13.10.3 Documenting BI-RADS® lexicon classification with images

A single image is not able to document all of the features contained in the BI-RADS® lexicon classification form. To assist in documentation of lesion features, the integrated BI-RADS® menu offers the advantage of documenting one or more BI-RADS® features with an associated image.

Step: 1 Define a new lesion and activate the BI-RADS® menu.

Step: 2 Using the cursor and the Touchpad, results may be selected in the BI-RADS® lexicon classification form.

Step: 3 Press Save Image.

Note: The image number appears to the right of the selected answer on the BI-RADS® lexicon classification form. You may exit BI-RADS® after each Save Image and return to the same lesion to enter additional data associated with new images. In the case that disparate results are entered for the same lesion, reconciliation of results can be done in the Report.

13.11 Description of the report feature

The Imagio® Breast Imaging System enables managing of all the information collected during the exam and generate a report.

The Report features is composed of two different views:

- the Report Builder, which allows creation and customization of the report
- the Report Preview, which allows viewing of the report as it will be printed or generated as pdf file

Press Report to access the report feature. The Report Builder is displayed.

13.12 Report builder

13.12.1 Description of the report builder

The Report Builder enables choosing of which elements of the examination are to be included in the final report.

The Report Builder is composed of the following tabs:

- Patient Info, for retrieval/editing of patient information
- Images, where the images acquired during the exam, associated measurements, BI-RADS® and comments are stored
- Measurements, where operators view all the labeled measurements and calculations of the exam
- Worksheet, where operators visualize the elements of the final report, modify patient information and patient data entry, use diagram or anatomical data to locate or
- view performed labeled measurements
- Conclusion, where operators add their conclusion of the exam

13.12.2 Patient information

The Patient Info tab displays the patient information, which is populated with the same information as the Patient Data screen.

As with the Patient Data screen, the Patient Information tab displays general information, as well as further information specific to each clinical application.

Operators can edit all the information from the Patient Information tab.

Operators can enter information in more than one tab if multiple applications have been used.

If Patient is pressed to return to Patient Data entry, all your edits from the Report Builder will be saved and updated.

Operators choose what information to show in the final report by using the Show/Hide Info icons:

- Selecting the Show Info icon will designate that information to appear in the final report
- Selecting the Hide Info icon will designate that information to not appear in the final report

Information which is hidden in the final report always remains part of the exam dataset as stored on the Imagio® Breast Imaging System.

13.12.3 Images

The Images tab in the Report Builder displays all of the images acquired in the exam. The images are organized into clinical application tabs.

Vertical tabs are dedicated to images associated with BI-RADS® table. Operators choose which images to show in the final report by selecting them.

Press the icon above the image to have it included in or excluded from the report.

Images which are not shown in the final report always remain part of the exam dataset as stored on the Imagio® Breast Imaging System.

Next to each image is a text box designated for comments.

Select the white text box next to the image for which comments need to be added. Enter comments using the Keyboard.

Measurements associated with each image are shown in the area to the right.

Operators choose to show/hide the measurements in the final report by using the Show/Hide Info icons:

- Selecting the Show Info icon will designate the measurements to appear in the final report
- Selecting the Hide Info icon will designate the measurements to not appear in the final report
- If no measurements were performed on the set of selected images, the region may appear empty.

13.12.4 Measurements

The Measurement tab enables viewing of the labeled measurement results and associated calculations performed during the exam.

In this tab, show or hide each label measurement result in the final report.

- Select the desired application sub-tab (available only for applications in which label measurements were performed during the exam).

Note: Even if the image on which the labeled measurements were taken was deleted, the corresponding labeled measurements are kept in the database.

Note: Show or hide each labeled measurement result individually with the check box next to each measurement.

Hidden labeled measurement results will not appear in the final report.

For multiple instances of the same measurement, select which instance will be displayed in the report. Press the + to display all the sub-results of the performed measurements.

13.12.5 Worksheet

The Worksheet tab is divided in sub-tabs corresponding to the applications and presets used to perform the exam. The Breast Worksheet displays the information contained in the Patient Data Entry and the Patient Information tab. It includes diagrams and clinical information for Breast. The measurements are also displayed and BI-RADS if they have been performed.

13.13 Exporting reports

An option enables exporting of the report as DICOM images. This can be used in particular when your PACS does not support viewing of pdf.

Step: 1 Touch Imagio® Breast Imaging System Config.

Step: 2 Select the Administration tab

Step: 3 Select the Devices sub-tab

Step: 4 Add or edit an existing DICOM store

Step: 5 In the Advanced Options tab, enable the Report Export

Step: 6 In the drop-down list, select the Ultrasound image option

Step: 7 Select OK

14 Laser Emitting Alerts, Laser State Messages, Status Codes, Information Messages, System Messages and Notifications Icons

14.1 Notifications Window

User information about the state of the laser is provided in the lower right corner of the Display and pink transient scrolling messages with laser status appear on the lower left/center area of the Display. Any pink messages that are scrolled on the Display may additionally be reviewed by selecting the Notifications button on the Laser Status Bar. The Notifications button turns pink when new notifications are available.



(1) Laser Status Messages

(2) Notifications Button on the Laser Status Bar

(3) Notifications Window shows history of Laser Status Messages

The Notifications Save Logs button allows the Operator to save logs encoded in a DICOM file which may be exported to a USB drive through the Review tab in the same manner as a normal patient image export.


Note: Two separate sets of system logs are generated. See the System Diagnostics section for instructions for the second log export.

The Notifications window About button provides:

- Software patents
- Software version
- Remote Support control (see the Remote Support section)

14.2 No Probe Contact Message

A message will flash on the bottom right of the display stating “No Probe Contact” when the probe face is not in contact with the surface.




14.3 Laser Emitting Alerts

14.2.1 Laser Emitting Intermittent Tone

Laser emission intermittent tone sounds to alert everyone in the scan room that laser light energy is being emitted from the Imagio® Breast Imaging System. Foot switch is pressed, and the OA/US L18-1 Probe is emitting laser energy.

14.2.2 Laser Emitting Laser State Message




Laser state message “Laser Emitting” flashes to alert everyone in the scan room that laser light energy is being emitted from the Imagio® Breast Imaging System.

14.4 Laser State Messages

An icon and associated text that indicates the readiness and emission status of the laser is displayed in the bottom right corner of the Display Monitor.


Icon and text	Message Description	Flashing
 Laser Heating	Laser is warming up and has not reached operating temperature. Warmup.. ##% status code will also be displayed. This laser status icon will change, and the status code will disappear when the laser reaches operating temperature.	No
 Preparing	Laser Adjustment Period. Designed to ensure that the laser attenuator has time to change prior to OA/US L18-1 Probe laser emission. Typically, 3-10 seconds <60 seconds	No
 Laser Standby	The laser enters standby, and the Laser Standby icon is displayed when: Not in OA mode, or the Console  Freeze button is pressed when in OA mode.	No
 Laser Starting	During initial startup of the lasers, the Laser Starting icon(s) maybe briefly displayed.	No
 Laser Ready	Laser has reached optimal operating temperature and an Imagio® Breast Imaging System exam can now be initiated.	No
 Laser Emitting	Foot switch is pressed, and the OA/US L18-1 Probe is emitting laser energy. Wearing laser protective eyewear is mandatory!	Yes
 Laser Standby Comm. Error	Laser communication errors are indicated when the laser state icon has a yellow background. Call Seno Service for assistance at (888) 978-8835.	No

Icon and text	Message Description	Flashing
	Laser communication errors are indicated when the laser state icon has a yellow background. Call Seno Service for assistance at (888) 978-8835.	No
	A condition has occurred that should be remedied as soon as possible. Call Seno Service for assistance at (888) 978-8835.	No
	This icon is briefly displayed at initial laser power on, and in the case of a laser malfunction. If the Laser Resetting icon is persistently displayed, call Seno Service for assistance at (888) 978-8835.	No

14.5 Status Codes

Status codes are additional messages that may appear with Laser State Messages that indicate other operational aspects of laser.

For example, when the laser is ready and distilled water (Add Coolant) is required to be added to the Imagio® Breast Imaging System, this message will appear:



Displayed Value	Description	Action
Cal. Required	Calibration required. (due to shot count, elapsed time, or change of firmware/software)	Calibrate the laser energy.
Laser Starting	Software is in process of attempting communication with the Laser.	Informative. No action needed.
Warmup.. ##%	The state of the laser temperature as the laser is heating to 40 deg. C operating temperature expressed as a percentage.	Informative. No action needed.
Add Coolant	Coolant has dropped below low mark in reservoir.	Fill the reservoir with coolant.
Probe Light Connector Disc.	Cable micro switch is not made.	Connect probe optic to laser head output port.
Check Light Connector	Cannot communicate with probe energy sensor.	The probe energy sensor pad on probe may be dirty and the contacts should be cleaned.
Probe Light Connector Seated	Cable micro switch is made.	Informative. No action needed.
Foot Switch Disc.	Foot Switch Interlock signal is not present.	Connect foot switch.
FF	FF Codes are 4-character codes beginning with FF that Seno uses to identify issues.	Call Seno Service for assistance at (888) 978-8835 should an FF status code appear.
Comm. Error	Communication with laser lost. This will cause a tint to appear to the icon.	Call Seno Service for assistance at (888) 978-8835.

14.6 Information Message

- Convey information of interest to the operator.
- Appear on the display monitor bottom left.
- Example Laser Heating Percentage or Reminder Everyone Wear Laser Protective Eyewear

14.7 System Messages

- System Messages appear on the display monitor top center.
- System Messages indicate an automatic response to address a potential condition.
- Record the message and call Seno Service for assistance at (888) 978-8835 for all of these messages except normal, operator initiated, and Imagio® Breast Imaging System power off.

Message	Type	Will the Imagio® Breast Imaging System Continue to Operate Correctly?	Actions to Take to Attempt to Resolve the Message
System is shutting down	Alert	Yes	Message will appear when the Imagio® Breast Imaging System is powering off. If normal power off has been attempted, no further action is necessary, and the Imagio® Breast Imaging System will resume normal operation when powered on. If the Imagio® Breast Imaging System is powering off but the operator did not attempt to power off the Imagio® Breast Imaging System call Seno Service for assistance at (888) 978-8835.
HIGH laser energy – (x) WAVELENGTH	Alert	No	Call Seno Service for assistance at (888) 978-8835.
UNABLE TO COMMUNICATE WITH LASER	Alert	No	Call Seno Service for assistance at (888) 978-8835.
LASER SHUT DOWN DUE TO HIGH OUTPUT POWER	Alert	No	Call Seno Service for assistance at (888) 978-8835.
UNABLE TO START LASER DUE TO HIGH TEMPERATURE	Alert	No	Call Seno Service for assistance at (888) 978-8835.
LASER HIGH TEMPERATURE SHUTDOWN	Alert	No	Call Seno Service for assistance at (888) 978-8835.
LASER SHUT DOWN DUE TO LOW OUTPUT POWER	Alert	No	Call Seno Service for assistance at (888) 978-8835.

14.8 General System Messages

The Imagio® Breast Imaging System general system messages utilizes either a pop-up window with suggested actions or a message in the information area of the main display monitor.

The medium priority messages are used to intervene automatically by halting the Imagio® Breast Imaging System upon detection of an alert condition that may result in harm to the patient, operator, or the Imagio® Breast Imaging System. The operator is instructed via a visual message. The course of operator's actions is limited to contacting a Seno representative and acknowledgment of the restarting the Imagio® Breast Imaging System.

The low priority messages are used as informational. The operator may be able to take action to resume Imagio® Breast Imaging System usage or the alert is strictly to inform the operator of a condition that may impact work flow.

The following alerts display pop-windows of a medium priority:

Alert	Title/message
HARDWARE ALERT	
HAL_ERROR_API	A problem in acoustic power intensity computation has occurred API error: please reboot and contact a Seno representative
Default case	Hardware control alert
Why is there nothing here?	Hardware controller board alert
HAL_ERROR_LOADING_SEQUENCE	A problem has occurred during sequence loading
SOFTWARE ALERT	
HWC_SP_MAY_HAVE_DIED	Hardware error: reboot is required Please contact a Seno representative
MGR_PROCESS_DEAD	An error occurred, the Imagio® Breast Imaging System will restart
MGR_PROCESS_NO_RESPONSE	Process Name not responding, wait a few seconds or force restart
HWC_OSAWA_WATCHDOG_USB	OSAWA watchdog error
HWC_SHORT_CIRCUIT	Short circuit error
HWC_US_BOARD_OVERHEAT	US board overheat
HWC_VOLTAGE_ERROR	Voltage error
HSPM_SOFTWARE_ERROR_SEIJI_ACCESS	SEIJI communication error
HSPM_SOFTWARE_ERROR_LOG_PROM_WRITE	Log EEPROM write error
HSPM_SOFTWARE_ERROR_LOGGING_INFORMATION	Log EEPROM logging information error
HSPM_SOFTWARE_ERROR_OSAWA_ERROR	OSAWA access error
HSPM_SOFTWARE_ERROR_MGR_MSG_TIMEOUT	Manager keep alive timeout
HSPM_SOFTWARE_ERROR_MGR_MSG_SEND_ERROR	Manager message send error
HSPM_SOFTWARE_ERROR_HSPM_START_ERROR	HSPM start error
HSPM_SOFTWARE_ERROR_OSAWA_WATCHDOG_ENABLE	OSAWA watchdog enable error
HSPM_SOFTWARE_ERROR_OSAWA_WATCHDOG_DISABLE	OSAWA watchdog disable error
HSPM_SOFTWARE_ERROR_EVENT_SERVICE_NOT_OPEN	SSI event service not opened
HSPM_SOFTWARE_ERROR_TEMPERATURE_SENSOR	Temperature alert
HSPM_SOFTWARE_ERROR_ECG	ECG monitoring error
MEAS_INVALID	ERROR

The following alerts display pop-windows of a low priority:

Alert	Title/message
SOFTWARE 3D ERROR	
E3D_CONTROLLER_ERROR_CONNECTION_INIT	The 3D controller is not connected
E3D_CONTROLLER_ERROR_NOT_CONNECT	
E3D_CONTROLLER_ERROR_NOT_GOOD_VERSION	The 3D controller is not identified
E3D_CONTROLLER_ERROR_USB_NOT_CONNECTED	The 3D motor controller USB is unplugged
E3D_CONTROLLER_ERROR_TIMEOUT	The 3D controller is not responding
E3D_CONTROLLER_ERROR_NONE	Try to change acquisition parameters
N/A	Try a new 3D acquisition
HARDWARE ALERT	
HAL_ERROR_IRQ_TIMEOUT_DMA_EXPORT	You may unfreeze to continue. DMA Sequencing warning - 0x000001
HAL_ERROR_IRQ_TIMEOUT_DMA_IMPORT	You may unfreeze to continue. DMA Sequencing warning - 0x000003
HAL_ERROR_IRQ_TIMEOUT_OTHER	You may unfreeze to continue. DMA Sequencing warning - 0x000003
HAL_ERROR_IRQ_TIMEOUT_TRIGGER	Press a mode button to continue. Trigger timeout - 0x000004
HAL_ERROR_INVALID_SQL	You may unfreeze to continue. DMA warning: invalid SGL Handle
HAL_ERROR_MAIN_LOOP_LOCK_TIMEOUT	You may unfreeze to continue. HWC main loop lock timeout
SOFTWARE ALERT	
COMPLUT_ALERT	Internal reconstruct LUT error
REVIEW_IMAGE_NOT_FOUND_ALERT	Review error: image cannot be loaded due to corrupted data
US_XML_VALIDATOR_ALERT	XML validator error: please contact an ultrasound engineer
MEAS_WARNING	ERROR
CONNECT_TOO_MANY_JOB	The job list contains more than 50 pending jobs. Please select and delete these jobs manually from the job list.
CONNECT_TOO_OLD_JOB	The job list contains some pending jobs older than one week. Please select and delete these jobs manually from the job list.
MGR_ENGINEERING_RELEASE	The software release that is currently running on the Imagio® Breast Imaging System is an engineering release that is not approved for human diagnosis.
ERROR_USB_TRANSFER	ERROR while transferring data to USB device
NOT_ENOUGH_SPACE	WARNING HARD DISK ALMOST FULL Please delete data on the Imagio® Breast Imaging System.
PARAMETER_BAD_USE	US parameter bad use

14.9 Main Display Notification Icons

Icons appear at the bottom left of the main display to indicate status or errors.

Icon	Meaning
	There is a USB mass storage device plugged in
	The Imagio® Breast Imaging System is exporting data to a USB device
	Error during USB export
	The Imagio® Breast Imaging System is connected to a network
	Network error
	The Imagio® Breast Imaging System is exporting data to a DICOM store SCP
	Error with the DICOM Modality Worklist SCP
	Error with the DICOM Store SCP
	The Imagio® Breast Imaging System is printing to a DICOM printer
	Error with a DICOM printer
	Error with a DICOM commit server
	The Imagio® Breast Imaging System is printing to a local printer
	Error with a local printer
	Modality Performed Procedure Step (MPPS) is busy
	Error with MPPS

Icon	Meaning
	DICOM Modality Worklist is refreshing
	Error with DICOM Modality Worklist
	DICOM Modality Worklist: Offline mode
	DICOM Modality Worklist filtered
	Check the air filters
	Available storage on the hard drive. <i>Do not try to acquire more data if the hard drive is full. Go to page XYZ to export existing data if full.</i>
	Query and Retrieve

15 Connecting and Disconnecting OA/US Probe, Ultrasound Probe, and Foot Switch

15.1 Removing and Connecting Probes

15.1.1 Removing/Connecting the OA/US L18-1 Probe

The OA/US L18-1 Probe should not have to be removed from the Imagio® Breast Imaging System. Examples of when the OA/US L18-1 Probe may need to be removed include situations when the Imagio® Breast Imaging System will be moved to a new location or if Seno has provided a new OA/US L18-1 Probe. Call Seno Service for assistance at (888) 978-8835 if needed.



Caution: *The Imagio® Breast Imaging System is designed for use with the Imagio® Breast Imaging System OA/US L18-1 and US L18-5 Linear Probes only. Other Probes will not be recognized or usable and may damage the system.*

- Do not attempt to attach any other probe to the Imagio® Breast Imaging System.

15.1.1.1 Connecting the OA/US 18-1 Probe

- Step: 1** Power off the Imagio® Breast Imaging System.
- Step: 2** Open the optic access panel located on the top right side of the cabinet.
- Step: 3** Feed the Optic Ferrule through the Optic Access opening on the front top right of the cabinet.
- Step: 4** Align the arrow on the Optic Ferrule to the top of the Laser Output Aperture.
- Step: 5** Gently push the Optic Ferrule into the Laser Output Aperture until it stops.
- Step: 6** Turn the Optic Ferrule in the direction of the arrow to lock it in place.
- Step: 7** Close the optic access panel.
- Step: 8** Push the ITT Cannon ZIF connector into the US connector port and lock handle left to secure.

15.1.1.2 Removing the OA/US L18-1 Probe

- Step: 1** Power off the Imagio® Breast Imaging System.
- Step: 2** Open the optic access panel.
- Step: 3** Turn the Optic Ferrule in the direction opposite of the arrow.
- Step: 4** Gently pull the Optic Ferrule out of the Laser Output Aperture.
- Step: 5** Feed the Optic Ferrule out of the Optic Access Opening located on the front top right of the cabinet.
- Step: 6** Unlock the lock handle to the right and gently pull the ITT Cannon ZIF connector from the port.

15.1.1.3 Optic Ferrule & Laser Output Aperture

Optic Ferrule with connecting/disconnecting arrow.



Optic Ferrule inserted into the Laser Output Aperture.


15.1.2 Connecting an Ultrasound Only Probe

Push the ZIF connector into its the compartment and push the lock handle all the way to the left to lock the connector.

Important: Prior to connecting a probe, ensure the lock is in the unlocked position (pushed to the right).



If a probe is not connected properly, an icon is displayed on the probe tab as follows:



When disconnecting the probe, push the lock all the way to the right to unlock the connector and then hold the connector and pull it straight out.

15.2 Cabinet Storage Areas

These areas may be used to store basic items such as notebooks, charts, and towels.



Caution: Liquids stored or placed on the Imagio® Breast Imaging System, outside of the recommended ultrasound gel, may spill and damage the Imagio® Breast Imaging System components rendering the device unusable.

- Do not place cups, drinks, or other fluid containers on the Imagio® Breast Imaging System.

15.3 Input/output Connectors

15.3.1 Removing/Connecting OA Foot Switch

15.3.1.1 To Remove the Foot Switch

- Disconnect the Foot Switch connector by pulling the outer ring.

15.3.1.2 To Connect the Foot Switch

- Line up the red orientation dot of the cable with the red notch on the connector port. Insert by pushing the connector in until the outer ring locks in place.

15.3.2 USB Ports



Caution: Externally powered USB devices are not permitted to be connected to the USB ports of the Imagio® Breast Imaging System. Any USB device may be connected to the system as long as it is SELF-powered through the USB port (two USB ports can be used) or powered from an external supply IEC 60601 (1 Means of Patient Protection MOPP) compliant.

The following ports are available on the system:

- Two USB ports are located on the left side of the console

15.4 Additional Video Devices

You may want to attach an additional video device such as an LCD monitor or video projector to the Imagio® Breast Imaging System.

The video output of the Imagio® Breast Imaging System is DisplayPort, with a native resolution of 1920 x 1080. You may connect video devices to the Imagio® Breast Imaging System which are DisplayPort compatible with the port type and support this video resolution. The external video device must comply with the IEC 60601-1.

To connect your video device with the Imagio® Breast Imaging System, perform the following steps:

- Power your external video device ON.
- Connect your external video device to the Imagio® Breast Imaging System by means of a DisplayPort cable no longer than 5 m* (15 ft*) in length.
- Turn on the power of the Imagio® Breast Imaging System. The main display will be mirrored on the system display and on the external monitor.

Note: External video devices which do not support the resolution of 1920x1080 may result in the following:

1. No signal displayed on the external video device
2. A "letterbox" effect where the video display is shown with a black border around the edges.
3. An aspect ratio adjusted signal is displayed. This may result in the image being stretched or compressed to fit the display.

If problems with the video quality or aspect ratio are observed, it is most likely caused by external video devices which do not support the native resolution of 1920x1080. For more information, call the Seno Service for assistance at (888) 978-8835.

16 System Configuration

The Imagio® Breast Imaging System can be customized in several ways to make it more useful for an operator's needs.

Presets designed specifically for the exams performed can be created, and Imagio® Breast Imaging System settings can be changed to optimize work flow.

System Configuration enables configuring of the entire Imagio® Breast Imaging System, its applications, and additional options.

To access System Configuration, touch the top left Imagio® logo on the Touch Screen while in any active imaging mode.

Any changes made to the System Configuration are automatically saved, unless otherwise specified.

Touch Image on the Touch Screen to exit System Configuration. A dialog may appear indicating the Imagio® Breast Imaging System is restarting due to configuration change. Accessible configuration features are dependent on log on credentials. Refer to "Login and Lock/Logout" on page 29.

The System Configuration displays:

- OA Settings configuration button
- System/Display tab
- Device Settings tab
- Administration tab
- Presets tab
- Measurements tab
- Online Services tab
- System Diagnostics tab

Refer to each section for detailed information.





Warning: *The Imagio® Breast Imaging System allows operators to configure Operator-defined Measurements, Calculations and Tables for diagnostic purposes, errors may lead to misdiagnosis or delay in treatment.*

- Seno does not endorse operator-defined Measurements, Calculations or Tables. These are utilized at the Operator's discretion and risk only.

16.1 OA Settings

OA Settings button is separate from the configuration tabs and located on the top right. Select the OA Settings button to enter the OA Settings configuration.



16.1.1 Laser Enable Duration

Laser Enable/Disable inactivity timeout occurs when there is no operator input detected. 15 minutes is the recommended duration and the default setting. Range is 1 to 30 minutes.

Image Parameters Prospective OA Clip Length may be adjusted from 3-15 seconds.

Cine or retrospective recording time is fixed at 30 seconds.


16.1.2 Laser Inactivity Sleep Time

Laser Heating inactivity timeout occurs when there is no operator input detected. The Laser will discontinue heating and enter laser sleep. 60 minutes is the recommended duration and default setting but can be adjusted to a longer duration to keep the Imagio® Breast Imaging System available for immediate use. Range is 60 to 540 minutes.

16.1.3 Registered and New Authorization Cards

The Imagio® Breast Imaging System authorization card database is initially empty when shipped. Seno Service will place the Admin authorization card on the System Console NFC reader. A message “Laser Admin Mode Enabled” appears.

Initially, the Imagio® Breast Imaging System will have no entries in the “Registered Authorization Cards” list. Level Operator Admin is selected by default and the message “Laser Admin Mode Enabled” is shown on bottom right. The first card added is Laser Admin.



OA Settings

Laser

Laser Enable Duration (min): 15
Range is 1 to 30

Laser Inactivity Sleep Time (min): 60
Range is 60 to 540

Registered Authorization Cards

ID	First Name	Last Name	Level
1 67675aa9	Admin	Guy	Laser Admin

Remove Card

New Authorization Card

First Name: Admin

Last Name: Guy

Card ID: 67675aa9
Scan card to populate

Level:

☐ Laser User

☒ Laser Admin

*Laser User can enable laser, Laser Admin can add/remove authorization cards

Add Card Laser Admin Mode Enabled

“Laser Admin Mode Enabled” permits any number NFC cards to then be added, without having to read a card prior to entry, either as a “Laser Admin” or “Laser Operator”.

Level:

- Laser Admin - Permits modifying registered cards, (unable to authorize laser).
- Laser Operators - Permits authorizing laser.

Read a card to populate the “Card ID” field and enter the desired “First Name” and “Last Name” entries.

Once the required fields (First Name, Last Name, and Card ID) have been populated then the Add Card button will be active. NOTE: make sure to select “Laser Admin” for the card level, otherwise the default selection is “Laser Operator”.

- Click “Add Card” to add the new Laser Admin card to the Registered Authorization Cards list.

Once a Laser Admin card has been enrolled then that card must be used in the future to enable Laser Admin Mode to remove and/or add new authorization cards to the Registered Authorization Cards list. Admin cards cannot be used to activate the laser; a separate authorization card is required.


After a “Laser Admin” card is read and the “Laser Admin Mode Enabled” text appears, repeat the process of reading a new Card ID and entering the “First Name” and “Last Name” to add “Laser Operator” or “Laser Admin” cards.

16.2 System/Display Tab

The System/Display tab is divided into tabs:

- System
- Regional parameters
- Exam
- Customization

16.2.1 System Configuration



Options:

- Enter the institution name and address
- Choose the patient name order

- Display/not display the sonographer's name in the image header
- Select the Thermal Index (TI) to display on the imaging screen
- Upload from a USB media and insert a logo which will be displayed on the report header or choose the default logo to be displayed on the report header
- Adjust the audio volume of the Imagio® Breast Imaging System


Note: To change the logo to be displayed in the printed report:

Make sure the «Display logo in report» check box is checked.

Do one of the following:

- Select «Use the default logo» if using the Imagio® Breast Imaging System logo
- Select «Use the custom logo» if using your own logo
 - Then plug a USB device with the logo into the Imagio® Breast Imaging System
 - Then select the logo to browse for your own logo
 - Enable/disable the screen saver, and configure it
- Adjust the system console brightness
- Adjust the Touch Screen brightness
- Activate a sound while for the following functions:
 - After each Save Image/Clip
 - For the Touch Screen Keyboard
- When the Imagio® Breast Imaging System boots up

16.2.2 Regional parameters



This section refers to the regional settings and language configuration of the Imagio® Breast Imaging System. Options:

- change the Imagio® Breast Imaging System interface language
- set the keyboard language

- change the date and time format
- adjust the Imagio® Breast Imaging System's date and time.

Note: To set date and time:


Step: 1 Select a time zone from the drop-down list

Step: 2 Do one of the following:

- Enter a Network Time Protocol (NTP) server (by default, enter ntp.ubuntu.com), and Press "Synchronize with NTP" in the Automatic setting box
- Pick a date and set the time in the Manual setting box, then press "Apply manual changes"

Note – the date and time cannot be adjusted while an exam is in progress.

16.2.3 Exam




Options:

- Customizing scanning preferences:
 - Adjust some preferences for the Color and PW modes
 - Steering Angle Step defines the incremental step for the steering of the color box
 - PW Fine Angle Step defines the incremental step for the Fine Angle Correction in PW
 - Choose to automatically invert the color flow map with the steering direction
 - Customize the knob rotation direction for the steering angle, focal zone, and depth
 - Configure the duration of the Manual TGC auto display on the Touch Screen after activating the Auto TGC. By default, the auto display is off.
- Customize some Freeze options:
 - Set time for the live mode's auto freeze
 - Choose the action to perform after freezing the image:
 - only freeze
 - display body markers
 - display annotations
 - display measurements

- Configure the limit duration of the prospective clip capture, for contrast mode and for the other modes.
- Customize the way the measurements are displayed
- Customize the Continue exam mode
- Choose to have a warning popup when saving an image for an anonymous patient
- Choose the Focal Zone Management to be manual or automatic
- Choose to activate or deactivate the Adaptive TGC mode
- Choose to enable looping in clips
- Customize the rotation direction for the Depth and Focus knobs
- Choose to send enlarged images to a printer or USB device (Enlarge mode). Check to flag an image to be exported at 133% zoom. Only images in single layout mode (like standard B-mode or the SWE mode with only SWE shown) support this feature. OA and SWE/B-mode dual top/bottom are examples of layouts that do not export as “enlarged” even if this setting is selected.

16.2.4 Customization



Options:

- Assign the system console buttons to desired software functions
- Configure the Touchpad gestures

16.2.4.1 System Consoles Keys

Assign software function position as desired. Options:

Note: Please note that only the key position (located on both sides of the Touchpad) can be changed.

Select the labeled buttons (which are the currently assigned buttons) and change the associated software control as needed.

Assign a function to the S1 and S2 keys.

Select the buttons and choose the desired function.

Customize the shortcut to access Body Markers or Annotation when pressing Annotate.

Select the drop-down list and choose to access to Body Markers or Annotation first when pressing Annotate on the system console.

16.2.4.1.1 Customizable Buttons

Control panel buttons are customizable according to your needs. You can assign functions that are not available on the control panel to S1 and S2 buttons in the System configuration. You can also switch button location to your convenience by doing the following: Put a screwdriver in one of the small notches located on each side of the buttons to remove them, then put them at the desired location.



16.2.4.2 Touchpad Gestures

Activate or deactivate gesture control according to operator's needs. To do so, check or uncheck the check boxes.

16.2.4.3 Foot Switch

Assign software functions to both foot switch pedals.

Select the drop-down lists to assign desired function to the left or right pedal.


16.3 Device Settings

16.3.1 Description of the Device Settings Tab

The Device Settings tab is divided in 3 tabs:

- Media settings
- Media association
- DICOM options

16.3.2 Media Settings



Configure all the media already added and associated. Options:

Select Media Settings to see the list of active devices.

For DICOM Stores devices, a menu enables defining of the method to send the data:

- Send after acquisition
- Send on end exam
- Send manually


Defining printing parameters for DICOM printers, Reports Printers, and Image printers, as well as JPEG/ AVI Removable Media Export parameters (Compression level, export LUTs (Look Up Tables),) are options.

A dedicated “Worklist Settings” menu is available for the DICOM Modality Worklist servers. Set the way of querying the Modality Worklist SCP:

- By Modality
- By AE Title
- By Date

and set the maximum responses to receive from the Modality Worklist server(s) when this device is available and configured in the Administration part.

16.3.3 Media Association



Associate any media added in the tab Administration.

Options:

This section enables adding of an active device to the working configuration.

Choose from the device list which has been pre-configured in the Administration part of the System Configuration.

To associate a device:

Step: 1 Navigate to the bottom of the screen to the “available devices” list.

Step: 2 Select the type of device, a list will be displayed.


Step: 3 Select the specific device to associate.

Step: 4 Press “associate”. The device will appear at the top of the screen in the “associated devices” list.

Note: Associate one or several DICOM Worklist Servers, to be able to query several Worklist servers at the same time.

Note: Remove a device from the associate list by selecting it and pressing the ‘unlink’ button.

16.3.4 DICOM options



Adjust the DICOM parameters. Options:

In the Main DICOM Options part, enter all the DICOM parameters that identify the Imagio® Breast Imaging System on a DICOM network.

Configure the Imagio® Breast Imaging System to communicate with known hosts .

Set TLS configuration to be the same for all services (Store SCP, DICOM Store, DICOM Printers...) of the Imagio® Breast Imaging System, by checking the Use global TLS configuration check box. The following options are also available:

- Press “Export our Certificate” to export the public key on a USB device
- Choose to use an anonymous TLS connection by checking the dedicated box
- Require a Peer certificate by checking the dedicated check box
- Choose the TLS security profile, between DICOM AES or DICOM BCP 195
- Press “Regenerate Key/Certificate” to generate a new Key/Certificate pair

To configure the TLS connection differently for each service, go to the Administration tab, Devices subtab refer to Devices to configure the different services as needed.

Important: Please be aware that the confidentiality and integrity of the Patient Identifying Information (PII) transmitted over non-DICOM TLS communication cannot be ensured.

In the Modality Worklist Options part, configure the way the Modality Worklist is refreshed with the “polling option”:

- Choose “Manual” to refresh the Worklist only manually
- Choose “Automatic” to set up a time to automatically refresh the Worklist
- Choose “Smart” to automatically refresh the Worklist each time End Exam is pressed on the system console.

Choose to use the Off-line mode, so that when the network is disconnected, the Worklist is still accessible, and by default this mode is on.

Set work flow preference for the end of the exam:

- Display the Modality Worklist
- Display the Patient Data Entry

Select the way of querying the Worklist when pressing End Exam in the “Query preference” drop-down list:

- Enable the Broad Query option
- Enable the Patient Query option
- Maintain the last selected option

In the Store SCP Options part, Choose the deletion method and whether or not to retrieve the multiframe images.

- In the Q/R SCU options part, configure the Query/Retrieve parameters.
- Configure the automatic query and automatic retrieve by selecting the queried fields and modalities.

16.4 Administration


16.4.1 Description of the Administration Tab

The Administration tab is divided into sub-tabs:

- Devices
- Ethernet configuration
- Wi-Fi Configuration
- Network tools
- Disk maintenance
- Medical Staff
- Operator Management

Important: Image compression may result in the loss of image information. While low levels of image compression are generally acceptable in medical imaging, using high levels of lossy compression may result in image degradation. It is the responsibility of the operator to set and maintain the degree of image compression which is diagnostically acceptable in exported images.

16.4.2 Devices



Devices refers to the initial device’s configuration.

A list of device types that can be added is displayed on the left side of the screen, available according to the options purchased:


- Add DICOM Store
- Add DICOM Modality Worklist
- Add DICOM Modality Performed Procedure Step (MPPS)
- Add DICOM Printer
- Add DICOM Query/Retrieve
- Add image Printer

Note: Use the recommended driver for the selected printer.

Press the appropriate button on the left to add a device and configure its parameters. Select a device in the middle of the screen to view and configure its parameters.

Note: When adding any DICOM device using the DICOM network connection, the following Basic Options must be fulfilled:

- Station Name, IP Address, Port Number and Application Entity Title (AET)
- Use a TLS connection for these devices when adding or editing them.



Please refer to the DICOM options for more information about DICOM TLS.

16.4.2.1 Configure a DICOM Store

To edit an existing DICOM Store device, select it on the devices list then press Edit Selected.

On the Basic Options tab, configure the DICOM device.

On the Advanced Options tab, manage the exportation settings.

1. In the General Export Settings choose:
 - to choose the patient's name to be shown or hidden on the exported files
 - the types of data to enable for exportation (single or multi frames, PDF reports)
 - the Look Up Table (LUT) to apply on the exported images/video clips
 - the type of character encoding
2. In the Pictures and Clips Settings, define:

- the transfer syntax with pre-set image compressions
- the compression quality of the images
- the support of Color or Monochrome modes at the exportation
- the level of image size reduction
- the duration of the retrospective clip

Note: Configure all these options the same way for any DICOM Removable Media.

In the Commit Option tab, choose to ask for a commitment from the storage service.

When all the desired options have been set, press OK to save the changes.

Note: The nine available LUTs are pre-set filters acting on specific Brightness and Contrast settings of the exported images and/or clips.

They are used to make the images appear on a DICOM Review workstation Monitor as they appear on the Imagio® Breast Imaging System Monitor.

Important: Applying a too bright or too contrasted LUT can have outcomes on the quality of the image's information. It is the operator's responsibility to carefully adjust all imaging parameters to avoid image saturation and obtain optimal quantitative time-intensity data.

16.4.2.2 Configure a DICOM Printer

To edit an existing DICOM Printer, select it on the devices list then press Edit Selected.


On the Basic Options tab configure the DICOM Printer.

On the Advanced Options tab manage the printing settings:

- adjust the brightness and contrast
- choose the printing media type
- define the priority, the destination, and the magnification type
- define the color and density of the borders and empty spaces of the document
- precise the session label, the configuration information, and the smoothing type
- define the reduction percentage of document

When all the desired options have been set, press OK to save the changes.

16.4.3 Ethernet Configuration



This section concerns the network configuration of Imagio® Breast Imaging System.

The Imagio® Breast Imaging System shall be connected to the network through an ethernet cable. The Imagio® Breast Imaging System can be configured with a specific IP address and network setting, with a cable connected or can be automatically configured via Dynamic Host Connection Protocol (DHCP) .

For any network configuration, the MAC address is displayed and once connected, the network status is also displayed (especially important for the DHCP configuration).

Note: The Imagio® Breast Imaging System network connection is only opened for DICOM and network printer service.


Note: The Imagio® Breast Imaging System does not require a minimum data rate to perform on the network.

Note: In case of network connection failure, the data can be exported using a DICOM removable media.



Caution: Integrating the Imagio® Breast Imaging System to a network with other connected devices may result in unpredictable risks. The operator is not permitted to install custom anti-virus software.

16.4.4 Network Tools




Once the network is configured, refer to “Ethernet Configuration” on page 150, these network tools are available:

- Ping
- Traceroute
- Dig and Reverse Dig
- MTR
- Duplicate address detection DAD
- IPcalc

Call Seno Service for assistance at (888) 978-8835 for further instructions on the use of these tools.

16.4.5 Disk Maintenance



This section concerns all of the options to manage the exams contained on the Imagio® Breast Imaging System hard drive.

Important: *The Imagio® should not be used as the primary source for long term storage.*

- Exams should be backed up either to an external drive or via network on a regular basis. If the Imagio® Breast Imaging System's HDD needs to be replaced or the Imagio® Breast Imaging System is upgraded, the HDD data will be erased and unrecoverable.

16.4.5.1 Automatic Local Image Removal

This section enables managing of the way to purge the local exam database.

Delete manually is the default option, it enables manual deleting of any exam from the Patient Directory.

Important: *Delete exams from time to time to prevent the HDD to be full.*

- Delete DICOM stored exams requires the Imagio® Breast Imaging System to be connected to a PACS. Configure the Imagio® Breast Imaging System to automatically delete exams already stored on the PACS, 24h, 2 days, 3 days, 5 days, 10 days, 30 days, 60 days, or 90 days after their creation.
- Delete DICOM stored & committed exams requires the Imagio® Breast Imaging System to be connected to a PACS. Configure the Imagio® Breast Imaging System to automatically delete exams already stored and committed on the PACS, 24h, 2 days, 3 days, 5 days, 10 days, 30 days, 60 days, or 90 days after their creation.
- Delete oldest exams automatically deletes exams 24h, 2 days, 3 days, 5 days, 10 days, 30 days, 60 days, or 90 days after their creation.
- Delete oldest exams to maintain sufficient free space on hard drive automatically deletes oldest exams after the data storage reaches less than 20Gb of free space. The Imagio® Breast Imaging System will delete exams until there is at least 50Gb of free space.

16.4.5.2 Backup/Restore

Backup and restore some elements from the Imagio® Breast Imaging System.

16.4.6.2.1 Back up

Step: 1 Plug a USB device to back up some elements from the Imagio® Breast Imaging System

Step: 2 Press the Backup/Restore button

Step: 3 Enter the password

Step: 4 Press OK

Step: 5 Select the USB device

Step: 6 Select the elements to back up

Step: 7 Press Start

Important: *Make sure to back up data from time to time to prevent the HDD from becoming full. Back up following the instructions above, or via DICOM.*

- Imagio® Breast Imaging System is an imaging modality device and shall not be used for archiving. In order to minimize the consequences of any security event, Seno strongly recommends archiving images to a PACS. In case of Imagio® Breast Imaging System failure, please note that the Imagio® Breast Imaging System disks are encrypted, which could make data recovery difficult.

16.4.6.2.2 Restore

Step: 1 Plug the USB device which was backed up with some elements of the Imagio® Breast Imaging System

Step: 2 Press the Backup/Restore button

Step: 3 Enter the password

Step: 4 Press OK

Step: 5 Select the USB device which was backed up with some elements of the Imagio® Breast Imaging System


Step: 6 Select the Restore tab

Step: 7 Select the backup file to restore on the left

Step: 8 Select the elements to restore on the right

Step: 9 Press Start

16.4.6 Medical Staff



System Configuration

System / Display | Device Settings | **Administration** | Presets | Measurements | Online Services | System Diagnostics

Devices | Ethernet configuration | Wifi configuration | Network tools | Disk maintenance | **Medical staff**

Last Name*
First Name*
Middle Name
Abbreviation
Name prefix
Name suffix

DICOM ID Sequence

Code value
Coding scheme designator
Coding scheme version
Code meaning

Name shall appear under: ☐ Performing physician / sonographer ☐ Referring physician ☐ Requesting physician

*Last Name or First Name is required
*Abbreviation of the performing physician / sonographer will be displayed during the exam and on the report

Name list

Name	Abbreviation

New Remove

This section enables managing of the list of physicians and sonographer's that are available to choose from in the Patient Data Entry.

In the list at the bottom of the screen, press a name to display the related information in the appropriate fields above the list.


To add a name:

Step: 1 Select New on the right side of the list

Step: 2 Enter the appropriate data

The new name is automatically saved. They can be selected from the Patient Data form.

16.4.7 Operator Management




In this section, Administrators can modify operator-related settings.

Step: 1 Select a profile.


- Operator information are now displayed.

Step: 2 Select Edit to modify or edit settings.



By doing so, the following settings can be changed:


- Auto-login: by enabling auto-login, operators won't have to enter their credentials at the Imagio® Breast Imaging System startup. Check or uncheck the dedicated check box to enable or disable auto-login.



Note: Please note that even if the auto-login is enabled, password entry is required when unlocking your session after a session lock. The auto-login can only enable one profile at a time.

Important: Please be informed that enabling the auto-login feature will alter the security parameters of the Imagio® Breast Imaging System.

- To change the password of the selected profile, follow the steps.
- Press Change Password to save the new password.



16.5 Presets

A Preset is a group of settings that optimizes the Imagio® Breast Imaging System for a specific type of exam. Presets establish many initial settings, such as gain value, color map, filter, TissueTuner™, or Flow Optimizations. When the Imagio® Breast Imaging System is turned on, the default Preset is active. Before beginning an exam, be sure that the appropriate Preset is active. Choose from several default Presets. These

default Presets cannot be deleted. Several Presets per probe/application combination can be created and stored, depending on the number of buttons available on the individual probe Touch Screen.


16.5.1 Description of the Presets Tab

The Presets tab is divided in 3 tabs:

- Annotations
- Bodymarkers
- Presets

It enables customization of annotations and bodymarker packages, as well as imaging presets.

16.5.2 Annotation



This section enables editing of the annotation library for the current language. To edit the annotation library:

Step: 1 Select a package

Step: 2 Select the page to edit by using the navigation controls, located above the annotation grid:

- First page/Last page
- Previous page/Next page
- Add page
- Remove page

Step: 3 To create a new package, select New Package and enter a name in the pop-up window.

Step: 4 To duplicate an existing package, select Duplicate Package and enter a new name in the pop-up window. A new package will be created with all the annotations from the initial package without modifying it.

Step: 5 To change the name of an annotation, select it and change it in the pop-up window

Step: 6 To create a new annotation on an empty button, select the empty button and enter its name in the pop-up window

Step: 7 To remove an annotation from a group, select Free Annot (the first color box on the left side) and select the annotation to remove from its group. The annotation button becomes “transparent” (no color) and will not be replaced by another one

Step: 8 To move an annotation from one group to another, select the new group color to apply, and select the

annotation. The annotation will change color


Step: 9 To change a rotary knob label or create a new one, select the Edit Title button located next to the knob to edit, and change it in the pop-up window

Step: 10 To delete a rotary knob value, select the value and then select the - button next to it

Step: 11 To add a rotary knob value, type the new value in the box and select the + Add to the list button next to it

Step: 12 To move a rotary knob value within the list, select it and then select the arrows next to it

16.5.3 Bodymarker



This section enables editing of the body marker libraries and to customize some options related to bodymarkers.

16.5.3.1 Options

Choose to remove the body marker every time the image is unfrozen by checking the “Remove bodymarker on unfreeze” option.

Choose to remove the probe marker from the body marker every time the image is unfrozen by checking the “Remove probe marker on unfreeze” option.

Set up a time out for the body marker Touch Screen. At the end of this time out, the body marker TouchScreen will be automatically closed.

16.5.3.2 Body Marker Library Customization

16.5.3.2.1 To view the body marker library available for a given preset

Step: 1 Select the package to modify on the right part

- Navigate through the different pages of the library by using the following buttons:
 - First displays the first page of the library
 - Previous displays the previous page of the library
 - Next displays the next page of the library
 - Last displays the last page of the library

Add and remove pages for the selected library by using the Add and Remove buttons.

16.5.3.2.2 To remove a body marker from the package

Step: 1 Select the body marker to remove

Step: 2 Select the Remove button below the arrow

16.5.3.2.3 To move a body marker on the package page

Step: 1 Select the body marker to move

Step: 2 Select the appropriate arrow

16.5.3.2.4 To set a given body marker as default for the package

Step: 1 Select the body marker to set as the default body marker for the library

Step: 2 Select Set as default below the arrows

The selected body marker will be displayed by default every time Body Marker is pressed on the system console for the selected preset.

16.5.3.2.5 To add a body marker to a package

All the available libraries on the Imagio® Breast Imaging System are displayed on the left part.

Step: 1 Select the appropriate package the available body markers are displayed below.

Step: 2 Select the body marker to add to the library

Step: 3 Select the Add button

The selected body marker will be added at the first available space in the library.

Note: all changes are automatically saved.

Reset a body marker library for a given preset as it was by default by selecting Reset to default.

16.5.3.2.6 To create a new body marker

Step: 1 Select New and enter the name of the package in the pop-up window.

Step: 2 Choose bodymarkers in the library and add them to the package. Refer to To Add a Body Marker to a Package.


16.5.3.2.7 To duplicate a package

Select Duplicate and enter a new name in the pop-up window.

A new package will be created with all the bodymarkers from the initial package without modifying it.

To add a bodymarker from another package, select the appropriate package and proceed as if adding a bodymarker.

16.5.4 Imaging Presets



16.5.4.1 Creating a Preset

Step: 1 Press Probe on the system console

Step: 2 Touch New Preset

- A dialog box will appear on the main display.

Step: 3 Select the application in which is needed to save the preset

Step: 4 Select an annotation library for the new preset

Step: 5 Select a body marker library for the new preset

Step: 6 Select a measurement library for the new preset

Step: 7 Using the keyboard, enter a name for the new Preset.

Step: 8 Select OK.

Step: 9 Select Probe on the system console to verify the Preset.

The operator-created presets appear in light gray on the Touch Screen. The factory presets appear in dark gray on the Touch Screen.

16.5.4.2 Managing Presets

Manage imaging presets in the Presets sub-tab.

Step: 1 Select the appropriate probe and application on the left side of the screen.

The available presets for this probe and this application will appear.

16.5.4.2.1 Adding presets

Selected Presets lists the presets as they will appear on the Touch Screen.

Step: 1 Select the preset needed to add from the Available Presets list on the left

Step: 2 Select the right arrow to add it to the Selected Presets list

16.5.4.2.2 Removing presets

Selected Presets lists the presets as they will appear on the Touch Screen.

Step: 1 Select the preset needed to remove from the Selected Presets list

Step: 2 Select the left arrow to add it to remove it

16.5.4.2.3 Changing the display order of the presets on the Touch Screen

Step: 1 Select the preset that needed to modify in the Selected Presets list.

Step: 2 Move a preset within the list with the Up and Down buttons.

They will be displayed in the same order on the Touch Screen preset page.

Note: To reset a factory preset to the default configuration, select Reset.

16.5.4.2.4 Modifying the annotation package(s) of a preset

Step: 1 Select the small annotation tab on the right side of the screen.

- You will see the available packages box on the left of the Add button and the selected packages box on the right.
- The selected packages are those already available in the preset.

Step: 2 To add an annotation package, select the appropriate package in the available packages box and select Add.

- It will appear in the selected packages box.

Step: 3 In the selected packages box, remove a package by selecting Remove.

Step: 4 Move a package within the list using the Up and Down buttons.

Step: 5 To edit the annotation package, select a package in the selected packages box and select Edit Package.

16.5.4.2.5 Modifying the body marker package(s) of a preset

Step: 1 Select the small bodymarker tab on the right side of the screen.

- You will see the available packages box on the left of the Add button and the selected packages box on the right.
- The selected packages are those already available in the preset.

Step: 2 To add a bodymarker package, select the appropriate package in the available packages box and select Add.

- It will appear in the selected packages box.

Step: 3 Select Remove in the packages box to remove a package.

Step: 4 Move a package within the list using the Up and Down buttons.

Step: 5 To edit the bodymarker package, select a package in the selected packages box and select Edit Package.

16.5.4.2.5.1 Modifying the measurement package(s) of a preset

Step: 1 Select the small measurement tab on the right side of the screen. You will see the available packages box on the left of the Add button and the selected packages box on the right.

- The selected packages are those already available in the preset.

Step: 2 To add a measurement package, select the appropriate package in the available packages box and select Add. It will appear in the selected packages box.

Step: 3 Select Remove in the selected packages box, to remove a package.

- Move a package within the list using the Up and Down buttons.

Step: 4 Select a package in the selected packages box and select Edit Package to edit the measurement package.

16.5.4.3 Changing a Preset Name

Step: 1 Do one of the following:

- Press Probe on the system console, then touch Manage Presets on the Touch Screen
- Touch Syst. Config. on the Touch Screen, then select Presets and Presets

Step: 2 Choose the probe and the application

- The list of operator-defined Presets appears.

Step: 3 Select the Preset needed to rename and select Rename

- A popup menu appears
- Step: 4** Enter a new name **Step: 5** Select OK to validate

16.5.4.4 Deleting a Preset

Delete the Presets created. The factory presets cannot be deleted.

Step: 1 Do one of the following:

- Press Probe on the system console, then touch Manage Presets on the Touch Screen
- Touch Syst. Config. on the Touch Screen, then select Presets and Presets

Step: 2 Choose the probe and the application

- The list of operator-defined Presets appears.

Step: 3 Select the Preset to delete and select Remove.

- A popup menu appears.

Step: 4 Select OK to validate

16.6 Measurements

16.6.1 Description of the Measurements Tab


The Measures tab is divided into tabs:

- Measures
- Labels and Calculations
- Packages

16.6.2 Measures

This section enables customization of the measurement tools.


Step: 1 Select a measurement in the left list



Step: 2 Modify the unit and the precision. Select the unit or the precision needed to modify and select the appropriate unit in the drop-down list.

Step: 3 Check the measurements that are needed to be displayed on screen and/or included in the report.

16.6.3 Labels and Calculations



This section enables configuration of each labeled measurement individually.

Step: 1 Select the appropriate package

Step: 2 Select the labeled measurement to customize

Note: You may use the white area as a search box to easily find a labeled measurement in the list.


- Select the broom icon to erase the text in the search box.

The information and options related to this labeled measurement appear on the right.

Step: 1 Modify the unit and the precision. Select the unit or the precision needed to modify and select the appropriate unit in the drop-down list

Step: 2 With the drop down, choose the measurements that are needed to be displayed on screen and/or included in the report.

16.6.4 Packages



This section enables configuration of the labeled measurements packages.

The left side box contains all the labels available on the Imagio® Breast Imaging System.

The right-side box contains the labels available for the selected application, as it appears on the main display when pressing Measurements on the system console.

Edit the labels that are available for a given application.

[16.6.4.1 To add a labeled measurement to a package:](#)

Step: 1 Select the package to modify on the right side of the screen

Step: 2 Select the package that contains the labeled measurement on the left side of the screen. Select **All** to see the labeled measurements available on the Imagio® Breast Imaging System.

Step: 3 Select the labeled measurement to add from the left side list

Step: 4 Select Add

[16.6.4.2 To remove a labeled measurement from a package:](#)

Step: 1 Select the package on the right side of the screen

Step: 2 Select the labeled measurement needed to delete from the right-side list

Step: 3 Select Remove

[16.6.4.3 To move a labeled measurement within the list in a package:](#)

Step: 1 Select the package

Step: 2 Select the labeled measurement needed to move from the right-side list

Step: 3 Select Move up or Move down until it reaches the desired place

[16.6.4.4 To rename a package:](#)

Step: 1 Select the package on the right side of the screen

Step: 2 Select Rename

Step: 3 Enter a new name in the pop-up window

[16.6.4.5 To reset a package to its default configuration:](#)

Step: 1 Select the package on the right side of the screen

Step: 2 Select Reset

[16.6.4.6 To create a new package:](#)

Step: 1 Select New

Step: 2 Enter a name in the pop-up window and select OK

Step: 3 Add labeled measurements to the new package. It will be available in the presets.


[16.6.4.7 To duplicate a package:](#)

Step: 1 Select Duplicate

Step: 2 Enter a name in the pop-up window and select OK


A new package will be created with all the labeled measurements.

16.7 Online Services



The Online Services tab is intended to be accessed by service personnel only. Password access to the some of the contents is required to configure the online services.

16.8 System Diagnostics



The Login area of the System Diagnostics tab provides access for authorized Seno Service personnel to perform system diagnostic tests.

A password is not required to access some basic Imagio® Breast Imaging System information, as follows:

- Product name, Software version, File format version, Ebox number, Cart number, Serial number, List of enabled options

In addition, the following buttons are available for the Operator to export logs to USB:

- 24h: export logs from the last 24 hours
- All: Export logs since installation or since logs were last cleared.

Note: Two separate sets of system logs are generated. See the Notifications Window section for instructions

for the second log export.

Reset network configuration is a password-controlled feature for Service personnel only.

16.9 Remote Support


The Remote Support feature allows service personnel to remotely access the system from an off-site support location through the Internet. The Remote Support feature:

- Requires that the Imagio be connected a network (≥ 10 Mbps) with Internet access
- Requires that the local user at the Imagio site to initiate the connection
- Requires the local user to confirm that they are responsible for the system during the support session
- Requires the local user to provide a system-generated (random) Session ID to the support personnel through another communication channel (e.g. phone at (888) 978-8835, email at service@senomedical.com)
- Requires that the service personnel have credentials for the session
- Uses the Secure Shell (SSH) protocol to encrypt the connection
- Always displays an option for the local user to terminate the connection
- Will automatically terminate the connection if the session is inactive for one hour (no cursor movement detected)
- Does not allow the laser to be fired remotely


To begin a Remote Support session, navigate to an Imaging screen, click “Notifications” on the Laser Status Bar, then “About”, then “Start Remote Support”:



Upon clicking “Start Remote Support”, a dialog will appear:




If the system successfully connects to the Remote Support server, several seconds after clicking “Agree”, a status area will appear at the top of the screen indicating the session is now in progress. “Waiting for Connection” means the remote support personnel has not yet connected, but the machine is open to a connection.



The local user must provide the Session ID number to the service personnel through another communication channel (e.g. phone, email) to enable the remote support connection.

When the service personnel successfully connects to the machine, the status area will change from “Waiting for ” to “Connected”:



The service personnel can now remotely access the machine. The local user may stop the session at any time, or the session will automatically stop if left idle (no cursor movement) for one hour.

A Troubleshooting

A.1 Troubleshooting

If difficulties are encountered in the operation of the system, use the information in this section to help correct the problem.

If the problem is not covered here, call the Seno Service for assistance at (888) 978-8835.

The troubleshooting table below contains a list of symptoms and the action to take to correct the problem.

Symptom	Corrective action
System does not power on	Ensure that the system is plugged in to a functioning outlet, and that the circuit breaker on the rear power is set. Turn on power.
Image quality is poor	Adjust ambient light Optimize Technical Parameters Adjust gain, TGC and Auto TGC Adjust position of LCD screen to improve viewing angle If there is no change, call Seno Service for assistance at (888) 978-8835.
Error message appears on screen	Note the error and call Seno Service for assistance at (888) 978-8835.
System does not recognize the probe	Disconnect, reconnect, and reselect the probe. If there is no change, call Seno Service for assistance at (888) 978-8835.
The system does not start, the fans do not spin	Ensure that the system is plugged in to a functioning outlet, and that the circuit breaker on the rear power is set. Turn on power. If there is no change, call Seno Service for assistance at (888) 978-8835.
The main monitor does not turn on (fans run, and system console is on)	Check the cables at the back of the monitor Reboot If there is no change, call Seno Service for assistance at (888) 978-8835.
The Touch Screen display is off when the system is ready	Reboot if no change, call Seno Service for assistance at (888) 978-8835.

Symptom	Corrective action
The system console is off	Reboot. If no change, call Seno Service for assistance at (888) 978-8835.
The system displays a banner indicating a HWproblem	Reboot. If banner stays on, call Seno Service for assistance at (888) 978-8835.
“GRUB error” is displayed at boot time.	Internal hard drive failure. Service is needed. Call Seno Service for assistance at (888) 978-8835.
Splash screen is up but the progress bar has not moved for more than 5 minutes	Operating system is damaged. Service is needed. Call Seno Service for assistance at (888) 978-8835.
A broom icon appears in the icon area of the screen	This icon appears once a month, it is to remind the operator to check the air filters. Stop the system and verify the filters. Select the icon A pop-up window appears Select yes on the pop-up window The icon disappears
Artifact, white cone(s) in the B-mode, and/or coloredcone(s) on the color flow image, starting from the skin line	Reboot the system. If there is no change, call Seno Service for assistance at (888) 978-8835.
An audible alert rings. A pop-up window may additionally appear.	Note the error code displayed at the rear of the system and call Seno Service for assistance at (888) 978-8835.

B Laser Classification and Warning Labels, Emissions & Immunity, EMC, and Declarations

B.1 Laser Classification

The Imagio® Breast Imaging System uses an Nd:YAG 1064 nm 5-25 ns pulse and an Alexandrite 757 nm 50-100 ns pulse lasers. Both lasers operate at 85 mJ +/-20% that correspond to a max energy out of the probe =102 mJ (0.51W) per laser (wavelength). Nominal operation is at 85 mJ, and the max energy is 102 mJ, that correspond to 20% over the operational energy output.

The Imagio® Breast Imaging System emitted laser beams are a Class 3B eye hazard if the direct or reflected beam is viewed. The lasers operate at two different wavelengths. Any energy transmitted by the Imagio® Breast Imaging System that enters the eye may be focused directly on the retina. Direct absorption of laser energy by the retina may result in temporary clouded vision, retina lesion, long term scotoma and long-term photophobia.

The potential for loss of sight exists in any case of:

- Direct laser radiation
- Reflected laser radiation
- Diffused laser radiation

The laser labeling which follows also has reference to the Class 4 laser. The laser is only a Class 4 if the lid is removed from the laser head assembly located within the cabinet of the system. This can only be done by the manufacturer. Seno has not trained any third party on the servicing of the laser head and as such, opening of the laser head is a safety issue and voids the warranty.

B.1.1 Laser Warning Labels

Label Name	Warning
System - Class 3B Laser	

Label Name	Warning
Explanatory - Class 4 Laser	 DURING SERVICE, WHEN THIS LID IS OFF AND THE SYSTEM IS POWERED, THE LASER IS CONSIDERED A CLASS 4 INVISIBLE LASER RADIATION AVOID EYE OR SKIN EXPOSURE TO DIRECT OR SCATTERED RADIATION Nd:YAG 1064nm 5-25ns pulse <200mJ (<1W) Alexandrite 757nm 50-100ns pulse <200mJ (<1W) IEC 60825-1:2014
High Voltage	
Interlocked Laser Panel - Class 4 Laser	 DANGER CLASS 4 INVISIBLE LASER RADIATION WITH LID REMOVED AND INTERLOCKS DEFEATED AVOID EYE OR SKIN EXPOSURE TO DIRECT OR SCATTERED RADIATION IEC 60825-1:2014
Attention High Voltage	 WARNING- HIGH VOLTAGE DEVICE CONTAINS A POTENTIALLY DANGEROUS ELECTRICAL CHARGE BEFORE WORKING ON THE DEVICE: -SWITCH OFF THE SUPPLY VOLTAGE -DISCONNECT EACH SOCKET -WAIT FOR 1 MIN. BEFORE OPENING THE DEVICE
Laser Beam	

B.1.2 The Imagio® Breast Imaging System and Output Specifications

Specified Parameter	Nd:YAG Flashlamp Pumped, QS SSL	Alexandrite Flashlamp Pumped, QS SSL
Wavelength (nm) (1)	1064	757
Beam Shape (cm ²) (2)	Rectangular 4.3 x 1.6	Rectangular 4.3 x 1.6

Specified Parameter	Nd:YAG Flashlamp Pumped, QS SSL	Alexandrite Flashlamp Pumped, QS SSL
Laser Mode	Pulsed	Pulsed
Repetition Rate (Hz)	5	5
Pulse Energy (mJ) (3)	102	102
Average Power (W)	0.51	0.51
Pulse Width (ns)	7	50
Beam Divergence (mrad)	560 x 980	560 x 980
Beam Diameter (cm) @ 10cm Viewer Distance	7.0 x 10.0	7.0 x 10.0

- (1)(+/-10nm)
- (2)Contiguous output windows: Total Area 4.3cm x 1.6cm. Note that the actual area of the probe is 4.3 cm x 3.3 cm
- (3)Maximum Energy Probe Output (85mJ+/-20%): 102mJ

Ocular Exposure		
Specified Parameter	Laser YAG (1064nm)	Laser Alexandrite (757nm)
Irradiance @ 10cm (mW/cm ²)	7.28	7.28
Radiant Exposure @ 10 cm (mJ/cm ²)	1.46	1.46
MPE_Repetitive-Pulse (μJ/cm ²)	1.88	0.23
Energy/Pulse Thru 7mm Aperture(J)	560 x 10 ⁻⁶	560 x 10 ⁻⁶
Class 3B AEL (J)	62.5 x 10 ⁻³	37.2 x 10 ⁻³
Laser Class	3B	3B
Viewing Condition	Unaided Viewing	Unaided Viewing
Necessary OD @ 10cm Viewer Distance	2.9	3.8
Nominal Ocular Hazard Distance (m)	3.14	8.9
Skin Exposure		
Specified Parameter	YAG Laser (1064nm)	Alexandrite Laser (757nm)
MPE Skin (mJ/cm ²)	100	25.2
Seno Level Radiant Exposure @ 0cm (mJ/cm ²) (4)	14.8	14.8
Necessary OD	None	None

- (4) Skin Exposure at the output of the Probe with Uniform Energy Distribution

B.1.3 General information

Electrostatic discharge (ESD) or static shock is a natural phenomenon. Static shock is a discharge of electrical energy that can be transferred to file cabinets, computer equipment, metal door knobs and other individuals. Static shock occurs most often during low humidity conditions that can be brought on by heating or air conditioning. To avoid damage to the system or probes resulting from an electrical energy discharge from a system operator or patient, the use of anti-static mats, anti-static sprays or a ground wire connection between the system and the patient table are recommended. Medical electrical equipment need special precautions regarding EMC and need to be installed and put into service according to the EMC information. The Imagio® Breast Imaging System must be installed only by Seno authorized personnel. This equipment has been tested and found to comply within the limits for medical devices in IEC 60601-1-2 Class A. These limits are designed to provide reasonable protection against harmful interference in a typical medical institution. This equipment can radiate radio frequency energy, and if not installed and used in accordance with the instructions, may

interfere with other devices in the vicinity. Powering the system Off and On can determine if the problem is caused by this unit. In addition, electromagnetic fields from fixed transmitters such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast, and TV broadcast may interfere with the use of this device. Portable and mobile RF communications equipment can affect the device. These fields may cause degradation of the image quality of the device. Sources of electromagnetic interference cannot be predicted theoretically with accuracy.

If this system causes harmful interference to other devices, or the system demonstrates an interference pattern on the image, the operator is encouraged to try to correct the interference by:

- Increasing the distance separating the pieces of equipment
- Ensuring the system is not connected to the same outlet as the other device(s)
- Using only shielded cabling when connecting the device to networks and peripherals
- Reorienting the device
- Consulting the manufacturer or field service representative for help

If abnormal performance persists, additional measures may be necessary, such as relocating the system.

B.1.4 Recommended separation distances between portable and mobile RF communications equipment and Imagio® Breast Imaging System

The Imagio® Breast Imaging System is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the operator of the Imagio® Breast Imaging System can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the Imagio® Breast Imaging System as recommended below, according to the maximum output power of the communications equipment.

Rated maximum output power of transmitter w	Separation distance according to frequency of transmitter		
	150 kHz to 80 MHz	80 MHz to 800 MHz	800 MHz to 2.5 GHz
	$d = 35\sqrt{P}$	$d = 1.2\sqrt{P}$	$d = 2.3\sqrt{P}$
0.01	35	0.12	0.23
0.1	11	0.40	0.73
1	35	1.2	2.3
10	111	3.8	7.3
100	350	12	23
For transmitters rated at a maximum output power not listed above, the recommended separation distance (d) in meters (m) can be estimated using the equation applicable to the frequency of the transmitter, where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer.			
<i>Note: At 80 MHz and 800 MHz, the separation distance for the higher frequency range applies</i>			
<i>Note: These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects, and people.</i>			

Imagio® Breast Imaging System is provided with the following cables:

- Power supply cable 3m
- Probe cables
- 2.10m
- Ethernet cable
- 5m

Imagio® Breast Imaging System has been EMC tested with the following accessories:

- Seno OA/US L18-1 Probe
- Seno US L18-5 Linear Probe
- Ethernet cable, CAT6a SSTP (PIMF) 500Mhz (shielded) AWG26/7, meets standards EIA/TIA 568 B2-ISO/IEC11801, maximum length 5 meters

B.1.5 Declaration of Electromagnetic Emissions

The Imagio® Breast Imaging System is suitable for use in the following environment. The operator must assure that it is used only in the electromagnetic environment as specified.

Emission test	Compliance	Electromagnetic environment - guidance
RF Emission CISPR 11	Group 1	Imagio® Breast Imaging System uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment.
RF Emission CISPR 11	Class A	Imagio® Breast Imaging System is suitable for use in all establishments other than domestic and those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes.
Harmonic emission IEC 61000-3-2	Class A	
Voltage fluctuation/Flicker emissions IEC 61000-3-3	Complies	

B.1.6 Declaration of Immunity

Guidance and manufacturer's declaration - electromagnetic immunity			
The Imagio® Breast Imaging System is suitable for use in the following environment. The operator must assure that the system is used according to the specified guidance and only in the electromagnetic environment listed in the electromagnetic environment as specified.			
Immunity Test	IEC 60601 Test	Compliance Level	Electromagnetic Environment Guidance
Conducted RF IEC 61000-4-6 Radiated RF IEC 61000-4-3	3 Vrms 150 kHz to 80 MHz 3 V/m 80 MHz to 2,5 GHz	0.10 Vrms 3.0 V/m	Potable and mobile RF communications equipment should be used no closer to any part of the Imagio® Breast Imaging System, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter.
Recommended separation distance $d = 35\sqrt{P}$ $d = 1.2\sqrt{P} \quad 80 \text{ MHz to } 800 \text{ MHz}$ $d = 2.3\sqrt{P} \quad 800 \text{ MHz to } 2.5 \text{ GHz}$ Where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in meters (m). Interference may occur in the vicinity of equipment marked with the following symbol: 			
<i>Note: At 80 MHz and 800 MHz, the higher frequency range applies</i> <i>Note: These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects, and people.</i> - Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which Imagio® Breast Imaging System is used exceeds the applicable RF compliance level above, Imagio® Breast Imaging System should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as re-orienting or relocating the Imagio® Breast Imaging System. - Over the frequency range 150 kHz to 80 MHz, field strengths should be less than [VI] V/m.			

Immunity test	IEC 60601 test level	Compliance level	Electromagnetic environment guidance
Electrostatic discharge (ESD) IEC 61000-4-2	±8 kV contact ±15 kV air	± 4 kV contact ± 8 kV air	Floors shall be wood, concrete, or ceramic tile. If floors are covered with synthetic material, the relative humidity shall be at least 30%
Electrical fast transient/burst IEC 61000-4-4	±2 kV for power supply line ±1 kV for input/output lines	±2 kV for power supply line ±1 kV for input/output lines	Main power quality shall be that of a typical commercial or hospital environment.
Surge IEC 61000-4-5	±1 kV differential mode ±2 kV common mode	±1 kV differential mode ±2 kV common mode	Main power quality should be that of a typical commercial or hospital environment.
Voltage dips, short interruptions, and voltage variations on power supply input line IEC 61000-4-11	<5% UT (>95% dip in UT) for 0.5 cycle 40% UT (60% dip in UT) for 5 cycles 70% UT (30% dip in UT) for 25 cycles <5% UT (>95% dip in UT) for 5 sec	<5% UT (>95% dip in UT) for 0.5 cycle 40% UT (60% dip in UT) for 5 cycles 70% UT (30% dip in UT) for 25 cycles <5% UT (>95% dip in UT) for 5 sec	Main power quality should be that of a typical commercial or hospital environment. If the operator of the Imagio® Breast Imaging System requires continued operation during power mains interruptions, it is recommended that the Imagio® Breast Imaging System be powered from an uninterruptible power supply.
Power frequency (50/60Hz) Magnetic field IEC 61000-4-8	3 A/m	3 A/m	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment.
<i>Note: UT is the A.C. main voltage prior to application of the test level.</i>			

B.2 Electromagnetic Compatibility Compliance

This section describes the Imagio® Breast Imaging System EMC compliance and Special EMC Precautions.

Medical Electrical Equipment needs special precautions regarding EMC and needs to be installed and put into service according to the EMC information provided in this section and Guidance and Manufacturer's Declaration.

Portable and Mobile RF Communications Equipment can affect Medical Electrical Equipment.

B.2.1 Electromagnetic Interference Testing

The Imagio® Breast Imaging System has undergone Electromagnetic Interference (EMI) testing and has been found to comply with the Electromagnetic Compatibility (EMC) Class A limits of the IEC/EN 60601-1-2.

The Imagio® Breast Imaging System can radiate radio frequency energy and may cause interference with other equipment in the same area.

B.2.2 Special EMC Precautions

Special precautions concerning electromagnetic compatibility (EMC) must be taken for all medical electrical equipment.

- The Imagio® Breast Imaging System complies with IEC/EN 60601-1-2. The Imagio® Breast Imaging System is intended only to be installed in professional medical environments, such as a hospital, clinic, or imaging center. The Imagio® Breast Imaging System is not to be connected directly to the public supply electrical network.
- The Imagio® Breast Imaging System complies with all applicable and required standards for electromagnetic interference.
- The Imagio® Breast Imaging System does not affect nearby equipment and devices and it is not normally affected by nearby equipment and devices.
- It is good practice to avoid using the Imagio® Breast Imaging System near other electronic equipment.

B.2.3 Interference from Other Devices

Portable and mobile RF communications equipment can affect the Medical Electrical Equipment. The quality of the Imagio® Breast Imaging System scan results can be negatively affected by interference from external electromagnetic fields generated by other electrical devices installed in proximity. These devices include portable and mobile RF communication devices. Take all necessary precautions to preclude EMI interference from other devices.

B.2.4 Separation Distances

The Imagio® Breast Imaging System is intended for use in the electromagnetic environment in which radiated disturbances are controlled. The customer or operator of the Imagio® Breast Imaging System can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF Communications Equipment and the Imagio® Breast Imaging System as recommended below, according to the maximum output power of the communications equipment.

Max Output Power (Watts)	Separation (m) 150kHz to 80MHz $D=(3.5/V1)(\text{Sqrt } P)$	Separation (m) 80 to 800MHz $D=(3.5/E1)(\text{Sqrt } P)$	Separation (m) 800MHz to 2.5GHz $D=(7/E1)(\text{Sqrt } P)$
0.01	0.11667	0.11667	0.23333
0.1	0.36894	0.36894	0.73785
1	1.1667	1.1667	2.3333
10	3.6894	3.6894	7.3785
100	11.667	11.667	23.333

C Product Specifications

Testing performed with a laser emission duty cycle of 5 minutes on and 10 minutes off.

System Cabinet	
Size (not including monitors)	79 cm x 104 cm x 135 - 170 cm (31" x 41" x 53 - 67")
Maximum Weight	500 pounds (227kg)
Power Requirements	208-240 VAC, 50-60 Hz, 10 A
Laser Warm-up Time	Up to 15 minutes
Environmental Requirements: Operational Temperature and Humidity	Ambient temperature 15 °C to 30 °C (59 °F to 86 °F), relative humidity 30% - 70% non-condensing
Environmental Requirements: Storage/Transport Temperature and Humidity	Ambient temperature 4 °C to 50 °C (39 °F to 122 °F), relative humidity 30% to 80% non-condensing
Acoustic Noise	<80 dB
Generated Heat	< 2500 watts (8530 BTUs)
Mode of Operation	OA duty cycle: 5 minutes ON / 10 minutes OFF
Ingress	IP20 cabinet IPX8 foot switch
Operator Interface	Flat 23-inch Full HD monitor with handle Full HD Touchscreen Multitouch Touchpad Brightness of Touchscreen and Console adjustable by the user multi-button console Easy to adjust the console height footswitch
Mobility	Wheels and locking casters

Laser System	
Laser Types	Alexandrite, Nd:YAG
Wavelengths	757 +/- 10 nm (Alexandrite); 1064 +/- 10 nm (Nd:YAG)
Maximum System Output Power	0.51 W (5 Hz) (Alexandrite); 0.51 W (5 Hz) (Nd:YAG). 1.02 W Combined Power (10 Hz)
Nominal System Nominal Output Energy	85 mJ +/- 20% (Alexandrite); 85 mJ +/- 20% (Nd:YAG)
System Output Fluence	< 20 mJ / cm ²

Laser System	
Pulse Width	75 ns nominal +/-25 ns (Alexandrite); 15 ns nominal +/-10 ns (Nd:YAG)
Repetition Rate	5 Hz per wavelength or 10 Hz “interlaced” operation. Both wavelengths are always selected during OA operation.
Laser Classification	Class 3B (out of the probe, external), Class 4 (internallaser source), IEC 60825 standard.
Activation	Laser Authorization Card, Interlock, Foot Switch
Cooling System	1.0 L of Distilled water required

OA/US L18-1 Probe	
Number of Elements	256
Type	OA/US L18-1 50 mm Linear Array
Biocompatibility	Parts intended for patient contact made of biocompatible materials; not made with natural rubber latex
Cable Length	1.88 m (74.0 in)
Minimum Cable Bend Radius	5 cm (2”)
Maximum Patient Contact Surface Size	70 1 mm x 42 mm
Mechanical Focal Length (Focal Depth)	21 ± 2 mm
Ultrasound Connector Type	ITT Cannon 127050-0595 Zero Insertion Force (ZIF)
Laser Connector Type	Seno custom
Probe Fluence	Skin Exposure <20 mJ / cm ²
Ingress	IPX1

US L18-5 Linear Probe	
Number of Elements	256
Type	US L18-5 (50 mm Linear Array
Biocompatibility	Parts intended for patient contact made of biocompatible materials; not made with natural rubber latex
Cable Length	2.10 m (82.7 in)
Maximum Patient Contact Surface Size	65.8 mm (2.6 in) x
Mechanical Focal Length (Focal Depth)	21 ± 2 mm
Ultrasound Connector Type	ITT Cannon 127050-0595 Zero Insertion Force (ZIF)
Ingress	IPX1

Imaging/Data Acquisition	
Imaging Modes	B-mode OA co-registered with B-Mode (OA/US) Color Doppler: Color Flow (CFI), Color Power (cPI), Direction Color Power (dCPI) Pulsed Wave Doppler (PW) ShearWave™ Elastography (SWE™)
Imaging Features	Panoramic Imaging Simultaneous Doppler (Duplex and Triplex) Wide Sector Imaging (Trapezoid) Tissue Harmonic Imaging SuperCompound (Spatial Compounding) SuperRes (Adaptive Filtering)

	TissueTuner (Speed of Sound control) B-Mode PRF (Reverberation Reduction)
Workflow	Prospective and Retrospective clip Capture DICOM 3.0 compatible
OA/US Mode Features	Maximum 40 mm imaging depth

D Probe Ranges, Accuracies, and Precision

D.1 Transducer information

Probe name	Geometry	Nominal center frequency	Imaging frequency range
US L18-5	Linear	8.5 MHz	4.0 – 15 MHz
OA/US L18-1	Linear	8.5 MHz	4.0 – 15 MHz

D.2 Measurements

D.2.1 2D measurements

The table below shows the measurement accuracy and range for the 2D measurements available on the Imagio® Breast Imaging System. Measurements were performed using both the US L18-5 and OA/US L18-1 linear probes.

Measurement type	Accuracy	Range	Notes
Axial distance	+/- 5% or 1mm	0.01 to 25 cm	Whichever is larger (1)
Lateral distance	+/- 5% or 2mm	0.01 to 20 cm	Whichever is larger (1)
Area	+/- 5% or 25mm ²	0.01 to 1000 cm ²	Whichever is larger (1, 2)

Note 1: Distance and area measurement accuracy test is performed using the CIRS Model 040GSE phantom.

Note 2: Continuous trace accuracy depends on operator.

Note: The tolerances specified apply only to measurements made in a homogeneous medium with a speed of sound of 1540m/s and with non-refractive surface and with insonification angle of 90 degrees relative to the surface of the probe.

D.2.2 Panoramic distance measurements

Panoramic distance accuracy is tested using an ATS 551 Small Parts phantom. A panoramic sweep is performed by translating the probe horizontally at the surface of the Phantom. Distance is measured between the first and last target structure of the Phantom in the panoramic image. Individual non-panoramic images are collected spanning the total area in the panoramic image are collected. Distance measurements are performed in the non-panoramic images and totaled. Panoramic distance measurement accuracy is the difference between the total distance measured in the individual images and the distance measured in the panoramic image.

Metric	Specification	Test level	Notes
Probe name / preset	US L18-5 Linear Probe/General	I	1
Panoramic distance measurement accuracy	+/- 10%	M	2
Metric	Specification	Test level	Notes
Probe name / preset	OA/US L18-1 Linear Probe/General	I	1
Panoramic distance measurement accuracy	+/- 10%	M	2

Test Level I: Verified by inspection or qualitative analysis.

Test Level M: Verified by measurement.

Note 1: Specification is for information purposes only and is not a testable requirement at the system level.

Note 2: The Panoramic distance measurement accuracy test is performed using the ATS 551 Small Parts phantom and covers a maximum distance of approximately 20 cm.

D.2.3 B-Mode spatial resolution measurements

Spatial resolution test is measured using a wire target at the elevational focal depth of the probe with the focal zone set as close as possible to the targets. The probe is placed across the wire targets such that they appear as points. The minimum distance at which the targets are individually distinguishable is the spatial resolution. Axial resolution is measured along the axis beam axis (vertically spaced targets), lateral resolution is perpendicular to the beam axis (horizontally spaced targets).

Metric	Specification	Test level	Notes
Probe name / preset	US L18-5	I	1
Maximum Axial Resolution	1 mm	M	2
Maximum Lateral Resolution	1.5 mm	M	2
Metric	Specification	Test level	Notes
Probe name / preset	OA/US L18-1	I	1
Maximum Axial Resolution	1 mm	M	2
Maximum Lateral Resolution	1.5 mm	M	2

Test Level I: verified by inspection or analysis

Test Level M: verified by measurement

Note 1: Specification is for information purposes only and is not a testable requirement at the system level.

Note 2: Measured using a custom wire phantom in water with two a pair of wires sloping towards each other at the focal point of the probe. The wires are vertically spaced for axial resolution measurement and horizontally spaced for lateral resolution measurement. The probe is placed orthogonal to the wires such that the appear as point targets. The probe is translated along the wire until the point targets are no longer individually discernable. Measurement may also be performed using a Phantom with fixed target spacing such as the ATS 551 Small Parts Phantom.

D.2.4 Doppler Accuracy and Sensitivity measurements

Doppler measurement calipers are positioned on a single pixel in a desired location. Velocity measurements are displayed in units of cm/s or m/s with at least 1 digit past the decimal point. The tables below show the measurement accuracy, range, and tolerance for the Doppler measurements on the ultrasound system in combination with the transducer.

Metric	Specification	Test level	Note
Probe name / preset	US L18-5 /Breast	I	1
PW Doppler spectral velocity accuracy	±15% error	M	2
PW Doppler minimum penetration depth	55 mm	M	3
Metric	Specification	Test level	Note
Probe name / preset	OA/US L18-1 / Breast	I	1
PW Doppler spectral velocity accuracy	±15% error	M	2
PW Doppler minimum penetration depth	55 mm	M	3

Test Level I: verified by inspection or analysis

Test Level M: verified by measurement

Note 1: Transmitter center frequencies for grayscale imaging

Note 2: Using CIRS 043 Doppler String Phantom

Note 3: Using Gammex 403TM Flow Phantom. PW gain is adjusted to the noise level ~1%. The depth limit value above which the vessel is poorly represented on the PW spectrum display is the penetration depth.

D.1.1 Color Flow Doppler sensitivity

The purpose of this test is to measure the penetration of color flow imaging performance. Each probe is placed in Color Flow Imaging mode. The Color Flow Doppler Phantom (Gammex 403™) is equipped with a 5 mm diameter sloping cylindrical vessel is imaged to measure the Color Doppler Sensitivity in centimeters. The flow pump is set to keep a flow rate at 7.5 ml/s and color gain is set to the noise level ~1%. The maximal depth (penetration) at which the Doppler signal can be detected on the display is measured using the calipers with one caliper at skin line and the other where the color flow signal starts to show dropouts.

Metric	Specification	Test level	Note
Probe name / preset	US L18-5 / Breast	I	1
Color penetration	60 mm	M	2
Metric	Specification	Test level	Note
Probe name / preset	OA/US L18-1 / Breast	I	1
Color penetration	60 mm	M	2

Test Level I: Verified by inspection or analysis

Test Level M: Verified by measurement

Note 1: Transmitter center frequencies for grayscale imaging

Note 2: Using Gammex 403 Phantom.

D.1.2 Elastography measurements

D.1.2.1 SWE penetration and spatial resolution

The tables below document the SWE elasticity estimation accuracy, SWE penetration depth, and SWE spatial resolution specifications for all probes where the SWE mode is available.

Metric (notes)	Specification	Test level	Notes
Probe name / preset	US L18-5 / General	I	1
Elasticity Estimation Accuracy	± 30% or 3 kPa (whichever is greater)	M	2
SWE minimum penetration depth	30 mm	M	3
SWE spatial resolution range	1.58 mm - 2.53 mm	M	4
Metric (notes)	Specification	Test level	Notes
Probe name / preset	OA/US L18-1 / General	I	1
Elasticity Estimation Accuracy	± 30% or 3 kPa (whichever is greater)	M	2
SWE minimum penetration depth	30 mm	M	3
SWE spatial resolution range	1.58 mm - 2.53 mm	M	4

Note: If the operator considers the shearwave velocity measurement, the error of 15% or m/s whichever is greater should be considered.

Note: If the operator considers the Young's modulus measurement, the error of 30% or 3 kPa whichever is greater should be considered.

Test Level I: Verified by inspection or analysis

Test Level M: Verified by measurement

Note 1: Specification is for information purposes only and is not a testable requirement at the system level.

Note 2: SWE elasticity estimation accuracy is defined as the difference between the SWE-based elasticity measurement from a given phantom target and the reference elasticity for this target provided by the phantom manufacturer. The elasticity estimation accuracy specifications were derived from SWE measurements on the four spherical targets (reference elasticities equal to 5.4, 7.8, 24.8, and 52.8 kPa) and the uniform-elasticity background (reference elasticity

equal to 16.3 kPa) contained in the CIRS 049 Quality Assurance Elasticity phantom.

Note 3: SWE penetration is defined as the maximum depth for which the SWE image exhibits good color filling and minimal noise. The SWE penetration specifications were derived by scanning regions of uniform-elasticity background material of the CIRS 040GSE Multi-Purpose Multi-Tissue Ultrasound Phantom.

Note 4: SWE spatial resolution range is measured using the CIRS 049A phantom. The smallest detectable inclusion corresponding to the highest elastic contrast (background vs type IV) is imaged. The procedure is to scan Type IV inclusions from largest size to smallest size until the inclusion cannot be detected on the elasticity map. The list of inclusion sizes is 16.668 mm, 10.404 mm, 6.49 mm, 4.05 mm, 2.53 mm, and 1.53 mm. The result is an interval for the resolution (smallest detected size < x < next smallest detected size).

E Acoustic Output Tables

E.1 Maximal Temperature Data

Transducer Name	Maximal Temperature Increase	Test Method
OA/US L18-1	21.6°C	Still Air
US L18-5	21.6°C	Still Air

E.2 OA/US L18-1 & US L18-5 Linear Probe B-MODE

Index Label		MI	TIS		TIB		TIC
			At surface	Below surface	At surface	Below surface	
Maximum Index Value		1.5	0.99		1.28		1.28
Index Component Value			0.99	0.99	1.28	0.99	
Associated Acoustic Parameters	pr.a@zMI (Mpa)	2.8					
	W0 (mW)		52.16		52.16		52.2
	W1x1 (mW)		3.43		3.43		
	ZS (cm)			0.2			
	zb (cm)					0.5	
	zMI (cm)	1.6					
	zpii.a (cm)	1.6					
	fawf (MHz)	3.75	4.25		4.25		4.25
Other Information	prf (Hz)	690					
	srr (Hz)	69					
	npps	10					
	lpa.a@zpii.a (W/cm2)	--					
	lspta.a@zpii.a (mW/cm2)	56.5					
	lspta@zpii (mW/cm2)	84.82					
	pr@zpii (Mpa)	3.49					
Operating control conditions	Condition 1	MI					
	Condition 2		TIS		TIB		TIC
	Condition 3						
	Condition 4						
Condition 1: B: General, B mode THI, Focal zone 37 mm, PEN, SuperCompound on, Acoustic Power 0 dB							
Condition 2: B: Thyroid, B mode Fundamental, Focal zone 4 mm, PEN, SuperCompound off, Acoustic Power 0 dB							

E.3 US L18-5 Linear Probe Color Doppler

Index Label		MI	TIS		TIB		TIC
			At surface	Below surface	At surface	Below surface	
Maximum Index Value		0,7	0,27		0,27		0,11
Index Component Value			0,27	0,27	0,15	0,27	
Associated Acoustic Parameters	pr,a@zMI (Mpa)	1,8					
	W0 (mW)		6,57		6,57		5,1
	W1x1 (mW)		0,29		0,29		
	zs (cm)			0,61			
	zb (cm)					0,59	
	zMI (cm)	1,8					
	zpII,a (cm)	2,0					
	fawf (MHz)	9,6	9,625		9,625		9,625
Other Information	prf (Hz)	135					
	srr (Hz)	135					
	npps	1					
	lpa,a@zpII,a (W/cm2)	124,9					
	lspta,a@zpII,a (mW/cm2)	54,11					
	lspta@zpII (mW/cm2)	105,88					
	pr@zpII (Mpa)	2,91					
Operating control conditions	Condition 1	MI	TIS		TIB		TIC
	Condition 2						
	Condition 3						
	Condition 4						
Condition 1: B: Thyroid, B mode Harmonic, Focal zone 14 mm, PEN, SuperCompound on, Acoustic Power 0 dB							

E.4 US L18-5 Linear Probe Shear Wave Elastography

Index Label		MI	TIS		TIB		TIC
			At surface	Below surface	At surface	Below surface	
Maximum Index Value		1,9	0,86		2,81		1,40
Index Component Value			0,86	0,51	1,00	2,81	
Associated Acoustic Parameters	pr,a@zMI (Mpa)	4,1					
	W0 (mW)		39,52		43,92		28,8
	W1x1 (mW)		39,52		43,92		
	zs (cm)			1,59			
	zb (cm)					2,7	
	zMI (cm)	1,3					
	zpii,a (cm)	1,6					
	fawf (MHz)	4,5	4,5		4,5		4,5
Other Information	pr (Hz)	19					
	srr (Hz)	1					
	npps	19,1,109					
	lpa,a@zpii,a (W/cm2)	367,7					
	lspta,a@zpii,a (mW/cm2)	4,36					
	lspta@zpii (mW/cm2)	0,00					
	pr@zpii (Mpa)	4,67					
Operating control conditions	Condition 1	MI					
	Condition 2		TIS		TIB		TIC
	Condition 3						
	Condition 4						
Condition 1: Breast, SWE Box position 20 mm, Bmode: Thyroid, B mode Harmonic, Focal zone 14 mm, PEN, SuperCompound on, Acoustic Power 0 dB							
Condition 2: Breast, SWE Box position 45 mm, Bmode: Thyroid, B mode Harmonic, Focal zone 14 mm, PEN, SuperCompound on, Acoustic Power 0 dB							

E.5 US L18-5 Linear Probe Pulsed Wave Doppler

Index Label		MI	TIS		TIB		TIC
			At surface	Below surface	At surface	Below surface	
Maximum Index Value		0,7	0,09		0,30		1,42
Index Component Value			0,09	0,07	0,17	0,30	
Associated Acoustic Parameters	pr,a@zMI (Mpa)	1,5					
	W0 (mW)		3,70		3,70		3,7
	W1x1 (mW)		3,70		3,70		
	zs (cm)			0,81			
	zb (cm)					1,2	
	zMI (cm)	1,2					
	zpii,a (cm)	1,3					
	fawf (MHz)	5,0	5		5		5
Other Information	prf (Hz)	805					
	srr (Hz)	N/A					
	npps	1					
	lpa,a@zpii,a (W/cm2)	59,9					
	lspta,a@zpii,a (mW/cm2)	112,00					
	lspta@zpii (mW/cm2)	174,25					
	pr@zpii (Mpa)	1,75					
Operating control conditions	Condition 1	MI					
	Condition 2		TIS		TIB		
	Condition 3						TIC
	Condition 4						
Condition 1: General, Focal zone 22 mm, SV 1 mm, Scale 8 cm/s, Acoustic Power 0 dB							
Condition 2: General, Focal zone 68 mm, SV 0.5 mm, Scale 110 cm/s, Acoustic Power 0 dB							
Condition 3: General, Focal zone 2 mm, SV 1.5 mm, Scale 190 cm/s, Acoustic Power 0 dB							

E.6 OA/US L18-1 B-MODE Elastography

Index Label	M	TIS			TIB			TIC	
		Scan	Non-scan		Scan	Non-scan		Scan	Non-scan
		At surface	At surface	Below surface	At surface	At surface	Below surface	At surface	At surface
Maximum Index Value	1.56		1.21			1.57		1.42	
Index Component Value	B: 1.56		B: 1.03	B: 1.03		B: 1.29	B: 1.29		B: 1.29
	P: 1.17		P: 0.05	P: 0.04		P: 0.07	P: 0.23		P: 0.07
	F: 1.17		F: 0.13	F: 0.04		F: 0.06	F: 0.03		F: 0.06
Associated Acoustic Parameters	pr_s@z0 (Mpa)	B: 3.17 P: 2.48 F: 2.54							
	W0 (mW)			B: 52.21 P: 2.54 F: 4.33			B: 52.21 P: 2.54 F: 4.33		B: 52.21 P: 2.54 F: 4.33
	W1x1 (mW)			B: 52.21 P: 2.54 F: 4.33			B: 52.21 P: 2.54 F: 4.33		
	minP (mW)			B: -- P: 1.09 F: 0.02					
	z0 (cm)			B: -- P: 1.34 F: 2.7					
	zdp (cm)			B: -- P: 1.34 F: 2.7			B: -- P: 1.34 F: 2.7		
	z0 (cm)						B: -- P: 1.54 F: 2.7		
	z0 (cm)	B: 1.27 P: 2.96 F: 1.53		B: -- P: 2.31 F: 1.53			B: -- P: 2.31 F: 1.53		
	z0 (cm)	B: 1.18 P: 2.89 F: 1.38		B: -- P: 1.65 F: 1.38					
	skq@z0 (cm)						B: -- P: 0.16 F: 1.06		
	z0w (Hz)	B: 4.13 P: 4.5 F: 6.25		B: 4.13 P: 4.5 F: 6.25			B: 4.13 P: 4.5 F: 6.25		B: 4.13 P: 4.5 F: 6.25
	Dim of Aspt			B: 1.62 P: 1.26 F: 5.12			B: 1.62 P: 1.26 F: 5.12		B: 1.62 P: 1.26 F: 5.12
				B: 0.05 P: 0.05 F: 0.05			B: 0.05 P: 0.05 F: 0.05		B: 0.05 P: 0.05 F: 0.05
Other Information	ld (us)	B: 0.41 P: 675.96 F: 0.52							
	err (Hz)	B: 690 P: 0.6 F: 25.2							
	err (Hz)	B: 69 P: 0.6 F: 0.6							
	pr@z0 (Mpa)	B: 3.75 P: 3.8 F: 3.32							
	skq@z0 (cm)						B: -- P: 0.14 F: 0.61		
	spe_s@z0 (W/cm2)	B: 356.54 P: 157 F: 174.75							
	Focal Length			B: -- P: 0.12 F: 5.2			B: -- P: 0.12 F: 5.2		
				B: -- P: 0.24 F: 0.18			B: -- P: 0.24 F: 0.18		

Index Label		MI	TIS		TIB		TIC
			At surface	Below surface	At surface	Below surface	
Maximum Index Value		1.56	1.21		1.57		1.42
Index Component Value			B : 1.03 P : 0.05 F : 0.13	B : 1.03 P : 0.04 F : 0.04	B : 1.29 P : 0.07 F : 0.06	B : 1.03 P : 0.23 F : 0.03	
Associated Acoustic Parameters	pr,a@zMI (Mpa)	B : 3.2 P : 2.5 F : 2.5					
	W0 (mW)		B : 52.21 P : 2.54 F : 4.33		B : 52.21 P : 2.54 F : 4.33		B : 52.21 P : 2.54 F : 4.33
	W1x1 (mW)		B : 52.21 P : 2.54 F : 4.33		B : 52.21 P : 2.54 F : 4.33		
	zs (cm)			B : -- P : 1.341 F : 2.703			
	zb (cm)				B : -- P : 1.544 F : 2.703		
	zMI (cm)	B : 1.2 P : 2.9 F : 1.4					
	zpII,a (cm)	B : 1.2 P : 2.9 F : 1.4					
	fawf (MHz)	B : 4.1 P : 4.5 F : 6.3	B : 4.13 P : 4.5 F : 6.25		B : 4.13 P : 4.5 F : 6.25		B : 4.13 P : 4.5 F : 6.25
Other Information	prf (Hz)	B : 690 P : 0.6 F : 25					
	srr (Hz)	B : 69 P : 0.6 F : 1					
	npps	B : 10 P : 1 F : 42					
	lpa,a@zpII,a (W/cm2)	B : 356.5 P : 157 F : 174.8					
	lspta,a@zpII,a (mW/cm2)	B : 262.2 P : 73.7 F : 2.3					
	lspta@zpII (mW/cm2)	B : 373.6 P : 182.6 F : 4.3					
	pr@zpII (Mpa)	B : 3.8 P : 3.8 F : 3.3					
Operating control Conditions	Condition 1	MI					
	Condition 2		TIS		TIB		
	Condition 3						TIC

Condition 1 : B : Breast preset ; Harmonic ; Optimization = Pen ; focal position 22 mm ; Acoustic Power 0 dB
P + F : Breast preset ; Optimization = Res ; SWE box position 4 mm ; Acoustic Power 0 dB

Condition 2 : B : Breast preset ; Harmonic ; Optimization = Pen ; focal position 52 mm ; Acoustic Power 0 dB
P + F : Breast preset ; Optimization = Std ; SWE box position 30 mm ; Acoustic Power 0 dB

Condition 3 : B : Breast preset ; Harmonic ; Optimization = Pen ; focal position 52 mm ; Acoustic Power 0 dB
P + F : Breast preset ; Optimization = Pen ; SWE box position 30 mm ; Acoustic Power 0 dB

E.7 OA/US L18-1 B-MODE Color Doppler

Index Label	MI	TIS			TIB			TIC	
		Scan	Non-scan		Scan	Non-scan		Scan	Non-scan
		At surface	At surface	Below surface	At surface	At surface	Below surface	At surface	At surface
Maximum Index Value	1.56		1.30			1.42		1.42	
Index Component Value	B: 1.56	B: 1.03			B: 1.29			B: 1.29	
	C: 1.42	C: 0.27			C: 0.14			C: 0.14	
Associated Acoustic Parameters	pr.a@zMi (Mpa)	B: 3.17							
		C: 1.73							
	W0 (mW)	B: 52.21			B: 52.21			B: 52.21	
		C: 6.65			C: 6.65			C: 6.65	
	W1x1 (mW)	B: 52.21			B: 52.21				
		C: 6.65			C: 6.65				
	minPj (mW)								
	za (cm)								
	ztp (cm)								
	zb (cm)								
	zpl (cm)	B: 1.27							
		C: 1.91							
	zBl (cm)	B: 1.18							
		C: 1.75							
	zba@zb (cm)								
	zawf (MHz)	B: 4.13			B: 4.13			B: 4.13	
		C: 8.88			C: 8.63			C: 8.63	
	Dim of Aspr								
Other Information	td (us)	B: 0.41							
		C: 0.36							
	prf (Hz)	B: 690							
		C: 290.32							
	prf (Hz)	B: 355.54							
		C: 26.29							
	pr@zpi (Mpa)	B: 3.75							
		C: 2.7							
	zba@zpi (cm)								
	zba@zpi.a (W/cm2)	B: 355.54							
		C: 99.84							
Focal Length	FLx (cm)	B: 0.24							
		C: 0.12							
	FLy (cm)	B: 0.21							
		C: 0.16							

Index Label	MI	TIS		TIB		TIC
		At surface	Below surface	At surface	Below surface	
Maximum Index Value	1.56	1.30		1.56		1.42
Index Component Value		B : 1.03 C : 0.27	B : 1.03 C : 0.27	B : 1.29 C : 0.14	B : 1.03 C : 0.27	
Associated Acoustic Parameters	pr.a@zMi (kPa)	B : 3.2 C : 1.7				
	WI (mW)		B : 52.21 C : 6.65	B : 52.21 C : 6.65		B : 52.21 C : 6.65
	W1x1 (mW)		B : 52.21 C : 6.65	B : 52.21 C : 6.65		
	zr (cm)					
	zb (cm)					
	zMi (cm)	B : 1.2 C : 1.8				
	zpi.a (cm)	B : 1.2 C : 1.8				
	fswf (MHz)	B : 4.1 C : 8.9	B : 4.13 C : 8.63	B : 4.13 C : 8.63		B : 4.13 C : 8.63
	prf (Hz)	B : 690 C : 290.3				
	frf (Hz)	B : 49 C : 26.4				
Other Information	pps	B : 10 C : 11				
	pe.a@zpi.a (W/cm2)	B : 356.5 C : 99.8				
	spia.a@zpi.a (mW/cm2)	B : 262.2 C : 14.8				
	spia@zpi (mW/cm2)	B : 373.6 C : 42.5				
	pr@zpi (kPa)	B : 3.8 C : 2.7				
	Condition 1	MI				
	Condition 2		TIS	TIB		
	Condition 3					TIC

Condition 1 : B : Breast preset ; Harmonic ; Optimization = Pen ; focal position 22 mm ; Acoustic Power 0 dB
 C : Breast preset ; Optimization = Res ; Boost = High Definition ; Color box position 22 mm ; scale 4 cm/s
 Condition 2 : B : Breast preset ; Harmonic ; Optimization = Pen ; focal position 52 mm ; Acoustic Power 0
 C : Breast preset ; Optimization = Res ; Boost = High Definition ; Color box position 68 mm ; scale 18 cm/s
 Condition 3 : B : Breast preset ; Harmonic ; Optimization = Pen ; focal position 52 mm ; Acoustic Power 0
 C : Breast preset ; Optimization = Gen ; Boost = High Definition ; Color box position 65 mm ; scale 21 cm/s

E.8 OA/US L18-1 B-MODE Pulsed Wave Doppler

Index Label	MI	TIS			TIB			TIC	
		Scan	Non-scan		Scan	Non-scan		Scan	Non-scan
		At surface	At surface	Below surface	At surface	At surface	Below surface	At surface	At surface
Maximum Index Value	1.18		1.01			2.50			1.43
Index Component Value	1.18	—	1.01	0.71	—	1.36	2.50	—	1.43
Associated Acoustic Parameters	pr.a@zMi (Mpa)	2.63							
	W0 (mW)			42.45			40.23		42.45
	W1x1 (mW)			42.45			40.23		
	W1x1 (mW)			0.01					
	zs (cm)			1.11					
	zdp (cm)			1.11			1.11		
	zdb (cm)						1.41		
	zpi (cm)	1.49		2.05			1.74		
	zMi (cm)	1.36		1.50					
	zMi@zdb (cm)						0.23		
	fawf (MHz)	5.00		5.00			5.00		5.00
	Dim of Aspt: X (cm)			0.86			0.86		0.86
	Dim of Aspt: Y (cm)			0.05			0.05		0.05
	td (us)	0.65							
Other Information	pr (Hz)	1392							
	srr (Hz)	1392							
	pr@zpi (Mpa)	3.15							
	zpi@zpi (cm)						0.22		
	lpa.a@zpi.a (W/cm2)	271.61							
	lspa.a@zpi.a (mW/cm2)			0.30			0.30		
	Focal Length: FLA (cm)	0.22							
	Focal Length: FLy (cm)	0.20		0.20			0.20		

Index Label	MI	TIS		TIB		TIC
		At surface	Below surface	At surface	Below surface	
Maximum Index Value	1.18		1.01		2.50	1.43
Index Component Value		1.01	0.71	1.36	2.50	
Associated Acoustic Parameters	pr.a@zMi (Mpa)	2.6				
	W0 (mW)		42.45		40.23	42.45
	W1x1 (mW)		42.45		40.23	
	zs (cm)		1.108			
	zdb (cm)				1.407	
	zMi (cm)	1.4				
	zpi.a (cm)	1.4				
	fawf (MHz)	5.0	5	5		5
Other Information	pr (Hz)	1392				
	srr (Hz)	1392				
	npps	1				
	lpa.a@zpi.a (W/cm2)	271.6				
	lspa.a@zpi.a (mW/cm2)	500.5				
	lspa.a@zpi (mW/cm2)	838.2				
	pr@zpi (Mpa)	3.1				
Operating control Conditions	Condition 1	MI				
	Condition 2		TIS			TIC
	Condition 3			TIB		

- Condition 1 : Breast preset ; Sample volume position 30 mm ; Sample volume size 1.5 mm ; Scale 10.3 cm/s
Condition 2 : Breast preset ; Sample volume position 52 mm ; Sample volume size 20 mm ; Scale 30.8 cm/s
Condition 3 : Breast preset ; Sample volume position 52 mm ; Sample volume size 20 mm ; Scale 15.4 cm/s

F Risk Benefit Analysis

Risk Benefit Analysis

Seno Medical assessed that the medical benefits of the device outweigh the residual risk associated with the potential for incorrect diagnosis of false positives/negatives and the other warnings documented in the Operator Manual as outlined in ISO 14971:2019 Medical devices – Application of risk management to medical devices.

G Guidelines for OA US Scanning

G.1 OA/US INSTRUMENTATION AND SCAN TECHNIQUE

The following image guidelines are for OA/US imaging and should not be used as a replacement for standalone ultrasound imaging clinical procedures and practices. The operator should minimally perform the OA/US captures listed for each application.

Important: Always refer to OA feature training resources for the complete range of reassuring to highly suspicious OA features.

OA provides information consistent with conventional diagnostic ultrasound and interleaves the Ultrasound (US) frames with OA frames containing functional data from the mass of interest. Our scan procedures are created in accordance with The American College of Radiology and the European Society of Breast Imaging Practice Guidelines.

G.1.1 Optimize technical factors and scanning techniques for OA/US

- Adjust the Optimization Setting (Frequency) to the highest available (Resolution) and optionally interrogating with harmonics optimizes the detail and contrast resolution of the OA/US image
- Adjust the Frequency (Penetration) to its lowest setting to better laterally resolve superficial <1cm depth tissue.
- Apply just enough gel to wet the skin surface to optimize OA/US transmission and reception. Too much gel may introduce a gel standoff, air gaps or air bubbles, which will all negatively affect light transmission.
- Apply normal to light probe pressure optimizes OA/US L18-1 Probe face skin contact.
- Adjust the Time Gain Compensation (TGC) so there is uniform gray scale from superficial to deep optimizes spatial brightness uniformity of the OA/US image. Press Auto TGC as the tissue profile changes. Press the touch screen TGC button then slide your finger on the touchscreen TGC curve to increase or decrease brightness at specific depths.
- Adjust depth so the field of view contains the entire area of breast tissue while capturing only the first few millimeters of non-breast structures deep to the area of interest (i.e. chest wall, implants).
- Adjust the gray scale Focus Height and Depth to match the region of interest or mass optimizing lateral resolution.

G.1.1.1 Recommended Parameter Values for B-Mode and OA Mode

The following are baseline optimized instrumentation settings which may be saved as an Imaging Preset. Refer to “Imaging Presets” on page 158

- Optimization - Resolution
- TissueTuner™ - 1480 m/s
- Dynamic Range - 70dB

- 2D Map - 10
- Sector - Large
- PRF - Max
- Persistence - Low
- AutoTGC offset - 0 dB
- SuperRes™ - 4
- Acoustic Power - 0.0 dB
- SuperCompound™ - On

G.1.2 How to Hold the OA/US L18-1 Probe

The technique for holding the OA/US L18-1 Probe is an Operator preference; however, the following conditions must be met:

The face of the OA/US L18-1 Probe must be perpendicular to the tissue on which it is placed so that uniform pressure is applied to the tissue across the entire surface of the OA/US L18-1 Probe face.

The amount of pressure should be just enough to make full contact with the tissue across full surface of the OA/US L18-1 Probe face and in accordance with established guidelines for Ultrasound imaging.

If too much pressure is applied to the skin tissue, the underlying blood vessels may be compressed causing diminished blood volume within those vessels resulting in poor imaging.

If too little pressure is applied to the skin tissue, the risk of accidental laser light emissions may occur in addition to poor image quality.

G.2 OA/US Breast Imaging

G.2.1 Interrogation

Some critical Opto-acoustic findings occur within the tissues that surround the primary breast mass. Therefore, it is important for the operator to interrogate the tissues and structures around the main portion of the lesion and to capture images and videos that will best demonstrate things such as vascular supply and other structures in the peri-tumoral bed. This does not necessarily need to be in strict radial or anti-radial and transverse and long scan planes but should be in two orthogonal scan planes that best demonstrates both boundary and peripheral zone structures.


Important: The feature score for each of the individual 5 OA characteristics internal (Vessels, Blush, Hemoglobin) and external (Boundary Zone Vessels, Peripheral Zone Vessels), is based on the internal and external features of the entire three-dimensional mass/tissue volume. Therefore, it is very important to make a mental note of the color, number and orientation of the individual features while interrogating the entire mass and surrounding tissue. This will assist in choosing a representative video frame in which to draw a region of interest (ROI) to segment between the internal zone, boundary zone, peripheral zone.

G.2.2 Breast OA/US Imaging and Recording

To capture a region of interest (ROI), tissue, palpable area, mass or structure, a scan technique that utilizes probe motion or “sweep” is used. As the probe is swept, a video recording is captured of the region of interest and surrounding volume of tissue. The recording is saved and can later be viewed in real time or frame by frame in order to analyze the opto-acoustic signals within the region of interest.


G.2.2.1 SAX or Short Axis Video Sweep

This technique is called a Short Axis Video sweep or SAX. It is created as follows: press unfreeze, position the probe approximately 2cm outside of the ROI, press freeze then unfreeze, (press the foot switch to emit laser in OA mode), then move the probe in one direction along an imaginary line that parallels the short axis of the transducer until the probe has reached 2 cm outside the ROI on the opposite side, press freeze, and finally press save clip. This should take a maximum of 10 seconds depending on the size of the ROI moving the probe no greater than 5 mm per second. Orthogonal SAX are taken.



G.2.2.2 LAX or Long Axis Video Sweep

This technique is called a Long Axis Video Sweep or LAX. It is created as follows: press unfreeze, position the probe in a plane midway through the mass and offset right or left so the mass is seen on one side of the field of view, press freeze then unfreeze, (press the foot switch to emit laser in OA mode), then move the probe in one direction along an imaginary line that parallels the long axis of the transducer until the mass is visualized on the opposite side of the field of view. press freeze, and finally press save clip. This should take a maximum of 10 seconds depending on the size of the ROI moving the probe no greater than 5 mm per second.



Note: To capture off axis or obliquely oriented OA/US features or avoid nipple or other artifacts, orthogonal scan planes other than true radial and anti-radial can be selected. Annotate either obliqueradial or oblique anti-radial

Note: To obtain a high-quality OA/US signal, use light probe pressure to avoid compressing small vessels but maintain contact so there is no signal drop out or gel standoff

Note: Transverse and longitudinal scan planes may be used in place of radial and anti-radial scan planes

G.2.2.3 B-Mode - Imagio® Breast Imaging System (Minimum) Recommended Video Recordings:

Step: 1 1 radial scan plane SAX

Step: 2 1 anti-radial scan plane SAX

For a mass with a partial thin capsule present on anterior and posterior surfaces, but not seen on sides, a complete thin capsule can be imaged by capturing a LAX.

G.2.2.4 OA-Mode - Imagio® Breast Imaging System (Minimum) Recommended Video Recordings:

Step: 1 1 radial scan plane SAX

Step: 2 1 anti-radial scan plane SAX

- For a mass with a partial thin capsule present on anterior and posterior surfaces, but not seen on sides, a complete thin capsule can be imaged by capturing a LAX.

G.2.3 Benign Appearing OA/US Features:

- The vessels feeding and draining hemoglobin in benign masses tend to be oriented parallel to the mass wall
- The vessels tend to be a parallel paired feeding artery and draining vein
- Benign masses are usually supplied and drained by no more than 2 vessels
- The feeding vessel(s) are typically oxygenated and appear green in OA, but may be red because in OA, colorization is relative

G.2.4 Malignant Appearing OA/US Features:

- The peripheral zone vessels and boundary zone neo-vessels in malignant masses tend to be oriented and enter or exit perpendicular to the mass wall.
- The polymorphic and tortuous neo-vascularity of malignant masses may appear in a dot-dash pattern on a static image
- The polymorphic and tortuous neo-vascularity of malignant masses tend to be deoxygenated and appear red, but may be green in part because in OA, colorization is relative
- The neo-vascularity (supply and drain) of malignant masses tend to be unevenly distributed in the boundary zone
- The distribution and corresponding higher density of vessels may occur in one or a combination of areas around the lesion's capsule. Additionally, these vessels are usually polymorphic and tortuous (vary in size, shape, and orientation)
- The neo-vessels tend to be oriented and course in a "snake like" pattern, therefore it is highly recommended that all three areas; the internal zone, the boundary zone, and the peritumoral area be thoroughly interrogated when performing an OA/US scan
- OA colorization is relative, meaning the OA colorized signal from oxy and deoxy hemoglobin in vessels is determined by different light absorption characteristics of each slice of tissue, and some out of plane tissue/light interactions. Additionally, the generated ultrasound from the Opto-acoustic effect in tissues is susceptible to and is affected in the same manner as returning conventional diagnostic

ultrasound. Therefore, a malignant mass's vascularity may not always be colorized red and benign vessels may not always be colorized green

Note: Neo-vessels in the boundary zone and radiating vessels in the peripheral zone of malignant appearing masses have unique OA features. First, while interrogating, it is important to make a mental note of the number, location, and orientation of vessels in the external zone. External OA features can be identified by the Relative and Total Hemoglobin map. Secondly, assess their relative color. Finally, note the OA features found in the Internal zone. Observing and noting malignant appearing OA features from the external zone first, then the internal zone, will assist in choosing a representative video frame in which to draw regions of interest (ROI) to segment between the internal zone, boundary zone, and peripheral zone.

G.2.4.1 Example: Highly Suspicious Breast Mass OA Features

External

- Multiple boundary zone red (deoxygenated) perpendicular polymorphic neo-vessels or red blush
- More than 2 peripheral zone radiating vessels on more than one side of the mass

Internal

- Red (deoxygenated) blush
- Many large and heterogenous vessels almost filling mass
- Multiple red (deoxygenated) vessels

G.3 Axillary Lymph Node OA/US Scanning

- The most frequent site of breast cancer metastasis are lymph nodes; so it is important to explore internal mammary lymph nodes and axillary lymph nodes when a breast mass is identified.

G.3.1 Axillary Lymph Node OA/US Imaging and Recording

Begin scanning in the inferior medial axillary tail (Tail of Spence) and move the probe superiorly to identify the Sentinel Node(s) (first node(s) "downstream" from the cancer in the lymph circulatory system).

G.3.1.1 Imagio® Breast Imaging System OA (Minimum) Recommended Video Recordings for Lymph Nodes:

Step: 1 SAX transverse or long scan plane.

Step: 2 LAX in the same plane as SAX

G.3.1.2 Example: Highly Suspicious Lymph Node OA Features

Internal

- Red (de-oxygenated) blush
- Multiple tortuous vessels

External

Non-hilar trans capsular vessels – red or green

H Tissue Phantom Study

The following tables indicate the acoustic output energy for the system and OA/US L18-1 Probe combination. These tables show the worst-case indices for the OA/US L18-1 Probe type and operating condition that must be reported. Note that all thermal and mechanical indices are below 1.0 for the OA/US L18-1 Probe and below 1.06 for the US L18-5 Linear Probe for all device settings.

The OA/US L18-1 Probe; B-mode + Color and Triplex (B-mode + Color + PW-mode) were evaluated, but neither produced higher $Ispta.3$, MI, or TI than the B-mode + PW-mode summarized below. The highest MI was produced in B-mode only, as this B-mode setting was not available in the combined mode that produced the highest TI.

Track 3 Reporting Tables are included to summarize the acquired acoustic data for each sample tested. The statistics analysis of worst-case measurements tables below shows the worst-case indices for each probe type and operating condition that must be reported. The values in the table are the average of the three samples measured.

Test	Purpose	Acceptance Criteria	Results
1. Spatial Resolution	Evaluate and quantify in-plane (axial, lateral), and out-of-plane (elevational) spatial resolution of OA images	None	Axial resolution, short wavelength = 0.58 ± 0.10 mm Axial resolution, long wavelength = 0.54 ± 0.10 mm Lateral resolution, short wavelength = 1.12 ± 0.35 mm Lateral resolution, long wavelength = 1.09 ± 0.25 mm Elevational resolution, short wavelength = 2.7 ± 0.4 mm (at ~4.5 mm depth), 5.6 ± 0.2 mm (at ~34 mm depth), respectively Elevational resolution, long wavelength = 2.7 ± 0.4 mm (at ~4.5 mm depth), 5.4 ± 0.5 mm (~34.5 mm depth), respectively
2. Geometric Accuracy and Precision	Evaluate accuracy of measured 1D distances and 2D areas in OA images. Assess OA image precision with repeated acquisitions.	None	Worst-case horizontal error = $+0.81/-0.54$ mm, std. deviation = 0.33 mm (over both wavelengths) Worst-case vertical error = $+0.19/-0.36$ mm, std. deviation = 0.13 mm (over both wavelengths) Worst-case Image precision = $+0.23$ mm/-0.90 mm (over both wavelengths) Average percent errors of drawn ellipses were 0.4% in diameter, 0.7% in area, and 0.3% in circumference.
3. OA/US Co-Registration	Evaluate co-alignment of fused OA, US images	None	Worst-case horizontal error, $-0.94/+0.85$ mm, std. deviation = 0.33 mm Worst-case vertical error = $-0.17/+0.21$ mm, std. deviation = 0.09 mm (over both wavelengths)

Test	Purpose	Acceptance Criteria	Results
1. Spatial Resolution	Evaluate and quantify in-plane (axial, lateral), and out-of-plane (elevational) spatial resolution of OA images	None	Axial resolution, short wavelength = 0.58 ± 0.10 mm Axial resolution, long wavelength = 0.54 ± 0.10 mm Lateral resolution, short wavelength = 1.12 ± 0.35 mm Lateral resolution, long wavelength = 1.09 ± 0.25 mm Elevational resolution, short wavelength = 2.7 ± 0.4 mm (at ~4.5 mm depth), 5.6 ± 0.2 mm (at ~34 mm depth), respectively Elevational resolution, long wavelength = 2.7 ± 0.4 mm (at ~4.5 mm depth), 5.4 ± 0.5 mm (~34.5 mm depth), respectively
4. Image uniformity	Quantify horizontal and vertical signal uniformity	None	With clinical contrast settings, Horizontal variation <2 dB (near-field) or <6 dB (far-field). Vertical uniformity < 31 dB (near-field), < 35 dB (far-field) due to expected effect of optical attenuation vs. depth.
5. Depth detection	Quantify the maximum penetration depth of OA imaging	None	Max imaging depth > 4 cm for OA Short and OA Long for detecting deoxygenated and oxygenated vessels, respectively.
6. Sensitivity-linearity	Evaluate and quantify OA sensitivity to and linearity vs. optical absorption	None	OA signal is highly linear vs. changes in target absorption coefficient using fixed contrast settings ($R^2 > 0.986$). Sensitivity = 0.046 cm^{-1} /grayscale for OA Short, 0.0343 cm^{-1} /grayscale for OA Long using fixed contrast settings. Clinical contrast settings can result in saturation effects, as expected.
7. OA Dynamic Range	Evaluate and quantify the OA	None	OA dynamic range (OA Short): 9.079-22.976 dB OA dynamic range (OA Long): 10.015-23.923 dB
8. Accuracy of vessel diameters for simulated blood vessels	Evaluate accuracy of measuring tube diameters with the Imagio® OA imaging	None	Device detects differences in vessel diameters correlated with known diameters, although measured diameter values are overestimated.

Test	Purpose	Acceptance Criteria	Results
1. Spatial Resolution	Evaluate and quantify in-plane (axial, lateral), and out-of-plane (elevational) spatial resolution of OA images	None	Axial resolution, short wavelength = 0.58 ± 0.10 mm Axial resolution, long wavelength = 0.54 ± 0.10 mm Lateral resolution, short wavelength = 1.12 ± 0.35 mm Lateral resolution, long wavelength = 1.09 ± 0.25 mm Elevational resolution, short wavelength = 2.7 ± 0.4 mm (at ~ 4.5 mm depth), 5.6 ± 0.2 mm (at ~ 34 mm depth), respectively Elevational resolution, long wavelength = 2.7 ± 0.4 mm (at ~ 4.5 mm depth), 5.4 ± 0.5 mm (~ 34.5 mm depth), respectively
9. Oxygen saturation variation	Quantify the ability of the Imagio® system to depict and resolve differences in blood oxygen saturation	None	Device can detect SO_2 differences of $\sim 3.5\% \pm 2.3\%$ for targets near 100% SO_2 , $\sim 11\% \pm 15\%$ for targets near 50% SO_2 .
10. Out-of-plane absorber effects	Evaluate and quantify the effects of out-of-plane absorbers on the OA colorization	None	Out-of-plane absorber artifacts can potentially affect OA image colorization. Effects are stronger when out-of-plane target has similar SO_2 and smaller vessel diameter relative to in-plane target.
11. Dual wave-length laser energy variation	Quantify the effects of varying laser energy on image quality	None	Device maintains accurate colorization even with reduced laser energy beyond allowable tolerances for operation.

I Clinical Study Summary

Reader-02 was a single arm, sequentially read, controlled, blinded, multi-reader, multi-case (MRMC) pivotal study to establish a reasonable assurance of Imagio® Breast Imaging System safety and effectiveness. The primary goal of the Reader-02 was to demonstrate that readers using the full functionality of Imagio® Breast Imaging System (IUS+OA) performed better compared to when using IUS alone in terms of specificity at a fixed sensitivity of 98%. The images used in Reader-02 were a subset of the images acquired in a previous study called the PIONEER Study. PIONEER Study data collection is described next.

The Reader-02 Pivotal Study's endpoints were as follows:

Primary endpoint: Evaluate reader specificity with IUS alone at 98% sensitivity versus Imagio® Breast Imaging System (IUS+OA) at the same sensitivity.

Secondary endpoints: Evaluate the following for IUS alone versus Imagio® Breast Imaging System (IUS+OA):

- a. Negative Likelihood Ratio (NLR) defined as $((1 - \text{sensitivity}) / \text{specificity})$
- b. Positive Likelihood Ratio (PLR) defined as $(\text{sensitivity} / (1 - \text{specificity}))$
- c. Partial ROC AUC (pAUC) corresponding to 95-100% sensitivity

Sensitivity and specificity for the secondary endpoints were based on reader's probability of malignancy (POM) scores with a positivity threshold of POM of 2%, i.e., $\text{POM} > 2\%$ was considered a positive read (positive result) and $\text{POM} \leq 2\%$ was considered a negative read (negative result).

The evaluation used a sequential hierarchical approach for hypothesis testing, with secondary endpoints tested in the order NLR, PLR, and pAUC.

Corresponding Hypothesis testing were as follow:

Primary hypothesis test:

$H_0: S_{\text{Imagio}^\circ} = S_{\text{IUS}}$ (no specificity [at 98% sensitivity] difference) $H_1:$

$S_{\text{Imagio}^\circ} \neq S_{\text{IUS}}$ (superior specificity [at 98% sensitivity])

where S_{Imagio° and S_{IUS} represent specificity [at 98% sensitivity], values associated with Imagio® Breast Imaging System (IUS + OA) and IUS, respectively.

Secondary hypothesis tests:

A 2% Probability of Malignancy (POM) cutoff was used as the positivity threshold. Specificity was defined as the proportion with a negative result ($POM \leq 2\%$) among all benign+TPB (truth panel benign) masses (to include all high-risk masses). Please refer to the description below for the definition of TPB and high-risk masses. Sensitivity was defined as the proportion with a positive result ($POM > 2\%$) among all malignant masses.

a. NLR:

$H_0: NLR_{IUS} = NLR_{Imagio®}$ vs

$H_A: NLR_{IUS} \neq NLR_{Imagio®}$ representing a reduction (improvement in NLR)

b. PLR:

$H_0: PLR_{IUS} = PLR_{Imagio®}$ vs

$H_A: PLR_{IUS} \neq PLR_{Imagio®}$ representing an increase (improvement in PLR)

c. Partial ROC AUC:

$H_0: pAUC_{IUS} = pAUC_{Imagio®}$ vs

$H_A: pAUC_{IUS} \neq pAUC_{Imagio®}$ representing an increase (improvement in partial ROC AUC) A sequential hierarchical approach for hypothesis testing was applied.

PIONEER Study Data Collection

The subjects for PIONEER Study were selected from women referred for a diagnostic breast ultrasound work-up who had a suspicious finding within the previous 45 business days, by palpation or by a screening mammogram or diagnostic methodology other than ultrasound, who were scheduled to undergo or already had a Clinical Diagnostic Ultrasound (CDU). The CDU at the time of Study enrollment was used to classify subjects as NDU (Negative Diagnostic Ultrasound) (CDU BI-RADS 3) or PDU (Positive Diagnostic Ultrasound) (CDU BI-RADS $\geq 4a$) to determine if follow-up (NDU patients) or biopsy (PDU patients) was required.

1 Clinical Inclusion and Exclusion Criteria

Enrollment in the PIONEER study was limited to patients who met the following inclusion criteria (all conditions are to be met to be included in the study):

- a. Has signed and dated informed consent, prior to initiation of any Study activities.

- b. Has had an undiagnosed suspicious finding within the previous 45 businessdays, by palpation or by a screening or diagnostic methodology other than ultrasound; this may have included more than one suspicious mass.
- c. Has at least 1 or up to 3 pre-selected and undiagnosed breast massesincluding suspicious solid masses and/or complex cystic and solid masses that the investigator had characterized as either BI-RADS 3, BI-RADS 4, or BI-RADS 5 that have been scheduled for either biopsy or follow-up.
- d. Has at least one undiagnosed breast mass that was detected by one of thefollowing 4 methodologies within 45 business days prior to enrollment with imaging results available for Study utilization:
 - Call backs for additional evaluation of suspicious area(s) identified by imaging other than ultrasound.
 - Diagnostic referral to assess focal physical symptoms and/ or signs that were either a chief complaint of the subject or were elicited by the healthcare practitioner (excluding focal breast pain in the absence of other positive clinical findings).
 - Interval clinical problems (symptoms or physical findings, excluding isolated focal breast pain, that had developed between yearly mammograms).
 - Other referrals to CDU including subjects younger than 30 years old for a clinically suspicious area, or subjects referred from a screening MRI because of an abnormality.
- e. Is at least 18 years of age.
- f. Has received a recommendation to either biopsy or not biopsy.
- g. Is willing and able to comply with protocol required procedures.

Patients were not permitted to enroll in the PIONEER study if they met any of the following exclusion criteria (any condition by itself is sufficient to exclude a subject):

- a. Subject is male.
- b. Has a condition or impediment which could interfere with the intended field-of-view (i.e., breast implants within the previous 12 months, or tattoos).
- c. Has or has had cancer in the ipsilateral breast or prior breast surgeries in the same quadrant of the ipsilateral breast that would have interfered with the ability to capture or interpret images.
- d. Prior benign excisional breast biopsy within the immediate vicinity of the currently evaluated suspicious mass within the past 18 months (benign excisional biopsy not within immediate Imagio® Breast Imaging System field-of-view will not exclude the subject from the Study).
- e. Has greater than 3 suspicious masses.

- f. Mass(es) of interest is greater than 4 cm.
- g. Has all mass(es) characterized as BI-RADS 1 and/or 2 as determined using a CDU.
- h. Currently has mastitis.
- i. Has focal pain without thickening or mass.
- j. Is pregnant or lactating.
- k. Has open sores including insect bites, rash, poison ivy, and chafing on the skin of the ipsilateral breast.
- l. Has an acute or a chronic hematoma and/or acute ecchymosis of the ipsilateral breast.
- m. Is experiencing photo-toxicity associated with currently taking, or having taken, photosensitizing agents within the previous 72 hours such as sulfonamides, ampicillin, tetracycline. Is currently undergoing phototherapy.
- n. Has a history of any photosensitive disease (e.g., porphyria, lupus erythematosus) or undergoing treatment for a photosensitive disease and was experiencing photosensitivity.
- o. Concurrent neoadjuvant therapy prior to the Imagio® Breast Imaging System evaluation or the biopsy.
- p. Has previously had image guided core biopsy, image guided DVAB, or surgical biopsy of the mass of interest.
- q. Has nipple rings that cannot be removed or are not removed during Imagio® Breast Imaging System evaluation.

2 **Methodology**

Subjects were prospectively enrolled upon confirmation of being BI-RADS 3, BI-RADS 4, or BI-RADS 5, as confirmed by CDU evaluation of the suspicious mass(es). Subjects who had all mass(es) characterized as BI-RADS 1 and/or 2 as determined using a CDU were regarded as screen failures. Of the remaining subjects who consented (n=2105), 23 were considered as enrollment failures because (1) mass was too deep for Imagio® Breast Imaging System to visualize, or (2) device malfunctioned prior to or during the scanning procedure, or (3) subject was withdrawn prior to scanning procedure initiated. Remaining subjects (n=2082) were exposed to OA.

The first 10 subjects scanned at the Cancer Therapy and Research Center (CTRC) and the next 100 subjects enrolled as pilot cases for the study were excluded from the Safety Population of the study. The PIONEER Safety Population thus consisted to 1972 subjects.

Subjects were scheduled for an Imagio® Breast Imaging System procedure after the decision whether to biopsy was made, but prior to any biopsy. The subjects must have had the Imagio®

Breast Imaging System procedure within 10 days of the enrollment visit and within 45 business days before the biopsy. Subjects in the NDU group who opted not to undergo a biopsy were to have second Imagio® Breast Imaging System procedure 12 months (± 30 days) after the initial imaging procedure.

All images, both OA and IUS, were to undergo QA review by the quality assurance radiologists (QARs). To ensure image reading consistency, the assessment of all Study imaging, including OA, was managed by an independent Imaging Core Laboratory (ICL).

Starting with the PIONEER Safety Population, 233 subjects were excluded from the PIONEER Intent-to-Diagnose (ITD) Population for the following reasons: QA failure of images, technical failures, PDU with no biopsy, NDU with no truth panel review, and protocol deviations.

3 Ground Truth Determination

For masses that underwent biopsy, the final diagnosis depended on the histopathological examination of the biopsy data from the suspected mass by an independent central histopathologist. Samples determined by the central histopathologist to be high-risk (HR), such as atypical ductal hyperplasia, atypical lobular neoplasia, and lobular carcinoma in situ were considered to be benign in the study effectiveness analysis. A Truth Panel was established to review and/or adjudicate NDU cases with follow-up at 12 months (≥ 11 months) to determine if the cases were True Negative (Truth Panel Benign [TPB]). TPB was determined for subjects who had masses classified as BI-RADS 3 with CDU by site investigators that were not biopsied and had $\leq 20\%$ increase in max diameter and no change in BI-RADS classification from 3 to 4a+ at the 12-month follow up. Panel members evaluated available imaging modalities in the following order: CDU, mammography, MRI and IUS. OA was not utilized for this determination. NDU subjects who were not biopsied within the 12-month follow-up window and showed an increase in mass size ($>20\%$) or a BI-RADS change to >3 were classified as truth panel change (TPC) masses by the Truth Panel. There were 8 such TPC masses, all of which were BI-RADS 3 by CDU at 12-month follow-up; none were recommended for biopsy by site investigators. The PIONEER ITD Population included 1739 subjects with a total of 652 biopsied cancer, 41 biopsied high risk, and 848 biopsied benign, 190 non-biopsied TPB and 8 non-biopsied "other" (truth panel change or

TPC) masses. PIONEER ITD analysis population excluded the 8 TPC cases because they could not be classified as TPB, high-risk, or cancer.

Case selection for Reader-02 Pivotal Study

Reader-02 Pivotal Study used a subset of the images acquired for the PIONEER Pivotal Study. No new subjects were enrolled for Reader-02 Study. The patient population for Reader-02 Study is described in the following.

Masses were selected at random for the Reader-02 Study in proportion to the original assignment distribution of BI-RADS classifications among subjects in the PIONEER Study by conventional diagnostic ultrasound (CDU). Masses selected from the PIONEER ITD analysis population for the Reader-02 Study could include up to 7 blocks of 120 images. The Reader-02 Pivotal Study was planned to consist of between 480 to 840 masses with complete imaging readsets from the original PIONEER ITD analysis population. The data were organized and presented to readers in blocks of 120 masses, each consisting of 72 benign plus 3 high risk (to be categorized as benign) and 45 malignant masses (reflecting a similar prevalence of cancer as the overall PIONEER ITD analysis population, approximately 38%). To facilitate the alignment of the PIONEER Pivotal Study data with the Reader-02 Pivotal Study data in terms of mammogram availability, the mass image set sampling plan selected a benign mass proportion with and without mammograms depending on availability of the mammograms to be the same as in PIONEER; this stratification did not apply for malignant masses where nearly all masses were previously evaluated using mammography. Sample size was assessed at a blinded interim analysis after the first 360 reads and could be adjusted as necessary based on inter- and intra- reader variance up to a maximum of 840 reads.

The dataset determined by the radiologists in the MRMC Study consisted of 480 (four blocks of 120 read sets each), and masses in total, comprised of the following:

- 180 malignant masses
- 300 benign masses (288 benign, 12 high risk defined as atypical ductal hyperplasia, atypical lobular neoplasia, and lobular carcinoma in situ, etc.)

To avoid any potential for bias, all four blocks had been read by all readers in advance of the database lock for the pre-planned interim analysis.

Mass Inclusion and Exclusion Criteria

The mass inclusion criteria for the Reader-02 Pivotal Study were as follows:

- a. One analyzable mass per patient: BI-RADS 3, 4a, 4b, 4c and 5 masses as declared by clinical site investigator via PIONEER study inclusion criteria and categorized as BI-RADS 3, 4a, 4b, 4c, and 5 by CDU
- b. Masses declared to be in the PIONEER ITD analysis population, including high risk cases per original PIONEER protocol
- c. Patient age, indication for study entry and available medical history
- d. Evaluable mammograms (when available) and IUS and OA video loops and still images for each mass

The mass Exclusion criteria for the Reader-02 Pivotal Study were as follows:

- a. Critical missing IUS or OA still image and/or video loop views or incorrect IUS or OA stills and video loops that would preclude a case from being evaluated by readers
- b. Reader-02 proficiency test and training cases
- c. Failure of quality assurance review, as described below.

Quality Assurance Review

The quality assurance radiologists (QARs) were two physicians selected by Seno Medical with knowledge and experience in breast imaging. In addition to exam quality checks, the QAR's role was to identify and label, with guidance from limited medical history data and breast MRI and CDU exams, the appropriate intent-to-diagnose mass on mammogram and IUS and Imagio® Breast Imaging System (IUS+OA) exams to be read and scored by the 15 independent readers on this study.

Reader Qualifications and Training

The study included 15 readers with an additional 5 back-up readers depending on qualifications and availability. Readers that participated in any previous study were not eligible to participate as independent readers in this Reader-02 Pivotal Study. The readers were required to undergo training before they were permitted to participate in the Reader-02 Pivotal Study.

Training consisted of three modules: Didactic training, interactive reading, and test. Didactic training included fundamentals of OA, OA feature scoring, correlation of OA features with core biopsy histopathology, OA artifacts, IUS feature scoring, Seno learnings from previous studies, and the use of the SenoGram. In the interactive reading module, the readers read and scored a mixture of up to 30 malignant and benign cases after being trained on how to use a reading

station, draw regions of interest (ROIs), score IUS and OA features, use the SenoGram to aid in predicting OA POM and BI-RADS category. In the test module, the readers had to pass a proficiency test involving the scoring and interpretation of 30 cases before starting their study reads.

Image Interpretation

Interpretation of each case consisted of two consecutive reads: Read 1, immediately followed by Read 2 within the same reading session. In both Read 1 and Read 2, the radiologist provided a POM rating between 0 and 100 using a custom graphical interface with zoomed-in gradations at low probabilities ($POM < 2\%$), and a corresponding BI-RADS score (BI-RADS 2, 3, 4a, 4b, 4c or 5).

Read 1 reflected the typical information available to a radiologist when evaluating standard ultrasound images, taking into consideration the mass, patient history and assessing mammogram BI-RADS results, when available. Read 1 (IUS reads) served as the control representing current clinical practice.

- Read 1:
 - o Data provided for the reader: Patient History (age) + Mammogram (if available) + IUS (stills and videos provided).
 - o Reader output: IUS Probability of Malignancy (POM) and BI-RADS category assigned in the data form, then locked.

Read 2 displayed the OA images in addition to the data provided in Read 1. Each reader entered 5 feature scores for IUS, 5 feature scores for OA, and four other features (patient age, mass size, depth to the posterior aspect of the mass and mammographic BI-RADS classification) into the SenoGram report form. Based on this information, the SenoGram displayed a predicted Likelihood of Malignancy (LOM) computed from reader input. Each reader then assigned final POM and BI-RADS scores.

- Read 2:
 - o Data initially provided for the reader: Patient History (age) + Mammogram (if available) + IUS (stills and videos provided), and Imagio® Breast Imaging System (IUS+OA) (stills and videos provided).
 - o Input provided by the reader to SenoGram: 5 feature scores for IUS, 5 feature scores for OA, and four other features.
 - o Reader output: Imagio® Breast Imaging System (IUS+OA) POM and BI-RADS category after viewing the SenoGram output. The data form is then locked.

Study Population

Of the 480 masses read in the Reader-02 study, the overall mean age was 49.9 ± 14.4 years; in the benign+TPB+High Risk (HR) group, the mean was 44.2 ± 12.5 years, and in the malignant group it was 59.4 ± 12.2 years, consistent with the age distribution of cancer diagnosis in the population. Almost ninety percent (89.6%) of subjects had available mammography, of which 84.0% were benign+TPB+HR subjects and 98.9% were cancer subjects.

In the overall Intent to Diagnose (ITD) population, two (0.4%) had CDU BI-RADS scores of 2; 74 (15.4%) were score 3; 129 (26.9%) were score 4a, 84 (17.5%) were score 4b, 87 (18.1%) were score 4c, and 102 (21.3%) were score 5. As expected, a greater proportion of subjects in the malignant group were scored BI-RADS 5 (96, 53.3%). Almost three-quarters of the masses (355, 74%) had a mammographic breast density of 2 or 3; 210 masses (43.8%) were palpable, and 199 masses (41.5%) were not. Almost half the subjects (219, 45.6%) were post-menopausal, and 14 (2.9%) had breast implants.

Safety and Effectiveness Results

1. Safety Results

The analysis of safety was based on the PIONEER Safety cohort of 1972 patients. Summary of the adverse effects that occurred in the PMA clinical study is as following:

Overall, 52 subjects (2.6%) in the PIONEER Safety Population experienced an Adverse Event (AE). Forty subjects (2.0%) experienced at least one AE assessed as mild and 11 subjects (0.6%) experienced at least one AE assessed as moderate in severity. One subject (0.1%) reported two AEs assessed as severe and serious. This subject experienced atrial fibrillation and congestive heart failure, both assessed as serious, severe, and not related to the procedure. The event of atrial fibrillation was attributed to renal failure and the event of congestive heart failure was attributed to atrial fibrillation.

The most common AEs by system organ class were injury, poisoning, and procedural complications (21 events) followed by nervous system disorders (16 events). The most common events by preferred term (PT) were paresthesia (10 events) and post procedural hematoma (all

post-biopsy; 5 events). Twenty-five events, including 10 events of paresthesia, had a start date of the same day as the OA procedure.

Of the Serious Adverse Events (SAE)s reported in 5 subjects, none were considered related to the OA procedure. There were no serious adverse device effects (SADEs), unanticipated adverse device effects (UADEs) or deaths reported during the Study.

Of the 52 subjects with AEs, 10 subjects (0.5%) reported 11 events that were considered related (possibly, probably, likely) to the OA procedure. All 11 procedure-related AEs were considered mild.

A single event of "burns second degree" occurred. This patient had a moderate skin rash of undetermined origin. This was deemed by the sponsor to be more likely due to contact dermatitis with post-biopsy tape than a photothermal skin injury due to laser exposure.

2. Effectiveness Results

The following conclusions are based upon the results of the pivotal study:

The Reader-02 Pivotal Study met its primary endpoint and demonstrated that Imagio® Breast Imaging System has better specificity than IUS at fixed sensitivity of 98%. Specificity at a fixed sensitivity of 98% (fSp) and partial area under the ROC curve (pAUC) between sensitivities of 95% and 100% are metrics that assess the part of the ROC curve where the clinical decisions are actually made. Diagnostic likelihood ratios (DLR) can be calculated from sensitivity and specificity based on POM scores and can be useful in comparing technologies in breast imaging.

Specificity at a fixed sensitivity of 98% was analyzed using MRMC analysis. Mean (average over all readers) fSp was found to be higher with statistical significance (two-sided $p=0.027$) for IUS+OA (47.2%, 95% CI=[35.9%,58.5%]) compared to IUS alone (38.2%, 95% CI= [24.9%, 51.6%]), with a difference in fSp of 9.0% with 95% CI=[1.0%, 17.0%]. Thus, IUS+OA achieved the primary endpoint in the Reader-02 Pivotal Study. When the empirical ROC curve is used, interpolation is typically required to estimate fSp because 98% sensitivity may not fall onto the intrinsic grid for the empirical ROC curve.

The observed mean NLR was 0.047 (95% CI: 0.032, 0.062) for IUS+OA suggesting that averaging over all 15 readers in the study a negative test read (POM $\leq 2\%$) was observed about 21 (i.e., $1/0.047$) times more often among non-cancer cases, compared to those with cancer, and the observed mean NLR was 0.053 (95% CI: 0.037, 0.070) for IUS alone suggesting that averaging over all 15 readers in the study a negative read (POM $\leq 2\%$) was observed about 19 (i.e., $1/0.053$) times more often among non-cancer cases, compared to those with cancer. The decrease (i.e., improvement with IUS+OA compared to IUS alone) in NLR could not be established with statistical significance because the observed relative NLR (the ratio in NLR for IUS+OA and IUS) was 0.896 with a 95% CI= (0.693, 1.11) which included 1 indicating that no evidence of a difference in NLR was found. The confidence intervals above for NLR and the relative NLR do not take into consideration the variability in the reader population and are therefore applicable only to the set of radiologists who took part in the reader study. Since a sequential hierarchical testing to control the study type I error rate was pre-specified, and this was the second hypothesis test in the sequence, and decrease in NLR was not met (i.e., an improvement in NLR for IUS+OA compared to IUS alone could not be shown), not only improvement in NLR for IUS+OA compared to IUS alone cannot be claimed, but hypotheses test results from subsequent hypotheses (for the remaining secondary endpoints of PLR and partial ROC AUC) are not

reported. The results reported below for PLR and Partial ROC AUC are considered descriptive or non-confirmatory.

Based on descriptive statistics that do not control type I error and that cannot be generalized outside this particular study, the observed mean PLR was 1.959 (95% CI: 1.870, 2.051) for IUS+OA (only as a descriptive result this suggests that averaging over all 15 readers in the study a positive test read (POM>2%) was observed about 2 times more often among cases with cancer, compared to those without cancer), and the mean PLR was reported as 1.548 (95% CI: 1.498, 1.597) for IUS alone (only as a descriptive result this suggests that averaging over all 15 readers in the study a positive test read (POM>2%) was observed about 1.5 times more often among cases with cancer, compared to those without cancer). The descriptive observed relative PLR was 1.281 (95% CI: 1.231, 1.298). The confidence intervals above for PLR and the relative PLR do not take into consideration the variability in the reader population and are therefore applicable only to the set of radiologists who took part in the reader study.

pAUC was lower in the hierarchical test order than NLR, which failed to achieve significance. Consequently, pAUC results are not part of the claim structure and are reported as descriptive statistics. The mean unscaled pAUC was 0.0244 (95% CI: 0.0230, 0.0258) for IUS+OA and 0.0205 (95% CI: 0.0191, 0.219) for IUS alone, a difference of 0.0039. All readers had a larger point estimate of pAUC for IUS+OA than for IUS alone.

The following results are not part of the claim structure and are reported as only descriptive and non-confirmatory results. POM scores were analyzed by computing the mean score among the 15 readers for a given mass with IUS only or with IUS+OA. For IUS+OA, the mean POM for malignant masses was 70.7 (95% CI: 66.86, 73.28) whereas the mean POM for benign masses was 15.90 (95% CI: 13.65, 18.15). For IUS alone, the mean POM for malignant masses was 65.23 (95% CI: 62.38, 68.08) whereas the mean POM for benign masses was 16.41 (95% CI: 14.54, 18.28). The confidence intervals provided above apply only to the average POM score of the 15 readers who participated in the study, and do not generalize to the average POM scores of other readers or the POM scores of individual readers.

SenoGram classification performance was assessed with descriptive (non-confirmatory) ROC metrics (fSp, pAUC) and classification metrics (sensitivity and specificity) observed on the Reader-02 dataset. fSp and pAUC were assessed based on the SenoGram inputs assigned by each reader in Reader02, and then averaged over the 15 readers. Sensitivity and specificity were assessed for each reader at a threshold of 2% for the SenoGram POM output and then averaged. Using endpoint interpolation, the average fSp was 44.1% (95% CI: 38.4%, 49.8%) and the

average pAUC was 0.0232 (95% CI: 0.0215, 0.0249). The average SenoGram sensitivity and specificity were 97.0% (95% CI: 96.0%, 97.9%) and 50.9% (95% CI: 46.0%, 55.8%), respectively.

The binary agreement between readers and SenoGram predictions was measured by comparing results for each case. Reader and SenoGram results agree if both predict positive (cancer) or both predict negative (benign). Readers agreed with the SenoGram results in 98.8% (2667/2700) of cancer cases, 93.3% (4197/4500) of benign+TPB+HR cases, and 95.3% (6864/7200) overall.