

# **User Manual**

# **Ambulatory ECG Monitor**

**Brand name: MEMO Patch 2**

**Model Number: MPT-E14R-UNA03**

**Document Number: UM-D-999**

**Revision Number: 1**

**Date: October 30, 2023**

## 1.0 PRODUCT WARRENTY

### 1.1 Warranty Claim

Contact the HUINNO customer service if you are unable to resolve the issue after reviewing the user manual.

### 1.2 Warranty Coverage

- The warranty provides at no extra cost to the user.
- The device is warranted for a period of 1 year.
- The device is warranted for the functional or performance defects arising when used for normal purposes in accordance with the user manual.

### 1.3 Warranty Exclusion

The warranty is not applicable in any of the following cases:

- Expiry of warranty period and/or inability to check warranty period.
- Failure or damage caused by user's negligent use, neglect, or careless operation of device.
- Failure caused by use of device not in accordance with the user manual.
- Failure or damage caused by using electricity of unauthorized voltage.
- Failure caused by using parts, accessories or consumables that are not approved by the manufacturer.
- Product or its parts arbitrarily removed, altered, modified or damaged.
- Product serviced and/or decomposed by unauthorized personnel that are not designated by the manufacturer.
- Service fee may apply for services irrelevant to product defects (e.g., product training, irregular inspection, Bluetooth connection problem due to external environment, defect due to using third-party products and/or software) regardless of the warranty period.

### 1.4 Warranty Period

- Warranty period refers to the period in which the manufacturer or authorized seller is obliged to replace the quality, performance, functional defects from normal use for free.
- The warranty becomes effective at the date of purchase. Please retain the product warranty card or the proof of purchase. If you do not have your warranty card or proof of purchase, your warranty will start 90 days after the date of manufacture, according to the manufacturer's records.
- The warranty is confined to the first purchaser of the product at an authorized dealer.
- The warranty is not applicable to second-hand products or products purchased from an unauthorized dealer. The manufacturer will not be responsible for the compensation of damage for the replacement and service of those products.

- The warranty for products delivered under a separate contract with the manufacturer follows the contents of the contract.

### 1.5 Warranty Card

<Warranty Card>

Product Name:

Model Number:

Serial Number:

Date of Purchase:

Place of Purchase:

Warranty Period: 1 year from date of purchase

Client Name:

Organization:

Phone Number:

This is to certify that this device has passed the strict quality control and comprehensive inspection.

Replacement and service may be denied in any of the following cases.

1. Unable to perform replace or provide service due to the user's intention and negligence
2. Unable to replace due to discontinuation of parts after the warranty period
3. Damage resulting from a force majeure event such as fire, explosion, storm, flood, earthquake, or other natural disasters
4. Removal, obliteration, or alteration of identification labels (Model Number, Serial Number etc.) of the product

## Table of Contents

|            |                                                      |           |
|------------|------------------------------------------------------|-----------|
| <b>1.0</b> | <b>PRODUCT WARRENTY .....</b>                        | <b>2</b>  |
| 1.1        | Warranty Claim.....                                  | 2         |
| 1.2        | Warranty Coverage .....                              | 2         |
| 1.3        | Warranty Exclusion .....                             | 2         |
| 1.4        | Warranty Period .....                                | 2         |
| 1.5        | Warranty Card .....                                  | 3         |
| <b>2.0</b> | <b>INSTRUCTION.....</b>                              | <b>7</b>  |
| <b>3.0</b> | <b>MANUFACTURER INFORMATION .....</b>                | <b>7</b>  |
| <b>4.0</b> | <b>DEVICE OVERVIEW .....</b>                         | <b>7</b>  |
| 4.1        | Product Description.....                             | 7         |
| 4.2        | Intended Use.....                                    | 7         |
| 4.3        | Indications for Use.....                             | 7         |
| 4.4        | Indication.....                                      | 7         |
| 4.5        | Intended Environment of Use.....                     | 8         |
| 4.6        | Intended Patient Population.....                     | 8         |
| 4.7        | Patient Contacting Components .....                  | 8         |
| <b>5.0</b> | <b>MARKS AND SYMBOLS .....</b>                       | <b>8</b>  |
| <b>6.0</b> | <b>SAFETY REQUIREMENTS.....</b>                      | <b>10</b> |
| 6.1        | Contraindications .....                              | 10        |
| 6.2        | Warnings.....                                        | 10        |
| 6.3        | Precautions.....                                     | 11        |
| 6.4        | IT Network .....                                     | 12        |
| 6.5        | Environmental Protection.....                        | 12        |
| <b>7.0</b> | <b>LABELING AND PACKAGING .....</b>                  | <b>12</b> |
| 7.1        | Device Label .....                                   | 12        |
| 7.2        | Electrode Bracket (60 mm and 100 mm same shape)..... | 13        |
| 7.3        | Package (Gift Box) Label .....                       | 14        |
| 7.4        | Accessory Label .....                                | 15        |
| 7.5        | Accessory Package (Gift Box) Label.....              | 15        |
| 7.6        | Package Components.....                              | 16        |
| <b>8.0</b> | <b>DEVICE COMPONENTS .....</b>                       | <b>17</b> |
| 8.1        | Ambulatory ECG Monitor .....                         | 17        |

|      |                                                                                                                                                                                                        |    |
|------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----|
| 8.2  | ECG Data Transferred Cradle (1port) .....                                                                                                                                                              | 18 |
| 8.3  | Attaching the 'Ambulatory ECG Monitor' .....                                                                                                                                                           | 19 |
| 8.4  | Applying the Ambulatory ECG Monitor: Step-by-step guide.....                                                                                                                                           | 20 |
| 8.5  | Check the quality of the ECG signal and finalize the examination.....                                                                                                                                  | 25 |
| 8.6  | Delivery of instructions and supplies to patients .....                                                                                                                                                | 26 |
| 9.0  | INSTRUCTION TO FOLLOW DURING MONITORING .....                                                                                                                                                          | 27 |
| 10.0 | PROCEDURES AFTER COMPLETING THE TEST.....                                                                                                                                                              | 27 |
| 10.1 | Patient: Detaching and returning the 'Ambulatory ECG Monitor' .....                                                                                                                                    | 27 |
| 10.2 | Checking the 'Ambulatory ECG Monitor', processing returns, and uploading data .....                                                                                                                    | 27 |
| 11.0 | PRODUCT OPERATING ENVIRONMENT .....                                                                                                                                                                    | 29 |
| 12.0 | TROUBLE SHOOTING GUIDE.....                                                                                                                                                                            | 30 |
| 12.1 | Trouble shooting.....                                                                                                                                                                                  | 30 |
| 12.2 | FAQs .....                                                                                                                                                                                             | 34 |
| 13.0 | DEVICE SPECIFICATION .....                                                                                                                                                                             | 37 |
| 13.1 | Performance Characteristics.....                                                                                                                                                                       | 37 |
| 13.2 | Electrical Characteristics.....                                                                                                                                                                        | 38 |
| 13.3 | Power Characteristics .....                                                                                                                                                                            | 38 |
| 13.4 | Physical Characteristics .....                                                                                                                                                                         | 39 |
| 13.5 | Environmental Characteristics.....                                                                                                                                                                     | 39 |
| 13.6 | Product Maintenance, Storage and Disposal Summary.....                                                                                                                                                 | 39 |
| 13.7 | Cleaning and Maintenance.....                                                                                                                                                                          | 40 |
| 13.8 | Storage and Transport.....                                                                                                                                                                             | 40 |
| 13.9 | Disposal (Accessory).....                                                                                                                                                                              | 40 |
| 14.0 | DEVICE LED SCENARIO .....                                                                                                                                                                              | 40 |
| 15.0 | HEART RATE CALCULATION.....                                                                                                                                                                            | 41 |
| 16.0 | PAUSE DETERMINATION .....                                                                                                                                                                              | 42 |
| 17.0 | ELECTRICAL SAFETY AND COMPATIBILITY .....                                                                                                                                                              | 42 |
| 17.1 | Table1 – ELECTROMAGNETIC EMISSIONS for MEMO Patch 2 (MPT-E14R-UNA03).....                                                                                                                              | 42 |
| 17.2 | Table2 – Electromagnetic IMMUNITY for MEMO Patch 2 (MPT-E14R-UNA03).....                                                                                                                               | 43 |
| 17.3 | Table3 – Electromagnetic IMMUNITY for MEMO Patch 2 (MPT-E14R-UNA03) that are not LIFE-SUPPORTING .....                                                                                                 | 44 |
| 17.4 | Table4 – Recommended separation distances between portable and mobile RF communication equipment and the ME EQUIPMENT or ME SYSTEM – for ME EQUIPMENT and ME SYSTEMS that are not LIFE-SUPPORTING..... | 46 |

|                                                                                                                                         |    |
|-----------------------------------------------------------------------------------------------------------------------------------------|----|
| 17.5 Table5 – Electromagnetic Compatibility Information for Ambulatory ECG Monitor (MPT-E14R-UNA03) & ECG Data transferred cradle ..... | 46 |
| 18.0 BUG LIST .....                                                                                                                     | 48 |
| 19.0 CUSTOMER SERVICE .....                                                                                                             | 48 |
| 20.0 REVISION HISTORY .....                                                                                                             | 49 |

## 2.0 INSTRUCTION

This user manual is intended for reference by Ambulatory ECG Monitor users. The manual includes information about components, instructions for use and safety precautions of Ambulatory ECG Monitor. Please read the user manual before using Ambulatory ECG Monitor. We recommend storing the manual in a safe place near the device.

## 3.0 MANUFACTURER INFORMATION

**HUINNO Co., Ltd.**

**TEL:** +82.2.2051.3161, **FAX:** +82.2.2051.3160

**Address:** 3F, 4F, 5F, 19, Apgujeong-ro 79-gil, Gangnam-gu, Seoul, Republic of Korea

**Webpage:** <https://www.huinno.com/en>

## 4.0 DEVICE OVERVIEW

### 4.1 Product Description

The Ambulatory ECG Monitor is a ECG monitor that provides a continuous, single channel recording for up to 14 days. The Ambulatory ECG Monitor records ECG without patient interaction, with the goal of improving patient compliance via simplicity of operation. Patients have the option of pressing a convenient event marking button which marks the time of symptom. The patient is encouraged to fill out a log to document symptomatic events, which will support symptom rhythm correlation in the Final Report.

At the conclusion of the wear period (up to 14 days), the patient removes the Ambulatory ECG Monitor and returns it by mail or directly to the doctor.

### 4.2 Intended Use

The Ambulatory ECG Monitor is intended to capture, analyze and report symptomatic and asymptomatic cardiac events and continuous electrocardiogram information for long-term monitoring. While continuously recording patient ECG, both patient-triggered and arrhythmia events are saved at the device.

### 4.3 Indications for Use

It is a prescription-only, continuously recording ECG monitor that can be worn up to 14 days. It is indicated for use on patients who may suffer from transient symptoms such as palpitations, shortness of breath, dizziness, light-headedness, pre-syncope, syncope, fatigue, and anxiety or patients who are asymptomatic.

### 4.4 Indication

Arrhythmia

#### 4.5 Intended Environment of Use

It is intended for use in a clinical environment. The product cannot be used in hospital emergency rooms or ambulances, but it is possible for patients to use the product at home during the monitoring period. In addition, the product can be used in places other than the patient's home.

#### 4.6 Intended Patient Population

Age: More than 18 years old

Weight: it's irrelevant

Health: The person who can recognize and record the symptoms that occur while using the product.

Nationality: it's irrelevant

Intended Operator: Patient & Physician

#### 4.7 Patient Contacting Components

ECG electrodes

### 5.0 MARKS AND SYMBOLS

| Symbol                                                                              | Definition           |
|-------------------------------------------------------------------------------------|----------------------|
|  | Serial Number        |
|  | Date of Manufacture  |
|  | Manufacturer         |
|  | BF Type              |
| IP27                                                                                | IP Classification    |
|  | Refer to User Manual |
|  | Caution              |
|  | Prohibitions         |
|  | Handle with care     |
|  | This way up          |
|  | Fragile              |

|                                                                                     |                                                                    |
|-------------------------------------------------------------------------------------|--------------------------------------------------------------------|
|    | Keep dry                                                           |
|    | Temperature Limit                                                  |
|    | Humidity Limit                                                     |
|    | Atmospheric Pressure Limitation                                    |
|    | MR, Magnetic Resonance Unsafe                                      |
|    | Medical Device                                                     |
|    | Authorized representative in the European community/European Union |
|    | Rx only                                                            |
|    | CE Mark                                                            |
|  | FCC Mark                                                           |
|  | UKCA-UK Mark                                                       |
|  | C-Tick Mark                                                        |
|  | Keep away from direct sunlight                                     |
|  | Keep Dry                                                           |
|  | Do not re-use                                                      |
|  | Unique Device Identification                                       |
|  | Non-ionizing electromagnetic radiation                             |
|  | Class II equipment                                                 |
|  | Direct Current                                                     |

## 6.0 SAFETY REQUIREMENTS

### 6.1 Contraindications

1. Please do not examine or treat using radiation such as ultrasound, CT, MRI, and X-ray interfere with the regular operation of the product and lead to inaccurate results of use.
2. Please do not use it at the same time as a defibrillator.
3. Please do not disassemble the device by the user or operator.
4. Please do not lay or cover electrical or magnetic products such as electric blankets, magnetic mats, electric stone beds, and jade mats during the examination.
5. Device is not intended for infants under 10kg.
6. Do not use the device for patients with symptomatic episodes where instance variant in cardiac performance could result in immediate danger to the patient
7. Do not use the device for patients with known history of life-threatening arrhythmias.
8. Do not use the device in combination with external cardiac defibrillators or high frequency surgical equipment near strong magnetic fields or devices such as MRI.
9. Do not use the device on patients who do not have the competency to wear the device for the prescribed monitoring period.

### 6.2 Warnings

1. Do not use the device on patients with known allergic reaction to adhesives or hydrogels or with family history of adhesive skin allergies. Patients may experience skin irritation.
2. Infants or people who cannot express themselves must not use the device.
3. Without the manufacturer's permission, you must not modify the device.
4. Please be aware of the risk of strangulation for infants and children due to batteries and battery opener.
5. Please be careful as children may swallow small parts such as batteries.
6. Direct use of electrical products such as electric blankets and heated mats is prohibited during the prescribed period of using the product, as it may affect the performance of the test results.
7. Do not use the device on patients with known allergic reaction to adhesives or hydrogels or with family history of adhesive skin allergies. Patient may experience skin irritation.
8. If skin irritation such as severe redness, itching or allergic symptoms develop, remove the device from the patient's chest. Call HUINNO Customer Service at +82.2.3443.3160
9. If a user who is not familiar with the proper usage replaces the battery of the product, there is a risk of battery ignition or explosion.

\* CAUTION: Federal law restricts this device to sale by or on the order of a physician

### 6.3 Precautions

1. Read all instructions and labels including this manual before starting to use the device system.
2. Do not use acetone or any other cleaning solvents to clean the device.
3. The device includes temperature and humidity limitations. If exposed, patients may experience degradation of adhesive performance causing the device to slip or fall off during the patient wear duration.
4. The device has a shelf-life data. Use of expired device may cause a degradation of ECG signal quality and/or low battery condition.
5. Keep device and packaging away from young children. Contents may be harmful if swallowed. Device contains button cell batteries that are not accessible during normal use but, if exposed, are known choking hazards and may cause severe tissue injury if ingested.
6. The battery life of the device may be shortened if the device is used frequently and/or for a prolonged period of time.
7. The battery life and capacity may decrease when the device is stored in a high-temperature environment.
8. The battery may self-discharge when the device is in storage.
9. Do not immerse the device into any liquid.
10. Do not expose the device to direct sunlight, heat source of thermal radiation, moisture, vibration, mechanical shock, excessive dust, or humidity.
11. The warranty will be void if the device is opened, disassembled, or altered by any unauthorized personnel.
12. Do not use caustic or abrasive cleaning agents to clean the device.
13. Do not excessively pull or overstress the device as it may break.
14. Do not sit or place a heavy object on the device.
15. Do not use the device if the package is damaged. Devices may not perform as intended.
16. Safety and effectiveness of the device on under 18 years old include pediatric patients has not been established.
17. Keep devices and packaging away from young children. Contents may be harmful if swallowed. The Device contains button cell batteries that are not accessible during normal use but, if exposed, are known choking hazards and may cause severe tissue injury if ingested.
18. Registration errors may result in limited functionality or erroneous ECG reporting. Utmost caution should be applied to ensure that patient registration is accurate and complete.
19. Ensure that the electrodes of the device do not come into contact with other conductive parts. If electrode is not contact well with the skin the performance and accuracy of the ECG test might not accurate.

20. When the abnormality is found during the examination, keep the patient in safe status and stop the examination.
21. Exposure of attachment parts or other accessories of the product for a long period of time may cause skin irritation. If the skin irritation is severe, please contact HUINNO Customer Service at +82.2.3443.3160 immediately.
22. Allergic reactions may occur to the tape on the electrodes used in the product.
23. Ensure that the electrodes and conductive portions of the BF type mounting part, including the neutral electrode, do not come into contact with other conductive parts that include grounding.

#### 6.4 IT Network

The Ambulatory ECG Monitor is connected to the mobile phone via Bluetooth. Through the mobile phone the following can be done:

1. Activate and connect the device with the patient
2. Check real time ECG data through 'MEMO Launcher' app


#### 6.5 Environmental Protection

Non-specialized organizations with responsibility should contact local authorities to determine the proper disposal methods for components and accessories that pose biological hazards.

### 7.0 LABELING AND PACKAGING

#### 7.1 Device Label

| Position      | Device                                                                              | Device Label              |
|---------------|-------------------------------------------------------------------------------------|---------------------------|
| Main Body Top |  | Serial Number<br>PXXXXXXX |

|                     |                                                                                   |                                                                                                                                                                                                                                                                                                                                |
|---------------------|-----------------------------------------------------------------------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Main Body<br>Bottom |  | <div> <b>Device Name</b><br/>           Ambulatory ECG Monitor<br/>           (MEMO Patch 2)         </div> <div> <b>Model Number</b><br/>           MPT-E14R-UNA03         </div> <div> <b>SN</b> PXXXXXXX<br/>  YYYY-MM-DD         </div> |
|                     |                                                                                   |                                                                                                                                                                                                                                             |

## 7.2 Electrode Bracket (60 mm and 100 mm same shape)

| Top                                                                                 | Bottom                                                                               |
|-------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------|
|  |  |

### 7.3 Package (Gift Box) Label

|                          |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                            |
|--------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <p>Device</p>            | <div data-bbox="411 257 906 956"> <h1>MEMO<sup>®</sup> Patch</h1> <p><b>Product Name</b><br/>Ambulatory ECG Monitor<br/>(MEMO Patch 2)</p> <p><b>Model Number</b><br/>MPT-E14R-UNA03</p>  <p>HUINNO Co., Ltd.<br/>3F, 4F, 5F, 19, Apgujeong-ro 79-gil, Gangnam-gu,<br/>Seoul, Republic of Korea (06011)</p> <p>Tel +82-2-3443-3160<br/>Fax +82-2-2051-3160</p> <p><b>R<sub>x</sub>only</b>      <b>Rating</b><br/>3.0VDC, 225mAh (CR2430)</p> <p>CAUTION: U.S. Federal<br/>law restricts this device to<br/>sale by or on the order of<br/>a licensed physician</p> <p><b>SN</b> PXXXXXXX      <b>IP27</b></p> <p><b>UDI</b>  (01)08800126500062<br/>(11)YYMMDD      <b>HUINNO</b></p> </div> <div data-bbox="932 819 1428 952">  </div>                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                               |
| <p>Electrode Bracket</p> | <div data-bbox="635 1001 1209 1281"> <h2>Electrode Bracket 60mm</h2> <p><b>Model Number</b><br/>ACC-B06P-01</p>  <p>HUINNO Co., Ltd.<br/>3F, 4F, 5F, 19, Apgujeong-ro 79-gil, Gangnam-gu,<br/>Seoul, Republic of Korea (06011)</p> <p>Tel +82-2-3443-3160      <b>MD</b>       Do Not Reuse             -20°C</p> <p><b>LOT</b>  (01)0880126500277<br/>(11)YYMMDD      <b>HUINNO</b></p> <p>YYY-MM-DD</p> </div> <div data-bbox="635 1292 1209 1565"> <h2>Electrode Bracket 100mm</h2> <p><b>Model Number</b><br/>ACC-B10P-01</p>  <p>HUINNO Co., Ltd.<br/>3F, 4F, 5F, 19, Apgujeong-ro 79-gil, Gangnam-gu,<br/>Seoul, Republic of Korea (06011)</p> <p>Tel +82-2-3443-3160      <b>MD</b>       Do Not Reuse             -20°C</p> <p><b>LOT</b>  (01)0880126500284<br/>(11)YYMMDD      <b>HUINNO</b></p> <p>YYY-MM-DD</p> </div> |

## 7.4 Accessory Label

|                                   |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                         |
|-----------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| ECG Data<br>transferred<br>Cradle | <div data-bbox="488 264 911 730"> <p><b>ECG Data Transferred Cradle</b></p> <p><b>Model Number</b><br/>ACC-C04P-03</p> <p><b>Serial Number</b><br/>CXXXXXXX</p> <p><b>Country of Manufacture</b><br/>Republic of Korea</p> <p><b>Rating</b><br/>5V<math>\overline{\text{---}}</math> , 0.5A</p>  (01)08800126500017<br/>(11)000000         </div> <div data-bbox="935 264 1350 730"> <p><b>ECG Data Transferred Cradle</b></p> <p><b>Model Number</b><br/>ACC-C01P-03</p> <p><b>Serial Number</b><br/>CXXXXXXX</p> <p><b>Country of Manufacture</b><br/>Republic of Korea</p> <p><b>Rating</b><br/>5V<math>\overline{\text{---}}</math> , 0.5A</p>  (01)08800126500017<br/>(11)000000         </div> |
|-----------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|

## 7.5 Accessory Package (Gift Box) Label

|                                   |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                            |
|-----------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| ECG Data<br>transferred<br>Cradle | <div data-bbox="408 866 871 1357"> <p><b>ECG Data Transferred Cradle</b></p> <p><b>Model Number</b><br/>ACC-C04P-03</p> <p><b>Serial Number</b><br/>CXXXXXXX</p> <p><b>Country of Manufacture</b><br/>Republic of Korea</p> <p><b>Rating</b><br/>5V<math>\overline{\text{---}}</math> , 0.5A</p> <p><b>HUINNO</b></p> </div> <div data-bbox="879 866 1430 1357"> <p><b>HUINNO Co., Ltd.</b></p> <p><b>Address</b><br/>3F, 4F, 5F, 19, Apgujeong-ro 79-gil,<br/>Gangnam-gu, Seoul,<br/>Republic of Korea (06011)</p>  (01)08800126500017<br/>(11)000000         </div>   |
|                                   | <div data-bbox="408 1435 871 1926"> <p><b>ECG Data Transferred Cradle</b></p> <p><b>Model Number</b><br/>ACC-C01P-03</p> <p><b>Serial Number</b><br/>CXXXXXXX</p> <p><b>Country of Manufacture</b><br/>Republic of Korea</p> <p><b>Rating</b><br/>5V<math>\overline{\text{---}}</math> , 0.5A</p> <p><b>HUINNO</b></p> </div> <div data-bbox="879 1435 1430 1926"> <p><b>HUINNO Co., Ltd.</b></p> <p><b>Address</b><br/>3F, 4F, 5F, 19, Apgujeong-ro 79-gil,<br/>Gangnam-gu, Seoul,<br/>Republic of Korea (06011)</p>  (01)08800126500017<br/>(11)000000         </div> |


## 7.6 Package Components

- ⑩ 1 Ambulatory ECG Monitor (MEMO Patch 2)
- ⑩ 1 Quick Guide
- ⑩ 1 Battery Opener
- ⑩ 1 ECG Monitoring Electrodes
- ⑩ Accessory: ECG Data transferred Cradle




## 8.0 DEVICE COMPONENTS

### 8.1 Ambulatory ECG Monitor



## 8.2 ECG Data Transferred Cradle (1port)



### 8.3 Attaching the 'Ambulatory ECG Monitor'



Prepare the device and ECG electrodes.

**NOTE:** The electrode is for one-time use only and should be disposed of after use. Reuse is strictly prohibited. For electrode, follow the instruction of the product.




Press and hold the power button for about 3 seconds to turn on the device.


When you turn on the power, the LED indicator will flash.

**NOTE:** If the device is not paired with the 'MEMO Launcher' App within 5 minutes before the start of examination, the device will be automatically power off.

#### 8.4 Applying the Ambulatory ECG Monitor: Step-by-step guide.




- Launch the 'MEMO Launcher' app and log in with the same credentials provided by HUIINNO.





- Click on '**Start a new examination**' and select the available patch serial for the patient.
- **NOTE:** If a patient is not displayed in the list, please check if the Bluetooth function of your mobile phone is activated.



- Check the serial number of the device to be attached to the patient.
- The serial number is written under the device label.
- NOTE: The serial number is a combination of an English letter "P" and numbers and can be found on the front of the device.







**Note:** Disinfect the attachment site with alcohol before attachment or if there is hair on the attachment site, shave the hair before attaching. If the device is not in close contact with the body due to body hair, it may interfere with the signal measurement.

- 1. Place the device in the middle of the clavicle and attach the electrode 45 degrees to the right
- (Right-handed side of the Medical Doctor). It is recommended to attach the patch in the direction of Lead II, and if the P wave is clear, the signal can be considered excellent.
- 
- 2. Check if the ECG signal appears well in the 'MEMO Launcher' App.
  - If the P wave and QRS complex are not clear, move the patch position slightly to find the optimal position
  - It may take some time for the signal to stabilize after each location change.
- 3. Click the button 'patch attachment completion' in the 'MEMO Launcher' App as well.
- 
- **NOTE:** The signal is for verification purposes and may differ from the actual electrocardiogram data that is being saved.
- 
- 4. To stop the display of real time ECG from the App, click the button 'pause'. If the signal does not come in, check the attachment status again.
- **NOTE:** The recommended interval for replacing the electrode is one day (24 hours).
- Since the patient may remove the patch after returning home, instructions should be given on how to reattach the patch at the location where it was originally applied in the hospital.

## 8.5 Check the quality of the ECG signal and finalize the examination.


- After checking the quality of the ECG signal, press the 'Start examination' button.





If you want to check the signal after starting the examination, log in to the app and select 'Check the signal' on the first screen.

You can check the signal by clicking on the patient's name in the list. However, since the patch and app are connected via Bluetooth, they may not be available if the patch is out of range or if the signal is weak.



## 8.6 Delivery of instructions and supplies to patients

- After completing the examination, provide the patient with the following items:
- Patients note: Record sleep time and symptoms.
- Ambulatory ECG Monitor storage envelope: Used to return the device to the hospital. Please put the detached device into the envelope and return it.

Patient guidebook: The patient guidebook contains instructions for recording symptoms, reattachment, precautions, and contact information.

## 9.0 INSTRUCTION TO FOLLOW DURING MONITORING

During monitoring, the Ambulatory ECG Monitor continuously records ECG information. In addition, to check the frequency and pattern of arrhythmia more accurately, if an arrhythmia symptom occurs, press the symptom record button on the device or record it in the patient note. Sleep time also should be recorded in the patient notes. If a problem occurs during monitoring, contact the Ambulatory ECG Monitor customer center at +82-2-3443-3160.

When undergoing X-ray, CT, or MRI tests, do not wear the device. For more details, please refer to the safety requirements.


## 10.0 PROCEDURES AFTER COMPLETING THE TEST




### 10.1 Patient: Detaching and returning the 'Ambulatory ECG Monitor'



- 1 After the test is completed, the patient removes the patch.

### 10.2 Checking the 'Ambulatory ECG Monitor', processing returns, and uploading data

- 1 To check the ECG data saved in the device, use the Dataloader program in the Windows PC.
- 2 Run MEMO Patch ECG Dataloader.



|                                                                                                                                                                                                                                                                                                                                                                                                                        |                                                                                                                                                                        |                                                                                      |
|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------|
|                                                                                                                                                                                                                                                                                                                                                                                                                        |   |                                                                                      |
| <ul style="list-style-type: none"> <li>• 3 After running the Ambulatory ECG Monitor ECG Dataloader, login with the same email and password provided by HUIINNO.</li> <li>• 4 Combine the device and the cradle and connect the cradle to the PC using the provided cable. If the connection is successful, the Serial No. of the Ambulatory ECG Monitor will be displayed. Then, click the download button.</li> </ul> |                                                                                                                                                                        |  |

|                                                                                                                                                                                                                                      |                                                                                      |
|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------|
|                                                                                                                                                                                                                                      |    |
| <ul style="list-style-type: none"> <li>5 When the download is successfully completed, a 'Done' message will appear.</li> </ul>                                                                                                       |   |
| <ul style="list-style-type: none"> <li>6 Click the 'Done' button, and raw ECG data will be displayed on the viewer.</li> <li>7 To export the data to a PDF file, click 'Save as PDF', and it will be saved in a PDF file.</li> </ul> |  |

## 11.0 PRODUCT OPERATING ENVIRONMENT

| MEMO Launcher |                            |                                  |
|---------------|----------------------------|----------------------------------|
|               | iOS                        | Android                          |
| OS            | - Model: iPhone 8 or above | - OS: Android OS 9(Pie) or above |

|         |                                                                                |                                                                  |
|---------|--------------------------------------------------------------------------------|------------------------------------------------------------------|
|         | - OS: iOS 13 or above                                                          |                                                                  |
| Screen  | - Resolution: 1344 x 750 or higher<br>- Optimized for portrait mode for iPhone | - Resolution: 720 x 1280 or higher - Optimized for portrait mode |
| Network | - Wireless Internet connection<br>- Bluetooth 5.0 or later                     | - Wireless Internet connection<br>- Bluetooth 5.0 or later       |

|                           |                       |
|---------------------------|-----------------------|
| MEMO Patch ECG Dataloader |                       |
| Operating environment     |                       |
| Above Window 10           |                       |
| Hardware requirement      |                       |
| Memory                    | above 512MB           |
| HDD                       | above 128GB           |
| Screen resolution         | above 1024 x 768      |
| External port             | above USB 2.0, 1 port |
| Network                   | Wi-Fi or Ethernet     |

## 12.0 TROUBLE SHOOTING GUIDE

### 12.1 Trouble shooting

|                     |          |
|---------------------|----------|
| 'MEMO Launcher' app |          |
| Issue               | Solution |

|                                                                                                                                                                                                                                     |                                                                                                           |
|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------|
| The e-mail or password you entered does not match. Please check your account information and try again.                                                                                                                             | To retrieve your account, email sales@huinno.com and we will help you.                                    |
| Your account is dormant. Your account has been inactive because you haven't used the service for over a year. You can use it again after applying for release from dormancy on your PC.                                             | To continue the service, contact sales@huinno.com to terminate the dormant account.                       |
| New version available. Please, update your app to the latest version to continue.                                                                                                                                                   | Update to the latest version of the app.                                                                  |
| You are automatically signed out. You were automatically signed out because there's no service use for 8hours.                                                                                                                      | The service will automatically log out when it is out for more than 8 hours. To continue service, log in. |
| Ambulatory ECG Monitor connection failed. Make sure the Ambulatory ECG Monitor is powered on and within the search range. If Ambulatory ECG Monitor connection continues to fail, please change Ambulatory ECG Monitor and proceed. | It notifies you if the connection fails because the test device is out of range or powered off.           |
| Don't see the patch on the list? Make sure the Ambulatory ECG Monitor is powered on and within the search range. Press and hold the button for 3 seconds to blink the green LED and turn on the power.                              | Press the power button of Ambulatory ECG Monitor for 3sec.                                                |
| Bluetooth is turned off. The connection failed because the Bluetooth function of the mobile device was turned off. Turn on Bluetooth and proceed again.                                                                             | Turn on the Bluetooth and recheck the connection.                                                         |
| Failed to start examination. There was a problem initializing Ambulatory ECG Monitor. Please try to                                                                                                                                 | Replace the patch or restart the program.                                                                 |

|                                                                                                                                                                                                                   |                         |                                                                                    |
|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------------|------------------------------------------------------------------------------------|
| initialize Ambulatory ECG Monitor again or change to another Ambulatory ECG Monitor and proceed.                                                                                                                  |                         |                                                                                    |
| There is no examination. Check if there is any patient who has started the examination.                                                                                                                           |                         | Verify the patient and start the examination again.                                |
| There's no search result. Make sure you search it correctly.                                                                                                                                                      |                         | Verify the patient's name or number.                                               |
| You signed into another mobile device. You signed in on another mobile device with the same ID and were automatically signed out of this device.                                                                  |                         | To continue service, log in.                                                       |
| The network is not available. Please check the network connection status of the mobile device and try again.                                                                                                      |                         | Check the network connection and access to the app again.                          |
| Ambulatory ECG Monitor is disconnected. Please reconnect Ambulatory ECG Monitor to continue. If Ambulatory ECG Monitor connection continues to fail, please change to another Ambulatory ECG Monitor and proceed. |                         | Turn on the Bluetooth and recheck the connection.                                  |
| 'MEMO Patch ECG Dataloader'                                                                                                                                                                                       |                         |                                                                                    |
| Login                                                                                                                                                                                                             | We couldn't log you in. | The email and password you entered did not match our records. Recheck the account. |
|                                                                                                                                                                                                                   | Couldn't connect        | There was a problem communicating with the server. Please try again later.         |
|                                                                                                                                                                                                                   | Import complete         | The data is entirely imported.                                                     |
|                                                                                                                                                                                                                   | System Error            | Restart the program to solve the unexpected error.                                 |

|               |                                                                            |                                                                            |
|---------------|----------------------------------------------------------------------------|----------------------------------------------------------------------------|
|               | Update required                                                            | The latest version is ready to install now.                                |
|               | Couldn't update                                                            | There was a problem communicating with the server. Please try again later. |
|               | Not connected Internet                                                     | Please check the Internet status.                                          |
| Data download | An error has occurred on the device. please try again.                     | Check communication error of patch.                                        |
|               | The device information is incorrect. Please check your device again.       | Check the saved information of the patch.                                  |
|               | Serial number is incorrect. Please check your device again.                | Check the serial number of the patch.                                      |
|               | CRC Error                                                                  | Check the CRC data of ECG Data.                                            |
|               | A system error has occurred on the device. please try again.               | Undefined error. inquiry email<br>sales@huinno.com                         |
| Upload        | There was a problem communicating with the server. please try again later. | Undefined error. inquiry email<br>sales@huinno.com                         |
|               | Please check the Internet status.                                          | Please check the Internet status.                                          |
|               | This device has not been returned.                                         | Check the status of the returned device.                                   |
|               | Response time exceeded. Please check the Internet status.                  | Check the internet connection and re-log in.                               |
|               | Unable to find file to upload.                                             | Please check the upload file.                                              |
|               | Your login time has expired. Please log in again...                        | To continue service, log in.                                               |

## 12.2 FAQs

|                               |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                |
|-------------------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <p>Ambulatory ECG Monitor</p> | <ol style="list-style-type: none"> <li>How do I turn off the power?<br/>This product cannot be powered off. If an emergency power off is required, use the included battery opener to open the battery cover and remove the built-in battery. When reusing the product, please insert the battery in the correct direction and turn on the product.</li> <li>How can you verify if the device is operating correctly?<br/>When you press and release the button for 1 second, the blue LED will flash once.</li> <li>How can you confirm if the connections are secure?<br/>If the connections are unstable, the yellow LED will blink twice at 6 second intervals.</li> </ol> |
| <p>MEMO Launcher App</p>      | <ol style="list-style-type: none"> <li>I cannot log in to the app, although I had no problem logging in before. What should I do?<br/>If you encounter a login error, try logging out and logging back in first. If it still doesn't work, check for an updated version on Google Play Store or Apple App Store and update it.<br/>(Android) If you still encounter a login error after the update, try clearing the data by going to Phone Settings &gt; Applications &gt; MEMO Launcher &gt; Storage &gt; Clear data.</li> </ol>                                                                                                                                             |
| <p>Data Upload</p>            | <ol style="list-style-type: none"> <li>Can I use a cable other than the provided cradle cable during data upload?<br/>You must use the provided cradle cable when performing the data upload process. Uploading may not be completed properly if a cable other than the provided one is used.</li> <li>I connected the memo patch to the cradle and then to the PC, but the patch is not showing a green light. What should I do?<br/>The cradle cable must be the provided cable in order to function properly.</li> </ol>                                                                                                                                                    |
| <p>Product Storage</p>        | <ol style="list-style-type: none"> <li>How should Ambulatory ECG Monitor be stored?<br/>Once the patient returns the device and the data upload is complete, it should be wiped down and stored according to the environmental conditions specified by the manufacturer.</li> </ol>                                                                                                                                                                                                                                                                                                                                                                                            |

|                   |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                 |
|-------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Patient Questions | <ol style="list-style-type: none"> <li>1. As a patient wearing an Ambulatory ECG Monitor, how can I check if the device is functioning properly during use?<br/> <p>The Patch has a symptom record button located on the power unit. Pressing the symptom record button for about 0.5 seconds and then releasing it will cause the green LED to flash. If the green LED flashes, it means that the device is functioning properly. If the green LED does not flash, please contact the HUINNO customer service at 02-3443-3160. If the device is not activated, the green LED will flash every 3 seconds, and the power will turn off after 5 minutes.</p> </li> <li>2. What should I do if I experience symptoms while wearing the Ambulatory ECG Monitor?<br/> <p>If you experience symptoms, you should press the symptom record button on the Ambulatory ECG Monitor. The symptom record button is located on the power unit of the Ambulatory ECG Monitor. Press the button for more about 1 second (Short Press button) and when the blue LED flashes, it indicates that the symptom has been recorded. (The symptom record button and the power button are the same button.)</p> </li> <li>3. What should I do if I forget to record my symptoms?<br/> <p>If you forget to record your symptoms, write down the approximate time when the symptoms occurred and the symptoms themselves in the patient note.</p> </li> <li>4. What should I do if I accidentally press the symptom record button multiple times or press it by mistake when no symptoms are present?<br/> <p>The symptom record button is used to compare whether the palpitation symptoms felt by the patient occurred simultaneously in the heart. Therefore, pressing the button several times or pressing it by mistake will not affect the test results.</p> </li> <li>5. What should I do if the Ambulatory ECG Monitor falls off my body?<br/> <p>Check the adhesive on the electrode before reattaching it to the same location as before. If the adhesive is sufficient, reattach the device to the same location. If the adhesive is insufficient, replace both electrodes and</p> </li> </ol> |
|-------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|

|  |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                      |
|--|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|  | <p>then reattach the device. If the skin is moist or sweaty, wipe it dry before reattaching the Ambulatory ECG Monitor. If the Ambulatory ECG Monitor falls off due to hair, shave the area before reattaching.</p> <p>6. Is it okay to use an electric blanket or electric corded bed while wearing the Ambulatory ECG Monitor?</p> <p>Do not use an electric blanket or electric bed with a power cord plugged in. Electrical devices can interfere with the signal and affect the accuracy of the test results.</p> <p>*However, it is okay to use the devices after unplugging the power cord.</p> <p>**Use of electric cars, microwave ovens, and induction cooktops do not affect the Ambulatory ECG Monitor and are okay to use.</p> <p>7. Can I exercise while wearing the Ambulatory ECG Monitor?</p> <p>Strenuous exercise can affect the electrocardiogram signal measurement, so it is recommended to avoid it. If sweating, the device can fall off, so be careful.</p> <p>8. Can I take a shower, bath, or swim while wearing the Ambulatory ECG Monitor?</p> <p>The Ambulatory ECG Monitor is waterproof, but it is recommended to remove the device before bathing, or swimming.</p> <p>9. Can I reattach the Ambulatory ECG Monitor to a slightly different location from the original attachment site?</p> <p>Yes, you can reattach the Ambulatory ECG Monitor to a slightly different location.</p> <p>10. What should I do if I experience skin irritation or itching?</p> <p>Some patients may experience mild skin irritation or itching. If you experience irritation or itching, or if it becomes severe, please contact the hospital where you received the prescription.</p> <p>11. What activities should be avoided during the test?</p> |
|--|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|

|  |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                 |
|--|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|  | <p>Please avoid vigorous exercise that can cause excessive sweating during the test. When sweating, the electrodes attached to the body may come off.</p> <p>12. Is it okay to travel while wearing the patch?<br/>It is generally okay to travel while undergoing the test, but it is recommended to avoid excessive physical activity that could potentially dislodge the patch.</p> <p>13. Is it okay to board a plane while wearing the patch? Are there any issues going through airport security?<br/>No, there is no issue. There might be some additional noise in the signal or alterations in the ECG waveform, which do not impact the analysis results. However, you might hear a warning sound while going through security.</p> <p>14. How do I return the patch after the test is finished?<br/>You can either send the patch by mail in a storage bag to the hospital where you received the prescription or visit the hospital in person to return it.</p> <p>15. If I have any questions during the test, where can I call?<br/>Call HUINNO customer service. Depend on the question, you may also need to contact the hospital directly.</p> |
|--|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|

To receive support for installation, use, or maintenance of the device, or to report unexpected behavior or events, please contact HUINNO's customer service.

Customer service operate from 09:00 to 18:00 on weekdays, with a one-hour lunch break from 12:00 to 13:00, closed on holidays and weekends (KST).

## 13.0 DEVICE SPECIFICATION

### 13.1 Performance Characteristics

|                 |           |
|-----------------|-----------|
| ECG Channels    | 1 channel |
| Memory capacity | 14 days   |

|                         |                  |
|-------------------------|------------------|
| Recording Format        | Continuous       |
| Service Life            | Up to 14 days    |
| Shelf Life              | 2 years          |
| Out-of-Pouch Shelf life | Use upon opening |

### 13.2 Electrical Characteristics

|                                                |                                                                                                                                  |
|------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------|
| Protection Against Electrical Shock            | Type BF Applied Part (Electrode)                                                                                                 |
|                                                | Internally powered ME equipment                                                                                                  |
| ECG Frequency Response                         | 0.5Hz to 50Hz                                                                                                                    |
| ECG Input Impedance                            | $\geq 10 \text{ M}\Omega$                                                                                                        |
| ECG Differential Range                         | $\pm 6.3 \text{ mV}$                                                                                                             |
| ECG A/D Sampling Rate                          | 250 Hz                                                                                                                           |
| ECG Resolution                                 | 12bit                                                                                                                            |
| Frequency Band                                 | <ul style="list-style-type: none"> <li>- 2.4 GHz radio transceiver</li> <li>- Number of Channels: 40 (2400-2483.5MHz)</li> </ul> |
| Bandwidth width of the receiving section       | 1M                                                                                                                               |
| Each transmission frequency or frequency band. | 2.4 GHz radio transceiver                                                                                                        |
| Modulation Type and Frequency Characteristics  | GFSK(Gaussian Frequency Shift Keying) modulation                                                                                 |
| Effective Radiated Power (ERP)                 | 2.5mW (4dB) or less.                                                                                                             |

### 13.3 Power Characteristics

|                    |                |
|--------------------|----------------|
| Patch Battery Type | CR2032, 225mAh |
| Battery Life       | 14 days        |

**13.4 Physical Characteristics**

|                                            |                                        |
|--------------------------------------------|----------------------------------------|
| Patch Dimensions                           | 29.20mm(W) X 48.40mm(D) X 8.30mm(H)    |
| Patch Weight                               | 12g (9g without battery)               |
| Data transferred cradle (1port) Dimensions | 44.75mm(W) x 41.47(D) x inch13.90mm(H) |
| Data transferred cradle (1port) Weight     | 25g                                    |

**13.5 Environmental Characteristics**

|                                                                                       |                             |
|---------------------------------------------------------------------------------------|-----------------------------|
| Operational Temperature                                                               | 10°C - 45°C                 |
| Operational Altitude                                                                  | 0 to 9,882 ft               |
| Operational & Storage Humidity                                                        | 10% to 95% (non-condensing) |
| Operational Atmospheric Pressure                                                      | 700hPa – 1060hPa            |
| Transport or Storage Pressure                                                         | 700hPa – 1060hPa            |
| Shipping (Short-term Storage) Temperature                                             | -40°C - 70°C                |
| Storage Altitude                                                                      | 0 to 9,882 ft               |
| Patch IP Classification                                                               | IP27                        |
| IP27) Protection against finger insertion and effects of temporary immersion in water |                             |

**13.6 Product Maintenance, Storage and Disposal Summary**

|                          |                                                                                                                                                                                                                                                                                                                                                                                                                |
|--------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Cleaning and Maintenance | Gently wipe with an alcohol swab<br>(with an alcohol content of at least 60%).                                                                                                                                                                                                                                                                                                                                 |
| Storage and Transport    | <p>When storing the device, avoid exposure to direct sunlight or heat.</p> <p>The device should be stored under the storage temperature range. If the device will be stored for a long time, the recommended conditions are:</p> <p>Temperature: -40°C to 70°C</p> <p>Relative humidity: 10% to 95% (non-condensing)</p> <p>The device should be stored in the room without acid, alkali, and harmful gas.</p> |

|          |                                                                                                                                             |
|----------|---------------------------------------------------------------------------------------------------------------------------------------------|
|          | Avoid exposure to violent vibration, rain, sunshine, and high humidity during transportation.                                               |
| Disposal | Follow local, state, and national governing ordinances and recycling instructions regarding disposal or recycling of the device components. |

### 13.7 Cleaning and Maintenance

- After use of the device gently wipe with an alcohol swab with an alcohol content of at least 60% and then clean your device with a soft and dry cloth.
- Do not use caustic or abrasive cleaning agents, or any cleaning agent containing ammonium chloride or isopropyl alcohol.
- Do not sterilize, autoclave, or immerse this device in liquid.
- Do not pour or spray any liquids onto the device.
- Do not repair, disassemble and modify this device.
- This device does not require calibration during its expected life cycle.

### 13.8 Storage and Transport

- When storing the device, avoid exposure to direct sunlight or heat.
- The device should be stored under the storage temperature range. If the device will be stored for a long time, the recommended conditions are:
- Temperature: -40°C to 70°C (MEMO Patch 2)
- Relative humidity: 10% to 95% (non-condensing) (MEMO Patch 2)
- The device should be stored in the room without acid, alkali and harmful gas.
- Avoid exposure to violent vibration, rain, sunshine and high humidity during transportation.

### 13.9 Disposal (Accessory)

- Follow local, state, and national governing ordinances and recycling instructions regarding disposal or recycling of the device components.

## 14.0 DEVICE LED SCENARIO

|                                            |                                                                                                                                                                            |
|--------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Power ON                                   | <ul style="list-style-type: none"> <li>• Green LED, flashing Once</li> </ul>                                                                                               |
| Discover mode after Power ON               | <ul style="list-style-type: none"> <li>• Before Operation: Green LED flashing once continuously (Every 5 seconds for 5 minutes)</li> <li>• After Operation: N/A</li> </ul> |
| Button before examination, lead off status | <ul style="list-style-type: none"> <li>• Green LED, flashing Once</li> </ul>                                                                                               |

|                                          |                                                                                                                                                                                                          |
|------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Start examination                        | <ul style="list-style-type: none"> <li>Green LED, Fast flashing Four times</li> </ul>                                                                                                                    |
| Connection of BLE                        | <ul style="list-style-type: none"> <li>Blue LED, Fast flashing Three times</li> </ul>                                                                                                                    |
| Disconnection of BLE                     | <ul style="list-style-type: none"> <li>Blue LED, Fast flashing Twice</li> </ul>                                                                                                                          |
| Event Marking (After Examination)        | <ul style="list-style-type: none"> <li>Green LED, flashing Once</li> </ul>                                                                                                                               |
| Lead OFF                                 | <ul style="list-style-type: none"> <li>Yellow LED, Fast flashing Twice every 6 seconds, Yellow LED, Fast flashing Twice</li> </ul>                                                                       |
| Low Battery                              | <ul style="list-style-type: none"> <li>Yellow LED, flashing Continuously every 6 seconds</li> </ul>                                                                                                      |
| When Device reset                        | <ul style="list-style-type: none"> <li>Reset due to malfunctioning: Red LED On -&gt; Reset-&gt; Power on and Green LED flashing once</li> <li>If Continuous reset occur, the Red LED flashes.</li> </ul> |
| Power OFF                                | <ul style="list-style-type: none"> <li>Yellow LED fast flashing Three times</li> </ul>                                                                                                                   |
| Attach cradle                            | <ul style="list-style-type: none"> <li>Green LED On</li> </ul>                                                                                                                                           |
| Launch MEMO Patch ECG Dataloader program | <ul style="list-style-type: none"> <li>Green LED On</li> </ul>                                                                                                                                           |
| Download                                 | <ul style="list-style-type: none"> <li>Green LED flashing</li> </ul>                                                                                                                                     |
| Download Completed                       | <ul style="list-style-type: none"> <li>LED off</li> </ul>                                                                                                                                                |

## 15.0 HEART RATE CALCULATION

|                             |     |                                                                                                                                         |
|-----------------------------|-----|-----------------------------------------------------------------------------------------------------------------------------------------|
| Episode Heart Rates         | Max | The maximum episode heart rate (ex. Maximum of all instantaneous heart rates within the episode)                                        |
|                             | Min | The minimum episode heart rate (ex. Minimum of all instantaneous heart rates within the episode)                                        |
|                             | Ave | The average episode heart rate (ex. Average of all instantaneous heart rates within the episode)                                        |
| Overall, Rhythm Heart Rates | Max | The maximum overall heart rate (ex. Maximum of all rhythm episode maximum heart rates within the record)                                |
|                             | Min | The minimum overall heart rate (ex. Minimum of all rhythm episode minimum heart rates exclusive of Pause heart rates within the record) |
|                             | Ave | The average overall heart rate (ex. Duration-weighted average of all rhythm episode heart rates within the record)                      |

After detecting the R peak in the ECG waveform, the heart rate is calculated according to the following formula:  $\text{current\_HR} = (60 \times \text{sample\_rate}) / (\text{current\_R\_peak\_index} - \text{prev\_R\_peak\_index})$

## 16.0 PAUSE DETERMINATION

Pause is defined as an RR interval greater than 3 seconds.

## 17.0 ELECTRICAL SAFETY AND COMPATIBILITY

Ambulatory ECG Monitor (MPT-E14R-UNA03) & ECG Data transferred cradle

WARNING: Use of this equipment adjacent to or stacked with other equipment should be avoided because it could result in improper operation. If such use is necessary, this equipment and the other equipment should be observed to verify that they are operating normally.

WARNING: Use of accessories, transducers and cables other than those specified or provided by the manufacturer of this equipment could result in increased electromagnetic emissions or decreased electromagnetic immunity of this equipment and result in improper operation.

WARNING: Portable RF communications equipment (including peripherals such as antenna cables and external antennas) should be used no closer than 30 cm (12 inches) to any part of the MPT-E14R-UNA03, including cables specified by the HUINNO.

### 17.1 Table1 – ELECTROMAGNETIC EMISSIONS for MEMO Patch 2 (MPT-E14R-UNA03)

| Guidance and manufacturer's declaration – electromagnetic emissions                                                                                                                                                  |            |                                                                                                                                                                                                                                 |
|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Ambulatory ECG Monitor (MPT-E14R-UNA03) is intended for use in the electromagnetic environment specified below. The customer or the user of the MPT-E14R-UNA03 should assure that it is used in such an environment. |            |                                                                                                                                                                                                                                 |
| Emissions test                                                                                                                                                                                                       | Compliance | Electromagnetic environment – guidance                                                                                                                                                                                          |
| RF emissions<br>CISPR 11                                                                                                                                                                                             | Group 1    | The MPT-E14R-UNA03 uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment.                                         |
| RF emissions<br>CISPR 11                                                                                                                                                                                             | Class B    | The MPT-E14R-UNA03 is suitable for use in all establishments, including domestic establishments and those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes. |
| Harmonic emissions<br>IEC 61000-3-2                                                                                                                                                                                  | Class A    |                                                                                                                                                                                                                                 |

|                                                             |          |  |
|-------------------------------------------------------------|----------|--|
| Voltage fluctuations/<br>flicker emissions<br>IEC 61000-3-3 | Complies |  |
|-------------------------------------------------------------|----------|--|

## 17.2 Table2 – Electromagnetic IMMUNITY for MEMO Patch 2 (MPT-E14R-UNA03)

| Guidance and manufacturer's declaration – electromagnetic immunity                                                                                                                                                       |                                                                                                                               |                                                                                                                               |                                                                                                                                                                                                                                                                                              |
|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| The Ambulatory ECG Monitor (MPT-E14R-UNA03) is intended for use in the electromagnetic environment specified below. The customer or the user of the MPT-E14R-UNA03 should assure that it is used in such an environment. |                                                                                                                               |                                                                                                                               |                                                                                                                                                                                                                                                                                              |
| IMMUNITY test                                                                                                                                                                                                            | IEC 60601 test level                                                                                                          | Compliance level                                                                                                              | Electromagnetic environment – guidance                                                                                                                                                                                                                                                       |
| Electrostatic discharge (ESD)<br><br>EN 61000-4-2                                                                                                                                                                        | ± 6 kV contact<br>± 8 kV air                                                                                                  | ± 6 kV contact<br>± 8 kV air                                                                                                  | Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30 %.                                                                                                                                               |
| Electrical fast transient/burst<br><br>EN 61000-4-4                                                                                                                                                                      | ± 2 kV for power supply lines<br>± 1 kV for input/output lines                                                                | ± 2 kV for power supply lines<br>± 1 kV for input/output lines                                                                | Mains power quality should be that of a typical commercial or hospital environment.                                                                                                                                                                                                          |
| Surge<br><br>EN 61000-4-5                                                                                                                                                                                                | ± 1 kV line(s) to line(s)<br>± 2 kV line(s) to earth                                                                          | ± 1 kV line(s) to line(s)<br>± 2 kV line(s) to earth                                                                          | Mains power quality should be that of a typical commercial or hospital environment.                                                                                                                                                                                                          |
| Voltage dips, short interruptions and voltage variations on power supply input lines<br><br>EN 61000-4-11                                                                                                                | <5 % UT<br>(>95 % dip in UT) for 0,5 cycle<br><br>40 % UT<br>(60 % dip in UT) for 5 cycles<br><br>70 % UT<br>(30 % dip in UT) | <5 % UT<br>(>95 % dip in UT) for 0,5 cycle<br><br>40 % UT<br>(60 % dip in UT) for 5 cycles<br><br>70 % UT<br>(30 % dip in UT) | Mains power quality should be that of a typical commercial or hospital environment. If the user of the MPT-E14R-UNA03 requires continued operation during power mains interruptions, it is recommended that the MPT-E14R-UNA03 be powered from an uninterruptible power supply or a battery. |

|                                                                            |                                                              |                                                              |                                                                                                                                                    |
|----------------------------------------------------------------------------|--------------------------------------------------------------|--------------------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------|
|                                                                            | for 25 cycles<br><br><5 % UT<br>(>95 % dip in UT)<br>for 5 s | for 25 cycles<br><br><5 % UT<br>(>95 % dip in UT)<br>for 5 s |                                                                                                                                                    |
| Power frequency<br>(50/60 Hz)<br>magnetic field<br><br>EN 61000-4-8        | 3 A/m                                                        | 3 A/m                                                        | Power frequency magnetic fields<br>should be at levels characteristic of a<br>typical location in a typical commercial<br>or hospital environment. |
| NOTE) UT is the a.c. mains voltage prior to application of the test level. |                                                              |                                                              |                                                                                                                                                    |

**17.3 Table3 – Electromagnetic IMMUNITY for MEMO Patch 2 (MPT-E14R-UNA03) that are not LIFE-SUPPORTING**

| Guidance and manufacturer's declaration – electromagnetic immunity                                                                                                                                                       |                         |                     |                                        |
|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------------|---------------------|----------------------------------------|
| The Ambulatory ECG Monitor (MPT-E14R-UNA03) is intended for use in the electromagnetic environment specified below. The customer or the user of the MPT-E14R-UNA03 should assure that it is used in such an environment. |                         |                     |                                        |
| IMMUNITY test                                                                                                                                                                                                            | IEC 60601 TEST<br>LEVEL | Compliance<br>level | Electromagnetic environment – guidance |

|                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                     |                             |        |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                       |
|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------------------------|--------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Conducted RF<br>EN 61000-4-6                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                        | 3 Vrms<br>150 kHz to 80 MHz | 3 Vrms | <p>Portable and mobile RF communications equipment should be used no closer to any part of the MPT-E14R-UNA03, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter.</p> <p>Recommended separation distance</p> $d = 1.17 \sqrt{P}$ <p>d = 1.17 80 MHz to 800 MHz</p> $d = 2.33 \sqrt{P}$ <p>80 MHz to 800 MHz</p> <p>where P is the maximum output power rating of the transmitter in watts(W) according to the transmitter manufacturer and d is the recommended separation distance in meters(m).</p> <p>Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey, a should be less than the compliance level in each frequency range. b</p> <p>Interference may occur in the vicinity of equipment marked with the following symbol:</p>  |
| Radiated RF<br>EN 61000-4-3                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                         | 3 V/m<br>80 MHz to 2,5 GHz  | 3 V/m  |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                       |
| <p>NOTE 1) At 80 MHz and 800 MHz, the higher frequency range applies.</p> <p>NOTE 2) These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.</p>                                                                                                                                                                                                                                                                                                                                                                |                             |        |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                       |
| <p>a) Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the MPT-E14R-UNA03 is used exceeds the applicable RF compliance level above, the MPT-E14R-UNA03 should be observed to verify normal operation. If abnormal performance is</p> |                             |        |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                       |

observed, additional measures may be necessary, such as re-orienting or relocating the MPT-E14R-UNA03.

b) Over the frequency range 150 kHz to 80 MHz, field strengths should be less than 3 V/m.

**17.4 Table4 – Recommended separation distances between portable and mobile RF communication equipment and the ME EQUIPMENT or ME SYSTEM – for ME EQUIPMENT and ME SYSTEMS that are not LIFE-SUPPORTING**

Recommended separation distances between portable and mobile RF communications equipment and the Ambulatory ECG Monitor (MPT-E14R-UNA03)

The Ambulatory ECG Monitor (MPT-E14R-UNA03) is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of the MPT-E14R-UNA03 can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the MPT-E14R-UNA03 as recommended below, according to the maximum output power of the communications equipment.

| Rated maximum output power of transmitter<br>W | Separation distance according to frequency of transmitter<br>m |                                             |                                              |
|------------------------------------------------|----------------------------------------------------------------|---------------------------------------------|----------------------------------------------|
|                                                | 150 kHz to 80 MHz<br>$\sqrt{P}$<br>d = 1.17                    | 80 MHz to 800 MHz<br>$\sqrt{P}$<br>d = 1.17 | 800 MHz to 2,5 GHz<br>$\sqrt{P}$<br>d = 2.33 |
| 0.01                                           | 0.117                                                          | 0.117                                       | 0.233                                        |
| 0.1                                            | 0.370                                                          | 0.370                                       | 0.736                                        |
| 1                                              | 1.17                                                           | 1.17                                        | 2.33                                         |
| 10                                             | 3.70                                                           | 3.70                                        | 7.36                                         |
| 100                                            | 11.7                                                           | 11.7                                        | 23.3                                         |

For transmitters rated at a maximum output power not listed above, the recommended separation distance d in meters (m) can be estimated using the equation applicable to the frequency of the transmitter, where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer.

NOTE 1) At 80 MHz and 800 MHz, the separation distance for the higher frequency range applies.

NOTE 2) These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

**17.5 Table5 – Electromagnetic Compatibility Information for Ambulatory ECG Monitor (MPT-E14R-UNA03) & ECG Data transferred cradle**

| Phenomenon | Basic EMC standard or test method | Test level/Requirement |
|------------|-----------------------------------|------------------------|
|------------|-----------------------------------|------------------------|

|                                                                                          |                                 |                                                                                                   |
|------------------------------------------------------------------------------------------|---------------------------------|---------------------------------------------------------------------------------------------------|
| Mains terminal disturbance voltage                                                       | CISPR 11<br>EN 55011            | Group1, Class B                                                                                   |
| Radiated disturbance                                                                     | CISPR 11<br>EN 55011            | Group1, Class B                                                                                   |
| Harmonic Current Emission                                                                | IEC 61000-3-2<br>EN 61000-3-2   | Class A                                                                                           |
| Voltage change,<br>Voltage fluctuations and<br>Flicker Emission                          | IEC 61000-3-3<br>EN 61000-3-3   | Pst: 1<br>Plt: 0.65<br>Tmax:0.5<br>dmax: 4%<br>dc: 3.3%                                           |
| Electrostatic Discharge Immunity                                                         | IEC 61000-4-2<br>EN 61000-4-2   | $\pm 8$ kV/Contact<br><br>$\pm 2, \pm 4, \pm 8, \pm 15$ kV/Air                                    |
| Radiated RF<br>Electromagnetic Field<br>Immunity                                         | IEC 61000-4-3<br>EN 61000-4-3   | 10 V/m<br>80 MHz - 2.7 GHz<br>80% AM at 1 kHz                                                     |
| Immunity to Proximity Fields<br>from RF wireless<br>Communications Equipment             | IEC 61000-4-3<br>EN 61000-4-3   | Table 9 in<br>IEC 60601-1-2: 2014<br>+A1:2020                                                     |
| Immunity to proximity<br>magnetic fields in the<br>frequency range 9 kHz to<br>13.56 MHz | IEC 61000-4-39<br>EN 61000-4-39 | 30 kHz: 8 A/m, CW<br><br>134.2 kHz: 65 A/m,<br>PM 2.1 kHz<br><br>13.56 MHz: 7.5 A/m,<br>PM 50 kHz |
| Electrical Fast<br>Transient/Burst Immunity                                              | IEC 61000-4-4<br>EN 61000-4-4   | $\pm 2$ kV, 100 kHz<br>repetition frequency                                                       |
| Surge Immunity                                                                           | IEC 61000-4-5<br>EN 61000-4-5   | Line to Line<br>$\pm 0.5$ kV, $\pm 1$ kV                                                          |
| Immunity to Conducted<br>Disturbances Induced by<br>RF fields                            | IEC 61000-4-6<br>EN 61000-4-6   | 3 V<br>0.15 MHz - 80 MHz                                                                          |

|                                         |                                 |                                                                                      |
|-----------------------------------------|---------------------------------|--------------------------------------------------------------------------------------|
|                                         |                                 | 6 V in ISM and Amateur radio band between 0.15 MHz and 80 MHz<br><br>80% AM at 1 kHz |
| Power Frequency Magnetic Field Immunity | IEC 61000-4-8<br>EN 61000-4-8   | 30 A/m<br>50 Hz and 60 Hz                                                            |
| Voltage dips                            | IEC 61000-4-11<br>EN 61000-4-11 | 0 % $U_T$ ; 0.5 cycle<br>At 0°, 45°, 90°, 135°, 180°, 225°, 270° and 315°            |
|                                         |                                 | 0 % $U_T$ ; 1 cycle and<br>70 % $U_T$ ; 25/30 cycles<br>Single phase: at 0°          |
| Voltage interruptions                   | IEC 61000-4-11<br>EN 61000-4-11 | 0 % $U_T$ ; 250/300 cycle                                                            |

## 18.0 BUG LIST

Defect free software

## 19.0 CUSTOMER SERVICE

If you encounter any issues during the inspection, please contact our customer service.

Our hours of operation are weekdays from 9am to 6pm, Closed for lunch from 12pm to 1pm and closed on weekends and holidays (KST).

TEL: +82.2.3443.3160

Address: 3F, 4F, 5F, 19, Apgujeong-ro 79-gil, Gangnam-gu, Seoul, Republic of Korea (06011)

E-Mail: [sales@huinno.com](mailto:sales@huinno.com)

- NOTE: Manufacturers have an obligation to disclose data on essential product performance to users whenever they request it.

**20.0 REVISION HISTORY**

| Rev. | Rev. Date        | Page(s) Affected                     | Revision Description                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                            |
|------|------------------|--------------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| 0    | January 5, 2023  | All                                  | Newly Established                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                               |
| 1    | October 30, 2023 | All (2, 3, 6, 61, 63, 64, 65, 68 76) | <p>Layout of the Contents</p> <p>Edited (Delete of 'SPI communication')</p> <p>ECG Resolution (16bit -&gt; 12bit)</p> <p>Edited (MEMO Patch Customer Center -&gt; HUIINNO Customer Service)</p> <p>Manufacture Contact Information (401, 301, 302, 19, Apgujeong-ro 79-gil, Gangnam-gu, Seoul, Republic of Korea -&gt; 3F, 4F, 5F, 19, Apgujeong-ro 79-gil, Gangnam-gu, Seoul, Republic of Korea) (cs@huinno.com -&gt; sales@huinno.com)</p> <p>Brand Name (MEMO Patch2 -&gt; MEMO Patch 2)</p> <ul style="list-style-type: none"> <li>Storage and Transport (TBD -&gt; When storing the device, avoid exposure to direct sunlight or heat.</li> <li>The device should be stored under the storage temperature range. If the device will be stored for a long time, the recommended conditions are:</li> <li>Temperature: -40°C to 70°C</li> <li>Relative humidity: 10% to 95% (non-condensing)</li> <li>The device should be stored in the room without acid, alkali, and harmful gas.</li> </ul> <p>Avoid exposure to violent vibration, rain, sunshine, and high humidity during transportation.)</p> <p>Edited (MemoPatch -&gt; Ambulatory ECG Monitor)</p> |

**FCC Compliance Statement**

This device complies with part 15 of the FCC rules. Operation is subject to the following two conditions:

(1) This device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

**FCC Interference Statement**

This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications.

However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to correct the interference by one of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit different from which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.

**FCC Caution**

Any changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate this equipment. This transmitter must not be co-located or operating in conjunction with any other antenna or transmitter.