

SBO Hearing A/S

Class II Permissive Change (C2PC)

2021-04-15

Federal Communication Commission

Equipment Authorization Division, Application Processing Branch
7435 Oakland Mills Road
Columbia, MD 21048

Certification and Engineering Bureau

Industry Canada
Spectrum Engineering Branch
3701 Carling Avenue, Building 94
Ottawa, Ontario K2H 8S2

TO WHOM IT MAY CONCERN

Pursuant to the applicable paragraphs in ISED's RSP-100 §7 and CFR 47 FCC part §2.1043(b)(2) we hereby declare the following changes under Class II Permissive Change to the product:

Hearing Instrument amplifier module Aurora miniRITE

FCC ID: 2ACAHSBMRTRC

IC: 11936A-SBMRTRC

Original grant: February 20th -2019

We, SBO Hearing A/S hereby declare the following Class 2 Permissive Change to the amplifier module Aurora miniRITE related to the FCC part 15.247 / RSS-247 DTS radio:

- The change of Bluetooth BAW filter and ESD protection diode due to "End-of-Life" components changes has been performed with a minor PCB layout change handled in correlation with KDB 178919 D01, section III, subclause D + Page 3, Note 5, bullet 2.
- Only passive components have been changed.
- The Limited Modular Approval (LMA) still apply after the C2PC product update.
- No mechanical changes has been applied to the product for this C2PC.
- The radio module performance is equal to the previously certified module performance.

We, SBO Hearing A/S declare that the hearing aid amplifier module changes as documented in the test report will have a similar module performance as established in the FCC/ISED original certification.

SBO Hearing A/S

If you have any questions, please feel free to contact me at the address shown below.

Sincerely,

A handwritten signature in blue ink, appearing to read 'Per K. Nielsen', written in a cursive style.

Per Klaus Nielsen, Certification Manager
SBO Hearing A/S
Kongebakken 9
DK-2765 Smørum, Denmark

Phone : +45 3917 7100

Email : pkni@sbohearing.com