

Declaration of Variant Category 1

Hereby,

Name of manufacturer: **Dyzle Services B.V.**
Address: **Transistorstraat 2D**
City: **1322 CE Almere**
Country: **The Netherlands**

declares that the equipment:

Product description: **Battery operated wireless thermocouple unit**
Type designation(s): **Y-TAG Thermocouple 2.0 / 020 1301 002 set**
Trademark: **Dyzle**
Hardware version: **2.0**
Software version: **0.26**

Product description: **Battery operated wireless Door monitor unit**
Type designation(s): **Y-TAG Door monitor 2.0 / 030 1001 002 set**
Trademark: **Dyzle**
Hardware version: **2.0**
Software version: **0.26**

are technical variants of:

Product description: **Battery operated wireless temperature unit**
Type designation(s): **Y-TAG temperature 2.0 / 022 1201 001**
Trademark: **Dyzle**
Hardware version: **2.0**
Software version: **0.27**

Full description of all changes:

Both units require a different software and regarding the Y-TAG Thermocouple there is also an adaptation to resistor R18. The resistance values in the Y-TAG temperature 2.0 is 4.7 kΩ. This should be changed to 1MΩ for Y-TAG Thermocouple 2.0.

Date: **November 28, 14**
City: **Almere**
Name: **Mr. René Tjong Tjin Tai**
Signature: 