



FCC ID Labeling

Federal Communications Commission
Authorization and Evaluation Division
1435 Oakland Mills Road
Columbia, MD 21046

Date: March 30, 2018

SUBJECT: Justification for Not Labeling FCC ID on Device

To Whom It May Concern:

The Kairos hearing aid is small and does not have adequate print area to allow printing of the FCC ID number. Current laser marking capability in manufacturing requires a flat surface to provide legible print, and the available area on the Kairos hearing aid case is only 0.0478 in².

Allowable Print Area —



As a result, we are requesting an allowable exclusion per FCC 15.19(a) regarding this labeling requirement since the FCC ID will be prominently marked on labeling of unit packaging, and in the User Guide provided with each device.

Sincerely,

A handwritten signature in blue ink, appearing to read 'J. Larson'.

James D. Larson
Quality and Regulatory Affairs Manager
Clarity, LLC