

eXact[®] Micro 20

Dual Wavelength Advanced Photometer System ***Instruction Manual***

**IDEAL FOR DRINKING WATER, POOLS & SPAS,
ENVIRONMENTAL, AND EDUCATIONAL TESTING**

U.S. Patent No. 7,333,194, U.S. Patent No. 7,491,546, South African Patent No. 2007/0628,
EU Patent #1,725,864, and International Patent Appl. No. PCT/US2005/033985




Contents

eXact® Micro 20 Specifications	3
eXact® Micro 20 with Bluetooth SMART Specifications	3
About your eXact® Micro 20 photometer	4
About Bluetooth® SMART	4
About eXact® Strip Micro	5
Compliance verification for Free & Total Chlorine Testing	5
About the eXact iDip® app for Micro 20 with Bluetooth® SMART	6
Managing data with the eXact iDip® app	7
eXact® Micro 20 Test Specifications	8 – 9
Select group overview	10
eXact Micro 20 meter messages	10
About the Sample Cell (built-in 4 mL)	11
How to install or replace "AAA" batteries	11
2-Year Limited Warranty	11
Using eXact®Strip Micro (reagent strips) Test Procedure	12 – 13
Alkali pH, Acid pH, Biguanide, Bromine, Calcium Hardness, Chloride, Chromium, Copper, High Range Chlorine, Hydrogen Peroxide, Nitrate, Nitrite, Ozone, Peracetic Acid, Permanganate, pH, Phosphate, Quaternary Ammonia, Sulfate, Total Alkalinity, and Total Hardness HR & LR	
Using eXact® Reagent (liquid drops) Test Procedure	14 – 15
Cyanuric Acid and Fluoride	
Free or Total Chlorine	16
Combined Chlorine	17
Free and Combined Chlorine (direct read)	18
Aluminum, Ammonia & Sulfide	19
Cyanide	20
Total Iron	21
High Range Chloride	22
Manganese	23
Chlorine Dioxide	24
Nitrate (salt water > 4000 ppm)	25
Accessories / Kits	26
About the Accuracy/Calibration of the eXact® Micro 20 System	27
eXact® Strip Micro (DPD–1) Accuracy	27
Tips for Best Accuracy	28
Available Reagents / Reorder Information	31
Contact Information	32



eXact® Micro 20 & eXact® Micro 20 with Bluetooth® SMART



*for eXact® Micro 20 with Bluetooth SMART only

Technical Specifications		eXact® Micro 20 (486700) eXact® Micro 20 with Bluetooth® SMART (486700BT)	
Measurement method	Photometric	Cell chamber	Custom-molded, proprietary, PET plastic fused into chamber, non-removable
Light source	Light Emitting Diode (LED) with precision filter	Sample required	4mL (0.13oz)
Wavelength	Dual—525nm & 638nm	Operating temp range	0°—50°C (32°—122°F)
Transmission range	100 — 0.00%T	Power supply	(4) AAA alkaline batteries
Photometric precision	+/- 0.1/0.01 %T	Battery life	>2000 tests
Automatic range selection	See specifications below	Electromagnetic compliance (EMC)	Emitted interference – EN61326 Immunity to interference – EN61326
Display	3-digit customized liquid crystal display with annunciators	Waterproof rating	Exceeds IP67
Cell path-length	20 nm	Weight	181g (6.4oz) with batteries
Reagent system	Utilizes patented eXact® strip micro reagent delivery system with our EZ-3™	Dimensions	5 x 3.5 x 16.5 cm (2 x 1.4 x 6.5) in W x D x H
Wireless	Bluetooth® SMART compatible with Bluetooth 4.0 (Android 4.3 and Apple iOS 6.1 to iOS7)		

About your eXact Micro 20 photometer

In order to save power, the meter is designed to turn off after 3 minutes (timed from the last button pressed). Should the meter turn off in the middle of a test, the last stored zero in the meter will remain valid when the meter is turned on again. Also, the test result is stored in memory for easy retrieval.

The eXact® Micro 20 meter is controlled by four buttons:



When first pressed, the **ZERO/ON** button powers the meter. When the meter is on and this button is pressed, it zeros the sample in the cell. It is recommended that each new water sample analyzed is zeroed before testing, to maximize sensitivity and accuracy.



With each press, the **SELECT** button advances through the Select Group 1 through 7. The current Select Group will appear as a small digit to the right of the selected **MENU** (example: [L 1]).



With each press, the **MENU** button advances through the tests available in the current Select Group. Each test menu can store up to 20 results. To retrieve the stored results, go to the desired test using the **MENU** key. When the desired test is displayed, press and hold down the **MENU** key. Continue holding down the **MENU** key to scroll the stored results for that test, starting with the most recent result. The meter will display, from memory, the last 20 readings in sequence beginning with -20, which is the latest result, followed by 19, which is the 2nd latest result, etc; and finally -01, which is the oldest result retained. Only the last 20 readings are stored in each menu.



When **READ** is pressed once, this button starts the timer for the parameter being tested. When pressed a second time the meter exits the timer and immediately prepares to measure the sample by colorimetric analysis. The meter will simultaneously store the measurement in memory.

If the parameter being measured is below or above the detection range, the display will show “**LO**” (Under Range) or “**HI**” (Over Range), respectively. This feature is menu specific and does not apply to all parameters.

About Bluetooth® SMART

Bluetooth® SMART is a low-power wireless networking standard which uses short radio waves to allow electronic devices to communicate with each other. The eXact® Micro 20 with Bluetooth® SMART comes standard with the latest Bluetooth® 4.0 technology (www.bluetooth.com/Pages/Bluetooth-Smart.aspx), a class 2 device with a wireless working distance of up to 30 feet (10 meters) and a 2.1 Mbps data transfer rate. This allows a seamless transfer of data between a smart device and the eXact® Micro 20 with Bluetooth® SMART.



About *eXact® Strip Micro*

The eXact® Micro 20 and Micro 20 with Bluetooth® SMART are designed to work with our line of eXact® Micro Strips. This type of reagent delivery method is designed to give the most precise accuracy reading for testing various water quality parameters.

eXact® Strip Micro has been designed to offer the user a more “Green” and cost-effective alternative to testing. Instead of using a 10 mL water sample, eXact® Strip Micro uses a 4mL water sample, which uses up to 60% less chemical per test. The accuracy of the meter is maintained by designing the photo cell with a 20 mm path-length.



For a complete list of eXact® Strips Micro we offer, please see **page 31**.

Compliance verification for Free & Total Chlorine testing

This DPD test system is accepted by most health departments because this test is USEPA (DIN Standard 38 408 G4, ISO 7393/2) accepted for testing requirements for Free and Total Chlorine. The eXact® Micro 20 meter uses a wavelength of 525 nm; and the compliance requirement is that the colorimeter wavelength is between 490 nm to 530 nm. The eXact® Strip Micro CL (DPD-1) uses the same reagents and proportions, and the resulting solution pH is maintained between 6.2 and 6.5 as specified by AWWA (American Water Works Association) method 4500-Cl G. It should be understood that the USEPA does not “approve” commercial DPD delivery systems such as reagent powder pillows, tablets, dispensers, or eXact® Strip DPD delivery devices. The eXact® Strip Micro CL (DPD-1) for Free Chlorine, and the eXact® Strip Micro CL (DPD-3) or the eXact® Strip Micro CL (DPD-4) for Total Chlorine meet your reportable testing requirements because the eXact® Strip Micro CL delivers the same chemicals in identical proportions (see table below); therefore, the system is compliant. Likewise, AWWA proportions are followed as required for Total Chlorine measurements using Potassium Iodide.

Component (Free Chlorine)	AWWA 4500-Cl G	eXact® DPD-1
Anhydrous DPD sulfate	1.5%	1.5%
Anhydrous Na ₂ HPO ₄	33.4%	33.4%
Anhydrous KH ₂ PO ₄ Na ₂	64.0%	64.0%
EDTA	1.1%	1.1%



MENU:
*Opens
slide-out for
easy access to
all app features*

HOME SCREEN

HISTORY:
*Accesses
saved results
which can be
sorted, edited,
and emailed.
Also accesses
History Map*

CALENDAR:
*Displays your
schedule/
appointments*

CUSTOMERS:
*Attaches
results to
people and/or
locations from
your smart
device contacts*

STORE:
*Opens store
to unlock
additional tests*



TEST:
*Initiates water
testing*

RESULTS:
*Accesses temporary
results that have
not been saved to
history*

**BLUETOOTH
DEVICE**

TEST RESULT SCREEN
M20B00007v69.02

**TEST
ABBREVIATION**

**TESTING
PARAMETER**

Alkalinity, Total AL

**PARAMETER
UNIT OF MEASURE**

*Scroll unit of measure values.
Some tests offer results in
multiple units of measure.*

81
ppm

TEST RESULT

**DOTS INDICATE
MULTIPLE UNITS OF MEASURE**
(Available for some tests)

Menu

The **Menu slide-out** is available from any screen within the app. The **Menu** allows you to access any of the app's features with ease.



HOW TO VIEW YOUR GPS LOCATION

In the **Settings** screen you can view your current GPS coordinates and/or refresh your current location coordinates.

ABOUT

Access the End-User License Agreement and contact information to reach our offices in the USA and Europe from the **About** section located in the **Menu** slide-out.

Located in the *About* section you can find which version of the app you are running. Be sure to check for updates and install the latest version before running a test as we are constantly updating and adding more features to the app!

Test

You can utilize two different testing methods under **Test**; Manual Entry or Bluetooth Device.

BLUETOOTH TEST

Tests will be performed with your eXact® Micro 20 Standard strip method or Standard drop method test procedure.

MANUAL TEST

This feature allows you to utilize other testing methods and manually enter your results into the app by selecting **Test**, **Manual entry**, then **Change value** to enter.



History

The **History** stores all your saved test result information and allows you to sort by date, customer name, or test type.

HOW TO SORT BY CUSTOMER

To sort by **Customer** begin by clicking **History**, then **Sort by**, **Customer**. You can then scroll through your list of customers, by name, to find a specific test result.

HOW TO SORT BY DATE RANGE

To sort by **Date** begin by clicking **History** then **Sort by**, **Date**. You can then scroll through a list of tests performed by date. You can also set a specific date range period by selecting **Date range**. Then set your **From** and **To** date ranges.

HOW TO SORT BY TESTS

To sort by **Test** begin by clicking **History** then **Sort by**, **Test**. You can then scroll through a list of tests sorted in alphabetical order.

HOW TO ACCESS HISTORY MAP

The History Map stores GPS locations of testing sites. See page 20 for instructions on how to utilize this feature.

eXact iDip® app for eXact® Micro 20 with Bluetooth® SMART




The benefits of purchasing an eXact Micro 20 with Bluetooth® SMART is the ability to connect your photometer to a smartphone or tablet and use our specially designed app to easily save, send, and share your test results. First, select your test using the menu selections on your Micro 20 then run the test and the results will simultaneously display on the app and on your Micro 20 with Bluetooth® SMART. From the app, you can save, send, and share the results directly linked to your customers information and GPS location to include date/time stamp. The eXact iDip® app is available for use on both Apple and Android devices. The app is compatible with Bluetooth 4.0 devices (Android 4.3 and Apple iOS 6.1 to 8.1) For a full list of compatible devices please visit www.sensafe.com/compatible-devices




When using your eXact® Micro 20 with Bluetooth® SMART and the eXact iDip app. You will need to complete the following steps after you have rinsed and filled your cell with the water sample and prior to zeroing your meter and dipping your strip.

HOW TO SELECT A CUSTOMER

From your smartphone/tablet, launch the eXact iDip® app. Select **`Customers`** from the home screen. From the Customers list **`Choose your customer from contacts`** or add a new customer by selecting the **`+`** in the upper right hand corner.




Please Note: When you add a new customer to your recent customers list, they will not appear in the list until you have ran and saved a test using the app and your eXact® Micro 20 with Bluetooth® SMART photometer.



HOW TO CONNECT VIA BLUETOOTH

Tap the menu slide out screen **`≡`** and select **`Bluetooth test`**. Select your eXact® Micro 20 from the list located at the bottom of the screen. Verify it has connected and tap **`OK`**.

The serial number is located on the back of your device, this will display in the app. Refer to the serial number to ensure you are connected to the correct device.




READ RESULTS

Results will display simultaneously on the Micro 20 photometer and app.

Please note that batch uploading tests saved on your eXact Micro 20 to a smartphone/tablet is not yet available. If you want to save a test result from your Micro 20 to the app, you will need to run the test while connected to the app or enter the result manually.

Managing data with the eExact® iDip app

After you have ran your test, you can save, send, and share your results, by following the steps below.





MANAGING DATA (SAVE/SEND/SHARE)

When all tests have been performed, select 'Results' at the bottom of the screen. To add notes tap the desired test result.


TYPE NOTES

Add any notes you wish into the 'Notes' box. The additions will be automatically saved.



MANAGING DATA (SAVE/SEND/SHARE)


Go back to 'Results' and select 'Save' to store the test results with notes into 'History'



HOW TO MANAGE DATA IN HISTORY

In History, you can edit, select, and email your results.

To email, you can either tap an individual result, or use the 'Select' button to access multiple data points.




HOW TO EMAIL RESULTS FROM HISTORY

Press the blue envelope icon from a single result selection. If multiple test results are selected, tap 'email' at the top to send the result information for selected results. Additional notes cab be added to the body of the email. Add recipients and tap send to complete.

ACCESSING RESULTS FROM HISTORY MAP

Tap 'Map' on 'History' page to access the History Map. Double tap or use fingers to zoom into an area. Select a pin by tapping to bring up results. From the specific location you can bring up the test results details page.



eXact® Micro 20 Test Specifications

Parameter / Test		Part No.	Pg	MENU & SELECT Group						
				①	②	③	④	⑤	⑥	⑦
1	Alkalinity, Total	486641		AL4 ₁		AL1 ₃	AL1 ₄			
2	Aluminum (Al ³⁺) ¹	486821							Al3 ₆	
3	Ammonia (NH ₃)	486654			NH3 ₂		NH4 ₄			
4	Biguanide	486810				bG8 ₃				
5	Bromine (DPD-1)	486637				bR6 ₃		bR2 ₅		
6	Calcium (CaCO ₃)	486629				CA5 ₃				
7	Chloride (NaCL) III	486757					CH5 ₄			
8	Chloride, HR (NaCL) III	486757				CHH ₃				
9	Chlorine Dioxide (DPD-1)	486633						Cd4 ₅		
10	Chlorine, Combined	486637 & 486638		CL1 ₁		CL3 ₃	CL6 ₄	CL1 ₅		
11	Chlorine, Free (DPD-1)	486637		CL1 ₁		CL3 ₃	CL6 ₄	CL1 ₅		
12	Chlorine, High Range	486672						CLH ₅		
13	Chlorine, Total (DPD-4)	486670		CL1 ₁		CL3 ₃	CL6 ₄	CL1 ₅		
14	Chromium (VI)	486614							CR6 ₆	
15	Copper (Cu ²⁺)	486632		CU6 ₁			Cu9 ₄			
16	Cyanide	486812			CN1 ₂					
17	Cyanuric Acid II	481652-II				CY7 ₃				
18	Fluoride	486643		F8 ₁					F1 ₆	
19	Hardness, Total HR (as CaCo ₃)	486656							THH ₆	
20	Hardness, Total LR (as CaCo ₃)	486630							THL ₆	
21	Hydrogen, Peroxide	486648						HP6 ₅		
23	Iron, Total	486650		FE3 ₁	FE2 ₂					
24	Manganese (Mn ²⁺)	486606		MN7 ₁					MN2 ₆	
25	Molybdate	486653							Mo7 ₆	
26	Nitrate (NO ₃)	486655					NO3 ₄			
27	Nitrate (salt water>400 ppm)	486655								TR1 ₇
28	Nitrite (NO ₂)	486623					NO2 ₄			
29	Ozone (DPD-4)	486634						O3 ₅		
30	Peracetic Acid (DPD-4)	486674						PA5 ₅		
31	Permanganate (DPD-1)	486626						PM7 ₅		
32	pH	486639		PH2 ₁		PH2 ₃	PH7 ₄			
33	pH, Acid	486624			PHA ₂					
34	pH, Alkali	486609			PHb ₂					
35	Phosphate (PO ₄)	486814			PO4 ₂	PO4 ₃	PO8 ₄			
34	Quaternary Ammonia	486823						QA8 ₅	QA5 ₅	
36	Sulfate (SO ₄)	486608							SO4 ₆	
37	Sulfide (S ²⁻)	486818			S5 ₂					
38	Turbidity	N/A		TU5 ₁						

For select group & menu overview, **see page 12.**

¹ Value provided represents best possible accuracy under laboratory conditions, but may vary throughout the detection range. For a complete list of accuracies throughout all ranges please visit [sensafe.com/micro20/specifications](https://www.sensafe.com/micro20/specifications).

Count-up Time	Reagents Used	Range (ppm)	Resolution Best Accuracy [†]	
Immediate	AL Strip	8 – 200	0.1(8–50), 1(51–200)	7.5
80 seconds	5 Drops AL Buffer & AL Strip	0.1 – 1.2	0.01	13
500 seconds	3 Drops NH (reg. water) or 10 Drops NH (salt water), & NH Strip	0.02 – 2.4	0.01	5
Immediate	BG Strip	1.6 – 210	0.1(0–20), 1(21 – 210)	7.5
Immediate	CL (DPD–1) Strip	0.01 – 12	0.01(0.1–2), 0.1(2.1 – 12)	5
Immediate	CA Strip	16 – 400	1	10
Immediate	CH Strip	1 – 330	1	5
Immediate	1:20 Dilution of sample & CH Strip	26 – 6600	20	8
Immediate	Glycine Strip & CL (DPD–1) Strip	0.01 – 10	0.01	5
Immediate	CL (DPD–1) Strip & CL (DPD–3) Strip	0.01 – 6.2	0.01	3
Immediate	CL (DPD–1) Strip	0.01 – 6.2	0.01	3
120 seconds	HR Strip	1 – 300	0.01(0.3–20), 1(21 – 300)	5
Immediate	CL (DPD–4) Strip	0.01 – 6.2	0.01	3
240 seconds	Cr Strip	0.01 – 1.8	0.01	5
120 seconds	CU Strip	0.01 – 11	0.01(0–4), 0.1(4.1 – 11)	2
600 seconds	CN–1 Strip & CN–2 Strip	0.01 – 1.1	0.01	13
60 seconds	5 Drops CY	7 – 110	1	8
Immediate	10 Drops F	0.04 – 1.5	0.01	15
Immediate	THH Strip	60 – 600	1	12
Immediate	THL Strip	1 – 80	1	10
120 seconds	HP Strip	0.3 – 100	0.01	8
40 seconds	EZ Open Reducer (Powder) & FE Strip	0.03 – 6	0.01(0.03–2.5), 0.1(2.51 – 6)	3
120 seconds	MN#1 Strip, MN#2 Strip, & 3 Drops MN	0.01 – 1.5	0.01	6
120 seconds	MO Strip, 5 drops of MO Reagent	0.01 – 3	0.01	15
600 seconds	NO3 Strip	0.1 – 30	0.01(0–5), 0.1(5.1 – 30)	15
580 seconds	NO3 Strip	0 – 90	0.01	15
360 seconds	NO2 Strip	0.01 – 1.8	0.01	5
Immediate	O3 (DPD–4) Strip	0.01 – 2	0.01	4
Immediate	PA (DPD–4) Strip	0.01 – 6	0.01	3
Immediate	PM (DPD–1) Strip	0.01 – 5	0.01	2
Immediate	PH Strip	5.8 – 8.5 pH	0.1	0.2
Immediate	Acid PH Strip	3.2 – 6 pH	0.1	0.3
Immediate	Alkali PH Strip	7.2 – 9.8 pH	0.1	0.3
120 seconds	PO4 Strip	0.03 – 4.4	0.01(0.03–2.5), 0.1(2.6 – 4.4)	4
Immediate	QA Strip	2 – 80	1	6
Immediate	SO4 Strip	2 – 210	1	10
180 seconds	4 Drops S & S2 Strip	0.01 – 1.7	0.01	6
Immediate	Distilled or DI water	4 – 800ntu	1	10

¹ Performance verified with various salt systems and water samples with optimal water temperature at 10°–40°C / 50°–104°F. Optimal water temperature for Total Alkalinity test is 15°–40°C / 59°–104°F.

² For example: If the sample has 1 ppm of Free Chlorine, the meter may read 0.97 ppm or 1.03 ppm. Contact sales department for detailed meter accuracy values.

Select group overview

Below is a list of Menu and selection choices with abbreviations. MENU SELECT

1 Water Quality	2 Miscellaneous	3 Pool & Spa	4 Environmental
CL1 — Free & Total Chlorine PH2 — pH FE3 — Iron AL4 — Total Alkalinity TU5 — Total Hardness CU6 — Copper MN7 — Manganese F8 — Fluoride	CN1 — Cyanide FE2 — Iron (II) & Total Iron NH3 — Ammonia PO4 — Phosphate S5 — Sulfide PHA — Acid pH PHb — Alkali pH	AL1 — Total Alkalinity PH2 — pH CL3 — Free & Total Chlorine PO4 — Phosphate CA5 — Calcium Hardness bR6 — Bromine Cy7 — Cyanuric Acid bG8 — Biguanide CHH — Chloride	AL1 — Total Alkalinity NO2 — Nitrite NO3 — Nitrate NH4 — Ammonia CH5 — Chloride CL6 — Free & Total Chlorine PH7 — Drinking Water PO8 — Phosphate CU9 — Copper

5 Oxidizers	6 Specialty	7 Transmission
CL1 — Free & Total Chlorine bR2 — Bromine O3 — Ozone Cd4 — Chlorine Dioxide PA5 — Peracetic Acid HP6 — Hydrogen Peroxide PM7 — Permanganate QA8 — Quaternary Ammonia CLH — High Range Chlorine	F1 — Fluoride MN2 — Manganese Al3 — Aluminum SO4 — Sulfate QA5 — Quaternary Ammonia CR6 — Chromium Mo7 — Molybdate THH — High Range Total Hardness THL — Low Range Total Hardness	TR1 — Transmission (525nm) TR2 — Transmission (638nm)

eXact® Micro 20 meter messages

The following are some of the common messages that may display on your photometer, including error messages. If an error message other than those listed below is displayed, please contact technical support in the USA at (803) 329-0162 (ext. 0).

LCD Message	Description	Corrective Action
HI	In READ mode: test sample concentration is above the measurement range (test specific).	Dilute and retest. Dilution kit available (Part No. 487200)
LO	In READ mode: test sample concentration is below the measurement range (test specific).	Sample value is below measurement range.
LO	In ZERO mode: sample absorbency (due to a cloudy or colored sample or a dirty cell) is too high to zero, the meter will read "LO" or low battery	Dilute sample, filter sample or clean cell. One of these options should remedy the problem. You may need to replace batteries if low battery indication.
ER	Excessive stray light detected. Normally this does not occur, even when testing in sunlight.	Place the light blocking CAP over the CELL for zeroing and for reading result. Moving to a shaded can also fix this problem.
 in lower left	Low battery indication during testing (meter may not zero)	Replace batteries.

About the Sample Cell (built-in 4 mL)


The built-in Sample Cell (CELL) is transparent plastic and, when filled to the top, contains 4 mL. The sturdy CELL design will last for over 20,000 readings. Scratches on the CELL will not interfere or compromise the accuracy of the readings because of its fixed position. For best accuracy, rinse cell with clean water immediately after a test is completed. Do not use solvents, such as acetone, to clean the cell. When the CELL becomes stained or cloudy from repeated testing, or when the meter does not blank when you press the ZERO/ON button, the cell needs to be cleaned. Clean as follows: Fill cell with clean water and move the Cell cleaning brush up-and-down and back-and-forth along the walls of the cell. Afterwards, rinse the cell and the meter is ready for use again. Cleaning the cell regularly is especially recommended after you run a test that is using turbidity or precipitation chemistry for analysis (Calcium Hardness and Cyanuric Acid).

How to install or replace "AAA" batteries

Batteries are not included. The meter requires (4) AAA in order to function.

- ① Unscrew the O-ring sealed battery cover counter-clockwise. Use proper sized pliers if necessary. Do not disturb the sealing O-ring.
- ② Remove the used batteries and install 4 new AAA batteries following the diagram for correct polarity (see diagram). We recommend high quality AAA alkaline batteries be used.
- ③ Replace the battery cover. Be sure to tighten the cover securely. This is necessary for meter to ensure it is waterproof.
- ④ Dispose of the used batteries in accordance with your local regulations.
- ⑤ Press  button to confirm the meter turns on. The meter is now ready for operation.

Meter will not work if battery orientation is incorrect.



2- Year limited warranty

Registration of your eXact® photometer must be received within 30 days from date of purchase to activate the warranty. The eXact® photometer is warranted to be free from defects in materials and workmanship for a period of two (2) years from the date of purchase by the customer. ITS will repair or replace any part of the product which is deemed to be faulty or otherwise defective. The non-transferable warranty does not cover product damage caused by abuse (such as crushing a tablet in the cell) or improper use. If the meter is faulty or otherwise defective contact ITS by phone (+1-803-329-9712 Ext. 0) or email (its@sensafe.com) to describe the problem and obtain a return authorization form before returning the photometer to ITS. Damage caused by improper packing of the photometer for return shipment to ITS will not be covered by the warranty. Customer is responsible for shipping charges to ITS. ITS pays postage when photometer is returned to customer. A maximum processing fee of \$75 will be charged for repair or replacement of non-registered photometers and damages not covered by this warranty. Registration is available over the phone (+1-803-329-9712 Ext. 0) or on-line at <http://www.sensafe.com/micro/warranty/> (Personal data is kept confidential).

Using eXact® Strip Micro Standard strip test procedure

Used for Acid pH, Alkali pH, Biguanide, Bromine, High Range Total Hardness¹, Low Range Total Hardness¹, Calcium Hardness², Chloride³, Chromium⁴, Copper, High Range Chlorine⁵, Hydrogen Peroxide, Nitrate⁶, Nitrite, Ozone, Paracetic Acid, Permanganate, pH, Phosphate⁷, Quaternary Ammonia, Sulfate, and Total Alkalinity⁸
(SEE PAGE 13 FOR SPECIAL NOTES)



REMOVE STRIP

Remove 1 **eXact® Strip Micro** from the bottle before beginning the test. Set the strip in a dry, convenient place and recap the bottle immediately.



TURN METER ON

Press the **ON ZERO** button to power the meter on; the display will show all annunciators, then the current **(MENU)** selection, followed by the last reading.



SELECT GROUP AND MENU

Press and re-press the **(SELECT)** button to Select your group. Next, to select the test parameter press and re-press the button to select **(MENU)**.



RINSE CELL AND FILL WITH SAMPLE

Rinse the **CELL** at least 3 times with the water sample you will be testing (rinsing minimizes the potential for cross-contamination from a previous test). Finally, fill the **CELL** to capacity (4 mL) with the water sample.

 If using eXact® Micro 20 with Bluetooth® SMART follow the steps to **SELECT CUSTOMER** & **CONNECT DEVICE VIA BLUETOOTH** before proceeding (see page 6).



ZERO METER

Press the **ON ZERO** button. The cursor will move across the display followed by a display reading of **0.00 PPM**. This indicates the sample is ready for testing.



DIP STRIP & PRESS READ

Dip the required strip into the **CELL**, and immediately press **(READ)**. This will start a **20 second** countdown timer. During this time, move the strip in a gentle back and forth motion (approx. 2 strokes per second). **Remove and discard the strip after '1' on the display disappears.**



RECORD RESULT

The cursor will move across the display while the meter prepares to measure the sample. Record the value displayed for the respective parameter. This value is automatically stored in the **MENU**.

eXact® Strip Micro test procedure (SPECIAL NOTES)

¹ **Total Hardness, HR & LR** — Positive interferences are observed if the test sample contains Barium. Interferences also observed if the test sample contains Copper, Lead, Cobalt, or Nickel.

² **Calcium Hardness** — This test is accurate in water with Chloride < 10,000 ppm as NaCl.

³ **Chloride** — If sample pH is high (>9), adjust pH to 5–6 using Vinegar.

⁴ **Chromium** — The strip needs to be angled in order to fit in the **CELL** because it is too wide and total alkalinity are in these ranges before running this test.

⁵ **High Range Chlorine** — Use a 10 second dip time if water temperature is above 40°C (113°F).

INTERFERENCES: Oxidizers such as Chloramine, Chlorine Dioxide, Bromine, Iodine, Ozone, Bromamines, and Permanganate will give false positive readings.

⁶ **Nitrate** — **A.** Use this procedure if NaCl is less than 400 ppm. Otherwise, use Nitrate (Salt Water) Procedure on page 26. **B.** The **CELL** needs to be cleaned with brush and distilled water after each test. If any zinc dust is adhering to the **CELL** wall, it will affect results.

⁷ **Phosphate** — **A.** Clean **CELL** with 0.1N HCl, Distilled Vinegar (5%), or Muriatic Acid before filling the meter with the sample to be tested. **B.** If running multiple tests in a row, using the same water sample, the **CELL** does not have to be rinsed or cleaned with acid between each test. It is recommended that the **CELL** be rinsed three times with the sample water. **C.** The calibration of the meter is based on a water temperature between 15°C (59°F) and 31°C (88°F). If temperature is below 15°C (59°F), your final Phosphate value may read low. This test can also be used for salt water testing.

⁸ **Total Alkalinity** — For water temperatures above 35°C (95°F, hot tubs), remove and discard the strip when the timer displays “10”; countdown continues.



The eXact® Micro 20 Dual Wavelength Advanced Photometer System is designed for use with the eXact® Strip Micro reagent delivery system. The eXact® Micro 20 Dual Wavelength Photometer is manufactured and tested in an ISO 9001 Facility.

Using eXact® reagent Standard liquid test procedure

Used for Cyanuric Acid¹ and Fluoride²



TURN METER ON

Press the **(ON ZERO)** button to power the meter on; the display will show all annunciators, then the current **(MENU)** selection, followed by the last reading.



SELECT GROUP AND MENU

Press and re-press the **(SELECT)** button to Select your group. Next, to select the test parameter press and re-press the button to select **(MENU)**.



RINSE CELL AND FILL WITH SAMPLE

Rinse the **CELL** at least 3 times with the water sample you will be testing (rinsing minimizes the potential for cross-contamination from a previous test). Finally, fill the **CELL** to capacity (4 mL) with the water sample.



If using eXact® Micro 20 with Bluetooth® SMART follow the steps to **`SELECT CUSTOMER`** & **`CONNECT DEVICE VIA BLUETOOTH`** before proceeding (see page 6).



ZERO METER

Press the **(ON ZERO)** button. The cursor will move across the display followed by a display reading of **0.00 PPM**. This indicates the sample is ready for testing.



ADD DROPS

Using the selected bottle of reagent add the required drops (see chart on page 8–9) and cover the **CELL** with the **CELL COVER**. *Precaution: Ensure that the bottle is straight when dispensing drops.*



PRESS READ & MIX

Press **(READ)** and a **20 second** countdown begins. During this time, turn the meter upside down repetitively. When the timer displays **`1`**, place the Micro 20 on a flat surface. Wait for count-up time.



RECORD RESULT

The cursor will move across the display while the meter prepares to measure the sample. Record the value displayed for the respective parameter. This value is automatically stored in its **MENU** and if using the eXact iDip® app, the result will be saved in the app's **`HISTORY`**.

eXact® reagent liquid drops (SPECIAL NOTES)

¹ **Cyanuric Acid** — Shake the bottle vigorously to mix before adding the drops to the sample

² **Fluoride** — The reagent contains acid, a stir bar may be used to mix the reagent.



The eXact® Micro 20 Dual Wavelength Advanced Photometer System is designed for use with the eXact® Strip Micro reagent delivery system. The eXact® Micro 20 Dual Wavelength Photometer is manufactured and tested in an ISO 9001 Facility.

*Some test procedures require the use combination of more than one test strip and/or liquid reagents. Please follow the step-by-step procedures as outlined in the following pages to ensure the best accuracy. For more tips on best accuracy, **see page 28.***

Any deviation from the outlined procedure, could result in inaccurate test results. Please take caution to testing notes. If your required procedure is not listed in this manual or if you have any questions, feel free to contact us at its@sensafe.com

Free or Total Chlorine test procedure

DPD–1 strip used for Free Chlorine detection, DPD–4 strip used for Total Chlorine detection



REMOVE STRIP

Remove 1 **eXact® Strip Micro CL (DPD–1)**, Part No. 486637 or **eXact® Strip Micro CL (DPD–4)**, Part No. 486670 from the bottle before beginning the test. Set the strip in a dry, convenient place and recap the bottle immediately.



TURN METER ON

Press the **ON ZERO** button to power the meter on; the display will show all annunciators, then the current **(MENU)** selection, followed by the last reading.



SELECT GROUP AND MENU

Press and re-press the **(SELECT)** button to Select your group (**Select 1, 4, or 5**). Next, to select the test parameter press and re-press the **(MENU)** button to select **CL1** or **CL6**.



RINSE CELL AND FILL WITH SAMPLE

Rinse the **CELL** at least 3 times with the water sample you will be testing (rinsing minimizes the potential for cross-contamination from a previous test). Finally, fill the **CELL** to capacity (4 mL) with the water sample.

 If using **eXact® Micro 20 with Bluetooth® SMART** follow the steps to **`SELECT CUSTOMER`** & **`CONNECT DEVICE VIA BLUETOOTH`** before proceeding (see page 6).



ZERO METER

Press the **ON ZERO** button. The cursor will move across the display followed by a display reading of **0.00 PPM**. This indicates the sample is ready for testing.



DIP STRIP & PRESS READ

Dip the required strip into the **CELL**, and immediately press **(READ)**. This will start a **20 second** countdown timer. During this time, move the strip in a gentle back and forth motion (approx. 2 strokes per second). **Remove and discard the strip after `1` on the display disappears.**



RECORD RESULT

The cursor will move across the display while the meter prepares to measure the sample. Record the value displayed for the respective parameter. This value is automatically stored in the **CL MENU**, and if using the **eXact iDip®** app, the result will be saved in the app's **`HISTORY`**

Combined Chlorine test procedure

IMPORTANT: DO NOT discard the sample from the Free Chlorine (DPD-1) test if you are planning to run eXact® Strip Micro DPD-3 (Total Chlorine) Procedure. Move directly to steps 8–10 below, otherwise immediately rinse the CELL.

This procedure is only valid when ran as a continuation of the eXact® Strip Micro CL (DPD-1 Free Chlorine). Test procedure located on the previous page.



REMOVE STRIP

Remove 1 eXact® Strip Micro CL (DPD-3), Part No. 486638 from the bottle before beginning the test. Set the strip in a dry, convenient place and recap the bottle immediately.



DIP STRIP & PRESS READ

Dip the eXact® strip micro (DPD-3) into the CELL, and immediately press **READ**. This will start a **20 second** countdown timer. During this time, move the strip in a gentle back and forth motion (approx. 2 strokes per second). **Remove and discard the strip after '1' on the display disappears.** The cursor will move across the display while the meter prepares to measure the sample. Record the value displayed and this value is automatically stored in the **CL MENU**, the result will be saved in the app's **'HISTORY'**.

(NOTE: The Iodide added with DPD-3 will, in the presence of Combined Chlorine or Chloramines, convert into Iodine).



PRESS READ AGAIN

Press **READ** again and the meter will count down and display the next reading. If this reading matches the previous result, then record this as the Total Chlorine result. This value is automatically stored in the **CL MENU**. After testing is completed, rinse **CELL** immediately. Record the highest value the meter displayed as your Total Chlorine result.

NOTE: Standard Method (4500–Cl G, procedure for total chlorine) requires the reading to be made after 2 minutes from the time the KI is added. For compliance testing, you must time the 2 minutes and then make your measurement.

From testing in our lab, water samples above 70°F (20°C), generally, reach a stabilized reading quicker than 2 minutes. Chlorine and Iodine react with N, N-diethyl-p-phenylenediamine as it is released from the strip to form a magenta color, directly proportional to the Chlorine concentration. (Ozone, Bromine, and Permanganate also form the color).



The eXact® Micro 20 Dual Wavelength Advanced Photometer System is designed for use with the eXact® Strip Micro reagent delivery system. The eXact® Micro 20 Dual Wavelength Photometer is manufactured and tested in an ISO 9001 Facility.

Aluminum, Ammonia & Sulfide test procedure

1 TURN METER ON

Press the  button to power the meter on; the display will show annunciators, followed by the current selection. It will then display the last reading.

2 SELECT GROUP & MENU

Press and re-press the  button to **Select Group**. Press and re-press the  button to select the test parameter (see chart on page 10–11).

3 RINSE & FILL CELL WITH SAMPLE

Rinse the **CELL** at least 3 times with the water sample you will be testing—rinsing minimizes the potential for cross-contamination from a previous test. Finally, fill **CELL** to capacity (4 mL) with the water sample.



If using eXact® Micro 20 with Bluetooth® SMART follow the steps to **SELECT CUSTOMER** & **CONNECT DEVICE VIA BLUETOOTH** before proceeding (see page 6).

4 ADD DROPS

Using the selected bottle of reagent add the required drops (see chart on page 10–11) and cover the **CELL** with the **CELL COVER**. *Precaution: Ensure that the bottle is straight when dispensing drops.*

5 ZERO METER

Press the  button. The cursor will move across the display followed by **0.00 PPM**. This will indicate that the sample is ready for testing.

6 DIP STRIP & PRESS READ

Using the required strip (see chart on page 10–11), dip strip into the **CELL**, and immediately press  to initiate a 20 second countdown. Move the strip using a gentle back and forth motion (approx. 2 strokes/sec). Remove and discard the strip after “1” on the display disappears. The meter will automatically start to count up. The count up time will vary for each parameter. At the end, the cursor will move across the display while the meter prepares to measure the sample. Record the value displayed for the respective parameter. This value is automatically stored in its MENU and if using the eXact iDip® app, the result will be saved in the app's **HISTORY**. After testing, rinse **CELL** immediately and clean with the brush. After Sulfide testing: rinse **CELL** with Distilled White Vinegar, 0.1N HCl, or Muriatic Acid and clean with brush.

Aluminum, Ammonia & Sulfide (SPECIAL NOTES)

¹ **Aluminum** — **A.** First, clean the **CELL** with 0.1N HCl, Distilled Vinegar (5%), or Muriatic Acid before filling the meter with the sample to be tested. **B.** If running multiple tests in a row, using the same water sample, the **CELL** does not have to be rinsed or cleaned with acid between each test. It is recommended that the **CELL** be rinsed 3 times with the sample water.

² **Ammonia** — The calibration of the meter is based on a water temperature between 14°C (57°F) and 28°C (82°F). If temperature is below 14°C (57°F), your final Ammonia value may read low.

³ **Sulfide** — **A.** For results as Hydrogen Sulfide (H₂S), multiply the resulting value by 1.06. **B.** The calibration of the meter is based on the water sample temperature above 20°C (68°F). If the water sample is below 20°C (68°F), the strip has to dip in the sample for an additional 10 seconds.

Cyanide test procedure

1 REMOVE STRIPS

Remove 1 **eXact® Strip Micro CN-1** Part No. 486812-A and **eXact® Strip Micro CN-2** Part No. 486812-B from the bottle before beginning the test. Set the strip in a dry, convenient place and recap the bottle immediately.

2 TURN METER ON

Press the  button to power the meter on; the display will show annunciators, followed by the current selection. It will then display the last reading.

3 SELECT GROUP & MENU

Press and re-press the  button to **Select Group 2**. Press and re-press the  button to select the test parameter (see chart on page 10–11).

4 RINSE & FILL CELL WITH SAMPLE

Rinse the **CELL** at least 3 times with the water sample you will be testing—rinsing minimizes the potential for cross-contamination from a previous test. Finally, fill **CELL** to capacity (4 mL) with the water sample.



*If using eXact® Micro 20 with Bluetooth® SMART follow the steps to **SELECT CUSTOMER` & CONNECT DEVICE VIA BLUETOOTH`** before proceeding (see page 6).*

5 ZERO METER

Press the  button. The cursor will move across the display followed by **0.00 PPM**. This will indicate that the sample is ready for testing.

6 DIP STRIP & PRESS READ

Dip the **CN-1** strip into the **CELL**, and immediately press . This starts the 30 Second countdown timer. Because the strip is 8 mm wide, the strip will need to be angled to fit in the cell. Be sure that the test pad is fully submerged. During this time, move the strip in a gentle back and forth motion (approx. 2 strokes/sec). **Remove and discard the strip after `1` on the display disappears.** The cursor will move across the display, at this point have your **CN-2** strip ready to dip into the **CELL**.

When the 30 Second countdown starts, immediately dip the **CN-2** strip into the **CELL**. During this time, with the strip angled slightly, move the strip in a gentle back and forth motion (approx. 2 strokes/sec). **Remove and discard the strip after `1` on the display disappears.**

The meter will automatically start to count up to 600 seconds. At 600 seconds, the cursor will move across the display while the meter prepares to measure the sample. Record the result displayed. This result is automatically stored in **CN MENU**, and if using the eXact iDlp® app, will be stored in the app's **‘HISTORY’**.

After testing, rinse **CELL** immediately and clean with the brush.

NOTE: The calibration of the meter is based on a water temperature between 20°C (68°F) and 25°C (77°F). If temperature is below 20°C (68°F), your final Cyanide value may read low.

Total Iron test procedure

To ensure accurate results, do not run this test immediately after a sulfide test

1 TURN METER ON

Press the  button to power the meter on; the display will show annunciators, followed by the current selection. It will then display the last reading.

2 SELECT GROUP & MENU

Press and re-press the  button to **Select Group**. Press and re-press the  button to select the test parameter (see chart on page 10–11).

3 RINSE & FILL CELL WITH SAMPLE

Rinse the **CELL** at least 3 times with the water sample you will be testing—rinsing minimizes the potential for cross-contamination from a previous test. Finally, fill **CELL** to capacity (4 mL) with the water sample.



If using eXact® Micro 20 with Bluetooth® SMART follow the steps to **SELECT CUSTOMER** & **CONNECT DEVICE VIA BLUETOOTH** before proceeding (for instructions, see page 6).

4 ADD REDUCER

Tilt the meter to discard about 0.2 mL sample in order to leave room for reagent. Add the contents of one **eXact® Reagent EZ Open REDUCER** (Part No. 486601) to the **CELL** and cover the **CELL** with the mixing cap. Press  to start the 20 second countdown timer. Place thumb over **CELL COVER** to secure in place and mix the sample by turning the meter upside-down repetitively. **When countdown displays `1`**, hold the meter upright and the cursor will flash. At this time the meter will begin a 40 second count up. After the count up, a result will be displayed (ignore this result).

5 ZERO METER

Press the  button. The cursor will move across the display followed by **0.00 PPM**. This will indicate that the sample is ready for testing.

6 DIP STRIP & PRESS READ

Dip the **eXact® Strip Micro FE (TPTZ)** (Part No. 486631) into the **CELL** and immediately press . This starts the 20 second countdown timer. During this time, move the strip in a gentle back and forth motion (approx. 2 strokes/sec). Remove and discard the strip after `1` on the display disappears. The meter will automatically start to count up for 40 seconds. At the end, the cursor will move across the display while the meter prepares to measure the sample. Record result displayed (this value is automatically stored in the **FE MENU**, and if using the eXact iDip® app, will be stored in the app's **HISTORY**).

After testing is completed, rinse **CELL** immediately and clean with brush.

Total Iron (SPECIAL NOTES)

Total Iron — A. First, clean the **CELL** with 0.1N HCl, Distilled Vinegar (5%), or Muriatic Acid before filling the meter with the sample to be tested. **B.** If running multiple tests in a row, using the same water sample, the **CELL** does not have to be rinsed or cleaned with acid between each test. It is recommended that the **CELL** be rinsed 3 times with the sample water.

High Range Chloride test procedure

This test requires a 1:20 dilution of the salt system sample – Mini Dilution Kit (Part No. 487202)

1 PREPARE SAMPLE FOR TESTING

Using the **Mini Dilution Kit** (Part No. 487202) and **Distilled** or **Deionized** (salt-free water) prepare a 1 to 20 (1:20) dilution of your sample.

PREPARE DILUTION SAMPLE

1. Rinse the syringe 3 times with salt system sample that you want to test by moving the plunger up and down.
2. Rinse 50 mL graduated conical tube with distilled or deionized (salt-free) water.
3. Rinse the 3.0 mL syringe with water sample to be tested. Fill the 3.0 mL syringe to the 2.0 mL line precisely (plunger ring should line up at the 2.0 mL line and little or no air bubble should be present).
4. Add the syringe content (2.0 mL salt system sample) to the clean 50 mL graduated conical tube by pushing the plunger all the way down to expel sample.
5. Fill the graduated conical tube to the 40 mL line with distilled or deionized (salt-free) water and place cap on top.
6. Mix content of graduated conical tube by turning upside down at least 3 times.

Sample is now ready for testing.

2 REMOVE STRIPS

Remove 1 **eXact® Strip Micro Chloride III** Part No. 486757 from the bottle before beginning the test. Set the strip in a dry, convenient place and recap the bottle immediately.

3 TURN METER ON

Press the  button to power the meter on; the display will show annunciators, followed by the current selection. It will then display the last reading.

4 SELECT GROUP & MENU

Press and re-press the  button to **Select Group**. Press and re-press the  button to select the test parameter (see chart on page 10–11).

5 RINSE & FILL CELL WITH SAMPLE

Using the 1:20 Dilution Sample prepared above, rinse the **CELL** 3 times. Then fill the **CELL** to capacity (4 mL) with the 1:20 Dilution Sample.



If using eXact® Micro 20 with Bluetooth® SMART follow the steps to **‘SELECT CUSTOMER’** & **‘CONNECT DEVICE VIA BLUETOOTH’** before proceeding (see page 6).

6 ZERO METER

Press the  button. The cursor will move across the display followed by **0.00 PPM**. This will indicate that the sample is ready for testing.

7 DIP STRIP & PRESS READ

Dip the **Chloride III** strip into the **CELL**, and immediately press . This starts a 20 second countdown timer. During this time, move the strip in a gentle back and forth motion (approx. 2 strokes/sec). Remove and discard the strip after ‘1’ on the display disappears. The cursor will move across the display, while the meter measures the sample. Record the result displayed (this result is automatically stored in CH MENU and, if using the eXact iDip® app, will be stored in ‘History’). After testing is completed, rinse **CELL** immediately and clean with the brush.

Chlorine Dioxide test procedure

1 REMOVE STRIPS

Remove 1 **eXact® Strip Micro Glycine** Part No. 484014 and **eXact® Strip Micro Cd** (DPD-1) Part No. 486633 strips from the bottle before beginning the test. Set the strip in a dry, convenient place and recap the bottle immediately.

2 TURN METER ON

Press the  button to power the meter on; the display will show annunciators, followed by the current selection. It will then display the last reading.

3 SELECT GROUP & MENU

Press and re-press the  button to **Select Group 5**. Press and re-press the  button to select the test parameter **Cd4** (see chart on page 8–9).

4 RINSE & FILL CELL WITH SAMPLE

Rinse the **CELL** at least 3 times with the water sample you will be testing—rinsing minimizes the potential for cross-contamination from a previous test. Finally, fill **CELL** to capacity (4 mL) with the water sample.



If using eXact® Micro 20 with Bluetooth® SMART follow the steps to 'SELECT CUSTOMER' & 'CONNECT DEVICE VIA BLUETOOTH' before proceeding (see page 6).

5 DIP STRIP & PRESS READ

Dip the **Glycine** strip into the **CELL**, and immediately press . This starts a 20 second countdown timer. During this time, move the strip in a gentle back and forth motion (approx. 2 strokes/sec). Remove and discard the strip after '1' on the display disappears. The cursor will move across the display, while the meter measures the sample, when a result appears, ignore value and continue to step 8.

6 ZERO METER

Press the  button. The cursor will move across the display followed by **0.00 PPM**. This will indicate that the sample is ready for testing.

7 DIP STRIP & PRESS READ

Dip the **CD (DPD-1)** strip into the **CELL**, and immediately press . This starts a 20 second countdown timer. During this time, move the strip in a gentle back and forth motion (approx. 2 strokes/sec). Remove and discard the strip after '1' on the display disappears. The cursor will move across the display, while the meter measures the sample. Record the result displayed (this result is automatically stored in Cd MENU and, if using the eXact iDip® app, will be stored in 'History'). After testing is completed, rinse **CELL** immediately and clean with the brush.

Manganese test procedure (combination strip and drop test)

1 REMOVE STRIPS

Remove 1 **eXact® Strip Micro Mn#1** Part No. 481020-1 and **eXact® Strip Micro Mn#2** Part No. 481020-2 strips from their foil packets before beginning the test. Also shake the bottle of **eXact® Reagent MN** and remove the cap. Set the strip in a dry, convenient place and recap the bottle immediately.

2 TURN METER ON

Press the  button to power the meter on; the display will show annunciators, followed by the current selection. It will then display the last reading.

3 SELECT GROUP & MENU

Press and re-press the  button to **Select Group 1 or 6**. Press and re-press the  button to select the test parameter **MN** (see chart on page 8-9).

4 RINSE & FILL CELL WITH SAMPLE

Rinse the **CELL** at least 3 times with the water sample you will be testing—rinsing minimizes the potential for cross-contamination from a previous test. Finally, fill **CELL** to capacity (4 mL) with the water sample.



If using **eXact® Micro 20 with Bluetooth® SMART** follow the steps to **‘SELECT CUSTOMER’ & ‘CONNECT DEVICE VIA BLUETOOTH’** before proceeding (see page 6).

5 DIP STRIP & PRESS READ

Dip the **Mn#1** strip into the **CELL**, and immediately press . This starts a 20 second countdown timer. During this time, move the strip in a gentle back and forth motion (approx. 2 strokes/sec). Remove and discard the strip after ‘1’ on the display disappears. The cursor will move across the display, informing you prepare to dip the **Mn#2** strip. When the next 20 second countdown starts, dip the **Mn#2** strip immediately the into the **CELL** using the same gentle back and forth motion (approx. 2 strokes/sec). Remove and discard the strip after ‘1’ on the display disappears. The meter will automatically start to count up to 20 seconds. After 20 seconds, the cursor will move across the display and the display will automatically zero.

6 ADD DROPS

Add 3 drops of **eXact® Reagent MN** to the **CELL** (Precaution: Ensure the bottle is straight while dispensing drops) and cover with the **CELL COVER**. When the 20 second countdown starts, place thumb over the cover and mix the sample by turning the meter upside-down repetitively during the countdown. When timer displays ‘1’, place the meter upright and the cursor will flash. The meter will begin a 120 second count up. After 120 seconds, the cursor will move across the display while the meter measures the sample. Record the result displayed (this result is automatically stored in MN MENU and, if using the **eXact iDip®** app, will be stored in ‘History’). After testing is completed, rinse **CELL** immediately and clean with the brush.

Nitrate (salt water > 4000 ppm) test procedure

- 1

REMOVE STRIPS

Remove 1 **eXact® Strip Micro NO₃** Part No. 486655 from the bottle before beginning the test. Set the strip in a dry, convenient place and recap the bottle immediately.
- 2

TURN METER ON

Press the  button to power the meter on; the display will show annunciators, followed by the current selection. It will then display the last reading.
- 3

SELECT GROUP & MENU

Press and re-press the  button to **Select Group 7**. Press and re-press the  button to select the test parameter TR1 (see chart on page 8–9).
- 4

RINSE & FILL CELL WITH SAMPLE

Rinse the **CELL** at least 3 times with the water sample you will be testing—rinsing minimizes the potential for cross-contamination from a previous test. Finally, fill **CELL** to capacity (4 mL) with the water sample.



If using eXact® Micro 20 with Bluetooth® SMART follow the steps to `SELECT CUSTOMER` & `CONNECT DEVICE VIA BLUETOOTH` before proceeding (see page 6).

- 5

ZERO METER

Press the  button. The cursor will move across the display followed by **100 %T**. This will indicate that the sample is ready for testing.
- 6

DIP STRIP & PRESS READ

Dip the **NO₃** strip into the **CELL**, and immediately press . This starts the 20 second countdown timer. Because the strip is 8 mm wide, the strip will need to be angled to fit in the cell. Be sure that the test pad is fully submerged. During this time, move the strip in a gentle back and forth motion (approx. 2 strokes/sec). Remove and discard the strip after `1` on the display disappears. Time the reaction in the **CELL** for 580 seconds (no timer provided). During this time, the meter will shut off. When 580 seconds have elapsed, turn meter on and wait for the display to show the last reading. Then press  to initiate a 20 second countdown.
- 7

RECORD RESULT

The cursor will move across the display while the meter measures the sample. Record the value displayed. This value is automatically stored in **TR1**, and if using the eXact iDip® app, the result will be stored in the app's **`HISTORY`**.
- 8

USE TABLE

Find the **TR1** result in the table below to determine the **Total Hardness** concentration in ppm.
***EXAMPLE: A TR1 result of 65.3 (round#) equals a Nitrate value of 23 ppm).**

eXact® Strip Micro Nitrate NO ₃ — for 4 mL samples										
%T	9	8	7	6	5	4	3	2	1	0
90	0	0	0	0	0	0	0	2	3	4
80	5	5	6	7	7	8	9	10	11	12
70	12	13	14	14	15	16	17	17	18	19
60	20	20	21	22	*23	24	25	26	27	27
50	28	29	30	31	32	33	33	34	35	36
40	37	38	39	40	41	42	43	44	45	46
30	47	48	49	50	51	52	53	54	55	56
20	57	59	60	61	62	64	65	66	67	68
10	70	71	72	74	75	76	78	79	80	82
83	83	85	86	87	89	90	>90	>90	>90	>90

Tips for Best Accuracy

1. Our lab testing with the Micro 20 meter has shown that zeroing and measuring of the sample normally does not require any cell cover for accurate results, except in sunlight. To obtain optimal accuracy when testing with the meter outdoors (sunlight), use the Mixing Cap/Cell Cover when zeroing and reading the sample.
2. Become familiar with the meter and the different tests by reading the instructions carefully.
3. The Free Chlorine, Combined Chlorine, and Total Chlorine reagents are compliant for meeting USEPA (4500-Cl G); ISO 7393/2; and German DIN 38408 G4-2 requirements.
4. Observe the dip time (as required for the test) for accurate results.
5. Test immediately after filling the CELL with water sample when testing for oxidizers such as Chlorine and Bromine (Ozone can be measured in CL3 MENU).
6. Be sure the CELL is filled to capacity (4 mL), especially for pH and Total Alkalinity.
7. Rinse the CELL with clean water immediately after completing each test. Some reagents may stain the CELL if not rinsed shortly after use. Other reagents including Cyanuric Acid, Chloride, and Calcium Hardness may coat the CELL wall. It is recommended, after these tests, to use the Cell Cleaning Brush with water to clean the CELL.
8. Just before testing, rinse the sample CELL with the sample water several times to get a representative sample. (Use deionized or distilled water for rinsing if you have a limited amount of sample).
9. Store the meter and all test materials out of direct sunlight and away from chemical storage areas.
10. Minimize exposure of meter and test reagents to heat above 100°F (38°C).
11. Dry the outside of the meter when testing is complete or before storage of the meter.
12. When running a DPD-1 Free Chlorine test AFTER a Total Chlorine DPD-3, a Total Chlorine DPD-4, or a HR Chlorine test, rinsing is very important to remove residual KI, which may interfere.
13. Each eXact® Strip Micro is valid for ONLY one test. Discard strip after single use in regular refuse that is inaccessible to children and pets.
14. Each bottle of eXact® Strip Micro contains the quantity of strips notated on the bottle. Due to the strip slitting process, you may find one or two strips that are noticeably smaller or larger in width than the normal strips in the bottle. These should be discarded. Using these strips may give unreliable results.
15. Each table supplied has a unique revision number located in the bottom right corner of the table. We recommended that you visit www.sensafe.com regularly for any updated revisions.
16. The eXact® Micro 20 Meter is not compatible for use with DPD-1, DPD-3, and DPD-4 powder pillows, tablets, and liquids available from other manufacturers. Accurate results can only be guaranteed by using genuine eXact® Micro strips or reagents (reorder information on page 19).
17. Remove batteries when meter is not used for more than a month (Warranty Requirement).
18. It is recommended that Pool and Spa samples for oxidizers (such as Chlorine) be taken 18 inches below the surface as follows: submerge meter with open cell facing down 18 inches, and then turn meter upright at that depth to fill the cell. Remove meter from water with the sample for testing.

About the Accuracy/Calibration of the eXact Micro 20 System

All tests have been calibrated using certified reference standards and standard analytical spectrophotometric methods. The algorithms in the software reflect the best correlation of the eXact® Micro 20 Systems against the AWWA, US EPA , DIN, and ISO reference test methods for chlorine. Studies show that the eXact® Micro 20 System repeatedly agrees with an EPA Compliant reference method greater than 99% (R2= 0.99948, 0 – 5.00 ppm – see below). The eXact® Micro 20 Advanced Photometric System has been factory calibrated for your convenience. You can expect the fixed calibrations in the meter to be valid for the life of the meter because of the quality, Long-Life LED, the photo **cell**, and the software as written into the meter. This is why the meter comes with a 2-Year Warranty.

Assigned Value for Ready Snap™ Solution				
Ready Snap™ Lot	Desired Value (%T)	Acceptable Value (%T)	Select Group	Menu
Red Dye #505	17.1	16.5 – 17.5	7	TR1
Blue Dye #506	30.1	28.0 – 32.8	7	TR2

NOTE: Values reflect current concentrations as found at time of manufacture and may change with consecutive lots.


eXact® Strip Micro DPD–1 Accuracy

Free Chlorine results are compared using the eXact® Strip Micro CL (DPD–1) with the eXact® Micro 20 Meter in MENU CL and Hach® DR890 Colorimeter in Program 9 and Program 12 using Hach® powder pillows.

DR890	Micro 20
0.00	0
0.27	0.27
0.58	0.57
1.10	1.10
1.64	1.62
2.16	2.21
2.8	2.73
3.6	3.53
4.2	4.14
4.6	4.46

Meter	Menu	Range (PPM)	Resolution
Micro 20	CL	0 to 5.00	0.01
DR890	Program 9	0.00 to 2.20	0.01
	Program 12	0.0 to 11.0	0.1

Hach® is a registered trademark of Danaher Corporation



Accessories / Kits

eXact Micro Carrying Case with Foam (486001)

Dilution Kit (487200)

Ready Snap™ 3 (486903)

Pool Kit (486700-KP)

Kit includes:

- 1 eXact® Micro 20 Meter (486700)
- eXact® Strip Micro DPD-1 (486637-25)
- eXact® Strip Micro DPD-3 (486638-25)
- eXact® Strip Micro pH (486639-25)
- eXact® Strip Micro Total Alkalinity (486641-25)
- eXact® Strip Micro Copper (486632-25)
- eXact® Strip Micro Nitrate (486655-25)
- eXact® Strip Micro Total Iron, TPTZ (486650-25)
- eXact® Strip Micro Calcium Hardness (486629-25)
- eXact® Strip Micro Phosphate (486814-25)
- eXact® Strip Micro Chloride (486757)
- eXact® Reagent Cyanuric Acid (481652-II)
- eXact® Strip Micro Biguanide (486810-25)
- Mini Dilution Kit II (487202)
- 1 Mixing Cap
- 1 Cell Cleaning Brush
- Instruction Booklet
- Plastic Carrying Case
- Plastic Stirrer



Part No. 486700-KP

Standard Kit (486700-K)

Kit includes:

- 1 eXact® Micro 20 Meter (486700)
- eXact® Strip Micro DPD-1 (486637-25)
- eXact® Strip Micro DPD-3 (486638-25)
- Mini Dilution Kit II (487202)
- 1 Mixing Cap
- 1 Cell Cleaning Brush
- Instruction Booklet
- Plastic Carrying Case
- Plastic Stirrer

Well Driller Kit (486700-WD)

Kit includes:

- 1 eXact® Micro 20 Meter (486700)
- eXact® Strip Micro DPD-1 (486637-25)
- eXact® Strip Micro DPD-3 (486638-25)
- eXact® Strip Micro pH (486639-25)
- eXact® Strip Micro Total Alkalinity (486641-25)
- eXact® Strip Micro Copper (486632-25)
- eXact® Strip Micro Nitrate (486655-25)
- eXact® Strip Micro Manganese (486606)
- eXact® Strip Micro Total Hardness, HR (486656-25)
- eXact® Strip Micro High Range Chlorine (486672-25)
- eXact® Strip Micro Total Iron, TPTZ (486650-25)
- Mini Dilution Kit II (487202)
- 1 Mixing Cap
- 1 Cell Cleaning Brush
- Instruction Booklet
- Plastic Carrying Case
- Plastic Stirrer

Mini Dilution Kit II (Part No. 487202) Instructions

HOW TO PREPARE A 1:20 SAMPLE USING THE 3 mL SYRINGE (DILUTION FACTOR OF 20)

1. Rinse the syringe 3 times with water sample that you want to test by moving the plunger up and down.
2. Rinse 50 mL graduated conical tube with distilled or deionized (salt-free) water.
3. Fill the 3 mL syringe to the 2 mL line by pulling up water sample to be tested with an upward motion of the plunger until you get to the 2 mL line. **NOTE:** The plunger ring should line up at the 2 mL line.
4. After adding sample to the cylinder, fill the graduated cylinder to the 40 mL line with distilled or deionized (salt-free) water. Securely put the cap on the cylinder.
5. Mix content of graduated conical tube by turning upside down at least 3 times.

Other dilutions possible with the 3 mL syringe

Volume in syringe	Volume filled in cylinder	Dilution factor
1.0 mL	40 mL	40
1.0 mL	30 mL	30
1.0 mL	20 mL	20
0.5 mL	25 mL	50
0.5 mL	50 mL	100
0.5 mL	50 mL	250

CALCULATION:
Test Result x Dilution Factor =
Actual Result

Available reagents / Reorder information

 = EPA COMPLIANT

PARAMETER / TEST	PART #	RANGE (ppm)	BEST ACCURACY †	# OF TESTS
Alkalinity, Total	486641	8 – 200	7.5	100
Aluminum	486821	0.01 – 1.2	13	50
Ammonia	486654	0.02 – 2.4	5	25
Biguanide	486810	1.6 – 210	7.5	50
Bromine (DPD-1)	486637	0.01 – 12	5	100
Calcium (as CaCO ₃)	486629	16 – 400	10	50
Chloride (as NaCl) III	486757	1 – 330	5	25
Chloride (as NaCl) III HR	486757	26 – 6600	8	25
Chlorine Dioxide (DPD-1)	486633	0.01 – 10	5	100
Chlorine, Free (DPD-1)	486637	0.01 – 6.2	3	100
Chlorine, High Range	486672	1 – 300	5	100
Chlorine, Combined (DPD-3)** (525 nm)	486638	0.01 – 6.2	3	50
Chlorine, Total (DPD-4)	486670	0.01 – 6.2	3	50
Chromium (VI)	486614	0.01 – 1.8	5	50
Copper (Cu ⁺²)	486632	0.01 – 11	2	50
Cyanide	486812	0.01 – 1.1	13	50
Cyanuric Acid II	481652 – II	7 – 110	8	60
Fluoride	486643	0.04 – 1.5	15	50
Glycine (used for Chlorine Dioxide)	484014	N/A	N/A	50
Hardness, Total HR (as CaCO ₃) (525 nm)	486656	60 – 600	12	100
Hardness, Total LR (as CaCO ₃) (525 nm)	486630	1 – 80	10	50
Hydrogen Peroxide	486648	0.3 – 100	8	100
Iron, Total (TPTZ)	486650	0.03 – 6	3	50
Manganese	486606	0.01 – 1.5	6	24
Molybdate	486653 – A	0.01 – 3	5	50
Nitrate (as NO ₃)	486655	0.1 – 30	15	50
Nitrate (as NO ₃ for Saltwater) (salt>400ppm)	486655	0 – 90	15	50
Nitrite (as NO ₂)	486623	0.01 – 1.8	5	50
Ozone (DPD-4)	486634	0.01 – 2	4	100
Peracetic Acid (DPD-4)	486674	0.01 – 6	3	100
Permanganate (DPD-1)	486626	0.01 – 5	2	100
pH	486639	5.8 – 8.5 pH	0.2 pH	100
Acid PH	486624	3.2 – 6 pH	0.3 pH	50
Alkali PH	486609	7.2 – 9.8 pH	0.3 pH	50
Phosphate	486814	0.03 – 4.4	4	50
Quaternary Ammonia	486823	2 – 80	6	50
Sulfate (as SO ₄)	486608	2 – 210	10	50
Sulfide (as S ²⁻)	486818	0.01 – 1.6	6	50
Turbidity	N/A	4 – 900 NTU	N/A	N/A
Sulfate (as SO ₄)	486608	2 – 210	10	50
Sulfide (as S ²⁻)	486818	0.01 – 1.6	6	50
Turbidity	N/A	4 – 800 ntu	N/A	N/A

† Value provided represents best possible accuracy under laboratory conditions, but may vary throughout the detection range. For a complete list of accuracies throughout all ranges please visit sensafe.com/micro20/specifications.

Because most of our products are test strips or use reagents that have little or no hazard in the quantity sold, MSDS sheets are not supplied with the test. The exceptions are the Manganese (486606) test, which comes with 2 strips and one liquid reagent (PAN); Fluoride (486643) test, which is a liquid reagent (SPADNS); and Iron (486650) test, which is a powder reagent.



Industrial Test Systems, Inc.

Innovators of Water Quality Testing

Visit us on-line at **sensafe.com/micro20**
for up-to-date product information and NEW tests available



SENSAFE



ITSSENSEAFE

MANUFACTURED BY

Revision 11.10.2014

Industrial Test Systems, Inc.

1875 Langston Street, Rock Hill, SC 29730

1-800-861-9712 — *Inside the U.S.A.*

1-803-329-9712 — *Outside the U.S.A.*

Fax: 1-803-329-9743

Email: ITS@SENSAFE.COM

WWW.SENSAFE.COM

ITS EUROPE, LTD

The UK Centre for Homeland Security

Building 7, Chilmark

Salisbury, Wiltshire, SP3 5DU, United Kingdom

Phone: +44 (0)1722 717911 Fax: +44 (0) 1722 717941

Email: ITSEUROPE@SENSAFE.COM

WWW.ITSEUROPE.CO.UK

FEDERAL COMMUNICATIONS COMMISSION INTERFERENCE STATEMENT

This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to Part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.

CAUTION:

Any changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate the equipment.

This device complies with part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) This device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.