

Declaration statement Letter

2015-03-10

MET Laboratories, Inc.
914 West Patapsco Avenue
Baltimore, MD 21230

RE: Declaration statement Letter

To Whom It May Concern:

We hereby declare that (Sleep wellness and breathing monitor/ STS100) fulfill the requirements of 47 CFR Part 15C Section 15.247(g) & 15.247(h).

Compliance for section 15.247(g)
According to Bluetooth Core Specification, the Bluetooth system transmits the packet with the pseudorandom hopping frequency with a continuous data and the short burst transmission from the Bluetooth system is also transmitted under the frequency hopping system with the pseudorandom hopping frequency system.
Compliance for section 15.247(h)
According to Bluetooth Core specification, the Bluetooth system incorporates with an adaptive system to detect other user within the spectrum band so that it individually and independently to avoid hopping on the occupied channels. According to the Bluetooth Core specification, the Bluetooth system is designed not have the ability to coordinated with other FHSS System in an effort to avoid the simultaneous occupancy of individual hopping frequencies by multiple transmitter.

Please contact me if there is any information you may need.

Sincerely,



Name: Frank Chang
Position: Sales Director
Company Name: Ospicon Technology Limited
Address: Unit 1025, 10 Floor, Metro Centre II, 21 Lam Hing Street, Kowloon
Bay Hong Kong
E-Mail: frank.chang@ospicon.com