

Medisign *MediKeto Plus 82 BT*
blood glucose and β -ketone monitoring system
User Manual

Introduction

Thank you for choosing the Medisign **MediKeto Plus 82 BT** blood glucose and β -ketone monitoring system to help monitor your blood glucose and blood β -ketone levels. The Medisign **MediKeto Plus 82 BT** blood glucose and β -ketone monitoring system is designed to be accurate and easy to use. This user manual contains all the information needed to use and maintain the Medisign **MediKeto Plus 82 BT** blood glucose and β -ketone monitoring system. Please read this User Manual carefully before you start the test and do a quality control test. If you have question about the control test, please consult with your healthcare professional. Please confirm the measuring unit on the display is correct with every test result.

If you need assistance, please contact Customer Service. If you have questions or need assistance outside the days and hours of operation, please contact your health care professional.

IMPORTANT SAFETY INFORMATION

- Meters that come into contact with blood present a potential risk of transmitting infectious diseases. To prevent infections and/or cross-contamination, maintain good hand hygiene and clean your Medisign **MediKeto Plus 82 BT** blood glucose and β -ketone meter as well as the corresponding accessories regularly.
- Healthcare Professionals performing the tests with this system on multiple patients must always wear gloves and should follow the infection control policies and procedures approved by their facility.
- The meter and lancing device are for single patient use only. Do **NOT** share them with anyone including family members.
- **NEVER** use lancets that has been used by someone else.
- All parts of the kit are considered biohazardous and can potentially transmit infectious diseases, even after you have performed cleaning.
- Do **NOT** leave the lancets in the lancing device after use, be sure to dispose of it according to instructions from your healthcare professional. Do **NOT** re-use a lancet.
- If you have taken the cap off of the lancet and the lancet has fallen to the ground, do **NOT** use it. Make sure you dispose of it and replace it with a new lancet.

CAUTION

Keep the test strip, meter, lancing device, lancets and control solution away from all children and pets. Batteries and test strips may be a choking hazard.

Sign of Trouble

If you encounter any issues including the ones mentioned below, see the trouble shooting. If problems persist, do **NOT** use and contact Customer Service.

- The meter does **NOT** power on when the new test strip is inserted into the meter.
- The meter does **NOT** power on when you replace the batteries.
- The meter does **NOT** show the test result after the test is completed.
- The LCD does **NOT** display correctly or is distorted after changing the batteries.

Intended Use

The Medisign **MediKeto Plus 82 BT** blood glucose and β -ketone monitoring system is intended to be used for the quantitative measurement of glucose and β -ketone (beta-hydroxybutyrate) in fresh capillary whole blood samples from the fingertips. Alternative site testing (palm and forearm) should be done only for glucose measurement and during a steady-state times. Venous whole blood testing is limited to healthcare professionals use only. It is intended for self-testing outside of the body (in vitro diagnostic use only) by people with diabetes at home and healthcare professionals in a clinical setting as an aid to monitor the diabetes control. This system is **NOT** for use in diagnosis of or screening for diabetes, or for neonatal use.

Test Principle

Glucose in the blood sample reacts with glucose dehydrogenase (FAD-GDH) on the test strip and a harmless current is produced. The strength of these currents change with the amount of glucose in the blood sample and the Medisign **MediKeto Plus 82 BT** blood glucose and β -ketone meter automatically interprets this reaction. Glucose measurements are reported as plasma equivalents. Results that use a plasma equivalent method are approximately 11% higher than those obtained with whole-blood equivalent test strips. This system is calibrated with capillary blood samples with various glucose concentrations. The reference method is traceable by using the NIST Standard Reference Materials (SRM). The method comparison was conducted by comparing the result from the system of Glucose Oxidase method using an automatic analyzer.

β -Ketone in the blood sample reacts with beta-hydroxybutyrate dehydrogenase on the test strip respectively. This reaction produces a harmless current and the Medisign **MediKeto Plus 82 BT** blood glucose and β -ketone meter measures the current, calculates the β -ketone level and displays the result. The strength of the current is dependent on the amount of β -ketone in the blood sample.

Important Information for the System

- Before using the Medisign **MediKeto Plus 82 BT** blood glucose and β -ketone monitoring system, please read this manual carefully.
- Use this system only for blood glucose and blood β -ketone test.
- The Medisign **MediKeto Plus 82 BT** blood glucose and β -ketone meter should be used with the Medisign **GH82** blood glucose test strips and the **MediKeto 82** blood β -ketone test strips.
- If results from alternative site testing do **NOT** match how you feel, perform a fingertip test to confirm the result. If the fingertip test result still does **NOT** match how you feel, call your physician or healthcare professional.
- If your test result does **NOT** match how you feel, call your physician or healthcare professional.
- **NEVER** ignore symptoms of high or low blood glucose and high blood β -ketone.
- If you see “HI” or “Lo” displayed, your blood glucose level may be above 600mg/dL or below 20mg/dL, repeat the blood glucose test. If the meter persists same results, contact your physician or healthcare professional.
- If you see “HI” displayed with **KETONE** symbol, your blood β -ketone level may be above 8.0 mmol/L, repeat the blood β -ketone test using a new test strip. If the meter persists same results, contact your physician or healthcare professional.
- This system could support your healthcare program but **NEVER** make major changes in your diabetes treatment program without consulting your physician.
- The results of this system should **NOT** be used for diabetic treatment or medications without consulting your doctor.
- Be sure to check the blood glucose and blood β -ketone measurement unit on the meter before the test. Use of the wrong blood glucose and blood β -ketone measurement unit may cause you to misinterpret the test results and result in incorrect treatment. Consult your healthcare professional if you have any questions about your blood glucose and blood β -ketone measurement units.
- Before the test you must make sure that the code number displayed on the meter matches the code number on the test strip vial or foil packet.
- If alcohol wipes are used to disinfect the fingertip, make sure the fingertip is completely dry before obtaining the blood sample.
- Before sampling blood, gently massage the hand and finger toward the puncture site to form a drop of blood. Do **NOT** squeeze excessively around the puncture site when you produce a drop of blood.
- If the sample drop of blood runs or spreads due to contact with hair or a crease in the palm, do **NOT** use that sample. Try puncturing again in a smoother area.
- Do **NOT** hold or shake the meter while it is testing.

Important Information for the Test Strip

- **NEVER** reuse a test strip that is either applied with blood or control solution. The test strips are for single use only.
- Before using, please check the test strip expiration date on the vial or the foil packet.
- Do **NOT** use test strips that have expired as it may cause inaccurate test results.
- Do **NOT** touch sampling end of the test strip.
- Do **NOT** smear the blood with sampling end of the test strip.
- Do **NOT** drop the blood directly on the sampling end of the test strip.
- Do **NOT** insert the test strip with force into the meter.
- Do **NOT** use wet, bent, scratched or visibly damaged test strips.
- You should use your test strip immediately after retrieving it from its vial or foil packet.
- Always keep the test strip vial closed immediately after retrieving the test strip.
- Do **NOT** use the test strip if the foil packet has a puncture or tear in it.
- For accurate results, fresh capillary whole blood and venous whole blood should be used. Arterial blood, plasma and serum are **NOT** suitable blood sample for the test.

Important Information for the Control Solution

- Use only the Medisign GH Glucose Control Solution and the Medisign Ketone Control Solution. Other brands of control solutions will produce inaccurate results.
- Use the Medisign GH Glucose Control Solution and the Medisign Ketone Control Solution only with their corresponding test strips.
- Shake the control solution well before using.
- Always check the expiration date of the control solution. Do **NOT** use expired control solution.
- Do **NOT** apply control solution to the test strip directly from the vial.
- Do **NOT** add or apply a second drop of blood or control solution. This may cause a false result.
- Replace the cap of the control solution immediately after use.
- If the control test result falls outside of the range provided, it indicates that the system may **NOT** be functioning properly. Do **NOT** use the meter and please contact Customer Service.
- Control solution is **NOT** for human consumption or injection. Keep away from eyes.
- The range printed on the test strip vial or box is for the control solution only. It has nothing to do with your blood glucose or blood β -ketone levels.
- The results obtained from control testing do **NOT** reflect your personal blood glucose or blood β -ketone levels in any way.

Limitation

- This system is only for in vitro diagnostics use and for self-testing.
- This system is **NOT** designed for use with arterial, serum or plasma samples. Do **NOT** put urine on the test strip.
- This system is **NOT** designed for use with neonatal.
- This system is **NOT** for screening or diagnosis of diabetes mellitus.
- Blood β -ketone testing must **NOT** use the forearm or palm site. Only use the fingertip for blood β -ketone testing.
- Hematocrit is the percentage of red blood cells in the blood. Testing outside of the Hematocrit level range of 20~60% may cause inaccurate result. If you do **NOT** know your hematocrit level, consult with your healthcare professional.
- Inaccurate results may occur if used at altitudes above 3,500m.
- This system is **NOT** for use on patients who are in shock, hyperosmolar state or dehydrated patients.
- Inaccurate results may occur in severely hypotensive individuals.
- Inaccurate low results may occur for individuals experiencing a hyperglycemic-hyperosmolar state, with or without ketosis.
- This system is **NOT** for use on critically ill patients.
- Xylose: Do **NOT** use during or soon after xylose absorption testing. Xylose in the blood will cause an interference.
- If you are taking acetaminophen containing drugs (e.g. Tylenol, certain cold and flu remedies etc.) in excess of the recommended levels (>20 mg/dL) then you may get inaccurate blood glucose results with this system. If you are unsure, then ask your doctor.
- If you are taking vitamin C (ascorbic acid) in excess of the recommended levels >3 mg/dL, then your blood glucose results may be inaccurate with this system. If you are unsure, then ask your doctor.
- If you have certain conditions that may cause your blood level of uric acid to rise (>10 mg/dL), such as gout or kidney disease then your blood glucose results may be inaccurate. If you are unsure, then ask your healthcare professional.
- Alternative site testing should be performed only during steady-state (when glucose is **NOT** changing rapidly).
- Alternative site measurements should **NEVER** be used to calibrate continuous glucose monitoring (CGMs).
- Alternative site measurements should **NEVER** be used for insulin dosing calculations.

Storage and Handling of the System

- Keep this system dry and keep away from the heat and direct sunlight.
- Avoid leaving the meter or test strips in extremely hot or cold places, such as near a heat source or in an extremely hot or cold car.
- Do **NOT** store or use the meter or test strips at high humidity levels, such as in the bathroom or kitchen.
- Store the test strips in the original vial between 4 ~ 30°C, at indoor area. Do **NOT** freeze.
- The test strip may be stored in a refrigerator between 4 ~ 8°C. But it must be brought to room temperature before using.
- When properly stored, unopened test strips are stable until the expiration date printed on the vial or the foil packet.
- The Medisign **GH82** test strips in the vial are good six (6) months after first opening date or until the expiration date, whichever comes first.
- The Medisign **MediKeto 82** test strips in the vial are good three (3) months after first opening date or until the expiration date, whichever comes first.
- Record the discard date on the test strip vial and discard any unused test strip after that date.
- The meter and the test strips are designed to be used within a temperature range of 10 ~ 40°C. Please keep the system operating temperature at a range between 10 ~ 40°C for more than twenty (20) minutes before testing.
- Do **NOT** test if there is condensation (water build-up) on your meter and the test strip vial or the foil packet. Move your meter or a test strip vial or a foil packet to a cool, dry spot. Wait for the meter and a test strip vial or a foil packet surface to dry before testing.
- The meter, test strips, and control solution should be handled at the same temperature.
- Do **NOT** handle the test strips with wet or dirty hands.
- For accurate results, make sure your hands are clean and dry before retrieving test strip from its foil packet.
- Do **NOT** allow dirt or dust in the test strip slot in order to avoid affecting the system's performance.
- Dispose of used test strips, lancets according to instructions from your healthcare professional.
- Dispose of used batteries according to local regulations.
- Do **NOT** destroy, repair or transform the system at any time. If the equipment is used in a manner **NOT** specified by the manufacturer, the protection provided may be impaired.
- Handle the meter with care – severe shock, such as dropping the meter, could damage the electronics.
- The Applicable Pollution Degree of the intended environment for this system is Pollution Degree 2.

Storage and Handling of the Control Solution

- The control solution should be stored at indoor area between 4 ~ 30°C. Do **NOT** freeze.
- The control solution may be stored in a refrigerator between 4 ~ 8°C. But it must be brought to room temperature before using.
- The control solution is good three (3) months after first opening date or until the expiration date, whichever comes first.
- Record the discard date on the control solution vial and discard remained control solution after that date.
- The control solution test is recommended to be done at room temperature (20 ~ 25°C).

Important Information for Caring the Meter

- Clean your Medisign **MediKeto Plus 82 BT** blood glucose and β -ketone meter once per week at a minimum.
- Do **NOT** get water inside the meter. **NEVER** immerse the meter or hold it under running water or spray any cleaning solution directly onto the meter.
- Do **NOT** try to get moisture on the test strip, test strip slot or serial port.
- Do **NOT** use glass cleaners or household cleaners or corrosive liquid (e.g. Benzene, Acetone) on the meter.
- Please use a soft cotton cloth to wipe the system.
- Squeeze the excess liquid from the cloth before you wipe the meter.
- The meter must be turned off when you clean the meter.

Table of Contents

1. About the Medisign <i>MediKeto Plus 82 BT</i> blood glucose and β -ketone monitoring system	10
■ Contents of Kit.....	10
■ Medisign <i>MediKeto Plus 82 BT</i> Blood Glucose and β -Ketone Meter	10
■ Meter Display Screen.....	11
■ Medisign <i>GH82</i> Blood Glucose Test Strip.....	12
■ Medisign <i>MediKeto 82</i> Blood β -Ketone Test Strip	12
■ Medisign GH Glucose Control Solution	13
■ Medisign Ketone Control Solution	13
■ Medisign Check Strip	13
2. Settings.....	14
■ Inserting (or Replacing) the Batteries	14
■ Set Date and Time.....	14
■ Acoustic Signal	15
■ Delete Test Results.....	15
■ Exit Settings.....	15
3. Control Solution Testing.....	16
■ Performing a Control Solution Test	17
■ Performing the Check Strip Test	18
4. Testing Your Blood Glucose	19
■ Preparing the Test	19
■ Inserting the Medisign <i>GH82</i> Blood Glucose Test Strip.....	19
■ Obtaining a Blood Drop.....	20
■ Applying Blood.....	21
■ Measuring Process	21
■ Flagging Test Results	21
■ Disposing Used Test Strips and Lancets	22
5. Understanding Your Blood Glucose Test Results	22
■ Blood Glucose Result with Smile Symbol.....	23
■ Ketone Alert.....	23
6. Testing Your Blood Glucose from the Palm and Forearm	23
■ Alternative Site Blood Sampling	24
7. Testing Your Blood β -Ketone	25

■ Preparing the Test	25
■ Inserting the Medisign <i>MediKeto 82</i> Blood β -Ketone Test Strip	25
■ Obtaining a Blood Drop	26
■ Applying Blood	26
■ Measuring Process	27
■ Disposing Used Test Strips and Lancets	27
8. Understanding Your Blood β -Ketone Test Results	28
■ Blood β -Ketone Results Interpretation	28
9. Displaying Stored Test Results	29
■ Display of Test Results	29
10. Downloading Your Test Results to a PC	30
11. Pairing Your Meter	30
■ Before Pairing Your Meter	30
■ How to Pair Your Meter	30
■ About Bluetooth	30
12. Cleaning Your Meter	31
13. Screen Messages and Troubleshooting	32
■ Screen Messages	32
■ Troubleshooting	33
14. System Specifications	33
15. Symbols	34
16. Disposing of the Meter	34
17. Additional Supplies	34
18. Product Warranty	34

1. About the Medisign *MediKeto Plus 82 BT* blood glucose and β -ketone monitoring system

■ Contents of Kit

The Medisign *MediKeto Plus 82 BT* blood glucose and β -ketone monitoring system includes the following items:

- Medisign *MediKeto Plus 82 BT* blood glucose and β -ketone meter
- Medisign *GH82* blood glucose test strips (10 Tests)
- Medisign *MediKeto 82* blood β -ketone test strips (10 Tests)
- Lancing Device
- Lancets
- Battery (CR2032)
- Carrying Bag
- User Manual


NOTE:

- The Medisign *GH82* blood glucose test strips, the *MediKeto 82* blood β -ketone test strips (for 25 Tests or 50 Tests), the *Check Strip*, the *GH Glucose Control Solution* and the *Ketone Control Solution* are required but **NOT** included and must be purchased separately. For additional information of other accessories, please contact your supplier, pharmacist, healthcare professional or Customer Service.
- The Test Strip may be provided with a foil packet for each or a vial for 10, 25 or 50 test strips.

■ Medisign *MediKeto Plus 82 BT* Blood Glucose and β -Ketone Meter



Meter Display Screen




1	KETONE	Appears when the <i>Medisign MediKeto 82</i> blood β -ketone test strip is inserted
2	KETONE?	Appears when your blood glucose result is equal to or higher than 240 mg/dL
3		Symbol to mark pre-meal or post-meal test results.
4		Symbol to mark the control solution test results
5		Appears when the meter pairs with a mobile device
6		Appears when the sound is set to OFF
7		Appears if the meter is out of operating temperature range
8		Appears when blood glucose test result falls in 80-120 mg/dL
9	SET	Segments to display the Set-up mode
10		Appears when the meter is ready for application of blood or control solution.
11		Appears when test strip is inserted. Blinks when test strip is removed during the meter is on.
12		Appears when test results stored in memory are displayed
13	88-88	Segments to display date and average day
14	DAY AVG	Appears when average result is displayed
15	88:88	Segments to display hour and minute of tests
16	mmgl/dL	Segments to display the unit of measurement
17		Appears when the meter detects low battery power
18	888	Segments to display test results or error messages


IMPORTANT

Be sure to check the blood glucose and blood β -ketone measurement unit on the meter before the test. Use of the wrong blood glucose and blood β -ketone measurement unit may cause you to misinterpret the test results and result in incorrect treatment. Consult your healthcare professional if you have any questions about your blood glucose and blood β -ketone measurement units.

Medisign GH82 Blood Glucose Test Strip



**Medisign GH82
Blood Glucose
Test Strip Vial**




**Medisign GH82
Blood Glucose
Test Strip**


Medisign MediKeto 82 Blood β -Ketone Test Strip



**Medisign MediKeto 82
Blood β -Ketone
Test Strip Vial**



**Medisign MediKeto 82
Blood β -Ketone
Test Strip Foil Packet**



**Medisign MediKeto 82
Blood β -Ketone
Test Strip**

- The expiration date of the Medisign **GH82** blood glucose test strip and the Medisign **MediKeto 82** blood β -ketone test strip can be found on the vial or the foil packet.

IMPORTANT

- The Medisign **MediKeto Plus 82 BT** blood glucose and β -ketone meter should be used with the Medisign **GH82** blood glucose test strips and **MediKeto 82** blood β -ketone test strips.
- Before using, please check the test strip expiration date on the vial or the foil packet.
- The Medisign **GH82** test strips in the vial are good six (6) months after first opening date or until the expiration date, whichever comes first.
- The Medisign **MediKeto 82** test strips in the vial are good three (3) months after first opening date or until the expiration date, whichever comes first.
- Record the discard date on the test strip vial and discard any unused test strip after that date.
- You should use your test strip immediately after retrieving it from its vial or foil packet.
- Always keep the test strip vial closed immediately after retrieving the test strip.
- Store the test strips in the original vial between 4 ~ 30°C, at indoor area.
- Keep the test strip away from the heat and direct sunlight. Do **NOT** freeze.
- The test strip may be stored in a refrigerator between 4 ~ 8°C. But it must be brought to room temperature before using.
- Do **NOT** test if there is condensation (water build-up) on your meter and the test strip vial or the foil packet. Move your meter or a test strip vial or a foil packet to a cool, dry spot. Wait for the meter and a test strip vial or a foil packet surface to dry before testing.
- Do **NOT** handle the test strips with wet or dirty hands.

■ Medisign GH Glucose Control Solution

The Medisign GH Glucose Control Solution is the standard glucose concentration solution required to perform a control test (See Control Solution Testing).

The Medisign GH Glucose Control Solution helps to validate the performance of the system. Run the control test to make sure the meter and test strips are working properly together, to practice testing procedure, and when using a new package of the Medisign **GH82** blood glucose test strips.

■ Medisign Ketone Control Solution

The Medisign Ketone Control Solution is the standard blood β -ketone concentration solution required to perform a control test (See Control Solution Testing).

The Medisign Ketone Control Solution helps to validate the performance of the system. Run the control test to make sure the meter and test strips are working properly together, to practice testing procedure and when using a new package of the Medisign **MediKeto 82** blood β -ketone test strips.

■ Medisign Check Strip

The Medisign GH Check Strip and the Medisign Ketone Check Strip verifies whether the Medisign **MediKeto Plus 82 BT** blood glucose and β -ketone meter is working properly.

2. Settings

■ Inserting (or Replacing) the Batteries

The batteries need to be inserted before using your Medisign **MediKeto Plus 82 BT** blood glucose and β -ketone meter for the first time or when the batteries are low. When the batteries are low the screen will display and blink the  on the bottom of the screen. It will be continuously shown until the meter becomes unusable due to low battery.



1. Open the battery compartment cover on the back of the meter.
2. Insert new batteries, with the (+) side facing up.
3. Replace the compartment cover back onto the meter.

⚠ CAUTION

- Risk of explosion if battery is replaced by an incorrect type.
- Small items such as the battery cover and batteries are choking hazards.

IMPORTANT

- Dispose of used batteries according to local regulations.
- If the meter is **NOT** to be used for a long time, remove the battery to eliminate the risk of battery leakage.

NOTE:


- The meter can also be used with a single battery. If you install two (2) batteries, you can use it for a longer period.
- Whenever replacing the batteries, you need to reset the current date and time. However, test results stored will **NOT** be deleted.

■ Set Date and Time



Press the  button to set the month, date, hours, minutes and year on after another. With the < or > button you can change the flashing value.

■ Acoustic Signal



Acoustic signals (**bEEP**) are activated by default. If you want to deactivate the acoustic signals, press the < or > button to change the status from ON (activated) to OFF (deactivated). Confirm your choice with the **⏻** button.

NOTE: Turning the beep sound on or off do **NOT** affect your blood glucose or blood β -ketone measurements.

■ Delete Test Results



You have the option to delete the entire test results from meter memory. You will see dEL flashing on the screen. Press the < or > button to choose between YES (delete all test results) or NO (keep all test results). Confirm your choice with the **⏻** button.

■ Exit Settings

Press **⏻** button for three (3) seconds to exit settings. No matter which steps the meter is set to, it will store the most recent setting and switches off.

NOTE: You can start testing by inserting a test strip at any step.

3. Control Solution Testing

The purpose of the control solution testing is to validate that the Medisign **MediKeto Plus 82 BT** blood glucose and β -ketone meter is working properly with the test strips. The control solution may **NOT** be included in the kit. Please contact your pharmacist or healthcare professional or Customer Service to purchase the Medisign GH Glucose Control Solution or the Medisign Ketone Control Solution.

You should perform a control solution test when:

- Using the meter for the first time.
- You open a new packing box of test strips.
- You want to check the meter and test strips.
- At least once per week to verify that the meter and test strips are working properly together.
- Your test strips were stored in extreme temperature or humidity where is out of the specified storage conditions.
- The meter was dropped or damaged or gotten wet.
- You suspect the meter and test strips are **NOT** working properly together.
- Your test results do **NOT** agree with how you feel.
- You want to practice your test technique.
- Your test results appear to be abnormally high or low.

IMPORTANT

- Use only the Medisign GH Glucose Control Solution and the Medisign Ketone Control Solution. Other brands of control solutions will produce inaccurate results.
- Use the Medisign GH Glucose Control Solution and the Medisign Ketone Control Solution only with their corresponding test strips.
- Always check the expiration date of the control solution. Do **NOT** use expired control solution.
- The control solution is good three (3) months after first opening date or until the expiration date, whichever comes first.
- Record the discard date on the control solution vial and discard remained control solution after that date.
- The control solution should be stored at indoor area between 4 ~ 30°C. Do **NOT** freeze.
- The control solution may be stored in a refrigerator between 4 ~ 8°C. But it must be brought to room temperature before using.
- Do **NOT** add or apply a second drop of control solution. This may cause a false result.
- Control solution is **NOT** for human consumption or injection. Keep away from eyes.
- Make sure the control solution vial is closed when **NOT** in use.
- If the control test result falls outside of the range provided, it indicates that the system may **NOT** be functioning properly. Do **NOT** use the meter and please contact Customer Service.
- The control solution test is recommended to be done at room temperature (20 ~ 25°C).



Performing a Control Solution Test



For the glucose control solution test



For the β -ketone control solution test




1. Insert the test strip with black lines facing upwards and pointing towards the test strip port. Make sure to firmly and completely insert the test strip into the meter. The meter will turn on automatically and the code number will appear on the screen. Verify that the code number displayed matches with the code number on the vial or the foil packet of the test strips. The code number on the picture above is for the example only. The code can be different for each test strip batch. If the two code numbers do **NOT** match, contact Customer Service.
2. When the blood drop symbol  appears on the screen, shake the control solution vial well. Remove the cap. Squeeze the vial, discard the first drop and wipe off the dispenser tip with a clean tissue to ensure an accurate result. If large bubbles appear at the tip of the vial, wipe the bubbles off before applying a drop of control solution to the test strip. Then squeeze the vial again to receive a drop. Apply this drop of control solution to a clean and non-absorbent base. Close the lid of the control solution vial immediately after use.
3. Bring the test strip gently to the drop of control solution so the sampling end of the test strip touches the control solution. The drop of control solution will be drawn in automatically by the test strip.

IMPORTANT


- Replace the cap of the control solution immediately after use.
- Do **NOT** apply control solution to the test strip directly from the vial.



For the glucose control solution test



For the β -ketone control solution test




4. When the test strip measurement chamber is full, the meter will beep. The meter will count down from six (6) seconds in case of the Medisign **GH82** blood glucose test strip and eight (8) seconds in case of the Medisign **MediKeto 82** blood β -ketone test strip and then show the control test result. Make sure the test result on the screen is between the ranges printed on your test strip box or vial.
5. You can flag the control solution symbol  by pressing the < or > button with the test result on the display and the test strip still in the meter.
6. Eject the test strip into a proper waste container.

NOTE: Mark all control solution tests with the control solution symbol  to distinguish them from blood glucose and blood β -ketone test results in the meter memory. Marked control solution results will **NOT** be included in the average result.

IMPORTANT

- The range printed on the test strip vial or box is for the control solution only. It has nothing to do with your blood glucose or blood β -ketone levels.
- The results obtained from control testing do **NOT** reflect your personal blood glucose or blood β -ketone levels in any way.
- If the control solution test result falls outside of the range provided, do the following:
 - Do **NOT** test your blood glucose or blood β -ketone.
 - Make sure you are using the Medisign GH Glucose Control Solution or the Medisign Ketone Control Solution.
 - Make sure the testing environment is between 10 ~ 40°C.
 - Make sure the control solution and test strips have **NOT** expired.
 - Repeat the test with a new test strip.
 - If the problem persists, contact Customer Service. If you have questions or need assistance outside the operational days and times, please contact your health care professional.

■ Performing the Check Strip Test



for the Medisign GH Check Strip



for the Medisign Ketone Check Strip


The Medisign GH Check Strip and the Medisign Ketone Check Strip verifies whether the meter is working properly. Insert the check strip into the meter. The meter switches on automatically. If the “OK” Message is **NOT** displayed on the screen after inserting the check strip, please contact Customer Service.

4. Testing Your Blood Glucose


IMPORTANT SAFETY INFORMATION

- Meters that come into contact with blood present a potential risk of transmitting infectious diseases. To prevent infections and/or cross-contamination, maintain good hand hygiene and clean your Medisign **MediKeto 82 BT** blood β -ketone meter as well as the corresponding accessories regularly.
- Healthcare Professionals performing the tests with this system on multiple patients must always wear gloves and should follow the infection control policies and procedures approved by their facility.
- The meter and lancing device are for single patient use only. Do **NOT** share them with anyone including family members.
- **NEVER** use lancets that has been used by someone else.
- All parts of the kit are considered biohazardous and can potentially transmit infectious diseases, even after you have performed cleaning.

■ Preparing the Test



1. The easiest way to collect a drop of blood is to wash your hands with warm water and soap first. This promotes the blood circulation in the fingertips. Then dry your hands properly. The puncture site can alternatively be wiped with an alcohol- soaked swab. Please ensure that the site is completely dry before collecting the blood sample.



2. Prepare the lancing device and turn the cap counterclockwise to remove it.
3. Insert the lancet into the lancet holder and push down firmly until it is fully seated. Do **NOT** twist the lancet.
4. Twist the protective disk until it separates from the lancet. Replace the lancing device cap. Turn it clockwise until it is snug fit.

NOTE: If you have your lancing device already, please read information in your lancing device instruction for use.

■ Inserting the Medisign GH82 Blood Glucose Test Strip




1. Insert a test strip to turn on the meter automatically. Make sure it is inserted completely without bending the test strip.

2. The meter identifies the code number automatically. Compare the code number displayed on the LCD with the code number shown on the vial. If they do **NOT** match, try again with another test strip. If the problem persists, please contact Customer Service.
3. When the blood symbol  appears, it indicates that the meter is ready for blood glucose testing.

IMPORTANT

- Before the test you must make sure that the code number displayed on the meter matches the code number on the test strip vial.

■ Obtaining a Blood Drop



1. Adjust the puncture depth setting if necessary, by turning the lancing device cap, number 1 is the shallowest depth while number 5 is the deepest. Slide the ejection/cocking control barrel back until it clicks. If it does **NOT** click, the lancing device may have been cocked when the lancet was inserted.
2. Hold the lancing device firmly against the side of your fingertip and lance the finger by pressing the release button of the lancing device.
3. Gently massage your finger towards the puncture site to produce a drop of blood.

NOTE: Prick the side of fingertip to avoid soreness. To avoid calluses, choose a different testing site each time.

IMPORTANT SAFETY INFORMATION

If you have taken the cap off of the lancet and the lancet has fallen to the ground, do **NOT** use it. Make sure you dispose of it and replace it with a new lancet.


IMPORTANT

- Do **NOT** squeeze excessively around the puncture site when you produce a drop of blood.
- If alcohol wipes are used to disinfect the fingertip, make sure the fingertip is completely dry before obtaining the blood sample.

■ Applying Blood



Place your finger on the tip of the test strip. Ensure that the blood is **NOT** applied from above. The measurement chamber of the test strip draws the blood of your finger automatically. Your finger should remain still until the measuring chamber is completely filled and you hear the “beep” sound.



IMPORTANT

- The sample channel at the end of the test strip should show full. When you hear the beep sound, you have enough blood in the test strip.
- Do **NOT** put urine on the test strip.
- Do **NOT** touch sampling end of the test strip.
- Do **NOT** smear the blood with sampling end of the test strip.
- Do **NOT** drop the blood directly on the sampling end of the test strip.
- Do **NOT** add or apply a second drop of blood. This may cause a false result.
- Do **NOT** hold or shake the meter while it is testing.

■ Measuring Process



After six (6) seconds you will hear a second beep sound and your test result will be displayed on the LCD screen. The Test result will be stored in meter memory automatically (See Viewing the Meter Memory).

■ Flagging Test Results



If you want to flag pre-meal, post-meal or control solution results, press the < or > button after you finish the test. The symbol for post-meal ¶ will be displayed on your screen. Press the < or > button to toggle between pre-meal ¶, post-meal ¶ and control solution ¶ symbol. Remove the test strip to store the test result with the flag which is displayed on the screen.

■ Disposing Used Test Strips and Lancets

Use a Strip Ejecting Button located on the right side of the meter to remove the used test strip. Used test strips should be safely discarded properly according to instructions from your healthcare professional. Wash your hands thoroughly with soap and water after handling the meter, lancing device or test strips.


IMPORTANT SAFETY INFORMATION

- Used lancets and strips are biohazardous material and can transmit blood-borne diseases. Dispose of used lancets and test strips according to instructions from your healthcare professional.
- Do **NOT** leave the lancets in the lancing device after use, be sure to dispose of it according to instructions from your healthcare professional.
- Do **NOT** re-use the lancet.

⚠ CAUTION

Keep the test strip, meter, lancing device, lancets and control solution away from all children and pets. Test strips may be a choking hazard.

5. Understanding Your Blood Glucose Test Results



Be sure to document your test result by writing it in your logbook after your test. You can then take your logbook to your physician who can review your results.

IMPORTANT

- If your test result does **NOT** match how you feel, call your physician or healthcare professional.
- **NEVER** ignore symptoms of high or low blood glucose.
- If you see “HI” or “Lo” displayed, your blood glucose level may be above 600mg/dL or below 20mg/dL, repeat the blood glucose test. If the meter persists same results, contact your physician or healthcare professional.
- **NEVER** change any treatment without first consulting with your physician or a healthcare professional.
- Hematocrit is the percentage of red blood cells in the blood. Testing outside of the Hematocrit level range of 20~60% may cause inaccurate result. If you do **NOT** know your hematocrit level, consult with your healthcare professional.
- This system is **NOT** for use on patients who are in shock, hyperosmolar state or dehydrated patients.
- Inaccurate results may occur in severely hypotensive individuals.
- Inaccurate low results may occur for individuals experiencing a hyperglycemic-hyperosmolar state, with or without ketosis.
- This system is **NOT** for use on critically ill patients.

■ Blood Glucose Result with Smile Symbol



The meter will display a 😊 sign if the result falls in the range of 80-120 mg/dL.

The default value is only for your reference, please consult your health care professional to find out your target blood glucose level.

■ Ketone Alert



When a blood glucose test result is 240 mg/dL or higher (**NOT** for a control solution test result), the **KETONE?** symbol will appear. It is recommended to check your blood β -ketone.

6. Testing Your Blood Glucose from the Palm and Forearm

The Medisign **MediKeto Plus 82 BT** blood glucose and β -ketone monitoring system can test for blood glucose in areas other than the fingertip, such as the palm and forearm (alternative site testing, or AST). Please consult your healthcare professional before using any alternative sites to test your blood glucose.

Alternative site testing can be less painful than fingertip testing. Because of the physiological differences between the fingertip, palm, and forearm, alternative site test results may be significantly different than results from the fingertip under certain conditions.

Do Alternative Site Testing Only the following intervals:

- In a pre-meal or fasting state (more than two (2) hours since the last meal).
- Two (2) hours or more after taking insulin.
- Two (2) hours or more after exercise.


Do NOT Alternative Site Testing when:

- If you sometimes do **NOT** notice when your blood glucose is low or high
- Within two (2) hours of a meal, exercise, or medication.
- During times of stress.
- You are sick.
- Your blood glucose seems to be low (hypoglycemia) or high (hyperglycemia).
- Your glucose results are often fluctuating.
- If your alternative site testing does **NOT** agree with how you feel.

IMPORTANT

- Alternative site testing should be performed only during steady-state (when glucose is **NOT** changing rapidly).
- Alternative site measurements should **NEVER** be used to calibrate continuous glucose monitoring (CGMs).
- Alternative site measurements should **NEVER** be used for insulin dosing calculations.
- If results from alternative site testing do **NOT** match how you feel, perform a fingertip test to confirm the result. If the fingertip test result still does **NOT** match how you feel, call your physician or healthcare professional.

■ Alternative Site Blood Sampling



Forearm sampling:

Choose a fleshy area of the forearm away from the bone, visible veins and hair. Sometimes there is less blood flow to the forearm than to the fingertips. To help you get a large enough drop of blood, you may gently massage or apply warmth to the site to increase blood flow. For forearm test sites, apply and release pressure to the lancing device for a few seconds while keeping the lancing device in constant contact with the skin. When a drop of blood is formed, lift the lancing device straight up without smearing the blood.



Palm sampling:

Choose a fleshy area on the palm below your thumb or little finger. Select an area away from visible veins or deep creases which may cause your blood sample to smear.

- Press the lancing device firmly against the test site and lance the site. Do **NOT** release the lancing device immediately.
- You may have to adjust the lancing device to a deeper setting to get a large enough drop of blood.
- You may use the Lancing Device Clear Cap for forearm and palm sampling. Replace the lancing device cap to the Clear Cap.

IMPORTANT


- You may need to wait a little longer to get a large enough drop of blood from the forearm or palm. Do **NOT** squeeze the site excessively.
- If the sample drop of blood runs or spreads due to contact with hair or a crease in the palm, do **NOT** use that sample. Try puncturing again in a smoother area.

7. Testing Your Blood β -Ketone


IMPORTANT SAFETY INFORMATION

- Meters that come into contact with blood present a potential risk of transmitting infectious diseases. To prevent infections and/or cross-contamination, maintain good hand hygiene and clean your Medisign **MediKeto 82 BT** blood β -ketone meter as well as the corresponding accessories regularly.
- Healthcare Professionals performing the tests with this system on multiple patients must always wear gloves and should follow the infection control policies and procedures approved by their facility.
- The meter and lancing device are for single patient use only. Do **NOT** share them with anyone including family members.
- **NEVER** use lancets that has been used by someone else.
- All parts of the kit are considered biohazardous and can potentially transmit infectious diseases, even after you have performed cleaning.

■ Preparing the Test



1. The easiest way to collect a drop of blood is to wash your hands with warm water and soap first. This promotes the blood circulation in the fingertips. Then dry your hands properly. The puncture site can alternatively be wiped with an alcohol- soaked swab. Please ensure that the site is completely dry before collecting the blood sample.



2. Prepare the lancing device and Turn the cap counterclockwise to remove it.
3. Insert the lancet into the lancet holder and push down firmly until it is fully seated. Do **NOT** twist the lancet.
4. Twist the protective disk until it separates from the lancet. Replace the lancing device cap. Turn it clockwise until it is snug fit.

NOTE: If you have your lancing device already, please read information in your lancing device instruction for use.

■ Inserting the Medisign **MediKeto 82** Blood β -Ketone Test Strip




1. Insert a test strip to turn on the meter automatically. Make sure it is inserted completely without bending the test strip. **KETONE** will appear on screen if the Medisign **MediKeto 82** blood β -ketone test strip is inserted.

2. The meter identifies the code number automatically. Compare the code number displayed on the LCD with the code number shown on the vial or the foil packet. If they do **NOT** match, try again with another test strip. If the problem persists, please contact Customer Service.
3. When the blood symbol  appears, it indicates that the meter is ready for blood β -ketone testing.

IMPORTANT

- Before the test you must make sure that the code number displayed on the meter matches the code number on the test strip vial or foil packet.

■ Obtaining a Blood Drop



1. Adjust the puncture depth setting if necessary, by turning the lancing device cap, number 1 is the shallowest depth while number 5 is the deepest. Slide the ejection/cocking control barrel back until it clicks. If it does **NOT** click, the lancing device may have been cocked when the lancet was inserted.
2. Hold the lancing device firmly against the side of your fingertip and lance the finger by pressing the release button of the lancing device.
3. Gently massage your finger towards the puncture site to produce a drop of blood.

NOTE: Prick the side of fingertip to avoid soreness. To avoid calluses, choose a different testing site each time.

IMPORTANT

- Do **NOT** squeeze excessively around the puncture site when you produce a drop of blood.
- If alcohol wipes are used to disinfect the fingertip, make sure the fingertip is completely dry before obtaining the blood sample.
- Blood β -ketone testing must **NOT** use the forearm or palm site. Only use the fingertip for blood β -ketone testing.


IMPORTANT SAFETY INFORMATION

If you have taken the cap off of the lancet and the lancet has fallen to the ground, do **NOT** use it. Make sure you dispose of it and replace it with a new lancet.

■ Applying Blood



Place your finger on the tip of the test strip. Ensure that the blood is **NOT** applied from above. The measurement chamber of the test strip draws the blood of your finger automatically. Your finger should remain still until the measuring chamber is completely filled and you hear the “beep” sound.



IMPORTANT

- The sample channel at the end of the test strip should show full. When you hear the beep sound, you have enough blood in the test strip.
- Do **NOT** put urine on the test strip.
- Do **NOT** touch sampling end of the test strip.
- Do **NOT** smear the blood with sampling end of the test strip.
- Do **NOT** drop the blood directly on the sampling end of the test strip.
- Do **NOT** add or apply a second drop of blood. This may cause a false result.
- Do **NOT** hold or shake the meter while it is testing.

Measuring Process



After eight (8) seconds you will hear a second beep sound and your test result will be displayed on the LCD screen.

The Test result will be stored in meter memory automatically (See Viewing the Meter Memory).

NOTE: The meter will display “0.0 mmol/L” when blood β -ketone test result is less than 0.1 mmol/L.

Disposing Used Test Strips and Lancets

Use a Strip Ejecting Button located on the right side of the meter to remove the used test strip.

Used test strips should be safely discarded properly according to instructions from your healthcare professional. Wash your hands thoroughly with soap and water after handling the meter, lancing device or test strips.

IMPORTANT SAFETY INFORMATION

- Used lancets and strips are biohazardous material and can transmit blood-borne diseases. Dispose of used lancets and test strips according to instructions from your healthcare professional.
- Do **NOT** leave the lancets in the lancing device after use, be sure to dispose of it according to instructions from your healthcare professional.
- Do **NOT** re-use the lancet.

CAUTION

Keep the test strip, meter, lancing device, lancets and control solution away from all children and pets. Test strips may be a choking hazard.

8. Understanding Your Blood β -Ketone Test Results

Blood β -Ketone Results Interpretation

β -Ketone Results	Interpretation
< 0.6 mmol/L	Normal, please consult with a healthcare professional for your appropriate range of blood β -ketone level.
0.6 ~ 1.5 mmol/L	With blood glucose level greater than 300 mg/dL or 16.7 mmol/L, medical consultation with a healthcare professional is recommended.
> 1.5 mmol/L	With blood glucose level greater than 300 mg/dL or 16.7 mmol/L, consult with a healthcare professional immediately as it indicates a risk of developing Diabetic Ketoacidosis (DKA).

NOTE: Consult with your healthcare professional for the blood β -ketone range that is appropriate for you.


IMPORTANT

- If your test result does **NOT** match how you feel, call your physician or healthcare professional.
- **NEVER** ignore symptoms of high blood β -ketone.
- If you see “HI” displayed with **KETONE**, your blood β -ketone level may be above 8.0 mmol/L, repeat the blood β -ketone test using a new test strip. If the meter persists same results, contact your physician or healthcare professional.
- **NEVER** change any treatment without first consulting with your physician or a healthcare professional.
- Hematocrit is the percentage of red blood cells in the blood. Testing outside of the Hematocrit level range of 20~60% may cause inaccurate result. If you do **NOT** know your hematocrit level, consult with your healthcare professional.
- This system is **NOT** for use on patients who are in shock or dehydrated.
- Inaccurate results may occur in severely hypotensive individuals.
- This system is **NOT** for use on critically ill patients.

9. Displaying Stored Test Results

The Medisign **MediKeto Plus 82 BT** blood glucose and β -ketone meter automatically stores up to five hundred (500) glucose and β -ketone test results together with the date and time. The meter also provides a seven (7), fourteen (14), thirty (30) days' average to help track a patient's blood glucose trend. The time and date must be set in order to use the memory and day averaging function (See Set Date and Time).

■ Display of Test Results



1. Begin with the meter turned off. Press the $\mathcal{U}$ button, you will hear a beep sound and the meter will turn on.
2. The screen will display the date and time. After two (2) seconds, the average blood glucose test result of last seven (7) days will be displayed with a sign $\mathcal{U}^{DAY} AVG$ and the number of test results for seven (7) days on the display. For example, **8n** above the picture means that there are eight (8) test results saved on the memory during the last seven (7) days. Each time pressing the < button, fourteen (14) days and thirty (30) days' average result will be displayed.
3. The latest test result is displayed first when you press the > button while the seven (7) days' average is on the screen.
4. Each time you continue to press the > button, the screen will show the test result from the most recent to the oldest. Use the < or > button to switch stored results.

NOTE:

- Control solution and blood β -ketone test results will **NOT** be included in the seven (7), fourteen (14), thirty (30) days' average calculation.
- After five hundred (500) data entries, every new test results replaces the oldest (first in, first out).
- Meter memory test results and day averages will **NOT** be erased when battery is removed.
- The Meter will automatically turn off after two (2) minutes of non-use.

10. Downloading Your Test Results to a PC

You can transfer test results from the Medisign **MediKeto Plus 82 BT** blood glucose and β -ketone meter to a computer. The program can help you to make a report with graphs and tables. To use this feature, you need the Medisign Link Diabetes Management Software and the Medisign Link Cable. For more information, please contact Customer Service.

NOTE: While the meter is connected to the PC, you are unable to perform a test.

11. Pairing Your Meter

■ Before Pairing Your Meter

Pairing must be completed between the Medisign **MediKeto Plus 82 BT** blood glucose and β -ketone meter and your Bluetooth mobile device before your meter can communicate with your mobile device. Both devices must be within ten meter (10 M) range apart of each other in order to sync properly. The meter can be paired with only one smart device at a time. To avoid malfunctions when transferring data, please remove the meter from the list of paired devices before setting up any new pairing.

The Medisign **MediKeto Plus 82 BT** blood glucose and β -ketone meter is compatible with:

- Android device (requires OS 4.3 or higher)
- Apple iOS device (requires iOS 8.0 or higher)
- Mobile devices compatible with Bluetooth 4.0

■ How to Pair Your Meter

1. Press the  button to turn on the meter.
2. If your mobile device is set to search for other devices, you can find the Medisign **MediKeto Plus 82 BT** blood glucose and β -ketone meter name along with the serial number displayed.
3. Select or tap the name of the meter displayed you wish to connect.
4. If your mobile device and the meter are successfully connected, the  will be shown on the meter display.
5. If you are prompted to pair your mobile device after connecting, you may need to enter the pairing password on the display. The pairing password is “000000” (six zeros).

■ About Bluetooth

This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions:

- (1) This device may not cause harmful interference, and
- (2) This device must accept any interference received, including interference that may cause undesired operations.

12. Cleaning Your Meter

Wash and dry your hands before use to prevent damage of the meter and test strips. The meter should be cleaned with a soft cloth or paper towel slightly dampened with one of these cleaning solutions:

- 70% isopropyl alcohol
- Mild dishwashing liquid mixed with water
- 10% household bleach solution made on the day

After cleaning, allow the meter to dry completely in a cool place away from direct sunlight.

IMPORTANT

- The meter must be turned off when you clean the meter.
- Clean your Medisign **MediKeto Plus 82 BT** blood glucose and β -ketone meter once per week at a minimum.
- Do **NOT** get water inside the meter. **NEVER** immerse the meter or hold it under running water or spray any cleaning solution directly onto the meter.
- Do **NOT** use glass cleaners or household cleaners or corrosive liquid (e.g. Benzene, Acetone) on the meter.
- Do **NOT** try to get moisture on the test strip, test strip slot or serial port.
- Do **NOT** destroy, repair or transform the system at any time. If the equipment is used in a manner NOT specified by the manufacturer, the protection provided may be impaired.
- Handle the meter with care – severe shock, such as dropping the meter, could damage the electronics.
- Squeeze the excess liquid from the cloth before you wipe the meter.

Sign of Deterioration

The list of things you should look for as signs of deterioration after cleaning include:





- Control Solution results is out of the range
- Clouding of the LCD display
- Corroding of the plastic housing or buttons
- Malfunction of any meter function

If these signs of deterioration are noted, do **NOT** use the meter and contact Customer Service.

13. Screen Messages and Troubleshooting

Screen Messages

This section explains the display screen messages and error messages you may encounter when using your Medisign **MediKeto Plus 82 BT** blood glucose and β -ketone meter and test strips. If your meter is still **NOT** working after you have followed the troubleshooting instructions, contact Customer Service.

What you see	What it means	What you should do
	The meter has detected a problem with the test strip. The test strip is either used or damaged.	Repeat the test with a new test strip.
	The blood sample was applied to the test strip before the blood symbol appeared on the screen.	Remove and discard the test strip and insert a new one. Apply blood to the test strip after the blood symbol  appear on the screen.
	The ambient temperature is too high or low to test.	Put the meter and test strips in a location within the operating temperature range (10 ~ 40°C) for at least twenty (20) minutes and repeat the test. Do NOT raise or lower the meter temperature artificially.
	Not enough blood sample was drawn into the test strip.	Repeat the test with a new test strip. If error persists, please contact Customer Service.
	The wrong test strip has been inserted into the meter.	Make sure you are using the Medisign GH82 blood glucose test strip or the Medisign MediKeto 82 blood β -ketone test strip and repeat the test.
	There is an internal error with the meter.	Please contact Customer Service.
	Temperature is out of the operating range during the test.	Repeat the test under the stable temperature condition.
	The test strip is flicked during the test.	Repeat test with a new test strip. Please do NOT touch the test strip during the test. If error persists, contact to Customer Service.
	Blood glucose test result is higher than 600 mg/dL.	Repeat the test with a new test strip. If the result is still “HI”, contact your physician or healthcare professional.
	Blood glucose test result is lower than 20 mg/dL.	Repeat the test with a new test strip. If the result is still “Lo”, contact your physician or healthcare professional.
	Blood β -ketone test result is higher than 8.0 mmol/L.	Repeat the test with a new test strip. If the result is still “HI”, contact your physician or healthcare professional.


Troubleshooting

What is happening	What you should do
The meter does NOT work after the blood is applied.	• Make sure that the test strip is insert into the test strip slot completely. • See if the test strip is covered in debris or lint. • Make sure the blood is applied to the test strip correctly. • Make sure the test strip is inserted correctly and repeat the test.
The test result is abnormal.	• See if the test strip is covered in debris or lint. • Check if the test strip is within the validity period. • Make sure the test strip has NOT been stored out of its vial. • Make sure the test strip has NOT been used before. • Make sure the code on the meter display matches with the test strip code.
The thermometer symbol  appears on the screen.	• Put the meter and test strips in a location within the operating temperature range (10 ~ 40°C) for at least twenty (20) minutes and repeat the test. Do NOT raise or lower the meter temperature artificially. • If the thermometer symbol is constantly showing, contact Customer Service.
The meter does NOT work after the test strip is inserted.	• Make sure the test strip is completely inserted into the test strip port. • Make sure the batteries are inserted correctly, with the + sign facing upwards.

14. System Specifications

Measuring Range (Glucose)	20 ~ 600 mg/dL
Measuring Range (β-ketone)	0.1 ~ 8.0 mmol/L
Measuring Time (Glucose)	6 seconds
Measuring Time (β-ketone)	8 seconds
Memory capacity	500 test results (include date and time)
Sample Type (Glucose)	Fresh capillary and venous whole blood
Sample Type (β-ketone)	Fresh capillary(only fingertip) and venous whole blood
Operating condition	10 ~ 40°C
Sample volume	0.5 μℓ
Screen type	Backlight LCD
Dimensions (mm)	90 (L) x 53 (W) x 18 (H)
Weight	Approximately 52g (with batteries)
Power supply	DC 3V (CR2032) Lithium Battery
Data Output	Bluetooth, USB Data Transfer Cable
Wireless Frequency	2.4 GHz Band

15. Symbols

						
Keep away from sunlight	Biological Risks	Manufacturer	LOT Number	Temperature Limitation	Catalogue Number	Authorized Representative in the European Community
						
Use By	In vitro Diagnostic Medical Device	Consult Instructions for Use	Date of Manufacture	Do not reuse	Serial Number	Direct Current
						
Sufficient For	Caution, refer to safety-related notes in the instructions for use accompanying this product					

16. Disposing of the Meter

IMPORTANT SAFETY INFORMATION

Objects (e. g. the meter) which had been in contact with blood potentially bear the risk for infectious diseases. Please dispose these materials according to local regulations for disposal of contaminated medical devices.



Please dispose of the contaminated meter – after removing the battery – according to the regulations applicable to your country. For information about correct disposal, please contact your local council or authority. The meter falls within the scope of European Directive 2012/19/EU (Directive on waste electrical and electronic equipment (WEEE)).

17. Additional Supplies

The following supplies and accessories are available from your supplier, for more information contact Customer Service.

- Medisign Link Cable
- Medisign Link Diabetes Management Software
- Medisign **GH82** blood glucose test strips
- Medisign **MediKeto 82** blood β -ketone test strips
- Medisign GH Glucose Control Solution
- Medisign Ketone Control Solution
- Medisign GH Check Strip
- Medisign Ketone Check Strip

18. Product Warranty

The Medisign **MediKeto Plus 82 BT** blood glucose and β -ketone meter is warranted to be free of defects in workmanship and materials for a period of three (3) years from the date of purchase. Manufacturer's liability for

warranty claims is limited to repair or replacement, and in no event shall manufacturer be liable for indirect or consequential damages, or for any loss arising from misuse, improper use, abuse, neglect, unauthorized repair or modification.

This warranty is void and of no force and effect in the event of product misuse, improper use, abuse, neglect, unauthorized repair or modification. This warranty specifically excludes the Medisign **GH82** blood glucose test strip, the Medisign **MediKeto 82** blood β -ketone test strip, the Medisign GH Glucose Control Solution and the Medisign Ketone Control Solution.

THIS WARRANTY IS THE SOLE AND EXCLUSIVE WARRANTY TO THE EXCLUSION OF ALL OTHER WARRANTIES, EXPRESS OR IMPLIED, ALL OF WHICH ARE WAIVED AND DISALAIMED.

All warranty claims must be directed to Customer Service. This warranty is extended only to the original purchaser of the meter.

References

1. Schade DS, Eaton RP. Metabolic and clinical significance of ketosis. *Special Topics in Endocrinology and Metabolism* 1982;4:1-27.
2. Wiggam MI, O’Kane MJ, Harper R, Atkinson AB, Hadden DR, Trimble ER, Bell PM. Treatment of diabetic ketoacidosis using normalization of blood 3-hydroxybutyrate concentration as the endpoint of emergency management. *Diabetes Care* 1997;20:1347-52.
3. Harano Y, Kosugi K, Hyosu T, Suzuki M, Hidaka H, Kashiwagi A, Uno S, Shigeta Y. Ketone bodies as markers for Type 1 (insulin-dependent) diabetes and their value in the monitoring of diabetes control. *Diabetologia* 1984;26:343-8.
4. Ubukata E. Diurnal variation of blood β -Ketone bodies in insulin-dependent diabetes mellitus and noninsulindependent diabetes mellitus patients: The relationship to serum C-peptide immunoreactivity and free insulin. *Ann Nutr Metab* 1990;34:333-42.
5. Luzi L, Barrett EJ, Groop LC, Ferrannini E, DeFronzo RA. Metabolic effects of low-dose insulin therapy on glucose metabolism in diabetic ketoacidosis. *Diabetes* 1988;37:1470-77.
6. Hale PJ, Crase J, Nattrass M. Metabolic effects of bicarbonate in the treatment of diabetic ketoacidosis. *Br Med J* 1984;289:1035-8.
7. Guideline for Isolation Precautions: Preventing Transmission of Infectious Agents in Healthcare Settings 2007'. <http://www.cdc.gov/hicpac/2007ip/2007isolationprecautions.html>

FCC Information to User

This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to Part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.

Caution

Modifications not expressly approved by the party responsible for compliance could void the user's authority to operate the equipment.

FCC Compliance Information : This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) This device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

This equipment complies with FCC radiation exposure limits set forth for an uncontrolled environment.