

#01_NFC_ASK13.56M_Back_0mm

Communication System: WPT; Frequency: 13.56 MHz; Duty Cycle: 1:1

Medium: HSL_13_220623 Medium parameters used : $f = 13.56$ MHz; $\sigma = 0.726$ S/m; $\epsilon_r = 54.247$; $\rho = 1000$ kg/m³

Ambient Temperature : 23.5 °C ; Liquid Temperature : 22.5 °C

DASY5 Configuration:

- Probe: EX3DV4 - SN3931; ConvF(18.36, 18.36, 18.36) @ 13.56 MHz; Calibrated: 2021/10/21
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn699; Calibrated: 2022/2/24
- Phantom: SAM_Left; Type: SAM; Serial: 1796
- Measurement SW: DASY52, Version 52.10 (4); SEMCAD X Version 14.6.14 (7501)

Area Scan (71x121x1): Interpolated grid: $dx=1.500$ mm, $dy=1.500$ mm

Maximum value of SAR (interpolated) = 0.180 W/kg


Zoom Scan (6x6x7)/Cube 0: Measurement grid: $dx=8$ mm, $dy=8$ mm, $dz=5$ mm

Reference Value = 13.81 V/m; Power Drift = -0.19 dB

Peak SAR (extrapolated) = 0.291 W/kg

SAR(1 g) = 0.070 W/kg; SAR(10 g) = 0.028 W/kg

Maximum value of SAR (measured) = 0.179 W/kg



0 dB = 0.180 W/kg = -7.45 dBW/kg