



	In 2024	2025		
Impedence for Head TSL	$45.0\Omega + 1.2j\Omega$	$46.97\Omega + 3.05j\Omega$	$1.97\Omega + 1.85j\Omega$	$\pm 5\Omega$
Return Loss for Head TSL	-25.24dB	-25.90dB	2.61%	$\pm 20\%$ (No less than 20 dB)

The return loss is <-20dB, within 20% of prior calibration; the impedance is within 5 ohm of prior calibration.

Therefore the verification result should support extended calibration.

※※END OF THE REPORT※※