

Measurement Report for Device

Exposure Conditions

Phantom Section	Position, Test Distance [mm]	Frequency [MHz], Channel Number	Conversion Factor
5G Air	Front, 10.00	10000.0Validation band, 10000	1.0

Hardware Setup


Phantom	Medium	Probe, Calibration Date	DAE, Calibration Date
mmWave- 1083	---Air	EUmmWV4 - SN9607_F1-55GHz, 2026-01-14	DAE4 Sn878, 2026-03-23

Scan Setup

	5G Scan
Grid Extents [mm]	25.0 x 25.0
Grid Steps [lambda]	0.25 x 0.25
Sensor Surface [mm]	10.0
MAIA	N/A

Measurement Results

	5G Scan
Date	2026-07-18
Avg. Area [cm²]	4.00
psPDn+ [W/m²]	174
psPDtot+ [W/m²]	177
psPDmod+ [W/m²]	181
E _{max} [V/m]	282
Power Drift [dB]	-0.03



Measurement Report for Device

Exposure Conditions

Phantom Section	Position, Test Distance [mm]	Frequency [MHz], Channel Number	Conversion Factor
5G Air	Front, 10.00	10000.0Validation band, 10000	1.0

Hardware Setup


Phantom	Medium	Probe, Calibration Date	DAE, Calibration Date
mmWave- 1083	---Air	EUmmWV4 - SN9607_F1-55GHz, 2026-01-14	DAE4 Sn878, 2026-03-23

Scan Setup

	5G Scan
Grid Extents [mm]	25.0 x 25.0
Grid Steps [lambda]	0.25 x 0.25
Sensor Surface [mm]	10.0
MAIA	N/A

Measurement Results

	5G Scan
Date	2026-07-19
Avg. Area [cm²]	4.00
psPDn+ [W/m²]	153
psPDtot+ [W/m²]	156
psPDmod+ [W/m²]	152
E _{max} [V/m]	268
Power Drift [dB]	0.08



Measurement Report for Device

Exposure Conditions

Phantom Section	Position, Test Distance [mm]	Frequency [MHz], Channel Number	Conversion Factor
5G Air	Front, 10.00	10000.0Validation band, 10000	1.0

Hardware Setup


Phantom	Medium	Probe, Calibration Date	DAE, Calibration Date
mmWave- 1083	---Air	EUmmWV4 - SN9607_F1-55GHz, 2026-01-14	DAE4 Sn878, 2026-03-23

Scan Setup

	5G Scan
Grid Extents [mm]	25.0 x 25.0
Grid Steps [lambda]	0.25 x 0.25
Sensor Surface [mm]	10.0
MAIA	N/A

Measurement Results

	5G Scan
Date	2026-07-20
Avg. Area [cm²]	4.00
psPDn+ [W/m²]	148
psPDtot+ [W/m²]	151
psPDmod+ [W/m²]	149
E _{max} [V/m]	266
Power Drift [dB]	0.08



Measurement Report for Device

Exposure Conditions

Phantom Section	Position, Test Distance [mm]	Frequency [MHz], Channel Number	Conversion Factor
5G Air	Front, 10.00	10000.0Validation band, 10000	1.0

Hardware Setup


Phantom	Medium	Probe, Calibration Date	DAE, Calibration Date
mmWave- 1083	---Air	EUmmWV4 - SN9607_F1-55GHz, 2026-01-14	DAE4 Sn878, 2026-03-23

Scan Setup

	5G Scan
Grid Extents [mm]	25.0 x 25.0
Grid Steps [lambda]	0.25 x 0.25
Sensor Surface [mm]	10.0
MAIA	N/A

Measurement Results

	5G Scan
Date	2026-07-21
Avg. Area [cm²]	4.00
psPDn+ [W/m²]	145
psPDtot+ [W/m²]	149
psPDmod+ [W/m²]	148
E _{max} [V/m]	249
Power Drift [dB]	0.02



Measurement Report for Device

Exposure Conditions

Phantom Section	Position, Test Distance [mm]	Frequency [MHz], Channel Number	Conversion Factor
5G Air	Front, 10.00	10000.0Validation band, 10000	1.0

Hardware Setup

Phantom	Medium	Probe, Calibration Date	DAE, Calibration Date
mmWave- 1083	---Air	EUmmWV4 - SN9607_F1-55GHz, 2026-01-14	DAE4 Sn878, 2026-03-23

Scan Setup

	5G Scan
Grid Extents [mm]	25.0 x 25.0
Grid Steps [lambda]	0.25 x 0.25
Sensor Surface [mm]	10.0
MAIA	N/A

Measurement Results

	5G Scan
Date	2026-07-22
Avg. Area [cm²]	4.00
psPDn+ [W/m²]	171
psPDtot+ [W/m²]	175
psPDmod+ [W/m²]	179
E _{max} [V/m]	274
Power Drift [dB]	0.04

