

Appendix B
CN24K76Q 003

Prüfbericht - Produkte
Test Report - Products

Page 218 of 311




Shenzhen PSI Testing Co., Ltd.
1-2F, Building 5, Yudafu Industrial Park, No. 10, Xingye West Road,
Shajing Street, Bao'an District, Shenzhen, Guangdong, China 518104

Test Site: PSI EMC Lab-3m Chamber Polarization: **Horizontal** Temperature: 26 °C
EUT: See note Distance: 3m Humidity: 54 %RH
M/N: See note Test Engineer: Felix Pang
Test Mode: ax40 5190MHz(MIMO)
Test Voltage: DC 24V
Note: Optical 3D Measuring and Dynamic Tracking System
FreeScan Trak ProW(FreeScan TE25 W)

Radiated Emission Measurement

Test Date: 2024/11/6 10:32:03



No.	Frequency (MHz)	Reading (dBuV)	Factor (dB/m)	Level (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector	Height (cm)	Azimuth (deg.)	P/F	Remark
1 *	5143.680	54.84	-18.32	36.52	54.00	-17.48	AVG	150	124	P	
2	5150.000	53.46	-18.31	35.15	54.00	-18.85	AVG	150	168	P	

Note: *:Maximum data x:Over limit !:over margin Level = Reading + Factor Margin = Level - Limit

Appendix B
CN24K76Q 003

Prüfbericht - Produkte
Test Report - Products

Page 219 of 311



Shenzhen PSI Testing Co., Ltd.
1-2F, Building 5, Yudafu Industrial Park, No. 10, Xingye West Road,
Shajing Street, Bao'an District, Shenzhen, Guangdong, China 518104

Test Site: PSI EMC Lab-3m Chamber Polarization: **Vertical** Temperature: 26 °C
EUT: See note Distance: 3m Humidity: 54 %RH
M/N: See note Test Engineer: Felix Pang
Test Mode: ax40 5190MHz(MIMO)
Test Voltage: DC 24V
Note: Optical 3D Measuring and Dynamic Tracking System
FreeScan Trak ProW(FreeScan TE25 W)

Radiated Emission Measurement

Test Date: 2024/11/6 10:38:52



No.	Frequency (MHz)	Reading (dBuV)	Factor (dB/m)	Level (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector	Height (cm)	Azimuth (deg.)	P/F	Remark
1	5133.760	65.85	-19.20	46.65	74.00	-27.35	peak	150	206	P	
2 *	5150.000	58.66	-19.16	39.50	54.00	-14.50	AVG	150	168	P	

Note: *:Maximum data x:Over limit !:over margin Level = Reading + Factor Margin = Level - Limit

Appendix B
CN24K76Q 003
Page 220 of 311

Prüfbericht - Produkte
Test Report - Products




Shenzhen PSI Testing Co., Ltd.
1-2F, Building 5, Yudafu Industrial Park, No. 10, Xingye West Road,
Shajing Street, Bao'an District, Shenzhen, Guangdong, China 518104

Test Site: PSI EMC Lab-3m Chamber Polarization: **Vertical** Temperature: 26 °C
EUT: See note Distance: 3m Humidity: 54 %RH
M/N: See note Test Engineer: Felix Pang
Test Mode: ax40 5230MHz(MIMO)
Test Voltage: DC 24V
Note: Optical 3D Measuring and Dynamic Tracking System
FreeScan Trak ProW(FreeScan TE25 W)

Radiated Emission Measurement

Test Date: 2024/11/6 10:42:00



No.	Frequency (MHz)	Reading (dBuV)	Factor (dB/m)	Level (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector	Height (cm)	Azimuth (deg.)	P/F	Remark
1 *	5350.000	50.22	-18.63	31.59	68.82	-37.23	peak	150	138	P	
2	5355.200	52.70	-18.60	34.10	74.00	-39.90	peak	150	127	P	

Note: *:Maximum data x:Over limit !:over margin Level = Reading + Factor Margin = Level - Limit

Appendix B
CN24K76Q 003

Prüfbericht - Produkte
Test Report - Products

Page 221 of 311




Shenzhen PSI Testing Co., Ltd.
1-2F, Building 5, Yudafu Industrial Park, No. 10, Xingye West Road,
Shajing Street, Bao'an District, Shenzhen, Guangdong, China 518104

Test Site: PSI EMC Lab-3m Chamber Polarization: **Horizontal** Temperature: 26 °C
EUT: See note Distance: 3m Humidity: 54 %RH
M/N: See note Test Engineer: Felix Pang
Test Mode: ax40 5230MHz(MIMO)
Test Voltage: DC 24V
Note: Optical 3D Measuring and Dynamic Tracking System
FreeScan Trak ProW(FreeScan TE25 W)

Radiated Emission Measurement

Test Date: 2024/11/6 10:47:11



No.	Frequency (MHz)	Reading (dBuV)	Factor (dB/m)	Level (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector	Height (cm)	Azimuth (deg.)	P/F	Remark
1 *	5350.000	50.89	-17.98	32.91	68.82	-35.91	peak	150	124	P	
2	5357.920	51.60	-17.96	33.64	74.00	-40.36	peak	150	204	P	

Note: *:Maximum data x:Over limit !:over margin Level = Reading + Factor Margin = Level - Limit

Appendix B
CN24K76Q 003

Prüfbericht - Produkte
Test Report - Products

Page 222 of 311




Shenzhen PSI Testing Co., Ltd.
1-2F, Building 5, Yudafu Industrial Park, No. 10, Xingye West Road,
Shajing Street, Bao'an District, Shenzhen, Guangdong, China 518104

Test Site: PSI EMC Lab-3m Chamber Polarization: **Horizontal** Temperature: 26 °C
EUT: See note Distance: 3m Humidity: 54 %RH
M/N: See note Test Engineer: Felix Pang
Test Mode: ac80 5210MHz(MIMO)
Test Voltage: DC 24V
Note: Optical 3D Measuring and Dynamic Tracking System
FreeScan Trak ProW(FreeScan TE25 W)

Radiated Emission Measurement

Test Date: 2024/11/6 10:52:30



No.	Frequency (MHz)	Reading (dBuV)	Factor (dB/m)	Level (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector	Height (cm)	Azimuth (deg.)	P/F	Remark
1	5148.000	52.22	-18.32	33.90	54.00	-20.10	AVG	150	125	P	
2 *	5150.000	52.85	-18.31	34.54	54.00	-19.46	AVG	150	247	P	

Note: *:Maximum data x:Over limit !:over margin Level = Reading + Factor Margin = Level - Limit

Appendix B
CN24K76Q 003

Prüfbericht - Produkte
Test Report - Products

Page 223 of 311




Shenzhen PSI Testing Co., Ltd.
1-2F, Building 5, Yudafu Industrial Park, No. 10, Xingye West Road,
Shajing Street, Bao'an District, Shenzhen, Guangdong, China 518104

Test Site: PSI EMC Lab-3m Chamber Polarization: **Vertical** Temperature: 26 °C
EUT: See note Distance: 3m Humidity: 54 %RH
M/N: See note Test Engineer: Felix Pang
Test Mode: ac80 5210MHz(MIMO)
Test Voltage: DC 24V
Note: Optical 3D Measuring and Dynamic Tracking System
FreeScan Trak ProW(FreeScan TE25 W)

Radiated Emission Measurement

Test Date: 2024/11/6 10:56:23



No.	Frequency (MHz)	Reading (dBuV)	Factor (dB/m)	Level (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector	Height (cm)	Azimuth (deg.)	P/F	Remark
1	5138.720	71.05	-19.19	51.86	74.00	-22.14	peak	150	208	P	
2 *	5150.000	69.66	-19.16	50.50	68.82	-18.32	peak	150	138	P	

Note: *:Maximum data x:Over limit !:over margin Level = Reading + Factor Margin = Level - Limit

Appendix B
CN24K76Q 003

Prüfbericht - Produkte
Test Report - Products

Page 224 of 311




Shenzhen PSI Testing Co., Ltd.
1-2F, Building 5, Yudafu Industrial Park, No. 10, Xingye West Road,
Shajing Street, Bao'an District, Shenzhen, Guangdong, China 518104

Test Site: PSI EMC Lab-3m Chamber Polarization: **Vertical** Temperature: 26 °C
EUT: See note Distance: 3m Humidity: 54 %RH
M/N: See note Test Engineer: Felix Pang
Test Mode: ax80 5210MHz(MIMO)
Test Voltage: DC 24V
Note: Optical 3D Measuring and Dynamic Tracking System
FreeScan Trak ProW(FreeScan TE25 W)

Radiated Emission Measurement

Test Date: 2024/11/6 10:59:08



No.	Frequency (MHz)	Reading (dBuV)	Factor (dB/m)	Level (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector	Height (cm)	Azimuth (deg.)	P/F	Remark
1	5137.760	54.31	-19.19	35.12	54.00	-18.88	AVG	150	138	P	
2 *	5150.000	55.44	-19.16	36.28	54.00	-17.72	AVG	150	247	P	

Note: *:Maximum data x:Over limit !:over margin Level = Reading + Factor Margin = Level - Limit

Appendix B
CN24K76Q 003

Prüfbericht - Produkte
Test Report - Products

Page 225 of 311




Shenzhen PSI Testing Co., Ltd.
1-2F, Building 5, Yudafu Industrial Park, No. 10, Xingye West Road,
Shajing Street, Bao'an District, Shenzhen, Guangdong, China 518104

Test Site: PSI EMC Lab-3m Chamber Polarization: **Horizontal** Temperature: 26 °C
EUT: See note Distance: 3m Humidity: 54 %RH
M/N: See note Test Engineer: Felix Pang
Test Mode: ax80 5210MHz(MIMO)
Test Voltage: DC 24V
Note: Optical 3D Measuring and Dynamic Tracking System
FreeScan Trak ProW(FreeScan TE25 W)

Radiated Emission Measurement

Test Date: 2024/11/6 11:02:16



No.	Frequency (MHz)	Reading (dBuV)	Factor (dB/m)	Level (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector	Height (cm)	Azimuth (deg.)	P/F	Remark
1	5118.720	54.07	-18.36	35.71	54.00	-18.29	AVG	150	158	P	
2 *	5150.000	55.26	-18.31	36.95	54.00	-17.05	AVG	150	247	P	

Note: *:Maximum data x:Over limit !:over margin Level = Reading + Factor Margin = Level - Limit

Appendix B
CN24K76Q 003

Prüfbericht - Produkte
Test Report - Products

Page 226 of 311




Shenzhen PSI Testing Co., Ltd.
1-2F, Building 5, Yudafu Industrial Park, No. 10, Xingye West Road,
Shajing Street, Bao'an District, Shenzhen, Guangdong, China 518104

Test Site: PSI EMC Lab-3m Chamber Polarization: **Horizontal** Temperature: 26 °C
EUT: See note Distance: 3m Humidity: 54 %RH
M/N: See note Test Engineer: Felix Pang
Test Mode: a 5180MHz(Ant 1)
Test Voltage: DC 24V
Note: Optical 3D Measuring and Dynamic Tracking System
FreeScan Trak ProW(FreeScan TE25 W)

Radiated Emission Measurement

Test Date: 2024/11/6 08:40:22



No.	Frequency (MHz)	Reading (dBuV)	Factor (dB/m)	Level (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector	Height (cm)	Azimuth (deg.)	P/F	Remark
1	5117.100	48.88	-18.36	30.52	74.00	-43.48	peak	150	78	P	
2 *	5150.000	48.20	-18.31	29.89	68.82	-38.93	peak	150	135	P	

Note: *:Maximum data x:Over limit !:over margin Level = Reading + Factor Margin = Level - Limit

Appendix B
CN24K76Q 003

Prüfbericht - Produkte
Test Report - Products

Page 227 of 311




Shenzhen PSI Testing Co., Ltd.
1-2F, Building 5, Yudafu Industrial Park, No. 10, Xingye West Road,
Shajing Street, Bao'an District, Shenzhen, Guangdong, China 518104

Test Site: PSI EMC Lab-3m Chamber Polarization: **Vertical** Temperature: 26 °C
EUT: See note Distance: 3m Humidity: 54 %RH
M/N: See note Test Engineer: Felix Pang
Test Mode: a 5180MHz(Ant 1)
Test Voltage: DC 24V
Note: Optical 3D Measuring and Dynamic Tracking System
FreeScan Trak ProW(FreeScan TE25 W)

Radiated Emission Measurement

Test Date: 2024/11/6 08:45:48



No.	Frequency (MHz)	Reading (dBuV)	Factor (dB/m)	Level (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector	Height (cm)	Azimuth (deg.)	P/F	Remark
1	5146.000	53.43	-19.17	34.26	74.00	-39.74	peak	150	78	P	
2 *	5150.000	49.32	-19.16	30.16	68.82	-38.66	peak	150	105	P	

Note: *:Maximum data x:Over limit !:over margin Level = Reading + Factor Margin = Level - Limit

Appendix B
CN24K76Q 003

Prüfbericht - Produkte
Test Report - Products

Page 228 of 311




Shenzhen PSI Testing Co., Ltd.
1-2F, Building 5, Yudafu Industrial Park, No. 10, Xingye West Road,
Shajing Street, Bao'an District, Shenzhen, Guangdong, China 518104

Test Site: PSI EMC Lab-3m Chamber Polarization: **Vertical** Temperature: 26 °C
EUT: See note Distance: 3m Humidity: 54 %RH
M/N: See note Test Engineer: Felix Pang
Test Mode: a 5240MHz(Ant 1)
Test Voltage: DC 24V
Note: Optical 3D Measuring and Dynamic Tracking System
FreeScan Trak ProW(FreeScan TE25 W)

Radiated Emission Measurement

Test Date: 2024/11/6 08:49:18



No.	Frequency (MHz)	Reading (dBuV)	Factor (dB/m)	Level (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector	Height (cm)	Azimuth (deg.)	P/F	Remark
1 *	5350.000	47.16	-18.63	28.53	68.82	-40.29	peak	150	24	P	
2	5353.100	48.75	-18.61	30.14	74.00	-43.86	peak	150	135	P	

Note: *:Maximum data x:Over limit !:over margin Level = Reading + Factor Margin = Level - Limit

Appendix B
CN24K76Q 003

Prüfbericht - Produkte
Test Report - Products

Page 229 of 311




Shenzhen PSI Testing Co., Ltd.
1-2F, Building 5, Yudafu Industrial Park, No. 10, Xingye West Road,
Shajing Street, Bao'an District, Shenzhen, Guangdong, China 518104

Test Site: PSI EMC Lab-3m Chamber Polarization: **Horizontal** Temperature: 26 °C
EUT: See note Distance: 3m Humidity: 54 %RH
M/N: See note Test Engineer: Felix Pang
Test Mode: a 5240MHz(Ant 1)
Test Voltage: DC 24V
Note: Optical 3D Measuring and Dynamic Tracking System
FreeScan Trak ProW(FreeScan TE25 W)

Radiated Emission Measurement

Test Date: 2024/11/6 08:53:38



No.	Frequency (MHz)	Reading (dBuV)	Factor (dB/m)	Level (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector	Height (cm)	Azimuth (deg.)	P/F	Remark
1 *	5350.000	47.23	-17.98	29.25	68.82	-39.57	peak	150	46	P	
2	5355.350	48.91	-17.96	30.95	74.00	-43.05	peak	150	217	P	

Note: *:Maximum data x:Over limit !:over margin Level = Reading + Factor Margin = Level - Limit

Appendix B
CN24K76Q 003

Prüfbericht - Produkte
Test Report - Products

Page 230 of 311




Shenzhen PSI Testing Co., Ltd.
1-2F, Building 5, Yudafu Industrial Park, No. 10, Xingye West Road,
Shajing Street, Bao'an District, Shenzhen, Guangdong, China 518104

Test Site: PSI EMC Lab-3m Chamber Polarization: **Horizontal** Temperature: 26 °C
EUT: See note Distance: 3m Humidity: 54 %RH
M/N: See note Test Engineer: Felix Pang
Test Mode: n20 5180MHz(MIMO)
Test Voltage: DC 24V
Note: Optical 3D Measuring and Dynamic Tracking System
FreeScan Trak ProW(FreeScan TE25 W)

Radiated Emission Measurement

Test Date: 2024/11/6 08:58:09



No.	Frequency (MHz)	Reading (dBuV)	Factor (dB/m)	Level (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector	Height (cm)	Azimuth (deg.)	P/F	Remark
1	5144.000	55.38	-18.32	37.06	74.00	-36.94	peak	150	24	P	
2 *	5150.000	60.01	-18.31	41.70	68.82	-27.12	peak	150	125	P	

Note: *:Maximum data x:Over limit !:over margin Level = Reading + Factor Margin = Level - Limit

Appendix B
CN24K76Q 003

Prüfbericht - Produkte
Test Report - Products

Page 231 of 311




Shenzhen PSI Testing Co., Ltd.
1-2F, Building 5, Yudafu Industrial Park, No. 10, Xingye West Road,
Shajing Street, Bao'an District, Shenzhen, Guangdong, China 518104

Test Site: PSI EMC Lab-3m Chamber Polarization: **Vertical** Temperature: 26 °C
EUT: See note Distance: 3m Humidity: 54 %RH
M/N: See note Test Engineer: Felix Pang
Test Mode: n20 5180MHz(MIMO)
Test Voltage: DC 24V
Note: Optical 3D Measuring and Dynamic Tracking System
FreeScan Trak ProW(FreeScan TE25 W)

Radiated Emission Measurement

Test Date: 2024/11/6 09:03:12



No.	Frequency (MHz)	Reading (dBuV)	Factor (dB/m)	Level (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector	Height (cm)	Azimuth (deg.)	P/F	Remark
1	5147.200	56.64	-19.17	37.47	74.00	-36.53	peak	150	78	P	
2 *	5150.000	55.75	-19.16	36.59	68.82	-32.23	peak	150	205	P	

Note: *:Maximum data x:Over limit !:over margin Level = Reading + Factor Margin = Level - Limit

Appendix B
CN24K76Q 003

Prüfbericht - Produkte
Test Report - Products

Page 232 of 311




Shenzhen PSI Testing Co., Ltd.
1-2F, Building 5, Yudafu Industrial Park, No. 10, Xingye West Road,
Shajing Street, Bao'an District, Shenzhen, Guangdong, China 518104

Test Site: PSI EMC Lab-3m Chamber Polarization: **Vertical** Temperature: 26 °C
EUT: See note Distance: 3m Humidity: 54 %RH
M/N: See note Test Engineer: Felix Pang
Test Mode: n20 5240MHz(MIMO)
Test Voltage: DC 24V
Note: Optical 3D Measuring and Dynamic Tracking System
FreeScan Trak ProW(FreeScan TE25 W)

Radiated Emission Measurement

Test Date: 2024/11/6 09:08:05



No.	Frequency (MHz)	Reading (dBuV)	Factor (dB/m)	Level (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector	Height (cm)	Azimuth (deg.)	P/F	Remark
1 *	5350.000	48.68	-18.63	30.05	68.82	-38.77	peak	150	205	P	
2	5352.050	51.49	-18.61	32.88	74.00	-41.12	peak	150	38	P	

Note: *:Maximum data x:Over limit !:over margin Level = Reading + Factor Margin = Level - Limit

Appendix B
CN24K76Q 003
Page 233 of 311

Prüfbericht - Produkte
Test Report - Products




Shenzhen PSI Testing Co., Ltd.
1-2F, Building 5, Yudafu Industrial Park, No. 10, Xingye West Road,
Shajing Street, Bao'an District, Shenzhen, Guangdong, China 518104

Test Site: PSI EMC Lab-3m Chamber Polarization: **Horizontal** Temperature: 26 °C
EUT: See note Distance: 3m Humidity: 54 %RH
M/N: See note Test Engineer: Felix Pang
Test Mode: n20 5240MHz(MIMO)
Test Voltage: DC 24V
Note: Optical 3D Measuring and Dynamic Tracking System
FreeScan Trak ProW(FreeScan TE25 W)

Radiated Emission Measurement

Test Date: 2024/11/6 09:13:06



No.	Frequency (MHz)	Reading (dBuV)	Factor (dB/m)	Level (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector	Height (cm)	Azimuth (deg.)	P/F	Remark
1 *	5350.000	48.50	-17.98	30.52	68.82	-38.30	peak	150	105	P	
2	5352.650	50.42	-17.96	32.46	74.00	-41.54	peak	150	214	P	

Note: *:Maximum data x:Over limit !:over margin Level = Reading + Factor Margin = Level - Limit

Appendix B
CN24K76Q 003

Prüfbericht - Produkte
Test Report - Products

Page 234 of 311




Shenzhen PSI Testing Co., Ltd.
1-2F, Building 5, Yudafu Industrial Park, No. 10, Xingye West Road,
Shajing Street, Bao'an District, Shenzhen, Guangdong, China 518104

Test Site: PSI EMC Lab-3m Chamber Polarization: **Horizontal** Temperature: 26 °C
EUT: See note Distance: 3m Humidity: 54 %RH
M/N: See note Test Engineer: Felix Pang
Test Mode: ac20 5180MHz(MIMO)
Test Voltage: DC 24V
Note: Optical 3D Measuring and Dynamic Tracking System
FreeScan Trak ProW(FreeScan TE25 W)

Radiated Emission Measurement

Test Date: 2024/11/6 09:18:16



No.	Frequency (MHz)	Reading (dBuV)	Factor (dB/m)	Level (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector	Height (cm)	Azimuth (deg.)	P/F	Remark
1	5144.600	54.95	-18.32	36.63	74.00	-37.37	peak	150	135	P	
2 *	5150.000	54.63	-18.31	36.32	68.82	-32.50	peak	150	274	P	

Note: *:Maximum data x:Over limit !:over margin Level = Reading + Factor Margin = Level - Limit

Appendix B
CN24K76Q 003

Prüfbericht - Produkte
Test Report - Products

Page 235 of 311




Shenzhen PSI Testing Co., Ltd.
1-2F, Building 5, Yudafu Industrial Park, No. 10, Xingye West Road,
Shajing Street, Bao'an District, Shenzhen, Guangdong, China 518104

Test Site: PSI EMC Lab-3m Chamber Polarization: **Vertical** Temperature: 26 °C
EUT: See note Distance: 3m Humidity: 54 %RH
M/N: See note Test Engineer: Felix Pang
Test Mode: ac20 5180MHz(MIMO)
Test Voltage: DC 24V
Note: Optical 3D Measuring and Dynamic Tracking System
FreeScan Trak ProW(FreeScan TE25 W)

Radiated Emission Measurement

Test Date: 2024/11/6 09:23:23



No.	Frequency (MHz)	Reading (dBuV)	Factor (dB/m)	Level (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector	Height (cm)	Azimuth (deg.)	P/F	Remark
1 *	5147.500	55.87	-19.17	36.70	74.00	-37.30	peak	150	78	P	
2	5150.000	49.82	-19.16	30.66	68.82	-38.16	peak	150	135	P	

Note: *:Maximum data x:Over limit !:over margin Level = Reading + Factor Margin = Level - Limit

Appendix B
CN24K76Q 003

Prüfbericht - Produkte
Test Report - Products

Page 236 of 311




Shenzhen PSI Testing Co., Ltd.
1-2F, Building 5, Yudafu Industrial Park, No. 10, Xingye West Road,
Shajing Street, Bao'an District, Shenzhen, Guangdong, China 518104

Test Site: PSI EMC Lab-3m Chamber Polarization: **Vertical** Temperature: 26 °C
EUT: See note Distance: 3m Humidity: 54 %RH
M/N: See note Test Engineer: Felix Pang
Test Mode: ac20 5240MHz(MIMO)
Test Voltage: DC 24V
Note: Optical 3D Measuring and Dynamic Tracking System
FreeScan Trak ProW(FreeScan TE25 W)

Radiated Emission Measurement

Test Date: 2024/11/6 09:28:49



No.	Frequency (MHz)	Reading (dBuV)	Factor (dB/m)	Level (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector	Height (cm)	Azimuth (deg.)	P/F	Remark
1 *	5350.000	48.41	-18.63	29.78	68.82	-39.04	peak	150	135	P	
2	5355.500	50.42	-18.60	31.82	74.00	-42.18	peak	150	147	P	

Note: *:Maximum data x:Over limit !:over margin Level = Reading + Factor Margin = Level - Limit

Appendix B
CN24K76Q 003

Prüfbericht - Produkte
Test Report - Products

Page 237 of 311




Shenzhen PSI Testing Co., Ltd.
1-2F, Building 5, Yudafu Industrial Park, No. 10, Xingye West Road,
Shajing Street, Bao'an District, Shenzhen, Guangdong, China 518104

Test Site: PSI EMC Lab-3m Chamber Polarization: **Horizontal** Temperature: 26 °C
EUT: See note Distance: 3m Humidity: 54 %RH
M/N: See note Test Engineer: Felix Pang
Test Mode: ac20 5240MHz(MIMO)
Test Voltage: DC 24V
Note: Optical 3D Measuring and Dynamic Tracking System
FreeScan Trak ProW(FreeScan TE25 W)

Radiated Emission Measurement

Test Date: 2024/11/6 09:32:57



No.	Frequency (MHz)	Reading (dBuV)	Factor (dB/m)	Level (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector	Height (cm)	Azimuth (deg.)	P/F	Remark
1 *	5350.000	48.90	-17.98	30.92	68.82	-37.90	peak	150	214	P	
2	5352.500	49.30	-17.96	31.34	74.00	-42.66	peak	150	138	P	

Note: *:Maximum data x:Over limit !:over margin Level = Reading + Factor Margin = Level - Limit

Appendix B
CN24K76Q 003
Page 238 of 311

Prüfbericht - Produkte
Test Report - Products

U-NII-3 Band:




Shenzhen PSI Testing Co., Ltd.
1-2F, Building 5, Yudafu Industrial Park, No. 10, Xingye West Road,
Shajing Street, Bao'an District, Shenzhen, Guangdong, China 518104

Test Site: PSI EMC Lab-3m Chamber Polarization: **Vertical** Temperature: 26 °C
EUT: See note Distance: 3m Humidity: 54 %RH
M/N: See note Test Engineer: Felix Pang
Test Mode: 11ax20 5785MHz(MIMO)
Test Voltage: DC 24V
Note: Optical 3D Measuring and Dynamic Tracking System
FreeScan Trak ProW(FreeScan TE25 W)

Radiated Emission Measurement

Test Date: 2024/11/5 12:59:36



No.	Frequency (MHz)	Reading (dBuV)	Factor (dB/m)	Level (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector	Height (cm)	Azimuth (deg.)	P/F	Remark
1	7007.375	45.65	-8.03	37.62	74.00	-36.38	peak	150	157	P	
2 *	8229.250	44.63	-6.94	37.69	74.00	-36.31	peak	150	48	P	
3	9999.375	41.86	-5.61	36.25	74.00	-37.75	peak	150	67	P	
4	11227.625	38.26	-4.27	33.99	74.00	-40.01	peak	150	209	P	
5	12411.250	36.45	-3.68	32.77	74.00	-41.23	peak	150	83	P	
6	14942.125	31.69	-2.19	29.50	74.00	-44.50	peak	150	256	P	

Note: *:Maximum data x:Over limit !:over margin Level = Reading + Factor Margin = Level - Limit

Appendix B
CN24K76Q 003

Prüfbericht - Produkte
Test Report - Products

Page 239 of 311




Shenzhen PSI Testing Co., Ltd.
1-2F, Building 5, Yudafu Industrial Park, No. 10, Xingye West Road,
Shajing Street, Bao'an District, Shenzhen, Guangdong, China 518104

Test Site: PSI EMC Lab-3m Chamber Polarization: **Horizontal** Temperature: 26 °C
EUT: See note Distance: 3m Humidity: 54 %RH
M/N: See note Test Engineer: Felix Pang
Test Mode: 11ax20 5785MHz(MIMO)
Test Voltage: DC 24V
Note: Optical 3D Measuring and Dynamic Tracking System
FreeScan Trak ProW(FreeScan TE25 W)

Radiated Emission Measurement

Test Date: 2024/11/5 13:04:03



No.	Frequency (MHz)	Reading (dBuV)	Factor (dB/m)	Level (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector	Height (cm)	Azimuth (deg.)	P/F	Remark
1	7304.875	44.80	-7.78	37.02	74.00	-36.98	peak	150	125	P	
2 *	9387.375	43.06	-5.97	37.09	74.00	-36.91	peak	150	324	P	
3	10770.750	38.57	-4.60	33.97	74.00	-40.03	peak	150	265	P	
4	12589.750	33.99	-3.43	30.56	74.00	-43.44	peak	150	207	P	
5	14491.625	31.05	-2.12	28.93	74.00	-45.07	peak	150	168	P	
6	16531.625	31.69	-2.97	28.72	74.00	-45.28	peak	150	78	P	

Note: *:Maximum data x:Over limit !:over margin Level = Reading + Factor Margin = Level - Limit

Appendix B

CN24K76Q 003

Prüfbericht - Produkte
Test Report - Products

Page 240 of 311




Shenzhen PSI Testing Co., Ltd.
1-2F, Building 5, Yudafu Industrial Park, No. 10, Xingye West Road,
Shajing Street, Bao'an District, Shenzhen, Guangdong, China 518104

Test Site: PSI EMC Lab-3m Chamber Polarization: **Horizontal** Temperature: 26 °C
EUT: See note Distance: 3m Humidity: 54 %RH
M/N: See note Test Engineer: Felix Pang
Test Mode: 11ax20 5825MHz(MIMO)
Test Voltage: DC 24V
Note: Optical 3D Measuring and Dynamic Tracking System
FreeScan Trak ProW(FreeScan TE25 W)

Radiated Emission Measurement

Test Date: 2024/11/5 13:09:27



No.	Frequency (MHz)	Reading (dBuV)	Factor (dB/m)	Level (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector	Height (cm)	Azimuth (deg.)	P/F	Remark
1	7226.250	45.12	-7.85	37.27	74.00	-36.73	peak	150	156	P	
2 *	8082.625	44.41	-7.08	37.33	74.00	-36.67	peak	150	78	P	
3	8907.125	42.57	-6.29	36.28	74.00	-37.72	peak	150	357	P	
4	10371.250	39.38	-5.13	34.25	74.00	-39.75	peak	150	205	P	
5	11229.750	37.70	-4.27	33.43	74.00	-40.57	peak	150	68	P	
6	12419.750	34.02	-3.67	30.35	74.00	-43.65	peak	150	157	P	

Note: *:Maximum data x:Over limit !:over margin Level = Reading + Factor Margin = Level - Limit

Appendix B
CN24K76Q 003

Prüfbericht - Produkte
Test Report - Products

Page 241 of 311




Shenzhen PSI Testing Co., Ltd.
1-2F, Building 5, Yudafu Industrial Park, No. 10, Xingye West Road,
Shajing Street, Bao'an District, Shenzhen, Guangdong, China 518104

Test Site: PSI EMC Lab-3m Chamber Polarization: **Vertical** Temperature: 26 °C
EUT: See note Distance: 3m Humidity: 54 %RH
M/N: See note Test Engineer: Felix Pang
Test Mode: 11ax20 5825MHz(MIMO)
Test Voltage: DC 24V
Note: Optical 3D Measuring and Dynamic Tracking System
FreeScan Trak ProW(FreeScan TE25 W)

Radiated Emission Measurement

Test Date: 2024/11/5 13:12:55



No.	Frequency (MHz)	Reading (dBuV)	Factor (dB/m)	Level (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector	Height (cm)	Azimuth (deg.)	P/F	Remark
1 *	7001.000	45.72	-8.04	37.68	74.00	-36.32	peak	150	135	P	
2	8318.500	44.45	-6.86	37.59	74.00	-36.41	peak	150	35	P	
3	9285.375	42.20	-6.03	36.17	74.00	-37.83	peak	150	257	P	
4	10645.375	39.24	-4.77	34.47	74.00	-39.53	peak	150	68	P	
5	11814.125	36.32	-4.23	32.09	74.00	-41.91	peak	150	157	P	
6	13801.000	30.26	-2.22	28.04	74.00	-45.96	peak	150	204	P	

Note: *:Maximum data x:Over limit !:over margin Level = Reading + Factor Margin = Level - Limit

Appendix B
CN24K76Q 003

Prüfbericht - Produkte
Test Report - Products

Page 242 of 311




Shenzhen PSI Testing Co., Ltd.
1-2F, Building 5, Yudafu Industrial Park, No. 10, Xingye West Road,
Shajing Street, Bao'an District, Shenzhen, Guangdong, China 518104

Test Site: PSI EMC Lab-3m Chamber Polarization: **Vertical** Temperature: 26 °C
EUT: See note Distance: 3m Humidity: 54 %RH
M/N: See note Test Engineer: Felix Pang
Test Mode: 11n40 5755MHz(MIMO)
Test Voltage: DC 24V
Note: Optical 3D Measuring and Dynamic Tracking System
FreeScan Trak ProW(FreeScan TE25 W)

Radiated Emission Measurement

Test Date: 2024/11/5 13:16:52



No.	Frequency (MHz)	Reading (dBuV)	Factor (dB/m)	Level (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector	Height (cm)	Azimuth (deg.)	P/F	Remark
1	4693.250	51.12	-13.65	37.47	74.00	-36.53	peak	150	124	P	
2 *	7249.625	46.04	-7.83	38.21	74.00	-35.79	peak	150	324	P	
3	8231.375	44.72	-6.95	37.77	74.00	-36.23	peak	150	257	P	
4	9223.750	43.39	-6.07	37.32	74.00	-36.68	peak	150	78	P	
5	10443.500	40.85	-5.03	35.82	74.00	-38.18	peak	150	167	P	
6	12368.750	34.63	-3.74	30.89	74.00	-43.11	peak	150	49	P	

Note: *:Maximum data x:Over limit !:over margin Level = Reading + Factor Margin = Level - Limit

Appendix B
CN24K76Q 003

Prüfbericht - Produkte
Test Report - Products

Page 243 of 311




Shenzhen PSI Testing Co., Ltd.
1-2F, Building 5, Yudafu Industrial Park, No. 10, Xingye West Road,
Shajing Street, Bao'an District, Shenzhen, Guangdong, China 518104

Test Site: PSI EMC Lab-3m Chamber Polarization: **Horizontal** Temperature: 26 °C
EUT: See note Distance: 3m Humidity: 54 %RH
M/N: See note Test Engineer: Felix Pang
Test Mode: 11n40 5755MHz(MIMO)
Test Voltage: DC 24V
Note: Optical 3D Measuring and Dynamic Tracking System
FreeScan Trak ProW(FreeScan TE25 W)

Radiated Emission Measurement

Test Date: 2024/11/5 13:21:18



No.	Frequency (MHz)	Reading (dBuV)	Factor (dB/m)	Level (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector	Height (cm)	Azimuth (deg.)	P/F	Remark
1	4516.875	50.61	-14.26	36.35	74.00	-37.65	peak	150	246	P	
2 *	8280.250	44.61	-6.89	37.72	74.00	-36.28	peak	150	57	P	
3	6314.625	46.43	-9.98	36.45	74.00	-37.55	peak	150	247	P	
4	9950.500	40.90	-5.64	35.26	74.00	-38.74	peak	150	168	P	
5	11652.625	37.34	-4.26	33.08	74.00	-40.92	peak	150	92	P	
6	13352.625	32.14	-2.59	29.55	74.00	-44.45	peak	150	154	P	

Note: *:Maximum data x:Over limit !:over margin Level = Reading + Factor Margin = Level - Limit

Appendix B
CN24K76Q 003

Prüfbericht - Produkte
Test Report - Products

Page 244 of 311




Shenzhen PSI Testing Co., Ltd.
1-2F, Building 5, Yudafu Industrial Park, No. 10, Xingye West Road,
Shajing Street, Bao'an District, Shenzhen, Guangdong, China 518104

Test Site: PSI EMC Lab-3m Chamber Polarization: **Horizontal** Temperature: 26 °C
EUT: See note Distance: 3m Humidity: 54 %RH
M/N: See note Test Engineer: Felix Pang
Test Mode: 11n40 5795MHz(MIMO)
Test Voltage: DC 24V
Note: Optical 3D Measuring and Dynamic Tracking System
FreeScan Trak ProW(FreeScan TE25 W)

Radiated Emission Measurement

Test Date: 2024/11/5 13:26:28



No.	Frequency (MHz)	Reading (dBuV)	Factor (dB/m)	Level (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector	Height (cm)	Azimuth (deg.)	P/F	Remark
1	6911.750	46.46	-8.31	38.15	74.00	-35.85	peak	150	138	P	
2 *	8278.125	45.21	-6.90	38.31	74.00	-35.69	peak	150	241	P	
3	9266.250	42.88	-6.05	36.83	74.00	-37.17	peak	150	267	P	
4	11195.750	38.30	-4.28	34.02	74.00	-39.98	peak	150	109	P	
5	13454.625	32.12	-2.50	29.62	74.00	-44.38	peak	150	98	P	
6	16501.875	31.30	-2.98	28.32	74.00	-45.68	peak	150	251	P	

Note: *:Maximum data x:Over limit !:over margin Level = Reading + Factor Margin = Level - Limit

Appendix B
CN24K76Q 003

Prüfbericht - Produkte
Test Report - Products

Page 245 of 311




Shenzhen PSI Testing Co., Ltd.
1-2F, Building 5, Yudafu Industrial Park, No. 10, Xingye West Road,
Shajing Street, Bao'an District, Shenzhen, Guangdong, China 518104

Test Site: PSI EMC Lab-3m Chamber Polarization: **Vertical** Temperature: 26 °C
EUT: See note Distance: 3m Humidity: 54 %RH
M/N: See note Test Engineer: Felix Pang
Test Mode: 11n40 5795MHz(MIMO)
Test Voltage: DC 24V
Note: Optical 3D Measuring and Dynamic Tracking System
FreeScan Trak ProW(FreeScan TE25 W)

Radiated Emission Measurement

Test Date: 2024/11/5 13:31:55



No.	Frequency (MHz)	Reading (dBuV)	Factor (dB/m)	Level (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector	Height (cm)	Azimuth (deg.)	P/F	Remark
1	4852.625	50.94	-13.93	37.01	74.00	-36.99	peak	150	157	P	
2 *	7001.000	46.84	-8.04	38.80	74.00	-35.20	peak	150	217	P	
3	8244.125	44.55	-6.93	37.62	74.00	-36.38	peak	150	168	P	
4	9419.250	42.35	-5.96	36.39	74.00	-37.61	peak	150	92	P	
5	10845.125	37.81	-4.51	33.30	74.00	-40.70	peak	150	205	P	
6	12795.875	32.51	-3.16	29.35	74.00	-44.65	peak	150	43	P	

Note: *:Maximum data x:Over limit !:over margin Level = Reading + Factor Margin = Level - Limit

Appendix B
CN24K76Q 003

Prüfbericht - Produkte
Test Report - Products

Page 246 of 311




Shenzhen PSI Testing Co., Ltd.
1-2F, Building 5, Yudafu Industrial Park, No. 10, Xingye West Road,
Shajing Street, Bao'an District, Shenzhen, Guangdong, China 518104

Test Site: PSI EMC Lab-3m Chamber Polarization: **Vertical** Temperature: 26 °C
EUT: See note Distance: 3m Humidity: 54 %RH
M/N: See note Test Engineer: Felix Pang
Test Mode: 11ac40 5755MHz(MIMO)
Test Voltage: DC 24V
Note: Optical 3D Measuring and Dynamic Tracking System
FreeScan Trak ProW(FreeScan TE25 W)

Radiated Emission Measurement

Test Date: 2024/11/5 13:34:36



No.	Frequency (MHz)	Reading (dBuV)	Factor (dB/m)	Level (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector	Height (cm)	Azimuth (deg.)	P/F	Remark
1	4038.750	52.54	-15.96	36.58	74.00	-37.42	peak	150	108	P	
2 *	7228.375	45.56	-7.85	37.71	74.00	-36.29	peak	150	92	P	
3	5662.250	48.44	-11.43	37.01	74.00	-36.99	peak	150	304	P	
4	9043.125	43.84	-6.19	37.65	74.00	-36.35	peak	150	247	P	
5	10796.250	38.89	-4.56	34.33	74.00	-39.67	peak	150	250	P	
6	13758.500	30.45	-2.25	28.20	74.00	-45.80	peak	150	189	P	

Note: *:Maximum data x:Over limit !:over margin Level = Reading + Factor Margin = Level - Limit

Appendix B
CN24K76Q 003

Prüfbericht - Produkte
Test Report - Products

Page 247 of 311




Shenzhen PSI Testing Co., Ltd.
1-2F, Building 5, Yudafu Industrial Park, No. 10, Xingye West Road,
Shajing Street, Bao'an District, Shenzhen, Guangdong, China 518104

Test Site: PSI EMC Lab-3m Chamber Polarization: **Horizontal** Temperature: 26 °C
EUT: See note Distance: 3m Humidity: 54 %RH
M/N: See note Test Engineer: Felix Pang
Test Mode: 11ac40 5755MHz(MIMO)
Test Voltage: DC 24V
Note: Optical 3D Measuring and Dynamic Tracking System
FreeScan Trak ProW(FreeScan TE25 W)

Radiated Emission Measurement

Test Date: 2024/11/5 13:39:03



No.	Frequency (MHz)	Reading (dBuV)	Factor (dB/m)	Level (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector	Height (cm)	Azimuth (deg.)	P/F	Remark
1 *	5649.500	49.30	-11.47	37.83	74.00	-36.17	peak	150	201	P	
2	7005.250	45.70	-8.04	37.66	74.00	-36.34	peak	150	247	P	
3	8061.375	44.58	-7.11	37.47	74.00	-36.53	peak	150	268	P	
4	9121.750	43.19	-6.13	37.06	74.00	-36.94	peak	150	92	P	
5	10458.375	39.49	-5.00	34.49	74.00	-39.51	peak	150	106	P	
6	11499.625	37.86	-4.26	33.60	74.00	-40.40	peak	150	45	P	

Note: *:Maximum data x:Over limit !:over margin Level = Reading + Factor Margin = Level - Limit

Appendix B
CN24K76Q 003

Prüfbericht - Produkte
Test Report - Products

Page 248 of 311




Shenzhen PSI Testing Co., Ltd.
1-2F, Building 5, Yudafu Industrial Park, No. 10, Xingye West Road,
Shajing Street, Bao'an District, Shenzhen, Guangdong, China 518104

Test Site: PSI EMC Lab-3m Chamber Polarization: **Horizontal** Temperature: 26 °C
EUT: See note Distance: 3m Humidity: 54 %RH
M/N: See note Test Engineer: Felix Pang
Test Mode: 11ac40 5795MHz(MIMO)
Test Voltage: DC 24V
Note: Optical 3D Measuring and Dynamic Tracking System
FreeScan Trak ProW(FreeScan TE25 W)

Radiated Emission Measurement

Test Date: 2024/11/5 13:43:01



No.	Frequency (MHz)	Reading (dBuV)	Factor (dB/m)	Level (dBuV/m)	Limit (dBuV/m)	Margin (dB)	Detector	Height (cm)	Azimuth (deg.)	P/F	Remark
1	7013.750	46.11	-8.03	38.08	74.00	-35.92	peak	150	125	P	
2 *	8558.625	45.34	-6.63	38.71	74.00	-35.29	peak	150	307	P	
3	9364.000	42.88	-5.99	36.89	74.00	-37.11	peak	150	219	P	
4	10464.750	40.05	-4.99	35.06	74.00	-38.94	peak	150	68	P	
5	11529.375	38.83	-4.25	34.58	74.00	-39.42	peak	150	247	P	
6	12521.750	35.82	-3.54	32.28	74.00	-41.72	peak	150	130	P	

Note: *:Maximum data x:Over limit !:over margin Level = Reading + Factor Margin = Level - Limit