

USER MANUAL

MODEL: CR3

Warning: Changes or modifications to this unit not expressly approved by the party responsible for compliance could void the user's authority to operate the equipment.

NOTE: This device complies with part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) This device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

FCC Statement: This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.

Content

1. Overview	1
2. Operating instructions	3
3. System function description	7
4. Daily maintenance	9
5. Technical parameters	10

1. Overview

Mini ColorReader is a portable small color measuring instrument with small size, convenient use and accurate measurement. It is widely used in industries such as plastic electronics, paint and coating, textile and garment printing and dyeing.

The instrument can be connected to the mobile phone APP and SQCX PC quality management software to achieve more functional expansion. The mobile APP comes with multiple sets of electronic color cards, which can realize color card retrieval and matching. The mobile phone APP and SQCX PC software have a variety of light source, color space, color index, and reflectance functions, which can realize massive chromaticity index testing.

This instrument has the following advantages:

- 1) It can be measured after power-on, no need to perform white calibration every time, making color measurement easier;
- 2) Separate design of calibration whiteboard and measuring port to ensure long service life of whiteboard and accurate measurement;
- 3) Accurate, stable, small size, easy to carry, APP built-in mobile electronic color card, it is a good helper for mobile color quality control;
- 4) A variety of chromaticity parameters can be set on the instrument side, test records can be checked, and the stand-alone function is powerful;
- 5) Built-in rechargeable lithium battery, Type-C USB interface, easy to use;
- 6) Connect the mobile phone APP and PC SQCX software to achieve more functional expansion.

1. Button and interface description




Figure 1 Schematic diagram of button interface (front)




Figure 2 Schematic diagram of button interface (rear)




Figure 3 Schematic diagram of calibration cover

"Operation button" icon description:

(←) Left button: swipe left, left button to record/up select:

(→) Right key: Swipe to the right, right key to record/down select:

(■) Confirm: Short press to return or switch sample/standard sample, long press to confirm.

2. Operating instructions

2.1 Battery instructions

This instrument uses a built-in lithium battery, the specification of the lithium battery is Li-ion 3.7V, and the capacity is 800mAh. Use a USB data cable to connect to the USB port on the PC or connect to a 5V==1A power adapter to charge the lithium battery. The maximum charging current can reach 500MA. When charging, the battery level is greater than 80%, the LED indicator is green, the battery level is less than 20%, the LED indicator is red, and the battery level is between 20% and 80% is yellow. The display shows a charging prompt, and the charging is complete, indicating that the battery is full.

2.2 Switch on and off

Power on: In the power off state, press the power switch, the instrument is powered on, the LED indicator flashes, the display screen lights up and enters the measurement interface.

Shutdown: In the power-on state, long press the power switch to shut down the instrument. When the automatic sleep mode is turned on, the instrument will automatically shut down.

Note: When not using the instrument for a long time, please turn off the power.

2.3 Black and white calibration

In the measurement interface, long press the "operation button" to enter the system function interface, slide the "operation button" to the left or right, position the cursor on the "black and white calibration", and long press the "operation button" to enter the calibration interface. The instrument has built-in calibration parameters, and calibration is needed when the environment changes drastically.

Black calibration: "black calibration" the instrument against the air, the surrounding must be a darker environment without bright light source, and there is no obstruction within 1 meter of the instrument against the air. Short press the "measurement button" to perform black calibration, and the calibration success prompts "pass". After the black calibration is completed, the white calibration interface is automatically entered.

White calibration: align the measurement aperture with the calibration cover. The calibration cover adopts a spring-type pressing structure, gently press the instrument to make the measurement aperture fit the whiteboard closely, short press the "measurement button" for white calibration, the calibration is successful, and the prompt "pass". After the white calibration is completed, the measurement interface is automatically entered.

2.4 Mobile APP installation and measurement

The mobile APP supports Android operating system version 5.0 and above. Use the mobile browser or WeChat to scan the QR code below, download and install as prompted.



Figure 4 QR code for APP installation

Figure 5 App icon after installation

After the installation is complete, open the mobile APP (see the mobile APP manual for details on the use of the mobile APP), click "Connect Bluetooth" or "Connect to the instrument", first check the instrument model and SN code in the "Device Information" of the instrument, and then in the mobile APP Select the instrument to be connected (instrument model: SN code) in the "Connected Instruments" list, and click to automatically connect. In the color measurement interface, click "standard sample measurement" or "take standard sample" to obtain the chromaticity data and simulated color of the standard sample. Click "Measure Color Difference" to obtain the chromaticity data, difference and artificial color of the sample, as shown in Figures 6 and 7.




Figure 6 APP measurement after connected



Figure 7 Instrument measurement interface after connected APP

2.5 Quality management software installation

The instrument can establish a connection with the PC-side SQCX quality management software through a USB data cable or a Bluetooth module (only for products equipped with a Bluetooth module) to communicate and achieve more functional expansion. The measurement interface of the instrument connected to the PC is shown in Figure 8.




Figure 8 The measurement interface of the instrument connected to the PC

2.6 Measurement

Standard sample measurement and sample measurement can be realized by instruments, and other chromaticity indexes can be realized by mobile phone APP and SQCX PC software.

2.6.1 Stand-alone standard sample measurement

In the standard sample measurement interface, align the measurement aperture to the standard sample and press the "measurement button" shortly. The instrument obtains the chromaticity data of the standard sample, as shown in Figure 9. The green indicator LED flashes once when the measurement is completed. The status bar at the top of the instrument screen displays the current measurement conditions, the icon at the bottom displays the USB and Bluetooth connection status, and "01" in the lower right corner indicates the current standard number. The number of the stand-alone standard samples is from 01 to 10. When the standard sample data is full of 10, the system will prompt whether to clear the standard sample again. After clearing, the measurement can be continued. After the standard sample is cleared, the sample associated with the standard sample will also be cleared. Slide the operation button to the left or right to view the target data.




Figure 9 Instrument standard sample measurement

2.6.2 Single machine sample measurement

In the standard sample measurement interface, short press the "operation button" to enter the sample measurement interface associated with the standard sample, as shown in Figure 10. Align the measuring aperture with the sample and press the "measurement button" shortly, and the instrument will obtain the chromaticity data of the sample. The green indicator LED flashes once when the measurement is completed. The "01" on the left in the lower right corner of the instrument screen represents the standard sample number associated with the sample, and the 01 on the right represents the current sample number, and the sample number ranges from 01-99. After 99 pieces of sample data are full, the system prompts whether to clear the sample again, and the measurement can be continued after the sample is cleared. Slide the operation button to the left or right to view the sample data.

The default DE tolerance of the instrument is 1.0. When the color difference is within the tolerance range, it will prompt "Qualified", otherwise, it will prompt "Bad". The instrument DE tolerance can be set via APP and SQCX.




Figure 10 Instrument sample measurement

2.7 Mini printer

The mini printer is a non-standard accessory and needs to be purchased separately. The instrument is connected to a specially equipped mini printer, which can print measurement data.

3. System function description

Long press the "operation button" in the measurement interface to enter the system function interface. Slide the "operating button" to the left or right to select the system function, and long press the "confirm" button to make the corresponding settings.

3.1 Data deletion

Data deletion can delete measurement data.

Select " Standard Delete" and long press the "Operation Button" to delete all standard sample records and sample records in the instrument.

Select "Sample Delete" and long press the "Operation Button" to delete all sample records in the instrument and keep the standard sample records.

Note: After the record is deleted, the data cannot be recovered, please operate carefully to prevent accidentally deleting the required history record.

3.2 Light source settings

The light source setting includes the observer angle and the light source setting.

The observer angle can choose 2° and 10° (different models have differences);

The light source can choose D65, A, C, D50, D55, D75, F1, F2, F3, F4, F5, F6, F7, F8, F9, F10, F11, F12 (different models have differences).

3.3 color space

The color space can choose CIE LAB, XYZ, Yxy, LCh, CIE LUV, s-RGB, HunterLab, β_{xy} , DIN Lab99. The default color space of the instrument is CIE LAB (different models are different).

3.4 Color difference formula

The color difference formula interface can choose DE*ab, DE*uv, DE*94, DE*cmc2:1, DE*cmc1:1, DE*00, DIN De99, DE(Hunter) (different models have differences).

3.5 System Settings

3.5.1 Auto save

When auto-save is turned on, the measurement result will be automatically stored in the instrument for each measurement. Otherwise, the sample measurement will end and the measurement record will not be automatically saved.

3.5.2 Sound

When the buzzer sound is turned on, a reminder sound will sound every time you operate, otherwise there will be no reminder sound.

3.5.3 Bluetooth

For instruments equipped with Bluetooth, it can communicate with mobile phone APP or PC-side quality management software through Bluetooth.

3.5.4 Automatic printing

When automatic printing is turned on, the sample will be automatically printed every time it is measured. Otherwise, the sample will not be automatically printed after the measurement is completed, and you need to manually click to print.

3.5.5 Calibration

Calibration (calibration threshold enable) is used to set whether to enable calibration threshold inspection. When set to off, when the instrument performs black and white calibration, the corrected data will not be checked; when set to on, when the instrument performs black and white calibration, the data collected by the detector will be checked to meet the verification rules. Calibration can pass.

The calibration threshold enable will not affect the test results, and the instrument does not need frequent black and white calibration. It is recommended to perform black and white calibration only when it is used for the first time, when the temperature difference environment has changed significantly, after a long period of unused use, or when the measurement data of the instrument is found to be inaccurate.

3.5.6 Language selection

The system language can be Chinese or English.

3.5.7 Sleep time

Set the automatic shutdown time after the screen is turned off. The sleep time is divided into "30 seconds", "60 seconds", "90 seconds", and "no sleep". If you choose not to sleep, the instrument will not automatically shut down. If it is set to "30 seconds", the instrument will count when the screen is off, and will shut down after 30 seconds and enter the power saving mode. The meanings of "60 seconds" and "90 seconds" are the same as above. When the instrument is sleeping, you can start the instrument by short pressing the operation button.

3.5.8 Backlight time

Set the display lighting time, the time is divided into "30 seconds", "60 seconds", "90 seconds", "normally open". If normally open is selected, the screen will always be on. If it is set to "30 seconds", the instrument will count from the last operation, and the display will turn off after 30 seconds. The meanings of "60 seconds" and "90 seconds" are the same as above.

3.5.9 Factory reset

Select "Restore Factory", the instrument will clear all measurement records and user settings, and restore to the factory state.

Note: After restoring the factory settings, all data cannot be restored, please operate with caution.

3.5.10 Device information

Display instrument model, SN code, software and hardware version number and other information.

4. Daily maintenance

1) This instrument is a precision optical instrument. Please keep and use the instrument properly. Avoid using and storing the instrument in humid, strong electromagnetic interference, strong light, and dusty environments. It is recommended to use and store the instrument in a standard laboratory environment (temperature 20 degrees Celsius, 1 standard atmosphere, humidity 50~70%RH).

2) The whiteboard is a precision optical element. It should be kept and used properly. Avoid hitting the working surface with sharp objects, avoid contaminating the working surface with dirt, and avoid exposing the whiteboard under strong light. Clean the working surface of the

whiteboard with a wiping cloth dipped in alcohol regularly, and remove the dust on the working surface in time when correcting.

3) In order to ensure the validity of the measurement data, it is recommended that the complete instrument and the whiteboard should be inspected by the manufacturer or a qualified metrology institute for one year from the date of purchase.

4) Please do not disassemble the instrument privately. If you have any questions, please contact the relevant after-sales staff, tearing off the easy-to-tear label will affect the after-sales maintenance of the instrument.

5. Technical parameters

Optical Geometry	D/8 (Diffuse illuminance, 8°reception, SCI) , Conforms to standards CIE No.15, GB/T 3978.
Features	Used for color difference quality control in plastic electronics, paint and coating, textile and garment printing and dyeing, ceramics and other industries; Small size, easy to carry, search for the color cards and multifunction expanded through smartphone APP
Integrating Sphere Size	Φ20mm
Illuminant Device	Full spectrum LED light source
Spectro Method	Filter spectro (Some models do not have this function)
Sensor	CMOS sensor
Wavelength Range	400 ~ 700nm
Measuring Aperture	Φ8mm
SCI/SCE	SCI
Color Space	CIE LAB, XYZ, Yxy, LCh, CIE LUV, s-RGB, HunterLab, βxy, DIN Lab99 (Different models are different)
Color Difference Formula	$\Delta E^*ab, \Delta E^*uv, \Delta E^*94, \Delta E^*cmc(2:1), \Delta E^*cmc(1:1), \Delta E^*00, \text{DIN}\Delta E99, \Delta E(\text{Hunter})$ (Different models are different)

Other Color Index	WI(ASTM E313, CIE/ISO, AATCC, Hunter), YI(ASTM D1925, ASTM 313), Metamerism Index Mt , Color fastness, color strength, Opacity, 555 color classification, Munsell(C/2) (implemented by mobile APP) (Some models do not have this function)
Display Accuracy	0.01
Observer Angle	2/10° (Some models only support 10°)
Illuminants	D65, A, C, D50, D55, D75, F1, F2(CWF), F3, F4 , F5 , F6, F7(DLF), F8, F9 , F10(TPL5), F11(TL84), F12(TL83/U30) (Different models are different)
Display	Sample chromaticity value, color difference value, pass/fail result, color simulation, color deviation, reflectance (partly through mobile APP) (Some models don't support reflectance)
Measuring Time	About 1.0s
Repeatability	Color : ΔE^*ab within 0.05 (After warm-up calibration, the average value of measuring the whiteboard 30 times at an interval of 5s)
Accuracy	Pass National Metrology(Some models are not guaranteed)
Measuring Mode	Single measurement , Average measurement (2~99times)) (Implemented through smartphone App)
Locating Method	Spot positioning (Some models do not have this function)
Dimension	Ø30X100mm
Weight	88g
Battery	Lithium battery, can do tests continuously for 12000 times in a single full charge
Light source	5-years more than 3 million times measurements

lifespan	
Screen	IPS full color screen, 1.14inch
Port	Type-C USB ; Bluetooth 5.0 ; Buttons
Data Storage	Standards: 10pcs, Samples: 100pcs. Mass storage can be expanded through mobile APP (Different models are different)
Language	Chinese , English
Working Temperature range	0~40℃ , 0~85%RH (No condensation) , Altitude: lower than 2000m
Storage Temperature Range	-20~50℃ , 0~85%RH (No condensation)
Standard Accessories	Data cable, manual, calibration box, SQCX PC Color quality management software, MOBCCS APP (download from official website) (Some models don't have SQCX software)
Optional Accessories	USB mini printer, powder test box
Note	Technical parameters are for reference only, subject to actual instruments.