

Getwell Health Technology(Wuhu) Co., Ltd.

FCC Class II Permissive Change Request Letter

Date: 2018-11-12

To FCC:

RE: FCC Permissive II Change Request for Company: Getwell Health Technology(Wuhu) Co., Ltd. **FCC ID:** **2ANB8-T1**

We are submitting an application for a class II permissive change to the FCC approval of the Company name: Getwell Health Technology(Wuhu) Co., Ltd., product description: getwell Aerobic capacity assessment (FCC ID: **2ANB8-T1**, Original Grant Date: 09/03/2017). The transmitter module itself has not changed. Here are the changes:

The magnetic charging head 5PIN is changed to 3PIN, the prototype name was changed to GETWELL Aerobic capacity assessment and muscle oxygen monitoring.

For more detail, please see page 2&3.

Sincerely,



Signature



Feng Chao/Production Manager

E-mail: fengchao@igetwell.cn

Phone: 0513-5111086

Fax: 0513-5111086

Address: 4th Building, Service outsourcing park, new and high-tech zone,
Yijiang District, Wuhu, Anhui



New

