


Impedance Plot for D2450V2-SN: 968 2450 Head

Calibrated impedance: $51.9\Omega + 3.63j\Omega$; Measurement impedance: $53.1\Omega - 1.6j\Omega$ (within 5Ω)



Calibrated return loss: -27.9dB; Measurement return loss: -28.36dB (within 20%)

