# MEASUREMENT REPORT

## FCC PART 15.247 BLE Test Report

| | |
|---|---|
| **FCC ID:** | 2APAPVGM09 |
| **APPLICANT:** | VivaChek Biotech (Hangzhou) Co., Ltd. |
| **Application Type:** | Certification |
| **Product:** | Blood Glucose Meter |
| **Model No.:** | VGM09, VGM22 |
| **FCC Classification:** | Digital Transmission System (DTS) |
| **FCC Rule Part(s):** | Part 15 Subpart C (Section 15.247) |
| **Test Procedure(s):** | ANSI C63.10-2013, KDB 558074 D01v04 |
| **Test Date:** | March 06 ~ March 29, 2018 |

Reviewed By : ( Kevin Guo )

Approved By : ( Marlin Chen )

ACCREDITED
TESTING LABORATORY
CERTIFICATE #3628.01

The test results relate only to the samples tested.

This equipment has been shown to be capable of compliance with the applicable technical standards as indicated in the measurement report and was tested in accordance with the measurement procedures specified in KDB 558074 D01v04. Test results reported herein relate only to the item(s) tested.

The test report shall not be reproduced except in full without the written approval of MRT Technology (Suzhou) Co., Ltd.

# Revision History

| Report No. | Version | Description | Issue Date | Note |
|---|---|---|---|---|
| 1803WSU007-U1 | Rev. 01 | Initial Report | 03-30-2018 | Valid |
| | | | | |

# CONTENTS

Description Page

**1. INTRODUCTION ..... 6**

1.1. Scope ..... 6
1.2. MRT Test Location ..... 6

**2. PRODUCT INFORMATION ..... 7**

2.1. Feature of Equipment under Test ..... 7
2.2. Product Specification Subjective to this Report ..... 7
2.3. Working Frequencies ..... 7
2.4. Device Capabilities ..... 8
2.5. Test Configuration ..... 8
2.6. EMI Suppression Device(s)/Modifications ..... 8
2.7. Labeling Requirements ..... 8
2.8. Test Software ..... 8

**3. DESCRIPTION OF TEST ..... 9**

3.1. Evaluation Procedure ..... 9
3.2. AC Line Conducted Emissions ..... 9
3.3. Radiated Emissions ..... 10

**4. ANTENNA REQUIREMENTS ..... 11**

**5. TEST EQUIPMENT CALIBRATION DATE ..... 12**

**6. MEASUREMENT UNCERTAINTY ..... 13**

**7. TEST RESULT ..... 14**

7.1. Summary ..... 14
7.2. 6dB Bandwidth Measurement ..... 15
7.2.1. Test Limit ..... 15
7.2.2. Test Procedure used ..... 15
7.2.3. Test Setting ..... 15
7.2.4. Test Setup ..... 15
7.2.5. Test Result ..... 16
7.3. Output Power Measurement ..... 17
7.3.1. Test Limit ..... 17
7.3.2. Test Procedure Used ..... 17
7.3.3. Test Setting ..... 17
7.3.4. Test Setup ..... 17
7.3.5. Test Result of Output Power ..... 18

7.4. Power Spectral Density Measurement ..... 19
7.4.1. Test Limit ..... 19
7.4.2. Test Procedure Used ..... 19
7.4.3. Test Setting ..... 19
7.4.4. Test Setup ..... 19
7.4.5. Test Result ..... 20
7.5. Conducted Band Edge and Out-of-Band Emissions ..... 21
7.5.1. Test Limit ..... 21
7.5.2. Test Procedure Used ..... 21
7.5.3. Test Settitng ..... 21
7.5.4. Test Setup ..... 22
7.5.5. Test Result ..... 23
7.6. Radiated Spurious Emission Measurement ..... 25
7.6.1. Test Limit ..... 25
7.6.2. Test Procedure Used ..... 25
7.6.3. Test Setting ..... 25
7.6.4. Test Setup ..... 27
7.6.5. Test Result ..... 29
7.7. Radiated Restricted Band Edge Measurement ..... 34
7.7.1. Test Limit ..... 34
7.7.2. Test Procedure Used ..... 35
7.7.3. Test Setting ..... 35
7.7.4. Test Setup ..... 36
7.7.5. Test Result ..... 37
7.8. AC Conducted Emissions Measurement ..... 45
7.8.1. Test Limit ..... 45
7.8.2. Test Setup ..... 45
7.8.3. Test Result ..... 46

**8. CONCLUSION ..... 48**

## §2.1033 General Information

| | |
|---|---|
| **Applicant:** | VivaChek Biotech (Hangzhou) Co., Ltd. |
| **Applicant Address:** | Level 2, Block 2, 146 East Chaofeng Rd., Yuhang Economy Development Zone Hangzhou, Zhejiang Province 311100, P.R. China |
| **Manufacturer:** | VivaChek Biotech (Hangzhou) Co., Ltd. |
| **Manufacturer Address:** | Level 2, Block 2, 146 East Chaofeng Rd., Yuhang Economy Development Zone Hangzhou, Zhejiang Province 311100, P.R. China |
| **Test Site:** | MRT Technology (Suzhou) Co., Ltd |
| **Test Site Address:** | D8 Building, Youxin Industrial Park, No.2 Tian'edang Rd., Wuzhong Economic Development Zone, Suzhou, China |
| **FCC Registration No.:** | 893164 |
| **Test Device Serial No.:** | N/A ☐ Production ☒ Pre-Production ☐ Engineering |

### Test Facility / Accreditations

**Measurements were performed at MRT Laboratory located in Tian'edang Rd., Suzhou, China.**

- **MRT facility is a FCC registered (MRT Reg. No. 893164) test facility with the site description report on file and has met all the requirements specified in Section 2.948 of the FCC Rules.**
- **MRT facility is an IC registered (MRT Reg. No. 11384A-1) test laboratory with the site description on file at Industry Canada.**
- **MRT facility is a VCCI registered (R-4179, G-814, C-4664, T-2206) test laboratory with the site description on file at VCCI Council.**
- **MRT Lab is accredited to ISO 17025 by the American Association for Laboratory Accreditation (A2LA) under the American Association for Laboratory Accreditation Program (A2LA Cert. No. 3628.01) in EMC, Telecommunications and Radio testing for FCC, Industry Canada, EU and TELEC Rules.**

# 1. INTRODUCTION

## 1.1. Scope

Measurement and determination of electromagnetic emissions (EMC) of radio frequency devices including intentional and/or unintentional radiators for compliance with the technical rules and regulations of the Federal Communications Commission and the Industry Canada Certification and Engineering Bureau.

## 1.2. MRT Test Location

The map below shows the location of the MRT LABORATORY, its proximity to the Taihu Lake. These measurement tests were conducted at the MRT Technology (Suzhou) Co., Ltd. Facility located at D8 Building, Youxin Industrial Park, No.2 Tian'edang Rd., Wuzhong Economic Development Zone, Suzhou, China. The detailed description of the measurement facility was found to be in compliance with the requirements of § 2.948 according to ANSI C63.4-2009 on September 30, 2013.

# 2. PRODUCT INFORMATION

## 2.1. Feature of Equipment under Test

| | |
|---|---|
| Product Name: | Blood Glucose Meter |
| Model No.: | VGM09, VGM22 |
| Work Voltage | DC 3.7V / Battery |
| Bluetooth Specification | Bluetooth LE |

Note 1: There are the same Housing materials, PCB&PCBA and technology between VGM09 and VGM22, except the VGM09 has the VOICE and Backlight Function, but VGM22 does not.

Note 2: This report relate only to VGM09 tested.

## 2.2. Product Specification Subjective to this Report

| | |
|---|---|
| Bluetooth Frequency | 2402~2480MHz |
| Channel Number | 40 |
| Data Rate | 1Mbps(GFSK) |
| Antenna Type | PIFA Antenna |
| Antenna Gain | 2dBi |

## 2.3. Working Frequencies

Channel List for BLE

| Channel | Frequency | Channel | Frequency | Channel | Frequency |
|---|---|---|---|---|---|
| 00 | 2402 MHz | 01 | 2404 MHz | 02 | 2406 MHz |
| 03 | 2408 MHz | 04 | 2410 MHz | 05 | 2412 MHz |
| 06 | 2414 MHz | 07 | 2416 MHz | 08 | 2418 MHz |
| 09 | 2420 MHz | 10 | 2422 MHz | 11 | 2424 MHz |
| 12 | 2426 MHz | 13 | 2428 MHz | 14 | 2430 MHz |
| 15 | 2432 MHz | 16 | 2434 MHz | 17 | 2436 MHz |
| 18 | 2438 MHz | 19 | 2440 MHz | 20 | 2442 MHz |
| 21 | 2444 MHz | 22 | 2446 MHz | 23 | 2448 MHz |
| 24 | 2450 MHz | 25 | 2452 MHz | 26 | 2454 MHz |
| 27 | 2456 MHz | 28 | 2458 MHz | 29 | 2460 MHz |
| 30 | 2462 MHz | 31 | 2464 MHz | 32 | 2466 MHz |
| 33 | 2468 MHz | 34 | 2470 MHz | 35 | 2472 MHz |
| 36 | 2474 MHz | 37 | 2476 MHz | 38 | 2478 MHz |
| 39 | 2480 MHz | -- | -- | -- | -- |

## 2.4. Device Capabilities

This device contains the following capabilities:
Bluetooth LE.

## 2.5. Test Configuration

The device was tested per the guidance of KDB 558074 D01v04. ANSI C63.10-2013 was used to reference the appropriate EUT setup for radiated spurious emissions testing and AC line conducted testing.

## 2.6. EMI Suppression Device(s)/Modifications

No EMI suppression device(s) were added and/or no modifications were made during testing.

## 2.7. Labeling Requirements

Per 2.1074 & 15.19; Docket 95-19
The label shall be permanently affixed at a conspicuous location on the device; instruction manual or pamphlet supplied to the user and be readily visible to the purchaser at the time of purchase. However, when the device is so small wherein placement of the label with specified statement is not practical, only the FCC ID must be displayed on the device per Section 15.19(a)(5). Please see attachment for FCC ID label and label location.

## 2.8. Test Software

The test utility software used during testing was “sscom_51”.

# 3. DESCRIPTION OF TEST

## 3.1. Evaluation Procedure

The measurement procedures described in the American National Standard for Testing Unlicensed Wireless Devices (ANSI C63.10-2013), and the guidance provided in KDB 558074 D01v04 were used in the measurement of the device.

**Deviation from measurement procedure…….....None**

## 3.2. AC Line Conducted Emissions

The line-conducted facility is located inside an 8'x4'x4' shielded enclosure. A 1m x 2m wooden table 80cm high is placed 40cm away from the vertical wall and 80cm away from the sidewall of the shielded room. Two 10kHz-30MHz, 50Ω/50uH Line-Impedance Stabilization Networks (LISNs) are bonded to the shielded room floor. Power to the LISNs is filtered by external high-current high-insertion loss power line filters. These filters attenuate ambient signal noise from entering the measurement lines. These filters are also bonded to the shielded enclosure.

The EUT is powered from one LISN and the support equipment is powered from the second LISN. All interconnecting cables more than 1 meter were shortened to a 1 meter length by non-inductive bundling (serpentine fashion) and draped over the back edge of the test table. All cables were at least 40cm above the horizontal reference ground-plane. Power cables for support equipment were routed down to the second LISN while ensuring that that cables were not draped over the second LISN.

Sufficient time for the EUT, support equipment, and test equipment was allowed in order for them to warm up to their normal operating condition. The RF output of the LISN was connected to the receiver and exploratory measurements were made to determine the frequencies producing the maximum emission from the EUT. The receiver was scanned from 150kHz to 30MHz. The detector function was set to peak mode for exploratory measurements while the bandwidth of the analyzer was set to 9kHz. The EUT, support equipment, and interconnecting cables were arranged and manipulated to maximize each emission. Each emission was also maximized by varying: power lines, the mode of operation or data exchange speed, or support equipment whichever determined the worst-case emission. Once the worst case emissions have been identified, the one EUT cable configuration/arrangement and mode of operation that produced these emissions were used for final measurements on the same test site. The analyzer is set to CISPR quasi-peak and average detectors with a 9kHz resolution bandwidth for final measurements.

An extension cord was used to connect to a single LISN which powered by EUT. The extension cord was calibrated with LISN, the impedance and insertion loss are compliance with the requirements as stated in ANSI C63.10-2013.

## 3.3. Radiated Emissions

The radiated test facilities consisted of an indoor 3 meter semi-anechoic chamber used for final measurements and exploratory measurements, when necessary. The measurement area is contained within the semi-anechoic chamber which is shielded from any ambient interference. For measurements above 1GHz absorbers are arranged on the floor between the turn table and the Antenna mast in such a way so as to maximize the reduction of reflections. For measurements below 1GHz, the absorbers are removed. A MF Model 210SS turntable is used for radiated measurement. It is a continuously rotatable, remote controlled, metallic turntable and 2 meters (6.56 ft.) in diameter. The turn table is flush with the raised floor of the chamber in order to maintain its function as a ground plane. An 80cm high PVC support structure is placed on top of the turntable. For all measurements, the spectrum was scanned through all EUT azimuths and from 1 to 4 meter receive Antenna height using a broadband Antenna from 30MHz up to the upper frequency shown in 15.33(b)(1) depending on the highest frequency generated or used in the device or on which the device operates or tunes. For frequencies above 1GHz, linearly polarized double ridge horn Antennas were used. For frequencies below 30MHz, a calibrated loop Antenna was used. When exploratory measurements were necessary, they were performed at 1 meter test distance inside the semi-anechoic chamber using broadband Antennas, broadband amplifiers, and spectrum analyzers to determine the frequencies and modes producing the maximum emissions. Sufficient time for the EUT, support equipment, and test equipment was allowed in order for them to warm up to their normal operating condition. The test set-up for frequencies below 1GHz was placed on top of the 0.8 meter high, 1 x 1.5 meter table; and test set-up for frequencies 1-40GHz was placed on top of the 1.5 meter high, 1 x 1.5 meter table. The EUT, support equipment, and interconnecting cables were arranged and manipulated to maximize each emission. Appropriate precaution was taken to ensure that all emissions from the EUT were maximized and investigated. The system configuration, clock speed, mode of operation or video resolution, if applicable, turntable azimuth, and receive Antenna height was noted for each frequency found.

Final measurements were made in the semi-anechoic chamber using calibrated, linearly polarized broadband and horn Antennas. The test setup was configured to the setup that produced the worst case emissions. The spectrum analyzer was set to investigate all frequencies required for testing to compare the highest radiated disturbances with respect to the specified limits. The turntable containing the EUT was rotated through 360 degrees and the height of the receive Antenna was varied 1 to 4 meters and stopped at the azimuth and height producing the maximum emission. Each emission was maximized by changing the orientation of the EUT through three orthogonal planes and changing the polarity of the receive Antenna, whichever produced the worst-case emissions. According to 3dB Beam-Width of horn Antenna, the horn Antenna should be always directed to the EUT when rising height.

# 4. ANTENNA REQUIREMENTS

**Excerpt from §15.203 of the FCC Rules/Regulations:**

"An intentional radiator antenna shall be designed to ensure that no antenna other than that furnished by the responsible party can be used with the device. The use of a permanently attached antenna or of an antenna that uses a unique coupling to the intentional radiator shall be considered sufficient to comply with the provisions of this section."

**Conclusion:**

The device unit complies with the requirement of §15.203.

# 5. TEST EQUIPMENT CALIBRATION DATE

Conducted Emissions - SR2

| Instrument | Manufacturer | Type No. | Asset No. | Cali. Interval | Cali. Due Date |
|---|---|---|---|---|---|
| EMI Test Receiver | R&S | ESR7 | MRTSUE06001 | 1 year | 2018/08/18 |
| Two-Line V-Network | R&S | ENV 216 | MRTSUE06002 | 1 year | 2018/06/21 |
| Two-Line V-Network | R&S | ENV 216 | MRTSUE06003 | 1 year | 2018/06/21 |
| Thermohygrometer | Testo | 608-H1 | MRTSUE06404 | 1 year | 2018/08/14 |
| Shielding Anechoic Chamber | Mikebang | Chamber-SR2 | MRTSUE06215 | 1 year | 2018/05/09 |

Radiated Emission - AC2

| Instrument | Manufacturer | Type No. | Asset No. | Cali. Interval | Cali. Due Date |
|---|---|---|---|---|---|
| MXE EMI Receiver | Agilent | N9038A | MRTSUE06125 | 1 year | 2018/08/18 |
| EXA Signal Analyzer | Agilent | N9010A | MRTSUE06195 | 1 year | 2018/04/22 |
| Loop Antenna | Schwarzbeck | FMZB 1519 | MRTSUE06025 | 1 year | 2018/11/20 |
| Bilog Period Antenna | Schwarzbeck | VULB 9162 | MRTSUE06022 | 1 year | 2018/10/21 |
| Broad Band Horn Antenna | Schwarzbeck | BBHA 9120D | MRTSUE06171 | 1 year | 2018/11/18 |
| Broadband Coaxial Preamplifier | Schwarzbeck | BBV 9718 | MRTSUE06176 | 1 year | 2018/11/17 |
| Digitial Thermometer & Hygrometer | MingGao | ETH529 | MRTSUE06170 | 1 year | 2018/12/12 |
| Anechoic Chamber | RIKEN | Chamber-AC2 | MRTSUE06213 | 1 year | 2018/05/09 |

Conducted Test Equipment - TR3

| Instrument | Manufacturer | Type No. | Asset No. | Cali. Interval | Cali. Due Date |
|---|---|---|---|---|---|
| Power Meter | Agilent | U2021XA | MRTSUE06030 | 1 year | 2018/12/06 |
| Temperature Humidity Chamber | BAOYT | BYH-150CL | MRTSUE06051 | 1 year | 2018/12/06 |
| EXA Signal Analyzer | Agilent | N9020A | MRTSUE06106 | 1 year | 2018/04/25 |
| Thermohygrometer | Testo | 608-H1 | MRTSUE06401 | 1 year | 2018/08/14 |

| Software | Version | Function |
|---|---|---|
| e3 | V8.3.5 | EMI Test Software |

# 6. MEASUREMENT UNCERTAINTY

Where relevant, the following test uncertainty levels have been estimated for tests performed on the EUT as specified in CISPR 16-4-2. This uncertainty represents an expanded uncertainty expressed at approximately the 95% confidence level using a coverage factor of k = 2.

| |
|---|
| AC Conducted Emission Measurement - SR2 |
| Measuring Uncertainty for a Level of Confidence of 95% (U=2Uc(y)):<br>150kHz~30MHz: 3.46dB |
| Radiated Emission Measurement - AC2 |
| Measuring Uncertainty for a Level of Confidence of 95% (U=2Uc(y)):<br>9kHz ~ 1GHz: 4.18dB<br>1GHz ~ 25GHz: 4.76dB |
| Spurious Emissions, Conducted - TR3 |
| Measuring Uncertainty for a Level of Confidence of 95% (U=2Uc(y)):<br>0.78dB |
| Output Power - TR3 |
| Measuring Uncertainty for a Level of Confidence of 95% (U=2Uc(y)):<br>1.13dB |
| Power Spectrum Density - TR3 |
| Measuring Uncertainty for a Level of Confidence of 95% (U=2Uc(y)):<br>1.15dB |
| Occupied Bandwidth - TR3 |
| Measuring Uncertainty for a Level of Confidence of 95% (U=2Uc(y)):<br>0.28% |

# 7. TEST RESULT

## 7.1. Summary

**Company Name:** **VivaChek Biotech (Hangzhou) Co., Ltd.**

**FCC ID:** **2APAPVGM09**

| FCC Section(s) | Test Description | Test Limit | Test Condition | Test Result | Reference |
|---|---|---|---|---|---|
| 15.247(a)(2) | 6dB Bandwidth | Refer to Section 7.2.1 | Conducted | Pass | Section 7.2 |
| 15.247(b)(3) | Output Power | Refer to Section 7.3.1 | | Pass | Section 7.3 |
| 15.247(e) | Power Spectral Density | Refer to Section 7.4.1 | | Pass | Section 7.4 |
| 15.247(d) | Band Edge / Out-of-Band Emissions | Refer to Section 7.5.1 | | Pass | Section 7.5 |
| 15.205<br>15.209 | General Field Strength Limits (Restricted Bands and Radiated Emission Limits) | Refer to Section 7.6.1 & 7.7.1 | Radiated | Pass | Section 7.6 & 7.7 |
| 15.207 | AC Conducted Emissions 150kHz - 30MHz | Refer to Section 7.8.1 | Line Conducted | Pass | Section 7.8 |

Notes: The analyzer plots shown in this section were all taken with a correction table loaded into the analyzer. The correction table was used to account for the losses of the cables and attenuators used as part of the system to connect the EUT to the analyzer at all frequencies of interest.

## 7.2. 6dB Bandwidth Measurement

### 7.2.1.Test Limit

The minimum 6dB bandwidth shall be at least 500 kHz.

### 7.2.2.Test Procedure used

KDB 558074 D01v04 - Section 8.2 Option 2

### 7.2.3.Test Setting

1. The Spectrum's automatic bandwidth measurement capability was used to perform the 6dB bandwidth measurement. The "X" dB bandwidth parameter was set to X = 6. The bandwidth measurement was not influenced by any intermediate power nulls in the fundamental emission.
2. Set RBW = 100 kHz
3. VBW ≥ 3 × RBW
4. Detector = Peak
5. Trace mode = max hold
6. Sweep = auto couple
7. Allow the trace was allowed to stabilize

### 7.2.4.Test Setup



### 7.2.5.Test Result

| Product | Blood Glucose Meter | Temperature | 25°C |
|---|---|---|---|
| Test Engineer | Hunk Li | Relative Humidity | 54% |
| Test Site | TR3 | Test Date | 2018/03/14 |

| Test Mode | Data Rate (Mbps) | Channel No. | Frequency (MHz) | 6dB Bandwidth (MHz) | Limit (MHz) | Result |
|---|---|---|---|---|---|---|
| BLE | 1 | 00 | 2402 | 0.74 | ≥ 0.5 | Pass |
| BLE | 1 | 19 | 2440 | 0.75 | ≥ 0.5 | Pass |
| BLE | 1 | 39 | 2480 | 0.74 | ≥ 0.5 | Pass |

**BLE 6dB Bandwidth**

**Channel 00 (2402MHz)**

**Channel 19 (2440MHz)**

**Channel 39 (2480MHz)**

## 7.3. Output Power Measurement

### 7.3.1.Test Limit

The maximum out power shall be less 1 Watt (30dBm).

### 7.3.2.Test Procedure Used

KDB 558074 D01v04 - Section 9.1.2 PKPM1 - Peak Power Method

### 7.3.3.Test Setting

**Method PKPM1 (Peak Power Measurement of Signals with DTS BW ≤ 50MHz)**

Peak power measurements were performed only when the EUT was transmitting at its maximum power control level using a broadband power meter with a pulse sensor. The pulse sensor employs a VBW = 50MHz so this method was only used for signals whose DTS bandwidth was less than or equal to 50MHz.

### 7.3.4.Test Setup



### 7.3.5.Test Result of Output Power

| Product | VivaChek Ino Sound | Temperature | 25°C |
|---|---|---|---|
| Test Engineer | Hunk Li | Relative Humidity | 54% |
| Test Site | TR3 | Test Date | 2018/03/14 |

**Test Result of Peak Output Power**

| Test Mode | Data Rate (Mbps) | Channel No. | Frequency (MHz) | Peak Power (dBm) | Limit (dBm) | Result |
|---|---|---|---|---|---|---|
| BLE | 1 | 00 | 2402 | -3.23 | ≤ 30 | Pass |
| BLE | 1 | 19 | 2440 | -3.44 | ≤ 30 | Pass |
| BLE | 1 | 39 | 2480 | -3.35 | ≤ 30 | Pass |

**Test Result of Average Output Power (Report Only)**

| Test Mode | Data Rate (Mbps) | Channel No. | Frequency (MHz) | Average Power (dBm) | Limit (dBm) | Result |
|---|---|---|---|---|---|---|
| BLE | 1 | 00 | 2402 | -3.30 | ≤ 30 | Pass |
| BLE | 1 | 19 | 2440 | -3.53 | ≤ 30 | Pass |
| BLE | 1 | 39 | 2480 | -3.42 | ≤ 30 | Pass |

## 7.4. Power Spectral Density Measurement

### 7.4.1.Test Limit

The maximum permissible power spectral density is 8dBm in any 3 kHz band.

### 7.4.2.Test Procedure Used

KDB 558074 D01v04 - Section 10.2 Method PKPSD

### 7.4.3.Test Setting

1. Analyzer was set to the center frequency of the DTS channel under investigation
2. Span = 1.5 times the DTS channel bandwidth
3. RBW = 3kHz
4. VBW = 10kHz
5. Detector = peak
6. Sweep time = auto couple
7. Trace mode = max hold
8. Trace was allowed to stabilize

### 7.4.4.Test Setup



### 7.4.5.Test Result

| Product | Blood Glucose Meter | Temperature | 25°C |
|---|---|---|---|
| Test Engineer | Hunk Li | Relative Humidity | 54% |
| Test Site | TR3 | Test Date | 2018/03/14 |

| Test Mode | Data Rate (Mbps) | Channel No. | Frequency (MHz) | PSD Result (dBm / 3kHz) | Limit (dBm / 3kHz) | Result |
|---|---|---|---|---|---|---|
| BLE | 1 | 00 | 2402 | -19.97 | ≤ 8.00 | Pass |
| BLE | 1 | 19 | 2440 | -20.10 | ≤ 8.00 | Pass |
| BLE | 1 | 39 | 2480 | -20.14 | ≤ 8.00 | Pass |

**BLE PSD**

**Channel 00 (2402MHz)**

**Channel 19 (2440MHz)**

**Channel 39 (2480MHz)**

## 7.5. Conducted Band Edge and Out-of-Band Emissions

### 7.5.1.Test Limit

The limit for out-of-band spurious emissions at the band edge is 30dB below the fundamental emission level, as determined from the in-band power measurement of the DTS channel performed in a 100kHz bandwidth per the PSD procedure.

### 7.5.2.Test Procedure Used

KDB 558074 D01v04 - Section 11.2 & Section 11.3

### 7.5.3.Test Settitng

**Reference level measurement**

1. Set instrument center frequency to DTS channel center frequency
2. Set the span to ≥ 1.5 times the DTS bandwidth
3. Set the RBW = 100 kHz
4. Set the VBW ≥ 3 x RBW
5. Detector = peak
6. Sweep time = auto couple
7. Trace mode = max hold
8. Allow trace to fully stabilize

**Emission level measurement**

1. Set the center frequency and span to encompass frequency range to be measured
2. RBW = 100kHz
3. VBW = 300kHz
4. Detector = Peak
5. Number of sweep points ≥ 2 x Span/RBW
6. Trace mode = max hold
7. Sweep time = auto couple
8. The trace was allowed to stabilize

### 7.5.4.Test Setup



### 7.5.5.Test Result

| Product | VivaChek Ino Sound | Temperature | 25°C |
|---|---|---|---|
| Test Engineer | Hunk Li | Relative Humidity | 54% |
| Test Site | TR3 | Test Date | 2018/03/14 |

| Test Mode | Data Rate (Mbps) | Channel No. | Frequency (MHz) | Limit | Result |
|---|---|---|---|---|---|
| BLE | 1 | 00 | 2402 | 20dBc | Pass |
| BLE | 1 | 19 | 2440 | 20dBc | Pass |
| BLE | 1 | 39 | 2480 | 20dBc | Pass |

**BLE Out-of-Band Emissions**

**Channel 00 (2402MHz)**

**100kHz PSD Reference Level**

**Low Band Edge**

**Spurious Emission 30MHz ~ 25GHz**

| Channel 19 (2440MHz) | |
|---|---|
| **100kHz PSD Reference Level** | **Spurious Emission 30MHz ~ 25GHz** |
| | |
| **Channel 39 (2480MHz)** | |
| **100kHz PSD Reference Level** | **High Band Edge** |
| |  |
| **Spurious Emission 30MHz ~ 25GHz** | |
| | |

## 7.6. Radiated Spurious Emission Measurement

### 7.6.1.Test Limit

All out of band emissions appearing in a restricted band as specified in Section 15.205 of the Title 47 CFR must not exceed the limits shown in Table per Section 15.209.

| FCC Part 15 Subpart C Paragraph 15.209 | | |
|---|---|---|
| Frequency [MHz] | Field Strength [uV/m] | Measured Distance [Meters] |
| 0.009 - 0.490 | 2400/F (kHz) | 300 |
| 0.490 - 1.705 | 24000/F (kHz) | 30 |
| 1.705 - 30 | 30 | 30 |
| 30 - 88 | 100 | 3 |
| 88 - 216 | 150 | 3 |
| 216 - 960 | 200 | 3 |
| Above 960 | 500 | 3 |

### 7.6.2.Test Procedure Used

KDB 558074 D01v04 - Section 12.2.3 (quasi-peak measurements)

KDB 558074 D01v04 - Section 12.2.4 (peak power measurements)

KDB 558074 D01v04 - Section 12.2.5 (average power measurements)

### 7.6.3.Test Setting

**Peak Field Strength Measurements per Section 12.2.4 of KDB 558074 D01v04**

1. Analyzer center frequency was set to the frequency of the radiated spurious emission of interest
2. RBW = as specified in Table 1
3. VBW = 3MHz
4. Detector = peak
5. Sweep time = auto couple

6. Trace mode = max hold
7. Trace was allowed to stabilize

**Table 1 - RBW as a function of frequency**

| Frequency | RBW |
|---|---|
| 9 ~ 150 kHz | 200 ~ 300 Hz |
| 0.15 ~ 30 MHz | 9 ~ 10 kHz |
| 30 ~ 1000 MHz | 100 ~ 120 kHz |
| > 1000 MHz | 1 MHz |

**Average Field Strength Measurements per Section 12.2.4 of KDB 558074 D01v04**

1. Analyzer center frequency was set to the frequency of the radiated spurious emission of interest
2. RBW = 1MHz
3. VBW ≥ 1/T
4. De As an alternative, the instrument may be set to linear detector mode. Ensure that video filtering is applied in linear voltage domain (rather than in a log or dB domain). Some instruments require linear display mode in order to accomplish this. Others have a setting for Average-VBW Type, which can be set to “Voltage” regardless of the display mode
5. Detector = Peak
6. Sweep time = auto
7. Trace mode = max hold
8. Allow max hold to run for at least 50 times (1/duty cycle) traces

### 7.6.4.Test Setup

9kHz ~ 30MHz Test Setup:




30MHz ~ 1GHz Test Setup:



1GHz ~ 18GHz Test Setup:




18GHz ~25GHz Test Setup:



### 7.6.5.Test Result

| Test Mode: | BLE | Test Site: | AC2 |
|---|---|---|---|
| Test Channel: | 00 | Test Engineer: | Dandy Li |
| Remark: | 1. Average measurement was not performed if peak level lower than average limit.<br>2. Other frequency was 20dB below limit line within 1-18GHz, there is not show in the report. | | |

| Mark | Frequency (MHz) | Reading Level (dBμV) | Factor (dB) | Measure Level (dBμV/m) | Limit (dBμV/m) | Margin (dB) | Detector | Polarization |
|---|---|---|---|---|---|---|---|---|
| | 3805.0 | 38.8 | 2.2 | 41.0 | 74.0 | -33.0 | Peak | Horizontal |
| | 4799.5 | 44.7 | 5.5 | 50.2 | 74.0 | -23.8 | Peak | Horizontal |
| * | 5692.0 | 35.8 | 7.0 | 42.8 | 74.0 | -31.2 | Peak | Horizontal |
| * | 6593.0 | 34.6 | 10.7 | 45.3 | 74.0 | -28.7 | Peak | Horizontal |
| | 4094.0 | 37.5 | 3.1 | 40.6 | 74.0 | -33.4 | Peak | Vertical |
| | 4799.5 | 38.5 | 5.5 | 44.0 | 74.0 | -30.0 | Peak | Vertical |
| * | 5998.0 | 34.2 | 8.1 | 42.3 | 74.0 | -31.7 | Peak | Vertical |
| * | 6848.0 | 34.9 | 11.3 | 46.2 | 74.0 | -27.8 | Peak | Vertical |

Note 1: “*” is not in restricted band, its limit is 20dBc of the fundamental emission level (87.8dBμV/m) or 15.209 which is higher.

Note 2: Measure Level (dBμV/m) = Reading Level (dBμV) + Factor (dB)

Factor (dB) = Cable Loss (dB) + Antenna Factor (dB/m) - Pre_Amplifier Gain (dB)

| Test Mode: | BLE | Test Site: | AC2 |
|---|---|---|---|
| Test Channel: | 19 | Test Engineer: | Dandy Li |
| Remark: | 1. Average measurement was not performed if peak level lower than average limit.<br>2. Other frequency was 20dB below limit line within 1-18GHz, there is not show in the report. | | |

| Mark | Frequency (MHz) | Reading Level (dBμV) | Factor (dB) | Measure Level (dBμV/m) | Limit (dBμV/m) | Margin (dB) | Detector | Polarization |
|---|---|---|---|---|---|---|---|---|
| | 4068.5 | 36.8 | 3.1 | 39.9 | 74.0 | -34.1 | Peak | Horizontal |
| | 4876.0 | 46.4 | 5.6 | 52.0 | 74.0 | -22.0 | Peak | Horizontal |
| * | 5887.5 | 33.8 | 7.9 | 41.7 | 74.0 | -32.3 | Peak | Horizontal |
| * | 6593.0 | 34.2 | 10.7 | 44.9 | 74.0 | -29.1 | Peak | Horizontal |
| | 4060.0 | 37.2 | 3.1 | 40.3 | 74.0 | -33.7 | Peak | Vertical |
| | 4884.5 | 42.4 | 5.6 | 48.0 | 74.0 | -26.0 | Peak | Vertical |
| * | 5760.0 | 35.5 | 7.5 | 43.0 | 74.0 | -31.0 | Peak | Vertical |
| * | 6797.0 | 34.4 | 10.9 | 45.3 | 74.0 | -28.7 | Peak | Vertical |

Note 1: "*" is not in restricted band, its limit is 20dBc of the fundamental emission level (87.9dBμV/m) or 15.209 which is higher.

Note 2: Measure Level (dBμV/m) = Reading Level (dBμV) + Factor (dB)

Factor (dB) = Cable Loss (dB) + Antenna Factor (dB/m) - Pre_Amplifier Gain (dB)

| Test Mode: | BLE | Test Site: | AC2 |
|---|---|---|---|
| Test Channel: | 39 | Test Engineer: | Dandy Li |
| Remark: | 1. Average measurement was not performed if peak level lower than average limit.<br>2. Other frequency was 20dB below limit line within 1-18GHz, there is not show in the report. | | |

| Mark | Frequency (MHz) | Reading Level (dBμV) | Factor (dB) | Measure Level (dBμV/m) | Limit (dBμV/m) | Margin (dB) | Detector | Polarization |
|---|---|---|---|---|---|---|---|---|
| | 3958.0 | 37.7 | 2.5 | 40.2 | 74.0 | -33.8 | Peak | Horizontal |
| | 4960.0 | 48.7 | 5.7 | 54.4 | 74.0 | -19.6 | Peak | Horizontal |
| * | 4960.0 | 44.9 | 5.7 | 50.6 | 54.0 | -3.4 | Peak | Horizontal |
| * | 6108.5 | 34.5 | 8.3 | 42.8 | 74.0 | -31.2 | Peak | Horizontal |
| | 6678.0 | 33.5 | 10.8 | 44.3 | 74.0 | -29.7 | Peak | Vertical |
| | 3779.5 | 37.5 | 2.1 | 39.6 | 74.0 | -34.4 | Peak | Vertical |
| * | 4961.0 | 42.2 | 5.7 | 47.9 | 74.0 | -26.1 | Peak | Vertical |
| * | 6023.5 | 34.0 | 8.1 | 42.1 | 74.0 | -31.9 | Peak | Vertical |

Note 1: "*" is not in restricted band, its limit is 20dBc of the fundamental emission level (87.72dBμV/m) or 15.209 which is higher.

Note 2: Measure Level (dBμV/m) = Reading Level (dBμV) + Factor (dB)

Factor (dB) = Cable Loss (dB) + Antenna Factor (dB/m) - Pre_Amplifier Gain (dB)

**The worst case of Radiated Emission below 1GHz:**

| | |
|---|---|
| Site: AC2 | Time: 2018/03/13 - 21:17 |
| Limit: FCC_Part15.209_RE(3m) | Engineer: Dandy Li |
| Probe: VULB9162_0.03-8GHz | Polarity: Horizontal |
| EUT: Blood Glucose Meter | Power: By Battery |
| **Note: There is the worst case within frequency range 30MHz~1GHz.** | |




| No | Flag | Mark | Frequency (MHz) | Measure Level (dBuV/m) | Reading Level (dBuV) | Over Limit (dB) | Limit (dBuV/m) | Factor (dB) | Type |
|---|---|---|---|---|---|---|---|---|---|
| 1 | | | 95.960 | 30.775 | 15.967 | -12.725 | 43.500 | 14.808 | QP |
| 2 | | | 120.210 | 30.728 | 17.224 | -12.772 | 43.500 | 13.504 | QP |
| 3 | | * | 168.225 | 33.279 | 21.020 | -10.221 | 43.500 | 12.259 | QP |
| 4 | | | 191.990 | 32.121 | 18.208 | -11.379 | 43.500 | 13.913 | QP |
| 5 | | | 215.755 | 31.820 | 17.346 | -11.680 | 43.500 | 14.474 | QP |
| 6 | | | 407.815 | 32.617 | 14.332 | -13.383 | 46.000 | 18.285 | QP |

Note 1: Measure Level (dBμV/m) = Reading Level (dBμV) + Factor (dB)

Factor (dB) = Cable Loss (dB) + Antenna Factor (dB/m)

Note 2: The test trace is same as the ambient noise and the amplitude of the emissions are attenuated more than 20dB below the permissible (the test frequency range: 9kHz ~ 30MHz, 18GHz ~ 25GHz), therefore no data appear in the report.

| | |
|---|---|
| Site: AC2 | Time: 2018/03/13 - 21:18 |
| Limit: FCC_Part15.209_RE(3m) | Engineer: Dandy Li |
| Probe: VULB9162_0.03-8GHz | Polarity: Vertical |
| EUT: VivaChek Ino Sound | Power: By Battery |
| **Note: There is the worst case within frequency range 30MHz~1GHz.** | |




| No | Flag | Mark | Frequency (MHz) | Measure Level (dBuV/m) | Reading Level (dBuV) | Over Limit (dB) | Limit (dBuV/m) | Factor (dB) | Type |
|---|---|---|---|---|---|---|---|---|---|
| 1 | | * | 30.970 | 27.263 | 12.334 | -12.737 | 40.000 | 14.929 | QP |
| 2 | | | 60.070 | 26.382 | 9.988 | -13.618 | 40.000 | 16.394 | QP |
| 3 | | | 120.210 | 27.208 | 13.704 | -16.292 | 43.500 | 13.504 | QP |
| 4 | | | 159.980 | 30.616 | 18.672 | -12.884 | 43.500 | 11.944 | QP |
| 5 | | | 455.830 | 32.263 | 13.412 | -13.737 | 46.000 | 18.851 | QP |
| 6 | | | 635.765 | 31.661 | 10.203 | -14.339 | 46.000 | 21.458 | QP |

Note 1: Measure Level (dBμV/m) = Reading Level (dBμV) + Factor (dB)

Factor (dB) = Cable Loss (dB) + Antenna Factor (dB/m)

Note 2: The test trace is same as the ambient noise and the amplitude of the emissions are attenuated more than 20dB below the permissible (the test frequency range: 9kHz ~ 30MHz, 18GHz ~ 25GHz), therefore no data appear in the report.

## 7.7. Radiated Restricted Band Edge Measurement

### 7.7.1.Test Limit

**For 15.205 requirement:**

Radiated emissions which fall in the restricted bands, as defined in Section 15.205(a) of FCC part 15, must also comply with the radiated emission limits specified in Section 15.209(a).

| Frequency (MHz) | Frequency (MHz) | Frequency (MHz) | Frequency (GHz) |
|---|---|---|---|
| 0.090 - 0.110 | 16.42-16.423 | 399.9 - 410 | 4.5-5.15 |
| [1] 0.495 - 0.505 | 16.69475-16.69525 | 608 - 614 | 5.35-5.46 |
| 2.1735-2.1905 | 16.80425-16.80475 | 960 - 1240 | 7.25-7.75 |
| 4.125-4.128 | 25.5 -25.67 | 1300 - 1427 | 8.25 - 8.5 |
| 4.17725-4.17775 | 37.5-38.25 | 1435-1626.5 | 9.0-9.2 |
| 4.20725-4.20775 | 73-74.6 | 1645.5-1646.5 | 9.3-9.5 |
| 6.215-6.218 | 74.8-75.2 | 1660 - 1710 | 10.6-12.7 |
| 6.26775-6.26825 | 108-121.94 | 1718.8-1722.2 | 13.25-13.4 |
| 6.31175-6.31225 | 123 - 138 | 2200 - 2300 | 14.47-14.5 |
| 8.291-8.294 | 149.9-150.05 | 2310–2390 | 15.35-16.2 |
| 8.362-8.366 | 156.52475-156.525 | 2483.5 - 2500 | 17.7-21.4 |
| 8.37625-8.38675 | 156.7-156.9 | 2690 - 2900 | 22.01-23.12 |
| 8.41425-8.41475 | 162.0125-167.17 | 3260 - 3267 | 23.6-24.0 |
| 12.29-12.293 | 167.72-173.2 | 3332 - 3339 | 31.2-31.8 |
| 12.51975-12.52025 | 240 - 285 | 3345.8 - 3358 | 36.43-36.5 |
| 12.57675-12.57725 | 322-335.4 | 3600 - 4400 | ([2]) |
| 13.36-13.41 | -- | -- | -- |

All out of band emissions appearing in a restricted band as specified in Section 15.205 of the Title 47CFR must not exceed the limits shown in Table per Section 15.209.

| FCC Part 15 Subpart C Paragraph 15.209 | | |
|---|---|---|
| Frequency [MHz] | Field Strength [uV/m] | Measured Distance [Meters] |
| 0.009 – 0.490 | 2400/F (kHz) | 300 |
| 0.490 – 1.705 | 24000/F (kHz) | 30 |
| 1.705 - 30 | 30 | 30 |
| 30 - 88 | 100 | 3 |
| 88 - 216 | 150 | 3 |
| 216 - 960 | 200 | 3 |
| Above 960 | 500 | 3 |

**7.7.2.Test Procedure Used**

KDB 558074 D01v04 - Section 12.2.4 (peak power measurements)

KDB 558074 D01v04 - Section 13.3.3 (average power measurements)

**7.7.3.Test Setting**

**Peak Field Strength Measurements**

1. Analyzer center frequency was set to the frequency of the radiated spurious emission of interest
2. RBW = as specified in Table 1
3. VBW = 3MHz
4. Detector = peak
5. Sweep time = auto couple
6. Trace mode = max hold
7. Trace was allowed to stabilize

**Table 1 - RBW as a function of frequency**

| Frequency | RBW |
|---|---|
| 9 ~ 150 kHz | 200 ~ 300 Hz |
| 0.15 ~ 30 MHz | 9 ~ 10 kHz |
| 30 ~ 1000 MHz | 100 ~ 120 kHz |
| > 1000 MHz | 1 MHz |

**Average Field Strength Measurements**

1. Analyzer center frequency was set to the frequency of the radiated spurious emission of interest
2. RBW = 1MHz
3. VBW ≥ 1/T
4. De As an alternative, the instrument may be set to linear detector mode. Ensure that video filtering is applied in linear voltage domain (rather than in a log or dB domain). Some instruments require linear display mode in order to accomplish this. Others have a setting for Average-VBW Type, which can be set to “Voltage” regardless of the display mode
5. Detector = Peak
6. Sweep time = auto
7. Trace mode = max hold
8. Allow max hold to run for at least 50 times (1/duty cycle) traces

### 7.7.4.Test Setup



### 7.7.5.Test Result

| | |
|---|---|
| Site: AC2 | Time: 2018/03/13 - 22:54 |
| Limit: FCC_Part15.209_RE(3m) | Engineer: Dandy Li |
| Probe: VULB9162_0.03-8GHz | Polarity: Horizontal |
| EUT: Blood Glucose Meter | Power: By Battery |
| Note: Transmit by BLE at channel 2402MHz | |




| No | Flag | Mark | Frequency (MHz) | Measure Level (dBuV/m) | Reading Level (dBuV) | Over Limit (dB) | Limit (dBuV/m) | Factor (dB) | Type |
|---|---|---|---|---|---|---|---|---|---|
| 1 | | | 2363.960 | 62.523 | 28.235 | -11.477 | 74.000 | 34.287 | PK |
| 2 | | | 2390.000 | 60.745 | 26.312 | -13.255 | 74.000 | 34.433 | PK |
| 3 | | * | 2402.150 | 87.792 | 53.285 | N/A | N/A | 34.508 | PK |

Note: Measure Level (dBμV/m) = Reading Level (dBμV) + Factor (dB)

Factor (dB) = Cable Loss (dB) + Antenna Factor (dB/m)

| | |
|---|---|
| Site: AC2 | Time: 2018/03/13 - 22:58 |
| Limit: FCC_Part15.209_RE(3m) | Engineer: Dandy Li |
| Probe: VULB9162_0.03-8GHz | Polarity: Horizontal |
| EUT: Blood Glucose Meter | Power: By Battery |
| Note: Transmit by BLE at channel 2402MHz | |




| No | Flag | Mark | Frequency (MHz) | Measure Level (dBuV/m) | Reading Level (dBuV) | Over Limit (dB) | Limit (dBuV/m) | Factor (dB) | Type |
|---|---|---|---|---|---|---|---|---|---|
| 1 | | | 2390.000 | 48.110 | 13.677 | -5.890 | 54.000 | 34.433 | AV |
| 2 | | * | 2402.008 | 86.811 | 52.305 | N/A | N/A | 34.506 | AV |

Note: Measure Level (dBμV/m) = Reading Level (dBμV) + Factor (dB)

Factor (dB) = Cable Loss (dB) + Antenna Factor (dB/m)

| | |
|---|---|
| Site: AC2 | Time: 2018/03/13 - 22:59 |
| Limit: FCC_Part15.209_RE(3m) | Engineer: Dandy Li |
| Probe: VULB9162_0.03-8GHz | Polarity: Vertical |
| EUT: Blood Glucose Meter | Power: By Battery |
| Note: Transmit by BLE at channel 2402MHz | |




| No | Flag | Mark | Frequency (MHz) | Measure Level (dBuV/m) | Reading Level (dBuV) | Over Limit (dB) | Limit (dBuV/m) | Factor (dB) | Type |
|---|---|---|---|---|---|---|---|---|---|
| 1 | | | 2377.925 | 62.227 | 27.860 | -11.773 | 74.000 | 34.367 | PK |
| 2 | | | 2390.000 | 60.810 | 26.377 | -13.190 | 74.000 | 34.433 | PK |
| 3 | | * | 2401.722 | 78.505 | 44.001 | N/A | N/A | 34.504 | PK |

Note: Measure Level (dBμV/m) = Reading Level (dBμV) + Factor (dB)

Factor (dB) = Cable Loss (dB) + Antenna Factor (dB/m)

| | |
|---|---|
| Site: AC2 | Time: 2018/03/13 - 23:01 |
| Limit: FCC_Part15.209_RE(3m) | Engineer: Dandy Li |
| Probe: VULB9162_0.03-8GHz | Polarity: Vertical |
| EUT: Blood Glucose Meter | Power: By Battery |
| Note: Transmit by BLE at channel 2402MHz | |




| No | Flag | Mark | Frequency (MHz) | Measure Level (dBuV/m) | Reading Level (dBuV) | Over Limit (dB) | Limit (dBuV/m) | Factor (dB) | Type |
|---|---|---|---|---|---|---|---|---|---|
| 1 | | | 2390.000 | 48.128 | 13.695 | -5.872 | 54.000 | 34.433 | AV |
| 2 | | * | 2402.008 | 76.801 | 42.295 | N/A | N/A | 34.506 | AV |

Note: Measure Level (dBμV/m) = Reading Level (dBμV) + Factor (dB)

Factor (dB) = Cable Loss (dB) + Antenna Factor (dB/m)

| | |
|---|---|
| Site: AC2 | Time: 2018/03/13 - 23:01 |
| Limit: FCC_Part15.209_RE(3m) | Engineer: Dandy Li |
| Probe: VULB9162_0.03-8GHz | Polarity: Horizontal |
| EUT: Blood Glucose Meter | Power: By Battery |
| Note: Transmit by BLE at channel 2480MHz | |




| No | Flag | Mark | Frequency (MHz) | Measure Level (dBuV/m) | Reading Level (dBuV) | Over Limit (dB) | Limit (dBuV/m) | Factor (dB) | Type |
|---|---|---|---|---|---|---|---|---|---|
| 1 | | * | 2480.059 | 87.721 | 52.686 | N/A | N/A | 35.036 | PK |
| 2 | | | 2483.500 | 62.042 | 26.986 | -11.958 | 74.000 | 35.056 | PK |
| 3 | | | 2486.798 | 63.454 | 28.378 | -10.546 | 74.000 | 35.077 | PK |

Note: Measure Level (dBμV/m) = Reading Level (dBμV) + Factor (dB)

Factor (dB) = Cable Loss (dB) + Antenna Factor (dB/m)

| | |
|---|---|
| Site: AC2 | Time: 2018/03/13 - 23:04 |
| Limit: FCC_Part15.209_RE(3m) | Engineer: Dandy Li |
| Probe: VULB9162_0.03-8GHz | Polarity: Horizontal |
| EUT: Blood Glucose Meter | Power: By Battery |
| Note: Transmit by BLE at channel 2480MHz | |




| No | Flag | Mark | Frequency (MHz) | Measure Level (dBuV/m) | Reading Level (dBuV) | Over Limit (dB) | Limit (dBuV/m) | Factor (dB) | Type |
|---|---|---|---|---|---|---|---|---|---|
| 1 | | * | 2479.979 | 86.992 | 51.957 | N/A | N/A | 35.035 | AV |
| 2 | | | 2483.500 | 48.887 | 13.831 | -5.113 | 54.000 | 35.056 | AV |

Note: Measure Level (dBμV/m) = Reading Level (dBμV) + Factor (dB)

Factor (dB) = Cable Loss (dB) + Antenna Factor (dB/m)

| | |
|---|---|
| Site: AC2 | Time: 2018/03/13 - 23:04 |
| Limit: FCC_Part15.209_RE(3m) | Engineer: Dandy Li |
| Probe: VULB9162_0.03-8GHz | Polarity: Vertical |
| EUT: Blood Glucose Meter | Power: By Battery |
| Note: Transmit by BLE at channel 2480MHz | |




| No | Flag | Mark | Frequency (MHz) | Measure Level (dBuV/m) | Reading Level (dBuV) | Over Limit (dB) | Limit (dBuV/m) | Factor (dB) | Type |
|---|---|---|---|---|---|---|---|---|---|
| 1 | | * | 2480.278 | 78.988 | 43.951 | N/A | N/A | 35.037 | PK |
| 2 | | | 2483.500 | 61.612 | 26.556 | -12.388 | 74.000 | 35.056 | PK |
| 3 | | | 2485.119 | 63.505 | 28.439 | -10.495 | 74.000 | 35.066 | PK |

Note: Measure Level (dBμV/m) = Reading Level (dBμV) + Factor (dB)

Factor (dB) = Cable Loss (dB) + Antenna Factor (dB/m)

| | |
|---|---|
| Site: AC2 | Time: 2018/03/13 - 23:06 |
| Limit: FCC_Part15.209_RE(3m) | Engineer: Dandy Li |
| Probe: VULB9162_0.03-8GHz | Polarity: Vertical |
| EUT: Blood Glucose Meter | Power: By Battery |
| Note: Transmit by BLE at channel 2480MHz | |




| No | Flag | Mark | Frequency (MHz) | Measure Level (dBuV/m) | Reading Level (dBuV) | Over Limit (dB) | Limit (dBuV/m) | Factor (dB) | Type |
|---|---|---|---|---|---|---|---|---|---|
| 1 | | * | 2480.059 | 77.639 | 42.604 | N/A | N/A | 35.036 | AV |
| 2 | | | 2483.500 | 48.853 | 13.797 | -5.147 | 54.000 | 35.056 | AV |

Note: Measure Level (dBμV/m) = Reading Level (dBμV) + Factor (dB)

Factor (dB) = Cable Loss (dB) + Antenna Factor (dB/m)

## 7.8. AC Conducted Emissions Measurement

### 7.8.1.Test Limit

| FCC Part 15 Subpart C Paragraph 15.207 Limits | | |
|---|---|---|
| Frequency (MHz) | QP (dBuV) | AV (dBuV) |
| 0.15 - 0.50 | 66 - 56 | 56 - 46 |
| 0.50 - 5.0 | 56 | 46 |
| 5.0 - 30 | 60 | 50 |

Note 1: The lower limit shall apply at the transition frequencies.

Note 2: The limit decreases linearly with the logarithm of the frequency in the range 0.15MHz to 0.5MHz.

### 7.8.2.Test Setup



### 7.8.3.Test Result

| | |
|---|---|
| Site: SR2 | Time: 2018/03/29 - 11:05 |
| Limit: FCC_Part15.207_CE_AC Power | Engineer: Polly Zong |
| Probe: ENV216_101683_Filter On | Polarity: Line |
| EUT: VGM 09 | Power: AC 120V/60Hz |
| **Note: There is the worst case within frequency range 150KHz~30MHz.** | |




| No | Flag | Mark | Frequency (MHz) | Measure Level (dBuV) | Reading Level (dBuV) | Over Limit (dB) | Limit (dBuV) | Factor (dB) | Type |
|---|---|---|---|---|---|---|---|---|---|
| 1 | | * | 0.166 | 50.455 | 40.368 | -14.703 | 65.158 | 10.087 | QP |
| 2 | | | 0.166 | 19.262 | 9.175 | -35.896 | 55.158 | 10.087 | AV |
| 3 | | | 0.186 | 47.387 | 37.348 | -16.826 | 64.213 | 10.039 | QP |
| 4 | | | 0.186 | 16.556 | 6.517 | -37.658 | 54.213 | 10.039 | AV |
| 5 | | | 0.270 | 38.644 | 28.664 | -22.474 | 61.118 | 9.980 | QP |
| 6 | | | 0.270 | 10.145 | 0.165 | -40.973 | 51.118 | 9.980 | AV |
| 7 | | | 0.590 | 32.270 | 22.150 | -23.730 | 56.000 | 10.120 | QP |
| 8 | | | 0.590 | 17.553 | 7.433 | -28.447 | 46.000 | 10.120 | AV |
| 9 | | | 0.766 | 15.135 | 5.106 | -40.865 | 56.000 | 10.029 | QP |
| 10 | | | 0.766 | 1.922 | -8.106 | -44.078 | 46.000 | 10.029 | AV |
| 11 | | | 12.146 | 17.158 | 7.063 | -42.842 | 60.000 | 10.095 | QP |
| 12 | | | 12.146 | 3.496 | -6.599 | -46.504 | 50.000 | 10.095 | AV |

Note: Measure Level (dBμV) = Reading Level (dBμV) + Factor (dB)

Factor (dB) = Cable Loss (dB) + LISN Factor (dB)

| | |
|---|---|
| Site: SR2 | Time: 2018/03/29 - 11:10 |
| Limit: FCC_Part15.207_CE_AC Power | Engineer: Polly Zong |
| Probe: ENV216_101683_Filter On | Polarity: Neutral |
| EUT: Blood Glucose Meter | Power: AC 120V/60Hz |
| **Note: There is the worst case within frequency range 150KHz~30MHz.** | |




| No | Flag | Mark | Frequency (MHz) | Measure Level (dBuV) | Reading Level (dBuV) | Over Limit (dB) | Limit (dBuV) | Factor (dB) | Type |
|---|---|---|---|---|---|---|---|---|---|
| 1 | | * | 0.166 | 50.644 | 40.573 | -14.514 | 65.158 | 10.071 | QP |
| 2 | | | 0.166 | 19.711 | 9.640 | -35.447 | 55.158 | 10.071 | AV |
| 3 | | | 0.214 | 43.882 | 33.894 | -19.167 | 63.049 | 9.988 | QP |
| 4 | | | 0.214 | 13.935 | 3.948 | -39.113 | 53.049 | 9.988 | AV |
| 5 | | | 0.330 | 34.802 | 24.742 | -24.649 | 59.451 | 10.060 | QP |
| 6 | | | 0.330 | 8.670 | -1.390 | -40.781 | 49.451 | 10.060 | AV |
| 7 | | | 0.590 | 36.433 | 26.297 | -19.567 | 56.000 | 10.137 | QP |
| 8 | | | 0.590 | 19.933 | 9.796 | -26.067 | 46.000 | 10.137 | AV |
| 9 | | | 1.194 | 19.784 | 9.881 | -36.216 | 56.000 | 9.903 | QP |
| 10 | | | 1.194 | 2.365 | -7.537 | -43.635 | 46.000 | 9.903 | AV |
| 11 | | | 12.458 | 23.888 | 13.784 | -36.112 | 60.000 | 10.104 | QP |
| 12 | | | 12.458 | 7.130 | -2.974 | -42.870 | 50.000 | 10.104 | AV |

Note: Measure Level (dBμV) = Reading Level (dBμV) + Factor (dB)

Factor (dB) = Cable Loss (dB) + LISN Factor (dB)

# 8. CONCLUSION

The data collected relate only the item(s) tested and show that the device is in compliance with Part 15C of the FCC Rules.

The End