

LABELING

Original Freigabeformular in TR archiviert.  
Original form stored in the technical writing department.

Label sticks inside of the device

FA. Nr. ##### - ##  
SN YYMMCE####  
Removing this label causes loss of warranty.  
10000021607 01

MAX. IRRIGATION RATE  
PERFORMANCE MAX. IRRIGATION

3.0 l/min

MAX. PUMPING PRESSURE  
PRESSION MAX. POMPE

450 mmHg

MAINS VOLTAGE RANGE  
TENSION SECTEUR

AC 100-240 V~

SUPPLY FREQUENCY  
FREQUENCE DU RESEAU

50/60 Hz

POWER CONSUMPTION  
CONSOMMATION DE PUISSANCE

AC 100 V~ 0.72 A  
AC 240 V~ 0.38 A

WARNING! Risk of fire. Replace fuse as marked.

99

YYYY-MM-DD

IP 21

COVIDIEN

LAP-Pump

Distributed by: Covidien llc,  
15 Hampshire Street,  
Mansfield, MA 02048 USA.  
W.O.M. WORLD OF MEDICINE GmbH  
Salzufer 8, 10587 Berlin, Germany  
Phone: +49 30 39981-550  
Fax: +49 30 39981-545,  
Email: info.berlin@wom.group

PT00101905

Issuing Agency: GS1

REF PP120

SN YYMMCE####

TYPE PP120

Contains FCC ID: 2A5SK-TSHW42  
Contains IC ID: 25004-TSHW42A  
HVIN: 1430

CE0197

GS1 EAN  
Datamatrix  
01/11/21

(01)#####  
(11)YYMMDD  
(21)YYMMCE####

Rear of device

Diagram of the rear of the device showing various components and their part numbers:

- 10000017229 XX (Top center)
- 1109291 (Top left)
- 10000016395 XX (Top right)
- 1600041 (Right side, top)
- 1600385 (Right side, middle)
- 1600374 (Bottom left)
- 1601793 (Bottom right)

Box

Image of the device box with part numbers 10000022524 XX and 10000021608 XX.

Revision history

| Rev. WOM | ECR     | Change                                                                                                                                                     |
|----------|---------|------------------------------------------------------------------------------------------------------------------------------------------------------------|
| 01       | 2019107 | added: RFID info on type label; replaced document no. with material no. for stickers on the rear of the device; update: design drawing of „Rear of Device“ |

Product Management / Customer

Project Management (R&D)

Regulatory Affairs

WORLD OF MEDICINE / Approval

Overview device label PP120 LAP-Pump Covidien

W.O.M. Item number

10000021607 01

File FMT

10000021607 01.fmt

Name

Leyer

File CDR

10000021607 01.cdr

Date

11.06.2019