

Compuray Scan

REF

SN



Manufactured

Product Name Intraoral Imaging Plate System

Model Name Compuray Scan

24V  2A

CLASS 1 LASER PRODUCT
APPAREIL A LASER DE CLASSE 1


COMPLIES WITH 21 CFR 1040.10
AND 1040.11 EXCEPT FOR
DEVIATIONS PURSUANT TO LASER
NOTICE NO. 50, DATED JUNE 24, 2007.



Carestream Dental LLC
3625 Cumberland Boulevard, Suite 700
Atlanta GA USA 30339
Made in China



AREA	REV	DATE	DESCRIPTION	SIGNATURE
	A	2019-04-02	RELEASE TO PRODUCTION.	HATA GAO



- NOTES:
- MATERIAL: GENERAL ELECTRIC LEXAN 8010, 0.254mm HIGH GLOSS, POLYCARBONATE WITH 3M 468MP ADHESIVE OR 3M #9505 ADHESIVE.
 - LABEL TO BE MADE PER ELECTRONIC ARTWORK SUPPLIED BY CARESTREAM DENTAL LLC.
 - OVERLAY MATERIAL MUST MEET THE REQUIREMENTS OF UL STANDARD 969 (MARKING AND LABELING SYSTEMS [PGDQ2]).
 - SUPPLIER ARTWORK TO BE APPROVED BY CUSTOMER PRIOR TO PRINTING.
 - INSPECTION REQUIREMENTS, IF SHOWN ON DRAWING:
 - INDICATES PRIMARY INSPECTION DIMENSION.
 - GEOMETRIC TOLERANCES ARE PRIMARY INSPECTION ATTRIBUTES.
 - REFERENCE DIMENSIONS ARE ENCLOSED IN PARENTHESES (). DO NOT INSPECT REFERENCE DIMENSIONS.

- SUPPLIER REQUIREMENTS:
- SUPPLIER MUST USE UL RECOGNIZED PRINTING PROCESS PER UL STANDARD 969 (MARKING AND LABELING SYSTEMS [PGDQ2]).
 - UL RECOGNIZED SUPPLIERS MUST INDICATE COMPLIANCE TO UL STANDARD 969 BY MARKING THE SMALLEST LOT SIZE WITH THE FOLLOWING:
 - UL FILE NUMBER
 - SUPPLIER'S NAME AND/OR TRADEMARK
 - CARESTREAM HEALTH INC. PART NUMBER AND REVISION
 - DATE OF PRODUCTION.

CARESTREAM DENTAL LLC GLOBAL R&D CENTER (SHANGHAI)				TRACE FILE :		MATERIAL				
				DESIGN SW:						
DATUM PLANE PRIORITY A, B, C UNLESS SPECIFIED				CONTRACT NO:		NAME PRODUCT LABEL-LABEL FCC STATEMENT, CS 7600 CSD				
				DESIGN HATA GAO	DATE 2019-04-02					
UNIT : mm		METER				BUY LEV		SIZE	PART NUMBER TA1313	REV A
UNSPECIFIED TOL.										
ANGLE ±1	0 ±0.5			AUDITING		DATE				
0.0 ±0.13	0.00 ±NA			APPROVE		DATE		SCALE: 1: 1		PROJECT: CS 7600
								SHEET 1 OF 1		