



SAR
Dipole Performance Measurement
Report

EUT Type: SAR Validation Dipole and Waveguide

Model Name: DIP0G750,DIP0G835,DIP1G800,DIP1G900, DIP2G450, DIP2G600

Brand Name: SATIMO

Test Conclusion: Pass

Test Date: 01 Jul . 2024 to 02 Jul. 2025

Date of Issue: 02 Jul. 2025

Testing Engineer :

Xin Liu

(Xin.Liu)

Technical Manager :

Shi fan-long

(Shifan. Long)

Authorized Signatory :

Bovey Yang

(Bovey Yang)



Any reproduction of this document must be done in full. No single part of this document may be reproduced without permission from STS, All Test Data Presented in this report is only applicable to presented Test sample.



ISSUED BY
Shenzhen STS Test Services Co., Ltd.

Kind of Equipment	Manufacturer	Type No.	Serial No.	Last Calibration	Calibrated Until
PC	Acer	N/A	N/A	N/A	N/A
E-Field Probe	MVG	SSE2	SN 07/21 EPGO352	2024.02.14	2025.02.13
Dielectric Probe Kit	MVG	SCLMP	SN 32/14 OCPG67	2023.10.25	2024.10.24
Phantom1	MVG	SAM	SN 32/14 SAM115	N/A	N/A
Phantom3	MVG	SAM	SN 21/21 ELLI48	N/A	N/A
Attenuator	Agilent	99899	DC-18GHz	N/A	N/A
Directional coupler	Narda	4226-20	3305	N/A	N/A
Network Analyzer	Agilent	E5071C	MY46520378	2023-09-26	2024-09-25
Multi Meter	Keithley	Multi Meter 2000	4050073	2023-09-26	2024-09-25
Signal Generator	Agilent	ZHL-42W	9638	2023.10.07	2024.10.06
Power Amplifier	DESAY	NRP	100510	2023-09-26	2024-09-25
Power Meter	R&S	NRP-Z11	101919	2023-09-26	2024-09-25
Power Sensor	R&S	U2021XA	MY56280002	2023-09-26	2024-09-25
Temperature hygrometer	SuWei	SW-108	N/A	2023.09.28	2024.09.27
Thermograph	Elitech	RC-4	S/N EF7176501537	2023.10.12	2024.10.11



ISSUED BY
Shenzhen STS Test Services Co., Ltd.

Kind of Equipment	Manufacturer	Type No.	Serial No.	Last Calibration	Calibrated Until
PC	Acer	N/A	N/A	N/A	N/A
E-Field Probe	MVG	SSE2	SN 08/21 EPGO352	2024.09.18	2025.09.17
Dielectric Probe Kit	MVG	SCLMP	SN 32/14 OCPG67	2024.09.18	2025.09.17
Phantom1	MVG	SAM	SN 32/14 SAM115	N/A	N/A
Phantom2	MVG	SAM	SN 32/14 SAM116	N/A	N/A
Attenuator	Agilent	99899	DC-18GHz	N/A	N/A
Directional coupler	Narda	4226-20	3305	N/A	N/A
Network Analyzer	Agilent	E5071C	MY46520378	2024-09-25	2025-09-26
Multi Meter	Keithley	Multi Meter 2000	4050073	2024-09-25	2025-09-26
Signal Generator	Agilent	N5182A	MY50140530	2024-09-25	2025-09-26
Power Amplifier	DESAY	ZHL-42W	9638	2024-09-25	2025-09-26
Power Meter	R&S	NRP	100510	2024-09-25	2025-09-26
Power Sensor	R&S	NRP-Z11	101919	2024-09-25	2025-09-26
Power Sensor	Keysight	U2021XA	MY56280002	2024-09-25	2025-09-26
Temperature hygrometer	SuWei	SW-108	N/A	2024.10.15	2025.10.14
Thermograph	Elitech	RC-4	S/N EF7176501537	2024.10.15	2025.10.14



ISSUED BY
Shenzhen STS Test Services Co., Ltd.

Kind of Equipment	Manufacturer	Type No.	Serial No.	Last Calibration	Calibrated Until
SAR Test System	SATIMO	N/A	V5.3.15.11	N/A	N/A
Dielectric Probe Kit	MVG	SCLMP	SN 32/14 OCPG67	2025.09.15	2026.09.14
Antenna	MVG	ANTA3	SN 07/13 ZNTA52	N/A	N/A
Phantom1	MVG	SAM	SN 32/14 SAM115	N/A	N/A
Phantom3	MVG	SAM	SN 21/21 ELLI48	N/A	N/A
Phone holder	MVG	N/A	SN 32/14 MSH97	N/A	N/A
Laptop holder	MVG	N/A	SN 32/14 LSH29	N/A	N/A
Attenuator	Agilent	HXT-10-8-SMA	240327017	2025-02-22	2026-02-21
Directional coupler	Xi'an Xingbo	XBOH-OA08-20dB	211123-4-3	2025-02-22	2026-02-21
Network Analyzer	Agilent	E5071C	MY46520378	2025-09-26	2026-09-24
Multi Meter	Keithley	Multi Meter 2000	4050073	2025-09-22	2026-09-21
Signal Generator	Agilent	N5182A	MY50140530	2025-09-26	2026-09-25
Wireless Communication Test Set	Agilent	8960-E5515C	MY48360751	2025-02-22	2026-02-21
Wireless Communication Test Set	R&S	CMW500	156324	2025-09-25	2026-09-24
Power Amplifier	DESAY	WCS-DLNA50M8G50G1	CSKJLNA231227A	2025-09-29	2026-09-28
Power Meter	R&S	NRP	100510	2025-09-25	2026-09-24
Power Sensor	R&S	NRP-Z11	101919	2025-09-25	2026-09-24
Power Sensor	Keysight	U2021XA	MY56280002	2025-09-25	2026-09-24
Temperature hygrometer	TOPRIE	TP502V4-P	6700000124198	2025.10.09	2026.10.08
Thermograph	Elitech	RC-4	S/N EF7176501537	2025.10.17	2026.10.16



2. <Justification of the extended calibration>


Referring to KDB 865664 D01, if dipoles are verified in return loss < -20dB, (within 20% of prior calibration), and in impedance (within 5 ohm of prior calibration), the annual calibration is not necessary and the calibration interval can be extended.

Head 750 MHz				
Date of Measurement	Return Loss (dB)	Delta (%)	Impedance	Delta(ohm)
2023.07.04	-25.47	-	55.2	-
2024.07.01	-25.98	2.00	54.86	-0.34
2025.07.01	-25.49	0.08	55.49	0.029

The return loss is < -20dB, within 20% of prior calibration; the impedance is within 5 ohm of prior calibration. Therefore the verification result should support extended calibration.

<Dipole Verification Data>

Head 750 MHz






ISSUED BY
Shenzhen STS Test Services Co., Ltd.

Head 835 MHz				
Date of Measurement	Return Loss (dB)	Delta (%)	Impedance	Delta(ohm)
2023.07.04	-28.13	-	51.4	-
2024.07.01	-28.43	1.07	52.62	1.22
2025.07.01	-27.70	-1.53	54.05	2.65

The return loss is <-20dB, within 20% of prior calibration; the impedance is within 5 ohm of prior calibration. Therefore the verification result should support extended calibration.

<Dipole Verification Data>

Head 835MHz






Head 1800 MHz				
Date of Measurement	Return Loss (dB)	Delta (%)	Impedance	Delta(ohm)
2023.12.20	-27.38	/	50.2	/
2024.07.01	-27.04	-1.23	48.48	-1.72
2025-07-01	-27.38	0	48.55	-1.65

The return loss is <-20dB, within 20% of prior calibration; the impedance is within 5 ohm of prior calibration. Therefore the verification result should support extended calibration.

<Dipole Verification Data>

Head 1800 MHz






ISSUED BY
Shenzhen STS Test Services Co., Ltd.

Head 1900 MHz				
Date of Measurement	Return Loss (dB)	Delta (%)	Impedance	Delta(ohm)
2023.09-14	-23.66	-	51.4	-
2024.07.01	-23.62	-0.17	50.73	-0.67
2025.07.02	-21.75	-8.07	47.01	-4.39

The return loss is <-20dB, within 20% of prior calibration; the impedance is within 5 ohm of prior calibration. Therefore the verification result should support extended calibration.

<Dipole Verification Data>

Head 1900 MHz





Head 2450 MHz				
Date of Measurement	Return Loss (dB)	Delta (%)	Impedance	Delta(ohm)
2023.07.04	-26.03	-	46.3	-
2024.07.01	-26.42	1.50	47.25	0.95
2025.07.01	-25.75	3.38	49.40	3.10

The return loss is <-20dB, within 20% of prior calibration; the impedance is within 5 ohm of prior calibration. Therefore the verification result should support extended calibration.

<Dipole Verification Data>

Head 2450 MHz





Head 2600 MHz				
Date of Measurement	Return Loss (dB)	Delta (%)	Impedance	Delta(ohm)
2023.07.04	-34.32	-	50.3	-
2024.07.01	-33.58	-2.16	49.61	-0.69
2025.07.01	-32.91	-4.11	49.19	-1.11

The return loss is <-20dB, within 20% of prior calibration; the impedance is within 5 ohm of prior calibration. Therefore the verification result should support extended calibration.

<Dipole Verification Data>

Head 2600 MHz

