



■ Report No.: DDT-R22081519-2E03

■ Issued Date: Sep. 15, 2022

RF EXPOSURE REPORT

FOR

Applicant	:	MicroTech Medical (Hangzhou) Co., Ltd.
Address	:	No.108 Liuze St., Cangqian, Yuhang District, Hangzhou 311121 Zhejiang P.R.China
Equipment under Test	:	Personal Diabetes Assistant
Model No.	:	G7-P02, MTM-2
Trade Mark	:	AIDEX [®]
FCC ID	:	2ATOV-P-G7
Manufacturer	:	MicroTech Medical (Hangzhou) Co., Ltd.
Address	:	No.108 Liuze St., Cangqian, Yuhang District, Hangzhou 311121 Zhejiang P.R.China

Issued By: Dongguan Dongdian Testing Service Co., Ltd.

Add.: No. 17, Zongbu Road 2, Songshan Lake Sci&Tech, Industry Park,
Dongguan City, Guangdong Province, China, 523808

Tel.: +86-0769-38826678, **E-mail:** ddt@dgddt.com, <http://www.dgddt.com>

REPORT

Table of Contents

	Test report declares.....	3
1.	General Information	5
1.1.	Description of equipment	5
1.2.	Assess laboratory.....	5
2.	RF Exposure evaluation for FCC.....	5

Test Report Declare

Applicant	:	MicroTech Medical (Hangzhou) Co., Ltd.
Address	:	No.108 Liuze St., Cangqian, Yuhang District, Hangzhou 311121 Zhejiang P.R.China
Equipment under Test	:	Personal Diabetes Assistant
Model No.	:	G7-P02, MTM-2
Trade mark	:	AIDEX [®]
Manufacturer	:	MicroTech Medical (Hangzhou) Co., Ltd.
Address	:	No.108 Liuze St., Cangqian, Yuhang District, Hangzhou 311121 Zhejiang P.R.China

Standard Used: KDB447498 D01 General RF Exposure Guidance v06

We Declare:

The equipment described above is assessed by Dongguan Dongdian Testing Service Co., Ltd. and in the configuration assessed the equipment complied with the standards specified above. The assessed results are contained in this report and Dongguan Dongdian Testing Service Co., Ltd. is assumed of full responsibility for the accuracy and completeness of these assess.

After evaluation, our opinion is that the equipment In Accordance with above standard.

Report No:	DDT-R22081519-2E03		
Date of Receipt:	Aug. 26, 2022	Date of Test:	Aug. 26, 2022 ~ Sep. 15, 2022

Prepared By:

Sanvin Zheng

Sanvin Zheng/Engineer

Approved By:



Damon Hu/EMC Manager

Note: This report applies to above tested sample only. This report shall not be reproduced in parts without written approval of Dongguan Dongdian Testing Service Co., Ltd.

Revision History

Rev.	Revisions	Issue Date	Revised By
---	Initial issue	Sep. 15, 2022	

1. General Information

1.1. Description of equipment

EUT* Name	: Personal Diabetes Assistant
Model Number	: G7-P02, MTM-2
Model difference	: The appearance, color and internal circuitry of all models are consistent, the only difference is their built-in software. Therefore, G7-P02 was tested
EUT function description	: Please reference user manual of this device
Power Supply	: DC 5V from external USB or DC 3.7V built-in lithium battery
Radio Specification	: Bluetooth V5.0
Operation Frequency	: 2402 MHz - 2480 MHz
Modulation	: GFSK
Data Rate	: 1 Mbps
Antenna Gain	: FPC antenna, maximum PK gain: 0.59 dBi
Sample Number	: S22081519-05 for conductive S22081519-06 for radiation

1.2. Assess laboratory

Dongguan Dongdian Testing Service Co., Ltd.

Add.: No. 17, Zongbu Road 2, Songshan Lake Sci&Tech, Industry Park, Dongguan City, Guangdong Province, China, 523808.

Tel.: +86-0769-38826678, <http://www.dgddt.com>, Email: ddt@dgddt.com.

CNAS Accreditation No. L6451; A2LA Accreditation Number: 3870.01

FCC Designation Number: CN1182, Test Firm Registration Number: 540522

Innovation, Science and Economic Development Canada Site Registration Number: 10288A

Conformity Assessment Body identifier: CN0048

VCCI facility registration number: C-20087, T-20088, R-20123, R-20155, G-20118

2. RF Exposure evaluation for FCC

According to 447498 D01 General RF Exposure Guidance v06

The 1-g and 10-g SAR test exclusion thresholds for 100 MHz to 6 GHz at test separation distances ≤ 50 mm are determined by:

$[(\text{max. power of channel, including tune-up tolerance, mW})/(\text{min. test separation distance, mm})] \cdot [\sqrt{f(\text{GHz})}] \leq 3.0$ for 1-g SAR and ≤ 7.5 for 10-g extremity SAR, where:

$f(\text{GHz})$ is the RF channel transmit frequency in GHz

Power and distance are rounded to the nearest mW and mm before calculation

The result is rounded to one decimal place for comparison

Manufacturing Tolerance**BLE**

GFSK (Peak) 1M			
Channel	Channel 0	Channel 39	Channel 78
Target (dBm)	-3.26	-3.81	-5.46
Tolerance ±(dB)	1	1	1

Estimtion Result

Worse case is as below: [2402 MHz, -2.26 dBm, (0.59 mW) output power]

$(0.59/5) \cdot [\sqrt{2.402(\text{GHz})}] = 0.18 < 3.0$ for 1-g SAR

Then SAR evaluation is not required.

END OF REPORT