


FCC Label and ID Label Location for Advanced Bionics Naida CI L Sound Processor

Pursuant to §2.925 of 47 C.F.R., the Advanced Bionics Naida CI L Sound Processor, which is the object of this filing, shall bear a permanently affixed, readily visible label listing the information as specified in §2.925 and §15.19(a) of 47 C.F.R. Due to the specific use of this product, its external surface is in permanent contact with the skin behind the ear. Therefore, it is not appropriate to permanently affix a label on the outer surface of the device which is in contact with the skin in order to prevent any possible skin irritations. The only possible location where the FCC ID, the hardware identification number and the serial number of the device can be placed is the battery interface shown below. This surface is smaller than 15mm X 4mm. In this case the size of the FCC ID font size will be smaller than two points, which does not satisfy the requirements for the size of the label as defined in §2.925(f).



According to §15.19(a)(5) of 47 C.F.R., the FCC ID will be placed in the Instruction for Use (IFU) document and on the device packing label for the sound processor.

The certification statement placed in the sound processor IFU is as follows:

This instrument is certified under:

- FCC ID: 2AU6O-ABBTE3