


EMC TEST REPORT FCC 47 CFR Part 15B, ISED ICES-003 Issue 6	
Report Reference No	G0M-1911-8589-EF0115B-V01
Testing Laboratory	Eurofins Product Service GmbH
Address	Storkower Str. 38c 15526 Reichenwalde Germany
Accreditation	    DAKKS - Registration number : D-PL-12092-01-03 (ISED) ISED Testing Laboratory site: 3470A-2 DAKKS - Registration number : D-PL-12092-01-04 (FCC) FCC Filed Test Laboratory, Reg.-No.: 96970
Applicant	UAB Axioma Metering
Address	Veterinaru str. 52, Biruliskes, Kaunas county 54469 Kaunas Lithuania
Test Specification Standard(s)	47 CFR Part 15 Subpart B ISED ICES-003 Issue 6 ANSI C63.4:2014
Non-Standard Test Method	None
Equipment under Test (EUT):	
Product Description	Ultrasonic water meter
Model(s)	Qalcosonic W1
Additional Model(s)	None
Brand Name(s)	None
Hardware Version(s)	v19
Software Version(s)	v1.02
FCC-ID	2AU8N-QW1
IC	25675-QW1
Test Result	PASSED

Possible test case verdicts:		
required by standard but not tested	N/T	
not required by standard	N/R	
required by standard but not appl. to test object	N/A	
test object does meet the requirement	P(PASS)	
test object does not meet the requirement	F(FAIL)	
Testing:		
Date of receipt of test item	2020-01-20	
Report:		
Compiled by	Stephan Liebich	
Tested by (+ signature) (Responsible for Test)	Stephan Liebich	
	Matthias Handrik	
Approved by (+ signature) (Deputy Head of Lab)	Jens Marquardt	
Date of Issue	2020-03-24	
Total number of pages	26	
General Remarks:		
The test results presented in this report relate only to the object tested. The results contained in this report reflect the results for this particular model and serial number. It is the responsibility of the manufacturer to ensure that all production models meet the intent of the requirements detailed within this report. This report shall not be reproduced, except in full, without the written approval of the Issuing testing laboratory.		
Additional Comments:		

ABBREVIATIONS AND ACRONYMS

Acronyms	
Acronym	Description
EUT	Equipment Under Test
FCC	Federal Communications Commission
ISED	Innovation, Science and Economic Development Canada
T _{NOM}	Nominal operating temperature
V _{NOM}	Nominal supply voltage

VERSION HISTORY

Version History			
Version	Issue Date	Remarks	Revised By
01	2020-03-24	Initial Release	-

REPORT INDEX

1	Equipment (Test Item) Under Test.....	6
1.1	Equipment Ports.....	7
1.2	Equipment Photos - Internal.....	8
1.3	Equipment Photos - External.....	10
1.4	Support Equipment.....	13
1.5	Operational Modes.....	13
1.6	EUT Configuration.....	13
1.7	Sample emission level calculation.....	14
2	Result Summary.....	15
2.1	Test Conditions and Results - Radiated emissions acc. to ANSI C63.4.....	16

1 Equipment (Test Item) Under Test

Description	Ultrasonic water meter	
Model	Qalcosonic W1	
Additional Model(s)	None	
Brand Name(s)	None	
Serial Number(s)	05064289	
Sample ID	27264	
Hardware Version(s)	v19	
Software Version(s)	v1.02	
FCC-ID	2AU8N-QW1	
IC	25675-QW1	
Class	Class B	
Equipment type	Table top	
Highest internal frequency [MHz]	928	
Radio Module	Type	NON-SPECIFIC SHORT RANGE DEVICE (LoRa)
	Model	unknown
	Manufacturer	unknown
	FCC-ID	unknown
	IC	unknown
Supply Voltage	V _{NOM}	3.3 V DC (internal battery)
AC/DC-Adaptor	None	
Manufacturer	UAB Axioma Metering Veterinaru str. 52, Biruliskes, Kaunas county 54469 Kaunas Lithuania	


1.1 Equipment Ports

Name	Type	Attributes	Comment
OPTICAL	NE	Count: 1 Direction: IO Service only: Yes	-
Description:			
AC	AC mains power input/output port		
DC	DC power input/output port		
BAT	DC power input port connected to external battery		
IO	Input/Output port		
TP	Telecommunication port		
NE	Non-electrical port		


1.2 Equipment Photos - Internal





EUT PCB TOP SIDE WITHOU DISPLAY



EUT PCB BOTTOM SIDE




1.3 Equipment Photos - External



EUT TOP SIDE



EUT BOTTOM SIDE



EUT LEFT SIDE



EUT RIGHT SIDE



1.4 Support Equipment

Product Type	Device	Manufacturer	Model	Comment
AE	Kerlink LoRa Gateway	Kerlink	Wirnet Station	Companion device
AE	PoE Adapter	Microsemi	3501G	-
AE	Optical communication interface	UAB Axioma Metering	OG-1-USB	-
MON	Laptop	lenovo	T450	-
Description:				
AE	Auxiliary Equipment			
SIM	Simulator			
MON	Monitoring Equipment			
CBL	Connecting Cable			
Comment:				

1.5 Operational Modes

Mode #	Description
1	EUT in normal working mode (measuring). Additional the EUT transmit every 6 s data packets via LoRa connection to companion device.
Comment:	

1.6 EUT Configuration

Configuration #	Description
1	EUT is powered via 3.3 V DC internal battery. EUT is connected to Kerlink LoRa Gateway via LoRa connection.
Comment:	

1.7 Sample emission level calculation

The following is a description of terms and a sample calculation, as appears in the radiated emissions data table. The numbers used in the calculation are for example only. There is no direct correlation to the specific data taken for the product described in this document:

Reading:

This is the reading obtained on the spectrum analyser in dBµV. Any external preamplifiers used are taken into account through internal analyser settings.

A.F.:

This is the antenna factor for the receiving antenna. It is a conversion factor, which converts electric fields strengths to voltages, which can be measured directly on the spectrum analyser. It is treated as a loss in dB. Cable losses have been included with the A.F. to simplify the calculations. The antenna factor is used in calculations as follows:

$$\text{Reading on Analyser (dB}\mu\text{V)} + \text{A.F. (dB/m)} = \text{Net field strength (dB}\mu\text{V/m)}$$

Net:

This is the net field strength measurement (as shown above).

Limit:

This is the FCC Class B radiated emission limit (in units of dBµV/m). The FCC limits are given in units of µV/m. The following formula is used to convert the units of µV/m to dBµV/m:

$$\text{Limit (dB}\mu\text{V/m)} = 20 \cdot \log(\mu\text{V/m})$$

Margin:

This is the margin of compliance below the FCC limit. The units are given in dB. A negative margin indicates the emission was below the limit. A positive margin indicates that the emission exceeds the limit.

Example only:

Reading + AF	= Net Reading	:	Net reading - FCC limit	= Margin
+21.5 dBµV + 26 dB/m	= 47.5 dBµV/m	:	47.5 dBµV/m - 57.0 dBµV/m	= -9.5 dB

2 Result Summary

FCC 47 CFR Part 15B, ISED ICES-003 Issue 6				
Reference	Requirement	Reference Method	Result	Remarks
Emission				
FCC 15.109 ICES-003, 8, 6.1	Radiated emissions	ANSI C63.4:2014	PASS	-
FCC 15.107 ICES-003, 8, 6.2	AC power line conducted emissions	ANSI C63.4:2014	N/R	No relevant port
Comment:				



Possible Test Case Verdicts	
PASS	Test object does meet the requirements
FAIL	Test object does not meet the requirements
N/T	Required by standard but not tested
N/R	Not required by standard for the test object



2.1 Test Conditions and Results - Radiated emissions acc. to ANSI C63.4

2.1.1 Information

Test Information	
Reference	FCC 15.109, ICES-003, 8, 6.1
Reference method	ANSI C63.4:2014 Section 8
Equipment class	Class B
Equipment type	Table top
Highest internal frequency [MHz]	928
Measurement range	30 MHz to 5 GHz
Temperature [°C]	22
Humidity [%]	27
Operator	Stephan Liebich
Date	2020-03-05

2.1.2 Setup





2.1.3 Equipment

Test Software			
Description	Manufacturer	Name	Version
EMC Software	DARE Instruments	Radimation	2016.1.10

Test Equipment					
Description	Manufacturer	Model	Identifier	Cal. Date	Cal. Due
Anechoic chamber	Frankonia	AC1	EF00062	2018-07	2021-07
EMI Test Receiver	Keysight	N9038A-526/WXP	EF01070	2019-09	2020-09
Biconical Antenna	R&S	HK 116	EF00030	2019-04	2022-04
LPD Antenna	R&S	HL 223	EF00187	2019-05	2022-05
Horn antenna	Schwarzbeck	BBHA 9120D	EF00018	2019-10	2022-10
Climatic Sensor	Embedded Data Systems, LLC.	2800100000254 17E	EF01054	2019-05	2020-05

2.1.4 Procedure

Exploratory measurement	
1.	The EUT was placed on a non-conductive table at a height of 0.8m.
2.	The EUT and support equipment, if needed, were set up to simulate typical usage.
3.	Cables, of type and length specified by the manufacturer, were connected to at least one port of each type and were terminated by a device or simulating load of actual usage.
4.	The antenna was placed at a distance of 3 or 10 m.
5.	The received signal was monitored at the measurement receiver.
6.	This procedure has to be performed in both antenna polarizations, horizontal and vertical.
7.	The arrangement of the equipment with the maximum emission level is shown on the setup picture at item 1.3

Final measurement	
1.	The EUT was placed on a 0.8 m non-conductive table at a 3 m distance from the receive antenna. The antenna output was connected to the measurement receiver.
2.	A biconical antenna was used for the frequency range 30 – 200 MHz, a logarithmic periodical antenna was used for the frequency range from 200 – 1000 MHz. Above one 1 GHz a Double Ridged Broadband Horn antenna was used. The antenna was placed on an adjustable height antenna mast.
3.	The EUT and cable arrangement were based on the exploratory measurement results.
4.	Emissions were maximized at each frequency by rotating the EUT and adjusting the receive antenna height and polarization. The maximum values were recorded.
5.	The test data of the worst-case conditions were recorded and shown on the next pages.

2.1.5 Limits



Class B @ 3 m		
Frequency [MHz]	Detector	Limit [dB μ V/m]
30 - 88	Quasi-peak	40
88 - 216	Quasi-peak	43.5
216 - 960	Quasi-peak	46
960 - 1000	Quasi-peak	54
> 1000	Peak	74
	Average	54

Class A @ 10 m		
Frequency [MHz]	Detector	Limit [dB μ V/m]
30 - 88	Quasi-peak	39
88 - 216	Quasi-peak	43.5
216 - 960	Quasi-peak	46.5
960 - 1000	Quasi-peak	49.5
> 1000	Peak	69.5
	Average	49.5


2.1.6 Results

Test Results			
Operational mode	EUT Configuration	Verdict	Remark
1	1	PASS	-


2.1.7 Setup Photos



TEST SETUP HL 223



TEST SETUP BBHA9120D




2.1.8 Records

Radiated emissions according to FCC part 15B

Project Number:	G0M-1911-8589
Applicant:	UAB Axioma Metering
Model Description:	Ultrasonic water meter
Model:	Qalcosonic W1
Test Sample ID:	27264
Test Site:	Eurofins Product Service Germany
Operator:	Mr. Liebich
Test Date:	2020-03-05
Operating Conditions:	ambient temperature: 22 power input: 3.3 V DC
Antenna:	Rohde & Schwarz HK 116, Vertical
Measurement Distance:	3m
Mode:	1
Note 1:	


Index 1



Radiated emissions according to FCC part 15B

Project Number: G0M-1911-8589
 Applicant: UAB Axioma Metering
 Model Description: Ultrasonic water meter
 Model: Qalcosonic W1
 Test Sample ID: 27264
 Test Site: Eurofins Product Service Germany
 Operator: Mr. Liebich
 Test Date: 2020-03-05
 Operating Conditions: ambient temperature: 22
 power input: 3.3 V DC
 Antenna: Rohde & Schwarz HK 116, Horizontal
 Measurement Distance: 3m
 Mode: 1
 Note 1:


Index 2



Radiated emissions according to FCC part 15B

Project Number: G0M-1911-8589
 Applicant: UAB Axioma Metering
 Model Description: Ultrasonic water meter
 Model: Qalcosonic W1
 Test Sample ID: 27264
 Test Site: Eurofins Product Service Germany
 Operator: Mr. Liebich
 Test Date: 2020-03-05
 Operating Conditions: ambient temperature: 22
 power input: 3.3 V DC
 Antenna: Rohde & Schwarz HL 223, Vertical
 Measurement Distance: 3m
 Mode: 1
 Note 1:

Index 3




Peak Number	Frequency	Quasi-Peak	Quasi-Peak Limit	Quasi-Peak Difference	Quasi-Peak Status	Angle	Height
1	902.789 MHz	LoRa-Carrier			Pass	0 Degree	1 m

Radiated emissions according to FCC part 15B

Project Number: G0M-1911-8589
 Applicant: UAB Axioma Metering
 Model Description: Ultrasonic water meter
 Model: Qalcosonic W1
 Test Sample ID: 27264
 Test Site: Eurofins Product Service Germany
 Operator: Mr. Liebich
 Test Date: 2020-03-05
 Operating Conditions: ambient temperature: 22
 power input: 3.3 V DC
 Antenna: Rohde & Schwarz HL 223, Horizontal
 Measurement Distance: 3m
 Mode: 1
 Note 1:

Index 4




Peak Number	Frequency	Quasi-Peak	Quasi-Peak Limit	Quasi-Peak Difference	Quasi-Peak Status	Angle	Height
1	902.366 MHz	LoRa-Carrier			Pass	0 Degree	1 m

Radiated emissions according to FCC part 15B

Project Number: G0M-1911-8589
 Applicant: UAB Axioma Metering
 Model Description: Ultrasonic water meter
 Model: Qalcosonic W1
 Test Sample ID: 27264
 Test Site: Eurofins Product Service Germany
 Operator: Mr. Liebich
 Test Date: 2020-03-05
 Operating Conditions: ambient temperature: 22
 power input: 3.3 V DC
 Antenna: Schwarzbeck BBHA 9120D, Vertical
 Measurement Distance: 3m
 Mode: 1
 Note 1:

Index 5



Peak Number	Frequency	Peak	Peak Limit	Peak Difference	Peak Status	Angle	Height
1	2.469 GHz	47.57 dBμV/m	73.98 dBμV/m	-26.4 dB	Pass	0 Degree	1 m
2	3.566 GHz	45.02 dBμV/m	73.98 dBμV/m	-28.96 dB	Pass	0 Degree	1 m

Peak Number	Frequency	Average	Average Limit	Average Difference	Average Status	Angle	Height
1	2.469 GHz	29.66 dBμV/m	53.98 dBμV/m	-24.31 dB	Pass	0 Degree	1 m
2	3.566 GHz	32.53 dBμV/m	53.98 dBμV/m	-21.45 dB	Pass	0 Degree	1 m


Test Report No.: G0M-1911-8589-EF0115B-V01

Eurofins Product Service GmbH
 Storkower Str. 38c, D-15526 Reichenwalde, Germany

Radiated emissions according to FCC part 15B

Project Number: G0M-1911-8589
 Applicant: UAB Axioma Metering
 Model Description: Ultrasonic water meter
 Model: Qalcosonic W1
 Test Sample ID: 27264
 Test Site: Eurofins Product Service Germany
 Operator: Mr. Liebich
 Test Date: 2020-03-05
 Operating Conditions: ambient temperature: 22
 power input: 3.3 V DC
 Antenna: Schwarzbeck BBHA 9120D, Horizontal
 Measurement Distance: 3m
 Mode: 1
 Note 1:

Index 6



Peak Number	Frequency	Peak	Peak Limit	Peak Difference	Peak Status	Angle	Height
1	3.569 GHz	45.6 dBµV/m	73.98 dBµV/m	-28.38 dB	Pass	0 Degree	1 m

Peak Number	Frequency	Average	Average Limit	Average Difference	Average Status	Angle	Height
1	3.569 GHz	32.45 dBµV/m	53.98 dBµV/m	-21.53 dB	Pass	0 Degree	1 m