

Maximum Permissible Exposure Report

Product : UASure II Link Blood Uric Acid Monitoring System

Model Name : UAM302

FCC ID : 2AV24UAM302

Test Regulation : 47 CFR FCC Part 2.1091

Received Date : 2021/3/20

Test Date : 2021/3/22 ~ 2021/3/29

Issued Date : 2023/11/3

Applicant : Apex Biotechnology Corp.
No. 7, Li-Hsin Rd. V, Hsinchu Science Park, Hsinchu, Taiwan,
R.O.C.


Issued By : Underwriters Laboratories Taiwan Co., Ltd.
Building A, B and E, No. 372-7, Sec. 4, Zhongxing Rd.,
Zhudong Township, Hsinchu County, Taiwan



The results reported herein have been performed in accordance with the laboratory's terms of accreditation. This report shall not be reproduced except in full without the written approval of the Laboratory. The results in this report are responsible of the test sample(s) provided by the client only and are not to be used to indicate applicability to other similar products.

Underwriters Laboratories Taiwan Co., Ltd.

Building A, B and E, No. 372-7, Sec. 4, Zhongxing Rd., Zhudong Township, Hsinchu County, Taiwan
Telephone : +886-2-7737-3000
Facsimile (FAX) : +886-3-583-7948



Original Test Report No.: 4790996517-US-R1-V0

[illegible]

Building A, B and E, No. 372-7, Sec. 4, Zhongxing Rd., Zhudong Township, Hsinchu County, Taiwan
Telephone :+886-2-7737-3000
Facsimile (FAX) :+886-3-583-7948 Doc No: Form-ULID-004725 (DCS:17-

Doc No: Form-ULID-004725 (DCS:17-EM-F0864) / 5.1

Table of Contents

1. Attestation of Test Results	4
2. Test Methodology and Reference Procedures.....	5
3. Facilities and Accreditation	5
4. Equipment Under Test.....	6
4.1. Description of EUT	6
4.2. Description of Available Antennas.....	7
5. Requirement.....	8
6. General RF Exposure Test Exemption.....	9
7. Radio Frequency Radiation Exposure Evaluation	11

1. Attestation of Test Results

APPLICANT: Apex Biotechnology Corp.
No. 7, Li-Hsin Rd. V, Hsinchu Science Park, Hsinchu, Taiwan, R.O.C.

MANUFACTURER: Apex Biotechnology Corp.
No. 7, Li-Hsin Rd. V, Hsinchu Science Park, Hsinchu, Taiwan, R.O.C.

EUT DESCRIPTION: UASure II Link Blood Uric Acid Monitoring System

BRAND: ApexBio

MODEL: UAM302

SAMPLE STAGE: Engineering Verification Test sample

APPLICABLE STANDARDS	
STANDARD	Test Results
47 CFR FCC Part 2.1091	PASS

Underwriters Laboratories Taiwan Co., Ltd. tested the above equipment in accordance with the requirements set forth in the above standards. All indications of Pass/Fail in this report are opinions expressed by Underwriters Laboratories Taiwan Co., Ltd. based on interpretations and/or observations of test results. The test results show that the equipment tested is capable of demonstrating compliance with the requirements as documented in this report.

Note: The results documented in this report apply only to the tested sample, under the conditions and modes of operation as described herein. This document may not be altered or revised in any way unless done so by Underwriters Laboratories Taiwan Co., Ltd. and all revisions are duly noted in the revisions section. Any alteration of this document not carried out by Underwriters Laboratories Taiwan Co., Ltd. will constitute fraud and shall nullify the document. This report must not be used by the client to claim product certification, approval, or endorsement by NVLAP, NIST, any agency of the Federal Government, or any agency of any government.

Prepared By:



Sally Lu
Project Handler

Date : 2023/11/3

Approved and Authorized By:



Eric Lee
Senior Laboratory Engineer

Date : 2023/11/3

Underwriters Laboratories Taiwan Co., Ltd.

Building A, B and E, No. 372-7, Sec. 4, Zhongxing Rd., Zhudong Township, Hsinchu County, Taiwan

Telephone : +886-2-7737-3000

Facsimile (FAX) : +886-3-583-7948

Doc No: Form-ULID-004725 (DCS:17-EM-F0864) / 5.1

2. Test Methodology and Reference Procedures

The tests documented in this report were performed in accordance with KDB 447498 D04 Interim General RF Exposure Guidance v01.

3. Facilities and Accreditation

Test Location	Underwriters Laboratories Taiwan Co., Ltd.
Address	Building A, B and E, No. 372-7, Sec. 4, Zhongxing Rd., Zhudong Township, Hsinchu County, Taiwan
Accreditation Certificate	Underwriters Laboratories Taiwan Co., Ltd. is accredited by TAF, Laboratory Code 3398.

Underwriters Laboratories Taiwan Co., Ltd.

Building A, B and E, No. 372-7, Sec. 4, Zhongxing Rd., Zhudong Township, Hsinchu County, Taiwan

Telephone :+886-2-7737-3000

Facsimile (FAX) :+886-3-583-7948

Doc No: Form-ULID-004725 (DCS:17-EM-F0864) / 5.1

4. Equipment Under Test

4.1. Description of EUT

Product Name	UASure II Link Blood Uric Acid Monitoring System	
Brand Name	ApexBio	
Model Name	UAM302	
Operating Frequency	Bluetooth LE	2402MHz ~ 2480MHz
Modulation	Bluetooth LE	GFSK
Number of Channel	Bluetooth LE	40
Normal Voltage	5Vdc from Host 3Vdc from battery	
S/N	3679494	

Note:

1. The EUT contains following accessory devices:

Product	Brand	Model	Description
Micro USB cable	iPro Technology Inc.	CP2104-F03-EK-MTMC	Length: 1 m
*Battery	TOSHIBA	CR2032	3V
*Battery	Mitsubishi	CR2032	3V
*Battery	MAXELL	CR2032	3V

*The battery's difference is only brand, specification are same. Therefore only evaluates TOSHIBA battery as representative.

2. For this report measurement uncertainty, statement of conformity, determining compliance, it is necessary to refer to the original measurement report of EUT.
3. The above EUT information is declared by manufacturer and for more detailed features description, please refer the manufacturer's or user's manual, the laboratory shall not be held responsible.

Underwriters Laboratories Taiwan Co., Ltd.

Building A, B and E, No. 372-7, Sec. 4, Zhongxing Rd., Zhudong Township, Hsinchu County, Taiwan

Telephone : +886-2-7737-3000

Facsimile (FAX) : +886-3-583-7948

Doc No: Form-ULID-004725 (DCS:17-EM-F0864) / 5.1

4.2. Description of Available Antennas

Ant. No.	Transmitter Circuit	Brand Name	Model Name	Ant. Type	Maximum Gain (dBi)
1	Chain (0)	APEX BIOTECHNOLOGY	APEX_UAM302 PCB ANT	PCB	0.87

Note: The above antenna information was provided from customer and for more detailed features description, please refer the manufacturer's specification or user's manual, the laboratory shall not be held responsible.

5. Requirement

Limits for General Population/Uncontrolled Exposure

Limits for General Population/Uncontrolled Exposure				
Frequency Range (MHz)	Electric Field Strength (E) (V/m)	Magnetic Field Strength (H) (A/m)	Power Density (S) (mW/cm ²)	Averaging Time E ² , H ² or S (minutes)
0.3-1.34	614	1.63	*100	30
1.34-30	824/f	2.19/f	*180/f ²	30
30-300	27.5	0.073	0.2	30
300-1500	--	--	f/1500	30
1500-100,000	--	--	1.0	30
Note 1: f = frequency in MHz, * means Plane-wave equivalent power density				
Note 2: General population/uncontrolled exposures apply in situations in which the general public may be exposed, or in which persons that are exposed as a consequence of their employment may not be fully aware of the potential for exposure or cannot exercise control over their exposure.				

Power Density (S) is calculated by the following formula:

$$S = (P \cdot G) / 4\pi R^2$$

where: S = power density (in appropriate units, e.g. mW/ cm²)

P = power input to the antenna (in appropriate units, e.g., mW)

G = power gain of the antenna in the direction of interest relative to an isotropic radiator R = distance to the center of radiation of the antenna (appropriate units, e.g., cm)

Underwriters Laboratories Taiwan Co., Ltd.

Building A, B and E, No. 372-7, Sec. 4, Zhongxing Rd., Zhudong Township, Hsinchu County, Taiwan

Telephone : +886-2-7737-3000

Facsimile (FAX) : +886-3-583-7948

Doc No: Form-ULID-004725 (DCS:17-EM-F0864) / 5.1

6. General RF Exposure Test Exemption

The corresponding Exclusion Threshold condition, listed below:

- 1) Blanket Exempt: Following 47 CFR 1.1307(b)(3)(i)(A), the available maximum time-averaged power is no more than 1 mW.
- 2) SAR Exempt: Following 47 CFR 1.1307(b)(3)(i)(B), the available maximum time-averaged power or effective radiated power (ERP), whichever is greater, is less than or equal to the threshold P_{th} (mW) described in the following formula. This method shall only be used at separation distances (cm) from 0.5 centimeters to 40 centimeters and at frequencies from 0.3 GHz to 6 GHz (inclusive). P_{th} is given by:

$$P_{th} \text{ (mW)} = \begin{cases} ERP_{20 \text{ cm}} (d/20 \text{ cm})^x & d \leq 20 \text{ cm} \\ ERP_{20 \text{ cm}} & 20 \text{ cm} < d \leq 40 \text{ cm} \end{cases}$$

Where

$$x = -\log_{10} \left(\frac{60}{ERP_{20 \text{ cm}} \sqrt{f}} \right) \text{ and } f \text{ is in GHz;}$$

and

$$ERP_{20 \text{ cm}} \text{ (mW)} = \begin{cases} 2040f & 0.3 \text{ GHz} \leq f < 1.5 \text{ GHz} \\ 3060 & 1.5 \text{ GHz} \leq f \leq 6 \text{ GHz} \end{cases}$$

d = the separation distance (cm);

Underwriters Laboratories Taiwan Co., Ltd.

Building A, B and E, No. 372-7, Sec. 4, Zhongxing Rd., Zhudong Township, Hsinchu County, Taiwan

Telephone : +886-2-7737-3000

Facsimile (FAX) : +886-3-583-7948

Doc No: Form-ULID-004725 (DCS:17-EM-F0864) / 5.1

- 3) MPE Exempt: Following 47 CFR 1.1307(b)(3)(i)(C), using Table 1 and the minimum separation distance (R in meters) from the body of a nearby person for the frequency (f in MHz) at which the source operates, the ERP (watts) is no more than the calculated value prescribed for that frequency. For the exemption in Table 1 to apply, R must be at least $\lambda/2\pi$, where λ is the free-space operating wavelength in meters. If the ERP of a single RF source is not easily obtained, then the available maximum time-averaged power may be used in lieu of ERP if the physical dimensions of the radiating structure(s) do not exceed the electrical length of $\lambda/4$ or if the antenna gain is less than that of a half-wave dipole (1.64 linear value).

Table 1 to § 1.1307(b)(3)(i)(C) - Single RF Sources Subject to Routine Environmental Evaluation

RF Source frequency (MHz)	Threshold ERP (watts)
0.3-1.34	$1,920 R^2$.
1.34-30	$3,450 R^2/f^2$.
30-300	$3.83 R^2$.
300-1,500	$0.0128 R^2 f$.
1,500-100,000	$19.2 R^2$.

Underwriters Laboratories Taiwan Co., Ltd.

Building A, B and E, No. 372-7, Sec. 4, Zhongxing Rd., Zhudong Township, Hsinchu County, Taiwan

Telephone : +886-2-7737-3000

Facsimile (FAX) : +886-3-583-7948

Doc No: Form-ULID-004725 (DCS:17-EM-F0864) / 5.1

7. Radio Frequency Radiation Exposure Evaluation

(1) General RF Exposure Test Exemption

Option	Evaluation Method	Clause
☐	Blanket Exempt	47 CFR 1.1307(b)(3)(i)(A)
☒	SAR Exempt	47 CFR 1.1307(b)(3)(i)(B)
☐	MPE Exempt	47 CFR 1.1307(b)(3)(i)(C)

Note: Max. ERP (dBm) = Max. Average power (dBm) + Antenna Gain (dBi) - 2.15 (dB)

Bluetooth LE

Operating Mode	Evaluation Frequency	Max. Average Power	Antenna Gain	Min. Test Separation Distance	Max. ERP Power	Limits of SAR Exempt	Result
	(MHz)	(mW)	(dBi)	(cm)	(mW)	(mW)	
BT LE(1Mbps)	2402 -2480	1.718	0.87	0.5	1.279	2.71	PASS
BT LE(2Mbps)	2402 -2480	2.377	0.87	0.5	1.77	2.71	PASS

Note:

1. Max. ERP (dBm) = Max. Average power (dBm) + Antenna Gain (dBi) -2.15
2. Max. ERP (mW) = $10^{(\text{Max. ERP (dBm)} / 10)}$
3. Limits of SAR Exempt = $ERP_{20cm} (d/20cm)^x$
 $x = -\log_{10} \left(\frac{60}{ERP_{20cm} \cdot \sqrt{f}} \right)$ and f is in GHz
 $ERP_{20cm} (mW) = 3060$

Conclusion:

According to 47 CFR §2.1091, the RF exposure analysis concludes that the RF Exposure is FCC compliant.

END OF REPORT

Underwriters Laboratories Taiwan Co., Ltd.

Building A, B and E, No. 372-7, Sec. 4, Zhongxing Rd., Zhudong Township, Hsinchu County, Taiwan
Telephone : +886-2-7737-3000
Facsimile (FAX) : +886-3-583-7948

Doc No: Form-ULID-004725 (DCS:17-EM-F0864) / 5.1