

## Cover Letter-Wireless Charger Approval

FCC ID: 2AV76-NMOK-301W

Date: 02 Sep, 2022

Gentlemen:

There's a NMOK-301W that would like to have your authorization as an Inductive wireless power transfer applications approval.

The specific product as below, NMOK-301W, with its designed features and specified description, considers special requirements for KDB 680106 D01 v03 section 5, b) requirements.

|                      |                                  |
|----------------------|----------------------------------|
| <b>Company:</b>      | NIDEC MOBILITY KOREA CORPORATION |
| <b>Product Name:</b> | WIRELESS POWER CHARGING SYSTEM   |
| <b>Model Number:</b> | NMOK-301W                        |
| <b>FCC ID:</b>       | 2AV76-NMOK-301W                  |

| KDB 680106 D01 v03 Section 5, b) Requirements:                                                                                                                                                      | Product Technical Specification:                                       | Result:  |
|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------|----------|
| (1) Power transfer frequency is less than 1 MHz                                                                                                                                                     | 0.120 MHz                                                              | Complied |
| (2) Output power from each primary coil is less than or equal 15 watts.                                                                                                                             | 10 Watts                                                               | Complied |
| (3) The system may consist of more than one source primary coils, charging one or more clients. If more than one primary coil is present, the coil pairs may be powered on at the same time.        | The transfer system includes only single primary coil.                 | Complied |
| (4) Client device is placed directly in contact with the transmitter                                                                                                                                | Client device is placed directly in contact with the transmitter.      | Complied |
| (5) Mobile exposure conditions only (portable exposure conditions are not covered by this exclusion).                                                                                               | This is a portable device that need PAG with 50mm separation distance. | N/A      |
| (6) The aggregate H-field strengths at 15 cm surrounding the device and 20 cm above the top surface from all simultaneous transmitting coils are demonstrated to be less than 50% of the MPE limit. | H-Field: 0.289 A/m (at 15 cm)<br>< 0.815 A/m (50% of the limit)        | Complied |

Sincerely,

By:



(Signature1)

Jinhyoung Cho

(Print name)

Title:

Technical Manager

On behalf of: SGS Korea Co., Ltd. (Company Name)

Telephone: +82-31-428-0903