

Appendix C for KSCR220900176501

Calibration Certificate

Object	Apply	No	Model	SN	Calibration Date
Dipole	☐	1	CLA150	4025	2021/04/26
	☐	2	D450V3	1103	2021/04/21
	☐	3	D750V3	1188	2022/03/29
	☐	4	D835V2	4d114	2022/03/31
	☐	5	D900V2	1d079	2022/06/07
	☐	6	D1800V2	2d170	2022/03/31
	☐	7	D1900V2	5d136	2022/06/07
	☐	8	D2000V2	1041	2022/06/06
	☐	9	D2300V2	1096	2022/03/31
	☒	10	D2450V2	817	2022/04/01
	☐	11	D2600V2	1158	2022/03/31
	☐	12	D5GHzV2	1095	2022/06/01
DAE	☒	13	DAE4	1245	2022/05/30
Probe	☒	14	EX3DV4	7346	2022/03/30



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

1 Dipole

1.1 CLA150 - SN 4025

Calibration Laboratory of Schmid & Partner Engineering AG Zugstrasse 43, 8004 Zurich, Switzerland		S Schweizerischer Kalibrierdienst Service suisse d'étalonnage Servizio svizzero di taratura S Swiss Calibration Service	
Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates		Accreditation No.: SCS 0108	
Client: SGS-CN (Audeon)		Certificate No.: CLA150-4025_Apr21	
CALIBRATION CERTIFICATE			
Object: CLA150 - SN: 4025			
Calibration procedure(s): QA CAL-15.v9 Calibration Procedure for SAR Validation Sources below 700 MHz			
Calibration date: April 26, 2021			
The calibration certificate documents the traceability to national standards, which require the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and on part of the certificate.			
All calibrations have been conducted in the closed laboratory facility; environment temperature (22 ± 3)°C and humidity < 70%.			
Calibration Equipment Used (M&E critical for calibration)			
Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration
Power meter NRP	SN: 10475	09-Apr-21 (No. 217-03001/03002)	Apr-22
Power sensor NRP Z01	SN: 100344	09-Apr-21 (No. 217-03001)	Apr-22
Power sensor NRP Z01	SN: 100345	09-Apr-21 (No. 217-03002)	Apr-22
Reference 20 dB Attenuator	SN: CC2002 (200)	09-Apr-21 (No. 217-03003)	Apr-22
Type-N mismatch combination	SN: 310002 / 0007	09-Apr-21 (No. 217-03004)	Apr-22
Reference Probe EX3004	SN: 3877	30-Dec-20 (No. 203-3077) Dec-20	Dec-21
EXE4	SN: 661	16-Jan-20 (No. 19459-659-Jan20)	Jan-21
Secondary Standards	ID #	Check Date (in house)	Scheduled Check
Power meter S44195	SN: G341303024	09-Apr-15 (in house check Jun-20)	In house check Jun-20
Power sensor E413A	SN: MY4169007	09-Apr-15 (in house check Jun-20)	In house check Jun-20
Power sensor E413A	SN: 00010010	09-Apr-15 (in house check Jun-20)	In house check Jun-20
RF generator HP 8594C	SN: U504001709	04-Aug-19 (in house check Jun-20)	In house check Jun-20
Network Analyzer Agilent T8368A	SN: U541000477	31-Mar-14 (in house check Oct-19)	In house check Oct-21
Calibrated by:	Johann Katschauer	Function: Laboratory Technician	Signature: [Signature]
Approved by:	Kajko Polakovic	Function: Technical Manager	Signature: [Signature]
This calibration certificate shall not be reproduced except in full without written approval of the laboratory. Issued: April 26, 2021			
Certificate No: CLA150-4025_Apr21 Page 1 of 5			

Calibration Laboratory of Schmid & Partner Engineering AG Zugstrasse 43, 8004 Zurich, Switzerland		S Schweizerischer Kalibrierdienst Service suisse d'étalonnage Servizio svizzero di taratura S Swiss Calibration Service	
Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates		Accreditation No.: SCS 0108	
Glossary:			
TSL	Issue simulating liquid		
CorrF	sensitivity in TSL / NORM x,y,z		
N/A	not applicable or not measured		
Calibration is Performed According to the Following Standards:			
a) IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Technique", June 2013			
b) IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016			
c) IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010			
d) KDB 865864, "SAR Measurement Requirements for 100 MHz to 6 GHz"			
Additional Documentation:			
e) DASY4/5 System Handbook			
Methods Applied and Interpretation of Parameters:			
• Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.			
• Antenna Parameters with TSL: The source is mounted in a touch configuration below the center marking of the flat phantom.			
• Return Loss: This parameter is measured with the source positioned under the liquid filled phantom (as described in the measurement condition clause). The Return Loss ensures low reflected power. No uncertainty required.			
• SAR measured: SAR measured at the stated antenna input power.			
• SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.			
• SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.			
The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.			
Certificate No: CLA150-4025_Apr21 Page 2 of 5			

Measurement Conditions			
DASY system configuration, as far as not given on page 1.			
DASY Version	DASY5	V52.10.4	
Extrapolation	Advanced Extrapolation		
Phantom	ELIA Flat Phantom	Shell thickness: 2 ± 0.2 mm	
EUT Positioning	Touch Position		
Zoom Scan Resolution	dx, dy = 4.0 mm, dz = 1.4 mm	Graded Ratio = 1.4 (Z direction)	
Frequency	150 MHz ± 1 MHz		

Head TSL parameters			
The following parameters and calculations were applied.			
	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	52.3	0.75 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	51.1 ± 6 %	0.75 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C	---	---

SAR result with Head TSL		
SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	1 W input power	3.90 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	3.88 W/kg ± 19.4 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Head TSL		
SAR measured	Condition	
SAR measured	1 W input power	2.60 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	2.59 W/kg ± 18.0 % (k=2)

Appendix (Additional assessments outside the scope of SCS 0108)	
Antenna Parameters with Head TSL	
Impedance, transformed to feed point	47.9 Ω ± 1.5 Ω
Return Loss	-31.4 dB

Additional EUT Data	
Manufactured by	SPLEAG

Certificate No: CLA150-4025_Apr21 Page 3 of 5

Certificate No: CLA150-4025_Apr21 Page 4 of 5




Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn

t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com



1.2 D450V3 - SN 1103

Calibration Laboratory of Schmid & Partner Engineering AG
Zugstrasse 43, 8004 Zurich, Switzerland

Accredited by the Swiss Accreditation Service (SAS)
The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Client: **SGS-CN (Aiden)** Certificate No: **D450V3-1103_Apr21**

CALIBRATION CERTIFICATE

Object: **D450V3 - SN: 1103**

Calibration procedure(s): **QA CAL-15-V9**
Calibration Procedure for SAR Validation Sources below 700 MHz

Calibration date: **April 21, 2021**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (20 ± 0.5) °C and humidity < 75%.

Calibration Equipment used (MPE critical for calibration)

Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration
Power meter NRP	SN 104779	09-Apr-21 (No. 217-030110300)	Apr-22
Power sensor MHP-Z91	SN 103344	09-Apr-21 (No. 217-03011)	Apr-22
Power sensor MHP-Z91	SN 103345	09-Apr-21 (No. 217-03015)	Apr-22
Reference 20 dB Attenuator	SN C2052 (200)	09-Apr-21 (No. 217-03049)	Apr-22
Type-N measurement combination	SN 310822 / 06327	09-Apr-21 (No. 217-03044)	Apr-22
Reference Probe EX3DV4	SN 3877	30-Dec-20 (No. EX3-2077 Dec20)	Dec-21
DA54	SN 454	06-Jun-20 (No. DA54-654-Jun20)	Jun-21

Secondary Standards

ID #	Cal Date (In House)	Scheduled Check
Power meter E41188	06-Apr-19 (In house check Jun-20)	In house check Jun-20
Power sensor E4112A	06-Apr-19 (In house check Jun-20)	In house check Jun-20
Power sensor E4112A	06-Apr-19 (In house check Jun-20)	In house check Jun-20
RF generator HP 8448C	06-Aug-19 (In house check Jun-20)	In house check Jun-20
Network Analyzer Agilent E8358A	31-Mar-14 (In house check Dec-20)	In house check Oct-21

Calibrated by: **Kuba Fomke** Function: **Laboratory Technician**

Approved by: **Kuba Fomke** Technical Manager

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Issued: April 23, 2021

Certificate No: D450V3-1103_Apr21 Page 1 of 6

Calibration Laboratory of Schmid & Partner Engineering AG
Zugstrasse 43, 8004 Zurich, Switzerland

Accredited by the Swiss Accreditation Service (SAS)
The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Client: **SGS-CN (Aiden)** Certificate No: **D450V3-1103_Apr21**

Glossary:

TSL: Issue simulating liquid sensitivity in TSL / NORM x,y,z

ConvF: not applicable or not measured

N/A: not applicable or not measured

Calibration is Performed According to the Following Standards:

- IEC 61010-1:2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2015
- IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- KDB 665664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:

e) DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: D450V3-1103_Apr21 Page 2 of 6



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx>; and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com
No.10, Weyue Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

Measurement Conditions		
DASY system configuration, as far as not given on page 1.		
DASY Version	DASY5	V82.10.4
Extrapolation	Advanced Extrapolation	
Phantom	ELJ Flat Phantom	Shell thickness: 2 ± 0.2 mm
Distance Dipole Center - TSL	15 mm	with Spacer
Zoom Scan Resolution	$dx, dy, dz \leq 5$ mm	
Frequency	450 MHz ± 1 MHz	

Head TSL parameters			
The following parameters and calculations were applied:			
	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	43.5	0.57 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	$43.1 \pm 6 \%$	0.07 mho/m $\pm 8 \%$
Head TSL temperature change during test	< 0.5 °C	---	---

SAR result with Head TSL		
SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	1.14 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	4.56 W/kg $\pm 18.1 \%$ (k=2)

SAR result with Head TSL		
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	
SAR measured	250 mW input power	0.757 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	3.06 W/kg $\pm 17.6 \%$ (k=2)

Appendix (Additional assessments outside the scope of SCS 0106)	
Antenna Parameters with Head TSL	
Impedance, transformed to feed point	57.1 $\Omega - 2.8 j\Omega$
Return Loss	-23.0 dB

General Antenna Parameters and Design	
Electrical Delay (one direction)	1.346 ns

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard overlight coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC signals. On some of the dipoles, small and caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data	
Manufactured by	SPEAG

Certificate No: D450V3-1103_Apr21

Page 5 of 6

Certificate No: D450V3-1103_Apr21

Page 4 of 6

DASY5 Validation Report for Head TSL

Date: 21.04.2021

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 450 MHz; Type: D450V3; Serial: D450V3 - SN:1103

Communication System: UTD 0 - CW; Frequency: 450 MHz

Medium parameters used: $f = 450$ MHz, $\alpha = 0.87$ S/m; $\epsilon_r = 43.1$; $\rho = 1000$ kg/m³

Phantom section: Flat Section

Measurement Standard: DASY5 (IEC/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 - SN3877; ConvF [10.64, 10.64] @ 450 MHz; Calibrated: 30.12.2020
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn654; Calibrated: 26.06.2020
- Phantom: ELJ v4.0; Type: QDOVA001BB; Serial: TP:1003
- DASY52 52.10.4 (1527); SEMCAD X (4.6.14/7483)

Dipole Calibration for Head Tissue/d=15mm, Pin=250mW/Zoom Scan (7x7x7)/Cube 0:

Measurement grid: $dx=5$ mm, $dy=5$ mm, $dz=5$ mm

Reference Value = 39.18 V/m; Power DUT = 0.08 dB


Peak SAR (extrapolated) = 1.76 W/kg

SAR(1 g) = 1.14 W/kg; SAR(10 g) = 0.767 W/kg

Smallest distance from peaks to all points 3 dB below: Larger than measurement grid

Ratio of SAR at M2 to SAR at M1 = 64.9%

Maximum value of SAR (measured) = 1.53 W/kg




Certificate No: D450V3-1103_Apr21

Page 5 of 6

Certificate No: D450V3-1103_Apr21

Page 6 of 6



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

1.3 D750V3 - SN 1188

TTL Speag CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanli Road, Haidian District, Beijing, 100191 Tel: +86-10-4239613-2172 Fax: +86-10-4239613-2564 E-mail: cti@china-test.com http://www.china-test.com		中国合格评定国家认可委员会 CALIBRATION CNAS L0570	
Client	SGS-CN	Certificate No.	Z22-60103
CALIBRATION CERTIFICATE			
Object	D750V3 - SN: 1188		
Calibration Procedure(s)	FF-Z11-003-01 Calibration Procedures for dipole validation kits		
Calibration date:	March 28, 2022		
This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.			
All calibrations have been conducted in the closed laboratory facility; environment temperature (22±3)°C and humidity <70%.			
Calibration Equipment used (M&TE critical for calibration)			
Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Power Meter NRP2	104277	24-Sep-21 (CTTL No. J21X08328)	Sep-22
Power sensor NRP88	104291	24-Sep-21 (CTTL No. J21X08328)	Sep-22
Reference Probe EX30V4	SN 7307	26-May-21 (SPEAG No. EX3-7307_May21)	May-22
DAE4	SN 1556	12-Jan-22 (CTTL-SPEAG No. Z22-60007)	Jan-23
Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Signal Generator S4436C	MY46071430	13-Jan-22 (CTTL No. J22X00406)	Jan-23
Network Analyzer E5071C	MY46110673	14-Jan-22 (CTTL No. J22X00406)	Jan-23
Calibrated by:	Name	Function	Signature
	Zhao Jing	SAR Test Engineer	
Reviewed by:	Lin Hao	SAR Test Engineer	
Approved by:	Qi Dianyun	SAR Project Leader	
Issued: April 3, 2022			
This calibration certificate shall not be reproduced except in full without written approval of the laboratory.			
Certificate No: Z22-60103		Page 1 of 6	

TTL Speag CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanli Road, Haidian District, Beijing, 100191, China Tel: +86-10-4239613-2079 Fax: +86-10-4239613-2564 E-mail: cti@china-test.com http://www.china-test.com		中国合格评定国家认可委员会 CALIBRATION CNAS L0570	
Measurement Conditions DASY system configuration, as far as not given on page 1:			
DASY Version	DASY52	V52.10.4	
Extrapolation	Advanced Extrapolation		
Phantom	Triple Flat Phantom 5.1C		
Distance Dipole Center - TSL	15 mm	with Spacer	
Zoom Scan Resolution	dx, dy, dz = 5 mm		
Frequency	750 MHz ± 1 MHz		
Head TSL parameters The following parameters and calculations were applied:			
	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	42.0	0.90 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	41.4 ± 6 %	0.89 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---
SAR result with Head TSL			
SAR averaged over 1 cm ² (1 g) of Head TSL	Condition		
SAR measured	250 mW input power	2.07 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	8.27 W/kg ± 18.6 % (k=2)	
SAR averaged over 10 cm ² (10 g) of Head TSL	Condition		
SAR measured	250 mW input power	1.37 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	5.48 W/kg ± 18.7 % (k=2)	
Certificate No: Z22-60103		Page 3 of 6	

TTL Speag CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanli Road, Haidian District, Beijing, 100191, China Tel: +86-10-4239613-2079 Fax: +86-10-4239613-2564 E-mail: cti@china-test.com http://www.china-test.com		中国合格评定国家认可委员会 CALIBRATION CNAS L0570	
Glossary:			
TSL	Issue simulating liquid		
ConvF	sensitivity in TSL / NORMx.y.z		
N/A	not applicable or not measured		
Calibration is Performed According to the Following Standards:			
a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020			
b) KDB 855654, "SAR Measurement Requirements for 100 MHz to 6 GHz"			
Additional Documentation:			
c) DASY4/5 System Handbook			
Methods Applied and Interpretation of Parameters:			
• Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.			
• Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.			
• Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.			
• Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.			
• SAR measured: SAR measured at the stated antenna input power.			
• SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.			
• SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.			
The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.			
Certificate No: Z22-60103		Page 1 of 6	

TTL Speag CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanli Road, Haidian District, Beijing, 100191, China Tel: +86-10-4239613-2079 Fax: +86-10-4239613-2564 E-mail: cti@china-test.com http://www.china-test.com		中国合格评定国家认可委员会 CALIBRATION CNAS L0570	
Appendix (Additional assessments outside the scope of CNAS L0570)			
Antenna Parameters with Head TSL			
Impedance, transformed to feed point	53.60- 1.13jΩ		
Return Loss	>28.7dB		
General Antenna Parameters and Design			
Electrical Delay (one direction)	0.947 ns		
After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.			
The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small and caps are added to the dipole arms in order to improve matching when placed according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.			
No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.			
Additional EUT Data			
Manufactured by	SPEAG		
Certificate No: Z22-60103		Page 4 of 6	





Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx>; and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.


Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com



1.4 D835V2 - SN 4d114

 in Collaboration with TTLSPEAG CALIBRATION LABORATORY Addr: No.32 HuaYuanJie, Haidian District, Beijing, 100191 Tel: +86-10-42304033-2312 Fax: +86-10-42304813-2504 E-mail: cttl@cttl.com.cn http://www.cttl.net.cn	  中国认证 认可 校准 实验室 CNAS L8070	CAICT 中国信息通信研究院
Client SOS-CN	Certificate No: 223-60104	
CALIBRATION CERTIFICATE		
Object	D835V2 - SN: 41114	
Calibration Procedure(s)	FF-Z11-003-01 Calibration Procedures for dipole validation kits	
Calibration date	March 31, 2022	
This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.		
All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3)°C and humidity<70%.		
Calibration Equipment used (MATE critical for calibration)		
Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)
Power Meter NRP/2	100277	24-Sep-21 (CTTL, No.J21X08326)
Power sensor NRP/BS	104291	24-Sep-21 (CTTL, No.J21X08326)
Reference Probe EX3D/VS	SN 7307	26-May-21 (SPEAG, No.EX3-7307_May21)
DAA4	SN 1556	12-Jan-22 (CTTL-SPEAG, No.Z22-60007)
Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)
Signal Generator E438C	MY49071430	13-Jan-22 (CTTL, No.J22X04059)
Network Analyzer ES071C	MY46110673	14-Jan-22 (CTTL, No.J22X04058)
Calibrated by:	Name	Function
	Zhao Jing	SAR Test Engineer
Reviewed by:	Lin Hao	SAR Test Engineer
Approved by:	Qi Dianyan	SAR Project Leader
Issued: April 6, 2022		Signature:   
This calibration certificate shall not be reproduced except in full without written approval of the laboratory.		

 In Collaboration with: 3 g TCALIBRATION LABORATORY Address: 82 Newmarket Road, Hadden (Hertford, England) HGF9 9TQ Tel: +44 (0) 1462 304633-2079 Fax: +44 (0) 1462 304633-2504 E-mail: info@tcalib.com http://www.tcalib.com									
Glossary: <table border="0"> <tr> <td>TSL</td> <td>issue simulating liquid</td> </tr> <tr> <td>ConvF</td> <td>sensitivity in TSL / NORM_{FS} y,z</td> </tr> <tr> <td>N/A</td> <td>not applicable or not measured</td> </tr> </table>				TSL	issue simulating liquid	ConvF	sensitivity in TSL / NORM _{FS} y,z	N/A	not applicable or not measured
TSL	issue simulating liquid								
ConvF	sensitivity in TSL / NORM _{FS} y,z								
N/A	not applicable or not measured								
Calibration is Performed According to the Following Standards: a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures" (Frequency range of 4 MHz to 10 GHz), October 2020 b) KDB 865864, "SAR Measurement Requirements for 100 MHz to 6 GHz"									
Additional Documentation: a) DASY4/5 System Handbook									
Methods Applied and Interpretation of Parameters: • Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated. • Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis. • Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required. • Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required. • SAR measured: SAR measured at the stated antenna input power. • SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector. • SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.									
The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.									



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed hereon, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx>, and for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of its intervention only and within the limits of the Company's instructions, if any. The Company's intervention is to be limited and its conclusions are to be restricted to the facts and transactions arising at their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: For any request for testing/inspection report or certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com.

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sqs.china@sqs.com

In Collaboration with TTL s p e a g CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanhui Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.com		Add: No.52 HuaYuanhui Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.com	
Measurement Conditions DASY system configuration, as far as not given on page 1			
DASY Version	DASY52	VS2 10.4	
Extrapolation	Advanced Extrapolation		
Phantom	Triple Flat Phantom 5.1C		
Distance Dipole Center - TSL	15 mm	with Spacer	
Zoom Scan Resolution	dx, dy, dz = 5 mm		
Frequency	835 MHz ± 1 MHz		
Head TSL parameters The following parameters and calculations were applied:			
	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	41.5	0.90 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	41.0 ± 0.5	0.91 mho/m ± 0.5%
Head TSL temperature change during last	< 1.0 °C	—	—
SAR result with Head TSL			
SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition		
SAR measured	250 mW input power	2.37 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	6.40 W/kg ± 18.8% (k=2)	
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition		
SAR measured	250 mW input power	1.54 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	6.12 W/kg ± 18.7% (k=2)	
Certificate No: Z22-60104 Page 1 of 6			
Appendix (Additional assessments outside the scope of CNAS L0570) Antenna Parameters with Head TSL		General Antenna Parameters and Design	
Impedance, transformed to feed point Return Loss		48.70-5.22jΩ -25.3dB	
Electrical Delay (one direction)		1.307 ns	
After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured. The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.			
Additional EUT Data			
Manufactured by		SPEAG	
Certificate No: Z22-60104 Page 1 of 6			
DASY5 Validation Report for Head TSL Test Laboratory: CTTL, Beijing, China DUT: Dipole 835 MHz; Type: D835V2; Serial: D835V2 - SN: 40114 Communication System: UTD 0, CW; Frequency: 835 MHz; Duty Cycle: 1:1 Medium parameters used: f = 835 MHz; σ = 0.907 S/m; ε = 40.98; ρ = 1000 kg/m ³ Phantom section: Right Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007) DASY5 Configuration: - Probe: EX3DV4 - SN7307; ConvF(10.13, 10.13, 10.13) @ 835 MHz; Calibrated: 2021-05-26 - Sensor-Surface: 1.4mm (Mechanical Surface Detection) - Electronics: DA64 Sn1556; Calibrated: 2022-01-12 - Phantom: MFP_V5.1C (20deg probe tilt); Type: QD 000 P51 Cx; Serial: 1062 - DASY52 52.10.4(153); SEMCAD X 14.6.14(7501)		Impedance Measurement Plot for Head TSL	
Dipole Calibration/Zoom Scan (7x7x7) (7x7x7) Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 57.88 V/m; Power Drift = 0.04 dB Peak SAR (extrapolated) = 3.56 W/kg SAR(1 g) = 2.37 W/kg; SAR(10 g) = 1.54 W/kg Smallest distance from peaks to all points 3 dB below = 15.8 mm Ratio of SAR at M2 to SAR at M1 = 66.2% Maximum value of SAR (measured) = 3.17 W/kg			
		Certificate No: Z22-60104 Page 6 of 6	



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx>; and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com
 No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
 中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com
 t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

1.5 D900V2 - SN 1d079

TTL Calibration Laboratory		CAICT																					
Add: No.52 Huayuan Road, Haidian District, Beijing, 100191 Tel: +86-10-62166323/3117 E-mail: cti@tts.com.cn		Add: No.52 Huayuan Road, Haidian District, Beijing, 100191 Tel: +86-10-62166323/3117 E-mail: cti@tts.com.cn																					
Client: SGS-CN		Certificate No: Z22-60184																					
CALIBRATION CERTIFICATE																							
Object: D900V2 - SN 1d079																							
Calibration Procedure(s): PF-Z11-003-01																							
Calibration date: June 7, 2022																							
This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.																							
All calibrations have been conducted in the closed laboratory facility, environment temperature (23±1)°C and humidity 70%.																							
Calibration Equipment used (MSTE critical for calibration)																							
<table border="1"><thead><tr><th>Primary Standards</th><th>ID #</th><th>Cal Date (Calibrated by Certificate No.)</th><th>Scheduled Calibration</th></tr></thead><tbody><tr><td>Power Meter: NRP2</td><td>106277</td><td>24-Sep-21 (CTTL No. J21X08326)</td><td>Sep-22</td></tr><tr><td>Power sensor: NRP85</td><td>104291</td><td>24-Sep-21 (CTTL No. J21X08326)</td><td>Sep-22</td></tr><tr><td>Reference Probe: EX32V4</td><td>SN 7484</td><td>26-Jan-22 (SPEAG No. EX3-7484_Jan22)</td><td>Jan-23</td></tr><tr><td>DAE4</td><td>SN 1586</td><td>12-Jan-22 (CTTL-SPEAG No. Z22-60007)</td><td>Jan-23</td></tr></tbody></table>				Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration	Power Meter: NRP2	106277	24-Sep-21 (CTTL No. J21X08326)	Sep-22	Power sensor: NRP85	104291	24-Sep-21 (CTTL No. J21X08326)	Sep-22	Reference Probe: EX32V4	SN 7484	26-Jan-22 (SPEAG No. EX3-7484_Jan22)	Jan-23	DAE4	SN 1586	12-Jan-22 (CTTL-SPEAG No. Z22-60007)	Jan-23
Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration																				
Power Meter: NRP2	106277	24-Sep-21 (CTTL No. J21X08326)	Sep-22																				
Power sensor: NRP85	104291	24-Sep-21 (CTTL No. J21X08326)	Sep-22																				
Reference Probe: EX32V4	SN 7484	26-Jan-22 (SPEAG No. EX3-7484_Jan22)	Jan-23																				
DAE4	SN 1586	12-Jan-22 (CTTL-SPEAG No. Z22-60007)	Jan-23																				
<table border="1"><thead><tr><th>Secondary Standards</th><th>ID #</th><th>Cal Date (Calibrated by Certificate No.)</th><th>Scheduled Calibration</th></tr></thead><tbody><tr><td>Signal Generator: E4438C</td><td>MY48071430</td><td>13-Jan-22 (CTTL No. J22X00406)</td><td>Jan-23</td></tr><tr><td>Network Analyzer: E5071C</td><td>MY48110473</td><td>14-Jan-22 (CTTL No. J22X00406)</td><td>Jan-23</td></tr></tbody></table>				Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration	Signal Generator: E4438C	MY48071430	13-Jan-22 (CTTL No. J22X00406)	Jan-23	Network Analyzer: E5071C	MY48110473	14-Jan-22 (CTTL No. J22X00406)	Jan-23								
Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration																				
Signal Generator: E4438C	MY48071430	13-Jan-22 (CTTL No. J22X00406)	Jan-23																				
Network Analyzer: E5071C	MY48110473	14-Jan-22 (CTTL No. J22X00406)	Jan-23																				
Calibrated by: Name: Zhao Jing, Function: SAR Test Engineer, Signature: [Signature]																							
Reviewed by: Lin Hao, SAR Test Engineer, Signature: [Signature]																							
Approved by: Qi Dianyan, SAR Project Leader, Signature: [Signature]																							
Issued: June 13, 2022																							
This calibration certificate shall not be reproduced except in full without written approval of the laboratory.																							
Certificate No: Z22-60184		Page 1 of 6																					

TTL Calibration Laboratory		CAICT	
Add: No.52 Huayuan Road, Haidian District, Beijing, 100191 Tel: +86-10-62166323/3117 E-mail: cti@tts.com.cn		Add: No.52 Huayuan Road, Haidian District, Beijing, 100191 Tel: +86-10-62166323/3117 E-mail: cti@tts.com.cn	
Glossary: TSL: Issue simulating liquid ConvF: sensitivity in TSL / NORM.y,z N/A: not applicable or not measured			
Calibration is performed according to the following standards: a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices-Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020 b) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz" Additional Documentation: c) DASYS4 System Handbook			
Methods Applied and Interpretation of Parameters: • Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated. • Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis. • Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required. • Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required. • SAR measured: SAR measured at the stated antenna input power. • SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector. • SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.			
The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.			
Certificate No: Z22-60184		Page 1 of 6	

TTL Calibration Laboratory		CAICT																			
Add: No.52 Huayuan Road, Haidian District, Beijing, 100191 Tel: +86-10-62166323/3117 E-mail: cti@tts.com.cn		Add: No.52 Huayuan Road, Haidian District, Beijing, 100191 Tel: +86-10-62166323/3117 E-mail: cti@tts.com.cn																			
Measurement Conditions DASYS system configuration, as far as not given on page 1:																					
<table border="1"><thead><tr><th>DASYS Version</th><th>DASYS2</th><th>52.10.4</th></tr></thead><tbody><tr><td>Extrapolation</td><td>Advanced Extrapolation</td><td></td></tr><tr><td>Phantom</td><td>Tight Fit Phantom 5.1C</td><td></td></tr><tr><td>Distance Dipole Center - TSL</td><td>15 mm</td><td>with Spacer</td></tr><tr><td>Zoom Scan Resolution</td><td>0x, dy, dz = 5 mm</td><td></td></tr><tr><td>Frequency</td><td>900 MHz ± 1 MHz</td><td></td></tr></tbody></table>				DASYS Version	DASYS2	52.10.4	Extrapolation	Advanced Extrapolation		Phantom	Tight Fit Phantom 5.1C		Distance Dipole Center - TSL	15 mm	with Spacer	Zoom Scan Resolution	0x, dy, dz = 5 mm		Frequency	900 MHz ± 1 MHz	
DASYS Version	DASYS2	52.10.4																			
Extrapolation	Advanced Extrapolation																				
Phantom	Tight Fit Phantom 5.1C																				
Distance Dipole Center - TSL	15 mm	with Spacer																			
Zoom Scan Resolution	0x, dy, dz = 5 mm																				
Frequency	900 MHz ± 1 MHz																				
Head TSL parameters The following parameters and calculations were applied:																					
<table border="1"><thead><tr><th></th><th>Temperature</th><th>Permittivity</th><th>Conductivity</th></tr></thead><tbody><tr><td>Nominal Head TSL parameters</td><td>22.0 °C</td><td>41.5</td><td>0.97 mho/m</td></tr><tr><td>Measured Head TSL parameters</td><td>(22.0 ± 0.2) °C</td><td>42.1 ± 6 %</td><td>0.98 mho/m ± 6 %</td></tr><tr><td>Head TSL temperature change during test</td><td><+1.0 °C</td><td></td><td></td></tr></tbody></table>					Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	41.5	0.97 mho/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	42.1 ± 6 %	0.98 mho/m ± 6 %	Head TSL temperature change during test	<+1.0 °C				
	Temperature	Permittivity	Conductivity																		
Nominal Head TSL parameters	22.0 °C	41.5	0.97 mho/m																		
Measured Head TSL parameters	(22.0 ± 0.2) °C	42.1 ± 6 %	0.98 mho/m ± 6 %																		
Head TSL temperature change during test	<+1.0 °C																				
SAR result with Head TSL																					
<table border="1"><thead><tr><th>SAR averaged over 1 cm³ (1 g) of Head TSL</th><th>Condition</th><th></th></tr></thead><tbody><tr><td>SAR measured</td><td>250 mW input power</td><td>2.76 W/kg</td></tr><tr><td>SAR for nominal Head TSL parameters</td><td>normalized to 1W</td><td>11.0 W/kg ± 18.8 % (k=2)</td></tr><tr><td>SAR averaged over 10 cm³ (10 g) of Head TSL</td><th>Condition</th><th></th></tr><tr><td>SAR measured</td><td>250 mW input power</td><td>1.75 W/kg</td></tr><tr><td>SAR for nominal Head TSL parameters</td><td>normalized to 1W</td><td>7.09 W/kg ± 18.7 % (k=2)</td></tr></tbody></table>				SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition		SAR measured	250 mW input power	2.76 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	11.0 W/kg ± 18.8 % (k=2)	SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition		SAR measured	250 mW input power	1.75 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	7.09 W/kg ± 18.7 % (k=2)
SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition																				
SAR measured	250 mW input power	2.76 W/kg																			
SAR for nominal Head TSL parameters	normalized to 1W	11.0 W/kg ± 18.8 % (k=2)																			
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition																				
SAR measured	250 mW input power	1.75 W/kg																			
SAR for nominal Head TSL parameters	normalized to 1W	7.09 W/kg ± 18.7 % (k=2)																			
Certificate No: Z22-60184		Page 1 of 6																			

TTL Calibration Laboratory		CAICT					
Add: No.52 Huayuan Road, Haidian District, Beijing, 100191 Tel: +86-10-62166323/3117 E-mail: cti@tts.com.cn		Add: No.52 Huayuan Road, Haidian District, Beijing, 100191 Tel: +86-10-62166323/3117 E-mail: cti@tts.com.cn					
Appendix (Additional assessments outside the scope of CNAS L0670)							
Antenna Parameters with Head TSL							
<table border="1"><thead><tr><th>Impedance, transformed to feed point</th><th>48 (1.0-9.45G)</th></tr><tr><th>Return Loss</th><th>-23.3 dB</th></tr></thead></table>				Impedance, transformed to feed point	48 (1.0-9.45G)	Return Loss	-23.3 dB
Impedance, transformed to feed point	48 (1.0-9.45G)						
Return Loss	-23.3 dB						
General Antenna Parameters and Design							
<table border="1"><thead><tr><th>Electrical Delay (one direction)</th><th>1.312 ns</th></tr></thead></table>				Electrical Delay (one direction)	1.312 ns		
Electrical Delay (one direction)	1.312 ns						
After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.							
The dipole is made of standard serringid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.							
Additional EUT Data							
<table border="1"><thead><tr><th>Manufactured by</th><th>SPEAG</th></tr></thead></table>				Manufactured by	SPEAG		
Manufactured by	SPEAG						
Certificate No: Z22-60184		Page 1 of 6					




Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx>, and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us by telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weyue Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

(86-512) 57355888 (86-512) 57370818 www.sgs.com.cn
(86-512) 57355888 (86-512) 57370818 sgs.china@sgs.com



1.6 D1800V2 - SN 2d170

TTL Calibration Laboratory
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42306033-2117
E-mail: cti@china.ttl.com

CAICT

Client: **SGS-CN** Certificate No: **Z22-60105**

CALIBRATION CERTIFICATE

Object: **D1800V2 - SN 2d170**

Calibration Procedure(s): **FF-Z11-003-01**
Calibration Procedures for dipole validation kits

Calibration date: **March 31, 2022**

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility, environment temperature (22±3)°C and humidity <70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Power Meter NRP2	106277	24-Sep-21 (CTTL No.J21X08326)	Sep-22
Power sensor NRP8S	104291	24-Sep-21 (CTTL No.J21X08326)	Sep-22
Reference Probe EX3DV4	SN 7307	26-May-21 (SPEAG No.EX3-7307_May21)	May-22
DAE4	SN 1556	12-Jan-22 (CTTL-SPEAG No.Z22-60007)	Jan-23

Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY48071430	13-Jan-22 (CTTL No.J22X00409)	Jan-23
Network Analyzer E5071C	MY48110673	14-Jan-22 (CTTL No.J22X00409)	Jan-23

Calibrated by: **Zhao Jing** SAR Test Engineer
Reviewed by: **Lin Hao** SAR Test Engineer
Approved by: **Qi Diqiyuan** SAR Project Leader

Issued: April 6, 2022
This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z22-60105 Page 1 of 6

TTL Calibration Laboratory
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42306033-2117
E-mail: cti@china.ttl.com

CAICT

Glossary:
TSL: tissue simulating liquid
ConvF: sensitivity in TSL / NORM_{x,y,z}
N/A: not applicable or not measured

Calibration is Performed According to the Following Standards:
a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020
b) KDB 855664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:
c) DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: Z22-60105 Page 2 of 6




Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com



1.7 D1900V2 - SN 5d136

TTL S p e a g CALIBRATION LABORATORY Address: No. 52 Huayuan Road, Haidian District, Beijing, 101910 Tel: +86-10-4226403-2117 E-mail: csl@sgs.com.cn Client: SGS-CN Certificate No: Z22-60185 CALIBRATION CERTIFICATE Object: D1900V2 - SN 5d136 Calibration Procedure(s): FF-Z11-003-01 Calibration Procedures for dipole validation kits Calibration date: June 7, 2022 This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate. All calibrations have been conducted in the closed laboratory facility, environment temperature (23±1)°C and humidity <70%. Calibration Equipment used (M&E critical for calibration) <table border="1"><thead><tr><th>Primary Standards</th><th>ID #</th><th>Cal Date (Calibrated by Certificate No.)</th><th>Scheduled Calibration</th></tr></thead><tbody><tr><td>Power Meter NRP2</td><td>106277</td><td>24-Sep-21 (CTTL No. J21X08326)</td><td>Sep-22</td></tr><tr><td>Power sensor NRP85</td><td>104291</td><td>24-Sep-21 (CTTL No. J21X08326)</td><td>Sep-22</td></tr><tr><td>Reference Probe EX3DV4</td><td>SN 7404</td><td>28-Jan-22 (SPEAG No. EX3-7404_Jan22)</td><td>Jan-23</td></tr><tr><td>DAEA</td><td>SN 1558</td><td>12-Jan-22 (CTTL No. J22-60007)</td><td>Jan-23</td></tr></tbody></table> <table border="1"><thead><tr><th>Secondary Standards</th><th>ID #</th><th>Cal Date (Calibrated by Certificate No.)</th><th>Scheduled Calibration</th></tr></thead><tbody><tr><td>Signal Generator E4438C</td><td>MY48071430</td><td>13-Jan-22 (CTTL No. J22X00409)</td><td>Jan-23</td></tr><tr><td>Network Analyser E5071C</td><td>MY48110073</td><td>14-Jan-22 (CTTL No. J22X00406)</td><td>Jan-23</td></tr></tbody></table> Calibrated by: Name: Zhao Jing, Function: SAR Test Engineer, Signature: [Signature] Reviewed by: Name: Lin Hao, Function: SAR Test Engineer, Signature: [Signature] Approved by: Name: Qi Danyuan, Function: SAR Project Leader, Signature: [Signature] Issued: June 13, 2022 The calibration certificate shall not be reproduced except in full without written approval of the laboratory. Certificate No: Z22-60185 Page 1 of 4	Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration	Power Meter NRP2	106277	24-Sep-21 (CTTL No. J21X08326)	Sep-22	Power sensor NRP85	104291	24-Sep-21 (CTTL No. J21X08326)	Sep-22	Reference Probe EX3DV4	SN 7404	28-Jan-22 (SPEAG No. EX3-7404_Jan22)	Jan-23	DAEA	SN 1558	12-Jan-22 (CTTL No. J22-60007)	Jan-23	Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration	Signal Generator E4438C	MY48071430	13-Jan-22 (CTTL No. J22X00409)	Jan-23	Network Analyser E5071C	MY48110073	14-Jan-22 (CTTL No. J22X00406)	Jan-23	TTL S p e a g CALIBRATION LABORATORY Address: No. 52 Huayuan Road, Haidian District, Beijing, 101910 Tel: +86-10-4226403-2117 E-mail: csl@sgs.com.cn Glossary: TSL: Issue simulating liquid sensitivity in TSL / NORM.y.z ConF: not applicable or not measured N/A: not applicable or not measured Calibration is Performed According to the Following Standards: a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020 b) KDB 855884, "SAR Measurement Requirements for 100 MHz to 6 GHz" Additional Documentation: c) DASY4/S System Handbook Methods Applied and Interpretation of Parameters: • Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated. • Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the fat phantom section, with the arms oriented parallel to the body axis. • Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required. • Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required. • SAR measured: SAR measured at the stated antenna input power. • SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector. • SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result. The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%. Certificate No: Z22-60185 Page 2 of 4																																		
Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration																																																																
Power Meter NRP2	106277	24-Sep-21 (CTTL No. J21X08326)	Sep-22																																																																
Power sensor NRP85	104291	24-Sep-21 (CTTL No. J21X08326)	Sep-22																																																																
Reference Probe EX3DV4	SN 7404	28-Jan-22 (SPEAG No. EX3-7404_Jan22)	Jan-23																																																																
DAEA	SN 1558	12-Jan-22 (CTTL No. J22-60007)	Jan-23																																																																
Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration																																																																
Signal Generator E4438C	MY48071430	13-Jan-22 (CTTL No. J22X00409)	Jan-23																																																																
Network Analyser E5071C	MY48110073	14-Jan-22 (CTTL No. J22X00406)	Jan-23																																																																
TTL S p e a g CALIBRATION LABORATORY Address: No. 52 Huayuan Road, Haidian District, Beijing, 101910 Tel: +86-10-4226403-2117 E-mail: csl@sgs.com.cn Measurement Conditions DASY system configuration, as far as not given on page 1: <table border="1"><thead><tr><th>DASY Version</th><th>DASY152</th><th>52.10.4</th></tr></thead><tbody><tr><td>Extrapolation</td><td>Advanced Extrapolation</td><td></td></tr><tr><td>Phantom</td><td>Triple Flat Phantom 5.1C</td><td></td></tr><tr><td>Distance Dipole Center - TSL</td><td>10 mm</td><td>with Spacer</td></tr><tr><td>Zoom Scan Resolution</td><td>dx, dy, dz = 5 mm</td><td></td></tr><tr><td>Frequency</td><td>1900 MHz ± 1 MHz</td><td></td></tr></tbody></table> Head TSL parameters The following parameters and calculations were applied: <table border="1"><thead><tr><th>Parameter</th><th>Temperature</th><th>Permittivity</th><th>Conductivity</th></tr></thead><tbody><tr><td>Nominal Head TSL parameters</td><td>22.0 °C</td><td>46.0</td><td>1.40 nS/m</td></tr><tr><td>Measured Head TSL parameters</td><td>(22.0 ± 0.5) °C</td><td>39.9 ± 0.5 %</td><td>1.39 nS/m ± 0.5 %</td></tr><tr><td>Head TSL temperature change during test</td><td>+1.0 °C</td><td></td><td></td></tr></tbody></table> SAR result with Head TSL <table border="1"><thead><tr><th>SAR averaged over 1 cm² (1 g) of Head TSL</th><th>Condition</th><th>Value</th></tr></thead><tbody><tr><td>SAR measured</td><td>250 mW input power</td><td>9.65 W/kg</td></tr><tr><td>SAR for nominal Head TSL parameters</td><td>normalized to 1W</td><td>40.0 W/kg ± 16.8 % (k=2)</td></tr></tbody></table> SAR averaged over 10 cm² (10 g) of Head TSL <table border="1"><thead><tr><th>SAR measured</th><th>Condition</th><th>Value</th></tr></thead><tbody><tr><td>SAR measured</td><td>250 mW input power</td><td>5.18 W/kg</td></tr><tr><td>SAR for nominal Head TSL parameters</td><td>normalized to 1W</td><td>20.8 W/kg ± 16.7 % (k=2)</td></tr></tbody></table> Certificate No: Z22-60185 Page 3 of 4	DASY Version	DASY152	52.10.4	Extrapolation	Advanced Extrapolation		Phantom	Triple Flat Phantom 5.1C		Distance Dipole Center - TSL	10 mm	with Spacer	Zoom Scan Resolution	dx, dy, dz = 5 mm		Frequency	1900 MHz ± 1 MHz		Parameter	Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	46.0	1.40 nS/m	Measured Head TSL parameters	(22.0 ± 0.5) °C	39.9 ± 0.5 %	1.39 nS/m ± 0.5 %	Head TSL temperature change during test	+1.0 °C			SAR averaged over 1 cm ² (1 g) of Head TSL	Condition	Value	SAR measured	250 mW input power	9.65 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	40.0 W/kg ± 16.8 % (k=2)	SAR measured	Condition	Value	SAR measured	250 mW input power	5.18 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	20.8 W/kg ± 16.7 % (k=2)	TTL S p e a g CALIBRATION LABORATORY Address: No. 52 Huayuan Road, Haidian District, Beijing, 101910 Tel: +86-10-4226403-2117 E-mail: csl@sgs.com.cn Appendix (Additional assessments outside the scope of CNAS L0570) Antenna Parameters with Head TSL <table border="1"><thead><tr><th>Parameter</th><th>Value</th></tr></thead><tbody><tr><td>Impedance, transformed to feed point</td><td>51.0Ω ± 1.5Ω (k=2)</td></tr><tr><td>Return Loss</td><td>22.4dB</td></tr></tbody></table> General Antenna Parameters and Design <table border="1"><thead><tr><th>Parameter</th><th>Value</th></tr></thead><tbody><tr><td>Electrical Delay (one direction)</td><td>1.108 ns</td></tr></tbody></table> After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured. The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC signals. On some of the dipoles, small caps are added to the dipole arms in order to improve matching when loaded according to the position as specified in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged. Additional EUT Data <table border="1"><thead><tr><th>Manufacturer by</th><th>Value</th></tr></thead><tbody><tr><td></td><td>SPEAG</td></tr></tbody></table> Certificate No: Z22-60185 Page 4 of 4	Parameter	Value	Impedance, transformed to feed point	51.0Ω ± 1.5Ω (k=2)	Return Loss	22.4dB	Parameter	Value	Electrical Delay (one direction)	1.108 ns	Manufacturer by	Value		SPEAG
DASY Version	DASY152	52.10.4																																																																	
Extrapolation	Advanced Extrapolation																																																																		
Phantom	Triple Flat Phantom 5.1C																																																																		
Distance Dipole Center - TSL	10 mm	with Spacer																																																																	
Zoom Scan Resolution	dx, dy, dz = 5 mm																																																																		
Frequency	1900 MHz ± 1 MHz																																																																		
Parameter	Temperature	Permittivity	Conductivity																																																																
Nominal Head TSL parameters	22.0 °C	46.0	1.40 nS/m																																																																
Measured Head TSL parameters	(22.0 ± 0.5) °C	39.9 ± 0.5 %	1.39 nS/m ± 0.5 %																																																																
Head TSL temperature change during test	+1.0 °C																																																																		
SAR averaged over 1 cm ² (1 g) of Head TSL	Condition	Value																																																																	
SAR measured	250 mW input power	9.65 W/kg																																																																	
SAR for nominal Head TSL parameters	normalized to 1W	40.0 W/kg ± 16.8 % (k=2)																																																																	
SAR measured	Condition	Value																																																																	
SAR measured	250 mW input power	5.18 W/kg																																																																	
SAR for nominal Head TSL parameters	normalized to 1W	20.8 W/kg ± 16.7 % (k=2)																																																																	
Parameter	Value																																																																		
Impedance, transformed to feed point	51.0Ω ± 1.5Ω (k=2)																																																																		
Return Loss	22.4dB																																																																		
Parameter	Value																																																																		
Electrical Delay (one direction)	1.108 ns																																																																		
Manufacturer by	Value																																																																		
	SPEAG																																																																		




Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No. 10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

(86-512) 57355888 (86-512) 57370818 www.sgs.com.cn
(86-512) 57355888 (86-512) 57370818 sgs.china@sgs.com



1.8 D2000V2 - SN 1041

TTLSPEAG Calibration Laboratory
 Add: No.52 Huaiyuan Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-62586317
 E-mail: cti@ttlspeag.com

CAICT

Client: SGS-CN Certificate No: Z22-60186

CALIBRATION CERTIFICATE

Object: D2000V2 - SN: 1041

Calibration Procedure(s): FF-Z11-003-01
 Calibration Procedure for dipole validation kits

Calibration date: June 5, 2022

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (S). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (23±3)°C and humidity<70%.

Calibration Equipment used (MATE critical for calibration)

Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Power Meter NRP2	106277	24-Sep-21 (CTTL No. J21X08328)	Sep-22
Power sensor NRP5	104291	24-Sep-21 (CTTL No. J21X08328)	Sep-22
Reference Probe EX3DV4	SN 7464	26-Jan-22 (SPEAG No. EX3-7464, Jan22)	Jan-23
DAE4	SN 1556	12-Jan-22 (CTTL-SPEAG No. Z22-60007)	Jan-23

Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL No. Z22X00408)	Jan-23
Network Analyzer E5071C	MY46110673	14-Jan-22 (CTTL No. Z22X00408)	Jan-23

Calibrated by: Zhao Jing SAR Test Engineer

Reviewed by: Lin Hao SAR Test Engineer

Approved by: Qi Dianyan SAR Project Leader

Issued: June 13, 2022

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z22-60186 Page 1 of 6

TTLSPEAG Calibration Laboratory
 Add: No.52 Huaiyuan Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-62586317
 E-mail: cti@ttlspeag.com

CAICT

Glossary:
 TSL: Issue simulating liquid
 ConvF: sensitivity in TSL / NORMx.y.z
 N/A: not applicable or not measured

Calibration is Performed According to the Following Standards:
 a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-Head and Body-Mounted Wireless Communication Devices: Part 1528 Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020
 b) KDB 865564, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:
 c) DASY4/S System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameter with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: Z22-60186 Page 2 of 6





Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx>; and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weyue Road, Innovation Park, Kunshan, Jiangsu, China 215300
 中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

(86-512)57355888 (86-512)57370818 www.sgsgroup.com.cn
 (86-512)57355888 (86-512)57370818 sgs.china@sgs.com

In Collaboration with TTL s p e a g CALIBRATION LABORATORY CAICT Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-6204832-2117 E-mail: emc@chinaet.com.cn http://www.caict.ac.cn Measurement Conditions DASY system configuration, as for as not given on page 1. <table border="1"><tr><td>DASY Version</td><td>DASYV2</td><td>52.10.4</td></tr><tr><td>Extrapolation</td><td>Advanced Extrapolation</td><td></td></tr><tr><td>Phantom</td><td>Triple Flat Phantom S.1C</td><td></td></tr><tr><td>Distance Dipole Center - TSL</td><td>10 mm</td><td>with Spacer</td></tr><tr><td>Zoom Scan Resolution</td><td>dx, dy, dz = 5 mm</td><td></td></tr><tr><td>Frequency</td><td>2000 MHz ± 1 MHz</td><td></td></tr></table> Head TSL parameters The following parameters and calculations were applied: <table border="1"><tr><th></th><th>Temperature</th><th>Permittivity</th><th>Conductivity</th></tr><tr><td>Nominal Head TSL parameters</td><td>22.9 °C</td><td>40.0</td><td>1.40 mho/cm</td></tr><tr><td>Measured Head TSL parameters</td><td>(22.0 ± 0.2) °C</td><td>40.2 ± 0.6 %</td><td>1.36 mho/cm ± 0.6 %</td></tr><tr><td>Head TSL temperature change during test</td><td><1.0 °C</td><td></td><td></td></tr></table> SAR result with Head TSL <table border="1"><tr><th></th><th>Condition</th><th></th></tr><tr><td>SAR averaged over 1 cm³ (1 g) of Head TSL</td><td></td><td></td></tr><tr><td>SAR measured</td><td>250 mW input power</td><td>10.4 W/kg</td></tr><tr><td>SAR for nominal Head TSL parameters</td><td>normalized to 1W</td><td>41.8 W/kg ± 15.8 % (n=2)</td></tr><tr><td>SAR averaged over 10 cm³ (10 g) of Head TSL</td><td></td><td></td></tr><tr><td>SAR measured</td><td>250 mW input power</td><td>5.30 W/kg</td></tr><tr><td>SAR for nominal Head TSL parameters</td><td>normalized to 1W</td><td>21.3 W/kg ± 18.7 % (n=2)</td></tr></table> Certificate No: Z22-60186 Page 3 of 6	DASY Version	DASYV2	52.10.4	Extrapolation	Advanced Extrapolation		Phantom	Triple Flat Phantom S.1C		Distance Dipole Center - TSL	10 mm	with Spacer	Zoom Scan Resolution	dx, dy, dz = 5 mm		Frequency	2000 MHz ± 1 MHz			Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.9 °C	40.0	1.40 mho/cm	Measured Head TSL parameters	(22.0 ± 0.2) °C	40.2 ± 0.6 %	1.36 mho/cm ± 0.6 %	Head TSL temperature change during test	<1.0 °C				Condition		SAR averaged over 1 cm ³ (1 g) of Head TSL			SAR measured	250 mW input power	10.4 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	41.8 W/kg ± 15.8 % (n=2)	SAR averaged over 10 cm ³ (10 g) of Head TSL			SAR measured	250 mW input power	5.30 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	21.3 W/kg ± 18.7 % (n=2)	In Collaboration with TTL s p e a g CALIBRATION LABORATORY CAICT Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-6204832-2117 E-mail: emc@chinaet.com.cn http://www.caict.ac.cn Appendix (Additional assessments outside the scope of CNAS L0570) Antenna Parameters with Head TSL <table border="1"><tr><td>Impedance, transformed to feed point</td><td>48.4Ω ± 0.74Ω</td></tr><tr><td>Return Loss</td><td>-34.8dB</td></tr></table> General Antenna Parameters and Design <table border="1"><tr><td>Electrical Delay (one direction)</td><td>1.08 ns</td></tr></table> After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured. The dipole is made of standard semi-rigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small and large are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged. Additional EUT Data <table border="1"><tr><td>Manufactured by</td><td>SPEAG</td></tr></table> Certificate No: Z22-60186 Page 8 of 8	Impedance, transformed to feed point	48.4Ω ± 0.74Ω	Return Loss	-34.8dB	Electrical Delay (one direction)	1.08 ns	Manufactured by	SPEAG
DASY Version	DASYV2	52.10.4																																																														
Extrapolation	Advanced Extrapolation																																																															
Phantom	Triple Flat Phantom S.1C																																																															
Distance Dipole Center - TSL	10 mm	with Spacer																																																														
Zoom Scan Resolution	dx, dy, dz = 5 mm																																																															
Frequency	2000 MHz ± 1 MHz																																																															
	Temperature	Permittivity	Conductivity																																																													
Nominal Head TSL parameters	22.9 °C	40.0	1.40 mho/cm																																																													
Measured Head TSL parameters	(22.0 ± 0.2) °C	40.2 ± 0.6 %	1.36 mho/cm ± 0.6 %																																																													
Head TSL temperature change during test	<1.0 °C																																																															
	Condition																																																															
SAR averaged over 1 cm ³ (1 g) of Head TSL																																																																
SAR measured	250 mW input power	10.4 W/kg																																																														
SAR for nominal Head TSL parameters	normalized to 1W	41.8 W/kg ± 15.8 % (n=2)																																																														
SAR averaged over 10 cm ³ (10 g) of Head TSL																																																																
SAR measured	250 mW input power	5.30 W/kg																																																														
SAR for nominal Head TSL parameters	normalized to 1W	21.3 W/kg ± 18.7 % (n=2)																																																														
Impedance, transformed to feed point	48.4Ω ± 0.74Ω																																																															
Return Loss	-34.8dB																																																															
Electrical Delay (one direction)	1.08 ns																																																															
Manufactured by	SPEAG																																																															
In Collaboration with TTL s p e a g CALIBRATION LABORATORY CAICT Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-6204832-2117 E-mail: emc@chinaet.com.cn http://www.caict.ac.cn DASY Validation Report for Head TSL Test Laboratory: CTTL, Beijing, China DUT: Dipole 2000 MHz; Type: D2000V2; Serial: D2000V2 - SN: 1041 Communication System: UTD 0; CW; Frequency: 2000 MHz; Duty Cycle: 1:1 Medium parameters used: f = 2000 MHz; σ = 1.392 S/m; ε = 40.21; ρ = 1000 kg/m³ Phantom section: Right Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007) DASY5 Configuration: - Probe: EX3DV4 - SN7464; ConvF(R2, R2, R2) @ 2000 MHz; Calibrated: 2022-01-26 - Sensor Surface: 1.4mm (Mechanical Surface Detection) - Electronics: DA64 Sn1556; Calibrated: 2022-01-12 - Phantom: MPP_V5.1C (20deg probe tilt); Type: QD 000 P51 Cx; Serial: 1062 - DASY5: S2.10.4(1555); SEMCAD X 14.6.14(7501) Dipole Calibration/Zoom Scan (7x7x7) (7x7x7) Cube (0; Measurement grid: dx=5mm, dy=5mm, dz=5mm) Reference Value = 103.4 V/m; Power Drift = 0.03 dB Peak SAR (extrapolated) = 19.6 W/kg SAR(1 g) = 10.4 W/kg; SAR(10 g) = 5.3 W/kg Smallest distance from peaks to all points 3 dB below = 9.1 mm Ratio of SAR at M2 to SAR at M1 = 51.0 % Maximum value of SAR (measured) = 16.3 W/kg  Certificate No: Z22-60186 Page 5 of 6	In Collaboration with TTL s p e a g CALIBRATION LABORATORY CAICT Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-6204832-2117 E-mail: emc@chinaet.com.cn http://www.caict.ac.cn Impedance Measurement Plot for Head TSL  Certificate No: Z22-60186 Page 6 of 6																																																															



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com
No. 10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

1.9 D2300V2 - SN 1096


TTL Speaq CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-6230433-2512 Fax: +86-10-6230433-2504 E-mail: cti@chinaul.com http://www.chinaul.cn		Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-6230433-2579 Fax: +86-10-6230433-2504 E-mail: cti@chinaul.com http://www.chinaul.cn	
Client: SGS-CN		Certificate No: Z22-60106	
CALIBRATION CERTIFICATE			
Object	D2300V2 - SN 1096		
Calibration Procedure(s)	FF-Z11-003-01 Calibration Procedures for dipole validation kits		
Calibration date:	March 31, 2022		
This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.			
All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3)°C and humidity <70%.			
Calibration Equipment used (M&TE critical for calibration)			
Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Power Meter NRP2	108277	24-Sep-21 (CTTL No.J21X08328)	Sep-22
Power sensor NRP8	104291	24-Sep-21 (CTTL No.J21X08328)	Sep-22
Reference Probe EX30V4	SN 7307	26-May-21 (SPEAG No.EK3-7307_May21)	May-22
DAE4	SN 1556	12-Jan-22 (CTTL-SPEAG No.Z22-60007)	Jan-23
Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL No.J22X00406)	Jan-23
Network Analyzer E5071C	MY48110673	14-Jan-22 (CTTL No.J22X00406)	Jan-23
Calibrated by:	Name	Function	Signature
	Zhao Jing	SAR Test Engineer	
Reviewed by:	Lin Hao	SAR Test Engineer	
Approved by:	Qi Diqian	SAR Project Leader	
Issued: April 6, 2022			
This calibration certificate shall not be reproduced except in full without written approval of the laboratory			
Certificate No: Z22-60106		Page 1 of 6	
TTL Speaq CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-6230433-2579 Fax: +86-10-6230433-2504 E-mail: cti@chinaul.com http://www.chinaul.cn		Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-6230433-2579 Fax: +86-10-6230433-2504 E-mail: cti@chinaul.com http://www.chinaul.cn	
Measurement Conditions DASY system configuration, as far as not given on page 1:			
DASY Version	DASY2	52.10.4	
Extrapolation	Advanced Extrapolation		
Phantom	Triple Flat Phantom 5.1C		
Distance Dipole Center - TSL	10 mm	with Spacer	
Zoom Scan Resolution	dx, dy, dz = 5 mm		
Frequency	2300 MHz ± 1 MHz		
Head TSL parameters The following parameters and calculations were applied:			
	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	39.5	1.67 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	39.8 ± 9 %	1.70 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	—	
SAR result with Head TSL			
SAR averaged over 1 cm ² (1 g) of Head TSL	Condition		
SAR measured	250 mW input power	12.4 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	49.2 W/kg ± 18.8 % (k=2)	
SAR averaged over 10 cm ² (10 g) of Head TSL	Condition		
SAR measured	250 mW input power	5.88 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	23.4 W/kg ± 18.7 % (k=2)	
Certificate No: Z22-60106		Page 3 of 6	
TTL Speaq CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-6230433-2579 Fax: +86-10-6230433-2504 E-mail: cti@chinaul.com http://www.chinaul.cn		Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-6230433-2579 Fax: +86-10-6230433-2504 E-mail: cti@chinaul.com http://www.chinaul.cn	
Appendix (Additional assessments outside the scope of CNAS L0570)			
Antenna Parameters with Head TSL			
Impedance, transformed to feed point	49.20 - 4.50j		
Return Loss	-26.6dB		
General Antenna Parameters and Design			
Electrical Delay (one direction)	1.083 ns		
After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.			
The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when backed according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.			
Additional EUT Data			
Manufactured by	SPEAG		
Certificate No: Z22-60106		Page 4 of 6	



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx>; and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com
No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

(86-512)57355888 (86-512)57370818 www.sgs.com.cn
(86-512)57355888 (86-512)57370818 sgs.china@sgs.com



1.10 D2450V2 - SN 817

TTL Speaq
CALIBRATION LABORATORY

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62504633-3079 Fax: +86-10-62504633-2504
E-mail: csl@china.ttl.com http://www.china.ttl.com

CAICT

Client: **SGS-CN** Certificate No: **Z22-60107**

CALIBRATION CERTIFICATE

Object: **D2450V2 - SN 817**

Calibration Procedure(s): **FF-Z11-003-3**
Calibration Procedures for dipole validation kits

Calibration date: **April 1, 2022**

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements [B]. The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3)°C and humidity <70%.

Calibration Equipment used (M&E critical for calibration)

Primary Standards	ID #	Cal Date (Calibrated by: Certificate No.)	Scheduled Calibration
Power Meter: NRP2	106277	24-Sep-21 (CTTL No.J21X08326)	Sep-22
Power sensor: NRP8S	104291	24-Sep-21 (CTTL No.J21X08326)	Sep-22
Reference Probe EX3DV4	SN 7307	25-May-21 (SPEAG No EX3-7307_May21)	May-22
DAE4	SN 1555	12-Jan-22 (CTTL-SPEAG No Z22-60007)	Jan-23

Secondary Standards	ID #	Cal Date (Calibrated by: Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL No.J22X00406)	Jan-23
Network Analyzer E5071C	MY46110673	14-Jan-22 (CTTL No.J22X00406)	Jan-23

Calibrated by: **Zhao Jing** SAR Test Engineer
Reviewed by: **Liu Hao** SAR Test Engineer
Approved by: **Qi Dianyan** SAR Project Leader

Issued: April 6, 2022

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z22-60107 Page 1 of 5

TTL Speaq
CALIBRATION LABORATORY

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62504633-3079 Fax: +86-10-62504633-2504
E-mail: csl@china.ttl.com http://www.china.ttl.com

CAICT

Glossary:
TSL: tissue simulating liquid
ConvF: sensitivity in TSL / NORMix.y.z
N/A: not applicable or not measured

Calibration is Performed According to the Following Standards:

a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020
b) KDB 865664, "SAR Measurement Requirements for 100 MHz to 5 GHz"

Additional Documentation:
c) DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor: $k=2$, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

Certificate No: Z22-60107 Page 2 of 5



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.


Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com


No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

TTL Speaq CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaul.com.cn http://www.chinaul.com.cn			
Measurement Conditions DASY system configuration, as far as not given on page 1:			
DASY Version	DASY52	52.10.4	
Extrapolation	Advanced Extrapolation		
Phantom	Triple Flat Phantom 5.1C		
Distance Dipole Center - TSL	10 mm	with Spacer	
Zoom Scan Resolution	dx, dy, dz = 5 mm		
Frequency	2450 MHz ± 1 MHz		
Head TSL parameters The following parameters and calculations were applied:			
Normal Head TSL parameters	Temperature	Permittivity	Conductivity
Measured Head TSL parameters	22.0 °C	39.2	1.80 mho/m
Head TSL temperature change during test	(22.0 ± 0.2) °C	39.5 ± 6 %	1.79 mho/m ± 8 %
SAR result with Head TSL			
SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition		
SAR measured	250 mW input power	13.2 W/kg	
SAR for normal Head TSL parameters	normalized to 1W	53.0 W/kg ± 18.8 % (k=2)	
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition		
SAR measured	250 mW input power	6.15 W/kg	
SAR for normal Head TSL parameters	normalized to 1W	24.7 W/kg ± 18.7 % (k=2)	
Certificate No: Z22-40107 Page 3 of 6			

TTL Speaq CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaul.com.cn http://www.chinaul.com.cn			
Appendix (Additional assessments outside the scope of CNAS L0570)			
Antenna Parameters with Head TSL			
Impedance, transformed to feed point	52.10 ± 3.20Ω		
Return Loss	~28.5dB		
General Antenna Parameters and Design			
Electrical Delay (one direction)	1.006 ns		
After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.			
The dipole is made of standard serringid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.			
Additional EUT Data			
Manufactured by	SPEAG		
Certificate No: Z22-40107 Page 4 of 6			

TTL Speaq CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaul.com.cn http://www.chinaul.com.cn			
DASY5 Validation Report for Head TSL Test Laboratory: CTTL, Beijing, China DUT: Dipole 2450 MHz; Type: D2450V2; Serial: D2450V2 - SN: 817 Communication System: UID 0, CW; Frequency: 2450 MHz; Duty Cycle: 1:1 Medium parameters used: f = 2450 MHz; σ = 1.79 S/m; ε = 39.52; ρ = 1000 kg/m ³ Phantom section: Right Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007) DASY5 Configuration:			
• Probe: EX3DV4 - SN7307; ConvF(7.75, 7.75) @ 2450 MHz; Calibrated: 2021-05-26			
• Sensor-Surface: 1.4mm (Mechanical Surface Detection)			
• Electronics: DA14 Snt 556; Calibrated: 2022-01-12			
• Phantom: MFP, V5.1C (20deg probe tilt); Type: QD 000 P51 Cx; Serial: 1062			
• DASY52 52.10.4(1535); SEMCAD X 14.6.14(7501)			
Dipole Calibration/Zoom Scan (7x7x7) (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 104.6 V/m; Power Drift = -0.03 dB Peak SAR (extrapolated) = 27.0 W/kg SAR(1 g) = 13.2 W/kg; SAR(10 g) = 6.15 W/kg Smallest distance from peaks to all points 3 dB below = 8.9 mm Ratio of SAR at M2 to SAR at M1 = -49.2% Maximum value of SAR (measured) = 22.1 W/kg			
			
Certificate No: Z22-40107 Page 5 of 6			

TTL Speaq CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaul.com.cn http://www.chinaul.com.cn			
Impedance Measurement Plot for Head TSL			
			
Certificate No: Z22-40107 Page 6 of 6			



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com
No.10, Weyue Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

1.11 D2600V2 - SN 1158

TTL S p e a g CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191 Tel: +86-10-42304633-2512 Fax: +86-10-42304633-2504 E-mail: cti@china-test.com.cn http://www.china-test.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191 Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china-test.com.cn http://www.china-test.cn	
Client: SGS-CN	Certificate No: Z22-60108		
CALIBRATION CERTIFICATE			
Object:	D2600V2 - SN: 1158		
Calibration Procedure(s):	FF-Z11-003-01 Calibration Procedures for dipole validation kits		
Calibration date:	March 31, 2022		
This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.			
All calibrations have been conducted in the closed laboratory facility: environment temperature (23±3)°C and humidity<70%.			
Calibration Equipment used (MSTE critical for calibration)			
Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Power Meter NRP2	108277	24-Sep-21 (CTTL No.J21X08326)	Sep-22
Power sensor NRP8S	104291	24-Sep-21 (CTTL No.J21X08326)	Sep-22
Reference Probe EX3/DVA	SN 7307	26-May-21 (SPEAG No EX3-7307_May21)	May-22
DAE4	SN 1556	12-Jan-22 (CTTL-SPEAG No Z22-60007)	Jan-23
Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY48071430	13-Jan-22 (CTTL No.Z22X00406)	Jan-23
Network Analyzer E5071C	MY48110673	14-Jan-22 (CTTL No.Z22X00406)	Jan-23
Calibrated by:	Name: Zhao Jing	Function: SAR Test Engineer	Signature: [Signature]
Reviewed by:	Name: Lin Hao	Function: SAR Test Engineer	Signature: [Signature]
Approved by:	Name: Qi Dianyuan	Function: SAR Project Leader	Signature: [Signature]
Issued: April 6, 2022			
This calibration certificate shall not be reproduced except in full without written approval of the laboratory.			
Certificate No: Z22-60108	Page 1 of 6		

TTL S p e a g CALIBRATION LABORATORY		CAICT					
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china-test.com.cn http://www.china-test.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china-test.com.cn http://www.china-test.cn					
Measurement Conditions DASY system configuration, as far as not given on page 1.		Appendix (Additional assessments outside the scope of CNAS L0570)					
DASY Version	DASY52	52.10.4					
Extrapolation	Advanced Extrapolation						
Phantom	Triple Flat Phantom S.1C						
Distance Dipole Center - TSL	10 mm	with Spacer					
Zoom Scan Resolution	dx, dy, dz = 5 mm						
Frequency	2600 MHz ± 1 MHz						
Head TSL parameters The following parameters and calculations were applied:							
	Temperature	Permittivity	Conductivity				
Nominal Head TSL parameters	22.0 °C	38.0	1.96 mho/m				
Measured Head TSL parameters	(22.0 ± 0.2) °C	38.7 ± 8 %	1.96 mho/m ± 8 %				
Head TSL temperature change during test	< ± 0.2 °C	—	—				
SAR result with Head TSL							
SAR averaged over 1 cm² (1 g) of Head TSL	Condition						
SAR measured	250 mW input power	13.7 W/kg					
SAR for nominal Head TSL parameters	normalized to 1W	54.8 W/kg ± 18.8 % (k=2)					
SAR averaged over 10 cm² (10 g) of Head TSL	Condition						
SAR measured	250 mW input power	8.12 W/kg					
SAR for nominal Head TSL parameters	normalized to 1W	24.5 W/kg ± 18.7 % (k=2)					
Antenna Parameters with Head TSL <table border="1"> <tr> <td>Impedance, transformed to feed point</td> <td>49.90 - j6.49Ω</td> </tr> <tr> <td>Return Loss</td> <td>-23.9dB</td> </tr> </table>				Impedance, transformed to feed point	49.90 - j6.49Ω	Return Loss	-23.9dB
Impedance, transformed to feed point	49.90 - j6.49Ω						
Return Loss	-23.9dB						
General Antenna Parameters and Design <table border="1"> <tr> <td>Electrical Delay (one direction)</td> <td>1.053 ns</td> </tr> </table>				Electrical Delay (one direction)	1.053 ns		
Electrical Delay (one direction)	1.053 ns						
After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured. The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.							
Additional EUT Data <table border="1"> <tr> <td>Manufactured by</td> <td>SPEAG</td> </tr> </table>				Manufactured by	SPEAG		
Manufactured by	SPEAG						
Certificate No: Z22-60108	Page 3 of 6						




Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx>; and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com



1.12 D5GHZV2 - SN 1095

TTL Speaq
CALIBRATION LABORATORY

Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504
E-mail: cti@ttspeaq.com http://www.ttspeaq.com

CAICT

Client: SGS-CN Certificate No: Z22-60187

CALIBRATION CERTIFICATE

Object: D5GHZV2 - SN 1095

Calibration Procedure(s): FF-Z11-003-01
Calibration Procedures for dipole validation kit

Calibration date: June 1, 2022

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (23±3)°C and humidity <70%.

Calibration Equipment used (MSTE critical for calibration)

Primary Standards	ID #	Cal Date (Calibrated by: Certificate No.)	Scheduled Calibration
Power Meter NRP2	106277	24-Sep-21 (CTTL No. J21008328)	Sep-22
Power sensor NRP8	104291	24-Sep-21 (CTTL No. J21008328)	Sep-22
Reference Probe EX3DV4	SN 7464	26-Jan-22 (SPEAG No. EX3-7464, Jan22)	Jan-23
DA64	SN 1556	12-Jan-22 (CTTL-SPEAG No. Z22-60007)	Jan-23

Secondary Standards	ID #	Cal Date (Calibrated by: Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY48071430	13-Jan-22 (CTTL No. J22000409)	Jan-23
Network Analyzer E5071C	MY46110673	14-Jan-22 (CTTL No. J22000406)	Jan-23

Calibrated by: Zhao Jing SAR Test Engineer

Reviewed by: Lin Hao SAR Test Engineer

Approved by: Qi Dianyan SAR Project Leader

Issued: June 9, 2022

This calibration Certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z22-60187 Page 1 of 10

TTL Speaq
CALIBRATION LABORATORY

Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504
E-mail: cti@ttspeaq.com http://www.ttspeaq.com

CAICT

Glossary:

TSL: Issue simulating liquid

CompF: sensitivity in TSL + NORMk,y,z

N/A: not applicable or not measured

Calibration is Performed According to the Following Standards:

a) IEC/IEEE 62209-1528, "Measurement Procedure for the Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices-Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020

b) KDB 665664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:

c) DASY5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance station is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

Certificate No: Z22-60187 Page 7 of 10



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx>; and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

TTL Calibration Laboratory		CAICT																																																					
In Collaboration with Address: No.52 HuaYuan Road, Haidian District, Beijing, 100191, China Tel: +86-10-62420217 E-mail: info@ttsl.com.cn http://www.ttsl.com.cn		In Collaboration with Address: No.52 HuaYuan Road, Haidian District, Beijing, 100191, China Tel: +86-10-62420217 E-mail: info@ttsl.com.cn http://www.ttsl.com.cn																																																					
Measurement Conditions EMC system configuration, as for as not given on page 1.																																																							
DASY Version: DASY2 Extrapolation: Advanced Extrapolation Phantom: Triple Flat Phantom S.1C Distance Dipole Center - TSL: 10 mm with Spacer Zoom Scan Resolution: $dx, dy = 4 \text{ mm}$, $dz = 1.4 \text{ mm}$ (Graded Ratio = 1.4 (Z direction)) Frequency: 5000 MHz $\pm 1 \text{ MHz}$ 6300 MHz $\pm 1 \text{ MHz}$ 3500 MHz $\pm 1 \text{ MHz}$ 5600 MHz $\pm 1 \text{ MHz}$ 5800 MHz $\pm 1 \text{ MHz}$		Head TSL parameters at 5200MHz The following parameters and calculations were applied: <table border="1"> <thead> <tr> <th></th> <th>Temperature</th> <th>Permittivity</th> <th>Conductivity</th> </tr> </thead> <tbody> <tr> <td>Nominal Head TSL parameters</td> <td>22.0 °C</td> <td>35.9</td> <td>4.76 mho/m</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>(22.0 $\pm$ 0.2) °C</td> <td>35.4 $\pm$ 5 %</td> <td>4.73 mho/m $\pm$ 5 %</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td><1.0 °C</td> <td>—</td> <td>—</td> </tr> </tbody> </table>			Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	35.9	4.76 mho/m	Measured Head TSL parameters	(22.0 $\pm$ 0.2) °C	35.4 $\pm$ 5 %	4.73 mho/m $\pm$ 5 %	Head TSL temperature change during test	<1.0 °C	—	—																																				
	Temperature	Permittivity	Conductivity																																																				
Nominal Head TSL parameters	22.0 °C	35.9	4.76 mho/m																																																				
Measured Head TSL parameters	(22.0 $\pm$ 0.2) °C	35.4 $\pm$ 5 %	4.73 mho/m $\pm$ 5 %																																																				
Head TSL temperature change during test	<1.0 °C	—	—																																																				
SAR result with Head TSL at 5200MHz SAR averaged over 1 cm^2 (1 g) of Head TSL <table border="1"> <thead> <tr> <th></th> <th>Condition</th> <th></th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>7.75 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>77.8 W/kg $\pm$ 24.4 % (k=2)</td> </tr> </tbody> </table> SAR averaged over 10 cm^2 (10 g) of Head TSL <table border="1"> <thead> <tr> <th></th> <th>Condition</th> <th></th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>2.22 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>22.1 W/kg $\pm$ 24.2 % (k=2)</td> </tr> </tbody> </table>			Condition		SAR measured	250 mW input power	7.75 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	77.8 W/kg $\pm$ 24.4 % (k=2)		Condition		SAR measured	250 mW input power	2.22 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	22.1 W/kg $\pm$ 24.2 % (k=2)	Head TSL parameters at 5300MHz The following parameters and calculations were applied: <table border="1"> <thead> <tr> <th></th> <th>Temperature</th> <th>Permittivity</th> <th>Conductivity</th> </tr> </thead> <tbody> <tr> <td>Nominal Head TSL parameters</td> <td>22.0 °C</td> <td>35.9</td> <td>4.76 mho/m</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>(22.0 $\pm$ 0.2) °C</td> <td>35.2 $\pm$ 5 %</td> <td>4.73 mho/m $\pm$ 5 %</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td><1.0 °C</td> <td>—</td> <td>—</td> </tr> </tbody> </table> SAR result with Head TSL at 5300MHz SAR averaged over 1 cm^2 (1 g) of Head TSL <table border="1"> <thead> <tr> <th></th> <th>Condition</th> <th></th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>100 mW input power</td> <td>7.84 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>78.1 W/kg $\pm$ 24.4 % (k=2)</td> </tr> </tbody> </table> SAR averaged over 10 cm^2 (10 g) of Head TSL <table border="1"> <thead> <tr> <th></th> <th>Condition</th> <th></th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>100 mW input power</td> <td>2.27 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>22.6 W/kg $\pm$ 24.2 % (k=2)</td> </tr> </tbody> </table>			Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	35.9	4.76 mho/m	Measured Head TSL parameters	(22.0 $\pm$ 0.2) °C	35.2 $\pm$ 5 %	4.73 mho/m $\pm$ 5 %	Head TSL temperature change during test	<1.0 °C	—	—		Condition		SAR measured	100 mW input power	7.84 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	78.1 W/kg $\pm$ 24.4 % (k=2)		Condition		SAR measured	100 mW input power	2.27 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	22.6 W/kg $\pm$ 24.2 % (k=2)
	Condition																																																						
SAR measured	250 mW input power	7.75 W/kg																																																					
SAR for nominal Head TSL parameters	normalized to 1W	77.8 W/kg $\pm$ 24.4 % (k=2)																																																					
	Condition																																																						
SAR measured	250 mW input power	2.22 W/kg																																																					
SAR for nominal Head TSL parameters	normalized to 1W	22.1 W/kg $\pm$ 24.2 % (k=2)																																																					
	Temperature	Permittivity	Conductivity																																																				
Nominal Head TSL parameters	22.0 °C	35.9	4.76 mho/m																																																				
Measured Head TSL parameters	(22.0 $\pm$ 0.2) °C	35.2 $\pm$ 5 %	4.73 mho/m $\pm$ 5 %																																																				
Head TSL temperature change during test	<1.0 °C	—	—																																																				
	Condition																																																						
SAR measured	100 mW input power	7.84 W/kg																																																					
SAR for nominal Head TSL parameters	normalized to 1W	78.1 W/kg $\pm$ 24.4 % (k=2)																																																					
	Condition																																																						
SAR measured	100 mW input power	2.27 W/kg																																																					
SAR for nominal Head TSL parameters	normalized to 1W	22.6 W/kg $\pm$ 24.2 % (k=2)																																																					
Certificate No: Z22-60187 Page 1 of 10		Certificate No: Z22-60187 Page 4 of 10																																																					

TTL Calibration Laboratory		CAICT																			
In Collaboration with Address: No.52 HuaYuan Road, Haidian District, Beijing, 100191, China Tel: +86-10-62420217 E-mail: info@ttsl.com.cn http://www.ttsl.com.cn		In Collaboration with Address: No.52 HuaYuan Road, Haidian District, Beijing, 100191, China Tel: +86-10-62420217 E-mail: info@ttsl.com.cn http://www.ttsl.com.cn																			
Head TSL parameters at 5600MHz The following parameters and calculations were applied: <table border="1"> <thead> <tr> <th></th> <th>Temperature</th> <th>Permittivity</th> <th>Conductivity</th> </tr> </thead> <tbody> <tr> <td>Nominal Head TSL parameters</td> <td>22.0 °C</td> <td>35.5</td> <td>5.07 mho/m</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>(22.0 $\pm$ 0.2) °C</td> <td>34.7 $\pm$ 5 %</td> <td>5.05 mho/m $\pm$ 5 %</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td><1.0 °C</td> <td>—</td> <td>—</td> </tr> </tbody> </table>					Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	35.5	5.07 mho/m	Measured Head TSL parameters	(22.0 $\pm$ 0.2) °C	34.7 $\pm$ 5 %	5.05 mho/m $\pm$ 5 %	Head TSL temperature change during test	<1.0 °C	—	—		
	Temperature	Permittivity	Conductivity																		
Nominal Head TSL parameters	22.0 °C	35.5	5.07 mho/m																		
Measured Head TSL parameters	(22.0 $\pm$ 0.2) °C	34.7 $\pm$ 5 %	5.05 mho/m $\pm$ 5 %																		
Head TSL temperature change during test	<1.0 °C	—	—																		
SAR result with Head TSL at 5600MHz SAR averaged over 1 cm^2 (1 g) of Head TSL <table border="1"> <thead> <tr> <th></th> <th>Condition</th> <th></th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>100 mW input power</td> <td>8.12 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>80.8 W/kg $\pm$ 24.4 % (k=2)</td> </tr> </tbody> </table> SAR averaged over 10 cm^2 (10 g) of Head TSL <table border="1"> <thead> <tr> <th></th> <th>Condition</th> <th></th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>100 mW input power</td> <td>2.30 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>22.9 W/kg $\pm$ 24.2 % (k=2)</td> </tr> </tbody> </table>					Condition		SAR measured	100 mW input power	8.12 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	80.8 W/kg $\pm$ 24.4 % (k=2)		Condition		SAR measured	100 mW input power	2.30 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	22.9 W/kg $\pm$ 24.2 % (k=2)
	Condition																				
SAR measured	100 mW input power	8.12 W/kg																			
SAR for nominal Head TSL parameters	normalized to 1W	80.8 W/kg $\pm$ 24.4 % (k=2)																			
	Condition																				
SAR measured	100 mW input power	2.30 W/kg																			
SAR for nominal Head TSL parameters	normalized to 1W	22.9 W/kg $\pm$ 24.2 % (k=2)																			
Head TSL parameters at 5800MHz The following parameters and calculations were applied: <table border="1"> <thead> <tr> <th></th> <th>Temperature</th> <th>Permittivity</th> <th>Conductivity</th> </tr> </thead> <tbody> <tr> <td>Nominal Head TSL parameters</td> <td>22.0 °C</td> <td>35.5</td> <td>5.27 mho/m</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>(22.0 $\pm$ 0.2) °C</td> <td>34.4 $\pm$ 5 %</td> <td>5.25 mho/m $\pm$ 5 %</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td><1.0 °C</td> <td>—</td> <td>—</td> </tr> </tbody> </table>					Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	35.5	5.27 mho/m	Measured Head TSL parameters	(22.0 $\pm$ 0.2) °C	34.4 $\pm$ 5 %	5.25 mho/m $\pm$ 5 %	Head TSL temperature change during test	<1.0 °C	—	—		
	Temperature	Permittivity	Conductivity																		
Nominal Head TSL parameters	22.0 °C	35.5	5.27 mho/m																		
Measured Head TSL parameters	(22.0 $\pm$ 0.2) °C	34.4 $\pm$ 5 %	5.25 mho/m $\pm$ 5 %																		
Head TSL temperature change during test	<1.0 °C	—	—																		
SAR result with Head TSL at 5800MHz SAR averaged over 1 cm^2 (1 g) of Head TSL <table border="1"> <thead> <tr> <th></th> <th>Condition</th> <th></th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>100 mW input power</td> <td>7.71 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>76.7 W/kg $\pm$ 24.4 % (k=2)</td> </tr> </tbody> </table> SAR averaged over 10 cm^2 (10 g) of Head TSL <table border="1"> <thead> <tr> <th></th> <th>Condition</th> <th></th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>100 mW input power</td> <td>2.16 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>21.5 W/kg $\pm$ 24.2 % (k=2)</td> </tr> </tbody> </table>					Condition		SAR measured	100 mW input power	7.71 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	76.7 W/kg $\pm$ 24.4 % (k=2)		Condition		SAR measured	100 mW input power	2.16 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	21.5 W/kg $\pm$ 24.2 % (k=2)
	Condition																				
SAR measured	100 mW input power	7.71 W/kg																			
SAR for nominal Head TSL parameters	normalized to 1W	76.7 W/kg $\pm$ 24.4 % (k=2)																			
	Condition																				
SAR measured	100 mW input power	2.16 W/kg																			
SAR for nominal Head TSL parameters	normalized to 1W	21.5 W/kg $\pm$ 24.2 % (k=2)																			
Certificate No: Z22-60187 Page 5 of 10		Certificate No: Z22-60187 Page 8 of 10																			

TTL Calibration Laboratory		CAICT							
In Collaboration with Address: No.52 HuaYuan Road, Haidian District, Beijing, 100191, China Tel: +86-10-62420217 E-mail: info@ttsl.com.cn http://www.ttsl.com.cn		In Collaboration with Address: No.52 HuaYuan Road, Haidian District, Beijing, 100191, China Tel: +86-10-62420217 E-mail: info@ttsl.com.cn http://www.ttsl.com.cn							
Appendix (Additional assessments outside the scope of CNAS L0570)									
Antenna Parameters with Head TSL at 5200MHz <table border="1"> <thead> <tr> <th></th> <th></th> </tr> </thead> <tbody> <tr> <td>Impedance, transformed to feed point</td> <td>48.10 - j0.03j</td> </tr> <tr> <td>Return Loss</td> <td>-23.8dB</td> </tr> </tbody> </table>						Impedance, transformed to feed point	48.10 - j0.03j	Return Loss	-23.8dB
Impedance, transformed to feed point	48.10 - j0.03j								
Return Loss	-23.8dB								
Antenna Parameters with Head TSL at 5300MHz <table border="1"> <thead> <tr> <th></th> <th></th> </tr> </thead> <tbody> <tr> <td>Impedance, transformed to feed point</td> <td>47.80 - 2.42j</td> </tr> <tr> <td>Return Loss</td> <td>-23.5dB</td> </tr> </tbody> </table>						Impedance, transformed to feed point	47.80 - 2.42j	Return Loss	-23.5dB
Impedance, transformed to feed point	47.80 - 2.42j								
Return Loss	-23.5dB								
Antenna Parameters with Head TSL at 5500MHz <table border="1"> <thead> <tr> <th></th> <th></th> </tr> </thead> <tbody> <tr> <td>Impedance, transformed to feed point</td> <td>50.30 - 4.26j</td> </tr> <tr> <td>Return Loss</td> <td>-27.4dB</td> </tr> </tbody> </table>						Impedance, transformed to feed point	50.30 - 4.26j	Return Loss	-27.4dB
Impedance, transformed to feed point	50.30 - 4.26j								
Return Loss	-27.4dB								
Antenna Parameters with Head TSL at 5600MHz <table border="1"> <thead> <tr> <th></th> <th></th> </tr> </thead> <tbody> <tr> <td>Impedance, transformed to feed point</td> <td>54.80 - 4.86j</td> </tr> <tr> <td>Return Loss</td> <td>-24.6dB</td> </tr> </tbody> </table>						Impedance, transformed to feed point	54.80 - 4.86j	Return Loss	-24.6dB
Impedance, transformed to feed point	54.80 - 4.86j								
Return Loss	-24.6dB								
Antenna Parameters with Head TSL at 5800MHz <table border="1"> <thead> <tr> <th></th> <th></th> </tr> </thead> <tbody> <tr> <td>Impedance, transformed to feed point</td> <td>51.50 - 5.61j</td> </tr> <tr> <td>Return Loss</td> <td>-24.9dB</td> </tr> </tbody> </table>						Impedance, transformed to feed point	51.50 - 5.61j	Return Loss	-24.9dB
Impedance, transformed to feed point	51.50 - 5.61j								
Return Loss	-24.9dB								
Certificate No: Z22-60187 Page 9 of 10		Certificate No: Z22-60187 Page 9 of 10							





Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx>; and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weyue Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

In Collaboration with TTL speaq CALIBRATION LABORATORY CAICT Address: No.52 Huayuan Road, Haidian District, Beijing, 100191, China Tel: +86-10-62302117 E-mail: csl@speaq.com.cn http://www.speaq.com.cn General Antenna Parameters and Design Electrical Delay (one direction): 1.021 ns After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured. The dipole is made of standard semirigid coaxial cable. The carrier conductor of the feeding line is directly connected to the second arm of the dipole. This antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged. Additional EUT Data Manufactured by: SPEAQ Certificate No: Z22-60187 Page 1 of 10	In Collaboration with TTL speaq CALIBRATION LABORATORY CAICT Address: No.52 Huayuan Road, Haidian District, Beijing, 100191, China Tel: +86-10-62302117 E-mail: csl@speaq.com.cn http://www.speaq.com.cn DASY5 Validation Report for Head TSL Test Laboratory: CTTL, Beijing, China Date: 2022-09-01 DUT: Dipole 5GHz; Type: D5GHzV2; Serial: D5GHzV2 - SN: 1095 Communication System: CW; Frequency: 5200 MHz; Frequency: 5300 MHz; Frequency: 5500 MHz; Frequency: 5600 MHz; Frequency: 5800 MHz; Duty Cycle: 1:1 Medium parameters used: $f = 5200$ MHz; $\sigma = 4.82$ S/m; $\epsilon = 35.39$; $\rho = 1000$ kg/m³ Medium parameters used: $f = 5300$ MHz; $\sigma = 4.73$ S/m; $\epsilon = 35.19$; $\rho = 1000$ kg/m³ Medium parameters used: $f = 5500$ MHz; $\sigma = 4.939$ S/m; $\epsilon = 34.83$; $\rho = 1000$ kg/m³ Medium parameters used: $f = 5600$ MHz; $\sigma = 5.051$ S/m; $\epsilon = 34.89$; $\rho = 1000$ kg/m³ Medium parameters used: $f = 5800$ MHz; $\sigma = 5.247$ S/m; $\epsilon = 34.42$; $\rho = 1000$ kg/m³ Phantom section: Right Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007) DASY5 Configuration: - Probe: EX3DV4 - SN7484; ConvF(5.6, 5.6, 5.6) @ 5200 MHz; ConvF(5.32, 5.32, 5.32) @ 5300 MHz; ConvF(5.11, 5.11, 5.11) @ 5500 MHz; ConvF(4.91, 4.91, 4.91) @ 5600 MHz; ConvF(5.5, 5.5, 5.5) @ 5800 MHz; Calibrated: 2022-01-28 - Sensor-Surface: 1.4mm (Mechanical Surface Detection) - Electronics: DA84 Sn1556; Calibrated: 2022-01-12 - Phantom: MPF_V5.1C (20deg probe tilt); Type: QD 000 P51 Cx; Serial: 1062 - DASY52 52.10.4(1535); SEMCAD X 14.6.14(7501) Dipole Calibration /Pin=100mW, d=10mm, f=5200 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm; Reference Value = 60.80 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 29.8 W/kg SAR(1 g) = 7.73 W/kg; SAR(10 g) = 2.22 W/kg Smallest distance from peaks to all points 3 dB below = 7.2 mm Ratio of SAR at M2 to SAR at M1 = 66.8% Maximum value of SAR (measured) = 19.3 W/kg Dipole Calibration /Pin=100mW, d=10mm, f=5300 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm; Reference Value = 61.08 V/m; Power Drift = -0.07 dB Peak SAR (extrapolated) = 31.5 W/kg SAR(1 g) = 7.94 W/kg; SAR(10 g) = 2.27 W/kg Smallest distance from peaks to all points 3 dB below = 7.2 mm Ratio of SAR at M2 to SAR at M1 = 65.5% Maximum value of SAR (measured) = 19.0 W/kg Certificate No: Z22-60187 Page 2 of 10
In Collaboration with TTL speaq CALIBRATION LABORATORY CAICT Address: No.52 Huayuan Road, Haidian District, Beijing, 100191, China Tel: +86-10-62302117 E-mail: csl@speaq.com.cn http://www.speaq.com.cn Dipole Calibration /Pin=100mW, d=10mm, f=5500 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm; Reference Value = 61.92 V/m; Power Drift = -0.08 dB Peak SAR (extrapolated) = 34.7 W/kg SAR(1 g) = 8.28 W/kg; SAR(10 g) = 2.34 W/kg Smallest distance from peaks to all points 3 dB below = 7.2 mm Ratio of SAR at M2 to SAR at M1 = 63.9% Maximum value of SAR (measured) = 20.2 W/kg Dipole Calibration /Pin=100mW, d=10mm, f=5600 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm; Reference Value = 65.08 V/m; Power Drift = -0.07 dB Peak SAR (extrapolated) = 35.2 W/kg SAR(1 g) = 8.12 W/kg; SAR(10 g) = 2.3 W/kg Smallest distance from peaks to all points 3 dB below = 7.2 mm Ratio of SAR at M2 to SAR at M1 = 62.5% Maximum value of SAR (measured) = 19.1 W/kg Dipole Calibration /Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm; Reference Value = 62.13 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 34.8 W/kg SAR(1 g) = 7.71 W/kg; SAR(10 g) = 2.16 W/kg Smallest distance from peaks to all points 3 dB below = 7.2 mm Ratio of SAR at M2 to SAR at M1 = 61.6% Maximum value of SAR (measured) = 18.7 W/kg  0 dB = 18.7 W/kg = 12.72 dBW/kg Certificate No: Z22-60187 Page 3 of 10	In Collaboration with TTL speaq CALIBRATION LABORATORY CAICT Address: No.52 Huayuan Road, Haidian District, Beijing, 100191, China Tel: +86-10-62302117 E-mail: csl@speaq.com.cn http://www.speaq.com.cn Impedance Measurement Plot for Head TSL  Certificate No: Z22-60187 Page 4 of 10



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx>; and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weyue Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

(86-512)57355888 (86-512)57370818 www.sgs.com.cn
(86-512)57355888 (86-512)57370818 sgs.china@sgs.com

2 DAE4 - SN 1245

Schmid & Partner Engineering AG Zugzwangstrasse 43, 8004 Zurich, Switzerland Phone +41 (0)43 9702, Fax +41 (0)43 9775 info@sgs.ch, info@sgs.cn IMPORTANT NOTICE USAGE OF THE DAE4 The DAE4 unit is a delicate, high precision instrument and requires careful treatment by the user. There are no serviceable parts inside the DAE4. Special attention shall be given to the following points: Battery Exchange: The battery cover of the DAE4 unit is fixed using a screw, over tightening the screw may cause the threads inside the DAE4 to wear out. Shipping of the DAE4: Before shipping the DAE4 to SPEAG for calibration, remove the batteries and pack the DAE4 in an anti-static bag. This anti-static bag shall then be packed into a larger box or container which protects the DAE4 from impacts during transportation. The package shall be marked to indicate that a fragile instrument is inside. E-stop Failures: Touch detection may be malfunctioning due to broken magnets in the E-stop. Rough handling of the E-stop may lead to damage of these magnets. Touch and collision errors are often caused by dust and dirt accumulated in the E-stop. To prevent E-stop failure, the customer shall always mount the probe to the DAE4 carefully and keep the DAE4 unit in a non-dusty environment if not used for measurements. Repair: Minor repairs are performed at no extra cost during the annual calibration. However, SPEAG reserves the right to charge for any repair especially if rough unprofessional handling caused the defect. DASY Configuration Files: Since the exact values of the DAE4 input resistances, as measured during the calibration procedure of a DAE4 unit, are not used by the DASY software, a nominal value of 200 MOhm is given in the corresponding configuration file. Important Note: Warranty and calibration is void if the DAE4 unit is disassembled partly or fully by the Customer. Important Note: Never attempt to grease or oil the E-stop assembly. Cleaning and readjusting of the E-stop assembly is allowed by certified SPEAG personnel only and is part of the annual calibration procedure. Important Note: To prevent damage of the DAE4 probe connector pins, use great care when installing the probe to the DAE4. Carefully connect the probe with the connector notch oriented in the mating position. Avoid any rotational movement of the probe body versus the DAE4 while turning the locking nut of the connector. The same care shall be used when disconnecting the probe from the DAE4. TN_EH190306AE_DAE4.docx 07.03.2019	Calibration Laboratory of Schmid & Partner Engineering AG Zugzwangstrasse 43, 8004 Zurich, Switzerland Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates Client: SGS-CN (Auden) Certificate No.: DAE4-1245_May22 CALIBRATION CERTIFICATE Object: DAE4 - SD 000 D04 BM - SN: 1245 Calibration procedure(s): QA CAL-06 v30 Calibration procedure for the data acquisition electronics (DAE) Calibration date: May 30, 2022 This calibration certificate documents the traceability to national standards, which include the physical units of measurement (SI). The measurement and the uncertainties with confidence probability are given on the following pages and are part of the certificate. All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 1) °C and humidity < 70%. Calibration Equipment used (MATE critical for calibration) <table border="1"><thead><tr><th>Primary Standards</th><th>Q#</th><th>Due Date (Certificate No.)</th><th>Scheduled Calibration</th></tr></thead><tbody><tr><td>Kathrein Multimeter Type 2001</td><td>SN: 0810278</td><td>31-Aug-21 (No: 31368)</td><td>Aug-22</td></tr></tbody></table> <table border="1"><thead><tr><th>Secondary Standards</th><th>Q#</th><th>Check Date (in hours)</th><th>Scheduled Check</th></tr></thead><tbody><tr><td>Auto DAE Calibration Unit</td><td>SE LWS 053 AA 1001</td><td>24-Jan-22 (in hours check)</td><td>In house check: Jan-22</td></tr><tr><td>Calibrator New V1</td><td>SE LWS 006 AA 1002</td><td>24-Jan-22 (in hours check)</td><td>In house check: Jan-22</td></tr></tbody></table> Calibrated by: Dominique Steiner Function: Laboratory Technician Signature: <i>[Signature]</i> Approved by: (Ben Kuhn) Technical Manager: <i>[Signature]</i> This calibration certificate shall not be reproduced except in full without written approval of the laboratory. Certificate No.: DAE4-1245_May22 Page 1 of 5	Primary Standards	Q#	Due Date (Certificate No.)	Scheduled Calibration	Kathrein Multimeter Type 2001	SN: 0810278	31-Aug-21 (No: 31368)	Aug-22	Secondary Standards	Q#	Check Date (in hours)	Scheduled Check	Auto DAE Calibration Unit	SE LWS 053 AA 1001	24-Jan-22 (in hours check)	In house check: Jan-22	Calibrator New V1	SE LWS 006 AA 1002	24-Jan-22 (in hours check)	In house check: Jan-22
Primary Standards	Q#	Due Date (Certificate No.)	Scheduled Calibration																		
Kathrein Multimeter Type 2001	SN: 0810278	31-Aug-21 (No: 31368)	Aug-22																		
Secondary Standards	Q#	Check Date (in hours)	Scheduled Check																		
Auto DAE Calibration Unit	SE LWS 053 AA 1001	24-Jan-22 (in hours check)	In house check: Jan-22																		
Calibrator New V1	SE LWS 006 AA 1002	24-Jan-22 (in hours check)	In house check: Jan-22																		
Calibration Laboratory of Schmid & Partner Engineering AG Zugzwangstrasse 43, 8004 Zurich, Switzerland Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates Glossary DAE: data acquisition electronics Connector angle: information used in DASY system to align probe sensor X to the robot coordinate system. Methods Applied and Interpretation of Parameters • DC Voltage Measurement: Calibration Factor assessed for use in DASY system by comparison with a calibrated instrument traceable to national standards. The figure given corresponds to the full scale range of the voltmeter in the respective range. • Connector angle: The angle of the connector is assessed measuring the angle mechanically by a tool inserted. Uncertainty is not required. • The following parameters as documented in the Appendix contain technical information as a result from the performance test and require no uncertainty. • DC Voltage Measurement Linearity: Verification of the Linearity at +10% and -10% of the nominal calibration voltage. Influence of offset voltage is included in this measurement. • Common mode sensitivity: Influence of a positive or negative common mode voltage on the differential measurement. • Channel separation: Influence of a voltage on the neighbor channels not subject to an input voltage. • AD Converter Values with inputs shorted: Values on the internal AD converter corresponding to zero input voltage. • Input Offset Measurement: Output voltage and statistical results over a large number of zero voltage measurements. • Input Offset Current: Typical value for information; Maximum channel input offset current, not considering the input resistance. • Input resistance: Typical value for information; DAE input resistance at the connector, during internal auto-zeroing and during measurement. • Low Battery Alarm Voltage: Typical value for information. Below this voltage, a battery alarm signal is generated. • Power consumption: Typical value for information. Supply currents in various operating modes. Certificate No.: DAE4-1245_May22 Page 2 of 5	DC Voltage Measurement AD - Converter Resolution nominal High Range: 1LSB = 6.1 μV, full range = -190...+320 mV Low Range: 1LSB = 61 μV, full range = -1...+32 mV DASY measurement parameters: Auto Zero-Time: 3 sec; Measuring time: 3 sec <table border="1"><thead><tr><th>Calibration Factors</th><th>X</th><th>Y</th><th>Z</th></tr></thead><tbody><tr><td>High Range</td><td>405.265 ± 0.02% (k=2)</td><td>403.974 ± 0.02% (k=2)</td><td>406.092 ± 0.02% (k=2)</td></tr><tr><td>Low Range</td><td>3.99534 ± 1.50% (k=2)</td><td>3.99508 ± 1.50% (k=2)</td><td>4.01015 ± 1.50% (k=2)</td></tr></tbody></table> Connector Angle <table border="1"><thead><tr><th>Connector Angle to be used in DASY system</th><th>30.0° ± 1°</th></tr></thead></table> Certificate No.: DAE4-1245_May22 Page 3 of 5	Calibration Factors	X	Y	Z	High Range	405.265 ± 0.02% (k=2)	403.974 ± 0.02% (k=2)	406.092 ± 0.02% (k=2)	Low Range	3.99534 ± 1.50% (k=2)	3.99508 ± 1.50% (k=2)	4.01015 ± 1.50% (k=2)	Connector Angle to be used in DASY system	30.0° ± 1°						
Calibration Factors	X	Y	Z																		
High Range	405.265 ± 0.02% (k=2)	403.974 ± 0.02% (k=2)	406.092 ± 0.02% (k=2)																		
Low Range	3.99534 ± 1.50% (k=2)	3.99508 ± 1.50% (k=2)	4.01015 ± 1.50% (k=2)																		
Connector Angle to be used in DASY system	30.0° ± 1°																				



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx>; and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com
No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

Appendix (Additional assessments outside the scope of SCS0108)

1. DC Voltage Linearity

High Range	Reading (µV)	Difference (µV)	Error (%)
Channel X + Input	19994.45	1.52	0.00
Channel X - Input	20004.58	2.22	0.01
Channel X + Input	-20000.14	1.12	-0.01
Channel Y + Input	19994.72	1.58	0.00
Channel Y - Input	20001.22	-1.00	-0.00
Channel Y + Input	-20003.05	-1.57	0.01
Channel Z + Input	19992.44	0.19	0.00
Channel Z - Input	20003.09	0.58	0.00
Channel Z + Input	-20001.73	-0.27	0.00

Low Range	Reading (µV)	Difference (µV)	Error (%)
Channel X + Input	2001.91	0.41	0.02
Channel X - Input	200.54	0.65	0.32
Channel X + Input	-197.86	0.07	-0.04
Channel Y + Input	2002.05	0.58	0.03
Channel Y - Input	201.27	-0.57	-0.28
Channel Y + Input	-196.23	-0.06	0.03
Channel Z + Input	2001.96	0.08	0.00
Channel Z - Input	200.09	-1.03	-0.76
Channel Z + Input	-199.85	-1.57	0.79

2. Common mode sensitivity

Channel	Common mode Input Voltage (mV)	High Range Average Reading (µV)	Low Range Average Reading (µV)
Channel X	200	-5.87	-7.69
	-200	9.12	7.79
Channel Y	200	-8.68	-9.28
	-200	8.52	6.36
Channel Z	200	-5.36	-5.60
	-200	3.58	3.06

3. Channel separation

Channel	Input Voltage (mV)	Channel X (µV)	Channel Y (µV)	Channel Z (µV)
Channel X	200	-	4.07	-3.14
Channel Y	200	9.36	-	4.07
Channel Z	200	10.11	7.14	-

Certificate No: DAE4-1245_May22

Page 4 of 5

4. AD-Converter Values with inputs shorted

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

	High Range (LSB)	Low Range (LSB)
Channel X	15984	17040
Channel Y	16562	16768
Channel Z	16035	15968

5. Input Offset Measurement

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

Input (mV)	Average (µV)	min. Offset (µV)	max. Offset (µV)	Std. Deviation (µV)
Channel X	1.00	-0.15	1.93	0.45
Channel Y	-0.18	-1.28	0.94	0.45
Channel Z	-0.58	-2.81	0.58	0.60

6. Input Offset Current

Nominal input circuitry offset current on all channels: <25nA

7. Input Resistance (Typical values for information)

	Zeroing (kOhm)	Measuring (MOhm)
Channel X	200	200
Channel Y	200	200
Channel Z	200	200

8. Low Battery Alarm Voltage (Typical values for information)

Typical values	Alarm Level (VDC)
Supply (+ Vcc)	+7.9
Supply (- Vcc)	-7.6

9. Power Consumption (Typical values for information)

Typical values	Switched off (mA)	Stand by (mA)	Transmitting (mA)
Supply (+ Vcc)	+0.01	+8	+14
Supply (- Vcc)	-0.01	-8	-9

Certificate No: DAE4-1245_May22

Page 5 of 5

3 EX3DV4 - SN 7346

Calibration Laboratory of Schmid & Partner Engineering AG
Zürcherstrasse 63, 8004 Zürich, Switzerland

Accredited by the Swiss Accreditation Service (SAS)
The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Client: **Auden** Certificate No: **EX3-7346_May22**

Accreditation No.: **SCS 0108**

Calibration Certificate

Object: **EX3DV4 - SN 7346**

Calibration procedure: **QA CAL-01 v6, QA CAL-14 v6, QA CAL-23 v5, QA CAL-25 v7**
Calibration procedure for domestic E-field probes

Calibration date: **March 30, 2022**

This calibration certificate documents the traceability to national standards, which require the physical units of measurement (SI). The measurement and the uncertainty with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility, environment temperature (22 ± 0.5 °C) and humidity < 70%.

Calibration Equipment used (MPE): (value for calibration):

Primary Standards	SI	Cal Date (Certificate No.)	Scheduled Calibration
Power meter M4P	86, 10479	08-Apr-21 (No. 217-0501-0303)	Apr-22
Power sensor HP-291	86, 10304	08-Apr-21 (No. 217-0501-0303)	Apr-22
Power sensor MPT-291	86, 10304	08-Apr-21 (No. 217-0501-0303)	Apr-22
Reference DUT attenuator	86, 10304 (20)	08-Apr-21 (No. 217-0501-0303)	Apr-22
DUT	86, 460	13-Dec-21 (No. DAE4-460, 0401)	Dec-22
Reference Probe (S202)	86, 3013	27-Dec-21 (No. E53-3013, 0401)	Dec-22

Secondary Standards

SI	Check Date (in house)	Scheduled Date
Power meter E413B	08-Apr-21 (in house check, Jan-22)	in house check, Jan-22
Power meter E413B	08-Apr-21 (in house check, Jan-22)	in house check, Jan-22
Power sensor E413A	08-Apr-21 (in house check, Jan-22)	in house check, Jan-22
Power sensor E413A	08-Apr-21 (in house check, Jan-22)	in house check, Jan-22
RF generator HP 8446A	08-Apr-21 (in house check, Jan-22)	in house check, Jan-22
RF generator HP 8446A	08-Apr-21 (in house check, Jan-22)	in house check, Jan-22
Network Analyzer B 8700A	08-Apr-21 (in house check, Jan-22)	in house check, Jan-22

Calibrated by: **Henri** Function: **Calibration Technician** Signature: *[Signature]*

Approved by: **Steven Koller** Function: **Quality Manager** Signature: *[Signature]*

This calibration certificate and not the reproduced image of it is valid without approval of the laboratory.

Certificate No: EX3-7346_May22 Page 1 of 24

Calibration Laboratory of Schmid & Partner Engineering AG
Zürcherstrasse 63, 8004 Zürich, Switzerland

Accredited by the Swiss Accreditation Service (SAS)
The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Client: **Auden** Certificate No: **EX3-7346_May22**

Accreditation No.: **SCS 0108**

Glossary:

- TSL: traceability to national standards
- NORM(x,y,z): sensitivity in free space
- Comp: correction factor
- OCF: correction factor (15 day, cycle) of the RF signal
- A, B, C, D: modulation dependent measurement parameters
- Polarization: rotation around probe axis
- Connector Angle: rotation around an axis that is in the plane normal to probe axis (in measurement center)

Information used in DASY system to align probe sensor X to the robot coordinate system.

Calibration is performed according to the following standards:

- IEC 61010-1:2011, Measurement Procedure for the Assessment of Specific Absorption Rate (SAR) Exposure to Radio Frequency Fields From Hand-Held and Body-Worn Wireless Communication Devices, Part 1: SAR Measurement, Implementation and Procedures (Frequency Range of 4 MHz to 10 GHz), October 2020
- ISO 9001:2015, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Methods Applied and Interpretation of Parameters:

- NORM(x,y,z): Assessed for E-field polarization: $E = 0$ (1 ± 100 MHz in TEM-cell; 1 ± 100 MHz R22 waveguide), NORM(x,y,z) are only reference values, i.e., the uncertainty of NORM(x,y,z) does not affect the E-field uncertainty model TSL (see below Comp).
- NORM(x,y,z) = NORM(x,y,z) * frequency response (see Frequency Response Chart). The frequency response is implemented in DASY software version 1.0. The uncertainty of the frequency response is included in the stated uncertainty of Comp.
- OCF(x,y,z): OCF is a numerical factor used to correct the data of power sweep with CW signal (no uncertainty required). OCF does not depend on frequency nor media.
- PAC: PAC is the Peak to Average Ratio that is not calculated but determined based on the signal characteristics.
- A(x,y,z), B(x,y,z), C(x,y,z), D(x,y,z), V(x,y,z), A, B, C, D are numerical measurement parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. V(x,y,z) is the maximum calibration range measured in RMS voltage across the probe.
- Comp: and Boundary Effect Parameters: Assessed in the probe using E-field (or Temperature Transfer Standard for f < 400 MHz) and inside waveguide using analytical field distributions based on power measurements for f > 400 MHz. The same values are used for assessment of the parameters applied by boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are NORM(x,y,z) * Comp. The uncertainty of NORM(x,y,z) is not given for Comp. A frequency dependence of Comp is used in DASY version 1.0 and higher which allows assessing the validity from a 50 MHz to 100 MHz.
- Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (in probe axis). No tolerance required.
- Connector Angle: The angle is assessed using the information given by determining the NORM(x,y,z) uncertainty required.

Certificate No: EX3-7346_May22 Page 2 of 24



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weyue Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

(86-512) 57355888 (86-512) 57370818 www.sgsgroup.com.cn
(86-512) 57355888 (86-512) 57370818 sgs.china@sgs.com

EX3DV4 - SN:7346										March 30, 2022									
DASY/EASY - Parameters of Probe: EX3DV4 - SN:7346										DASY/EASY - Parameters of Probe: EX3DV4 - SN:7346									
Basic Calibration Parameters										Sensor Model Parameters									
Norm. $\Delta V/V$ (mV/V) ¹										C1 C2 C3 C4 C5 C6 C7 C8 C9 C10 C11 C12 C13 C14 C15 C16									
DOP (mV/V) ¹										ms.V ¹ ms.V ¹ ms.V ¹ ms.V ¹ ms.V ¹ ms.V ¹ ms.V ¹ ms.V ¹ ms.V ¹ ms.V ¹ ms.V ¹ ms.V ¹ ms.V ¹ ms.V ¹ ms.V ¹									
101.4 0.45 0.47 0.61 108.9 ± 10.1 %										X 39.2 291.80 35.10 5.63 0.33 5.02 1.42 0.12 1.01 37.1 270.84 34.12 8.29 0.00 5.01 1.82 0.05 1.01 9.7 69.74 33.37 4.96 0.00 4.94 0.61 0.00 1.00									
Calibration Results for Modulation Response										Other Probe Parameters									
UID Communication System Name A dB B dB-μV C dB D dB VR mV Max dBm Max dBm (dBS)										Sensor Arrangement Triangular									
T1 C1V X 0.00 0.00 0.00 0.00 143.8 ± 5.0 %										Connector Angle (°) -166.1									
T2005 AAA X 3.33 68.99 11.68 10.00 65.0 ± 6.5 %										Mechanical Surface Detection Mode enabled									
T1035 AAA X 3.33 68.99 11.68 10.00 65.0 ± 6.5 %										Optical Surface Detection Mode disabled									
T1035 AAA X 3.33 68.99 11.68 10.00 65.0 ± 6.5 %										Probe Overall Length 237 mm									
T1035 AAA X 3.33 68.99 11.68 10.00 65.0 ± 6.5 %										Probe Body Diameter 9 mm									
T1035 AAA X 3.33 68.99 11.68 10.00 65.0 ± 6.5 %										Tip Length 9 mm									
T1035 AAA X 3.33 68.99 11.68 10.00 65.0 ± 6.5 %										Tip Diameter 2.5 mm									
T1035 AAA X 3.33 68.99 11.68 10.00 65.0 ± 6.5 %										Probe Tip to Sensor X Calibration Point 1 mm									
T1035 AAA X 3.33 68.99 11.68 10.00 65.0 ± 6.5 %										Probe Tip to Sensor Y Calibration Point 1 mm									
T1035 AAA X 3.33 68.99 11.68 10.00 65.0 ± 6.5 %										Probe Tip to Sensor Z Calibration Point 1 mm									
T1035 AAA X 3.33 68.99 11.68 10.00 65.0 ± 6.5 %										Recommended Measurement Distance from Surface 1.4 mm									
Note: For details on UID parameters see Appendix										Note: Measurement distance from surface can be increased to 3-4 mm for an Area Scan job.									
The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.																			
Footnote: 1) The uncertainties of Norm. $\Delta V/V$ and DOP are not affected by the field uncertainty within 10% (see Page 2) of the field.																			
Footnote: 2) Humidity measurement parameter, uncertainty not required.																			
Footnote: 3) Uncertainty is determined using the value of the field uncertainty multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.																			
Certificate No: EX3-7346_Mar22										Page 3 of 34									

EX3DV4 - SN:7346										March 30, 2022									
DASY/EASY - Parameters of Probe: EX3DV4 - SN:7346										DASY/EASY - Parameters of Probe: EX3DV4 - SN:7346									
Calibration Parameter Determined in Head Tissue Simulating Media										Calibration Parameter Determined in Head Tissue Simulating Media									
f (MHz) Relative Permittivity Conductivity (S/m) Coref X Coref Y Coref Z Alpha ¹ Depth (mm) Unc. (k=2)										f (MHz) Relative Permittivity Conductivity (S/m) Coref X Coref Y Coref Z Alpha ¹ Depth (mm) Unc. (k=2)									
750 41.9 0.89 10.56 10.56 10.56 0.55 0.85 ± 12.0 %										6500 34.5 6.07 5.30 5.30 5.30 0.20 2.50 ± 18.0 %									
835 41.5 0.90 10.12 10.12 10.12 0.42 0.96 ± 12.0 %																			
900 41.5 0.97 10.10 10.10 10.10 0.53 0.80 ± 12.0 %																			
1450 40.5 1.20 9.26 9.26 9.26 0.50 0.80 ± 12.0 %																			
1750 40.1 1.37 8.83 8.83 8.83 0.34 0.86 ± 12.0 %																			
1900 40.0 1.40 8.48 8.48 8.48 0.35 0.86 ± 12.0 %																			
2000 40.0 1.40 8.35 8.35 8.35 0.34 0.86 ± 12.0 %																			
2300 39.5 1.67 7.86 7.86 7.86 0.39 0.90 ± 12.0 %																			
2450 39.2 1.80 7.63 7.63 7.63 0.41 0.90 ± 12.0 %																			
2600 39.0 1.96 7.33 7.33 7.33 0.44 0.90 ± 12.0 %																			
3300 38.2 2.71 7.15 7.15 7.15 0.30 1.35 ± 13.1 %																			
3500 37.8 2.81 7.14 7.14 7.14 0.30 1.35 ± 13.1 %																			
3750 37.7 3.12 6.85 6.85 6.85 0.30 1.35 ± 13.1 %																			
3900 37.5 3.32 6.71 6.71 6.71 0.40 1.60 ± 13.1 %																			
4100 37.2 3.53 6.58 6.58 6.58 0.40 1.60 ± 13.1 %																			
4200 37.1 3.63 6.30 6.30 6.30 0.40 1.70 ± 13.1 %																			
4400 36.9 3.84 6.24 6.24 6.24 0.40 1.70 ± 13.1 %																			
4600 36.7 4.04 6.11 6.11 6.11 0.40 1.70 ± 13.1 %																			
4800 36.4 4.25 5.98 5.98 5.98 0.40 1.80 ± 13.1 %																			
4900 36.3 4.40 5.84 5.84 5.84 0.40 1.80 ± 13.1 %																			
5200 36.0 4.66 5.25 5.25 5.25 0.40 1.80 ± 13.1 %																			
5300 35.9 4.78 5.12 5.12 5.12 0.40 1.80 ± 13.1 %																			
5500 35.6 4.96 4.85 4.85 4.85 0.40 1.80 ± 13.1 %																			
5600 35.5 5.07 4.70 4.70 4.70 0.40 1.80 ± 13.1 %																			
5800 35.3 5.27 4.75 4.75 4.75 0.40 1.80 ± 13.1 %																			
Footnote: 1) Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the Coref uncertainty at calibration frequency and the uncertainty for the indicated frequency band.																			
Footnote: 2) As frequencies > 30 GHz, the validity of tissue parameters (ε and σ) can be related to a 10% if liquid compensation formula is applied to measured GHz values. The uncertainty is the RSS of the Coref uncertainty for indicated target tissue parameters.																			
Footnote: 3) Applicable for frequencies below 3 GHz. Below 3 GHz, the remaining deviation due to the boundary effect after compensation is always less than 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.																			
Certificate No: EX3-7346_Mar22										Page 5 of 24									

EX3DV4 - SN:7346										March 30, 2022									
DASY/EASY - Parameters of Probe: EX3DV4 - SN:7346										DASY/EASY - Parameters of Probe: EX3DV4 - SN:7346									
Calibration Parameter Determined in Head Tissue Simulating Media										Calibration Parameter Determined in Head Tissue Simulating Media									
f (MHz) Relative Permittivity Conductivity (S/m) Coref X Coref Y Coref Z Alpha ¹ Depth (mm) Unc. (k=2)										f (MHz) Relative Permittivity Conductivity (S/m) Coref X Coref Y Coref Z Alpha ¹ Depth (mm) Unc. (k=2)									
6500 34.5 6.07 5.30 5.30 5.30 0.20 2.50 ± 18.0 %																			
Footnote: 1) Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the Coref uncertainty at calibration frequency and the uncertainty for the indicated frequency band.																			
Footnote: 2) As frequencies > 30 GHz, the validity of tissue parameters (ε and σ) can be related to a 10% if liquid compensation formula is applied to measured GHz values. The uncertainty is the RSS of the Coref uncertainty for indicated target tissue parameters.																			
Footnote: 3) Applicable for frequencies below 3 GHz. Below 3 GHz, the remaining deviation due to the boundary effect after compensation is always less than 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.																			
Certificate No: EX3-7346_Mar22										Page 6 of 24									




Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

(86-512) 57355888 (86-512) 57370818 www.sgs.com.cn
(86-512) 57355888 (86-512) 57370818 sgs.china@sgs.com



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

DC33042-EN7348			March 31, 2022			EX33046-EN7348			March 31, 2022		
Appendix: Modulation Calibration Parameters											
Line	Ref	Communication System Name	Group	PAE (dB)	Use?	Line	Ref	Communication System Name	Group	PAE (dB)	Use?
1001	CAE	SW	SW	0.00	0.0%	10050	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10012	CAE	SW	SW	0.00	0.0%	10051	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10013	CAE	SW	SW	0.00	0.0%	10052	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10014	CAE	SW	SW	0.00	0.0%	10053	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10015	CAE	SW	SW	0.00	0.0%	10054	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10016	CAE	SW	SW	0.00	0.0%	10055	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10017	CAE	SW	SW	0.00	0.0%	10056	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10018	CAE	SW	SW	0.00	0.0%	10057	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10019	CAE	SW	SW	0.00	0.0%	10058	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10020	CAE	SW	SW	0.00	0.0%	10059	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10021	CAE	SW	SW	0.00	0.0%	10060	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10022	CAE	SW	SW	0.00	0.0%	10061	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10023	CAE	SW	SW	0.00	0.0%	10062	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10024	CAE	SW	SW	0.00	0.0%	10063	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10025	CAE	SW	SW	0.00	0.0%	10064	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10026	CAE	SW	SW	0.00	0.0%	10065	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10027	CAE	SW	SW	0.00	0.0%	10066	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10028	CAE	SW	SW	0.00	0.0%	10067	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10029	CAE	SW	SW	0.00	0.0%	10068	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10030	CAE	SW	SW	0.00	0.0%	10069	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10031	CAE	SW	SW	0.00	0.0%	10070	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10032	CAE	SW	SW	0.00	0.0%	10071	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10033	CAE	SW	SW	0.00	0.0%	10072	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10034	CAE	SW	SW	0.00	0.0%	10073	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10035	CAE	SW	SW	0.00	0.0%	10074	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10036	CAE	SW	SW	0.00	0.0%	10075	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10037	CAE	SW	SW	0.00	0.0%	10076	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10038	CAE	SW	SW	0.00	0.0%	10077	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10039	CAE	SW	SW	0.00	0.0%	10078	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10040	CAE	SW	SW	0.00	0.0%	10079	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10041	CAE	SW	SW	0.00	0.0%	10080	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10042	CAE	SW	SW	0.00	0.0%	10081	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10043	CAE	SW	SW	0.00	0.0%	10082	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10044	CAE	SW	SW	0.00	0.0%	10083	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10045	CAE	SW	SW	0.00	0.0%	10084	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10046	CAE	SW	SW	0.00	0.0%	10085	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10047	CAE	SW	SW	0.00	0.0%	10086	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10048	CAE	SW	SW	0.00	0.0%	10087	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10049	CAE	SW	SW	0.00	0.0%	10088	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10050	CAE	SW	SW	0.00	0.0%	10089	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10051	CAE	SW	SW	0.00	0.0%	10090	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10052	CAE	SW	SW	0.00	0.0%	10091	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10053	CAE	SW	SW	0.00	0.0%	10092	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10054	CAE	SW	SW	0.00	0.0%	10093	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10055	CAE	SW	SW	0.00	0.0%	10094	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10056	CAE	SW	SW	0.00	0.0%	10095	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10057	CAE	SW	SW	0.00	0.0%	10096	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10058	CAE	SW	SW	0.00	0.0%	10097	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10059	CAE	SW	SW	0.00	0.0%	10098	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10060	CAE	SW	SW	0.00	0.0%	10099	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10061	CAE	SW	SW	0.00	0.0%	10100	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10062	CAE	SW	SW	0.00	0.0%	10101	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10063	CAE	SW	SW	0.00	0.0%	10102	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10064	CAE	SW	SW	0.00	0.0%	10103	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10065	CAE	SW	SW	0.00	0.0%	10104	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10066	CAE	SW	SW	0.00	0.0%	10105	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10067	CAE	SW	SW	0.00	0.0%	10106	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10068	CAE	SW	SW	0.00	0.0%	10107	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10069	CAE	SW	SW	0.00	0.0%	10108	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10070	CAE	SW	SW	0.00	0.0%	10109	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10071	CAE	SW	SW	0.00	0.0%	10110	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10072	CAE	SW	SW	0.00	0.0%	10111	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10073	CAE	SW	SW	0.00	0.0%	10112	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10074	CAE	SW	SW	0.00	0.0%	10113	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10075	CAE	SW	SW	0.00	0.0%	10114	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10076	CAE	SW	SW	0.00	0.0%	10115	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10077	CAE	SW	SW	0.00	0.0%	10116	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10078	CAE	SW	SW	0.00	0.0%	10117	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10079	CAE	SW	SW	0.00	0.0%	10118	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10080	CAE	SW	SW	0.00	0.0%	10119	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10081	CAE	SW	SW	0.00	0.0%	10120	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10082	CAE	SW	SW	0.00	0.0%	10121	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10083	CAE	SW	SW	0.00	0.0%	10122	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10084	CAE	SW	SW	0.00	0.0%	10123	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10085	CAE	SW	SW	0.00	0.0%	10124	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10086	CAE	SW	SW	0.00	0.0%	10125	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10087	CAE	SW	SW	0.00	0.0%	10126	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10088	CAE	SW	SW	0.00	0.0%	10127	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10089	CAE	SW	SW	0.00	0.0%	10128	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10090	CAE	SW	SW	0.00	0.0%	10129	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10091	CAE	SW	SW	0.00	0.0%	10130	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10092	CAE	SW	SW	0.00	0.0%	10131	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10093	CAE	SW	SW	0.00	0.0%	10132	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10094	CAE	SW	SW	0.00	0.0%	10133	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10095	CAE	SW	SW	0.00	0.0%	10134	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10096	CAE	SW	SW	0.00	0.0%	10135	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10097	CAE	SW	SW	0.00	0.0%	10136	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10098	CAE	SW	SW	0.00	0.0%	10137	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10099	CAE	SW	SW	0.00	0.0%	10138	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10100	CAE	SW	SW	0.00	0.0%	10139	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10101	CAE	SW	SW	0.00	0.0%	10140	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10102	CAE	SW	SW	0.00	0.0%	10141	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10103	CAE	SW	SW	0.00	0.0%	10142	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10104	CAE	SW	SW	0.00	0.0%	10143	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10105	CAE	SW	SW	0.00	0.0%	10144	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10106	CAE	SW	SW	0.00	0.0%	10145	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10107	CAE	SW	SW	0.00	0.0%	10146	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10108	CAE	SW	SW	0.00	0.0%	10147	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10109	CAE	SW	SW	0.00	0.0%	10148	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	8.87	0.6%
10110	CAE	SW	SW	0.00	0.0%	10149	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz			



EX3034-SN-7346							March 30, 2022	
10414	AAA	WLAN CDF4 64-QAM 40MHz	Generic	8.54	± 0.6 %			
10415	AAA	IEEE 802.11b WFI 2.4 GHz (DSSS, 1 Mbps, R90c-d)	WLAN	1.54	± 0.6 %			
10416	AAA	IEEE 802.11g WFI 2.4 GHz (DSSS-OFDM, 5.5 Mbps, R90c-d)	WLAN	8.23	± 0.6 %			
10417	AAC	IEEE 802.11n WFI 5 GHz (OFDM, 6 Mbps, R90c-d)	WLAN	8.23	± 0.6 %			
10418	AAA	IEEE 802.11g WFI 2.4 GHz (DSSS-OFDM, 5.5 Mbps, R90c-d)	WLAN	8.14	± 0.6 %			
10419	AAA	IEEE 802.11g WFI 2.4 GHz (DSSS-OFDM, 5.5 Mbps, R90c-d)	WLAN	8.19	± 0.6 %			
10420	AAC	IEEE 802.11n HT Overhead, 7.2 Mbps, R90c-d	WLAN	8.32	± 0.6 %			
10421	AAC	IEEE 802.11n HT Overhead, 6.3 Mbps, 16-QAM	WLAN	8.47	± 0.6 %			
10424	AAC	IEEE 802.11n HT Overhead, 7.2 Mbps, 64-QAM	WLAN	8.40	± 0.6 %			
10425	AAC	IEEE 802.11n HT Overhead, 14 Mbps, R90c-d	WLAN	8.41	± 0.6 %			
10426	AAC	IEEE 802.11n HT Overhead, 30 Mbps, 16-QAM	WLAN	8.45	± 0.6 %			
10427	AAC	IEEE 802.11n HT Overhead, 130 Mbps, 64-QAM	WLAN	8.41	± 0.6 %			
10428	AAD	LTE-FDD (FDDMA, 5 MHz, E-TRP 3.1)	LTE-FDD	8.28	± 0.6 %			
10431	AAD	LTE-FDD (FDDMA, 15 MHz, E-TRP 3.1)	LTE-FDD	8.38	± 0.6 %			
10432	AAC	LTE-FDD (FDDMA, 15 MHz, E-TRP 3.1)	LTE-FDD	8.34	± 0.6 %			
10433	AAC	LTE-FDD (FDDMA, 20 MHz, E-TRP 3.1)	LTE-FDD	8.34	± 0.6 %			
10434	AAA	WCDMA (BS Test Model 1, 64 QPSK)	WCDMA	8.60	± 0.6 %			
10435	AAF	LTE-TDD (SC-FDMA, 1 RB, 20 Mbps, QPSK, UL Sub)	LTE-TDD	7.82	± 0.6 %			
10447	AAD	LTE-FDD (FDDMA, 5 MHz, E-TRP 3.1, Clipping 44%)	LTE-FDD	7.86	± 0.6 %			
10448	AAD	LTE-FDD (FDDMA, 10 MHz, E-TRP 3.1, Clipping 44%)	LTE-FDD	7.93	± 0.6 %			
10449	AAC	LTE-FDD (FDDMA, 15 MHz, E-TRP 3.1, Clipping 44%)	LTE-FDD	7.91	± 0.6 %			
10450	AAC	LTE-FDD (FDDMA, 20 MHz, E-TRP 3.1, Clipping 44%)	LTE-FDD	7.88	± 0.6 %			
10451	AAA	WCDMA (BS Test Model 1, 64 QPSK, Clipping 44%)	WCDMA	7.89	± 0.6 %			
10452	AAD	Validation (Spurs, 10ms, 1ms)	Test	10.00	± 0.6 %			
10453	AAC	IEEE 802.11ac WFI (160MHz, 64-QAM, R90c-d)	WLAN	8.63	± 0.6 %			
10457	AAA	UMTS FDD (SC-HSPA)	WCDMA	6.62	± 0.6 %			
10458	AAA	CDMA2000 (1xEV-DO, Rev. 8, 3.1c)	CDMA2000	8.18	± 0.6 %			
10459	AAA	CDMA2000 (1xEV-DO, Rev. 8, 3.1c)	CDMA2000	8.25	± 0.6 %			
10462	AAA	UMTS FDD (HSPA, R90c-d)	WCDMA	2.39	± 0.6 %			
10463	AAB	LTE-TDD (SC-FDMA, 1 RB, 1.4 Mbps, QPSK, UL Sub)	LTE-TDD	7.82	± 0.6 %			
10462	AAB	LTE-TDD (SC-FDMA, 1 RB, 1.4 Mbps, 16-QAM, UL Sub)	LTE-TDD	8.30	± 0.6 %			
10463	AAB	LTE-TDD (SC-FDMA, 1 RB, 1.4 Mbps, 64-QAM, UL Sub)	LTE-TDD	8.30	± 0.6 %			
10464	AAC	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, QPSK, UL Sub)	LTE-TDD	7.82	± 0.6 %			
10465	AAC	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 16-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10466	AAC	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 64-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10467	AAF	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, QPSK, UL Sub)	LTE-TDD	7.82	± 0.6 %			
10468	AAF	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 16-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10469	AAF	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 64-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10470	AAF	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, QPSK, UL Sub)	LTE-TDD	7.82	± 0.6 %			
10471	AAF	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 16-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10472	AAF	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 64-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10473	AAB	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, QPSK, UL Sub)	LTE-TDD	7.82	± 0.6 %			
10474	AAB	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 16-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10475	AAB	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 64-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10476	AAF	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, QPSK, UL Sub)	LTE-TDD	7.82	± 0.6 %			
10477	AAF	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 16-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10478	AAF	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 64-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10479	AAB	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, QPSK, UL Sub)	LTE-TDD	7.82	± 0.6 %			
10480	AAB	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 16-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10481	AAB	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 64-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10482	AAC	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, QPSK, UL Sub)	LTE-TDD	7.82	± 0.6 %			
10483	AAC	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 16-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10484	AAC	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 64-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10485	AAF	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, QPSK, UL Sub)	LTE-TDD	7.82	± 0.6 %			
10486	AAF	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 16-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10487	AAF	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 64-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10488	AAF	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, QPSK, UL Sub)	LTE-TDD	7.82	± 0.6 %			

Certificate No. EX3-7346_Mar22 Page 15 of 24

EX3034-SN-7346							March 30, 2022	
10489	AAF	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, QPSK, UL Sub)	LTE-TDD	7.82	± 0.6 %			
10490	AAF	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 16-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10491	AAF	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 64-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10492	AAC	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, QPSK, UL Sub)	LTE-TDD	7.82	± 0.6 %			
10493	AAC	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 16-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10494	AAF	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, QPSK, UL Sub)	LTE-TDD	7.82	± 0.6 %			
10495	AAF	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 16-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10496	AAF	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 64-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10497	AAB	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, QPSK, UL Sub)	LTE-TDD	7.82	± 0.6 %			
10498	AAB	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 16-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10499	AAB	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 64-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10500	AAC	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, QPSK, UL Sub)	LTE-TDD	7.82	± 0.6 %			
10501	AAC	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 16-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10502	AAC	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 64-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10503	AAF	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, QPSK, UL Sub)	LTE-TDD	7.82	± 0.6 %			
10504	AAF	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 16-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10505	AAF	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 64-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10506	AAF	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, QPSK, UL Sub)	LTE-TDD	7.82	± 0.6 %			
10507	AAF	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 16-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10508	AAF	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 64-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10509	AAC	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, QPSK, UL Sub)	LTE-TDD	7.82	± 0.6 %			
10510	AAC	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 16-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10511	AAC	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 64-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10512	AAF	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, QPSK, UL Sub)	LTE-TDD	7.82	± 0.6 %			
10513	AAF	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 16-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10514	AAF	LTE-TDD (SC-FDMA, 1 RB, 3.1 Mbps, 64-QAM, UL Sub)	LTE-TDD	8.32	± 0.6 %			
10515	AAA	IEEE 802.11b WFI 2.4 GHz (DSSS, 1 Mbps, R90c-d)	WLAN	1.54	± 0.6 %			
10516	AAA	IEEE 802.11b WFI 2.4 GHz (DSSS, 5.5 Mbps, R90c-d)	WLAN	8.23	± 0.6 %			
10517	AAA	IEEE 802.11g WFI 2.4 GHz (DSSS, 11 Mbps, R90c-d)	WLAN	8.23	± 0.6 %			
10518	AAC	IEEE 802.11n WFI 5 GHz (OFDM, 6 Mbps, R90c-d)	WLAN	8.23	± 0.6 %			
10519	AAC	IEEE 802.11n WFI 5 GHz (OFDM, 12 Mbps, R90c-d)	WLAN	8.23	± 0.6 %			
10520	AAC	IEEE 802.11n WFI 5 GHz (OFDM, 18 Mbps, R90c-d)	WLAN	8.12	± 0.6 %			
10521	AAC	IEEE 802.11n WFI 5 GHz (OFDM, 24 Mbps, R90c-d)	WLAN	7.97	± 0.6 %			
10522	AAC	IEEE 802.11n WFI 5 GHz (OFDM, 36 Mbps, R90c-d)	WLAN	8.45	± 0.6 %			
10523	AAC	IEEE 802.11n WFI 5 GHz (OFDM, 48 Mbps, R90c-d)	WLAN	8.48	± 0.6 %			
10524	AAC	IEEE 802.11n WFI 5 GHz (OFDM, 54 Mbps, R90c-d)	WLAN	8.27	± 0.6 %			
10525	AAC	IEEE 802.11ac WFI (160MHz, MCS9, R90c-d)	WLAN	8.36	± 0.6 %			
10526	AAC	IEEE 802.11ac WFI (160MHz, MCS1, R90c-d)	WLAN	8.42	± 0.6 %			
10527	AAC	IEEE 802.11ac WFI (160MHz, MCS3, R90c-d)	WLAN	8.21	± 0.6 %			
10528	AAC	IEEE 802.11ac WFI (160MHz, MCS5, R90c-d)	WLAN	8.31	± 0.6 %			
10529	AAC	IEEE 802.11ac WFI (160MHz, MCS7, R90c-d)	WLAN	8.36	± 0.6 %			
10530	AAC	IEEE 802.11ac WFI (160MHz, MCS9, R90c-d)	WLAN	8.36	± 0.6 %			
10531	AAC	IEEE 802.11ac WFI (160MHz, MCS1, R90c-d)	WLAN	8.29	± 0.6 %			
10532	AAC	IEEE 802.11ac WFI (160MHz, MCS3, R90c-d)	WLAN	8.29	± 0.6 %			
10533	AAC	IEEE 802.11ac WFI (160MHz, MCS5, R90c-d)	WLAN	8.30	± 0.6 %			
10534	AAC	IEEE 802.11ac WFI (160MHz, MCS7, R90c-d)	WLAN	8.45	± 0.6 %			
10535	AAC	IEEE 802.11ac WFI (160MHz, MCS9, R90c-d)	WLAN	8.45	± 0.6 %			
10536	AAC	IEEE 802.11ac WFI (160MHz, MCS1, R90c-d)	WLAN	8.32	± 0.6 %			
10537	AAC	IEEE 802.11ac WFI (160MHz, MCS3, R90c-d)	WLAN	8.44	± 0.6 %			
10538	AAC	IEEE 802.11ac WFI (160MHz, MCS5, R90c-d)	WLAN	8.34	± 0.6 %			
10539	AAC	IEEE 802.11ac WFI (160MHz, MCS7, R90c-d)	WLAN	8.39	± 0.6 %			
10540	AAC	IEEE 802.11ac WFI (160MHz, MCS9, R90c-d)	WLAN	8.39	± 0.6 %			
10541	AAC	IEEE 802.11ac WFI (160MHz, MCS1, R90c-d)	WLAN	8.46	± 0.6 %			
10542	AAC	IEEE 802.11ac WFI (160MHz, MCS3, R90c-d)	WLAN	8.45	± 0.6 %			
10543	AAC	IEEE 802.11ac WFI (160MHz, MCS5, R90c-d)	WLAN	8.45	± 0.6 %			
10544	AAC	IEEE 802.11ac WFI (160MHz, MCS7, R90c-d)	WLAN	8.47	± 0.6 %			
10545	AAC	IEEE 802.11ac WFI (160MHz, MCS9, R90c-d)	WLAN	8.55	± 0.6 %			
10546	AAC	IEEE 802.11ac WFI (160MHz, MCS1, R90c-d)	WLAN	8.35	± 0.6 %			

Certificate No. EX3-7346_Mar22 Page 16 of 24

EX3034-SN-7346							March 30, 2022	
10547	AAC	IEEE 802.11ac WFI (160MHz, MCS9, R90c-d)	WLAN	8.49	± 0.6 %			
10548	AAC	IEEE 802.11ac WFI (160MHz, MCS7, R90c-d)	WLAN	8.37	± 0.6 %			
10549	AAC	IEEE 802.11ac WFI (160MHz, MCS5, R90c-d)	WLAN	8.39	± 0.6 %			
10550	AAC	IEEE 802.11ac WFI (160MHz, MCS3, R90c-d)	WLAN	8.39	± 0.6 %			
10551	AAC	IEEE 802.11ac WFI (160MHz, MCS1, R90c-d)	WLAN	8.37	± 0.6 %			
10552	AAC	IEEE 802.11ac WFI (160MHz, MCS9, R90c-d)	WLAN	8.42	± 0.6 %			
10553	AAC	IEEE 802.11ac WFI (160MHz, MCS7, R90c-d)	WLAN	8.38	± 0.6 %			
10554	AAC	IEEE 802.11ac WFI (160MHz, MCS5, R90c-d)	WLAN	8.48	± 0.6 %			
10555	AAC	IEEE 802.11ac WFI (160MHz, MCS3, R90c-d)						

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained herein reflects the company's findings at the time of the test and does not constitute an offer of insurance. Client's instructions, if any, The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced, except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this document are for information only to the sample(s) tested and such sample(s) are retained for 8 days of the test date. Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No. 10, Weiyue Road, Innovation Park, Kunshan, Jiangsu, China 215300
 中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

(86-512) 57355888 (86-512) 57370818 www.sgsgroup.com.cn
 (86-512) 57355888 (86-512) 57370818 sgs.china@sgs.com

[illegible]

4 Impedance and return loss

Dipole CLA150 SN 4025				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2021/4/26	-31.4	/	47.8	/
Dipole D450V3 SN 1103				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2021/4/21	-23	/	57.1	/



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is deemed to have read, understood and accepted the Terms and Conditions of the document and the Company's policy and the Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or expiration date of this document will render it void and null and any such use will be prosecuted to the fullest extent of the law. The results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.
Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443,
or email: CN.Doccheck@sgs.com
No.10, Weijue Road, Innovation Park, Kunshan, Jiangsu, China 215300 t(86-512)6735888 f(86-512)67370818 www.sgsgroup.com.cn
中国·江苏·昆山市留学生创业园佳业路10号 邮编 215300 t(86-512)6735888 f(86-512)67370818 sgs.china@sgs.com