

Manufacturer's Statement

FCC ID location justification

2026-07-24

FCC ID: 2AXDT-RFM030

To whom it may concern

We, the undersigned, hereby submit the justification for placing the FCC-ID in the associated User Manual for the Hearing Aids containing the module with the above-mentioned FCC-ID.

According to KDB Publication 784748 D01, when the device is so small, or for such use that it is impracticable to label with a font size that is four points or larger (and the device does not utilize electronic labeling), then the FCC-ID shall be placed in the User Manual and the device packaging or on a removable label attached to the device.

Hearing Aids having the above-mentioned FCC-ID are so small and considering the dimensions of the equipment and the label area, it is impracticable to label FCC-ID with a font size required on the devices. Therefore, the FCC-ID label is placed in the User Manual and on the device packaging of the end product. Instruction for module integration regarding the FCC-ID labelling is included in the integration manual.


If you have any questions, feel free to contact us at the address shown below.

Yours sincerely,



Dr. Bernd Matschkal
Head of Wireless Platform Core Development

bernd.matschkal@wsa.com



Udo Reitingger
Head of QA (System)

udo.reitingger@wsa.com