


單位:mm

外觀等級		A	B	C	D	E
表面	碰撞傷	≤ 0.2	≤ 0.5	≤ 0.2	≤ 1.0	不可見底
	壓痕					
	皺折					
	汙點雜質					
	水痕					
	砂孔	≤ 2.75				
針孔	≤ 0.05	≤ 0.1	≤ 0.3	≤ 0.7		
印刷	溢墨	≤ 0.05	≤ 0.1	≤ 0.3	≤ 0.7	- -
	漏印					
	針孔					



ibiomedi

Model:ES-2020

Electronic Stethoscope

Sound Land Corp.

No.32,Keji 1st Rd.,Guishan Dist.,

Taouyan City 33383,Taiwan

Two AAA battery

FCC ID:2AYQY-ES-2020

CCAM21LP0370T7



MADE IN TAIWAN