



|                         |                                                                           |
|-------------------------|---------------------------------------------------------------------------|
| <b>DOCUMENT NUMBER:</b> | Memo-0039                                                                 |
| <b>DOCUMENT TITLE:</b>  | Sitter On Cue PRO Wireless Nurse Call Adapter FCC ID number location memo |
| <b>DOCUMENT NOTES:</b>  | Rev 00: Initial release, DHF-0082                                         |

#### Document Information

|                |                              |
|----------------|------------------------------|
| Revision: 00   | Vault: Quality Assurance-rel |
| Doc Type: Memo | Status: Release              |

#### Date Information

|                             |                   |
|-----------------------------|-------------------|
| Effective Date: 12 Aug 2021 | Next Review Date: |
| Release Date: 12 Aug 2021   | Expiration Date:  |

#### Control Information

|                   |                  |
|-------------------|------------------|
| Author: SMCCARTHY | Previous Number: |
| Owner: SMCCARTHY  | Change Number:   |

All dates and times are in Central Time

EFFECTIVE



**To:** TIMCO

**From:** TIDI Products

Samantha McCarthy, Senior Product Development Engineer

**Subject:** 8282WL – Nurse Call Adapter FCC ID Location on Cord Warning Labeling

The FCC ID is currently located on the non-removable ¼" mono male jack cord warning label, as indicated in Figure 1. This device is intended to be attached to a hospital's nurse call system and remain hanging at that location with limited end user interaction. End user interactions would be limited to set up, installation, battery replacement, and intermittent troubleshooting if needed. There are no physical buttons or end use interface activities with this specific wireless adapter device.

Due to the intended clinical use of this device and its limited handling during use, this warning label shall be able to last the life of the product (4-year useful life) with a low risk to the warning label (where the FCC ID is located) being removed.




*Figure 1: 8282WL warning label and FCC ID location permanently located on the device.*

The 8282WL warning label, PN L9324C is made of the same label stock as other labels attached to Posey devices with the same intended use, use environment and end user interactions. These devices include the 8282 and the 8282SL Nurse Call Cables, which utilize the same cord diameter and material as the 8282WL. See Figure 2.

EFFECTIVE



*Figure 2. Images of 8282SL SKU – representative equivalent product*

A complaint review was evaluated for the calendar period from 2013 to 2021 and there are no complaints related to the warning label being removed, falling off, or becoming illegible.

**Conclusion:** The FCC ID on the 8282WL warning label is able to last the life of the product.

EFFECTIVE

**Signature Manifest****Document Number:** Memo-0039**Revision:** 00**Title:** Sitter On Cue PRO Wireless Nurse Call Adapter FCC ID number location memo**Effective Date:** 12 Aug 2021

All dates and times are in Central Time.

**Sitter On Cue PRO Wireless Nurse Call Adapter FCC ID number location memo****Edit / Collaborate**

| Name/Signature                | Title                        | Date                     | Meaning/Reason |
|-------------------------------|------------------------------|--------------------------|----------------|
| Samantha McCarthy (SMCCARTHY) | Sr Prod Development Engineer | 12 Aug 2021, 12:03:36 PM | Complete       |

**Doc Control QC Step**

| Name/Signature        | Title       | Date                     | Meaning/Reason |
|-----------------------|-------------|--------------------------|----------------|
| Mary Baccia (MBACCIA) | Doc Control | 12 Aug 2021, 12:11:51 PM | Complete       |

**Content Approval**

| Name/Signature                | Title                        | Date                     | Meaning/Reason |
|-------------------------------|------------------------------|--------------------------|----------------|
| Samantha McCarthy (SMCCARTHY) | Sr Prod Development Engineer | 12 Aug 2021, 12:38:20 PM | Approved       |
| Joseph Samz (JSAMZ)           | Quality Engineer             | 12 Aug 2021, 01:36:38 PM | Approved       |

**Final Release**

| Name/Signature                | Title                        | Date                     | Meaning/Reason |
|-------------------------------|------------------------------|--------------------------|----------------|
| Samantha McCarthy (SMCCARTHY) | Sr Prod Development Engineer | 12 Aug 2021, 01:49:58 PM | Approved       |