

FCC statement
location

Label location



Importer:

URIT

3A-800002

PRODUCT: Ultra Analyzer
MODEL: UC-300
POWER: Two AA alkaline
Batteries (not included)
FCC ID: B0442829

CE

RoHS

WEEE




REACH

URIT Medical Electronics Co., Ltd. Made in China
Address: No. 307 Information Industry Building, High-Tech Zone,
Guilin, Guangxi 541004, P.R. China
Guangzhou 510660, P.R. China
Singapore 119 001, S. Island, S.A.

KWtriO

No.20G-50

序号	时间	更改内容	更改人	标记	数量



技术要求：
黑色：C0 M0 Y0 K100
产品编号(SN) 从00010开始,对应条码形式为code-128
XXXX-XX 为生产日期年月， XXXX-XX 为使用期限年月。

材料为：铝箔不干胶。

耐久性要求：
用布沾上酒精，用手不加过分压力地擦拭30秒，标签内容仍应当清晰可辨，且不得出现松脱或卷边。

保密：此图及其全部知识产权（含著作权）归桂林优利特医疗电子有限公司所有。未经桂林优利特医疗电子有限公司预先书面许可，严禁出于任何目的，对此图的全部或者部分内容（包括但不限于图中信息、数据、运算结果等）泄漏、使用、拷贝或者复制。

拟 制		UC32A_英文后标签	代号：UCP050000813 图号：URT32A-05-008		
审 核			版 本	第 1 张	共 1 张
工 艺			V1.01		
批 准					

48mm

36mm

FCC statement: This device complies with part 15 of the FCC Rules. Operation is subject to the following two conditions:

(1) This device may not cause harmful interference, and

(2) This device must accept any interference received, including interference that may cause undesired operation.

Importer: