

Company name:URIT Medical Electronic Co., Ltd.
Address:No.D-07 Information Industry District, High-Tech Zone, Guilin, Guangxi 541004, P.R.China

Model difference declaration

Date (2021-06-09)

TUV SUD BABT UNLIMITED
Octagon House, Segensworth North,
Fareham, Hampshire, PO15 5RL,
United Kingdom

We,
(URIT Medical Electronic Co., Ltd.
Address:No.D-07 Information Industry District, High-Tech Zone, Guilin, Guangxi 541004,
P.R.China),

Here declare

Product: Urine analyzer
Model: UC-32A, UC-32B, UC-32A Vet, UC-32B Vet
FCC ID: 2AZS3UYUC32AB

The two models UC-32A, UC-32A Vet are identical in circuit design, PCB layout and components used but only different in model name.

The two models UC-32B, UC-32B Vet are identical in circuit design, PCB layout and components used but only different in model name.

The two models UC-32A, UC-32B are identical in circuit design, PCB layout but different in model name and components used.

All models used same Bluetooth module.

The difference between UC-32A and UC-32B in PCBA is that UC-32B has lithium battery charging management circuit and soft reset circuit, while UC-32A does not. In addition, UC-32A uses entity keys and UC-32B uses touch keys.

Sincerely,

Company Officer:Ma dejia

Ma dejia

Telephone Number:+86 18877380717

Email:abc1085@163.com