



## 4 MEASUREMENT METHOD

The IEC/IEEE 62209-1528 and FCC KDB865664 D01 standards provide requirements for reference dipoles used for system validation measurements. The following measurements were performed to verify that the product complies with the fore mentioned standards.

### 4.1 RETURN LOSS REQUIREMENTS

The dipole used for SAR system validation measurements and checks must have a return loss of -20 dB or better. The return loss measurement shall be performed against a liquid filled flat phantom, with the phantom constructed as outlined in the fore mentioned standards. A direct method is used with a network analyser and its calibration kit, both with a valid ISO17025 calibration.

### 4.2 MECHANICAL REQUIREMENTS

The IEC/IEEE 62209-1528 and FCC KDB865664 D01 standards specify the mechanical components and dimensions of the validation dipoles, with the dimension's frequency and phantom shell thickness dependent. The COMOSAR test bench employs a 2 mm phantom shell thickness therefore the dipoles sold for use with the COMOSAR test bench comply with the requirements set forth for a 2 mm phantom shell thickness. A direct method is used with a ISO17025 calibrated caliper.

## 5 MEASUREMENT UNCERTAINTY

All uncertainties listed below represent an expanded uncertainty expressed at approximately the 95% confidence level using a coverage factor of k=2, traceable to the Internationally Accepted Guides to Measurement Uncertainty.

### 5.1 RETURN LOSS

The following uncertainties apply to the return loss measurement:

| Frequency band | Expanded Uncertainty on Return Loss |
|----------------|-------------------------------------|
| 400-6000MHz    | 0.08 LIN                            |

### 5.2 DIMENSION MEASUREMENT

The following uncertainties apply to the dimension measurements:

| Length (mm) | Expanded Uncertainty on Length |
|-------------|--------------------------------|
| 0 - 300     | 0.20 mm                        |
| 300 - 450   | 0.44 mm                        |

### 5.3 VALIDATION MEASUREMENT

The guidelines outlined in the IEC/IEEE 62209-1528 and FCC KDB865664 D01 standards were followed to generate the measurement uncertainty for validation measurements.



**SAR REFERENCE DIPOLE CALIBRATION REPORT**

Ref: ACR.329.15.21.BES.A

| Scan Volume | Expanded Uncertainty |
|-------------|----------------------|
| 1 g         | 19 % (SAR)           |
| 10 g        | 19 % (SAR)           |


## 6 CALIBRATION MEASUREMENT RESULTS

### 6.1 RETURN LOSS AND IMPEDANCE IN HEAD LIQUID



| Frequency (MHz) | Return Loss (dB) | Requirement (dB) | Impedance                   |
|-----------------|------------------|------------------|-----------------------------|
| 2450            | -21.77           | -20              | $49.1 \Omega + 8.1 j\Omega$ |

### 6.2 RETURN LOSS AND IMPEDANCE IN BODY LIQUID



| Frequency (MHz) | Return Loss (dB) | Requirement (dB) | Impedance                   |
|-----------------|------------------|------------------|-----------------------------|
| 2450            | -21.55           | -20              | $54.7 \Omega + 6.8 j\Omega$ |

Page: 6/13

**Template\_ACR.DDD.N.YY.MVGB.ISSUE\_SAR Reference Dipole vJ**

This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.


**SAR REFERENCE DIPOLE CALIBRATION REPORT**

Ref: ACR.329.15.21.BES.A

### 6.3 MECHANICAL DIMENSIONS

| Frequency MHz | L mm             |          | h mm             |          | d mm            |          |
|---------------|------------------|----------|------------------|----------|-----------------|----------|
|               | required         | measured | required         | measured | required        | measured |
| 300           | 420.0 $\pm 1$ %. |          | 250.0 $\pm 1$ %. |          | 6.35 $\pm 1$ %. |          |
| 450           | 290.0 $\pm 1$ %. |          | 166.7 $\pm 1$ %. |          | 6.35 $\pm 1$ %. |          |
| 750           | 176.0 $\pm 1$ %. |          | 100.0 $\pm 1$ %. |          | 6.35 $\pm 1$ %. |          |
| 835           | 161.0 $\pm 1$ %. |          | 89.8 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 900           | 149.0 $\pm 1$ %. |          | 83.3 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 1450          | 89.1 $\pm 1$ %.  |          | 51.7 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 1500          | 86.2 $\pm 1$ %.  |          | 50.0 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 1640          | 79.0 $\pm 1$ %.  |          | 45.7 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 1750          | 75.2 $\pm 1$ %.  |          | 42.9 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 1800          | 72.0 $\pm 1$ %.  |          | 41.7 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 1900          | 68.0 $\pm 1$ %.  |          | 39.5 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 1950          | 66.3 $\pm 1$ %.  |          | 38.5 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 2000          | 64.5 $\pm 1$ %.  |          | 37.5 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 2100          | 61.0 $\pm 1$ %.  |          | 35.7 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 2300          | 55.5 $\pm 1$ %.  |          | 32.6 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 2450          | 51.5 $\pm 1$ %.  | 51.37    | 30.4 $\pm 1$ %.  | 30.45    | 3.6 $\pm 1$ %.  | 3.60     |
| 2600          | 48.5 $\pm 1$ %.  |          | 28.8 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 3000          | 41.5 $\pm 1$ %.  |          | 25.0 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 3300          | -                |          | -                |          | -               |          |
| 3500          | 37.0 $\pm 1$ %.  |          | 26.4 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 3700          | 34.7 $\pm 1$ %.  |          | 26.4 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 3900          | -                |          | -                |          | -               |          |
| 4200          | -                |          | -                |          | -               |          |
| 4600          | -                |          | -                |          | -               |          |
| 4900          | -                |          | -                |          | -               |          |

## 7 VALIDATION MEASUREMENT

The IEC/IEEE 62209-1528 and FCC KDB865664 D01 standards state that the system validation measurements must be performed using a reference dipole meeting the fore mentioned return loss and mechanical dimension requirements. The validation measurement must be performed against a liquid filled flat phantom, with the phantom constructed as outlined in the fore mentioned standards. Per the standards, the dipole shall be positioned below the bottom of the phantom, with the dipole length centered and parallel to the longest dimension of the flat phantom, with the top surface of the dipole at the described distance from the bottom surface of the phantom.

Page: 7/13

**Template\_ACR.DDD.N.YY.MVGB.ISSUE\_SAR Reference Dipole vJ**

*This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.*


**SAR REFERENCE DIPOLE CALIBRATION REPORT**

Ref: ACR.329.15.21.BES.A

**7.1 HEAD LIQUID MEASUREMENT**

| Frequency<br>MHz | Relative permittivity ( $\epsilon_r'$ ) |          | Conductivity ( $\sigma$ ) S/m |          |
|------------------|-----------------------------------------|----------|-------------------------------|----------|
|                  | required                                | measured | required                      | measured |
| 300              | 45.3 $\pm 10$ %                         |          | 0.87 $\pm 10$ %               |          |
| 450              | 43.5 $\pm 10$ %                         |          | 0.87 $\pm 10$ %               |          |
| 750              | 41.9 $\pm 10$ %                         |          | 0.89 $\pm 10$ %               |          |
| 835              | 41.5 $\pm 10$ %                         |          | 0.90 $\pm 10$ %               |          |
| 900              | 41.5 $\pm 10$ %                         |          | 0.97 $\pm 10$ %               |          |
| 1450             | 40.5 $\pm 10$ %                         |          | 1.20 $\pm 10$ %               |          |
| 1500             | 40.4 $\pm 10$ %                         |          | 1.23 $\pm 10$ %               |          |
| 1640             | 40.2 $\pm 10$ %                         |          | 1.31 $\pm 10$ %               |          |
| 1750             | 40.1 $\pm 10$ %                         |          | 1.37 $\pm 10$ %               |          |
| 1800             | 40.0 $\pm 10$ %                         |          | 1.40 $\pm 10$ %               |          |
| 1900             | 40.0 $\pm 10$ %                         |          | 1.40 $\pm 10$ %               |          |
| 1950             | 40.0 $\pm 10$ %                         |          | 1.40 $\pm 10$ %               |          |
| 2000             | 40.0 $\pm 10$ %                         |          | 1.40 $\pm 10$ %               |          |
| 2100             | 39.8 $\pm 10$ %                         |          | 1.49 $\pm 10$ %               |          |
| 2300             | 39.5 $\pm 10$ %                         |          | 1.67 $\pm 10$ %               |          |
| 2450             | 39.2 $\pm 10$ %                         | 36.4     | 1.80 $\pm 10$ %               | 1.96     |
| 2600             | 39.0 $\pm 10$ %                         |          | 1.96 $\pm 10$ %               |          |
| 3000             | 38.5 $\pm 10$ %                         |          | 2.40 $\pm 10$ %               |          |
| 3300             | 38.2 $\pm 10$ %                         |          | 2.71 $\pm 10$ %               |          |
| 3500             | 37.9 $\pm 10$ %                         |          | 2.91 $\pm 10$ %               |          |
| 3700             | 37.7 $\pm 10$ %                         |          | 3.12 $\pm 10$ %               |          |
| 3900             | 37.5 $\pm 10$ %                         |          | 3.32 $\pm 10$ %               |          |
| 4200             | 37.1 $\pm 10$ %                         |          | 3.63 $\pm 10$ %               |          |
| 4600             | 36.7 $\pm 10$ %                         |          | 4.04 $\pm 10$ %               |          |
| 4900             | 36.3 $\pm 10$ %                         |          | 4.35 $\pm 10$ %               |          |

**7.2 SAR MEASUREMENT RESULT WITH HEAD LIQUID**

The IEC/IEEE 62209-1528 and FCC KDB865664 D01 standards state that the system validation measurements should produce the SAR values shown below (for phantom thickness of 2 mm), within the uncertainty for the system validation. All SAR values are normalized to 1 W forward power. In bracket, the measured SAR is given with the used input power.

Page: 8/13

**Template ACR.DDD.N.YY.MVGB.ISSUE SAR Reference Dipole vJ**

*This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.*


**SAR REFERENCE DIPOLE CALIBRATION REPORT**

Ref: ACR.329.15.21.BES.A

|                                           |                                                             |
|-------------------------------------------|-------------------------------------------------------------|
| Software                                  | OPENSAR V5                                                  |
| Phantom                                   | SN 13/09 SAM68                                              |
| Probe                                     | SN 41/18 EPGO333                                            |
| Liquid                                    | Head Liquid Values: $\epsilon_{ps}'$ : 36.4 $\sigma$ : 1.96 |
| Distance between dipole center and liquid | 10.0 mm                                                     |
| Area scan resolution                      | $dx=8mm/dy=8mm$                                             |
| Zoon Scan Resolution                      | $dx=5mm/dy=5mm/dz=5mm$                                      |
| Frequency                                 | 2450 MHz                                                    |
| Input power                               | 20 dBm                                                      |
| Liquid Temperature                        | 20 +/- 1 °C                                                 |
| Lab Temperature                           | 20 +/- 1 °C                                                 |
| Lab Humidity                              | 30-70 %                                                     |

| Frequency<br>MHz | 1 g SAR (W/kg/W) |              | 10 g SAR (W/kg/W) |              |
|------------------|------------------|--------------|-------------------|--------------|
|                  | required         | measured     | required          | measured     |
| 300              | 2.85             |              | 1.94              |              |
| 450              | 4.58             |              | 3.06              |              |
| 750              | 8.49             |              | 5.55              |              |
| 835              | 9.56             |              | 6.22              |              |
| 900              | 10.9             |              | 6.99              |              |
| 1450             | 29               |              | 16                |              |
| 1500             | 30.5             |              | 16.8              |              |
| 1640             | 34.2             |              | 18.4              |              |
| 1750             | 36.4             |              | 19.3              |              |
| 1800             | 38.4             |              | 20.1              |              |
| 1900             | 39.7             |              | 20.5              |              |
| 1950             | 40.5             |              | 20.9              |              |
| 2000             | 41.1             |              | 21.1              |              |
| 2100             | 43.6             |              | 21.9              |              |
| 2300             | 48.7             |              | 23.3              |              |
| 2450             | 52.4             | 55.16 (5.52) | 24                | 24.15 (2.41) |
| 2600             | 55.3             |              | 24.6              |              |
| 3000             | 63.8             |              | 25.7              |              |
| 3300             | -                |              | -                 |              |
| 3500             | 67.1             |              | 25                |              |
| 3700             | 67.4             |              | 24.2              |              |
| 3900             | -                |              | -                 |              |
| 4200             | -                |              | -                 |              |
| 4600             | -                |              | -                 |              |
| 4900             | -                |              | -                 |              |


Page: 9/13

**Template ACR.DDD.N.YY.MVGB.ISSUE SAR Reference Dipole vJ**

*This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.*


**SAR REFERENCE DIPOLE CALIBRATION REPORT**

Ref: ACR.329.15.21.BES.A



Page: 10/13

**Template\_ACR.DDD.N.YY.MVGB.ISSUE\_SAR Reference Dipole vJ**

*This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.*


**SAR REFERENCE DIPOLE CALIBRATION REPORT**

Ref: ACR.329.15.21.BES.A

**7.3 BODY LIQUID MEASUREMENT**

| Frequency<br>MHz | Relative permittivity ( $\epsilon_r'$ ) |          | Conductivity ( $\sigma$ ) S/m |          |
|------------------|-----------------------------------------|----------|-------------------------------|----------|
|                  | required                                | measured | required                      | measured |
| 150              | 61.9 $\pm 10$ %                         |          | 0.80 $\pm 10$ %               |          |
| 300              | 58.2 $\pm 10$ %                         |          | 0.92 $\pm 10$ %               |          |
| 450              | 56.7 $\pm 10$ %                         |          | 0.94 $\pm 10$ %               |          |
| 750              | 55.5 $\pm 10$ %                         |          | 0.96 $\pm 10$ %               |          |
| 835              | 55.2 $\pm 10$ %                         |          | 0.97 $\pm 10$ %               |          |
| 900              | 55.0 $\pm 10$ %                         |          | 1.05 $\pm 10$ %               |          |
| 915              | 55.0 $\pm 10$ %                         |          | 1.06 $\pm 10$ %               |          |
| 1450             | 54.0 $\pm 10$ %                         |          | 1.30 $\pm 10$ %               |          |
| 1610             | 53.8 $\pm 10$ %                         |          | 1.40 $\pm 10$ %               |          |
| 1800             | 53.3 $\pm 10$ %                         |          | 1.52 $\pm 10$ %               |          |
| 1900             | 53.3 $\pm 10$ %                         |          | 1.52 $\pm 10$ %               |          |
| 2000             | 53.3 $\pm 10$ %                         |          | 1.52 $\pm 10$ %               |          |
| 2100             | 53.2 $\pm 10$ %                         |          | 1.62 $\pm 10$ %               |          |
| 2300             | 52.9 $\pm 10$ %                         |          | 1.81 $\pm 10$ %               |          |
| 2450             | 52.7 $\pm 10$ %                         | 53.4     | 1.95 $\pm 10$ %               | 2.14     |
| 2600             | 52.5 $\pm 10$ %                         |          | 2.16 $\pm 10$ %               |          |
| 3000             | 52.0 $\pm 10$ %                         |          | 2.73 $\pm 10$ %               |          |
| 3300             | 51.6 $\pm 10$ %                         |          | 3.08 $\pm 10$ %               |          |
| 3500             | 51.3 $\pm 10$ %                         |          | 3.31 $\pm 10$ %               |          |
| 3700             | 51.0 $\pm 10$ %                         |          | 3.55 $\pm 10$ %               |          |
| 3900             | 50.8 $\pm 10$ %                         |          | 3.78 $\pm 10$ %               |          |
| 4200             | 50.4 $\pm 10$ %                         |          | 4.13 $\pm 10$ %               |          |
| 4600             | 49.8 $\pm 10$ %                         |          | 4.60 $\pm 10$ %               |          |
| 4900             | 49.4 $\pm 10$ %                         |          | 4.95 $\pm 10$ %               |          |
| 5200             | 49.0 $\pm 10$ %                         |          | 5.30 $\pm 10$ %               |          |
| 5300             | 48.9 $\pm 10$ %                         |          | 5.42 $\pm 10$ %               |          |
| 5400             | 48.7 $\pm 10$ %                         |          | 5.53 $\pm 10$ %               |          |
| 5500             | 48.6 $\pm 10$ %                         |          | 5.65 $\pm 10$ %               |          |
| 5600             | 48.5 $\pm 10$ %                         |          | 5.77 $\pm 10$ %               |          |
| 5800             | 48.2 $\pm 10$ %                         |          | 6.00 $\pm 10$ %               |          |

Page: 11/13

**Template ACR.DDD.N.YY.MVGB.ISSUE SAR Reference Dipole vJ**

*This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.*



**SAR REFERENCE DIPOLE CALIBRATION REPORT**

Ref: ACR.329.15.21.BES.A

**7.4 SAR MEASUREMENT RESULT WITH BODY LIQUID**

|                                           |                                                             |
|-------------------------------------------|-------------------------------------------------------------|
| Software                                  | OPENSAR V5                                                  |
| Phantom                                   | SN 13/09 SAM68                                              |
| Probe                                     | SN 41/18 EPG0333                                            |
| Liquid                                    | Body Liquid Values: $\epsilon_{ps}'$ : 53.4 $\sigma$ : 2.14 |
| Distance between dipole center and liquid | 10.0 mm                                                     |
| Area scan resolution                      | $dx=8mm/dy=8mm$                                             |
| Zoon Scan Resolution                      | $dx=5mm/dy=5mm/dz=5mm$                                      |
| Frequency                                 | 2450 MHz                                                    |
| Input power                               | 20 dBm                                                      |
| Liquid Temperature                        | 20 +/- 1 °C                                                 |
| Lab Temperature                           | 20 +/- 1 °C                                                 |
| Lab Humidity                              | 30-70 %                                                     |

| Frequency<br>MHz | 1 g SAR (W/kg/W) | 10 g SAR (W/kg/W) |
|------------------|------------------|-------------------|
|                  | measured         | measured          |
| 2450             | 52.28 (5.23)     | 22.68 (2.27)      |



Page: 12/13

**Template ACR.DDD.N.YY.MVGB.ISSUE SAR Reference Dipole vJ**

This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.



## 8 LIST OF EQUIPMENT

| Equipment Summary Sheet            |                         |                    |                                               |                                               |
|------------------------------------|-------------------------|--------------------|-----------------------------------------------|-----------------------------------------------|
| Equipment Description              | Manufacturer / Model    | Identification No. | Current Calibration Date                      | Next Calibration Date                         |
| SAM Phantom                        | MVG                     | SN 13/09 SAM68     | Validated. No cal required.                   | Validated. No cal required.                   |
| COMOSAR Test Bench                 | Version 3               | NA                 | Validated. No cal required.                   | Validated. No cal required.                   |
| Network Analyzer                   | Rohde & Schwarz ZVM     | 100203             | 08/2021                                       | 08/2024                                       |
| Network Analyzer                   | Agilent 8753ES          | MY40003210         | 10/2022                                       | 10/2025                                       |
| Network Analyzer – Calibration kit | Rohde & Schwarz ZV-Z235 | 101223             | 05/2012                                       | 05/2025                                       |
| Network Analyzer – Calibration kit | HP 85033D               | 3423A08186         | 06/2021                                       | 06/2027                                       |
| Calipers                           | Mitutoyo                | SN 0009732         | 10/2022                                       | 10/2025                                       |
| Reference Probe                    | MVG                     | SN 41/18 EPG0333   | 10/2022                                       | 10/2025                                       |
| Multimeter                         | Keithley 2000           | 1160271            | 02/2023                                       | 02/2026                                       |
| Signal Generator                   | Rohde & Schwarz SMB     | 106589             | 04/2022                                       | 04/2025                                       |
| Amplifier                          | MVG                     | MODU-023-C-0002    | Characterized prior to test. No cal required. | Characterized prior to test. No cal required. |
| Power Meter                        | NI-USB 5680             | 170100013          | 06/2021                                       | 06/2024                                       |
| Power Meter                        | Rohde & Schwarz NRVD    | 832839-056         | 11/2022                                       | 11/2025                                       |
| Directional Coupler                | Krytar 158020           | 131467             | Characterized prior to test. No cal required. | Characterized prior to test. No cal required. |
| Temperature / Humidity Sensor      | Testo 184 H1            | 44225320           | 06/2021                                       | 06/2024                                       |

Page: 13/13

**Template\_ACR.DDD.N.YY.MVGB.ISSUE\_SAR Reference Dipole vJ**

*This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.*



## SAR Reference Dipole Calibration Report

Ref : ACR.329.16.21.BES.A

### **SHENZHEN BCTC TECHNOLOGY CO., LTD.**

**1 ~2/ F, NO. B FACTORY BUILDING, PENGZHOU  
INDUSTRIAL PARK, FUYUAN 1ST ROAD,  
TANGWEI COMMUNITY, FUHAI STREET, BAO'AN  
DISTRICT, SHENZHEN, GUANGDONG, CHINA  
MVG COMOSAR REFERENCE DIPOLE**

**FREQUENCY: 2600 MHZ**

**SERIAL NO.: SN 47/21 DIP 2G600-628**

**Calibrated at MVG**

**Z.I. de la pointe du diable**

**Technopôle Brest Iroise – 295 avenue Alexis de Rochon**

**29280 PLOUZANE - FRANCE**

**Calibration date: 11/25/2021**



Accreditations #2-6789 and #2-6814  
Scope available on [www.cofrac.fr](http://www.cofrac.fr)

**The use of the Cofrac brand and the accreditation references is prohibited from any reproduction.**

#### *Summary:*

This document presents the method and results from an accredited SAR reference dipole calibration performed in MVG using the COMOSAR test bench. All calibration results are traceable to national metrology institutions.


**SAR REFERENCE DIPOLE CALIBRATION REPORT**

Ref: ACR.329.16.21.BES.A

|                      | <i>Name</i>  | <i>Function</i>     | <i>Date</i> | <i>Signature</i>                                                                                                      |
|----------------------|--------------|---------------------|-------------|-----------------------------------------------------------------------------------------------------------------------|
| <i>Prepared by :</i> | Jérôme Luc   | Technical Manager   | 11/25/2021  |                                    |
| <i>Checked by :</i>  | Jérôme Luc   | Technical Manager   | 11/25/2021  |                                    |
| <i>Approved by :</i> | Yann Toutain | Laboratory Director | 11/25/2021  | <br>2021.11.25<br>11:57:32 +01'00' |

|                       | <i>Customer Name</i>                     |
|-----------------------|------------------------------------------|
| <i>Distribution :</i> | Shenzhen BCTC<br>Technology Co.,<br>Ltd. |

| <i>Issue</i> | <i>Name</i> | <i>Date</i> | <i>Modifications</i> |
|--------------|-------------|-------------|----------------------|
| A            | Jérôme Luc  | 11/25/2021  | Initial release      |
|              |             |             |                      |
|              |             |             |                      |
|              |             |             |                      |

Page: 2/13

**Template ACR.DDD.N.YY.MVGB.ISSUE SAR Reference Dipole vJ**

*This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.*



## TABLE OF CONTENTS

|     |                                                |    |
|-----|------------------------------------------------|----|
| 1   | Introduction.....                              | 4  |
| 2   | Device Under Test .....                        | 4  |
| 3   | Product Description .....                      | 4  |
| 3.1 | General Information .....                      | 4  |
| 4   | Measurement Method .....                       | 5  |
| 4.1 | Return Loss Requirements .....                 | 5  |
| 4.2 | Mechanical Requirements .....                  | 5  |
| 5   | Measurement Uncertainty .....                  | 5  |
| 5.1 | Return Loss .....                              | 5  |
| 5.2 | Dimension Measurement .....                    | 5  |
| 5.3 | Validation Measurement .....                   | 5  |
| 6   | Calibration Measurement Results .....          | 6  |
| 6.1 | Return Loss and Impedance In Head Liquid ..... | 6  |
| 6.2 | Return Loss and Impedance In Body Liquid ..... | 6  |
| 6.3 | Mechanical Dimensions .....                    | 7  |
| 7   | Validation measurement .....                   | 7  |
| 7.1 | Head Liquid Measurement .....                  | 8  |
| 7.2 | SAR Measurement Result With Head Liquid .....  | 8  |
| 7.3 | Body Liquid Measurement .....                  | 11 |
| 7.4 | SAR Measurement Result With Body Liquid .....  | 12 |
| 8   | List of Equipment .....                        | 13 |



## 1 INTRODUCTION

This document contains a summary of the requirements set forth by the IEC/IEEE 62209-1528 and FCC KDB865664 D01 standards for reference dipoles used for SAR measurement system validations and the measurements that were performed to verify that the product complies with the fore mentioned standards.

## 2 DEVICE UNDER TEST

| Device Under Test              |                                   |
|--------------------------------|-----------------------------------|
| Device Type                    | COMOSAR 2600 MHz REFERENCE DIPOLE |
| Manufacturer                   | MVG                               |
| Model                          | SID2600                           |
| Serial Number                  | SN 47/21 DIP 2G600-628            |
| Product Condition (new / used) | New                               |

## 3 PRODUCT DESCRIPTION

### 3.1 GENERAL INFORMATION

MVG's COMOSAR Validation Dipoles are built in accordance to the IEC/IEEE 62209-1528 and FCC KDB865664 D01 standards. The product is designed for use with the COMOSAR test bench only.



**Figure 1 – MVG COMOSAR Validation Dipole**



## 4 MEASUREMENT METHOD

The IEC/IEEE 62209-1528 and FCC KDB865664 D01 standards provide requirements for reference dipoles used for system validation measurements. The following measurements were performed to verify that the product complies with the fore mentioned standards.

### 4.1 RETURN LOSS REQUIREMENTS

The dipole used for SAR system validation measurements and checks must have a return loss of -20 dB or better. The return loss measurement shall be performed against a liquid filled flat phantom, with the phantom constructed as outlined in the fore mentioned standards. A direct method is used with a network analyser and its calibration kit, both with a valid ISO17025 calibration.

### 4.2 MECHANICAL REQUIREMENTS

The IEC/IEEE 62209-1528 and FCC KDB865664 D01 standards specify the mechanical components and dimensions of the validation dipoles, with the dimension's frequency and phantom shell thickness dependent. The COMOSAR test bench employs a 2 mm phantom shell thickness therefore the dipoles sold for use with the COMOSAR test bench comply with the requirements set forth for a 2 mm phantom shell thickness. A direct method is used with a ISO17025 calibrated caliper.

## 5 MEASUREMENT UNCERTAINTY

All uncertainties listed below represent an expanded uncertainty expressed at approximately the 95% confidence level using a coverage factor of k=2, traceable to the Internationally Accepted Guides to Measurement Uncertainty.

### 5.1 RETURN LOSS

The following uncertainties apply to the return loss measurement:

| Frequency band | Expanded Uncertainty on Return Loss |
|----------------|-------------------------------------|
| 400-6000MHz    | 0.08 LIN                            |

### 5.2 DIMENSION MEASUREMENT

The following uncertainties apply to the dimension measurements:

| Length (mm) | Expanded Uncertainty on Length |
|-------------|--------------------------------|
| 0 - 300     | 0.20 mm                        |
| 300 - 450   | 0.44 mm                        |

### 5.3 VALIDATION MEASUREMENT

The guidelines outlined in the IEC/IEEE 62209-1528 and FCC KDB865664 D01 standards were followed to generate the measurement uncertainty for validation measurements.



**SAR REFERENCE DIPOLE CALIBRATION REPORT**

Ref: ACR.329.16.21.BES.A

| Scan Volume | Expanded Uncertainty |
|-------------|----------------------|
| 1 g         | 19 % (SAR)           |
| 10 g        | 19 % (SAR)           |


## 6 CALIBRATION MEASUREMENT RESULTS

### 6.1 RETURN LOSS AND IMPEDANCE IN HEAD LIQUID



| Frequency (MHz) | Return Loss (dB) | Requirement (dB) | Impedance                      |
|-----------------|------------------|------------------|--------------------------------|
| 2600            | -28.07           | -20              | 52.8 $\Omega$ - 2.8 j $\Omega$ |

### 6.2 RETURN LOSS AND IMPEDANCE IN BODY LIQUID



| Frequency (MHz) | Return Loss (dB) | Requirement (dB) | Impedance                      |
|-----------------|------------------|------------------|--------------------------------|
| 2600            | -26.44           | -20              | 46.7 $\Omega$ - 3.4 j $\Omega$ |

Page: 6/13

Template ACR.DDD.N.YY.MVGB.ISSUE SAR Reference Dipole vJ

This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.


**SAR REFERENCE DIPOLE CALIBRATION REPORT**

Ref: ACR.329.16.21.BES.A

### 6.3 MECHANICAL DIMENSIONS

| Frequency MHz | L mm             |          | h mm             |          | d mm            |          |
|---------------|------------------|----------|------------------|----------|-----------------|----------|
|               | required         | measured | required         | measured | required        | measured |
| 300           | 420.0 $\pm 1$ %. |          | 250.0 $\pm 1$ %. |          | 6.35 $\pm 1$ %. |          |
| 450           | 290.0 $\pm 1$ %. |          | 166.7 $\pm 1$ %. |          | 6.35 $\pm 1$ %. |          |
| 750           | 176.0 $\pm 1$ %. |          | 100.0 $\pm 1$ %. |          | 6.35 $\pm 1$ %. |          |
| 835           | 161.0 $\pm 1$ %. |          | 89.8 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 900           | 149.0 $\pm 1$ %. |          | 83.3 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 1450          | 89.1 $\pm 1$ %.  |          | 51.7 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 1500          | 86.2 $\pm 1$ %.  |          | 50.0 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 1640          | 79.0 $\pm 1$ %.  |          | 45.7 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 1750          | 75.2 $\pm 1$ %.  |          | 42.9 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 1800          | 72.0 $\pm 1$ %.  |          | 41.7 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 1900          | 68.0 $\pm 1$ %.  |          | 39.5 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 1950          | 66.3 $\pm 1$ %.  |          | 38.5 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 2000          | 64.5 $\pm 1$ %.  |          | 37.5 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 2100          | 61.0 $\pm 1$ %.  |          | 35.7 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 2300          | 55.5 $\pm 1$ %.  |          | 32.6 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 2450          | 51.5 $\pm 1$ %.  |          | 30.4 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 2600          | 48.5 $\pm 1$ %.  | 48.19    | 28.8 $\pm 1$ %.  | 28.80    | 3.6 $\pm 1$ %.  | 3.59     |
| 3000          | 41.5 $\pm 1$ %.  |          | 25.0 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 3300          | -                |          | -                |          | -               |          |
| 3500          | 37.0 $\pm 1$ %.  |          | 26.4 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 3700          | 34.7 $\pm 1$ %.  |          | 26.4 $\pm 1$ %.  |          | 3.6 $\pm 1$ %.  |          |
| 3900          | -                |          | -                |          | -               |          |
| 4200          | -                |          | -                |          | -               |          |
| 4600          | -                |          | -                |          | -               |          |
| 4900          | -                |          | -                |          | -               |          |

## 7 VALIDATION MEASUREMENT

The IEC/IEEE 62209-1528 and FCC KDB865664 D01 standards state that the system validation measurements must be performed using a reference dipole meeting the fore mentioned return loss and mechanical dimension requirements. The validation measurement must be performed against a liquid filled flat phantom, with the phantom constructed as outlined in the fore mentioned standards. Per the standards, the dipole shall be positioned below the bottom of the phantom, with the dipole length centered and parallel to the longest dimension of the flat phantom, with the top surface of the dipole at the described distance from the bottom surface of the phantom.

Page: 7/13

**Template ACR.DDD.N.YY.MVGB.ISSUE SAR Reference Dipole vJ**

*This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.*


**SAR REFERENCE DIPOLE CALIBRATION REPORT**

Ref: ACR.329.16.21.BES.A

**7.1 HEAD LIQUID MEASUREMENT**

| Frequency<br>MHz | Relative permittivity ( $\epsilon_r'$ ) |          | Conductivity ( $\sigma$ ) S/m |          |
|------------------|-----------------------------------------|----------|-------------------------------|----------|
|                  | required                                | measured | required                      | measured |
| 300              | 45.3 $\pm 10$ %                         |          | 0.87 $\pm 10$ %               |          |
| 450              | 43.5 $\pm 10$ %                         |          | 0.87 $\pm 10$ %               |          |
| 750              | 41.9 $\pm 10$ %                         |          | 0.89 $\pm 10$ %               |          |
| 835              | 41.5 $\pm 10$ %                         |          | 0.90 $\pm 10$ %               |          |
| 900              | 41.5 $\pm 10$ %                         |          | 0.97 $\pm 10$ %               |          |
| 1450             | 40.5 $\pm 10$ %                         |          | 1.20 $\pm 10$ %               |          |
| 1500             | 40.4 $\pm 10$ %                         |          | 1.23 $\pm 10$ %               |          |
| 1640             | 40.2 $\pm 10$ %                         |          | 1.31 $\pm 10$ %               |          |
| 1750             | 40.1 $\pm 10$ %                         |          | 1.37 $\pm 10$ %               |          |
| 1800             | 40.0 $\pm 10$ %                         |          | 1.40 $\pm 10$ %               |          |
| 1900             | 40.0 $\pm 10$ %                         |          | 1.40 $\pm 10$ %               |          |
| 1950             | 40.0 $\pm 10$ %                         |          | 1.40 $\pm 10$ %               |          |
| 2000             | 40.0 $\pm 10$ %                         |          | 1.40 $\pm 10$ %               |          |
| 2100             | 39.8 $\pm 10$ %                         |          | 1.49 $\pm 10$ %               |          |
| 2300             | 39.5 $\pm 10$ %                         |          | 1.67 $\pm 10$ %               |          |
| 2450             | 39.2 $\pm 10$ %                         |          | 1.80 $\pm 10$ %               |          |
| 2600             | 39.0 $\pm 10$ %                         | 36.0     | 1.96 $\pm 10$ %               | 2.12     |
| 3000             | 38.5 $\pm 10$ %                         |          | 2.40 $\pm 10$ %               |          |
| 3300             | 38.2 $\pm 10$ %                         |          | 2.71 $\pm 10$ %               |          |
| 3500             | 37.9 $\pm 10$ %                         |          | 2.91 $\pm 10$ %               |          |
| 3700             | 37.7 $\pm 10$ %                         |          | 3.12 $\pm 10$ %               |          |
| 3900             | 37.5 $\pm 10$ %                         |          | 3.32 $\pm 10$ %               |          |
| 4200             | 37.1 $\pm 10$ %                         |          | 3.63 $\pm 10$ %               |          |
| 4600             | 36.7 $\pm 10$ %                         |          | 4.04 $\pm 10$ %               |          |
| 4900             | 36.3 $\pm 10$ %                         |          | 4.35 $\pm 10$ %               |          |

**7.2 SAR MEASUREMENT RESULT WITH HEAD LIQUID**

The IEC/IEEE 62209-1528 and FCC KDB865664 D01 standards state that the system validation measurements should produce the SAR values shown below (for phantom thickness of 2 mm), within the uncertainty for the system validation. All SAR values are normalized to 1 W forward power. In bracket, the measured SAR is given with the used input power.

Page: 8/13

**Template ACR.DDD.N.YY.MVGB.ISSUE SAR Reference Dipole vJ**

*This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.*