



AMETEK

Crystal
pressure



XP3i

Digital Test Gauge



peak

zero

units

3000 PSI

325698
7/16-20 MP



WARNING

DO NOT USE USB INTERFACE IN
HAZARDOUS LOCATIONS (U_m = 6V)



CE 0344



Ex ia IIC T3...T4 Ga  II 1G
Ta = -20°C to 50°C
ATEX: FTZU 23 ATEX 0015X
IECEx FTZU 23.0017X



SUBSTITUTION OF COMPONENTS
MAY IMPAIR INTRINSIC SAFETY

Ex ia
187869 Class I, Div 1, Grp A, B, C, D
T4/T3C
 Class I, Zone 0, AEx ia IIC T3C...T4 Ga
C US
CSA24CA80188655X

REPLACE 3 AA BATTERIES WITH
APPROVED TYPES IN NON-
HAZARDOUS LOCATIONS ONLY

Refer to Operation Manual
Reportez-vous au Mode d'emploi



R205-240004

FCC ID: 2BD5G001
IC: 31893-001

REFER TO MANUAL FOR
ENTITY PARAMETERS

Crystal Engineering Corporation
805.595.5477 • 800.444.1850
San Luis Obispo, CA U.S.A.

ametekcalibration.com

Made in U.S.A.









