

EMC TEST REPORT

Applicant Shenzhen D-Health Medical
Technology Co., Ltd.

FCC ID 2BD8CPM6100

Product Patient Monitor

Brand **BERRY®**

Model PM6100; PM6100A; AM6200

Report No. R2404A0388-E1V1

Issue Date June 6, 2024

Eurofins TA Technology (Shanghai) Co., Ltd. tested the above equipment in accordance with the requirements in **FCC Code CFR47 Part15B (2023)/ ANSI C63.4-2014**. The test results show that the equipment tested is capable of demonstrating compliance with the requirements as documented in this report.

Prepared by: Liu Wei

Approved by: Fan Guangchang

Eurofins TA Technology (Shanghai) Co., Ltd.

Building 3, No.145, Jintang Rd, Pudong Shanghai, P.R.China

TEL: +86-021-50791141/2/3

FAX: +86-021-50791141/2/3-8000

Table of Contents

1	Test Laboratory	5
1.1	Notes of the Test Report.....	5
1.2	Test Facility.....	5
1.3	Testing Location.....	5
2	General Description of Equipment Under Test	6
2.1	Applicant and Manufacturer Information	6
2.2	General Information	6
2.3	Applied Standards	8
2.4	Test Mode	9
3	Test Case Results	10
3.1	Radiated Emission.....	10
3.2	Conducted Emission.....	15
4	Uncertainty Measurement.....	18
5	Main Test Instruments	19
	ANNEX A: The EUT Appearance	20
	ANNEX B: Test Setup Photos	21

Version	Revision Description	Issue Date
Rev.0	Initial issue of report.	May 28, 2024
Rev.1	Update information in Page 6.	June 6, 2024
Note: This revised report (Report No.: R2404A0388-E1V1) supersedes and replaces the previously issued report (Report No.: R2404A0388-E1). Please discard or destroy the previously issued report and dispose of it accordingly.		

Summary of measurement results

Number	Test Case	Clause in FCC Rules	Conclusion
1	Radiated Emission	FCC Part15.109, ANSI C63.4-2014	PASS
2	Conducted Emission	FCC Part15.107, ANSI C63.4-2014	PASS
Date of Testing: April 18, 2024 ~ May 23, 2024			
Date of Sample Received: April 15, 2024			
Note: All indications of Pass/Fail in this report are opinions expressed by Eurofins TA Technology (Shanghai) Co., Ltd. based on interpretations and/or observations of test results. Measurement Uncertainties were not taken into account and are published for informational purposes only.			

1 Test Laboratory

1.1 Notes of the Test Report

This report shall not be reproduced in full or partial, without the written approval of **Eurofins TA Technology (Shanghai) Co., Ltd.** The results documented in this report apply only to the tested sample, under the conditions and modes of operation as described herein. Measurement Uncertainties were not taken into account and are published for informational purposes only. This report is written to support regulatory compliance of the applicable standards stated above.

1.2 Test Facility

FCC (Designation number: CN1179, Test Firm Registration Number: 446626)

Eurofins TA Technology (Shanghai) Co., Ltd. has been listed on the US Federal Communications Commission list of test facilities recognized to perform measurements.

A2LA (Certificate Number: 3857.01)

Eurofins TA Technology (Shanghai) Co., Ltd. has been listed by American Association for Laboratory Accreditation to perform measurement.

1.3 Testing Location

Company:	Eurofins TA Technology (Shanghai) Co., Ltd.
Address:	Building 3, No.145, Jintang Rd, Pudong Shanghai, P.R.China
City:	Shanghai
Post code:	201201
Country:	P. R. China
Contact:	Fan Guangchang
Telephone:	+86-021-50791141/2/3
Fax:	+86-021-50791141/2/3-8000
Website:	https://www.eurofins.com/electrical-and-electronics
E-mail:	Jack.Fan@cpt.eurofinscn.com

2 General Description of Equipment Under Test

2.1 Applicant and Manufacturer Information

Applicant	Shenzhen D-Health Medical Technology Co., Ltd.
Applicant address	Unit 501, No. 1 Juyin Technology Industrial Plant, Jihua Street, Longgang District, Shenzhen, China
Manufacturer	Shanghai Berry Electronic Tech Co., Ltd
Manufacturer address	Unit 104, 1st Floor, 7th Building, No.1188 Lianhang Road, Minhang District, Shanghai

2.2 General Information

EUT Description			
Device Type	Portable Device		
Model	PM6100; PM6100A; AM6200		
SN	M6000450		
HW Version	V3.3		
SW Version	PM6100; PM6100A: V2.53.00.00 AM6200: V1.55.00.00		
Power Rating	DC 3.7V from battery; DV 5V from Adapter		
Connecting I/O Port(s)	Please refer to the User's Manual.		
Antenna Type	Ceramic Chip Antenna		
Frequency	Band	Tx (MHz)	Rx (MHz)
	Bluetooth LE	2400 ~ 2483.5	2400 ~ 2483.5
EUT Accessory			
Battery	Manufacturer: Shanghai Berry Electronic Tech Co., Ltd Model: / DC 3.7V, 1800mAh		
Blood Pressur Cuff	Manufacturer: Shanghai Berry Electronic Tech Co., Ltd Model: MH002RC12HP1		
Auxiliary Test Equipment			
Adapter	Manufacturer: Xiaomi Communications Co., Ltd. Model: MDY-09-EQ Input: 100-240V~50/60Hz 0.3A Output: 5.0V ---2.0A		
Note: 1. The EUT is sent from the applicant to Eurofins TA and the information of the EUT is declared by the applicant.			

The differences between models are shown in the table below:

Model Item	PM6100	PM6100A	AM6200
SW Version	V2.53.00.00	V2.53.00.00	V1.55.00.00
SpO ₂ (Arterial oxygen saturation in the blood)	√	√	√
PR/HR (Heart Rate or Pulse Rate)	√	√	√
NIBP (non-invasive blood pressure)	√	√	√
TEMP (body temperature)	√	×	√
Communication (Bluetooth)	√	√	√
RESP (Respiration rate)	√	√	√
Appearance	The same		Different
Others	The same		
Note: This report only tests PM6100.			

2.3 Applied Standards

According to the specifications of the manufacturer, it must comply with the requirements of the following standards:

Test standards

FCC Code CFR47 Part15B (2023)

ANSI C63.4-2014

2.4 Test Mode

Test Mode	
Mode 1	Adapter + EUT + Auxiliary accessories + Working
Mode 2	Adapter + EUT + Auxiliary accessories + Standby
Mode 3	Battery powered + EUT + Auxiliary accessories + Working

Test Type	Test Mode	Worst Mode
Radiated Emission	Mode 1, 2, 3	Mode 3
Conducted Emission	Mode 1, 2	Mode 2
During the test, the preliminary test was performed in all modes, the test data of the worst-case condition was recorded in this report.		

3 Test Case Results

3.1 Radiated Emission

Ambient Condition

Temperature	Relative humidity
15°C ~ 35°C	30% ~ 60%

Methods of Measurement

The EUT is placed on a non-metallic table 0.8m above the horizontal metal reference ground plane. The distance between EUT and receive antenna should be 3 meters. During the test, the EUT was operating in its typical mode. The test method is according to ANSI C63.4-2014. Sweep the whole frequency band through the range from 30MHz to the 5th harmonic of the carrier. During the test, the height of receive antenna shall be moved from 1 to 4 meters, and the antenna shall be performed under horizontal and vertical polarization. The turn table shall be rotated from 0 to 360 degrees for detecting the maximum of radiated signal level.

The data of cable loss and antenna factor has been calibrated in full testing frequency range before the testing. During the test, the EUT is worked at maximum output power.

Set the spectrum analyzer in the following:

Below 1GHz:

RBW=100 kHz / VBW=300 kHz / Sweep=AUTO

Above 1GHz:


(a) PEAK Detector: RBW=1MHz / VBW=3MHz/ Sweep=AUTO

(b) AVERAGE Detector: RBW=1MHz / VBW=3MHz / Sweep=AUTO


The radiated emission was measured in the following position: EUT stand-up position (Z axis), lie-down position (X, Y axis). The worst emission was found in lie-down position (X axis) and the worst case was recorded.

Test Setup

Below 1GHz



Above 1GHz



Note: Area side: 2.4mX3.6m

Antenna Tower meets ANSI C63.4 requirements for measurements above 1 GHz by keeping the antenna aimed at the EUT during the antenna's ascent/ descent along the antenna mast.

Limits

Class B

Frequency (MHz)	Field Strength (dB μ V/m)	Detector
30 -88	40.0	Quasi-peak
88-216	43.5	Quasi-peak
216 – 960	46.0	Quasi-peak
960-1000	54.0	Quasi-peak
1000-5 th harmonic of the highest frequency or 40GHz, which is lower	54 74	Average Peak

Frequency range of radiated measurements


Highest frequency generated or used in the device or on which the device operates or tunes (MHz)	Upper frequency of measurement range (MHz)
Below 1.705	30
1.705-108	1000
108-500	2000
500-1000	5000
Above 1000	5th harmonic of the highest frequency or 40 GHz, whichever is lower.

Test Results

Sweep the whole frequency band through the range from 30MHz to the 5th harmonic of the carrier.

The following graphs display the maximum values of horizontal and vertical by software.

For above 1GHz, Blue trace uses the peak detection, Green trace uses the average detection.




Radiated Emission from 30MHz to 1GHz

Frequency (MHz)	Quasi-Peak (dBμV/m)	Limit (dBμV/m)	Margin (dB)	Height (cm)	Polarization	Azimuth (deg)	Correct Factor (dB)
44.27	26.46	40.00	13.54	100.0	V	268.00	20
53.93	23.94	40.00	16.06	206.0	H	22.00	20
148.30	21.58	43.50	21.92	185.0	H	324.00	15
198.01	26.37	43.50	17.13	325.0	V	140.00	18
337.65	30.75	46.00	15.25	100.0	H	120.00	22
618.87	32.16	46.00	13.84	285.0	H	15.00	27

Remark: 1. Correction Factor = Antenna factor + Insertion loss(cable loss+amplifier gain)

2. Margin = Limit – Quasi-Peak



Radiated Emission from 1GHz to 18GHz

Frequency (MHz)	MaxPeak (dBμV/m)	Average (dBμV/m)	Limit (dB μ V/m)	Margin (dB)	Meas. Time (ms)	Height (cm)	Pol	Azimuth (deg)	Corr. (dB/m)
1231.63	---	22.19	54.00	31.81	1000.00	106.0	H	47.00	-21
1446.25	34.22	---	74.00	39.78	1000.00	106.0	H	9.00	-20
1983.88	---	22.11	54.00	31.89	1000.00	100.0	H	174.00	-19
2045.50	34.45	---	74.00	39.55	1000.00	198.0	H	6.00	-19
2498.13	---	23.19	54.00	30.81	1000.00	198.0	V	35.00	-17
2500.25	35.37	---	74.00	38.63	1000.00	190.0	V	42.00	-17
4017.50	---	26.61	54.00	27.39	1000.00	110.0	V	0.00	-13
4019.63	39.44	---	74.00	34.56	1000.00	206.0	H	130.00	-13
6854.38	44.68	---	74.00	29.33	1000.00	106.0	H	5.00	-5
7664.00	---	33.80	54.00	20.20	1000.00	210.0	V	180.00	-3
10687.88	---	39.70	54.00	14.30	1000.00	210.0	V	207.00	0
10989.63	51.74	---	74.00	22.26	1000.00	106.0	H	16.00	1

Remark: 1. Correction Factor = Antenna factor + Insertion loss (cable loss + amplifier gain)

2. Peak Margin = Limit –MAX Peak/ Average

3.2 Conducted Emission


Ambient Condition

Temperature	Relative humidity
15°C ~ 35°C	30% ~ 60%

Methods of Measurement

The EUT is placed on a non-metallic table of 80cm height above the horizontal metal reference ground plane. During the test, the EUT was operating in its typical mode. The test method is according to ANSI C63.4-2014. Connect the AC power line of the EUT to the L.I.S.N. Use EMI receiver to detect the average and Quasi-peak value. RBW is set to 9 kHz, VBW is set to 30kHz. The measurement result should include both L line and N line.

Test Setup



Note: Power Supply is AC Power source and it is used to change the voltage 120V/60Hz.


Limits

Frequency (MHz)	Class A (dBμV)		Class B (dBμV)	
	Quasi-peak	Average	Quasi-peak	Average
0.15 - 0.5	79	66	66 to 56 *	56 to 46*
0.5 - 5	73	60	56	46
5 - 30	73	60	60	50
*: Decreases with the logarithm of the frequency.				

Note: The EUT should meet CLASS B limit.

Test Results

Following plots, Blue trace uses the peak detection; Green trace uses the average detection.




Frequency (MHz)	QuasiPeak (dBμV)	Average (dBμV)	Limit (dBμV)	Margin (dB)	Meas. Time (ms)	Bandwidth (kHz)	Line	Filter	Corr. (dB)
0.35	---	25.54	48.96	23.42	1000.0	9.000	L1	ON	21.0
0.36	33.79	---	58.80	25.01	1000.0	9.000	L1	ON	21.0
0.43	---	37.45	47.23	9.78	1000.0	9.000	L1	ON	20.9
0.44	46.61	---	57.14	10.53	1000.0	9.000	L1	ON	20.9
0.98	32.77	---	56.00	23.23	1000.0	9.000	L1	ON	20.3
0.98	---	27.20	46.00	18.80	1000.0	9.000	L1	ON	20.3
2.14	---	23.07	46.00	22.93	1000.0	9.000	L1	ON	19.7
3.19	27.32	---	56.00	28.68	1000.0	9.000	L1	ON	19.5
5.27	---	19.48	50.00	30.52	1000.0	9.000	L1	ON	19.5
5.32	23.59	---	60.00	36.41	1000.0	9.000	L1	ON	19.5
13.11	20.20	---	60.00	39.80	1000.0	9.000	L1	ON	19.6
13.26	---	17.82	50.00	32.18	1000.0	9.000	L1	ON	19.6

Remark: Correct factor=cable loss + LISN factor

L line

Conducted Emission from 150 kHz to 30 MHz



Frequency (MHz)	QuasiPeak (dBμV)	Average (dBμV)	Limit (dBμV)	Margin (dB)	Meas. Time (ms)	Bandwidth (kHz)	Line	Filter	Corr. (dB)
0.36	---	23.69	48.75	25.06	1000.0	9.000	N	ON	21.0
0.36	34.06	---	58.75	24.69	1000.0	9.000	N	ON	21.0
0.43	46.41	---	57.27	10.86	1000.0	9.000	N	ON	20.9
0.43	---	35.44	47.23	11.79	1000.0	9.000	N	ON	20.9
0.96	30.26	---	56.00	25.74	1000.0	9.000	N	ON	20.3
0.96	---	21.47	46.00	24.53	1000.0	9.000	N	ON	20.3
3.37	---	19.39	46.00	26.61	1000.0	9.000	N	ON	19.5
3.39	26.95	---	56.00	29.05	1000.0	9.000	N	ON	19.5
5.27	---	17.99	50.00	32.01	1000.0	9.000	N	ON	19.5
5.35	24.49	---	60.00	35.51	1000.0	9.000	N	ON	19.5
13.06	---	16.84	50.00	33.16	1000.0	9.000	N	ON	19.6
13.11	19.40	---	60.00	40.60	1000.0	9.000	N	ON	19.6

Remark: Correct factor=cable loss + LISN factor

N line

Conducted Emission from 150 kHz to 30 MHz

4 Uncertainty Measurement

Case	Uncertainty	Factor k
Radiated Emission 30MHz – 200MHz	4.17 dB	1.96
Radiated Emission 200MHz – 1GHz	4.84 dB	1.96
Radiated Emission 1GHz – 18GHz	4.35 dB	1.96
Conducted Emission	2.57 dB	2

5 Main Test Instruments

Name of Equipment	Manufacturer	Type/Model	Serial Number	Calibration Date	Expiration Time
Radiated Emission					
EMI Test Receiver	R&S	ESCI3	100948	2023-05-12	2024-05-11
				2024-05-07	2025-05-06
Signal Analyzer	R&S	FSV3044	103495	2023-09-19	2024-09-18
TRILOG Broadband Antenna	SCHWARZBECK	VULB 9163	1023	2023-07-14	2026-07-13
Horn Antenna	R&S	HF907	102723	2021-07-24	2024-07-23
Amplifier	R&S	SCU18	10034	2023-05-12	2024-05-11
				2024-05-08	2025-05-07
Software	R&S	EMC32	9.26.01	/	/
Conducted Emission					
Artificial main network	R&S	ENV216	102191	2022-12-10	2024-12-09
EMI Test Receiver	R&S	ESR	101667	2023-05-12	2024-05-11
				2024-05-07	2025-05-06
Software	R&S	EMC32	10.35.10	/	/

ANNEX A: The EUT Appearance

The EUT Appearance are submitted separately.

ANNEX B: Test Setup Photos

The Test Setup Photos are submitted separately.

***** END OF REPORT *****