

Chen, Henry (SZX-PS:EMC)

From: oetech@fcc.gov
Sent: Friday, May 24, 2024 7:30 PM
To: Bu, Cookies
Subject: Response to Inquiry to FCC (Tracking Number 263319)

"Attention! External Mail. Be careful with Links/Attachments!"

Inquiry on 05/21/2024 :

Inquiry:

Dear
FCC,

Good
day!

The product is a Continuous Glucose Monitoring System, it include three parts in the package, Sensor Pack, Sensor Applicator and Sensor, information in Page 9 in user manual for your reference, Sensor Pack and Sensor Applicator don't include any electrical components inside, Sensor supports BLE function (FCC ID:2BEC2-GS1) and it will be kept into Sensor Pack before use, Sensor Pack will be used with the Sensor Applicator to prepare the Sensor for use, as the sensor will be used to applies to human body to monitor continuous glucose , so it need to be kept as sterilization, so it is impossible to print any label information include FCC ID on it, would you please advise if it is acceptable to print label information into user manual and package? Is it necessary to print label information on Sensor Pack also? Please help to advise and thank you very much.

Attachment
user manual which include FCC ID and warning information in page 5 for your reference.

Attachment
outer packing which include FCC ID for your reference.

You
can access to below website to see a video about how the system work, as the video is too big to send out and sorry to the inconveniences caused to you.

<https://cgm-ce-public.s3.eu-central-1.amazonaws.com/video/introduce/Sijoy%E4%BD%A9%E6%88%B4%E8%A7%86%E9%A2%91%EF%BC%88%E7%BA%AF%E8%>

Kind regards,

Bu, Cookies

FCC response on 05/23/2024

Hello,

Yes, if your device meets the requirement in Section 2.935(f) of the FCC rules.

---Reply from Customer on 05/23/2024---

Dear FCC,

Thank you very much to your reply. § 2.935 is the requirement about electronic labeling of radiofrequency devices. But this product will not use electrical label. The sensor will be used to apply to human body to monitor continuous glucose, so it needs to be kept as sterilization, so it is impossible to print any label information on it, would you please advise if it is acceptable to print label information into user manual and package?

Kind regards,

Bu, Cookies

FCC response on 05/24/2024

Hello,

Please see 2.925(f) of the FCC Rules.

If a device is so small that it is impractical to label it with the FCC Identifier in a font that is four-point or larger, and the device does not have a display that can show electronic labeling, then the FCC Identifier shall be placed in the user manual and must also either be placed on the device packaging or on a removable label attached to the device.

Attachment Details:

Do not reply to this message. Please select the [Reply to an Inquiry Response](#) link from the OET Inquiry System to add any additional information pertaining to this inquiry.