

Dipole D5GHzV2 SN 1174						
5250MHz Head Liquid						
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$	Impedance imaginary ($j\Omega$)	$\Delta j\Omega$
2023-08-23	-22.1	/	47.3	/	-7.19	/
2024-08-22	-21.9	0.90%	47.1	0.2	-7.35	0.16
2025-08-21	-22	0.45%	47.2	0.1	-7.28	0.09
5600MHz Head Liquid						
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$	Impedance imaginary ($j\Omega$)	$\Delta j\Omega$
2023-08-23	-27.4	/	53.9	/	-2.14	/
2024-08-22	-27.1	1.09%	53.7	0.2	-2.35	0.21
2025-08-21	-27.2	1.09%	53.8	0.1	-2.23	0.09
5750MHz Head Liquid						
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$	Impedance imaginary ($j\Omega$)	$\Delta j\Omega$
2023-08-23	-24.6	/	53.2	/	-5.17	/
2024-08-22	-24.2	1.63%	52.9	0.3	-5.39	0.22
2025-08-21	-24.4	1.63%	53.1	0.1	-5.27	0.1

Dipole D6.5GHzV2- SN 1030						
Head Liquid						
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$	Impedance imaginary ($j\Omega$)	$\Delta j\Omega$
2025/1/13	-26.1	/	52.1	/	-4.6	/
2026/1/10	-27.5	5.36%	55.3	3.2	-2.6	2.0

5G Verification Source 10 GHZ						
Ari						
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$	Impedance imaginary ($j\Omega$)	$\Delta j\Omega$
2024-08-20	-21.4	/	42.2	/	1.1	/
2025-08-19	-21.2	0.93%	42.1	0.1	0.9	0.2



Report No.: SUCR260400034210
Amendment 01
Page: 138 of 138

- End of the Appendix -