



Huxley Medical
1465 Northside Dr. NW, Suite 217
Atlanta, Georgia 30318

Date: 09/27/2024

Federal Communications Commission
Authorization and Evaluation Division

Contains FCC ID: 2BHX6-SANSALTE1000

Request for Class II Permissive Change

To whom it may concern,

This is a request to file a Class II permissive change for FCC ID: 2BHX6-SANSALTE1000, originally granted on 09/26/2024.

In the updated host, the module is which are being used in a portable device (Product Name: SANSA, Model: 1000-1, Brand Name: Huxley Medical).

There are no hardware changes in the module.

The SANSA 1000-1 only supports band 2, 4, and 12. The rest of the bands are not supported through firmware.

The summary changes filed under this application are:

	Original Module 1#	Present Portable Device
FCC ID	FCC ID: 2BHX6-SANSALTE1000	Contain FCC ID: 2BHX6-SANSALTE1000
	Original Grant Date: September 26 th , 2024	Only band 2, 4, 12

We have performed the SAR evaluation of this portable device and also re-measured the radiated spurious emissions.

Sincerely,

Signature:

Daniel Burnham

Daniel Burnham
CTO