

October 06, 2025

Federal Communications Commission
Authorization and Evaluation Division
7435 Oakland Mills Road
Columbia, MD 21046

Subject: Justification for Alternate FCC ID Labeling
FCC ID: 2BRSV-FR1RICRA


To Whom It May Concern:

Fortell Research Inc. respectfully submits this letter to justify the placement of the FCC ID label for the above-referenced device.

Our Hearing Aids have physical dimensions too small to apply a permanent, legible label with the required FCC ID information in a 4-point font or larger. The critical dimensions and limited surface area available on the product's exterior make it impractical to meet standard on-device labeling requirements.


Therefore, in accordance with FCC regulations, specifically KDB Publication 784748 D01 and 47 CFR 2.925(f), we will use the following alternative labeling method:

1. **FCC ID in User Manual:** The FCC Identifier will be clearly and prominently included in the device's user manual.
2. **FCC ID on Packaging:** The full FCC ID will be printed on the device's retail packaging.



3. **Mobile App:** The full FCC ID will be listed in the mobile app device information.

- Account > Device Information:



×

Device information

Hearing aid system info

Hearing aid firmware

-

Type

FCC ID (Left RIC)

2BRSV-FR1RICLA

FCC ID (Right RIC)

2BRSV-FR1RICRA

FCC

The FCC ID will be easily accessible to the user through the provided manual, mobile app, and the product packaging, ensuring that all regulatory requirements are met.

Thank you for your consideration of this request.

Sincerely,



John Becker
Principal Electrical Engineer
Fortell Research Inc.
jbecker@fortell.com