



Picture 8: Labels on back of debritom+



MEDICAL-GENERAL MEDICAL EQUIPMENT
AS TO ELECTRICAL SHOCK, FIRE
AND MECHANICAL HAZARDS ONLY
ANSI/AAMI ES60601-1 (2005)
+ AMD 1 (2012), CSA CAN/CSA-C22.2
No.60601-1:14, IEC 60601-1-6 (2010)
+ AMD 1 (2013) CSA-C22.2
No. 60601-1-6:11 + AMD1:2015

FCC ID:2BS2Y10000000A
Contains FCC ID:XPYNORAB2

This device complies with Part 15 of the
FCC Rules. Operation is subject to the
following two conditions: (1) this device
may not cause harmful interference, and
(2) this device must accept any
interference received, including
interference that may cause undesired
operation.

9000.0030/B

debritom⁺ basic unit

REF 1000.0000

UDI

SN 00106



(01)07 649 996 166 531
(21)00 106

[VAC] 100-240

[Hz] 50 / 60

[VA] 200

 T 6.3 AH, 250 V

IP 20



CE 0123 **R_X** only



2026-03-19

MD

REP

Emergo Europe B.V.
Westervoortsedijk 60
6827 AT Arnhem
The Netherlands

EU



Medaxis AG
Bahnhofstrasse 9
CH-6340 Baar

9000.0024/H