

Prüfbericht-Nr.: <i>Test report no.:</i>	DE26E8YH 001	Auftrags-Nr.: <i>Order no.:</i>	1184687 70	Seite 1 von 16 Page 1 of 16
Kunden-Referenz-Nr.: <i>Client reference no.:</i>	Tim Jessat	Auftragsdatum: <i>Order date:</i>	2025-01-27	
Auftraggeber: <i>Client:</i>	STORZ MEDICAL AG, Lohstampfestr. 8, 8274 Tägerwilen, Schweiz			
Prüfgegenstand: <i>Test item:</i>	Applied part for Radial Shock Wave device			
Bezeichnung / Typ-Nr.: <i>Identification / Type no.:</i>	Handpiece R-SW (Endurance) including RFID			
Auftrags-Inhalt: <i>Order content:</i>	Prüfung der Funkparameter nach FCC & ISCED <i>Test of radio parameters acc. to FCC & ISCED</i>			
Prüfgrundlage: <i>Test specification:</i>	Vollprüfung / Complete test RF exposure evaluation acc. to FCC CFR 47 Part 1 Subpart I §1.1307 (b) acc. to RSS-102 issue 6			
Wareneingangsdatum: <i>Date of sample receipt:</i>	2025-04-11	For reasons of confidentiality, no photos of the EUT are included in this report. For EUT and test setup photos please see the Photo Documentation to this report Number.		
Prüfmuster-Nr.: <i>Test sample no.:</i>	A003964660-001			
Prüfzeitraum: <i>Testing period:</i>	2025-05-13 - 2025-05-13			
Ort der Prüfung: <i>Place of testing:</i>	Nürnberg Nuremberg			
Prüflaboratorium: <i>Testing laboratory:</i>	Wireless Labor Wireless Test Lab			
Prüfergebnis*: <i>Test result*:</i>	Pass			
geprüft von: <i>tested by:</i>	16.04.2026 X 		genehmigt von: <i>authorized by:</i>	17.04.2026 X 
Datum: <i>Date:</i>	Signiert von: Primoz Erzen		Ausstellungsdatum: <i>Issue date:</i>	Signiert von: Matthias Kraeutlein
Stellung / Position:	Sachverständige(r)/Expert		Stellung / Position:	Sachverständige(r)/Expert
Sonstiges / <i>Other:</i>				
Zustand des Prüfgegenstandes bei Anlieferung: <i>Condition of the test item at delivery:</i>	Prüfmuster vollständig und unbeschädigt <i>Test item complete and undamaged</i>			
* Legende: P(ass) = entspricht o.g. Prüfgrundlage(n) F(ail) = entspricht nicht o.g. Prüfgrundlage(n) N/A = nicht anwendbar N/T = nicht getestet * Legend: P(ass) = passed a.m. test specification(s) F(ail) = failed a.m. test specification(s) N/A = not applicable N/T = not tested				
Dieser Prüfbericht bezieht sich nur auf das o.g. Prüfmuster und darf ohne Genehmigung der Prüfstelle nicht auszugsweise vervielfältigt werden. Dieser Bericht berechtigt nicht zur Verwendung eines Prüfzeichens. <i>This test report only relates to the above mentioned test sample. Without permission of the test center this test report is not permitted to be duplicated in extracts. This test report does not entitle to carry any test mark.</i>				

Prüfbericht-Nr.: DE26E8YH 001
Test report no.:

Seite 2 von 16
Page 2 of 16

Anmerkungen Remarks

1	Alle eingesetzten Prüfmittel waren zum angegebenen Prüfzeitraum gemäß eines festgelegten Kalibrierungsprogramms unseres Prüfhauses kalibriert. Sie entsprechen den in den Prüfprogrammen hinterlegten Anforderungen. Die Rückverfolgbarkeit der eingesetzten Prüfmittel ist durch die Einhaltung der Regelungen unseres Managementsystems gegeben. Detaillierte Informationen bezüglich Prüfkonditionen, Prüfequipment und Messunsicherheiten sind im Prüflabor vorhanden und können auf Wunsch bereitgestellt werden. <i>The equipment used during the specified testing period was calibrated according to our test laboratory calibration program. The equipment fulfils the requirements included in the relevant standards. The traceability of the test equipment used is ensured by compliance with the regulations of our management system. Detailed information regarding test conditions, equipment and measurement uncertainty is available in the test laboratory and could be provided on request.</i>
2	Wie vertraglich vereinbart, wurde dieses Dokument nur digital unterzeichnet. Der TÜV Rheinland hat nicht überprüft, welche rechtlichen oder sonstigen diesbezüglichen Anforderungen für dieses Dokument gelten. Diese Überprüfung liegt in der Verantwortung des Benutzers dieses Dokuments. Auf Verlangen des Kunden kann der TÜV Rheinland die Gültigkeit der digitalen Signatur durch ein gesondertes Dokument bestätigen. Diese Anfrage ist an unseren Vertrieb zu richten. Eine Umweltgebühr für einen solchen zusätzlichen Service wird erhoben. Informationen zur Verifizierung der Authentizität unserer Dokumente erhalten Sie auf folgender Webseite: go.tuv.com/digital-signature <i>As contractually agreed, this document has been signed digitally only. TUV Rheinland has not verified and unable to verify which legal or other pertaining requirements are applicable for this document. Such verification is within the responsibility of the user of this document. Upon request by its client, TUV Rheinland can confirm the validity of the digital signature by a separate document. Such request shall be addressed to our Sales department. An environmental fee for such additional service will be charged. For information on verifying the authenticity of our documents, please visit the following website: go.tuv.com/digital-signature</i>
3	Prüfklausel mit der Note * wurden an qualifizierte Unterauftragnehmer vergeben und sind unter der jeweiligen Prüfklausel des Berichts beschrieben. Abweichungen von Prüfspezifikation(en) oder Kundenanforderungen sind in der jeweiligen Prüfklausel im Bericht aufgeführt. <i>Test clauses with remark of * are subcontracted to qualified subcontractors and described under the respective test clause in the report.</i> <i>Deviations of testing specification(s) or customer requirements are listed in specific test clause in the report.</i>
4	Die Entscheidungsregel für Konformitätserklärungen basierend auf numerischen Messergebnissen in diesem Prüfbericht basiert auf der "Null-Grenzwert-Regel" und der "Einfachen Akzeptanz" gemäß ILAC G8:2019 und IEC Guide 115:2023, es sei denn, in der auf Seite 1 dieses Berichts genannten angewandten Norm ist etwas anderes festgelegt oder vom Kunden gewünscht. Dies bedeutet, dass die Messunsicherheit nicht berücksichtigt wird und daher auch nicht im Prüfbericht angegeben wird. Zu weiteren Informationen bezüglich des Risikos durch diese Entscheidungsregel siehe ILAC G8:2019. <i>The decision rule for statements of conformity, based on numerical measurement results, in this test report is based on the "Zero Guard Band Rule" and "Simple Acceptance" in accordance with ILAC G8:2019 and IEC Guide 115:2023, unless otherwise specified in the applied standard mentioned on Page 1 of this report or requested by the customer. This means that measurement uncertainty is not taken in account and hence also not declared in the test report. For additional information to the resulting risk based of this decision rule please refer to ILAC G8:2019.</i>
5	Wenn auf dem Bericht kein Akkreditierungshinweis aufgebracht ist, wurde der Bericht nicht im akkreditierten Bereich erstellt und ist folglich auch nicht vom EA MLA abgedeckt. Unabhängig davon wurde der Bericht auf Basis der allgemeinen Regeln der ISO/IEC 17000er Reihe erstellt. Mit "#" gekennzeichnete Prüfungen sind nicht Bestandteil der Akkreditierung D-PL-14169-03-00. <i>If there is no accreditation notice on the report, the report has not been produced in the accredited area and is consequently not covered by the EA MLA. Regardless of this, the report has been prepared based on the general rules of the ISO/IEC 17000 series. Tests marked with "#" are not covered by the accreditation D-PL-14169-03-00.</i>

Prüfbericht-Nr.: DE26E8YH 001
Test report no.:

Seite 3 von 16
Page 3 of 16

Produktbeschreibung
Product description

1	Test item	Applied part for Radial Shock Wave device
2	Typ-No.	Handpiece R-SW (Endurance) including RFID
3	Test sample obtaining	☒ Sampling by customer ☐ Sampling by TÜV Rheinland Group ☐ others:
4	Serial-No.	A003964660-001 (rad. sample): 37700.00015
5	Supported radio technologies evaluated in this report	RFID 13.56MHz
6	Supported radio technologies not evaluated in this report	-
7	Description of EUT	Handpiece including RFID Reader and RFID Tag
8	Rated Voltage / Frequency	12V -24V DC provided by Host device
9	Antenna Type	☒ Integral antenna (designed as a fixed part of the EUT) ☐ Dedicated antenna (removable, obligation to be used) ☐ Permanent antenna connector
10	Antenna number of chains	1
11	Software / Firmware	RFID_Test, Standard software with all functions deactivated but RFID functionality
12	Hardware	37700_03
13	Provided Samples	A003964660-001
14	Radiated measurements sample	A003964660-001 (rad. sample)
15	Companion Device used for radio connection with the EUT	RFID always active. Therefore, no Companion Device was used.
16	Auxiliary equipment used to set up the EUT	-
17	Accessory Devices used together with the EUT	A003964660-003 12V-Battery with EUT-Connector
18	Temperature Range	-20 to 50 °C
19	Environment	Indoor

Prüfbericht-Nr.: DE26E8YH 001
Test report no.:

Seite 4 von 16
Page 4 of 16

Prüfdokumentation
Test documentation

Contents

1	Registration	5
2	EUT Classification	5
2.1	Wireless technologies and frequencies supported by the EUT	5
2.2	Standard specific classification of the EUT	5
3	Test conditions and configurations	6
3.1	Normal environmental test conditions	6
3.2	Equipment modifications	6
3.3	Test Modes	7
3.4	Measurement Uncertainty	8
4	Test results in detail FCC RF exposure evaluation	9
4.1	Maximum Permissible Exposure Evaluation (MPE) Mobile device or fixed RF source	9
5	Test results in detail ISED RF exposure Compliance	11
5.1	ISED Maximum Exposure Requirements	11
6	Test Setup Photo	13
7	Application form	14
8	Equipment List	15
8.1	Hardware	15
9	Change history	16

Prüfbericht-Nr.: DE26E8YH 001
Test report no.:

Seite 5 von 16
Page 5 of 16

Prüfdokumentation
Test documentation

1 Registration

The measurement facilities for conducted and radiated disturbance measurements of the TÜV Rheinland LGA Products have been found to be in compliance with the requirements of Section 2.948 of the FCC Rules. Measurement data will be accepted in conjunction with applications for Certification under Parts 15 and 18 of the Commission's Rules.

FCC Registration Number: 939976

Bundesnetzagentur Registration Number: BNetzA-CAB-17/21-16

The measurement facilities are also recognized by Innovation, Science and Economic Development (ISED) Canada to test to Canadian radio equipment requirements.

Company Number: 11235A

CAB Identifier: DE0018

2 EUT Classification

2.1 Wireless technologies and frequencies supported by the EUT.

The named technologies are only those falling in the specification of the applied standard.

Technology	Frequency Range (TX)	TX Function	Supported by the EUT	Evaluated in this report
RFID	13.11 – 14.01 MHz (fc=13.56 MHz)	SISO	☒	☒

SISO: Single Input Single Output

2.2 Standard specific classification of the EUT

2.2.1 Applied FCC standards

FCC CFR 47 Part 1 Subpart I §1.1307 (b)

FCC CFR 47 Part 1 Subpart I §1.1310 (limits)

FCC CFR 47 Part 2 Subpart J §2.1091 (mobile)

FCC CFR 47 Part 2 Subpart J §2.1093 (portable)

FCC KDB 447498 D04 v01 – Interim General RF Exposure procedures Guidance (Procedures and Equipment Authorization Policies for Mobile and Portable Devices)

2.2.2 Applied ISED standards

RSS-GEN (General Requirements for Compliance of Radio Apparatus)

RSS-102 issue 6 (Radio Frequency (RF) Exposure Compliance of Radiocommunication Apparatus)

Health Canada's Safety Code 6 (2015)

Prüfbericht-Nr.: DE26E8YH 001
Test report no.:

Seite 6 von 16
Page 6 of 16

Prüfdokumentation
Test documentation

2.2.3 Exposure conditions

	Type of exposure
☒	General public exposure
☐	Occupational exposure

	Separation distance – Intended usage location
☐	< 20 cm – portable
☐	Body-worn device -
☐	Limb-worn device -
☐	Head exposure device -
☐	Implanted device
☒	≥ 20 cm
☐	Fixed device (physically secured at one location, even temporarily) specific mitigation actions are required
☒	Mobile device

3 Test conditions and configurations

3.1 Normal environmental test conditions

Environmental condition	Parameter	Range during test
Temperature	°C	21-24
Relative humidity	%	20-60
Supply voltage	V DC	12

3.2 Equipment modifications

No modifications were found to be necessary to perform the tests or to achieve compliance.

Prüfbericht-Nr.: DE26E8YH 001
Test report no.:

Seite 7 von 16
Page 7 of 16

Prüfdokumentation
Test documentation

3.3 Test Modes

Following configurations were used for the test cases:

Mode description	Mode configuration
Mode 1: TX-Mode external Power supply	Normal operation. EUT is connected to Accessory device A003964660-003 12V-Battery with EUT-Connector. RFID always active.
Mode 2: N/A	-
Mode 3: N/A	-
Mode 4: N/A	-

Prüfbericht-Nr.: DE26E8YH 001
Test report no.:

Seite 8 von 16
Page 8 of 16

Prüfdokumentation
Test documentation

3.4 Measurement Uncertainty

The data and results referenced in this document are true and accurate. The reader is cautioned that there may be errors within the calibration limits of the equipment and facilities. The measurement uncertainty was calculated for all measurements listed in this test report acc. to ETSI TR 100 028-1/-2 and ETSI TS 103 051 (radiated measurements) and is documented in the quality system acc. to ISO/IEC 17025:2017.

Furthermore, component and process variability of devices similar to that tested may result in additional deviation.

The stated values are the expanded uncertainty values, the measured value lies within the assigned range of values with a probability of 95% ($k = 1.96$).

The manufacturer has the sole responsibility of continued compliance of the device.

Test	Value	Unit	Range
Frequency Error	$2.7 \cdot 10^{-8}$	Hz	-
Frequency Stability under low voltage conditions			
Time	$1.1 \cdot 10^{-9}$	s	-
Conducted Carrier Power	1.0	dB	9k-1GHz
Conducted Spurious Emissions (RX/TX)	1.7	dB	1GHz-6GHz
	2.8	dB	6GHz-40GHz
Occupied Bandwidth (OBW)	0.1	%	-
TX Power Spectral Density	4.2	mW	9kHz - 6GHz
	1.6	dB	9kHz - 6GHz
Dwell Time	$4.6 \cdot 10^{-7}$	Hz	-
Frequency Separation			-
Measurement of conducted emissions at the power supply connection to LISN	2.3	dB	9kHz – 150kHz
	2.2	dB	150kHz – 30MHz
Measurement of the field strength at distance 3m	1.6	dB	9k-150kHz
	4.5	dB	30-1000MHz
	5	dB	1-6GHz
	5.3	dB	6-40GHz
Temperature	0.8	K	-
Humidity	4	%	-
Voltage (AC/DC)	1.0	%	-

Measured value	Limit	Extended measurement uncertainty ($k=1.96$)	Confidence interval (95%)
19 dBm @ 2.48GHz	20 dBm	1.7 dB	17.3 dBm – 20.7 dBm

Prüfbericht-Nr.: DE26E8YH 001
Test report no.:

Seite 9 von 16
Page 9 of 16

Prüfdokumentation
Test documentation

4 Test results in detail FCC RF exposure evaluation

Radio frequency devices operating under the provisions of FCC CFR 47 Part 15 are subject to the radio frequency radiation exposure requirements specified in §§ 1.1307(b), 1.1310, 2.1091, and 2.1093, as appropriate.

4.1 Maximum Permissible Exposure Evaluation (MPE) Mobile device or fixed RF source

4.1.1 Description

For operations within the frequency range of 300 kHz and 6 GHz (inclusive), the limits for maximum permissible exposure (MPE), derived from whole-body SAR limits and listed in Table 1 to § 1.1310(e)(1), may be used instead of whole-body SAR limits to evaluate the environmental impact of human exposure to RF radiation as specified in § 1.1307(b) of this part, except for portable devices as defined in § 2.1093 of this chapter as these evaluations shall be performed according to the SAR provisions in § 2.1093.

4.1.2 Limits

Limits for Maximum Permissible Exposure (MPE) to radiofrequency electromagnetic fields from §1.1310(e)(1) are shown in the table below.

Table 1 to § 1.1310(e)(1) - Limits for Maximum Permissible Exposure (MPE)				
Frequency range (MHz)	Electric field strength (V/m)	Magnetic field strength (A/m)	Power density (mW/cm ²)	Averaging time (minutes)
(i) Limits for Occupational/Controlled Exposure				
0.3-3.0	614	1.63	*(100)	≤6
3.0-30	1842/f	4.89/f	*(900/f ²)	<6
30-300	61.4	0.163	1.0	<6
300-1,500			f/300	<6
1,500-100,000			5	<6
(ii) Limits for General Population/Uncontrolled Exposure				
0.3-1.34	614	1.63	*(100)	<30
1.34-30	824/f	2.19/f	*(180/f ²)	<30
30-300	27.5	0.073	0.2	<30
300-1,500			f/1500	<30
1,500-100,000			1.0	<30

f = frequency in MHz. * = Plane-wave equivalent power density.

Prüfbericht-Nr.: DE26E8YH 001
Test report no.:

Seite 10 von 16
Page 10 of 16

Prüfdokumentation Test documentation

4.1.3 Evaluation for Technologies assumed to be in near field at the distance d

In the near-field zone, both the unperturbed electric- and magnetic-field strengths shall be measured, since there is no simple relationship between these two quantities.

At 50 MHz, the near field extends out to about 1 m. For frequencies below 50 MHz and within this distance, the magnetic field is the predominant field component of loop antennas. Therefore, only the H-field was measured.

As some limits are frequency dependent, for each Band the limit is calculated based of the lowest band edge frequency f_{Lowest} .

4.1.3.1 Test setup

Test method	Radiated		
EUT	A003964660-001 (rad. sample)		
Environmental Cond.	Normal		
Chamber details	☐ Shielded room ☒ 3m Fully Anechoic Chamber (FAC) ☐ 10m Semi Anechoic Chamber (SAC)		
Measurement positioning	Distance: ☒ 20cm ☐ 3m ☐ 10m	EUT height: ☐ 1.5m ☒ 0.8m ☐ floor standing	Antenna height: ☐ 1.5m ☐ 1m to 4m – height scan ☒ around the EUT
Companion device	RFID always active. Therefore, no Companion Device was used.		
Operation mode	☒ 1: TX-Mode external Power supply ☐ 3: N/A ☐ 2: N/A ☐ 4: N/A		
Further parameters	-		
Test engineer	Dipl. Ing. Primoz Erzen		

4.1.3.2 Test results for General Population

f [MHz]	RBW [MHz]	Integration [MHz]	Sweep Time [ms]	AVG	H [A/m]	f_{Lowest} [MHz]	H_L [A/m]	H / H_L
13.56	1	8.58-18.58	30ms	4	0.04899	11.810	0.414	0.118

4.1.4 Final evaluation

4.1.4.1 Single RF source

The equipment meets the requirements according to §1.1307(b) since the magnetic field strength is below the applicable limit from §1.1310(e)(1)

Final test result	Pass
-------------------	------

Prüfbericht-Nr.: DE26E8YH 001
Test report no.:

Seite 11 von 16
Page 11 of 16

Prüfdokumentation
Test documentation

5 Test results in detail ISED RF exposure Compliance

Radio frequency devices operating under the provisions of RSS-Gen are subject to the radio frequency radiation exposure requirements specified in RSS-102.

5.1 ISED Maximum Exposure Requirements

According to RSS-102 issue 6, Industry Canada has adopted the SAR and RF field strength limits established in Health Canada's RF exposure guideline, Safety Code 6 (2015) and its Notice. Any updates to Safety Code 6 and/or its Notice will supersede the values outlined in RSS-102 issue 6

5.1.1 Limits

To ensure compliance with the basic restrictions at frequencies between 10 MHz and 300 GHz, the reference levels for electric- and magnetic-field strength and power density must be complied with.

5.1.1.1 Electric Field Strength, Magnetic Field Strength and Power Density (10 MHz - 300 GHz) general public

According to the limits of human exposure in Safety Code 6 clause 2.2.2:

To ensure compliance with the basic restrictions at frequencies between 10 MHz and 300 GHz, the reference levels for electric- and magnetic-field strength and power density must be complied with.

Table 7: RF field strength and power density limits for devices used by the general public (uncontrolled environment)

Frequency range (MHz)	Electric field (V _{RMS} /m)	Magnetic field (A _{RMS} /m)	Power density (W/m ²)	Reference period (minutes)
10-20	27.46	0.0728	2	6
20-48	$58.07 / f^{0.25}$	$0.1540 / f^{0.25}$	$8.944 / f^{0.5}$	6
48-300	22.06	0.05852	1.291	6
300-6000	$3.142 f^{0.3417}$	$0.008335 f^{0.3417}$	$0.02619 f^{0.6834}$	6
6000-15000	61.4	0.163	10	6
15000-150000	61.4	0.163	10	$616000 / f^{1.2}$
150000-300000	$0.158 f^{0.5}$	$4.21 \times 10^{-4} f^{0.5}$	$6.67 \times 10^{-5} f$	$616000 / f^{1.2}$

Note: f is frequency in MHz.

Prüfbericht-Nr.: DE26E8YH 001
Test report no.:

Seite 12 von 16
Page 12 of 16

Prüfdokumentation
Test documentation

5.1.2 Evaluation for Technologies assumed to be in near-field at the distance d

In the near-field zone, both the unperturbed electric- and magnetic-field strengths shall be measured, since there is no simple relationship between these two quantities.

At 50 MHz, the near field extends out to about 1 m. For frequencies below 50 MHz and within this distance, the magnetic field is the predominant field component of loop antennas. Therefore, only the H-field was measured.

As some limits are frequency dependent, for each Band the limit is calculated based of the lowest band edge frequency f_{Lowest} .

5.1.2.1 Test setup

Test method	Radiated		
EUT	A003964660-001 (rad. sample)		
Environmental Cond.	Normal		
Chamber details	☐ Shielded room ☒ 3m Fully Anechoic Chamber (FAC) ☐ 10m Semi Anechoic Chamber (SAC)		
Measurement positioning	Distance: ☒ 20cm ☐ 3m ☐ 10m	EUT height: ☐ 1.5m ☒ 0.8m ☐ floor standing	Antenna height: ☐ 1.5m ☐ 1m to 4m – height scan ☒ around the EUT
Companion device	RFID always active. Therefore, no Companion Device was used.		
Operation mode	☒ 1: TX-Mode external Power supply ☐ 3: N/A ☐ 2: N/A ☐ 4: N/A		
Further parameters	-		
Test engineer	Dipl. Ing. Primoz Erzen		

5.1.2.2 Test results for general public

f [MHz]	RBW [MHz]	Integration [MHz]	Sweep Time [ms]	AVG [ms]	H [A/m]	f_{Lowest} [MHz]	H_L [A/m]	H / H_L
13.56	1	8.58	18.58	120	0.04899	11.810	0.073	0.671

5.1.3 Final evaluation

5.1.3.1 Single RF source

The equipment meets the requirements according to Safety Code 6 clause 2.2.2 since the magnetic field strength is below the applicable limit.

5.1.4 Final test result

The equipment is compliant according to Health Canada Safety Code 6 (2015).

Final test result	Pass
-------------------	------

Prüfbericht-Nr.: DE26E8YH 001
Test report no.:

Seite 13 von 16
Page 13 of 16

Fotodokumentation
Photo documentation

6 Test Setup Photo

For reasons of confidentiality, no photos of the EUT are included in this report.
For EUT and test setup photos please see the Photo Documentation to this report Number.

Prüfbericht-Nr.: DE26E8YH 001
Test report no.:

Seite 14 von 16
Page 14 of 16

Zusatzdokumentation
Additional documentation

7 Application form

The following information was provided by the customer and form the basis for the execution of the tests and the assessment of conformity. The given information can affect the results of both.

No application form was provided.

Prüfbericht-Nr.: DE26E8YH 001
Test report no.:

Seite 15 von 16
Page 15 of 16

Liste der verwendeten Prüfmittel
List of used test equipment

8 Equipment List

8.1 Hardware

Prüfmittel Test equipment		Prüfmittel-Nr. / ID- Nr. Equipment-No. / ID- No.	Letzte Kalibrierung Last calibration	Nächste Kalibrierung Next calibration
E/H Field Meter SRM-3006 9kHz-6GHz	Narda	2729050	03.07.2023	03.07.2026
H-Field Antenna Three-Axis 9kHz - 250MHz	Narda	2728696	03.07.2023	03.07.2026

Prüfbericht-Nr.: DE26E8YH 001
Test report no.:

Seite 16 von 16
Page 16 of 16

Änderungsverzeichnis
Change history

9 Change history

Revision Number	List of revisions
001	Initial Release

Note: Latest revision report will replace all previous reports.

Ende des Prüfberichts
End of Test Report