





Compliance label
location on meter



FOR REFERENCE ONLY

Colour:

- PMS Black C
- Pantone 303C
- Pantone 1525 C
- Pantone 3125 C
- Pantone 7530 C
- Tech Info
- Dimensions

NOTES:

1. All dims. on this drawing are for reference only. Refer to M0676 for dimensional information.
2. Refer to M0676 for material information.
3. All printing to be letter pressed.
4. Roboto font for all text.
5. White coloured branding, text, and symbols (CMYK 0/0/0/0) or equivalent.
6. Serial Number (S/N) PMS Black C (CMYK 67/64/66/67)

- 7.** UDI Barcode prints in PMS Black C (CMYK 67/64/66/67). Manufacturer to build barcode variably. Magenta to be replaced with PMS Black C once real barcode is created and placed on artwork. UDI Barcode Build Details, see below:

Barcode Print Details

Barcode Dimension: 8.890 mm Square (Including Quiet Zones)
Barcode Type: GS1 Data Matrix
Verification Grade Required: C Or Better
Minimum Module Size: 0.340 mm
Maximum Module Size: 0.495 mm
Print Requirements: Direct Print

Variable Print Details

Device Identifier:
GTIN (GS1 Code 01): 9360136000161
Production Identifier:
Serial (GS1 Code 21) Controlled by Supplier (Refer to Note 8.)
Human Readable Form: (21) XXXXXX
Model (GS1 Code 240): A0419
Analyzer (GS1 Code 251): A0418CRW

- 8.** Do not print magenta.
Area reserved for Serial Number (S/N), which is variably printed.
Magenta to be replaced with PMS Black C once Serial Number is assigned and placed on artwork.
Serial Number will start and increment as follows:
XXXXXX is incrementing, for example, 000001,000002

 Aquascout Testing Pty Ltd 1 CORPORATE AVE, ROWVILLE, VIC 3178, AUSTRALIA TEL: +61 3 9213 9000	
TITLE: AquaScout Analyzer Label Artwork	
DRAWING NO.: 79400	VERSION: 1.3
DRAWN DATE: 18-03-2026	
SCALE: 1:1	SHEET: 1 OF 1
Commercial in Confidence © Copyright 2026 Property of Aquascout Testing Pty Ltd	
A4	