TEST REPORT

Product Name : Wearable Breast Pump
Model Number : M21-APP, K10-APP
FCC ID : 2BW2F-K10M21LVB

Prepared for	:	HuiZhou LVB maternal and infant supplies Co,Ltd
Address	:	1st floor, Luotang leprosy ridge section, Shuikou Street Office, Huicheng District, Huizhou City, Guangdong, China
Prepared by	:	EMTEK (DONGGUAN) CO., LTD.
Address	:	1&2/F.,Building 2, Zone A, Zhongda Marine Biotechnology Research and Development Base, No.9, Xincheng Avenue, Songshanhu High-technology Industrial Development Zone, Dongguan, Guangdong, China TEL: +86-0769-22807078 FAX: +86-0769-22807079

Report Number	:	EDG2607060282E01203R
Date of sample receipt	:	july 06, 2026
Date(s) of Tests	:	july 06, 2026 to August 14, 2026
Date of issue	:	August 14, 2026

Table of Contents

1. TEST RESULT CERTIFICATION 3
2. EUT SPECIFICATION 5
3. TEST REQUIREMENT 6
4. MEASUREMENT RESULT 7

1. TEST RESULT CERTIFICATION

Applicant	:	HuiZhou LVB maternal and infant supplies Co,Ltd
Address	:	1st floor, Luotang leprosy ridge section, Shuikou Street Office, Huicheng District, Huizhou City, Guangdong, China
Manufacturer	:	HuiZhou LVB maternal and infant supplies Co,Ltd
Address	:	1st floor, Luotang leprosy ridge section, Shuikou Street Office, Huicheng District, Huizhou City, Guangdong, China
Factory	:	HuiZhou LVB maternal and infant supplies Co,Ltd
Address	:	1st floor, Luotang leprosy ridge section, Shuikou Street Office, Huicheng District, Huizhou City, Guangdong, China
EUT	:	Wearable Breast Pump
Model Name	:	M21-APP, K10-APP
Trademark	:	N/A

Measurement Procedure Used:

APPLICABLE STANDARDS	
STANDARD	**TEST RESULT**
§ 15.247(i), § 2.1093	**PASS**

The above equipment was tested by EMTEK (DONGGUAN) CO., LTD. The test data, data evaluation, test procedures, and equipment configurations shown in this report were made in accordance with the procedures given in ANSI C63.10 (2020) and the energy emitted by the sample EUT tested as described in this report is in compliance with the requirements of FCC Rules FCC § 15.247(i), § 2.1093.

The test results of this report relate only to the tested sample identified in this report

Date of Test : july 06, 2026 to August 14, 2026

Prepared by : Jessica Zhang /Editor

Reviewer : Warren Deng /Supervisor

Approve & Authorized Signer : Sam Lv / Manager

Modified History

Version	Report No.	Revision Date	Summary
	EDG2607060282E01203R	/	Original Report

2. EUT Specification

Characteristics	Description
Product:	Wearable Breast Pump
Model Number:	M21-APP, K10-APP Note: All models are the same, only the model name and color are different. Model M21-APP is selected here to complete all the tests.
Sample:	1#
Data Rate:	1Mbps for GFSK modulation 2Mbps for π/4-DQPSK modulation 1Mbps for GFSK modulation 2Mbps for GFSK modulation
Modulation:	GFSK, π/4-DQPSK, 8DPSK
Operating Frequency Range(s) :	2402-2480MHz
Number of Channels:	BT: 79 channels BLE:40 channels
Transmit Power Max:	BT: -5.06 dBm(0.000312 W) BLE: -4.92 dBm(0.000396 W)
Antenna Gain:	3.08 dBi
Power supply:	DC5V from USB, DC3.7V from battery
Evaluation applied:	☐ MPE Evaluation ☒ SAR Evaluation

3. Test Requirement

SAR Evaluation

According to 447498 D01 V06, systems operating under the provisions of this section shall be operated in a manner that ensures that the public is not exposed to radio frequency energy level in excess of the Commission's guidelines.

The 1-g and 10-g SAR test exclusion thresholds for 100 MHz to 6 GHz at *test separation distances* ≤ 50 mm are determined by:

[(*max. power of channel, including tune-up tolerance, mW*)/(*min. test separation distance, mm*)] · [$\sqrt{f_{(GHz)}}$] ≤ 3.0 for 1-g SAR and ≤ 7.5 for 10-g extremity SAR,[24] where

- $f_{(GHz)}$ is the RF channel transmit frequency in GHz
- Power and distance are rounded to the nearest mW and mm before calculation[25]
- The result is rounded to one decimal place for comparison
- 3.0 and 7.5 are referred to as the numeric thresholds in the step 2 below

The test exclusions are applicable only when the minimum *test separation distance* is ≤ 50 mm and for transmission frequencies between 100 MHz and 6 GHz. When the minimum *test separation distance* is < 5 mm, a distance of 5 mm according to 5) in section 4.1 is applied to determine SAR test exclusion.

Routine SAR evaluation refers to that specifically required by § 2.1093, using measurements or computer simulation. When routine SAR evaluation is not required, portable transmitters with output power greater than the applicable low threshold require SAR evaluation to quality for TCB approval.
One antenna is available for the EUT. The minimum separation distance is 5mm.

4. Measurement Result

Antenna gain:3.0 dBi

Transmit Frequency (MHz)	Mode	Measured Power (dBm)	Tune up Power (dBm)	Max tune up power (dBm)	Calculation Result	Calculation threshold(1-g SAR)
2402	GFSK	-5.52	-6±1	-5	0.0980204	3
2441	GFSK	-7.47	-8±1	-7	0.0623468	3
2480	GFSK	-6.56	-7±1	-6	0.0791145	3
2402	Π/4-DQPSK	-5.06	-6±1	-5	0.0980204	3
2441	Π/4-DQPSK	-6.94	-7±1	-6	0.0784899	3
2480	Π/4-DQPSK	-6.11	-7±1	-6	0.0791145	3
2402	GFSK(1M)	-4.02	-2±1	-4	0.1234004	3
2440	GFSK(1M)	-5.85	-6±1	-5	0.0988130	3
2480	GFSK(1M)	-5.96	-6±1	-5	0.0995992	3
2402	GFSK(2M)	-4.63	-5±1	-4	0.1234004	3
2440	GFSK(2M)	-5.65	-6±1	-5	0.0988130	3
2480	GFSK(2M)	-5.77	-6±1	-5	0.0995992	3

According to KDB 447498 D01 V06, no stand-alone required for BT/BLE antenna, and no simultaneous SAR measurement is required.

*** End of Report ***