835MHz Brain Dipole Validation

Generic Twin Phantom; Flat Section; Probe: ET3DV6 - SN1560 -- Probe Cal Date 20/02/01; ConvF(6.78,6.78,6.78)
Med. Parameters 835 MHz Brain: σ = 0.90 mho/m ε_r = 41.5 ρ = 1.00 g/cm^3; Antenna Position -- Out; Crest Factor 1.0
SAR (1g): 2.43 mW/g, SAR (10g): 1.60 mW/g

835MHz Brain Dipole Validation (D835V2 S/N: 406)
Frequency: 835 MHz; Antenna Input Power: 250 [mW]
PCTEST Brain Tissue Simulating Liquid [01-23-2002]