

RF System Check Test Report

Measurement performed on March 4, 2026

Hardware Setup

Probe Name	Probe Calibration Date	DAE Name	DAE Calibration Date
EF3DV3 - SN4047	December 03, 2025	DAE4 Sn316	January 15, 2026

Communication Systems

Band Name	Communication Systems Name	Channel	Frequency [MHz]
CD2450	CW	50	2450.0

Grid Settings

Extent X [mm]	Extent Y [mm]	Step X [mm]	Step Y [mm]	Distance [mm]
20.0	90.0	5.0	5.0	15.0

Results

Dipole Type	Dipole Serial Number	E _{max} [V/m]	Drift [dB]
CD2450	1039	87.0	-0.08

