

Appendix A - System Verification

Test Laboratory: TUV Inc.

Date: 2026/9/2

System Check_HSL2450MHz

DUT: D2450V2 - SN804

Communication System: UID 0, CW; Frequency: 2450 MHz; Duty Cycle: 1:1

Medium: HSL2450_260902 Medium parameters used: $f = 2450$ MHz; $\sigma = 1.812$ S/m; $\epsilon_r = 39.95$; $\rho = 1000$ kg/m³

Ambient Temperature : 23.5 °C; Liquid Temperature : 22.4°C

DASY Configuration:

- Electronics: DAE4 Sn855; Calibrated: 2026/4/20
- Probe: EX3DV4 - SN7400; ConvF(6.98, 6.88, 8.06) @ 2450 MHz; Calibrated: 2026/4/23
- Sensor-Surface: 1.4mm (Mechanical Surface Detection), $z = -9.0, 31.0$
- Phantom: ELI V5.0 (20deg probe tilt); Type: QD OVA 002 AA; Serial: 1153
- DASY52 52.10.4(1535); SEMCAD X 14.6.14(7501)

Pin=250mW/Area Scan (9x9x1): Measurement grid: $dx=12$ mm, $dy=12$ mm

Maximum value of SAR (measured) = 19.3 W/kg


Pin=250mW/Zoom Scan (7x7x7)/Cube 0: Measurement grid: $dx=5$ mm, $dy=5$ mm, $dz=5$ mm

Reference Value = 110.2 V/m; Power Drift = -0.02 dB

Peak SAR (extrapolated) = 23.4 W/kg

SAR(1 g) = 12.5 W/kg; SAR(10 g) = 5.9 W/kg

Maximum value of SAR (measured) = 19.5 W/kg



0 dB = 19.3 W/kg = 12.86 dBW/kg